



## **Report of Analysis**

Client: ENTACT

540 Degraw St, Brooklyn, NY - E9309

Client Sample ID: WC-A2-04-C

Lab Sample ID: Q2236-07

Analytical Method: 8081B

Sample Wt/Vol: 100 Units: mL

Soil Aliquot Vol: uL

Extraction Type:

PD088861.D

Project:

GPC Factor: 1.0 PH:

Prep Method: SW3541B

File ID/Qc Batch: Dilution:

Prep Date

Date Analyzed

Injection Volume:

Date Collected:

Date Received:

SDG No.:

Matrix:

% Solid:

Final Vol:

Test:

06/04/25

06/04/25

Q2236

**TCLP** 

10000

TCLP Pesticide

Prep Batch ID

Decanted:

иL

06/06/25 12:32 06/10/25 13:41 PB168330

| CAS Number | Parameter            | Conc.    | Qualifier | MDL                 | LOQ / CRQL | Units   |
|------------|----------------------|----------|-----------|---------------------|------------|---------|
| TARGETS    |                      |          |           |                     |            |         |
| 58-89-9    | gamma-BHC (Lindane)  | 0.000037 | U         | 0.000037            | 0.00050    | mg/L    |
| 76-44-8    | Heptachlor           | 0.000027 | U         | 0.000027            | 0.00050    | mg/L    |
| 1024-57-3  | Heptachlor epoxide   | 0.000096 | U         | 0.000096            | 0.00050    | mg/L    |
| 72-20-8    | Endrin               | 0.000032 | U         | 0.000032            | 0.00050    | mg/L    |
| 72-43-5    | Methoxychlor         | 0.00011  | U         | 0.00011             | 0.00050    | mg/L    |
| 8001-35-2  | Toxaphene            | 0.0017   | U         | 0.0017              | 0.010      | mg/L    |
| 57-74-9    | Chlordane            | 0.00088  | U         | 0.00088             | 0.0050     | mg/L    |
| SURROGATES |                      |          |           |                     |            |         |
| 2051-24-3  | Decachlorobiphenyl   | 23.9     |           | 30 (57) - 150 (171) | 120%       | SPK: 20 |
| 877-09-8   | Tetrachloro-m-xylene | 23.1     |           | 30 (61) - 150 (148) | 116%       | SPK: 20 |

## Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit