

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-03	WC-A4-05A-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-07	WC-A2-04-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-11	WC-A2-05-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-15	WC-A2-06-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-19	WC-A2-07-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/10/25	06/04/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2236
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								
				0.000				
			Total Svoc :			0.00		
			Total Concentration:			0.00		



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2236

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168311TB	PB168311TB	2-Fluorophenol	150	141	94		15 (23)	110 (138)
		Phenol-d6	150	132	88		15 (10)	110 (134)
		Nitrobenzene-d5	100	86.1	86		30 (67)	130 (132)
		2-Fluorobiphenyl	100	83.7	84		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	142	94		15 (44)	110 (137)
		Terphenyl-d14	100	81.7	82		30 (42)	130 (152)
PB168352BL	PB168352BL	2-Fluorophenol	150	151	101		15 (23)	110 (138)
		Phenol-d6	150	142	94		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.5	92		30 (67)	130 (132)
		2-Fluorobiphenyl	100	91.6	92		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	156	104		15 (44)	110 (137)
		Terphenyl-d14	100	88.3	88		30 (42)	130 (152)
PB168352BS	PB168352BS	2-Fluorophenol	150	146	97		15 (23)	110 (138)
		Phenol-d6	150	140	93		15 (10)	110 (134)
		Nitrobenzene-d5	100	88.6	89		30 (67)	130 (132)
		2-Fluorobiphenyl	100	86.1	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	153	102		15 (44)	110 (137)
		Terphenyl-d14	100	86.8	87		30 (42)	130 (152)
Q2226-04MS	TP06-MHI-WCMS	2-Fluorophenol	150	125	83		15 (23)	110 (138)
		Phenol-d6	150	111	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	83.6	84		30 (67)	130 (132)
		2-Fluorobiphenyl	100	80.0	80		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	88		15 (44)	110 (137)
		Terphenyl-d14	100	83.3	83		30 (42)	130 (152)
Q2226-04MSD	TP06-MHI-WCMSD	2-Fluorophenol	150	132	88		15 (23)	110 (138)
		Phenol-d6	150	118	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.0	90		30 (67)	130 (132)
		2-Fluorobiphenyl	100	86.3	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	143	95		15 (44)	110 (137)
		Terphenyl-d14	100	90.5	91		30 (42)	130 (152)
Q2236-03	WC-A4-05A-C	2-Fluorophenol	150	129	86		15 (23)	110 (138)
		Phenol-d6	150	111	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.1	90		30 (67)	130 (132)
		2-Fluorobiphenyl	100	85.0	85		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	150	100		15 (44)	110 (137)
		Terphenyl-d14	100	85.1	85		30 (42)	130 (152)
Q2236-07	WC-A2-04-C	2-Fluorophenol	150	137	92		15 (23)	110 (138)
		Phenol-d6	150	121	81		15 (10)	110 (134)
		Nitrobenzene-d5	100	89.7	90		30 (67)	130 (132)
		2-Fluorobiphenyl	100	81.9	82		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	149	99		15 (44)	110 (137)
		Terphenyl-d14	100	83.5	84		30 (42)	130 (152)
Q2236-11	WC-A2-05-C	2-Fluorophenol	150	138	92		15 (23)	110 (138)
		Phenol-d6	150	119	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.7	93		30 (67)	130 (132)
		2-Fluorobiphenyl	100	85.6	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	143	95		15 (44)	110 (137)
		Terphenyl-d14	100	91.3	91		30 (42)	130 (152)
Q2236-15	WC-A2-06-C	2-Fluorophenol	150	146	97		15 (23)	110 (138)
		Phenol-d6	150	126	84		15 (10)	110 (134)
		Nitrobenzene-d5	100	97.0	97		30 (67)	130 (132)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SW-846

SDG No.: Q2236

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2236-15	WC-A2-06-C	2-Fluorobiphenyl	100	90.3	90		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	152	101		15 (44)	110 (137)
		Terphenyl-d14	100	93.6	94		30 (42)	130 (152)
Q2236-19	WC-A2-07-C	2-Fluorophenol	150	142	95		15 (23)	110 (138)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	94.4	94		30 (67)	130 (132)
		2-Fluorobiphenyl	100	86.5	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	148	99		15 (44)	110 (137)
		Terphenyl-d14	100	93.9	94		30 (42)	130 (152)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2236

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2226-04MS	Client Sample ID:	TP06-MHI-WCMS					DataFile:	BM050250.D		
Pyridine	500	0	360	ug/L	72				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	400	ug/L	80				70 (55)	130 (125)	
2-Methylphenol	500	0	430	ug/L	86				70 (60)	130 (131)	
3+4-Methylphenols	500	0	430	ug/L	86				20 (54)	160 (136)	
Hexachloroethane	500	0	420	ug/L	84				20 (19)	160 (146)	
Nitrobenzene	500	0	470	ug/L	94				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	430	ug/L	86				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	480	ug/L	96				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	480	ug/L	96				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	520	ug/L	104				70 (74)	130 (137)	
Hexachlorobenzene	500	0	490	ug/L	98				70 (72)	130 (115)	
Pentachlorophenol	1000	0	990	ug/L	99				20 (52)	160 (162)	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2236

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2226-04MSD	Client Sample ID:	TP06-MHI-WCMSD					DataFile:	BM050251.D		
Pyridine	500	0	380	ug/L	76	5			20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	440	ug/L	88	10			70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	460	ug/L	92	7			70 (60)	130 (131)	20 (20)
3+4-Methylphenols	500	0	460	ug/L	92	7			20 (54)	160 (136)	20 (20)
Hexachloroethane	500	0	450	ug/L	90	7			20 (19)	160 (146)	20 (20)
Nitrobenzene	500	0	500	ug/L	100	6			70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	490	ug/L	98	13			70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	530	ug/L	106	10			70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	530	ug/L	106	10			70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	550	ug/L	110	6			70 (74)	130 (137)	20 (20)
Hexachlorobenzene	500	0	530	ug/L	106	8			70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	1100	ug/L	110	11			20 (52)	160 (162)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2236

Client: ENTACT

Analytical Method: 8270E DataFile: BM050264.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168352BS	Pyridine	50	43.6	ug/L	87				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	46.9	ug/L	94				70 (76)	130 (103)	
	2-Methylphenol	50	49.2	ug/L	98				70 (69)	130 (109)	
	3+4-Methylphenols	50	49.6	ug/L	99				20 (67)	160 (106)	
	Hexachloroethane	50	49.5	ug/L	99				20 (76)	160 (118)	
	Nitrobenzene	50	51.3	ug/L	103				70 (58)	130 (106)	
	Hexachlorobutadiene	50	48.2	ug/L	96				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	52.5	ug/L	105				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	52.9	ug/L	106				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	58.9	ug/L	118				70 (60)	130 (115)	
	Hexachlorobenzene	50	52.0	ug/L	104				70 (73)	130 (106)	
	Pentachlorophenol	100	110	ug/L	110				20 (47)	160 (114)	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168352BL

Lab Name: CHEMTECH

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236

SAS No.: Q2236 SDG NO.: Q2236

Lab File ID: BM050263.D

Lab Sample ID: PB168352BL

Instrument ID: BNA_M

Date Extracted: 06/09/2025

Matrix: (soil/water) water

Date Analyzed: 06/10/2025

Level: (low/med) LOW

Time Analyzed: 05:28

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
TP06-MHI-WCMS	Q2226-04MS	BM050250.D	06/09/2025
TP06-MHI-WCMSD	Q2226-04MSD	BM050251.D	06/09/2025
WC-A4-05A-C	Q2236-03	BM050253.D	06/09/2025
PB168352BS	PB168352BS	BM050264.D	06/10/2025
WC-A2-04-C	Q2236-07	BM050248.D	06/09/2025
WC-A2-05-C	Q2236-11	BM050254.D	06/09/2025
WC-A2-06-C	Q2236-15	BM050255.D	06/09/2025
WC-A2-07-C	Q2236-19	BM050256.D	06/10/2025
PB168311TB	PB168311TB	BM050246.D	06/09/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: ENTA05Lab Code: CHEMSAS No.: Q2236SDG NO.: Q2236Lab File ID: BM050193.DDFTPP Injection Date: 06/05/2025Instrument ID: BNA_MDFTPP Injection Time: 08:40

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	20.3
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	37.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.6
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	74.4
443	15.0 - 24.0% of mass 442	14.8 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050194.D	06/05/2025	09:20
SSTDICC005	SSTDICC005	BM050195.D	06/05/2025	09:59
SSTDICC010	SSTDICC010	BM050196.D	06/05/2025	10:38
SSTDICC020	SSTDICC020	BM050197.D	06/05/2025	11:17
SSTDICCC040	SSTDICCC040	BM050198.D	06/05/2025	11:57
SSTDICC050	SSTDICC050	BM050199.D	06/05/2025	12:36
SSTDICC060	SSTDICC060	BM050200.D	06/05/2025	13:16
SSTDICC080	SSTDICC080	BM050201.D	06/05/2025	13:56



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: ENTA05Lab Code: CHEMSAS No.: Q2236SDG NO.: Q2236Lab File ID: BM050244.DDFTPP Injection Date: 06/09/2025Instrument ID: BNA_MDFTPP Injection Time: 16:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	20.2
68	Less than 2.0% of mass 69	0.5 (1.4) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	46.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.3
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	12.7
442	Greater than 50% of mass 198	82.6
443	15.0 - 24.0% of mass 442	15.8 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050245.D	06/09/2025	17:03
PB168311TB	PB168311TB	BM050246.D	06/09/2025	17:43
WC-A2-04-C	Q2236-07	BM050248.D	06/09/2025	19:01
TP06-MHI-WCMS	Q2226-04MS	BM050250.D	06/09/2025	20:19
TP06-MHI-WCMSD	Q2226-04MSD	BM050251.D	06/09/2025	20:58
WC-A4-05A-C	Q2236-03	BM050253.D	06/09/2025	22:17
WC-A2-05-C	Q2236-11	BM050254.D	06/09/2025	22:56
WC-A2-06-C	Q2236-15	BM050255.D	06/09/2025	23:35
WC-A2-07-C	Q2236-19	BM050256.D	06/10/2025	00:15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: ENTA05Lab Code: CHEMSAS No.: Q2236 SDG NO.: Q2236Lab File ID: BM050261.DDFTPP Injection Date: 06/10/2025Instrument ID: BNA_MDFTPP Injection Time: 04:09

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.8
68	Less than 2.0% of mass 69	0.7 (1.5) 1
69	Mass 69 relative abundance	45.4
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	52.3
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	11.3
442	Greater than 50% of mass 198	70.3
443	15.0 - 24.0% of mass 442	14.2 (20.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050262.D	06/10/2025	04:48
PB168352BL	PB168352BL	BM050263.D	06/10/2025	05:28
PB168352BS	PB168352BS	BM050264.D	06/10/2025	06:07

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/09/2025

Lab File ID: BM050245.D Time Analyzed: 17:03

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	360068	7.769	1478210	10.56	892439	14.40
UPPER LIMIT	720136	8.269	2956420	11.057	1784880	14.898
LOWER LIMIT	180034	7.269	739105	10.057	446220	13.898
EPA SAMPLE NO.						
01 TP06-MHI-WCMS	392950	7.77	1521760	10.56	827375	14.40
02 TP06-MHI-WCMSD	395507	7.77	1511860	10.56	835521	14.40
03 WC-A2-04-C	334436	7.77	1312450	10.56	761229	14.40
04 PB168311TB	348706	7.77	1311820	10.55	755301	14.40
05 WC-A4-05A-C	339236	7.77	1291740	10.56	734983	14.40
06 WC-A2-05-C	372657	7.77	1419600	10.56	799341	14.40
07 WC-A2-06-C	313628	7.77	1214920	10.55	691337	14.40
08 WC-A2-07-C	372220	7.77	1448450	10.56	829081	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/09/2025

Lab File ID: BM050245.D Time Analyzed: 17:03

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1715620	17.139	1793890	21.374	1989460	24.35
UPPER LIMIT	3431240	17.639	3587780	21.874	3978920	24.85
LOWER LIMIT	857810	16.639	896945	20.874	994730	23.85
EPA SAMPLE NO.						
01 TP06-MHI-WCMS	1462390	17.14	1332940	21.37	1379950	24.35
02 TP06-MHI-WCMSD	1452450	17.14	1291560	21.37	1337170	24.34
03 WC-A2-04-C	1459850	17.13	1563300	21.36	1685600	24.34
04 PB168311TB	1467650	17.13	1523440	21.37	1647710	24.34
05 WC-A4-05A-C	1365220	17.13	1409430	21.36	1492060	24.34
06 WC-A2-05-C	1429370	17.13	1316640	21.36	1352310	24.34
07 WC-A2-06-C	1261240	17.13	1230720	21.36	1312540	24.34
08 WC-A2-07-C	1500990	17.13	1402390	21.36	1434270	24.34

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/10/2025

Lab File ID: BM050262.D Time Analyzed: 04:48

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	460715	7.763	1877130	10.55	1070950	14.40
UPPER LIMIT	921430	8.263	3754260	11.051	2141900	14.898
LOWER LIMIT	230358	7.263	938565	10.051	535475	13.898
EPA SAMPLE NO.						
01 PB168352BL	334540	7.76	1240840	10.55	698531	14.39
02 PB168352BS	333914	7.76	1311550	10.55	754658	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/10/2025

Lab File ID: BM050262.D Time Analyzed: 04:48

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1957350	17.133	1968380	21.368	2090330	24.338
UPPER LIMIT	3914700	17.633	3936760	21.868	4180660	24.838
LOWER LIMIT	978675	16.633	984190	20.868	1045170	23.838
EPA SAMPLE NO.						
01 PB168352BL	1381510	17.13	1501560	21.36	1574970	24.34
02 PB168352BS	1423730	17.13	1520710	21.37	1658220	24.34

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	06/09/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/09/25
Client Sample ID:	PB168311TB	SDG No.:	Q2236
Lab Sample ID:	PB168311TB	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050246.D	1	06/09/25 10:45	06/09/25 17:43	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	141		15 (23) - 110 (138)	94%	SPK: 150
13127-88-3	Phenol-d6	132		15 (10) - 110 (134)	88%	SPK: 150
4165-60-0	Nitrobenzene-d5	86.1		30 (67) - 130 (132)	86%	SPK: 100
321-60-8	2-Fluorobiphenyl	83.7		30 (52) - 130 (132)	84%	SPK: 100
118-79-6	2,4,6-Tribromophenol	142		15 (44) - 110 (137)	94%	SPK: 150
1718-51-0	Terphenyl-d14	81.7		30 (42) - 130 (152)	82%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	349000	7.769			
1146-65-2	Naphthalene-d8	1310000	10.551			
15067-26-2	Acenaphthene-d10	755000	14.398			
1517-22-2	Phenanthrene-d10	1470000	17.133			
1719-03-5	Chrysene-d12	1520000	21.368			
1520-96-3	Perylene-d12	1650000	24.344			

Report of Analysis

Client:	ENTACT		Date Collected:	06/09/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/09/25	
Client Sample ID:	PB168311TB		SDG No.:	Q2236	
Lab Sample ID:	PB168311TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050246.D	1	06/09/25 10:45	06/09/25 17:43	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050246.D
Acq On : 09 Jun 2025 17:43
Operator : RC/JU
Sample : PB168311TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168311TB

Quant Time: Jun 10 01:25:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	348706	20.000	ng	0.00
21) Naphthalene-d8	10.551	136	1311818	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	755301	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1467647	20.000	ng	0.00
76) Chrysene-d12	21.368	240	1523440	20.000	ng	0.00
86) Perylene-d12	24.344	264	1647706	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	2951685	141.203	ng	0.00
7) Phenol-d6	6.934	99	3637251	132.177	ng	0.00
23) Nitrobenzene-d5	8.916	82	2171348	86.085	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1216044	141.548	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	4669156	83.693	ng	0.00
79) Terphenyl-d14	19.768	244	6569575	81.720	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050246.D
Acq On : 09 Jun 2025 17:43
Operator : RC/JU
Sample : PB168311TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168311TB


Quant Time: Jun 10 01:25:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050246.D
Acq On : 09 Jun 2025 17:43
Operator : RC/JU
Sample : PB168311TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168311TB

Quant Time: Jun 10 01:25:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. -0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43




Instrument :
BNA_M
ClientSampleId :
PB168311TB


Tgt Ion:152 Resp: 348706
Ion Ratio Lower Upper
152 100
150 162.5 122.2 183.4
115 59.7 46.3 69.5



#5
2-Fluorophenol
Concen: 141.203 ng
RT: 5.357 min Scan# 403
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43

Tgt Ion:112 Resp: 2951685
Ion Ratio Lower Upper
112 100
64 47.9 39.8 59.6
63 24.5 19.8 29.8









#7
Phenol-d6
Concen: 132.177 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43




Instrument :
BNA_M
ClientSampleId :
PB168311TB


Tgt Ion: 99 Resp: 3637251
Ion Ratio Lower Upper
99 100
42 9.1 6.9 10.3
71 31.9 25.3 37.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1286
Delta R.T. -0.006 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43

Tgt Ion: 136 Resp: 1311818
Ion Ratio Lower Upper
136 100
137 10.9 8.6 13.0
54 5.3 3.8 5.8
68 6.2 4.9 7.3









#23
Nitrobenzene-d5
Concen: 86.085 ng
RT: 8.916 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43




Instrument :
BNA_M
ClientSampleId :
PB168311TB


Tgt Ion: 82 Resp: 2171348
Ion Ratio Lower Upper
82 100
128 44.2 35.2 52.8
54 31.7 24.5 36.7



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43

Tgt Ion:164 Resp: 755301
Ion Ratio Lower Upper
164 100
162 99.3 80.2 120.4
160 45.2 35.7 53.5









#42
2,4,6-Tribromophenol
Concen: 141.548 ng
RT: 15.880 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43




Instrument :
BNA_M
ClientSampleId :
PB168311TB


Tgt Ion:330 Resp: 1216044
Ion Ratio Lower Upper
330 100
332 97.1 77.5 116.3
141 35.5 28.8 43.2



#45
2-Fluorobiphenyl
Concen: 83.693 ng
RT: 13.022 min Scan# 1706
Delta R.T. -0.006 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43

Tgt Ion:172 Resp: 4669156
Ion Ratio Lower Upper
172 100
171 34.4 27.8 41.6
170 23.7 18.6 27.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43




Instrument :
BNA_M
ClientSampleId :
PB168311TB


Tgt Ion:188 Resp: 1467647
Ion Ratio Lower Upper
188 100
94 10.7 8.1 12.1
80 11.1 8.6 13.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43

Tgt Ion:240 Resp: 1523440
Ion Ratio Lower Upper
240 100
120 11.2 9.0 13.4
236 25.2 20.7 31.1









#79
Terphenyl-d14
Concen: 81.720 ng
RT: 19.768 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43




Instrument :
BNA_M
ClientSampleId :
PB168311TB

Tgt Ion:244 Resp: 6569575
Ion Ratio Lower Upper
244 100
212 7.1 6.0 9.0
122 10.3 8.6 12.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.344 min Scan# 3631
Delta R.T. 0.000 min
Lab File: BM050246.D
Acq: 09 Jun 2025 17:43

Tgt Ion:264 Resp: 1647706
Ion Ratio Lower Upper
264 100
260 23.1 18.6 28.0
265 21.6 16.8 25.2



Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A4-05A-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-03	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050253.D	1	06/09/25 10:45	06/09/25 22:17	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	129		15 (23) - 110 (138)	86%	SPK: 150
13127-88-3	Phenol-d6	111		15 (10) - 110 (134)	74%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.1		30 (67) - 130 (132)	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	85.0		30 (52) - 130 (132)	85%	SPK: 100
118-79-6	2,4,6-Tribromophenol	150		15 (44) - 110 (137)	100%	SPK: 150
1718-51-0	Terphenyl-d14	85.1		30 (42) - 130 (152)	85%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	339000	7.769			
1146-65-2	Naphthalene-d8	1290000	10.557			
15067-26-2	Acenaphthene-d10	735000	14.398			
1517-22-2	Phenanthrene-d10	1370000	17.133			
1719-03-5	Chrysene-d12	1410000	21.362			
1520-96-3	Perylene-d12	1490000	24.339			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A4-05A-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-03		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050253.D	1	06/09/25 10:45	06/09/25 22:17	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050253.D
 Acq On : 09 Jun 2025 22:17
 Operator : RC/JU
 Sample : Q2236-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 WC-A4-05A-C

Quant Time: Jun 10 01:28:07 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	339236	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	1291741	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	734983	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1365218	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1409427	20.000	ng	-0.01
86) Perylene-d12	24.339	264	1492062	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	2623964	129.029	ng	0.00
7) Phenol-d6	6.934	99	2974342	111.105	ng	0.00
23) Nitrobenzene-d5	8.916	82	2238948	90.145	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1255343	150.162	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	4616642	85.039	ng	0.00
79) Terphenyl-d14	19.763	244	6326070	85.056	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	10.604	128	6271348	93.902	ng	100
37) 2-Methylnaphthalene	12.216	142	655891	16.279	ng	100
38) 1-Methylnaphthalene	12.434	142	1041234	24.349	ng	100
46) 1,1'-Biphenyl	13.234	154	145035	2.635	ng	99
52) Acenaphthene	14.457	154	812188	18.764	ng	98
55) Dibenzofuran	14.792	168	145848	2.303	ng	99
58) Fluorene	15.445	166	430582	8.877	ng	99
71) Phenanthrene	17.174	178	1073820	13.769	ng	99
72) Anthracene	17.263	178	166472	2.134	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050253.D
Acq On : 09 Jun 2025 22:17
Operator : RC/JU
Sample : Q2236-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Quant Time: Jun 10 01:28:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050253.D
Acq On : 09 Jun 2025 22:17
Operator : RC/JU
Sample : Q2236-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C

Quant Time: Jun 10 01:28:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:152 Resp: 339236
Ion Ratio Lower Upper
152 100
150 154.8 122.2 183.4
115 59.7 46.3 69.5



#5
2-Fluorophenol
Concen: 129.029 ng
RT: 5.358 min Scan# 403
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:112 Resp: 2623964
Ion Ratio Lower Upper
112 100
64 49.2 39.8 59.6
63 24.9 19.8 29.8









#7
Phenol-d6
Concen: 111.105 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion: 99 Resp: 2974342
Ion Ratio Lower Upper
99 100
42 8.7 6.9 10.3
71 31.9 25.3 37.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.557 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion: 136 Resp: 1291741
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 5.1 3.8 5.8
68 5.9 4.9 7.3









#23
Nitrobenzene-d5
Concen: 90.145 ng
RT: 8.916 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion: 82 Resp: 2238948
Ion Ratio Lower Upper
82 100
128 43.0 35.2 52.8
54 30.7 24.5 36.7



#31
Naphthalene
Concen: 93.902 ng
RT: 10.604 min Scan# 1295
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:128 Resp: 6271348
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 12.8 10.2 15.4









#37
2-Methylnaphthalene
Concen: 16.279 ng
RT: 12.216 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:142 Resp: 655891
Ion Ratio Lower Upper
142 100
141 89.4 71.5 107.3
115 34.9 27.8 41.8



#38
1-Methylnaphthalene
Concen: 24.349 ng
RT: 12.434 min Scan# 1606
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:142 Resp: 1041234
Ion Ratio Lower Upper
142 100
141 90.8 72.9 109.3
116 3.8 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:164 Resp: 734983
Ion Ratio Lower Upper
164 100
162 99.6 80.2 120.4
160 44.7 35.7 53.5



#42
2,4,6-Tribromophenol
Concen: 150.162 ng
RT: 15.880 min Scan# 2192
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:330 Resp: 1255343
Ion Ratio Lower Upper
330 100
332 96.7 77.5 116.3
141 36.4 28.8 43.2









#45
2-Fluorobiphenyl
Concen: 85.039 ng
RT: 13.022 min Scan# 1706
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:172 Resp: 4616642
Ion Ratio Lower Upper
172 100
171 34.4 27.8 41.6
170 22.8 18.6 27.8



#46
1,1'-Biphenyl
Concen: 2.635 ng
RT: 13.234 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:154 Resp: 145035
Ion Ratio Lower Upper
154 100
153 39.7 19.2 59.2
76 13.1 0.0 32.0









#52
Acenaphthene
Concen: 18.764 ng
RT: 14.457 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:154 Resp: 812188
Ion Ratio Lower Upper
154 100
153 112.5 88.7 133.1
152 55.1 42.6 63.8



#55
Dibenzofuran
Concen: 2.303 ng
RT: 14.792 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:168 Resp: 145848
Ion Ratio Lower Upper
168 100
139 36.8 29.9 44.9
169 13.8 10.6 16.0









#58
Fluorene
Concen: 8.877 ng
RT: 15.445 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:166 Resp: 430582
Ion Ratio Lower Upper
166 100
165 97.4 77.0 115.4
167 13.1 10.7 16.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:188 Resp: 1365218
Ion Ratio Lower Upper
188 100
94 10.5 8.1 12.1
80 11.1 8.6 13.0









#71
Phenanthrene
Concen: 13.769 ng
RT: 17.174 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:178 Resp: 1073820
Ion Ratio Lower Upper
178 100
176 19.1 15.6 23.4
179 15.4 12.3 18.5



#72
Anthracene
Concen: 2.134 ng
RT: 17.263 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:178 Resp: 166472
Ion Ratio Lower Upper
178 100
176 18.7 14.8 22.2
179 16.1 12.3 18.5









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3124
Delta R.T. -0.011 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17




Instrument :
BNA_M
ClientSampleId :
WC-A4-05A-C


Tgt Ion:240 Resp: 1409427
Ion Ratio Lower Upper
240 100
120 11.4 9.0 13.4
236 25.6 20.7 31.1



#79
Terphenyl-d14
Concen: 85.056 ng
RT: 19.763 min Scan# 2852
Delta R.T. -0.006 min
Lab File: BM050253.D
Acq: 09 Jun 2025 22:17

Tgt Ion:244 Resp: 6326070
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 11.2 8.6 12.8





#86

Perylene-d12

Concen: 20.000 ng

RT: 24.339 min Scan# 30

Delta R.T. -0.006 min

Lab File: BM050253.D


Acq: 09 Jun 2025 22:17

Instrument :

BNA_M

ClientSampleId :

WC-A4-05A-C





Tgt Ion:264 Resp: 1492062

Ion Ratio Lower Upper

264 100

260 23.8 18.6 28.0

265 22.1 16.8 25.2



Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-07	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050248.D	1	06/09/25 10:45	06/09/25 19:01	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	137		15 (23) - 110 (138)	92%	SPK: 150
13127-88-3	Phenol-d6	121		15 (10) - 110 (134)	81%	SPK: 150
4165-60-0	Nitrobenzene-d5	89.7		30 (67) - 130 (132)	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	81.9		30 (52) - 130 (132)	82%	SPK: 100
118-79-6	2,4,6-Tribromophenol	149		15 (44) - 110 (137)	99%	SPK: 150
1718-51-0	Terphenyl-d14	83.5		30 (42) - 130 (152)	84%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	334000	7.769			
1146-65-2	Naphthalene-d8	1310000	10.557			
15067-26-2	Acenaphthene-d10	761000	14.398			
1517-22-2	Phenanthrene-d10	1460000	17.133			
1719-03-5	Chrysene-d12	1560000	21.362			
1520-96-3	Perylene-d12	1690000	24.344			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-04-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-07		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050248.D	1	06/09/25 10:45	06/09/25 19:01	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050248.D
 Acq On : 09 Jun 2025 19:01
 Operator : RC/JU
 Sample : Q2236-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 WC-A2-04-C

Quant Time: Jun 10 01:26:00 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	334436	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	1312450	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	761229	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1459849	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1563302	20.000	ng	-0.01
86) Perylene-d12	24.344	264	1685599	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	2752148	137.275	ng	0.00
7) Phenol-d6	6.934	99	3189329	120.845	ng	0.00
23) Nitrobenzene-d5	8.916	82	2263913	89.712	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1287864	148.741	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	4606639	81.929	ng	0.00
79) Terphenyl-d14	19.762	244	6889305	83.512	ng	0.00
Target Compounds						
						Qvalue
37) 2-Methylnaphthalene	12.222	142	4908486	119.906	ng	98
38) 1-Methylnaphthalene	12.439	142	3260277	75.038	ng	99
46) 1,1'-Biphenyl	13.233	154	580651	10.187	ng	99
49) Acenaphthylene	14.116	152	258538	3.607	ng	# 96
52) Acenaphthene	14.463	154	1714018	38.233	ng	98
58) Fluorene	15.445	166	693734	13.810	ng	99
71) Phenanthrene	17.174	178	1412684	16.939	ng	99
72) Anthracene	17.268	178	225734	2.706	ng	99
73) Carbazole	17.533	167	198686	2.614	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050248.D
Acq On : 09 Jun 2025 19:01
Operator : RC/JU
Sample : Q2236-07
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Quant Time: Jun 10 01:26:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050248.D
Acq On : 09 Jun 2025 19:01
Operator : RC/JU
Sample : Q2236-07
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C

Quant Time: Jun 10 01:26:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:152 Resp: 334436
Ion Ratio Lower Upper
152 100
150 155.2 122.2 183.4
115 58.3 46.3 69.5



#5
2-Fluorophenol
Concen: 137.275 ng
RT: 5.357 min Scan# 403
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:112 Resp: 2752148
Ion Ratio Lower Upper
112 100
64 51.2 39.8 59.6
63 25.7 19.8 29.8









#7
Phenol-d6
Concen: 120.845 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion: 99 Resp: 3189329
Ion Ratio Lower Upper
99 100
42 8.8 6.9 10.3
71 32.8 25.3 37.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.557 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion: 136 Resp: 1312450
Ion Ratio Lower Upper
136 100
137 10.6 8.6 13.0
54 5.5 3.8 5.8
68 6.1 4.9 7.3









#23
Nitrobenzene-d5
Concen: 89.712 ng
RT: 8.916 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion: 82 Resp: 2263913
Ion Ratio Lower Upper
82 100
128 42.0 35.2 52.8
54 31.4 24.5 36.7



#37
2-Methylnaphthalene
Concen: 119.906 ng
RT: 12.222 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion: 142 Resp: 4908486
Ion Ratio Lower Upper
142 100
141 87.5 71.5 107.3
115 35.2 27.8 41.8









#38
1-Methylnaphthalene
Concen: 75.038 ng
RT: 12.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:142 Resp: 3260277
Ion Ratio Lower Upper
142 100
141 90.2 72.9 109.3
116 3.9 3.0 4.6



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:164 Resp: 761229
Ion Ratio Lower Upper
164 100
162 101.2 80.2 120.4
160 45.0 35.7 53.5









#42
2,4,6-Tribromophenol
Concen: 148.741 ng
RT: 15.880 min Scan# 2192
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:330 Resp: 1287864
Ion Ratio Lower Upper
330 100
332 96.8 77.5 116.3
141 36.4 28.8 43.2



#45
2-Fluorobiphenyl
Concen: 81.929 ng
RT: 13.022 min Scan# 1706
Delta R.T. -0.006 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:172 Resp: 4606639
Ion Ratio Lower Upper
172 100
171 34.4 27.8 41.6
170 22.9 18.6 27.8









#46
1,1'-Biphenyl
Concen: 10.187 ng
RT: 13.233 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion	Ratio	Lower	Upper
154	100		
153	39.8	19.2	59.2
76	13.1	0.0	32.0



#49
Acenaphthylene
Concen: 3.607 ng
RT: 14.116 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.8	23.6
153	16.7	10.4	15.6









#52
Acenaphthene
Concen: 38.233 ng
RT: 14.463 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:154 Resp: 1714018
Ion Ratio Lower Upper
154 100
153 113.2 88.7 133.1
152 52.9 42.6 63.8



#58
Fluorene
Concen: 13.810 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:166 Resp: 693734
Ion Ratio Lower Upper
166 100
165 95.6 77.0 115.4
167 13.1 10.7 16.1









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:188 Resp: 1459849
Ion Ratio Lower Upper
188 100
94 10.8 8.1 12.1
80 11.8 8.6 13.0



#71
Phenanthrene
Concen: 16.939 ng
RT: 17.174 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:178 Resp: 1412684
Ion Ratio Lower Upper
178 100
176 18.9 15.6 23.4
179 15.5 12.3 18.5









#72
Anthracene
Concen: 2.706 ng
RT: 17.268 min Scan# 2431
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:178 Resp: 225734
Ion Ratio Lower Upper
178 100
176 18.3 14.8 22.2
179 14.9 12.3 18.5



#73
Carbazole
Concen: 2.614 ng
RT: 17.533 min Scan# 2473
Delta R.T. 0.000 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:167 Resp: 198686
Ion Ratio Lower Upper
167 100
166 21.4 17.0 25.6
139 13.1 10.2 15.4









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3128
Delta R.T. -0.012 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01




Instrument :
BNA_M
ClientSampleId :
WC-A2-04-C


Tgt Ion:240 Resp: 1563302
Ion Ratio Lower Upper
240 100
120 11.9 9.0 13.4
236 25.0 20.7 31.1



#79
Terphenyl-d14
Concen: 83.512 ng
RT: 19.762 min Scan# 2852
Delta R.T. -0.006 min
Lab File: BM050248.D
Acq: 09 Jun 2025 19:01

Tgt Ion:244 Resp: 6889305
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 11.5 8.6 12.8





#86

Perylene-d12

Concen: 20.000 ng

RT: 24.344 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM050248.D


Acq: 09 Jun 2025 19:01

Instrument :

BNA_M

ClientSampleId :

WC-A2-04-C





Tgt Ion:264 Resp: 1685599

Ion	Ratio	Lower	Upper

264	100		

260	23.0	18.6	28.0

265	21.7	16.8	25.2



Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-05-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-11	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050254.D	1	06/09/25 10:45	06/09/25 22:56	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	138		15 (23) - 110 (138)	92%	SPK: 150
13127-88-3	Phenol-d6	119		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.7		30 (67) - 130 (132)	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	85.6		30 (52) - 130 (132)	86%	SPK: 100
118-79-6	2,4,6-Tribromophenol	143		15 (44) - 110 (137)	95%	SPK: 150
1718-51-0	Terphenyl-d14	91.3		30 (42) - 130 (152)	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	373000	7.769			
1146-65-2	Naphthalene-d8	1420000	10.557			
15067-26-2	Acenaphthene-d10	799000	14.398			
1517-22-2	Phenanthrene-d10	1430000	17.133			
1719-03-5	Chrysene-d12	1320000	21.362			
1520-96-3	Perylene-d12	1350000	24.339			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-05-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-11		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050254.D	1	06/09/25 10:45	06/09/25 22:56	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050254.D
 Acq On : 09 Jun 2025 22:56
 Operator : RC/JU
 Sample : Q2236-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 WC-A2-05-C

Quant Time: Jun 10 01:28:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	372657	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	1419602	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	799341	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1429372	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1316639	20.000	ng	-0.01
86) Perylene-d12	24.339	264	1352313	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	3073333	137.573	ng	0.00
7) Phenol-d6	6.934	99	3489531	118.659	ng	0.00
23) Nitrobenzene-d5	8.916	82	2529986	92.688	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1300437	143.032	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	5056880	85.649	ng	0.00
79) Terphenyl-d14	19.762	244	6346327	91.342	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.604	128	7308166	99.571	ng	100
37) 2-Methylnaphthalene	12.216	142	763922	17.253	ng	99
38) 1-Methylnaphthalene	12.439	142	1094227	23.284	ng	99
46) 1,1'-Biphenyl	13.228	154	154552	2.582	ng	98
52) Acenaphthene	14.457	154	801278	17.021	ng	98
55) Dibenzofuran	14.792	168	169828	2.466	ng	99
58) Fluorene	15.439	166	416718	7.900	ng	100
71) Phenanthrene	17.174	178	985702	12.072	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050254.D
Acq On : 09 Jun 2025 22:56
Operator : RC/JU
Sample : Q2236-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Quant Time: Jun 10 01:28:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050254.D
Acq On : 09 Jun 2025 22:56
Operator : RC/JU
Sample : Q2236-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C

Quant Time: Jun 10 01:28:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:152 Resp: 372657
Ion Ratio Lower Upper
152 100
150 156.7 122.2 183.4
115 58.8 46.3 69.5



#5
2-Fluorophenol
Concen: 137.573 ng
RT: 5.357 min Scan# 403
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:112 Resp: 3073333
Ion Ratio Lower Upper
112 100
64 50.0 39.8 59.6
63 25.5 19.8 29.8









#7
Phenol-d6
Concen: 118.659 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion: 99 Resp: 3489531
Ion Ratio Lower Upper
99 100
42 8.6 6.9 10.3
71 32.3 25.3 37.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.557 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion: 136 Resp: 1419602
Ion Ratio Lower Upper
136 100
137 11.0 8.6 13.0
54 5.1 3.8 5.8
68 5.9 4.9 7.3









#23
Nitrobenzene-d5
Concen: 92.688 ng
RT: 8.916 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion: 82 Resp: 2529986
Ion Ratio Lower Upper
82 100
128 42.2 35.2 52.8
54 30.9 24.5 36.7



#31
Naphthalene
Concen: 99.571 ng
RT: 10.604 min Scan# 1295
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion: 128 Resp: 7308166
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.8 10.2 15.4









#37
2-Methylnaphthalene
Concen: 17.253 ng
RT: 12.216 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:142 Resp: 763922
Ion Ratio Lower Upper
142 100
141 88.1 71.5 107.3
115 34.4 27.8 41.8



#38
1-Methylnaphthalene
Concen: 23.284 ng
RT: 12.439 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:142 Resp: 1094227
Ion Ratio Lower Upper
142 100
141 90.1 72.9 109.3
116 3.9 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:164 Resp: 799341
Ion Ratio Lower Upper
164 100
162 100.4 80.2 120.4
160 43.8 35.7 53.5



#42
2,4,6-Tribromophenol
Concen: 143.032 ng
RT: 15.880 min Scan# 2192
Delta R.T. 0.000 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:330 Resp: 1300437
Ion Ratio Lower Upper
330 100
332 96.4 77.5 116.3
141 36.9 28.8 43.2









#45
2-Fluorobiphenyl
Concen: 85.649 ng
RT: 13.022 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:172 Resp: 5056880
Ion Ratio Lower Upper
172 100
171 34.9 27.8 41.6
170 23.4 18.6 27.8



#46
1,1'-Biphenyl
Concen: 2.582 ng
RT: 13.228 min Scan# 1741
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:154 Resp: 154552
Ion Ratio Lower Upper
154 100
153 39.9 19.2 59.2
76 13.1 0.0 32.0









#52
Acenaphthene
Concen: 17.021 ng
RT: 14.457 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:154 Resp: 801278
Ion Ratio Lower Upper
154 100
153 113.3 88.7 133.1
152 54.7 42.6 63.8



#55
Dibenzofuran
Concen: 2.466 ng
RT: 14.792 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:168 Resp: 169828
Ion Ratio Lower Upper
168 100
139 36.9 29.9 44.9
169 13.9 10.6 16.0









#58
Fluorene
Concen: 7.900 ng
RT: 15.439 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:166 Resp: 416718
Ion Ratio Lower Upper
166 100
165 96.6 77.0 115.4
167 13.3 10.7 16.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:188 Resp: 1429372
Ion Ratio Lower Upper
188 100
94 10.5 8.1 12.1
80 10.8 8.6 13.0









#71
Phenanthrene
Concen: 12.072 ng
RT: 17.174 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C


Tgt Ion:178 Resp: 985702
Ion Ratio Lower Upper
178 100
176 19.1 15.6 23.4
179 15.4 12.3 18.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3124
Delta R.T. -0.012 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:240 Resp: 1316639
Ion Ratio Lower Upper
240 100
120 11.7 9.0 13.4
236 25.0 20.7 31.1









#79
Terphenyl-d14
Concen: 91.342 ng
RT: 19.762 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56




Instrument :
BNA_M
ClientSampleId :
WC-A2-05-C

Tgt Ion:244 Resp: 6346327
Ion Ratio Lower Upper
244 100
212 7.3 6.0 9.0
122 11.3 8.6 12.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.339 min Scan# 3630
Delta R.T. -0.006 min
Lab File: BM050254.D
Acq: 09 Jun 2025 22:56

Tgt Ion:264 Resp: 1352313
Ion Ratio Lower Upper
264 100
260 23.2 18.6 28.0
265 21.5 16.8 25.2



Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-06-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-15	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050255.D	1	06/09/25 10:45	06/09/25 23:35	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	146		15 (23) - 110 (138)	97%	SPK: 150
13127-88-3	Phenol-d6	126		15 (10) - 110 (134)	84%	SPK: 150
4165-60-0	Nitrobenzene-d5	97.0		30 (67) - 130 (132)	97%	SPK: 100
321-60-8	2-Fluorobiphenyl	90.3		30 (52) - 130 (132)	90%	SPK: 100
118-79-6	2,4,6-Tribromophenol	152		15 (44) - 110 (137)	101%	SPK: 150
1718-51-0	Terphenyl-d14	93.6		30 (42) - 130 (152)	94%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	314000	7.769			
1146-65-2	Naphthalene-d8	1210000	10.551			
15067-26-2	Acenaphthene-d10	691000	14.398			
1517-22-2	Phenanthrene-d10	1260000	17.133			
1719-03-5	Chrysene-d12	1230000	21.362			
1520-96-3	Perylene-d12	1310000	24.338			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-06-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-15		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050255.D	1	06/09/25 10:45	06/09/25 23:35	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050255.D
 Acq On : 09 Jun 2025 23:35
 Operator : RC/JU
 Sample : Q2236-15
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 WC-A2-06-C

Quant Time: Jun 10 01:28:43 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	313628	20.000	ng	0.00
21) Naphthalene-d8	10.551	136	1214917	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	691337	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1261240	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1230719	20.000	ng	-0.01
86) Perylene-d12	24.338	264	1312536	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	2745871	146.049	ng	0.00
7) Phenol-d6	6.934	99	3109740	125.647	ng	0.00
23) Nitrobenzene-d5	8.916	82	2265879	96.998	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1196426	152.150	ng	0.00
45) 2-Fluorobiphenyl	13.021	172	4612245	90.322	ng	0.00
79) Terphenyl-d14	19.762	244	6079199	93.606	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.604	128	1202299	19.141	ng	100
37) 2-Methylnaphthalene	12.216	142	204859	5.406	ng	99
38) 1-Methylnaphthalene	12.433	142	309802	7.703	ng	99
52) Acenaphthene	14.457	154	408261	10.027	ng	99
58) Fluorene	15.439	166	201762	4.422	ng	97
71) Phenanthrene	17.174	178	760701	10.558	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050255.D
Acq On : 09 Jun 2025 23:35
Operator : RC/JU
Sample : Q2236-15
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Quant Time: Jun 10 01:28:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050255.D
Acq On : 09 Jun 2025 23:35
Operator : RC/JU
Sample : Q2236-15
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C

Quant Time: Jun 10 01:28:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. -0.000 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion:152 Resp: 313628
Ion Ratio Lower Upper
152 100
150 161.1 122.2 183.4
115 60.5 46.3 69.5



#5
2-Fluorophenol
Concen: 146.049 ng
RT: 5.357 min Scan# 403
Delta R.T. -0.000 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:112 Resp: 2745871
Ion Ratio Lower Upper
112 100
64 49.7 39.8 59.6
63 25.3 19.8 29.8









#7
Phenol-d6
Concen: 125.647 ng
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion: 99 Resp: 3109740
Ion Ratio Lower Upper
99 100
42 8.5 6.9 10.3
71 32.3 25.3 37.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1286
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion: 136 Resp: 1214917
Ion Ratio Lower Upper
136 100
137 10.6 8.6 13.0
54 5.5 3.8 5.8
68 6.4 4.9 7.3









#23
Nitrobenzene-d5
Concen: 96.998 ng
RT: 8.916 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion: 82 Resp: 2265879
Ion Ratio Lower Upper
82 100
128 42.9 35.2 52.8
54 31.1 24.5 36.7



#31
Naphthalene
Concen: 19.141 ng
RT: 10.604 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:128 Resp: 1202299
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 12.7 10.2 15.4









#37
2-Methylnaphthalene
Concen: 5.406 ng
RT: 12.216 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion:142 Resp: 204859
Ion Ratio Lower Upper
142 100
141 88.2 71.5 107.3
115 34.4 27.8 41.8



#38
1-Methylnaphthalene
Concen: 7.703 ng
RT: 12.433 min Scan# 1606
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:142 Resp: 309802
Ion Ratio Lower Upper
142 100
141 92.3 72.9 109.3
116 3.7 3.0 4.6





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.398 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BM050255.D

Acq: 09 Jun 2025 23:35

Instrument :

BNA_M

ClientSampleId :

WC-A2-06-C





Tgt Ion:164 Resp: 691337

Ion Ratio Lower Upper

164 100

162 100.2 80.2 120.4

160 45.0 35.7 53.5



#42

2,4,6-Tribromophenol

Concen: 152.150 ng

RT: 15.880 min Scan# 2192

Delta R.T. -0.000 min

Lab File: BM050255.D

Acq: 09 Jun 2025 23:35




Tgt Ion:330 Resp: 1196426


Ion Ratio Lower Upper

330 100

332 97.0 77.5 116.3

141 37.5 28.8 43.2









#45
2-Fluorobiphenyl
Concen: 90.322 ng
RT: 13.021 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion:172 Resp: 4612245
Ion Ratio Lower Upper
172 100
171 34.4 27.8 41.6
170 23.1 18.6 27.8



#52
Acenaphthene
Concen: 10.027 ng
RT: 14.457 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:154 Resp: 408261
Ion Ratio Lower Upper
154 100
153 113.1 88.7 133.1
152 53.2 42.6 63.8









#58
Fluorene
Concen: 4.422 ng
RT: 15.439 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion:166 Resp: 201762
Ion Ratio Lower Upper
166 100
165 99.2 77.0 115.4
167 14.1 10.7 16.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:188 Resp: 1261240
Ion Ratio Lower Upper
188 100
94 10.0 8.1 12.1
80 11.1 8.6 13.0









#71
Phenanthrene
Concen: 10.558 ng
RT: 17.174 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C


Tgt Ion:178 Resp: 760701
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.7 12.3 18.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3124
Delta R.T. -0.012 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:240 Resp: 1230719
Ion Ratio Lower Upper
240 100
120 12.1 9.0 13.4
236 24.9 20.7 31.1









#79
Terphenyl-d14
Concen: 93.606 ng
RT: 19.762 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35




Instrument :
BNA_M
ClientSampleId :
WC-A2-06-C

Tgt Ion:244 Resp: 6079199
Ion Ratio Lower Upper
244 100
212 7.2 6.0 9.0
122 11.2 8.6 12.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.338 min Scan# 3630
Delta R.T. -0.006 min
Lab File: BM050255.D
Acq: 09 Jun 2025 23:35

Tgt Ion:264 Resp: 1312536
Ion Ratio Lower Upper
264 100
260 23.5 18.6 28.0
265 22.2 16.8 25.2



Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-19	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050256.D	1	06/09/25 10:45	06/10/25 00:15	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	12.8	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.30	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	11.2	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	11.0	100	ug/L
67-72-1	Hexachloroethane	6.50	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	7.60	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	6.20	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	12.2	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	5.20	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	15.8	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	142		15 (23) - 110 (138)	95%	SPK: 150
13127-88-3	Phenol-d6	124		15 (10) - 110 (134)	83%	SPK: 150
4165-60-0	Nitrobenzene-d5	94.4		30 (67) - 130 (132)	94%	SPK: 100
321-60-8	2-Fluorobiphenyl	86.5		30 (52) - 130 (132)	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	148		15 (44) - 110 (137)	99%	SPK: 150
1718-51-0	Terphenyl-d14	93.9		30 (42) - 130 (152)	94%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	372000	7.769			
1146-65-2	Naphthalene-d8	1450000	10.557			
15067-26-2	Acenaphthene-d10	829000	14.398			
1517-22-2	Phenanthrene-d10	1500000	17.133			
1719-03-5	Chrysene-d12	1400000	21.362			
1520-96-3	Perylene-d12	1430000	24.338			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-07-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-19		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050256.D	1	06/09/25 10:45	06/10/25 00:15	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050256.D
 Acq On : 10 Jun 2025 00:15
 Operator : RC/JU
 Sample : Q2236-19
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 WC-A2-07-C

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 07:46:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	372220	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	1448451	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	829081	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1500989	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1402387	20.000	ng	-0.01
86) Perylene-d12	24.338	264	1434274	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	3174322	142.261	ng	0.00
7) Phenol-d6	6.934	99	3651067	124.298	ng	0.00
23) Nitrobenzene-d5	8.916	82	2629870	94.428	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1394367	147.862	ng	0.00
45) 2-Fluorobiphenyl	13.021	172	5297450	86.505	ng	0.00
79) Terphenyl-d14	19.762	244	6946519	93.867	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.604	128	23545755m	314.413	ng	
37) 2-Methylnaphthalene	12.221	142	5285669	116.996	ng	99
38) 1-Methylnaphthalene	12.439	142	3546936	73.970	ng	100
46) 1,1'-Biphenyl	13.233	154	635306	10.234	ng	98
49) Acenaphthylene	14.115	152	279738	3.583	ng	# 95
52) Acenaphthene	14.462	154	1884550	38.596	ng	98
58) Fluorene	15.445	166	770683	14.086	ng	99
71) Phenanthrene	17.174	178	1541989	17.983	ng	99
72) Anthracene	17.262	178	245778	2.865	ng	99
73) Carbazole	17.527	167	191114	2.446	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050256.D
Acq On : 10 Jun 2025 00:15
Operator : RC/JU
Sample : Q2236-19
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 07:46:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050256.D
Acq On : 10 Jun 2025 00:15
Operator : RC/JU
Sample : Q2236-19
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025


Quant Time: Jun 10 07:46:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:152 Resp: 372220
Ion Ratio Lower Upper
152 100
150 158.0 122.2 183.4
115 60.6 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#5
2-Fluorophenol
Concen: 142.261 ng
RT: 5.357 min Scan# 403
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:112 Resp: 3174322
Ion Ratio Lower Upper
112 100
64 50.2 39.8 59.6
63 25.2 19.8 29.8





#7
Phenol-d6
Concen: 124.298 ng
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion: 99 Resp: 365106
Ion Ratio Lower Upper
99 100
42 8.9 6.9 10.3
71 33.1 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.557 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:136 Resp: 1448451
Ion Ratio Lower Upper
136 100
137 11.0 8.6 13.0
54 5.3 3.8 5.8
68 6.0 4.9 7.3





#23
Nitrobenzene-d5
Concen: 94.428 ng
RT: 8.916 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion: 82 Resp: 2629870
Ion Ratio Lower Upper
82 100
128 42.4 35.2 52.8
54 31.5 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#31
Naphthalene
Concen: 314.413 ng m
RT: 10.604 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:128 Resp:23545755
Ion Ratio Lower Upper
128 100
129 12.4 8.6 12.8
127 14.0 10.2 15.4





#37
2-Methylnaphthalene
Concen: 116.996 ng
RT: 12.221 min Scan# 1570
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C




Tgt Ion:142 Resp: 528566
Ion Ratio Lower Upper
142 100
141 87.9 71.5 107.3
115 35.3 27.8 41.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#38
1-Methylnaphthalene
Concen: 73.970 ng
RT: 12.439 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Tgt Ion:142 Resp: 3546936
Ion Ratio Lower Upper
142 100
141 90.8 72.9 109.3
116 4.0 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1940
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:164 Resp: 82908
Ion Ratio Lower Upper
164 100
162 99.4 80.2 120.4
160 45.6 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#42
2,4,6-Tribromophenol
Concen: 147.862 ng
RT: 15.880 min Scan# 2192
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:330 Resp: 1394367
Ion Ratio Lower Upper
330 100
332 96.3 77.5 116.3
141 35.9 28.8 43.2





#45
2-Fluorobiphenyl
Concen: 86.505 ng
RT: 13.021 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:172 Resp: 5297450
Ion Ratio Lower Upper
172 100
171 34.8 27.8 41.6
170 23.1 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#46
1,1'-Biphenyl
Concen: 10.234 ng
RT: 13.233 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:154 Resp: 635306
Ion Ratio Lower Upper
154 100
153 40.3 19.2 59.2
76 13.1 0.0 32.0





#49
Acenaphthylene
Concen: 3.583 ng
RT: 14.115 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:152 Resp: 279738
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 18.0 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#52
Acenaphthene
Concen: 38.596 ng
RT: 14.462 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:154 Resp: 1884550
Ion Ratio Lower Upper
154 100
153 113.3 88.7 133.1
152 53.3 42.6 63.8





#58
Fluorene
Concen: 14.086 ng
RT: 15.445 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:166 Resp: 77068
Ion Ratio Lower Upper
166 100
165 97.2 77.0 115.4
167 13.1 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15

Tgt Ion:188 Resp: 1500989
Ion Ratio Lower Upper
188 100
94 10.3 8.1 12.1
80 10.9 8.6 13.0









#71
Phenanthrene
Concen: 17.983 ng
RT: 17.174 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15

Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C

Tgt Ion:178 Resp: 1541989
Ion Ratio Lower Upper
178 100
176 18.6 15.6 23.4
179 15.1 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#72
Anthracene
Concen: 2.865 ng
RT: 17.262 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:178 Resp: 245778
Ion Ratio Lower Upper
178 100
176 19.5 14.8 22.2
179 15.3 12.3 18.5





#73
Carbazole
Concen: 2.446 ng
RT: 17.527 min Scan# 2476
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:167 Resp: 191114
Ion Ratio Lower Upper
167 100
166 21.1 17.0 25.6
139 13.4 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3124
Delta R.T. -0.012 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15


Tgt Ion:240 Resp: 1402387
Ion Ratio Lower Upper
240 100
120 11.8 9.0 13.4
236 25.1 20.7 31.1





#79
Terphenyl-d14
Concen: 93.867 ng
RT: 19.762 min Scan# 2855
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15




Instrument :
BNA_M
ClientSampleId :
WC-A2-07-C



Tgt Ion:244 Resp: 6946519
Ion Ratio Lower Upper
244 100
212 7.3 6.0 9.0
122 11.3 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.338 min Scan# 3630
Delta R.T. -0.006 min
Lab File: BM050256.D
Acq: 10 Jun 2025 00:15

Tgt Ion:264 Resp: 1434274
Ion Ratio Lower Upper
264 100
260 22.9 18.6 28.0
265 21.7 16.8 25.2





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG No.: Q2236
 Instrument ID: BNA_M Calibration Date(s): 06/05/2025 06/05/2025
 Calibration Time(s): 09:20 13:56

LAB FILE ID:		RRF2.5 = BM050194.D		RRF005 = BM050195.D		RRF010 = BM050196.D		
		RRF020 = BM050197.D		RRF040 = BM050198.D		RRF050 = BM050199.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.353	1.352	1.370	1.311	1.425	1.361	2.8
2-Fluorophenol		1.251	1.193	1.192	1.145	1.238	1.199	3.5
Phenol-d6		1.569	1.521	1.617	1.565	1.699	1.578	4.0
1,4-Dichlorobenzene		1.619	1.544	1.564	1.450	1.591	1.534	4.3
2-Methylphenol		1.078	1.042	1.133	1.108	1.207	1.102	5.0
3+4-Methylphenols		1.367	1.375	1.533	1.509	1.661	1.474	7.0
Nitrobenzene-d5		0.359	0.362	0.392	0.377	0.414	0.385	5.4
Hexachloroethane	0.553	0.564	0.544	0.552	0.533	0.593	0.555	3.3
Nitrobenzene		0.333	0.337	0.359	0.339	0.370	0.350	4.1
Hexachlorobutadiene		0.188	0.184	0.189	0.182	0.202	0.190	3.7
2,4,6-Trichlorophenol		0.346	0.347	0.379	0.375	0.414	0.380	6.9
2-Fluorobiphenyl		1.589	1.536	1.532	1.377	1.463	1.477	5.7
2,4,5-Trichlorophenol		0.371	0.382	0.424	0.414	0.457	0.417	7.5
2,4-Dinitrotoluene		0.250	0.283	0.365	0.396	0.447	0.360	19.3
2,4,6-Tribromophenol		0.210	0.213	0.238	0.231	0.254	0.227	7.0
Hexachlorobenzene		0.249	0.240	0.253	0.238	0.261	0.249	3.3
Pentachlorophenol		0.133	0.138	0.165	0.164	0.187	0.162	12.2
Terphenyl-d14		1.148	1.088	1.165	1.010	1.064	1.055	8.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM060525.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 05 16:20:25 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BM050194.D 5 =BM050195.D 10 =BM050196.D 20 =BM050197.D 40 =BM050198.D 50 =BM050199.D 60 =BM050200.D 80 =BM050201.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.597	0.552	0.516	0.479	0.511	0.536	0.488	0.526	7.69	
3)	Pyridine	1.353	1.352	1.370	1.311	1.425	1.389	1.326	1.361	2.80	
4)	n-Nitrosodimet...	0.283	0.265	0.279	0.272	0.297	0.293	0.282	0.281	3.95	
5) S	2-Fluorophenol	1.251	1.193	1.192	1.145	1.238	1.224	1.150	1.199	3.45	
6)	Aniline	2.033	2.016	2.147	2.065	2.250	2.064	2.031	2.086	4.02	
7) S	Phenol-d6	1.569	1.521	1.617	1.565	1.699	1.560	1.516	1.578	4.00	
8)	2-Chlorophenol	1.289	1.259	1.344	1.295	1.418	1.338	1.288	1.319	4.02	
9)	Benzaldehyde	1.188	1.080	1.096	0.878	0.975	0.804		1.003	14.43	
10) C	Phenol	1.673	1.644	1.710	1.644	1.782	1.637	1.600	1.670	3.58	
11)	bis(2-Chloroet...	1.396	1.319	1.380	1.315	1.428	1.295	1.269	1.343	4.36	
12)	1,3-Dichlorobe...	1.516	1.462	1.491	1.412	1.545	1.485	1.409	1.474	3.46	
13) C	1,4-Dichlorobe...	1.619	1.544	1.564	1.450	1.591	1.525	1.444	1.534	4.34	
14)	1,2-Dichlorobe...	1.503	1.456	1.492	1.402	1.540	1.450	1.392	1.462	3.68	
15)	Benzyl Alcohol	1.072	1.045	1.152	1.141	1.267	1.102	1.101	1.126	6.43	
16)	2,2'-oxybis(1-...	1.012	0.949	0.974	0.909	0.978	0.863	0.845	0.933	6.68	
17)	2-Methylphenol	1.078	1.042	1.133	1.108	1.207	1.077	1.068	1.102	4.96	
18)	Hexachloroethane	0.564	0.544	0.552	0.533	0.593	0.559	0.543	0.556	3.51	
19) P	n-Nitroso-di-n...	0.847	0.944	0.947	1.072	1.001	1.096	0.906	0.913	0.966	8.82
20)	3+4-Methylphenols		1.367	1.375	1.533	1.509	1.661	1.437	1.438	1.474	6.98
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.515	0.506	0.517	0.476	0.519	0.495	0.470	0.500	4.03	
23) S	Nitrobenzene-d5	0.359	0.362	0.392	0.377	0.414	0.406	0.382	0.385	5.42	
24)	Nitrobenzene	0.333	0.337	0.359	0.339	0.370	0.363	0.347	0.350	4.10	
25)	Isophorone	0.647	0.624	0.685	0.659	0.733	0.652	0.646	0.664	5.33	
26) C	2-Nitrophenol	0.116	0.124	0.146	0.155	0.178	0.169	0.169	0.151	15.65	
27)	2,4-Dimethylph...	0.311	0.298	0.315	0.303	0.332	0.313	0.299	0.310	3.80	
28)	bis(2-Chloroet...	0.426	0.413	0.447	0.427	0.470	0.432	0.420	0.434	4.42	
29) C	2,4-Dichloroph...	0.265	0.262	0.289	0.286	0.319	0.300	0.289	0.287	6.83	
30)	1,2,4-Trichlor...	0.324	0.307	0.321	0.305	0.337	0.330	0.312	0.319	3.72	
31)	Naphthalene	1.111	1.041	1.056	0.977	1.061	1.025	0.969	1.034	4.79	
32)	Benzoic acid		0.127	0.176	0.204	0.233	0.211	0.219	0.195	19.69	
33)	4-Chloroaniline	0.438	0.423	0.454	0.433	0.477	0.440	0.422	0.441	4.37	
34) C	Hexachlorobuta...	0.188	0.184	0.189	0.182	0.202	0.197	0.188	0.190	3.74	
35)	Caprolactam	0.077	0.077	0.093	0.097	0.112	0.092	0.090	0.091	13.48	
36) C	4-Chloro-3-met...	0.293	0.276	0.313	0.313	0.347	0.302	0.299	0.306	7.19	
37)	2-Methylnaphth...	0.625	0.592	0.643	0.615	0.679	0.619	0.594	0.624	4.78	
38)	1-Methylnaphth...	0.676	0.637	0.689	0.649	0.714	0.648	0.622	0.662	4.86	

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM060525.M


39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.586	0.561	0.585	0.541	0.589	0.607	0.574	0.578	3.71	
41) P	Hexachlorocycl...	0.295	0.305	0.339	0.341	0.382	0.401	0.392	0.351	12.02	
42) S	2,4,6-Tribromo...	0.210	0.213	0.238	0.231	0.254	0.233	0.214	0.227	6.99	
43) C	2,4,6-Trichlor...	0.346	0.347	0.379	0.375	0.414	0.402	0.396	0.380	6.93	
44)	2,4,5-Trichlor...	0.371	0.382	0.424	0.414	0.457	0.442	0.429	0.417	7.46	
45) S	2-Fluorobiphenyl	1.589	1.536	1.532	1.377	1.463	1.480	1.364	1.477	5.67	
46)	1,1'-Biphenyl	1.582	1.513	1.514	1.404	1.526	1.508	1.437	1.498	3.96	
47)	2-Chloronaphth...	1.221	1.194	1.180	1.084	1.171	1.177	1.123	1.164	3.94	
48)	2-Nitroaniline	0.200	0.206	0.252	0.268	0.302	0.290	0.277	0.256	15.56	
49)	Acenaphthylene	1.863	1.834	1.940	1.819	1.978	1.924	1.824	1.883	3.39	
50)	Dimethylphthalate	1.339	1.285	1.398	1.343	1.474	1.344	1.284	1.353	4.92	
51)	2,6-Dinitrotol...	0.204	0.231	0.284	0.289	0.320	0.297	0.284	0.273	14.82	
52) C	Acenaphthene	1.217	1.176	1.185	1.132	1.228	1.181	1.127	1.178	3.24	
53)	3-Nitroaniline	0.231	0.252	0.312	0.323	0.360	0.336	0.314	0.304	15.16	
54) P	2,4-Dinitrophenol		0.093	0.131	0.152	0.179	0.160	0.160	0.146	20.67	
55)	Dibenzofuran	1.782	1.710	1.775	1.662	1.797	1.719	1.617	1.723	3.87	
56) P	4-Nitrophenol	0.199	0.216	0.265	0.274	0.307	0.288	0.262	0.259	14.93	
57)	2,4-Dinitrotol...	0.250	0.283	0.365	0.396	0.447	0.398	0.380	0.360	19.28	
58)	Fluorene	1.413	1.358	1.394	1.267	1.348	1.289	1.169	1.320	6.40	
59)	2,3,4,6-Tetrac...	0.338	0.330	0.370	0.365	0.401	0.371	0.350	0.361	6.60	
60)	Diethylphthalate	1.314	1.272	1.416	1.346	1.490	1.317	1.238	1.342	6.43	
61)	4-Chlorophenyl...	0.675	0.656	0.667	0.611	0.660	0.623	0.573	0.638	5.81	
62)	4-Nitroaniline	0.213	0.236	0.291	0.308	0.342	0.319	0.285	0.285	16.00	
63)	Azobenzene	1.342	1.330	1.415	1.313	1.443	1.316	1.230	1.341	5.25	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.075	0.101	0.111	0.128	0.120	0.121	0.109	17.72	
66) c	n-Nitrosodiphe...	0.630	0.626	0.639	0.594	0.646	0.639	0.625	0.628	2.73	
67)	4-Bromophenyl-...	0.207	0.204	0.215	0.204	0.228	0.217	0.218	0.213	4.13	
68)	Hexachlorobenzene	0.249	0.240	0.253	0.238	0.261	0.255	0.249	0.249	3.29	
69)	Atrazine	0.166	0.177	0.206	0.206	0.231	0.212	0.204	0.200	10.86	
70) C	Pentachlorophenol	0.133	0.138	0.165	0.164	0.187	0.177	0.171	0.162	12.21	
71)	Phenanthrene	1.239	1.156	1.160	1.071	1.161	1.144	1.068	1.143	5.14	
72)	Anthracene	1.167	1.135	1.175	1.086	1.181	1.165	1.092	1.143	3.46	
73)	Carbazole	1.041	1.003	1.063	1.009	1.103	1.085	0.985	1.041	4.30	
74)	Di-n-butylphth...	0.998	1.007	1.187	1.159	1.300	1.187	1.093	1.133	9.53	
75) C	Fluoranthene	1.180	1.139	1.231	1.178	1.301	1.284	1.121	1.205	5.77	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine		0.381	0.568	0.617	0.689	0.600	0.520	0.562	18.64	
78)	Pyrene	1.331	1.307	1.444	1.324	1.423	1.274	1.334	1.348	4.61	
79) S	Terphenyl-d14	1.148	1.088	1.165	1.010	1.064	0.955	0.957	1.055	8.06	
80)	Butylbenzylpht...	0.322	0.347	0.460	0.490	0.563	0.480	0.485	0.450	18.95	
81)	Benzo(a)anthra...	1.271	1.231	1.298	1.218	1.338	1.280	1.225	1.266	3.47	
82)	3,3'-Dichlorob...	0.265	0.301	0.379	0.415	0.476	0.450	0.416	0.386	19.97	
83)	Chrysene	1.235	1.184	1.225	1.151	1.256	1.217	1.161	1.204	3.27	
84)	Bis(2-ethylhex...	0.525	0.574	0.702	0.740	0.851	0.741	0.714	0.693	15.84	
85) c	Di-n-octyl pht...	0.641	0.729	0.936	1.114	1.359	1.245	1.169	1.028	26.08	

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM060525.M

		-----ISTD-----								
86) I	Perylene-d12									
87)	Indeno(1,2,3-c...	1.189	1.183	1.254	1.191	1.318	1.472	1.407	1.288	8.99
88)	Benzo(b)fluora...	1.104	1.082	1.205	1.170	1.269	1.178	1.132	1.163	5.46
89)	Benzo(k)fluora...	1.141	1.149	1.222	1.167	1.295	1.190	1.144	1.187	4.71
90) C	Benzo(a)pyrene	0.994	1.007	1.111	1.081	1.197	1.153	1.111	1.093	6.73
91)	Dibenzo(a,h)an...	0.954	0.966	1.021	0.972	1.075	1.188	1.141	1.045	8.83
92)	Benzo(g,h,i)pe...	1.002	0.978	1.013	0.954	1.037	1.194	1.146	1.046	8.57

(#) = Out of Range

2,4-Dinitrophenol



$R = -9.297e-003 A^2 + 2.144e-001 A - 6.582e-002$
 Coef of Det (r^2) = 0.992237 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060525.M
 Calibration Table Last Updated: Thu Jun 05 16:20:25 2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050194.D
Acq On : 05 Jun 2025 09:20
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Jun 05 15:11:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.786	152	244555	20.000	ng	0.00
21) Naphthalene-d8	10.575	136	945759	20.000	ng	0.00
39) Acenaphthene-d10	14.415	164	534696	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	982061	20.000	ng	0.00
76) Chrysene-d12	21.386	240	992954	20.000	ng	0.00
86) Perylene-d12	24.374	264	1067519	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
18) Hexachloroethane	8.869	117	16914	2.490	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050194.D
Acq On : 05 Jun 2025 09:20
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5


Quant Time: Jun 05 15:11:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050194.D
Acq On : 05 Jun 2025 09:20
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Jun 05 15:11:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.786 min Scan# 81
Delta R.T. -0.007 min
Lab File: BM050194.D
Acq: 05 Jun 2025 09:20

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5


Tgt Ion:152 Resp: 244555
Ion Ratio Lower Upper
152 100
150 150.0 122.2 183.4
115 59.6 46.3 69.5



#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.37 min


Lab File: BM050194.D
Acq: 05 Jun 2025 09:20

Tgt Ion: 112
Sig Exp Ratio
112 100
64 49.7
63 24.8



Ion 64.00 (63.70 to 64.70): BM050194.D\data.ms

Ion 63.00 (62.70 to 63.70): BM050194.D\data.ms



#7

Phenol-d6

Concen: 0.000 ng

Expected RT: 6.95 min

Lab File: BM050194.D

Acq: 05 Jun 2025 09:20

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5



Tgt Ion: 99

Sig Exp Ratio

99 100

42 8.6

71 31.6



#18

Hexachloroethane

Concen: 2.490 ng

RT: 8.869 min Scan# 1000

Delta R.T. -0.001 min

Lab File: BM050194.D

Acq: 05 Jun 2025 09:20




Tgt Ion: 117 Resp: 16914


Ion Ratio Lower Upper

117 100

119 96.5 77.0 115.6

201 87.0 74.6 111.8





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.575 min Scan# 1

Delta R.T. -0.001 min

Lab File: BM050194.D


Acq: 05 Jun 2025 09:20

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5



Tgt Ion:136 Resp: 945759




Ion Ratio Lower Upper

136 100

137 10.8 8.6 13.0

54 5.0 3.8 5.8

68 6.0 4.9 7.3



#23

Nitrobenzene-d5

Concen: 0.000 ng

Expected RT: 8.94 min

Lab File: BM050194.D

Acq: 05 Jun 2025 09:20


Tgt Ion: 82

Sig Exp Ratio

82 100


128 44.0

54 30.6



Ion 128.00 (127.70 to 128.70): BM050194.D\data.ms

Ion 54.00 (53.70 to 54.70): BM050194.D\data.ms



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.415 min Scan# 1943

Delta R.T. -0.001 min

Lab File: BM050194.D


Acq: 05 Jun 2025 09:20

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5






Tgt Ion:164 Resp: 534696

Ion Ratio Lower Upper

164 100

162 100.1 80.2 120.4

160 45.0 35.7 53.5



#42

2,4,6-Tribromophenol

Concen: 0.000 ng

Expected RT: 15.90 min

Lab File: BM050194.D

Acq: 05 Jun 2025 09:20


Tgt Ion: 330


Sig Exp Ratio

330 100

332 96.9

141 36.0







#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 13.05 min
Lab File: BM050194.D
Acq: 05 Jun 2025 09:20




Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5


Tgt Ion: 172
Sig Exp Ratio
172 100
171 34.7
170 23.2



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. -0.001 min
Lab File: BM050194.D
Acq: 05 Jun 2025 09:20

Tgt Ion: 188 Resp: 982061
Ion Ratio Lower Upper
188 100
94 10.3 8.1 12.1
80 11.1 8.6 13.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 31
Delta R.T. -0.001 min
Lab File: BM050194.D
Acq: 05 Jun 2025 09:20

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5


Tgt Ion:240 Resp: 992954
Ion Ratio Lower Upper
240 100
120 11.5 9.0 13.4
236 25.4 20.7 31.1




#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.78 min

Lab File: BM050194.D
Acq: 05 Jun 2025 09:20

Tgt Ion: 244
Sig Exp Ratio
244 100
212 7.5
122 10.7





#86

Perylene-d12

Concen: 20.000 ng

RT: 24.374 min Scan# 30

Delta R.T. -0.001 min

Lab File: BM050194.D


Acq: 05 Jun 2025 09:20

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5





Tgt Ion:264 Resp: 1067519

Ion Ratio Lower Upper

264 100

260 23.1 18.6 28.0

265 21.0 16.8 25.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050195.D
 Acq On : 05 Jun 2025 09:59
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 15:12:05 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	231885	20.000	ng	0.00
21) Naphthalene-d8	10.575	136	935638	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	551980	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	1022508	20.000	ng	0.00
76) Chrysene-d12	21.386	240	964608	20.000	ng	0.00
86) Perylene-d12	24.374	264	941239	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	145075	10.436	ng	0.00
7) Phenol-d6	6.946	99	181940	9.943	ng	0.00
23) Nitrobenzene-d5	8.934	82	167883	9.332	ng	0.00
42) 2,4,6-Tribromophenol	15.892	330	57984	9.236	ng	0.00
45) 2-Fluorobiphenyl	13.039	172	438411	10.753	ng	0.00
79) Terphenyl-d14	19.780	244	553477	10.873	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	34616	5.681	ng	98
3) Pyridine	3.687	79	78460	4.972	ng	95
4) n-Nitrosodimethylamine	3.599	42	16377	5.019	ng	# 94
6) Aniline	7.110	93	117867	4.872	ng	99
8) 2-Chlorophenol	7.351	128	74751	4.889	ng	98
9) Benzaldehyde	6.922	77	67142m	5.771	ng	
10) Phenol	6.969	94	97004	5.010	ng	98
11) bis(2-Chloroethyl)ether	7.210	93	80951	5.198	ng	98
12) 1,3-Dichlorobenzene	7.681	146	87889	5.142	ng	98
13) 1,4-Dichlorobenzene	7.822	146	93841	5.276	ng	99
14) 1,2-Dichlorobenzene	8.140	146	87131	5.139	ng	99
15) Benzyl Alcohol	8.016	79	62171	4.764	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.316	45	58670	5.423	ng	97
17) 2-Methylphenol	8.222	107	62467	4.890	ng	99
18) Hexachloroethane	8.869	117	32689	5.075	ng	96
19) n-Nitroso-di-n-propyla...	8.587	70	54735	4.888	ng	98
20) 3+4-Methylphenols	8.545	107	79245	4.636	ng	98
22) Acetophenone	8.604	105	120380	5.150	ng	# 97
24) Nitrobenzene	8.975	77	77828	4.757	ng	98
25) Isophorone	9.504	82	151295	4.872	ng	99
26) 2-Nitrophenol	9.687	139	27114	3.839	ng	94
27) 2,4-Dimethylphenol	9.751	122	72768	5.010	ng	99
28) bis(2-Chloroethoxy)met...	9.992	93	99703	4.915	ng	100
29) 2,4-Dichlorophenol	10.216	162	61926	4.611	ng	99
30) 1,2,4-Trichlorobenzene	10.439	180	75854	5.076	ng	97
31) Naphthalene	10.628	128	259768	5.370	ng	99
33) 4-Chloroaniline	10.728	127	102387	4.964	ng	98
34) Hexachlorobutadiene	10.922	225	43914	4.940	ng	95
35) Caprolactam	11.463	113	17903	4.205	ng	87
36) 4-Chloro-3-methylphenol	11.845	107	68614	4.788	ng	94
37) 2-Methylnaphthalene	12.239	142	146149	5.008	ng	98
38) 1-Methylnaphthalene	12.457	142	158007	5.101	ng	97
40) 1,2,4,5-Tetrachloroben...	12.604	216	80884	5.075	ng	99
41) Hexachlorocyclopentadiene	12.598	237	40674	4.202	ng	97
43) 2,4,6-Trichlorophenol	12.839	196	47806	4.561	ng	100
44) 2,4,5-Trichlorophenol	12.910	196	51196	4.449	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050195.D
 Acq On : 05 Jun 2025 09:59
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 15:12:05 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.251	154	218302	5.282	ng	99
47) 2-Chloronaphthalene	13.286	162	168466	5.243	ng	96
48) 2-Nitroaniline	13.480	65	27545	3.895	ng	83
49) Acenaphthylene	14.133	152	257055	4.946	ng	99
50) Dimethylphthalate	13.875	163	184777	4.950	ng	99
51) 2,6-Dinitrotoluene	13.980	165	28146	3.739	ng	88
52) Acenaphthene	14.480	154	167876	5.164	ng	97
53) 3-Nitroaniline	14.304	138	31831	3.794	ng	# 88
55) Dibenzofuran	14.810	168	245911	5.171	ng	99
56) 4-Nitrophenol	14.604	139	27420	3.840	ng	97
57) 2,4-Dinitrotoluene	14.763	165	34457	3.470	ng	87
58) Fluorene	15.463	166	195018	5.354	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.033	232	46644m	4.685	ng	
60) Diethylphthalate	15.239	149	181363	4.897	ng	98
61) 4-Chlorophenyl-phenyle...	15.463	204	93141	5.291	ng	98
62) 4-Nitroaniline	15.463	138	29436	3.743	ng	99
63) Azobenzene	15.751	77	185154	5.001	ng	96
66) n-Nitrosodiphenylamine	15.669	169	161093	5.014	ng	99
67) 4-Bromophenyl-phenylether	16.351	248	52965	4.857	ng	99
68) Hexachlorobenzene	16.463	284	63698	5.000	ng	98
69) Atrazine	16.616	200	42483	4.150	ng	96
70) Pentachlorophenol	16.798	266	33971	4.101	ng	95
71) Phenanthrene	17.192	178	316683	5.421	ng	99
72) Anthracene	17.286	178	298346	5.106	ng	99
73) Carbazole	17.545	167	266214	5.001	ng	99
74) Di-n-butylphthalate	18.133	149	255200	4.406	ng	99
75) Fluoranthene	19.209	202	301599	4.896	ng	99
78) Pyrene	19.568	202	320928	4.936	ng	99
80) Butylbenzylphthalate	20.480	149	77663	3.581	ng	92
81) Benzo(a)anthracene	21.368	228	306613	5.022	ng	98
82) 3,3'-Dichlorobenzidine	21.286	252	64021	3.438	ng	94
83) Chrysene	21.427	228	297925	5.130	ng	99
84) Bis(2-ethylhexyl)phtha...	21.321	149	126652	3.791	ng	# 97
85) Di-n-octyl phthalate	22.462	149	154664	3.120	ng	# 97
87) Indeno(1,2,3-cd)pyrene	27.732	276	279736	4.616	ng	# 90
88) Benzo(b)fluoranthene	23.433	252	259803	4.748	ng	99
89) Benzo(k)fluoranthene	23.492	252	268597	4.808	ng	99
90) Benzo(a)pyrene	24.233	252	233858	4.545	ng	99
91) Dibenzo(a,h)anthracene	27.809	278	224374	4.561	ng	98
92) Benzo(g,h,i)perylene	28.779	276	235785	4.789	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050195.D
Acq On : 05 Jun 2025 09:59
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 15:12:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050195.D
Acq On : 05 Jun 2025 09:59
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025


Quant Time: Jun 05 15:12:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.787 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:152 Resp: 23188
Ion Ratio Lower Upper
152 100
150 154.0 122.2 183.4
115 58.3 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#2
1,4-Dioxane
Concen: 5.681 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion: 88 Resp: 34616
Ion Ratio Lower Upper
88 100
58 55.0 42.9 64.3
43 13.7 11.0 16.6









#3
Pyridine
Concen: 4.972 ng
RT: 3.687 min Scan# 118
Delta R.T. 0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 78460
Ion Ratio Lower Upper
79 100
52 42.5 32.1 48.1
51 23.0 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#4
n-Nitrosodimethylamine
Concen: 5.019 ng
RT: 3.599 min Scan# 104
Delta R.T. 0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion: 42 Resp: 16377
Ion Ratio Lower Upper
42 100
74 269.3 222.0 333.0
44 25.9 7.0 10.6





#5
2-Fluorophenol
Concen: 10.436 ng
RT: 5.369 min Scan# 405
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion: 112 Resp: 145079
Ion Ratio Lower Upper
112 100
64 47.5 39.8 59.6
63 25.0 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#6
Aniline
Concen: 4.872 ng
RT: 7.110 min Scan# 701
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion: 93 Resp: 117867
Ion Ratio Lower Upper
93 100
66 34.8 27.4 41.2
65 18.1 13.6 20.4









#7
Phenol-d6
Concen: 9.943 ng
RT: 6.946 min Scan# 674
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Tgt Ion: 99 Resp: 181940
Ion Ratio Lower Upper
99 100
42 8.4 6.9 10.3
71 31.7 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#8
2-Chlorophenol
Concen: 4.889 ng
RT: 7.351 min Scan# 742
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:128 Resp: 74751
Ion Ratio Lower Upper
128 100
130 30.7 12.6 52.6
64 39.9 19.8 59.8





#9

Benzaldehyde

Concen: 5.771 ng m

RT: 6.922 min Scan# 6

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion: 77 Resp: 6714

Ion Ratio Lower Upper

77 100

105 97.9 80.9 120.9





106 92.0 77.1 117.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#10

Phenol

Concen: 5.010 ng

RT: 6.969 min Scan# 677

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion: 94 Resp: 97004


Ion Ratio Lower Upper

94 100

65 25.6 4.3 44.3

66 33.5 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 5.198 ng
RT: 7.210 min Scan# 718
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 8095
Ion Ratio Lower Upper
93 100
63 64.3 43.9 83.9
95 30.3 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#12
1,3-Dichlorobenzene
Concen: 5.142 ng
RT: 7.681 min Scan# 798
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:146 Resp: 87889
Ion Ratio Lower Upper
146 100
148 62.9 51.9 77.9
75 29.3 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 5.276 ng

RT: 7.822 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion:146 Resp: 9384

Ion Ratio Lower Upper

146 100

148 63.4 51.4 77.0





111 39.7 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#14

1,2-Dichlorobenzene

Concen: 5.139 ng

RT: 8.140 min Scan# 876

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion:146 Resp: 87131


Ion Ratio Lower Upper

146 100

148 61.7 50.4 75.6

111 42.2 34.3 51.5





#15

Benzyl Alcohol

Concen: 4.764 ng

RT: 8.016 min Scan# 855

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:	79	Resp:	6217
Ion	Ratio	Lower	Upper
79	100		
108	79.7	64.0	96.0
77	68.5	52.2	78.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 5.423 ng




RT: 8.316 min Scan# 906


Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Tgt Ion:	45	Resp:	58670
Ion	Ratio	Lower	Upper
45	100		
77	26.4	8.7	48.7
79	21.9	2.7	42.7





#17

2-Methylphenol

Concen: 4.890 ng

RT: 8.222 min Scan# 890

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion:107 Resp: 6246

Ion Ratio Lower Upper

107 100

108 114.1 90.8 136.2

77 45.7 35.7 53.5





79 43.2 35.0 52.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#18

Hexachloroethane

Concen: 5.075 ng

RT: 8.869 min Scan# 1000

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion:117 Resp: 32689


Ion Ratio Lower Upper

117 100

119 101.1 77.0 115.6

201 89.6 74.6 111.8





#19

n-Nitroso-di-n-propylamine

Concen: 4.888 ng

RT: 8.587 min Scan# 91

Delta R.T. -0.012 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion: 70 Resp: 5473

Ion Ratio Lower Upper

70 100

42 28.5 22.6 33.8

101 9.5 8.3 12.5





130 20.6 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#20

3+4-Methylphenols

Concen: 4.636 ng

RT: 8.545 min Scan# 945

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Tgt Ion:107 Resp: 79245




Ion Ratio Lower Upper


107 100

108 88.8 70.4 110.4

77 35.1 12.2 52.2


79 28.3 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.575 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:136 Resp: 935638
Ion Ratio Lower Upper
136 100
137 10.9 8.6 13.0
54 5.4 3.8 5.8
68 5.9 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#22
Acetophenone
Concen: 5.150 ng
RT: 8.604 min Scan# 955
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:105 Resp: 120380
Ion Ratio Lower Upper
105 100
71 2.2 5.0 7.4#
51 14.8 11.4 17.2
120 21.9 18.2 27.4





#23

Nitrobenzene-d5

Concen: 9.332 ng

RT: 8.934 min Scan# 1011

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion: 82 Resp: 16788

Ion Ratio Lower Upper

82 100

128 42.4 35.2 52.8





54 30.8 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#24

Nitrobenzene

Concen: 4.757 ng

RT: 8.975 min Scan# 1018

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion: 77 Resp: 77828


Ion Ratio Lower Upper

77 100

123 46.3 38.2 57.2


65 12.3 9.5 14.3





#25
Isophorone
Concen: 4.872 ng
RT: 9.504 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion: 82 Resp: 15129
Ion Ratio Lower Upper
82 100
95 7.0 5.7 8.5
138 17.2 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#26
2-Nitrophenol
Concen: 3.839 ng
RT: 9.687 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:139 Resp: 27114
Ion Ratio Lower Upper
139 100
109 30.3 23.3 34.9
65 49.2 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 5.010 ng

RT: 9.751 min Scan# 1150

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion:122 Resp: 72768

Ion Ratio Lower Upper

122 100

107 112.9 91.0 136.6





121 55.1 45.8 68.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#28

bis(2-Chloroethoxy)methane

Concen: 4.915 ng

RT: 9.992 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion: 93 Resp: 99703


Ion Ratio Lower Upper

93 100

95 33.1 26.2 39.4

123 11.2 8.9 13.3





#29

2,4-Dichlorophenol

Concen: 4.611 ng

RT: 10.216 min Scan# 1229

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion:162 Resp: 61920

Ion Ratio Lower Upper

162 100

164 64.0 45.0 85.0





98 35.7 16.3 56.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#30

1,2,4-Trichlorobenzene

Concen: 5.076 ng

RT: 10.439 min Scan# 1267

Delta R.T. -0.000 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion:180 Resp: 75854


Ion Ratio Lower Upper

180 100

182 91.9 76.6 114.8


145 29.0 24.2 36.4





#31
Naphthalene
Concen: 5.370 ng
RT: 10.628 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:128 Resp: 259768
Ion Ratio Lower Upper
128 100
129 11.4 8.6 12.8
127 13.2 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#33
4-Chloroaniline
Concen: 4.964 ng
RT: 10.728 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:127 Resp: 102387
Ion Ratio Lower Upper
127 100
129 31.0 25.4 38.0
65 27.5 21.0 31.4
92 17.1 14.8 22.2





#34
Hexachlorobutadiene
Concen: 4.940 ng
RT: 10.922 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:225 Resp: 43914
Ion Ratio Lower Upper
225 100
223 67.4 50.2 75.2
227 66.4 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#35
Caprolactam
Concen: 4.205 ng
RT: 11.463 min Scan# 1441
Delta R.T. -0.036 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:113 Resp: 17903
Ion Ratio Lower Upper
113 100
55 116.6 83.5 123.5
56 91.9 60.6 100.6





#36
4-Chloro-3-methylphenol
Concen: 4.788 ng
RT: 11.845 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:107 Resp: 68614
Ion Ratio Lower Upper
107 100
144 24.4 21.0 31.6
142 74.6 64.8 97.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#37
2-Methylnaphthalene
Concen: 5.008 ng
RT: 12.239 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Tgt Ion:142 Resp: 146149
Ion Ratio Lower Upper
142 100
141 90.8 71.5 107.3
115 36.8 27.8 41.8





#38
1-Methylnaphthalene
Concen: 5.101 ng
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:142 Resp: 15800
Ion Ratio Lower Upper
142 100
141 88.2 72.9 109.3
116 3.6 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:164 Resp: 551980
Ion Ratio Lower Upper
164 100
162 101.5 80.2 120.4
160 45.9 35.7 53.5









#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.075 ng
RT: 12.604 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Tgt Ion:216 Resp: 80884
Ion Ratio Lower Upper
216 100
214 77.9 63.0 94.4
179 21.5 16.6 25.0
108 20.1 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#41
Hexachlorocyclopentadiene
Concen: 4.202 ng
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:237 Resp: 40674
Ion Ratio Lower Upper
237 100
235 67.0 44.6 84.6
272 13.6 0.0 33.1





#42
2,4,6-Tribromophenol
Concen: 9.236 ng
RT: 15.892 min Scan# 2195
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :





SSTDICC005

Tgt Ion:330 Resp: 57984
Ion Ratio Lower Upper
330 100
332 97.0 77.5 116.3
141 38.0 28.8 43.2

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#43
2,4,6-Trichlorophenol
Concen: 4.561 ng
RT: 12.839 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:196 Resp: 47806
Ion Ratio Lower Upper
196 100
198 96.0 76.6 115.0
200 31.1 24.5 36.7





#44
2,4,5-Trichlorophenol
Concen: 4.449 ng
RT: 12.910 min Scan# 1709
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:196 Resp: 51190
Ion Ratio Lower Upper
196 100
198 96.9 76.1 114.1
97 54.5 39.3 58.9
132 28.7 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#45
2-Fluorobiphenyl
Concen: 10.753 ng
RT: 13.039 min Scan# 1709
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:172 Resp: 438411
Ion Ratio Lower Upper
172 100
171 34.0 27.8 41.6
170 22.8 18.6 27.8





#46
1,1'-Biphenyl
Concen: 5.282 ng
RT: 13.251 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:154 Resp: 218302
Ion Ratio Lower Upper
154 100
153 38.7 19.2 59.2
76 12.6 0.0 32.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#47
2-Chloronaphthalene
Concen: 5.243 ng
RT: 13.286 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Tgt Ion:162 Resp: 168466
Ion Ratio Lower Upper
162 100
127 36.7 31.1 46.7
164 30.7 26.5 39.7





#48
2-Nitroaniline
Concen: 3.895 ng
RT: 13.480 min Scan# 1784
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion: 65 Resp: 2754
Ion Ratio Lower Upper
65 100
92 68.2 64.2 96.4
138 104.7 101.0 151.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#49
Acenaphthylene
Concen: 4.946 ng
RT: 14.133 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:152 Resp: 257055
Ion Ratio Lower Upper
152 100
151 19.2 15.8 23.6
153 12.9 10.4 15.6





#50

Dimethylphthalate

Concen: 4.950 ng

RT: 13.875 min Scan# 1851

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion:163 Resp: 18477

Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9





164 10.1 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#51

2,6-Dinitrotoluene

Concen: 3.739 ng

RT: 13.980 min Scan# 1869

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59




Tgt Ion:165 Resp: 28146


Ion Ratio Lower Upper

165 100

63 52.9 35.8 53.6


89 64.5 44.8 67.2





#52
Acenaphthene
Concen: 5.164 ng
RT: 14.480 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:154 Resp: 167870
Ion Ratio Lower Upper
154 100
153 114.5 88.7 133.1
152 55.6 42.6 63.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#53
3-Nitroaniline
Concen: 3.794 ng
RT: 14.304 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Tgt Ion:138 Resp: 31831
Ion Ratio Lower Upper
138 100
108 14.4 8.8 13.2#
92 129.9 92.9 139.3





#55
Dibenzenofuran
Concen: 5.171 ng
RT: 14.810 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:168 Resp: 24591
Ion Ratio Lower Upper
168 100
139 38.2 29.9 44.9
169 13.6 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#56
4-Nitrophenol
Concen: 3.840 ng
RT: 14.604 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:139 Resp: 27420
Ion Ratio Lower Upper
139 100
109 72.2 52.6 92.6
65 88.6 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 3.470 ng
RT: 14.763 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:165 Resp: 3445
Ion Ratio Lower Upper
165 100
63 42.5 28.4 42.6
89 84.3 58.0 87.0
182 2.8 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#58
Fluorene
Concen: 5.354 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:166 Resp: 195018
Ion Ratio Lower Upper
166 100
165 93.9 77.0 115.4
167 13.3 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 4.685 ng m
RT: 15.033 min Scan# 2049
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:232 Resp: 4664
Ion Ratio Lower Upper
232 100
131 42.0 32.9 49.3
130 2.5 1.8 2.8
166 24.8 20.8 31.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#60
Diethylphthalate
Concen: 4.897 ng
RT: 15.239 min Scan# 2083
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Tgt Ion:149 Resp: 181363
Ion Ratio Lower Upper
149 100
177 20.5 17.2 25.8
150 12.0 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 5.291 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:204 Resp: 9314
Ion Ratio Lower Upper
204 100
206 32.9 25.4 38.0
141 58.1 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#62
4-Nitroaniline
Concen: 3.743 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.012 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:138 Resp: 29436
Ion Ratio Lower Upper
138 100
92 49.4 28.8 68.8
108 76.5 58.1 98.1





#63
Azobenzene
Concen: 5.001 ng
RT: 15.751 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion: 77 Resp: 185154
Ion Ratio Lower Upper
77 100
182 23.5 6.8 46.8
105 22.3 1.7 41.7
51 14.9 0.0 36.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Tgt Ion:188 Resp: 1022508
Ion Ratio Lower Upper
188 100
94 10.0 8.1 12.1
80 11.1 8.6 13.0





#66
n-Nitrosodiphenylamine
Concen: 5.014 ng
RT: 15.669 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:169 Resp: 16109
Ion Ratio Lower Upper
169 100
168 67.9 53.1 79.7
167 35.9 29.0 43.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#67
4-Bromophenyl-phenylether
Concen: 4.857 ng
RT: 16.351 min Scan# 2272
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:248 Resp: 52965
Ion Ratio Lower Upper
248 100
250 95.4 78.0 117.0
141 75.7 60.1 90.1





#68
Hexachlorobenzene
Concen: 5.000 ng
RT: 16.463 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:284 Resp: 63698
Ion Ratio Lower Upper
284 100
142 35.6 28.8 43.2
249 30.3 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#69
Atrazine
Concen: 4.150 ng
RT: 16.616 min Scan# 2317
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59



Tgt Ion:200 Resp: 42483
Ion Ratio Lower Upper
200 100
173 29.5 6.0 46.0
215 50.0 28.1 68.1









#70
Pentachlorophenol
Concen: 4.101 ng
RT: 16.798 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Tgt Ion:266 Resp: 3397
Ion Ratio Lower Upper
266 100
268 57.5 51.0 76.6
264 61.8 50.8 76.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#71
Phenanthrene
Concen: 5.421 ng
RT: 17.192 min Scan# 2415
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:178 Resp: 316683
Ion Ratio Lower Upper
178 100
176 19.8 15.6 23.4
179 15.6 12.3 18.5





#72
Anthracene
Concen: 5.106 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:178 Resp: 298340
Ion Ratio Lower Upper
178 100
176 18.3 14.8 22.2
179 15.8 12.3 18.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#73
Carbazole
Concen: 5.001 ng
RT: 17.545 min Scan# 2475
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Tgt Ion:167 Resp: 266214
Ion Ratio Lower Upper
167 100
166 20.4 17.0 25.6
139 12.9 10.2 15.4





#74
Di-n-butylphthalate
Concen: 4.406 ng
RT: 18.133 min Scan# 2575
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:149 Resp: 255200
Ion Ratio Lower Upper
149 100
150 8.8 7.4 11.0
104 5.6 4.0 6.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#75
Fluoranthene
Concen: 4.896 ng
RT: 19.209 min Scan# 2758
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:202 Resp: 301599
Ion Ratio Lower Upper
202 100
101 11.1 0.0 31.1
203 16.7 0.0 37.0





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:240 Resp: 964608
Ion Ratio Lower Upper
240 100
120 10.9 9.0 13.4
236 25.8 20.7 31.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#78
Pyrene
Concen: 4.936 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:202 Resp: 320928
Ion Ratio Lower Upper
202 100
200 20.3 16.7 25.1
203 17.4 14.1 21.1





#79
 Terphenyl-d14
 Concen: 10.873 ng
 RT: 19.780 min Scan# 2855
 Delta R.T. -0.000 min
 Lab File: BM050195.D
 Acq: 05 Jun 2025 09:59




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005



Tgt Ion:244 Resp: 55347
 Ion Ratio Lower Upper
 244 100
 212 6.8 6.0 9.0
 122 11.4 8.6 12.8




Manual Integrations
 APPROVED


Reviewed By :Rahul Chavli 06/06/2025
 Supervised By :Jagrut Upadhyay 06/06/2025



#80
 Butylbenzylphthalate
 Concen: 3.581 ng
 RT: 20.480 min Scan# 2974
 Delta R.T. -0.000 min
 Lab File: BM050195.D
 Acq: 05 Jun 2025 09:59


Tgt Ion:149 Resp: 77663
 Ion Ratio Lower Upper
 149 100
 91 78.8 56.9 85.3
 206 19.3 16.1 24.1





#81
Benzo(a)anthracene
Concen: 5.022 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:228 Resp: 306611
Ion Ratio Lower Upper
228 100
226 25.8 21.7 32.5
229 19.5 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#82
3,3'-Dichlorobenzidine
Concen: 3.438 ng
RT: 21.286 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:252 Resp: 64021
Ion Ratio Lower Upper
252 100
254 59.5 51.4 77.0
126 12.1 8.6 12.8





#83
Chrysene
Concen: 5.130 ng
RT: 21.427 min Scan# 3135
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:228 Resp: 297925
Ion Ratio Lower Upper
228 100
226 30.0 24.3 36.5
229 19.9 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 3.791 ng
RT: 21.321 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59


Tgt Ion:149 Resp: 126652
Ion Ratio Lower Upper
149 100
167 26.6 22.3 33.5
279 3.3 3.5 5.3#





#85
Di-n-octyl phthalate
Concen: 3.120 ng
RT: 22.462 min Scan# 3311
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005



Tgt Ion:149 Resp: 154664
Ion Ratio Lower Upper
149 100
167 1.7 1.2 1.8
43 5.9 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:264 Resp: 941239
Ion Ratio Lower Upper
264 100
260 23.6 18.6 28.0
265 21.6 16.8 25.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 4.616 ng
RT: 27.732 min Scan# 4211
Delta R.T. -0.024 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 279730
Ion Ratio Lower Upper
276 100
138 16.0 20.6 30.8
277 24.6 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#88
Benzo(b)fluoranthene
Concen: 4.748 ng
RT: 23.433 min Scan# 3476
Delta R.T. -0.000 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59

Tgt Ion:252 Resp: 259803
Ion Ratio Lower Upper
252 100
253 22.0 17.2 25.8
125 10.5 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 4.808 ng

RT: 23.492 min Scan# 34

Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59

Instrument :

BNA_M

ClientSampleId :





SSTDICC005

Tgt Ion:252 Resp: 26859
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.2
125 11.0 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#90

Benzo(a)pyrene

Concen: 4.545 ng




RT: 24.233 min Scan# 3612


Delta R.T. -0.006 min

Lab File: BM050195.D

Acq: 05 Jun 2025 09:59


Tgt Ion:252 Resp: 233858
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 12.5 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 4.561 ng
RT: 27.809 min Scan# 41
Delta R.T. -0.012 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59




Instrument :
BNA_M
ClientSampleId :
SSTDICC005




Tgt Ion:278 Resp: 224374
Ion Ratio Lower Upper
278 100
139 17.8 12.8 19.2
279 22.7 18.6 28.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#92
Benzo(g,h,i)perylene
Concen: 4.789 ng
RT: 28.779 min Scan# 4385
Delta R.T. -0.018 min
Lab File: BM050195.D
Acq: 05 Jun 2025 09:59



Tgt Ion:276 Resp: 235785
Ion Ratio Lower Upper
276 100
277 22.5 18.6 27.8
138 22.2 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050196.D
 Acq On : 05 Jun 2025 10:38
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:04 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	257321	20.000	ng	0.00
21) Naphthalene-d8	10.575	136	1031213	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	597701	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	1092132	20.000	ng	0.00
76) Chrysene-d12	21.386	240	1044209	20.000	ng	0.00
86) Perylene-d12	24.374	264	1027174	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	306899	19.895	ng	0.00
7) Phenol-d6	6.946	99	391479	19.279	ng	0.00
23) Nitrobenzene-d5	8.934	82	373577	18.841	ng	0.00
42) 2,4,6-Tribromophenol	15.898	330	127081	18.693	ng	0.00
45) 2-Fluorobiphenyl	13.039	172	918043	20.795	ng	0.00
79) Terphenyl-d14	19.780	244	1136477	20.625	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	71029	10.505	ng	97
3) Pyridine	3.687	79	173970	9.935	ng	98
4) n-Nitrosodimethylamine	3.599	42	34098	9.417	ng	# 99
6) Aniline	7.110	93	259330	9.660	ng	99
8) 2-Chlorophenol	7.351	128	161982	9.547	ng	99
9) Benzaldehyde	6.928	77	139011	10.767	ng	99
10) Phenol	6.969	94	211509	9.844	ng	99
11) bis(2-Chloroethyl)ether	7.210	93	169669	9.818	ng	99
12) 1,3-Dichlorobenzene	7.681	146	188081	9.915	ng	98
13) 1,4-Dichlorobenzene	7.822	146	198637	10.065	ng	99
14) 1,2-Dichlorobenzene	8.140	146	187370	9.959	ng	99
15) Benzyl Alcohol	8.022	79	134400	9.280	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.316	45	122149	10.175	ng	97
17) 2-Methylphenol	8.222	107	134124	9.461	ng	100
18) Hexachloroethane	8.869	117	70045	9.800	ng	98
19) n-Nitroso-di-n-propyla...	8.592	70	121867	9.807	ng	97
20) 3+4-Methylphenols	8.545	107	176847	9.324	ng	99
22) Acetophenone	8.604	105	260808	10.123	ng	# 98
24) Nitrobenzene	8.975	77	173945	9.646	ng	99
25) Isophorone	9.504	82	321944	9.407	ng	99
26) 2-Nitrophenol	9.687	139	64131	8.239	ng	96
27) 2,4-Dimethylphenol	9.751	122	153659	9.598	ng	99
28) bis(2-Chloroethoxy)met...	9.992	93	213076	9.531	ng	99
29) 2,4-Dichlorophenol	10.216	162	135013	9.122	ng	95
30) 1,2,4-Trichlorobenzene	10.439	180	158337	9.614	ng	99
31) Naphthalene	10.628	128	536516	10.063	ng	99
32) Benzoic acid	9.810	122	65412	6.508	ng	99
33) 4-Chloroaniline	10.728	127	218016	9.591	ng	99
34) Hexachlorobutadiene	10.922	225	94772	9.674	ng	98
35) Caprolactam	11.469	113	39590	8.437	ng	95
36) 4-Chloro-3-methylphenol	11.845	107	142482	9.022	ng	98
37) 2-Methylnaphthalene	12.239	142	305385	9.495	ng	97
38) 1-Methylnaphthalene	12.457	142	328309	9.617	ng	99
40) 1,2,4,5-Tetrachloroben...	12.610	216	167598	9.711	ng	98
41) Hexachlorocyclopentadiene	12.598	237	91203	8.701	ng	95
43) 2,4,6-Trichlorophenol	12.845	196	103556	9.124	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050196.D
 Acq On : 05 Jun 2025 10:38
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:04 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.910	196	114017	9.149	ng	98
46) 1,1'-Biphenyl	13.251	154	452133	10.102	ng	100
47) 2-Chloronaphthalene	13.286	162	356687	10.251	ng	99
48) 2-Nitroaniline	13.480	65	61491	8.031	ng	89
49) Acenaphthylene	14.133	152	548052	9.738	ng	99
50) Dimethylphthalate	13.874	163	384143	9.503	ng	99
51) 2,6-Dinitrotoluene	13.980	165	69112	8.478	ng	95
52) Acenaphthene	14.480	154	351485	9.985	ng	99
53) 3-Nitroaniline	14.304	138	75418	8.301	ng	97
54) 2,4-Dinitrophenol	14.510	184	27888	10.741	ng	98
55) Dibenzofuran	14.816	168	511101	9.925	ng	98
56) 4-Nitrophenol	14.604	139	64450	8.335	ng	99
57) 2,4-Dinitrotoluene	14.763	165	84481	7.856	ng	91
58) Fluorene	15.463	166	405716	10.286	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.033	232	98715m	9.156	ng	
60) Diethylphthalate	15.239	149	380099	9.477	ng	100
61) 4-Chlorophenyl-phenyle...	15.463	204	196034	10.283	ng	100
62) 4-Nitroaniline	15.463	138	70517	8.281	ng	97
63) Azobenzene	15.751	77	397398	9.913	ng	97
65) 4,6-Dinitro-2-methylph...	15.527	198	40699	6.823	ng	94
66) n-Nitrosodiphenylamine	15.668	169	341695	9.957	ng	99
67) 4-Bromophenyl-phenylether	16.351	248	111429	9.568	ng	96
68) Hexachlorobenzene	16.463	284	130925	9.622	ng	99
69) Atrazine	16.621	200	96720	8.846	ng	98
70) Pentachlorophenol	16.798	266	75315	8.513	ng	100
71) Phenanthrene	17.192	178	630983	10.114	ng	99
72) Anthracene	17.286	178	619530	9.926	ng	99
73) Carbazole	17.551	167	547511	9.629	ng	100
74) Di-n-butylphthalate	18.133	149	549702	8.885	ng	99
75) Fluoranthene	19.204	202	622189	9.457	ng	98
77) Benzidine	19.392	184	199082	6.780	ng	99
78) Pyrene	19.568	202	682132	9.692	ng	99
80) Butylbenzylphthalate	20.480	149	181298	7.723	ng	96
81) Benzo(a)anthracene	21.368	228	642785	9.725	ng	98
82) 3,3'-Dichlorobenzidine	21.292	252	157266	7.803	ng	99
83) Chrysene	21.427	228	618398	9.836	ng	99
84) Bis(2-ethylhexyl)phtha...	21.321	149	299821	8.291	ng	96
85) Di-n-octyl phthalate	22.462	149	380681	7.094	ng	99
87) Indeno(1,2,3-cd)pyrene	27.738	276	607722	9.189	ng	# 92
88) Benzo(b)fluoranthene	23.433	252	555747	9.306	ng	98
89) Benzo(k)fluoranthene	23.497	252	589937	9.677	ng	99
90) Benzo(a)pyrene	24.233	252	517175	9.211	ng	99
91) Dibenzo(a,h)anthracene	27.809	278	496324	9.245	ng	98
92) Benzo(g,h,i)perylene	28.785	276	502232	9.347	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050196.D
 Acq On : 05 Jun 2025 10:38
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025


Quant Time: Jun 05 14:58:04 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050196.D
Acq On : 05 Jun 2025 10:38
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025


Quant Time: Jun 05 14:58:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.787 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:152 Resp: 25732
Ion Ratio Lower Upper
152 100
150 156.3 122.2 183.4
115 59.6 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#2
1,4-Dioxane
Concen: 10.505 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion: 88 Resp: 71029
Ion Ratio Lower Upper
88 100
58 55.7 42.9 64.3
43 13.5 11.0 16.6









#3
Pyridine
Concen: 9.935 ng
RT: 3.687 min Scan# 118
Delta R.T. 0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 173970
Ion Ratio Lower Upper
79 100
52 41.0 32.1 48.1
51 20.2 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#4
n-Nitrosodimethylamine
Concen: 9.417 ng
RT: 3.599 min Scan# 104
Delta R.T. 0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion: 42 Resp: 34098
Ion Ratio Lower Upper
42 100
74 277.9 222.0 333.0
44 13.2 7.0 10.6





#5
2-Fluorophenol
Concen: 19.895 ng
RT: 5.369 min Scan# 405
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion: 112 Resp: 306899
Ion Ratio Lower Upper
112 100
64 50.5 39.8 59.6
63 25.5 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#6
Aniline
Concen: 9.660 ng
RT: 7.110 min Scan# 701
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion: 93 Resp: 259330
Ion Ratio Lower Upper
93 100
66 34.9 27.4 41.2
65 17.0 13.6 20.4









#7
Phenol-d6
Concen: 19.279 ng
RT: 6.946 min Scan# 674
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 391479
Ion Ratio Lower Upper
99 100
42 8.2 6.9 10.3
71 32.1 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#8
2-Chlorophenol
Concen: 9.547 ng
RT: 7.351 min Scan# 742
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:128 Resp: 161982
Ion Ratio Lower Upper
128 100
130 32.1 12.6 52.6
64 40.9 19.8 59.8









#9
Benzaldehyde
Concen: 10.767 ng
RT: 6.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 13901
Ion Ratio Lower Upper
77 100
105 102.1 80.9 120.9
106 97.0 77.1 117.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#10
Phenol
Concen: 9.844 ng
RT: 6.969 min Scan# 677
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion: 94 Resp: 211509
Ion Ratio Lower Upper
94 100
65 24.3 4.3 44.3
66 33.5 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 9.818 ng
RT: 7.210 min Scan# 718
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 169669
Ion Ratio Lower Upper
93 100
63 64.7 43.9 83.9
95 32.8 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#12
1,3-Dichlorobenzene
Concen: 9.915 ng
RT: 7.681 min Scan# 798
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:146 Resp: 188081
Ion Ratio Lower Upper
146 100
148 63.8 51.9 77.9
75 30.6 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 10.065 ng

RT: 7.822 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Tgt Ion:146 Resp: 19863

Ion Ratio Lower Upper

146 100

148 63.3 51.4 77.0





111 41.0 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#14

1,2-Dichlorobenzene

Concen: 9.959 ng

RT: 8.140 min Scan# 876

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38




Tgt Ion:146 Resp: 187370


Ion Ratio Lower Upper

146 100

148 64.2 50.4 75.6


111 43.0 34.3 51.5





#15
Benzyl Alcohol
Concen: 9.280 ng
RT: 8.022 min Scan# 856
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010




Tgt Ion: 79 Resp: 134400
Ion Ratio Lower Upper
79 100
108 79.3 64.0 96.0
77 66.3 52.2 78.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.175 ng
RT: 8.316 min Scan# 906
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38



Tgt Ion: 45 Resp: 122149
Ion Ratio Lower Upper
45 100
77 29.8 8.7 48.7
79 20.5 2.7 42.7





#17

2-Methylphenol

Concen: 9.461 ng

RT: 8.222 min Scan# 890

Delta R.T. -0.000 min

Lab File: BM050196.D


Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010



Tgt Ion:107 Resp: 134124

Ion Ratio Lower Upper

107 100

108 113.5 90.8 136.2

77 45.1 35.7 53.5




79 44.1 35.0 52.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#18

Hexachloroethane

Concen: 9.800 ng

RT: 8.869 min Scan# 1000

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38




Tgt Ion:117 Resp: 70045


Ion Ratio Lower Upper

117 100

119 100.5 77.0 115.6


201 93.4 74.6 111.8





#19
n-Nitroso-di-n-propylamine
Concen: 9.807 ng
RT: 8.592 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010




Tgt Ion: 70 Resp: 12186
Ion Ratio Lower Upper
70 100
42 27.4 22.6 33.8
101 9.4 8.3 12.5
130 19.8 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#20
3+4-Methylphenols
Concen: 9.324 ng
RT: 8.545 min Scan# 945
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Tgt Ion:107 Resp: 176847
Ion Ratio Lower Upper
107 100
108 89.5 70.4 110.4
77 33.1 12.2 52.2
79 28.9 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.575 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:136 Resp: 1031211
Ion Ratio Lower Upper
136 100
137 11.0 8.6 13.0
54 5.1 3.8 5.8
68 5.9 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#22
Acetophenone
Concen: 10.123 ng
RT: 8.604 min Scan# 955
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:105 Resp: 260808
Ion Ratio Lower Upper
105 100
71 2.6 5.0 7.4#
51 14.4 11.4 17.2
120 22.1 18.2 27.4





#23
Nitrobenzene-d5
Concen: 18.841 ng
RT: 8.934 min Scan# 1011
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010




Tgt Ion: 82 Resp: 37357
Ion Ratio Lower Upper
82 100
128 41.8 35.2 52.8
54 31.1 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#24
Nitrobenzene
Concen: 9.646 ng
RT: 8.975 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Tgt Ion: 77 Resp: 173945
Ion Ratio Lower Upper
77 100
123 46.7 38.2 57.2
65 11.9 9.5 14.3





#25
Isophorone
Concen: 9.407 ng
RT: 9.504 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion: 82 Resp: 321944
Ion Ratio Lower Upper
82 100
95 7.6 5.7 8.5
138 17.9 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#26
2-Nitrophenol
Concen: 8.239 ng
RT: 9.687 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:139 Resp: 64131
Ion Ratio Lower Upper
139 100
109 29.4 23.3 34.9
65 47.8 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 9.598 ng

RT: 9.751 min Scan# 1150

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Tgt Ion:122 Resp: 153659

Ion Ratio Lower Upper

122 100

107 111.9 91.0 136.6





121 56.9 45.8 68.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#28

bis(2-Chloroethoxy)methane

Concen: 9.531 ng

RT: 9.992 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38




Tgt Ion: 93 Resp: 213076


Ion Ratio Lower Upper

93 100

95 33.2 26.2 39.4


123 11.2 8.9 13.3





#29
2,4-Dichlorophenol
Concen: 9.122 ng
RT: 10.216 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:162 Resp: 13501
Ion Ratio Lower Upper
162 100
164 60.2 45.0 85.0
98 38.7 16.3 56.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#30
1,2,4-Trichlorobenzene
Concen: 9.614 ng
RT: 10.439 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:180 Resp: 158337
Ion Ratio Lower Upper
180 100
182 94.0 76.6 114.8
145 30.5 24.2 36.4





#33
4-Chloroaniline
Concen: 9.591 ng
RT: 10.728 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:127 Resp: 218010
Ion Ratio Lower Upper
127 100
129 33.0 25.4 38.0
65 26.0 21.0 31.4
92 18.2 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#34
Hexachlorobutadiene
Concen: 9.674 ng
RT: 10.922 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:225 Resp: 94772
Ion Ratio Lower Upper
225 100
223 60.2 50.2 75.2
227 62.5 51.0 76.4





#35
Caprolactam
Concen: 8.437 ng
RT: 11.469 min Scan# 1447
Delta R.T. -0.030 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:113 Resp: 39590
Ion Ratio Lower Upper
113 100
55 108.8 83.5 123.5
56 85.0 60.6 100.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#36
4-Chloro-3-methylphenol
Concen: 9.022 ng
RT: 11.845 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:107 Resp: 142482
Ion Ratio Lower Upper
107 100
144 24.5 21.0 31.6
142 82.1 64.8 97.2





#37
2-Methylnaphthalene
Concen: 9.495 ng
RT: 12.239 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:142 Resp: 305385
Ion Ratio Lower Upper
142 100
141 86.7 71.5 107.3
115 33.5 27.8 41.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#38
1-Methylnaphthalene
Concen: 9.617 ng
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:142 Resp: 328309
Ion Ratio Lower Upper
142 100
141 91.8 72.9 109.3
116 3.8 3.0 4.6





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.416 min Scan# 1943

Delta R.T. -0.000 min

Lab File: BM050196.D


Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010



Tgt Ion:164 Resp: 59770

Ion Ratio Lower Upper

164 100

162 100.7 80.2 120.4




160 45.5 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 9.711 ng

RT: 12.610 min Scan# 1636

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38

Tgt Ion:216 Resp: 167598




Ion Ratio Lower Upper


216 100

214 80.8 63.0 94.4

179 21.5 16.6 25.0


108 19.9 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 8.701 ng
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010




Tgt Ion: 237 Resp: 9120
Ion Ratio Lower Upper
237 100
235 60.2 44.6 84.6
272 12.6 0.0 33.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#42
2,4,6-Tribromophenol
Concen: 18.693 ng
RT: 15.898 min Scan# 2195
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38



Tgt Ion: 330 Resp: 127081
Ion Ratio Lower Upper
330 100
332 96.3 77.5 116.3
141 37.3 28.8 43.2





#43

2,4,6-Trichlorophenol

Concen: 9.124 ng

RT: 12.845 min Scan# 1676

Delta R.T. -0.000 min

Lab File: BM050196.D


Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010



Tgt Ion:196 Resp: 103550

Ion Ratio Lower Upper

196 100

198 96.1 76.6 115.0




200 30.0 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#44

2,4,5-Trichlorophenol


Concen: 9.149 ng

RT: 12.910 min Scan# 1687

Delta R.T. -0.006 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38



Tgt Ion:196 Resp: 114017



Ion Ratio Lower Upper


196 100

198 97.1 76.1 114.1

97 51.1 39.3 58.9


132 27.6 21.4 32.0





#45
2-Fluorobiphenyl
Concen: 20.795 ng
RT: 13.039 min Scan# 1709
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:172 Resp: 91804
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.6
170 22.7 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#46
1,1'-Biphenyl
Concen: 10.102 ng
RT: 13.251 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:154 Resp: 452133
Ion Ratio Lower Upper
154 100
153 39.3 19.2 59.2
76 12.6 0.0 32.0





#47

2-Chloronaphthalene

Concen: 10.251 ng

RT: 13.286 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM050196.D


Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010



Tgt Ion:162 Resp: 35668

Ion Ratio Lower Upper

162 100

127 39.6 31.1 46.7




164 32.9 26.5 39.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#48

2-Nitroaniline


Concen: 8.031 ng

RT: 13.480 min Scan# 1784

Delta R.T. -0.006 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38





Tgt Ion: 65 Resp: 61491


Ion Ratio Lower Upper

65 100

92 75.4 64.2 96.4


138 110.0 101.0 151.4





#49
Acenaphthylene
Concen: 9.738 ng
RT: 14.133 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:152 Resp: 548052
Ion Ratio Lower Upper
152 100
151 18.8 15.8 23.6
153 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#50
Dimethylphthalate
Concen: 9.503 ng
RT: 13.874 min Scan# 1851
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:163 Resp: 384143
Ion Ratio Lower Upper
163 100
194 3.7 3.3 4.9
164 10.1 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 8.478 ng
RT: 13.980 min Scan# 1869
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:165 Resp: 69112
Ion Ratio Lower Upper
165 100
63 48.2 35.8 53.6
89 58.8 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#52
Acenaphthene
Concen: 9.985 ng
RT: 14.480 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:154 Resp: 351485
Ion Ratio Lower Upper
154 100
153 110.1 88.7 133.1
152 54.3 42.6 63.8





#53
3-Nitroaniline
Concen: 8.301 ng
RT: 14.304 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:138 Resp: 75418
Ion Ratio Lower Upper
138 100
108 12.7 8.8 13.2
92 118.6 92.9 139.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#54
2,4-Dinitrophenol
Concen: 10.741 ng
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:184 Resp: 27888
Ion Ratio Lower Upper
184 100
63 58.6 45.3 67.9
154 68.6 54.3 81.5





#55
Dibenzofuran
Concen: 9.925 ng
RT: 14.816 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:168 Resp: 51110
Ion Ratio Lower Upper
168 100
139 36.4 29.9 44.9
169 12.8 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#56
4-Nitrophenol
Concen: 8.335 ng
RT: 14.604 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:139 Resp: 64450
Ion Ratio Lower Upper
139 100
109 71.9 52.6 92.6
65 85.6 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 7.856 ng
RT: 14.763 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:165 Resp: 8448
Ion Ratio Lower Upper
165 100
63 40.6 28.4 42.6
89 80.7 58.0 87.0
182 2.2 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#58
Fluorene
Concen: 10.286 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:166 Resp: 405716
Ion Ratio Lower Upper
166 100
165 97.6 77.0 115.4
167 12.9 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 9.156 ng m
RT: 15.033 min Scan# 2049
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010




Tgt Ion:232 Resp: 98715
Ion Ratio Lower Upper
232 100
131 42.6 32.9 49.3
130 2.4 1.8 2.8
166 25.3 20.8 31.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#60
Diethylphthalate
Concen: 9.477 ng
RT: 15.239 min Scan# 2083
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Tgt Ion:149 Resp: 380099
Ion Ratio Lower Upper
149 100
177 21.5 17.2 25.8
150 12.3 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 10.283 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:204 Resp: 196034
Ion Ratio Lower Upper
204 100
206 31.8 25.4 38.0
141 59.4 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#62
4-Nitroaniline
Concen: 8.281 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.012 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:138 Resp: 70517
Ion Ratio Lower Upper
138 100
92 50.6 28.8 68.8
108 75.4 58.1 98.1





#63
Azobenzene
Concen: 9.913 ng
RT: 15.751 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion: 77 Resp: 397398
Ion Ratio Lower Upper
77 100
182 24.5 6.8 46.8
105 22.5 1.7 41.7
51 15.4 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:188 Resp: 1092132
Ion Ratio Lower Upper
188 100
94 10.1 8.1 12.1
80 10.9 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 6.823 ng
RT: 15.527 min Scan# 2132
Delta R.T. -0.007 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:198 Resp: 40695
Ion Ratio Lower Upper
198 100
51 26.7 4.5 44.5
105 49.2 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#66
n-Nitrosodiphenylamine
Concen: 9.957 ng
RT: 15.668 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:169 Resp: 341695
Ion Ratio Lower Upper
169 100
168 66.0 53.1 79.7
167 35.0 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 9.568 ng
RT: 16.351 min Scan# 2272
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:248 Resp: 111429
Ion Ratio Lower Upper
248 100
250 100.9 78.0 117.0
141 78.2 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#68
Hexachlorobenzene
Concen: 9.622 ng
RT: 16.463 min Scan# 2291
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:284 Resp: 130925
Ion Ratio Lower Upper
284 100
142 36.3 28.8 43.2
249 31.1 25.8 38.6





#69
Atrazine
Concen: 8.846 ng
RT: 16.621 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:200 Resp: 96720
Ion Ratio Lower Upper
200 100
173 26.4 6.0 46.0
215 49.6 28.1 68.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#70
Pentachlorophenol
Concen: 8.513 ng
RT: 16.798 min Scan# 2348
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:266 Resp: 75315
Ion Ratio Lower Upper
266 100
268 63.6 51.0 76.6
264 63.1 50.8 76.2





#71
Phenanthrene
Concen: 10.114 ng
RT: 17.192 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:178 Resp: 63098
Ion Ratio Lower Upper
178 100
176 19.1 15.6 23.4
179 15.2 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#72
Anthracene
Concen: 9.926 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:178 Resp: 619530
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.2 12.3 18.5





#73
Carbazole
Concen: 9.629 ng
RT: 17.551 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:167 Resp: 54751
Ion Ratio Lower Upper
167 100
166 21.4 17.0 25.6
139 12.8 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#74
Di-n-butylphthalate
Concen: 8.885 ng
RT: 18.133 min Scan# 2575
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:149 Resp: 549702
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.0
104 5.7 4.0 6.0





#75
Fluoranthene
Concen: 9.457 ng
RT: 19.204 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:202 Resp: 622189
Ion Ratio Lower Upper
202 100
101 12.6 0.0 31.1
203 17.3 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:240 Resp: 1044209
Ion Ratio Lower Upper
240 100
120 11.1 9.0 13.4
236 24.8 20.7 31.1





#77
Benzidine
Concen: 6.780 ng
RT: 19.392 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:184 Resp: 19908
Ion Ratio Lower Upper
184 100
185 14.8 11.5 17.3
183 11.0 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#78
Pyrene
Concen: 9.692 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:202 Resp: 682132
Ion Ratio Lower Upper
202 100
200 20.8 16.7 25.1
203 17.1 14.1 21.1





#79
Terphenyl-d14
Concen: 20.625 ng
RT: 19.780 min Scan# 2855
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:244 Resp: 113647
Ion Ratio Lower Upper
244 100
212 7.0 6.0 9.0
122 11.7 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#80
Butylbenzylphthalate
Concen: 7.723 ng
RT: 20.480 min Scan# 2974
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:149 Resp: 181298
Ion Ratio Lower Upper
149 100
91 74.5 56.9 85.3
206 18.8 16.1 24.1





#81
Benzo(a)anthracene
Concen: 9.725 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:228 Resp: 642785
Ion Ratio Lower Upper
228 100
226 25.9 21.7 32.5
229 19.3 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#82
3,3'-Dichlorobenzidine
Concen: 7.803 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:252 Resp: 157266
Ion Ratio Lower Upper
252 100
254 63.4 51.4 77.0
126 11.5 8.6 12.8





#83

Chrysene

Concen: 9.836 ng

RT: 21.427 min Scan# 3135

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Tgt Ion:228 Resp: 618398

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5





229 19.4 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 8.291 ng

RT: 21.321 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38




Tgt Ion:149 Resp: 299821


Ion Ratio Lower Upper

149 100

167 25.6 22.3 33.5


279 3.8 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 7.094 ng
RT: 22.462 min Scan# 3311
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:149 Resp: 38068
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 5.2 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38


Tgt Ion:264 Resp: 1027174
Ion Ratio Lower Upper
264 100
260 23.4 18.6 28.0
265 21.4 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 9.189 ng
RT: 27.738 min Scan# 4211
Delta R.T. -0.018 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38




Instrument :
BNA_M
ClientSampleId :
SSTDICC010



Tgt Ion:276 Resp: 60772
Ion Ratio Lower Upper
276 100
138 17.5 20.6 30.8
277 24.7 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#88
Benzo(b)fluoranthene
Concen: 9.306 ng
RT: 23.433 min Scan# 3476
Delta R.T. -0.000 min
Lab File: BM050196.D
Acq: 05 Jun 2025 10:38

Tgt Ion:252 Resp: 555747
Ion Ratio Lower Upper
252 100
253 22.6 17.2 25.8
125 10.1 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 9.677 ng

RT: 23.497 min Scan# 34

Delta R.T. -0.000 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Tgt Ion:252 Resp: 58993

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2





125 11.1 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#90

Benzo(a)pyrene

Concen: 9.211 ng

RT: 24.233 min Scan# 3612

Delta R.T. -0.006 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38




Tgt Ion:252 Resp: 517175


Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 12.2 9.0 13.4





#91

Dibenzo(a,h)anthracene

Concen: 9.245 ng

RT: 27.809 min Scan# 41

Delta R.T. -0.012 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38

Instrument :

BNA_M

ClientSampleId :

SSTDICC010

Tgt Ion:278 Resp: 496324

Ion Ratio Lower Upper

278 100

139 16.7 12.8 19.2





279 22.6 18.6 28.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#92

Benzo(g,h,i)perylene

Concen: 9.347 ng

RT: 28.785 min Scan# 4386

Delta R.T. -0.012 min

Lab File: BM050196.D

Acq: 05 Jun 2025 10:38




Tgt Ion:276 Resp: 502232

Ion Ratio Lower Upper

276 100

277 23.7 18.6 27.8

138 23.8 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050197.D
 Acq On : 05 Jun 2025 11:17
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:25 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	277976	20.000	ng	0.00
21) Naphthalene-d8	10.575	136	1171357	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	720692	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	1364843	20.000	ng	0.00
76) Chrysene-d12	21.386	240	1240830	20.000	ng	0.00
86) Perylene-d12	24.374	264	1156806	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	662532	39.759	ng	0.00
7) Phenol-d6	6.946	99	899139	40.989	ng	0.00
23) Nitrobenzene-d5	8.934	82	917335	40.730	ng	0.00
42) 2,4,6-Tribromophenol	15.898	330	342402	41.770	ng	0.00
45) 2-Fluorobiphenyl	13.045	172	2208837	41.494	ng	0.00
79) Terphenyl-d14	19.780	244	2892092	44.169	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	143316	19.620	ng	99
3) Pyridine	3.681	79	380901	20.136	ng	99
4) n-Nitrosodimethylamine	3.593	42	77508	19.815	ng	# 97
6) Aniline	7.110	93	596745	20.578	ng	100
8) 2-Chlorophenol	7.352	128	373594	20.382	ng	99
9) Benzaldehyde	6.928	77	299039m	21.441	ng	
10) Phenol	6.975	94	475205	20.474	ng	100
11) bis(2-Chloroethyl)ether	7.210	93	383515	20.543	ng	99
12) 1,3-Dichlorobenzene	7.681	146	414543	20.230	ng	98
13) 1,4-Dichlorobenzene	7.822	146	434797	20.394	ng	99
14) 1,2-Dichlorobenzene	8.140	146	414769	20.407	ng	99
15) Benzyl Alcohol	8.022	79	320142	20.462	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.316	45	270693	20.873	ng	98
17) 2-Methylphenol	8.222	107	314991	20.568	ng	97
18) Hexachloroethane	8.869	117	153532	19.885	ng	99
19) n-Nitroso-di-n-propyla...	8.593	70	297971	22.197	ng	98
20) 3+4-Methylphenols	8.546	107	426134	20.798	ng	98
22) Acetophenone	8.605	105	605945	20.706	ng	# 98
24) Nitrobenzene	8.975	77	420036	20.506	ng	97
25) Isophorone	9.504	82	802320	20.639	ng	100
26) 2-Nitrophenol	9.687	139	170594	19.294	ng	97
27) 2,4-Dimethylphenol	9.752	122	368697	20.275	ng	98
28) bis(2-Chloroethoxy)met...	9.993	93	523393	20.610	ng	99
29) 2,4-Dichlorophenol	10.222	162	338221	20.118	ng	99
30) 1,2,4-Trichlorobenzene	10.440	180	375562	20.074	ng	99
31) Naphthalene	10.628	128	1236626	20.419	ng	100
32) Benzoic acid	9.840	122	205878	18.032	ng	99
33) 4-Chloroaniline	10.728	127	531493	20.585	ng	99
34) Hexachlorobutadiene	10.922	225	221558	19.910	ng	99
35) Caprolactam	11.481	113	109175	20.483	ng	98
36) 4-Chloro-3-methylphenol	11.851	107	366554	20.433	ng	99
37) 2-Methylnaphthalene	12.240	142	752886	20.607	ng	98
38) 1-Methylnaphthalene	12.457	142	806687	20.803	ng	99
40) 1,2,4,5-Tetrachloroben...	12.610	216	421559	20.257	ng	100
41) Hexachlorocyclopentadiene	12.598	237	244251	19.326	ng	98
43) 2,4,6-Trichlorophenol	12.845	196	273306	19.971	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050197.D
 Acq On : 05 Jun 2025 11:17
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:25 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jun 05 14:55:18 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.910	196	305541	20.334	ng	98
46) 1,1'-Biphenyl	13.251	154	1091201	20.221	ng	98
47) 2-Chloronaphthalene	13.292	162	850349	20.269	ng	99
48) 2-Nitroaniline	13.481	65	181276	19.634	ng	96
49) Acenaphthylene	14.139	152	1398350	20.607	ng	99
50) Dimethylphthalate	13.875	163	1007787	20.677	ng	100
51) 2,6-Dinitrotoluene	13.981	165	204779	20.834	ng	96
52) Acenaphthene	14.481	154	854295	20.128	ng	99
53) 3-Nitroaniline	14.304	138	224884	20.528	ng	97
54) 2,4-Dinitrophenol	14.510	184	94248	19.130	ng	98
55) Dibenzofuran	14.816	168	1279459	20.605	ng	98
56) 4-Nitrophenol	14.604	139	191313	20.518	ng	100
57) 2,4-Dinitrotoluene	14.769	165	263119	20.293	ng	99
58) Fluorene	15.463	166	1004752	21.126	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.034	232	266423	20.493	ng	92
60) Diethylphthalate	15.245	149	1020737	21.107	ng	99
61) 4-Chlorophenyl-phenyle...	15.463	204	480887	20.921	ng	99
62) 4-Nitroaniline	15.463	138	210030	20.456	ng	98
63) Azobenzene	15.751	77	1019944	21.101	ng	97
65) 4,6-Dinitro-2-methylph...	15.528	198	138015	18.514	ng	97
66) n-Nitrosodiphenylamine	15.669	169	872651	20.349	ng	99
67) 4-Bromophenyl-phenylether	16.351	248	293588	20.172	ng	99
68) Hexachlorobenzene	16.463	284	345929	20.343	ng	99
69) Atrazine	16.622	200	281360	20.591	ng	98
70) Pentachlorophenol	16.798	266	225494	20.396	ng	98
71) Phenanthrene	17.192	178	1583183	20.305	ng	99
72) Anthracene	17.286	178	1603637	20.560	ng	99
73) Carbazole	17.551	167	1450582	20.414	ng	99
74) Di-n-butylphthalate	18.133	149	1619797	20.950	ng	100
75) Fluoranthene	19.210	202	1679948	20.431	ng	99
77) Benzidine	19.392	184	704620	20.194	ng	99
78) Pyrene	19.569	202	1791963	21.427	ng	100
80) Butylbenzylphthalate	20.480	149	570581	20.453	ng	98
81) Benzo(a)anthracene	21.363	228	1610203	20.502	ng	100
82) 3,3'-Dichlorobenzidine	21.286	252	470294	19.636	ng	98
83) Chrysene	21.427	228	1519821	20.343	ng	99
84) Bis(2-ethylhexyl)phtha...	21.315	149	871329	20.277	ng	98
85) Di-n-octyl phthalate	22.457	149	1161953	18.223	ng	99
87) Indeno(1,2,3-cd)pyrene	27.739	276	1450383	19.472	ng	# 91
88) Benzo(b)fluoranthene	23.427	252	1393549	20.720	ng	99
89) Benzo(k)fluoranthene	23.492	252	1413975	20.594	ng	100
90) Benzo(a)pyrene	24.233	252	1284647	20.315	ng	98
91) Dibenzo(a,h)anthracene	27.809	278	1181274	19.538	ng	99
92) Benzo(g,h,i)perylene	28.791	276	1171360	19.358	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050197.D
 Acq On : 05 Jun 2025 11:17
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDICC020


Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050197.D
Acq On : 05 Jun 2025 11:17
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025


Quant Time: Jun 05 14:58:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.787 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:152 Resp: 277970
Ion Ratio Lower Upper
152 100
150 155.9 122.2 183.4
115 59.9 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#2
1,4-Dioxane
Concen: 19.620 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion: 88 Resp: 143316
Ion Ratio Lower Upper
88 100
58 54.4 42.9 64.3
43 13.7 11.0 16.6





#3
Pyridine
Concen: 20.136 ng
RT: 3.681 min Scan# 118
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion: 79 Resp: 38090
Ion Ratio Lower Upper
79 100
52 41.1 32.1 48.1
51 19.2 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#4
n-Nitrosodimethylamine
Concen: 19.815 ng
RT: 3.593 min Scan# 103
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion: 42 Resp: 77508
Ion Ratio Lower Upper
42 100
74 272.5 222.0 333.0
44 10.8 7.0 10.6





#5
2-Fluorophenol
Concen: 39.759 ng
RT: 5.369 min Scan# 405
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:112 Resp: 66253
Ion Ratio Lower Upper
112 100
64 49.2 39.8 59.6
63 24.5 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#6
Aniline
Concen: 20.578 ng
RT: 7.110 min Scan# 701
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion: 93 Resp: 596745
Ion Ratio Lower Upper
93 100
66 34.5 27.4 41.2
65 16.9 13.6 20.4









#7
Phenol-d6
Concen: 40.989 ng
RT: 6.946 min Scan# 673
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 899139
Ion Ratio Lower Upper
99 100
42 8.4 6.9 10.3
71 32.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#8
2-Chlorophenol
Concen: 20.382 ng
RT: 7.352 min Scan# 742
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:128 Resp: 373594
Ion Ratio Lower Upper
128 100
130 32.9 12.6 52.6
64 39.1 19.8 59.8





#9

Benzaldehyde

Concen: 21.441 ng m

RT: 6.928 min Scan# 670

Delta R.T. 0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion: 77 Resp: 299039

Ion Ratio Lower Upper

77 100

105 99.2 80.9 120.9





106 95.0 77.1 117.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#10

Phenol

Concen: 20.474 ng

RT: 6.975 min Scan# 678

Delta R.T. -0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17




Tgt Ion: 94 Resp: 475205


Ion Ratio Lower Upper

94 100

65 24.1 4.3 44.3

66 32.7 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 20.543 ng
RT: 7.210 min Scan# 718
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 38351
Ion Ratio Lower Upper
93 100
63 65.3 43.9 83.9
95 32.4 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#12
1,3-Dichlorobenzene
Concen: 20.230 ng
RT: 7.681 min Scan# 798
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:146 Resp: 414543
Ion Ratio Lower Upper
146 100
148 66.4 51.9 77.9
75 30.9 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 20.394 ng

RT: 7.822 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion:146 Resp: 43479

Ion Ratio Lower Upper

146 100

148 64.4 51.4 77.0





111 41.0 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#14

1,2-Dichlorobenzene

Concen: 20.407 ng

RT: 8.140 min Scan# 876

Delta R.T. -0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17




Tgt Ion:146 Resp: 414769


Ion Ratio Lower Upper

146 100

148 63.9 50.4 75.6

111 43.2 34.3 51.5





#15

Benzyl Alcohol

Concen: 20.462 ng

RT: 8.022 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion: 79 Resp: 32014

Ion Ratio Lower Upper

79 100

108 81.8 64.0 96.0





77 65.2 52.2 78.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 20.873 ng

RT: 8.316 min Scan# 906

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17




Tgt Ion: 45 Resp: 270693


Ion Ratio Lower Upper

45 100

77 30.0 8.7 48.7


79 21.9 2.7 42.7





#17
2-Methylphenol
Concen: 20.568 ng
RT: 8.222 min Scan# 890
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:107 Resp: 31499
Ion Ratio Lower Upper
107 100
108 118.9 90.8 136.2
77 44.9 35.7 53.5
79 43.9 35.0 52.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#18
Hexachloroethane
Concen: 19.885 ng
RT: 8.869 min Scan# 1000
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:117 Resp: 153532
Ion Ratio Lower Upper
117 100
119 95.4 77.0 115.6
201 92.4 74.6 111.8





#19

n-Nitroso-di-n-propylamine

Concen: 22.197 ng

RT: 8.593 min Scan# 91

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion: 70 Resp: 29797

Ion Ratio Lower Upper

70 100

42 27.4 22.6 33.8

101 9.6 8.3 12.5





130 20.7 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#20

3+4-Methylphenols

Concen: 20.798 ng

RT: 8.546 min Scan# 945

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Tgt Ion:107 Resp: 426134




Ion Ratio Lower Upper


107 100

108 88.2 70.4 110.4

77 31.6 12.2 52.2


79 27.8 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.575 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:136 Resp: 117135
Ion Ratio Lower Upper
136 100
137 10.6 8.6 13.0
54 5.1 3.8 5.8
68 5.9 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#22
Acetophenone
Concen: 20.706 ng
RT: 8.605 min Scan# 955
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:105 Resp: 605945
Ion Ratio Lower Upper
105 100
71 3.5 5.0 7.4#
51 13.7 11.4 17.2
120 22.8 18.2 27.4





#23

Nitrobenzene-d5

Concen: 40.730 ng

RT: 8.934 min Scan# 1011

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion: 82 Resp: 91733

Ion Ratio Lower Upper

82 100

128 42.6 35.2 52.8





54 31.1 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#24

Nitrobenzene

Concen: 20.506 ng

RT: 8.975 min Scan# 1018

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17




Tgt Ion: 77 Resp: 420036


Ion Ratio Lower Upper

77 100

123 45.4 38.2 57.2


65 11.3 9.5 14.3





#25
Isophorone
Concen: 20.639 ng
RT: 9.504 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion: 82 Resp: 802320
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 17.5 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#26
2-Nitrophenol
Concen: 19.294 ng
RT: 9.687 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:139 Resp: 170594
Ion Ratio Lower Upper
139 100
109 29.4 23.3 34.9
65 47.1 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 20.275 ng

RT: 9.752 min Scan# 1150

Delta R.T. -0.000 min

Lab File: BM050197.D


Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020



Tgt Ion:122 Resp: 36869

Ion Ratio Lower Upper

122 100

107 110.9 91.0 136.6




121 56.7 45.8 68.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#28

bis(2-Chloroethoxy)methane


Concen: 20.610 ng

RT: 9.993 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17





Tgt Ion: 93 Resp: 523393


Ion Ratio Lower Upper

93 100

95 31.9 26.2 39.4


123 11.1 8.9 13.3





#29
2,4-Dichlorophenol
Concen: 20.118 ng
RT: 10.222 min Scan# 1230
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:162 Resp: 33822
Ion Ratio Lower Upper
162 100
164 64.8 45.0 85.0
98 37.0 16.3 56.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#30
1,2,4-Trichlorobenzene
Concen: 20.074 ng
RT: 10.440 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:180 Resp: 375562
Ion Ratio Lower Upper
180 100
182 96.2 76.6 114.8
145 30.6 24.2 36.4





#31
Naphthalene
Concen: 20.419 ng
RT: 10.628 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:128 Resp: 1236620
Ion Ratio Lower Upper
128 100
129 10.5 8.6 12.8
127 12.8 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#32
Benzoic acid
Concen: 18.032 ng
RT: 9.840 min Scan# 1165
Delta R.T. -0.035 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:122 Resp: 205878
Ion Ratio Lower Upper
122 100
105 123.3 103.4 143.4
77 93.5 70.3 110.3





#33
4-Chloroaniline
Concen: 20.585 ng
RT: 10.728 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:127 Resp: 53149
Ion Ratio Lower Upper
127 100
129 32.3 25.4 38.0
65 25.6 21.0 31.4
92 18.5 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#34
Hexachlorobutadiene
Concen: 19.910 ng
RT: 10.922 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:225 Resp: 221558
Ion Ratio Lower Upper
225 100
223 62.3 50.2 75.2
227 64.3 51.0 76.4





#35
Caprolactam
Concen: 20.483 ng
RT: 11.481 min Scan# 1447
Delta R.T. -0.018 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:113 Resp: 109175
Ion Ratio Lower Upper
113 100
55 106.2 83.5 123.5
56 82.3 60.6 100.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#36
4-Chloro-3-methylphenol
Concen: 20.433 ng
RT: 11.851 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:107 Resp: 366554
Ion Ratio Lower Upper
107 100
144 26.4 21.0 31.6
142 80.0 64.8 97.2





#37
2-Methylnaphthalene
Concen: 20.607 ng
RT: 12.240 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:142 Resp: 752880
Ion Ratio Lower Upper
142 100
141 87.7 71.5 107.3
115 34.3 27.8 41.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#38
1-Methylnaphthalene
Concen: 20.803 ng
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:142 Resp: 806687
Ion Ratio Lower Upper
142 100
141 91.7 72.9 109.3
116 4.1 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:164 Resp: 720692
Ion Ratio Lower Upper
164 100
162 99.0 80.2 120.4
160 44.8 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.257 ng
RT: 12.610 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:216 Resp: 421559
Ion Ratio Lower Upper
216 100
214 79.0 63.0 94.4
179 20.9 16.6 25.0
108 19.3 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 19.326 ng
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020




Tgt Ion: 237 Resp: 24425
Ion Ratio Lower Upper
237 100
235 62.6 44.6 84.6
272 12.5 0.0 33.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#42
2,4,6-Tribromophenol
Concen: 41.770 ng
RT: 15.898 min Scan# 2195
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Tgt Ion: 330 Resp: 342402
Ion Ratio Lower Upper
330 100
332 96.7 77.5 116.3
141 36.8 28.8 43.2





#43
2,4,6-Trichlorophenol
Concen: 19.971 ng
RT: 12.845 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:196 Resp: 273300
Ion Ratio Lower Upper
196 100
198 98.7 76.6 115.0
200 31.9 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#44
2,4,5-Trichlorophenol
Concen: 20.334 ng
RT: 12.910 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:196 Resp: 305541
Ion Ratio Lower Upper
196 100
198 95.6 76.1 114.1
97 52.4 39.3 58.9
132 27.8 21.4 32.0





#45
2-Fluorobiphenyl
Concen: 41.494 ng
RT: 13.045 min Scan# 1710
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:172 Resp: 220883
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.6
170 23.3 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#46
1,1'-Biphenyl
Concen: 20.221 ng
RT: 13.251 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:154 Resp: 1091201
Ion Ratio Lower Upper
154 100
153 40.1 19.2 59.2
76 12.9 0.0 32.0





#47
2-Chloronaphthalene
Concen: 20.269 ng
RT: 13.292 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:162 Resp: 850349
Ion Ratio Lower Upper
162 100
127 37.9 31.1 46.7
164 33.0 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#48
2-Nitroaniline
Concen: 19.634 ng
RT: 13.481 min Scan# 1784
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion: 65 Resp: 181276
Ion Ratio Lower Upper
65 100
92 79.9 64.2 96.4
138 119.4 101.0 151.4





#49
Acenaphthylene
Concen: 20.607 ng
RT: 14.139 min Scan# 1896
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:152 Resp: 1398350
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 12.7 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#50
Dimethylphthalate
Concen: 20.677 ng
RT: 13.875 min Scan# 1851
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:163 Resp: 1007787
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 20.834 ng
RT: 13.981 min Scan# 1870
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:165 Resp: 204779
Ion Ratio Lower Upper
165 100
63 47.7 35.8 53.6
89 59.1 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#52
Acenaphthene
Concen: 20.128 ng
RT: 14.481 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:154 Resp: 854295
Ion Ratio Lower Upper
154 100
153 112.4 88.7 133.1
152 54.0 42.6 63.8





#53
3-Nitroaniline
Concen: 20.528 ng
RT: 14.304 min Scan# 1925
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:138 Resp: 224884
Ion Ratio Lower Upper
138 100
108 11.4 8.8 13.2
92 119.3 92.9 139.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#54
2,4-Dinitrophenol
Concen: 19.130 ng
RT: 14.510 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:184 Resp: 94248
Ion Ratio Lower Upper
184 100
63 56.8 45.3 67.9
154 64.7 54.3 81.5





#55

Dibenzenofuran

Concen: 20.605 ng

RT: 14.816 min Scan# 20

Delta R.T. -0.000 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion:168 Resp: 127945

Ion Ratio Lower Upper

168 100

139 36.1 29.9 44.9





169 13.2 10.6 16.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#56

4-Nitrophenol

Concen: 20.518 ng

RT: 14.604 min Scan# 1975

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17




Tgt Ion:139 Resp: 191313


Ion Ratio Lower Upper

139 100

109 72.5 52.6 92.6


65 84.2 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 20.293 ng
RT: 14.769 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:165 Resp: 263119
Ion Ratio Lower Upper
165 100
63 35.0 28.4 42.6
89 73.2 58.0 87.0
182 2.9 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#58
Fluorene
Concen: 21.126 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:166 Resp: 1004752
Ion Ratio Lower Upper
166 100
165 96.3 77.0 115.4
167 13.3 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 20.493 ng
RT: 15.034 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:232 Resp: 266421
Ion Ratio Lower Upper
232 100
131 45.6 32.9 49.3
130 2.3 1.8 2.8
166 30.7 20.8 31.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#60
Diethylphthalate
Concen: 21.107 ng
RT: 15.245 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:149 Resp: 1020737
Ion Ratio Lower Upper
149 100
177 21.9 17.2 25.8
150 12.0 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 20.921 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:204 Resp: 48088
Ion Ratio Lower Upper
204 100
206 32.0 25.4 38.0
141 58.6 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#62
4-Nitroaniline
Concen: 20.456 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.012 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:138 Resp: 210030
Ion Ratio Lower Upper
138 100
92 51.7 28.8 68.8
108 77.8 58.1 98.1





#63
Azobenzene
Concen: 21.101 ng
RT: 15.751 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion: 77 Resp: 1019944
Ion Ratio Lower Upper
77 100
182 24.0 6.8 46.8
105 21.6 1.7 41.7
51 15.4 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:188 Resp: 1364843
Ion Ratio Lower Upper
188 100
94 9.7 8.1 12.1
80 10.9 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 18.514 ng
RT: 15.528 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:198 Resp: 138015
Ion Ratio Lower Upper
198 100
51 26.6 4.5 44.5
105 46.0 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#66
n-Nitrosodiphenylamine
Concen: 20.349 ng
RT: 15.669 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:169 Resp: 872651
Ion Ratio Lower Upper
169 100
168 66.8 53.1 79.7
167 36.8 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 20.172 ng
RT: 16.351 min Scan# 2272
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020




Tgt Ion:248 Resp: 29358
Ion Ratio Lower Upper
248 100
250 96.0 78.0 117.0
141 75.1 60.1 90.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#68
Hexachlorobenzene
Concen: 20.343 ng
RT: 16.463 min Scan# 2291
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Tgt Ion:284 Resp: 345929
Ion Ratio Lower Upper
284 100
142 36.5 28.8 43.2
249 31.7 25.8 38.6





#69
Atrazine
Concen: 20.591 ng
RT: 16.622 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:200 Resp: 281360
Ion Ratio Lower Upper
200 100
173 27.4 6.0 46.0
215 49.3 28.1 68.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#70
Pentachlorophenol
Concen: 20.396 ng
RT: 16.798 min Scan# 2348
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:266 Resp: 225494
Ion Ratio Lower Upper
266 100
268 62.3 51.0 76.6
264 61.9 50.8 76.2





#71
Phenanthrene
Concen: 20.305 ng
RT: 17.192 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:178 Resp: 158318
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.5 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#72
Anthracene
Concen: 20.560 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:178 Resp: 1603637
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 14.9 12.3 18.5





#73
Carbazole
Concen: 20.414 ng
RT: 17.551 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:167 Resp: 1450582
Ion Ratio Lower Upper
167 100
166 21.0 17.0 25.6
139 12.5 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#74
Di-n-butylphthalate
Concen: 20.950 ng
RT: 18.133 min Scan# 2575
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:149 Resp: 1619797
Ion Ratio Lower Upper
149 100
150 9.2 7.4 11.0
104 5.2 4.0 6.0





#75
Fluoranthene
Concen: 20.431 ng
RT: 19.210 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:202 Resp: 167994
Ion Ratio Lower Upper
202 100
101 10.7 0.0 31.1
203 16.8 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:240 Resp: 1240830
Ion Ratio Lower Upper
240 100
120 10.9 9.0 13.4
236 25.5 20.7 31.1





#77
Benzidine
Concen: 20.194 ng
RT: 19.392 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:184 Resp: 704620
Ion Ratio Lower Upper
184 100
185 14.1 11.5 17.3
183 11.0 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#78
Pyrene
Concen: 21.427 ng
RT: 19.569 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:202 Resp: 1791963
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.5 14.1 21.1





#79
Terphenyl-d14
Concen: 44.169 ng
RT: 19.780 min Scan# 2855
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:244 Resp: 2892092
Ion Ratio Lower Upper
244 100
212 7.0 6.0 9.0
122 11.1 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#80
Butylbenzylphthalate
Concen: 20.453 ng
RT: 20.480 min Scan# 2974
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:149 Resp: 570581
Ion Ratio Lower Upper
149 100
91 72.4 56.9 85.3
206 19.5 16.1 24.1





#81
Benzo(a)anthracene
Concen: 20.502 ng
RT: 21.363 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:228 Resp: 161020
Ion Ratio Lower Upper
228 100
226 27.0 21.7 32.5
229 19.2 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#82
3,3'-Dichlorobenzidine
Concen: 19.636 ng
RT: 21.286 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:252 Resp: 470294
Ion Ratio Lower Upper
252 100
254 65.4 51.4 77.0
126 12.4 8.6 12.8





#83
Chrysene
Concen: 20.343 ng
RT: 21.427 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:228 Resp: 151982
Ion Ratio Lower Upper
228 100
226 29.5 24.3 36.5
229 19.6 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 20.277 ng
RT: 21.315 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:149 Resp: 871329
Ion Ratio Lower Upper
149 100
167 26.8 22.3 33.5
279 3.9 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 18.223 ng
RT: 22.457 min Scan# 311
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:149 Resp: 116195
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 5.1 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. -0.000 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17


Tgt Ion:264 Resp: 1156806
Ion Ratio Lower Upper
264 100
260 23.2 18.6 28.0
265 21.4 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 19.472 ng
RT: 27.739 min Scan# 41
Delta R.T. -0.018 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020



Tgt Ion:276 Resp: 145038
Ion Ratio Lower Upper
276 100
138 16.7 20.6 30.8
277 25.0 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#88
Benzo(b)fluoranthene
Concen: 20.720 ng
RT: 23.427 min Scan# 3475
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17

Tgt Ion:252 Resp: 1393549
Ion Ratio Lower Upper
252 100
253 22.5 17.2 25.8
125 10.3 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 20.594 ng

RT: 23.492 min Scan# 3487

Delta R.T. -0.006 min

Lab File: BM050197.D


Acq: 05 Jun 2025 11:17

Instrument :

BNA_M

ClientSampleId :

SSTDICC020



Tgt Ion:252 Resp: 141397

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.2




125 10.3 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/06/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#90

Benzo(a)pyrene

Concen: 20.315 ng

RT: 24.233 min Scan# 3612

Delta R.T. -0.006 min

Lab File: BM050197.D

Acq: 05 Jun 2025 11:17




Tgt Ion:252 Resp: 1284647


Ion Ratio Lower Upper

252 100

253 20.6 17.6 26.4


125 11.0 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 19.538 ng
RT: 27.809 min Scan# 41
Delta R.T. -0.012 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17




Instrument :
BNA_M
ClientSampleId :
SSTDICC020




Tgt Ion:278 Resp: 1181274
Ion Ratio Lower Upper
278 100
139 15.7 12.8 19.2
279 23.6 18.6 28.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/06/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#92
Benzo(g,h,i)perylene
Concen: 19.358 ng
RT: 28.791 min Scan# 4387
Delta R.T. -0.006 min
Lab File: BM050197.D
Acq: 05 Jun 2025 11:17



Tgt Ion:276 Resp: 1171360
Ion Ratio Lower Upper
276 100
277 24.0 18.6 27.8
138 21.8 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050198.D
 Acq On : 05 Jun 2025 11:57
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICCC040

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/05/2025
 Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:45 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	308806	20.000	ng	0.00
21) Naphthalene-d8	10.575	136	1316349	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	819036	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	1590674	20.000	ng	0.00
76) Chrysene-d12	21.386	240	1484493	20.000	ng	0.00
86) Perylene-d12	24.374	264	1410376	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	1414880	76.431	ng	0.00
7) Phenol-d6	6.952	99	1932648	79.307	ng	0.00
23) Nitrobenzene-d5	8.940	82	1984901	78.422	ng	0.00
42) 2,4,6-Tribromophenol	15.898	330	756605	81.216	ng	0.00
45) 2-Fluorobiphenyl	13.045	172	4510866	74.564	ng	0.00
79) Terphenyl-d14	19.780	244	5997941	76.566	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	296051	36.484	ng	100
3) Pyridine	3.681	79	809743	38.534	ng	100
4) n-Nitrosodimethylamine	3.593	42	167736	38.601	ng	100
6) Aniline	7.116	93	1275182	39.583	ng	100
8) 2-Chlorophenol	7.352	128	799983	39.288	ng	100
9) Benzaldehyde	6.928	77	542319	35.003	ng	100
10) Phenol	6.975	94	1015393	39.380	ng	100
11) bis(2-Chloroethyl)ether	7.211	93	812045	39.155	ng	100
12) 1,3-Dichlorobenzene	7.681	146	871882	38.301	ng	100
13) 1,4-Dichlorobenzene	7.822	146	895582	37.813	ng	100
14) 1,2-Dichlorobenzene	8.140	146	865951	38.353	ng	100
15) Benzyl Alcohol	8.022	79	704555	40.536	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.322	45	561618	38.982	ng	100
17) 2-Methylphenol	8.222	107	684260	40.220	ng	100
18) Hexachloroethane	8.869	117	329145	38.374	ng	100
19) n-Nitroso-di-n-propyla...	8.599	70	618395	41.468	ng	100
20) 3+4-Methylphenols	8.552	107	932041	40.948	ng	100
22) Acetophenone	8.605	105	1252780	38.094	ng	100
24) Nitrobenzene	8.981	77	893607	38.820	ng	100
25) Isophorone	9.510	82	1735374	39.724	ng	100
26) 2-Nitrophenol	9.687	139	407943	41.055	ng	100
27) 2,4-Dimethylphenol	9.752	122	798142	39.057	ng	100
28) bis(2-Chloroethoxy)met...	9.993	93	1123606	39.371	ng	100
29) 2,4-Dichlorophenol	10.222	162	752942	39.853	ng	100
30) 1,2,4-Trichlorobenzene	10.440	180	803693	38.227	ng	100
31) Naphthalene	10.628	128	2571805	37.788	ng	100
32) Benzoic acid	9.875	122	538152	41.942	ng	100
33) 4-Chloroaniline	10.728	127	1139540	39.273	ng	100
34) Hexachlorobutadiene	10.922	225	480290	38.406	ng	100
35) Caprolactam	11.499	113	256391	42.805	ng	100
36) 4-Chloro-3-methylphenol	11.851	107	824664	40.906	ng	100
37) 2-Methylnaphthalene	12.240	142	1619990	39.456	ng	100
38) 1-Methylnaphthalene	12.457	142	1708673	39.210	ng	100
40) 1,2,4,5-Tetrachloroben...	12.610	216	886232	37.472	ng	100
41) Hexachlorocyclopentadiene	12.598	237	558871	38.910	ng	100
43) 2,4,6-Trichlorophenol	12.845	196	613571	39.450	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050198.D
 Acq On : 05 Jun 2025 11:57
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICCC040

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/05/2025
 Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:45 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.916	196	678534	39.736	ng	100
46) 1,1'-Biphenyl	13.257	154	2299241	37.491	ng	100
47) 2-Chloronaphthalene	13.293	162	1775515	37.239	ng	100
48) 2-Nitroaniline	13.487	65	439212	41.860	ng	100
49) Acenaphthylene	14.140	152	2979847	38.640	ng	100
50) Dimethylphthalate	13.881	163	2200039	39.718	ng	100
51) 2,6-Dinitrotoluene	13.987	165	473177	42.359	ng	100
52) Acenaphthene	14.481	154	1854951	38.456	ng	100
53) 3-Nitroaniline	14.310	138	529333	42.518	ng	100
54) 2,4-Dinitrophenol	14.510	184	249778	37.659	ng	100
55) Dibenzofuran	14.816	168	2721863	38.571	ng	100
56) 4-Nitrophenol	14.610	139	449268	42.399	ng	100
57) 2,4-Dinitrotoluene	14.769	165	648297	43.996	ng	100
58) Fluorene	15.463	166	2075880	38.407	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.039	232	597496m	40.441	ng	
60) Diethylphthalate	15.245	149	2205514	40.130	ng	100
61) 4-Chlorophenyl-phenyle...	15.463	204	1001298	38.331	ng	100
62) 4-Nitroaniline	15.475	138	503785	43.174	ng	100
63) Azobenzene	15.757	77	2151453	39.165	ng	100
65) 4,6-Dinitro-2-methylph...	15.534	198	353237	40.658	ng	100
66) n-Nitrosodiphenylamine	15.675	169	1888442	37.784	ng	100
67) 4-Bromophenyl-phenylether	16.351	248	648140	38.210	ng	100
68) Hexachlorobenzene	16.463	284	755736	38.133	ng	100
69) Atrazine	16.622	200	654294	41.086	ng	100
70) Pentachlorophenol	16.804	266	520982	40.433	ng	100
71) Phenanthrene	17.198	178	3406564	37.488	ng	100
72) Anthracene	17.286	178	3455118	38.008	ng	100
73) Carbazole	17.551	167	3209816	38.758	ng	100
74) Di-n-butylphthalate	18.133	149	3688687	40.935	ng	100
75) Fluoranthene	19.210	202	3748435	39.116	ng	100
77) Benzidine	19.392	184	1831158	43.865	ng	100
78) Pyrene	19.569	202	3930978	39.288	ng	100
80) Butylbenzylphthalate	20.480	149	1455291	43.604	ng	100
81) Benzo(a)anthracene	21.368	228	3616438	38.488	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	1232348	43.008	ng	100
83) Chrysene	21.427	228	3415842	38.218	ng	100
84) Bis(2-ethylhexyl)phtha...	21.321	149	2198448	42.764	ng	100
85) Di-n-octyl phthalate	22.462	149	3308886	43.375	ng	100
87) Indeno(1,2,3-cd)pyrene	27.756	276	3359194	36.991	ng	100
88) Benzo(b)fluoranthene	23.433	252	3299020	40.232	ng	100
89) Benzo(k)fluoranthene	23.498	252	3292030	39.327	ng	100
90) Benzo(a)pyrene	24.239	252	3048619	39.542	ng	100
91) Dibenzo(a,h)anthracene	27.821	278	2740891	37.184	ng	100
92) Benzo(g,h,i)perylene	28.797	276	2690355	36.467	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050198.D
Acq On : 05 Jun 2025 11:57
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Quant Time: Jun 05 14:58:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050198.D
Acq On : 05 Jun 2025 11:57
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 14:58:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.793 min Scan# 817
Delta R.T. -0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 308800
Ion Ratio Lower Upper
152 100
150 152.8 122.2 183.4
115 57.9 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#2
1,4-Dioxane
Concen: 36.484 ng
RT: 3.293 min Scan# 52
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion: 88 Resp: 296051
Ion Ratio Lower Upper
88 100
58 53.6 42.9 64.3
43 13.8 11.0 16.6





#3
Pyridine
Concen: 38.534 ng
RT: 3.681 min Scan# 118
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion: 79 Resp: 80974
Ion Ratio Lower Upper
79 100
52 40.1 32.1 48.1
51 19.2 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#4
n-Nitrosodimethylamine
Concen: 38.601 ng
RT: 3.593 min Scan# 103
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion: 42 Resp: 167736
Ion Ratio Lower Upper
42 100
74 277.5 222.0 333.0
44 8.8 7.0 10.6





#5

2-Fluorophenol

Concen: 76.431 ng

RT: 5.369 min Scan# 405

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion:112 Resp: 1414880

Ion Ratio Lower Upper

112 100

64 49.7 39.8 59.6





63 24.8 19.8 29.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#6

Aniline

Concen: 39.583 ng

RT: 7.116 min Scan# 702

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57




Tgt Ion: 93 Resp: 1275182


Ion Ratio Lower Upper

93 100

66 34.3 27.4 41.2

65 17.0 13.6 20.4









#7
Phenol-d6
Concen: 79.307 ng
RT: 6.952 min Scan# 674
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 193264
Ion Ratio Lower Upper
99 100
42 8.6 6.9 10.3
71 31.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#8
2-Chlorophenol
Concen: 39.288 ng
RT: 7.352 min Scan# 742
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion:128 Resp: 799983
Ion Ratio Lower Upper
128 100
130 32.6 12.6 52.6
64 39.8 19.8 59.8









#9
Benzaldehyde
Concen: 35.003 ng
RT: 6.928 min Scan# 670
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 542319
Ion Ratio Lower Upper
77 100
105 100.9 80.9 120.9
106 97.1 77.1 117.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#10
Phenol
Concen: 39.380 ng
RT: 6.975 min Scan# 678
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion: 94 Resp: 1015393
Ion Ratio Lower Upper
94 100
65 24.3 4.3 44.3
66 32.9 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 39.155 ng
RT: 7.211 min Scan# 718
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 812045
Ion Ratio Lower Upper
93 100
63 63.9 43.9 83.9
95 32.7 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#12
1,3-Dichlorobenzene
Concen: 38.301 ng
RT: 7.681 min Scan# 798
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion:146 Resp: 871882
Ion Ratio Lower Upper
146 100
148 64.9 51.9 77.9
75 29.2 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 37.813 ng

RT: 7.822 min Scan# 822

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion:146 Resp: 89558

Ion Ratio Lower Upper

146 100

148 64.2 51.4 77.0





111 40.4 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#14

1,2-Dichlorobenzene

Concen: 38.353 ng

RT: 8.140 min Scan# 876

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57




Tgt Ion:146 Resp: 865951


Ion Ratio Lower Upper

146 100

148 63.0 50.4 75.6

111 42.9 34.3 51.5





#15

Benzyl Alcohol

Concen: 40.536 ng

RT: 8.022 min Scan# 856

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion: 79 Resp: 70455

Ion Ratio Lower Upper

79 100

108 80.0 64.0 96.0





77 65.2 52.2 78.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 38.982 ng

RT: 8.322 min Scan# 907

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57




Tgt Ion: 45 Resp: 561618


Ion Ratio Lower Upper

45 100

77 28.7 8.7 48.7

79 22.7 2.7 42.7





#17

2-Methylphenol

Concen: 40.220 ng

RT: 8.222 min Scan# 890

Delta R.T. 0.000 min

Lab File: BM050198.D


Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040



Tgt Ion:107 Resp: 684260

Ion Ratio Lower Upper

107 100

108 113.5 90.8 136.2

77 44.6 35.7 53.5




79 43.7 35.0 52.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#18

Hexachloroethane


Concen: 38.374 ng

RT: 8.869 min Scan# 1000

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57





Tgt Ion:117 Resp: 329145


Ion Ratio Lower Upper

117 100

119 96.3 77.0 115.6

201 93.2 74.6 111.8





#19

n-Nitroso-di-n-propylamine

Concen: 41.468 ng

RT: 8.599 min Scan# 954

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion: 70 Resp: 61839

Ion Ratio Lower Upper

70 100

42 28.2 22.6 33.8

101 10.4 8.3 12.5





130 21.8 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#20

3+4-Methylphenols

Concen: 40.948 ng

RT: 8.552 min Scan# 946

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Tgt Ion:107 Resp: 932041




Ion Ratio Lower Upper


107 100

108 90.4 70.4 110.4

77 32.2 12.2 52.2


79 27.7 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.575 min Scan# 1290
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:136 Resp: 1316349
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 4.8 3.8 5.8
68 6.1 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#22
Acetophenone
Concen: 38.094 ng
RT: 8.605 min Scan# 955
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion:105 Resp: 1252780
Ion Ratio Lower Upper
105 100
71 6.2 5.0 7.4
51 14.3 11.4 17.2
120 22.8 18.2 27.4





#23

Nitrobenzene-d5

Concen: 78.422 ng

RT: 8.940 min Scan# 1012

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion: 82 Resp: 198490

Ion Ratio Lower Upper

82 100

128 44.0 35.2 52.8





54 30.6 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#24

Nitrobenzene

Concen: 38.820 ng

RT: 8.981 min Scan# 1019

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57




Tgt Ion: 77 Resp: 893607


Ion Ratio Lower Upper

77 100

123 47.7 38.2 57.2


65 11.9 9.5 14.3





#25
Isophorone
Concen: 39.724 ng
RT: 9.510 min Scan# 1109
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion: 82 Resp: 1735374
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 17.7 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#26
2-Nitrophenol
Concen: 41.055 ng
RT: 9.687 min Scan# 1139
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:139 Resp: 407943
Ion Ratio Lower Upper
139 100
109 29.1 23.3 34.9
65 44.0 35.2 52.8





#27
2,4-Dimethylphenol
Concen: 39.057 ng
RT: 9.752 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040




Tgt Ion:122 Resp: 79814
Ion Ratio Lower Upper
122 100
107 113.8 91.0 136.6
121 57.2 45.8 68.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#28
bis(2-Chloroethoxy)methane
Concen: 39.371 ng
RT: 9.993 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Tgt Ion: 93 Resp: 1123606
Ion Ratio Lower Upper
93 100
95 32.8 26.2 39.4
123 11.1 8.9 13.3





#29
2,4-Dichlorophenol
Concen: 39.853 ng
RT: 10.222 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:162 Resp: 75294
Ion Ratio Lower Upper
162 100
164 65.0 45.0 85.0
98 36.3 16.3 56.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#30
1,2,4-Trichlorobenzene
Concen: 38.227 ng
RT: 10.440 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:180 Resp: 803693
Ion Ratio Lower Upper
180 100
182 95.7 76.6 114.8
145 30.3 24.2 36.4





#31
Naphthalene
Concen: 37.788 ng
RT: 10.628 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:128 Resp: 257180
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.8 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#32
Benzoic acid
Concen: 41.942 ng
RT: 9.875 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:122 Resp: 538152
Ion Ratio Lower Upper
122 100
105 123.4 103.4 143.4
77 90.3 70.3 110.3





#33
4-Chloroaniline
Concen: 39.273 ng
RT: 10.728 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:127 Resp: 1139540
Ion Ratio Lower Upper
127 100
129 31.7 25.4 38.0
65 26.2 21.0 31.4
92 18.5 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#34
Hexachlorobutadiene
Concen: 38.406 ng
RT: 10.922 min Scan# 1349
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:225 Resp: 480290
Ion Ratio Lower Upper
225 100
223 62.7 50.2 75.2
227 63.7 51.0 76.4





#35
Caprolactam
Concen: 42.805 ng
RT: 11.499 min Scan# 1447
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:113 Resp: 25639
Ion Ratio Lower Upper
113 100
55 103.5 83.5 123.5
56 80.6 60.6 100.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#36
4-Chloro-3-methylphenol
Concen: 40.906 ng
RT: 11.851 min Scan# 1507
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:107 Resp: 824664
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.6
142 81.0 64.8 97.2





#37
2-Methylnaphthalene
Concen: 39.456 ng
RT: 12.240 min Scan# 1573
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:142 Resp: 1619990
Ion Ratio Lower Upper
142 100
141 89.4 71.5 107.3
115 34.8 27.8 41.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#38
1-Methylnaphthalene
Concen: 39.210 ng
RT: 12.457 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:142 Resp: 1708673
Ion Ratio Lower Upper
142 100
141 91.1 72.9 109.3
116 3.8 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040




Tgt Ion:164 Resp: 819030
Ion Ratio Lower Upper
164 100
162 100.3 80.2 120.4
160 44.6 35.7 53.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.472 ng
RT: 12.610 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Tgt Ion:216 Resp: 886232
Ion Ratio Lower Upper
216 100
214 78.7 63.0 94.4
179 20.8 16.6 25.0
108 19.9 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 38.910 ng
RT: 12.598 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:237 Resp: 55887
Ion Ratio Lower Upper
237 100
235 64.6 44.6 84.6
272 13.1 0.0 33.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#42
2,4,6-Tribromophenol
Concen: 81.216 ng
RT: 15.898 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion:330 Resp: 756605
Ion Ratio Lower Upper
330 100
332 96.9 77.5 116.3
141 36.0 28.8 43.2





#43

2,4,6-Trichlorophenol

Concen: 39.450 ng

RT: 12.845 min Scan# 1676

Delta R.T. 0.000 min

Lab File: BM050198.D


Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040



Tgt Ion:196 Resp: 61357

Ion Ratio Lower Upper

196 100

198 95.8 76.6 115.0




200 30.6 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#44

2,4,5-Trichlorophenol

Concen: 39.736 ng

RT: 12.916 min Scan# 1688

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Tgt Ion:196 Resp: 678534




Ion Ratio Lower Upper


196 100

198 95.1 76.1 114.1

97 49.1 39.3 58.9

132 26.7 21.4 32.0





#45

2-Fluorobiphenyl

Concen: 74.564 ng

RT: 13.045 min Scan# 1710

Delta R.T. 0.000 min

Lab File: BM050198.D


Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040



Tgt Ion:172 Resp: 451086

Ion Ratio Lower Upper

172 100

171 34.7 27.8 41.6




170 23.2 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#46

1,1'-Biphenyl


Concen: 37.491 ng

RT: 13.257 min Scan# 1746

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57





Tgt Ion:154 Resp: 2299241


Ion Ratio Lower Upper

154 100

153 39.2 19.2 59.2


76 12.0 0.0 32.0





#47
2-Chloronaphthalene
Concen: 37.239 ng
RT: 13.293 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:162 Resp: 177551
Ion Ratio Lower Upper
162 100
127 38.9 31.1 46.7
164 33.1 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#48
2-Nitroaniline
Concen: 41.860 ng
RT: 13.487 min Scan# 1785
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion: 65 Resp: 439212
Ion Ratio Lower Upper
65 100
92 80.3 64.2 96.4
138 126.2 101.0 151.4





#49
Acenaphthylene
Concen: 38.640 ng
RT: 14.140 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:152 Resp: 297984
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.0 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#50
Dimethylphthalate
Concen: 39.718 ng
RT: 13.881 min Scan# 1852
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:163 Resp: 2200039
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 9.9 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 42.359 ng
RT: 13.987 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:165 Resp: 47317
Ion Ratio Lower Upper
165 100
63 44.7 35.8 53.6
89 56.0 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#52
Acenaphthene
Concen: 38.456 ng
RT: 14.481 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:154 Resp: 1854951
Ion Ratio Lower Upper
154 100
153 110.9 88.7 133.1
152 53.2 42.6 63.8





#53
3-Nitroaniline
Concen: 42.518 ng
RT: 14.310 min Scan# 1925
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040




Tgt Ion:138 Resp: 52933
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 116.1 92.9 139.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#54
2,4-Dinitrophenol
Concen: 37.659 ng
RT: 14.510 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Tgt Ion:184 Resp: 249778
Ion Ratio Lower Upper
184 100
63 56.6 45.3 67.9
154 67.9 54.3 81.5





#55
Dibenzenofuran
Concen: 38.571 ng
RT: 14.816 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:168 Resp: 272186
Ion Ratio Lower Upper
168 100
139 37.4 29.9 44.9
169 13.3 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#56
4-Nitrophenol
Concen: 42.399 ng
RT: 14.610 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:139 Resp: 449268
Ion Ratio Lower Upper
139 100
109 72.6 52.6 92.6
65 84.3 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 43.996 ng
RT: 14.769 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:165 Resp: 64829
Ion Ratio Lower Upper
165 100
63 35.5 28.4 42.6
89 72.5 58.0 87.0
182 2.7 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#58
Fluorene
Concen: 38.407 ng
RT: 15.463 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:166 Resp: 2075880
Ion Ratio Lower Upper
166 100
165 96.2 77.0 115.4
167 13.4 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 40.441 ng m
RT: 15.039 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:232 Resp: 597490
Ion Ratio Lower Upper
232 100
131 40.9 32.9 49.3
130 2.3 1.8 2.8
166 25.9 20.8 31.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#60
Diethylphthalate
Concen: 40.130 ng
RT: 15.245 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:149 Resp: 2205514
Ion Ratio Lower Upper
149 100
177 21.5 17.2 25.8
150 12.5 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 38.331 ng
RT: 15.463 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:204 Resp: 1001298
Ion Ratio Lower Upper
204 100
206 31.7 25.4 38.0
141 59.9 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#62
4-Nitroaniline
Concen: 43.174 ng
RT: 15.475 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:138 Resp: 503785
Ion Ratio Lower Upper
138 100
92 48.8 28.8 68.8
108 78.1 58.1 98.1





#63
Azobenzene
Concen: 39.165 ng
RT: 15.757 min Scan# 2171
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion: 77 Resp: 215145
Ion Ratio Lower Upper
77 100
182 26.8 6.8 46.8
105 21.7 1.7 41.7
51 16.2 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:188 Resp: 1590674
Ion Ratio Lower Upper
188 100
94 10.1 8.1 12.1
80 10.8 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 40.658 ng
RT: 15.534 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:198 Resp: 353231
Ion Ratio Lower Upper
198 100
51 24.5 4.5 44.5
105 44.9 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#66
n-Nitrosodiphenylamine
Concen: 37.784 ng
RT: 15.675 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:169 Resp: 1888442
Ion Ratio Lower Upper
169 100
168 66.4 53.1 79.7
167 36.3 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 38.210 ng
RT: 16.351 min Scan# 2272
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:248 Resp: 648140
Ion Ratio Lower Upper
248 100
250 97.5 78.0 117.0
141 75.1 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#68
Hexachlorobenzene
Concen: 38.133 ng
RT: 16.463 min Scan# 2291
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:284 Resp: 755736
Ion Ratio Lower Upper
284 100
142 36.0 28.8 43.2
249 32.2 25.8 38.6





#69
Atrazine
Concen: 41.086 ng
RT: 16.622 min Scan# 2318
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040




Tgt Ion:200 Resp: 654294
Ion Ratio Lower Upper
200 100
173 26.0 6.0 46.0
215 48.1 28.1 68.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#70
Pentachlorophenol
Concen: 40.433 ng
RT: 16.804 min Scan# 2349
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Tgt Ion:266 Resp: 520982
Ion Ratio Lower Upper
266 100
268 63.8 51.0 76.6
264 63.5 50.8 76.2





#71
Phenanthrene
Concen: 37.488 ng
RT: 17.198 min Scan# 2416
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:178 Resp: 3406564
Ion Ratio Lower Upper
178 100
176 19.5 15.6 23.4
179 15.4 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#72
Anthracene
Concen: 38.008 ng
RT: 17.286 min Scan# 2431
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:178 Resp: 3455118
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.4 12.3 18.5





#73
 Carbazole
 Concen: 38.758 ng
 RT: 17.551 min Scan# 2476
 Delta R.T. 0.000 min
 Lab File: BM050198.D
 Acq: 05 Jun 2025 11:57




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC040




Tgt Ion:167 Resp: 3209810
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.0 25.6
 139 12.8 10.2 15.4

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 06/05/2025
 Supervised By :Jagrut Upadhyay 06/06/2025




#74
 Di-n-butylphthalate
 Concen: 40.935 ng
 RT: 18.133 min Scan# 2575
 Delta R.T. 0.000 min
 Lab File: BM050198.D
 Acq: 05 Jun 2025 11:57




Tgt Ion:149 Resp: 3688687
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.0
 104 5.0 4.0 6.0





#75
Fluoranthene
Concen: 39.116 ng
RT: 19.210 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:202 Resp: 374843
Ion Ratio Lower Upper
202 100
101 11.1 0.0 31.1
203 17.0 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:240 Resp: 1484493
Ion Ratio Lower Upper
240 100
120 11.2 9.0 13.4
236 25.9 20.7 31.1





#77
Benzidine
Concen: 43.865 ng
RT: 19.392 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:184 Resp: 1831158
Ion Ratio Lower Upper
184 100
185 14.4 11.5 17.3
183 11.2 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#78
Pyrene
Concen: 39.288 ng
RT: 19.569 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:202 Resp: 3930978
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.6 14.1 21.1





#79
Terphenyl-d14
Concen: 76.566 ng
RT: 19.780 min Scan# 2855
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:244 Resp: 599794
Ion Ratio Lower Upper
244 100
212 7.5 6.0 9.0
122 10.7 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#80
Butylbenzylphthalate
Concen: 43.604 ng
RT: 20.480 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:149 Resp: 1455291
Ion Ratio Lower Upper
149 100
91 71.1 56.9 85.3
206 20.1 16.1 24.1





#81
Benzo(a)anthracene
Concen: 38.488 ng
RT: 21.368 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:228 Resp: 3616438
Ion Ratio Lower Upper
228 100
226 27.1 21.7 32.5
229 19.6 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#82
3,3'-Dichlorobenzidine
Concen: 43.008 ng
RT: 21.292 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion:252 Resp: 1232348
Ion Ratio Lower Upper
252 100
254 64.2 51.4 77.0
126 10.7 8.6 12.8





#83

Chrysene

Concen: 38.218 ng

RT: 21.427 min Scan# 3135

Delta R.T. 0.000 min

Lab File: BM050198.D


Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040



Tgt Ion:228 Resp: 341584

Ion Ratio Lower Upper

228 100

226 30.4 24.3 36.5




229 19.8 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#84

Bis(2-ethylhexyl)phthalate


Concen: 42.764 ng

RT: 21.321 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57





Tgt Ion:149 Resp: 2198448


Ion Ratio Lower Upper

149 100

167 27.9 22.3 33.5


279 4.4 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 43.375 ng
RT: 22.462 min Scan# 3311
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:149 Resp: 3308880
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.8 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57


Tgt Ion:264 Resp: 1410376
Ion Ratio Lower Upper
264 100
260 23.3 18.6 28.0
265 21.0 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 36.991 ng
RT: 27.756 min Scan# 4211
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040




Tgt Ion:276 Resp: 3359194
Ion Ratio Lower Upper
276 100
138 25.8 20.6 30.8
277 24.8 19.8 29.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#88
Benzo(b)fluoranthene
Concen: 40.232 ng
RT: 23.433 min Scan# 3476
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57



Tgt Ion:252 Resp: 3299020
Ion Ratio Lower Upper
252 100
253 21.5 17.2 25.8
125 10.3 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 39.327 ng

RT: 23.498 min Scan# 3487

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion:252 Resp: 3292030

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2





125 10.3 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#90

Benzo(a)pyrene

Concen: 39.542 ng

RT: 24.239 min Scan# 3613

Delta R.T. 0.000 min

Lab File: BM050198.D

Acq: 05 Jun 2025 11:57




Tgt Ion:252 Resp: 3048619


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4


125 11.2 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 37.184 ng
RT: 27.821 min Scan# 4222
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040



Tgt Ion:278 Resp: 274089
Ion Ratio Lower Upper
278 100
139 16.0 12.8 19.2
279 23.3 18.6 28.0




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#92
Benzo(g,h,i)perylene
Concen: 36.467 ng
RT: 28.797 min Scan# 4388
Delta R.T. 0.000 min
Lab File: BM050198.D
Acq: 05 Jun 2025 11:57

Tgt Ion:276 Resp: 2690355
Ion Ratio Lower Upper
276 100
277 23.2 18.6 27.8
138 22.2 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050199.D
 Acq On : 05 Jun 2025 12:36
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC050

Quant Time: Jun 05 14:59:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	291057	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	1245308	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	784298	20.000	ng	0.00
64) Phenanthrene-d10	17.157	188	1523841	20.000	ng	0.00
76) Chrysene-d12	21.392	240	1446151	20.000	ng	0.00
86) Perylene-d12	24.374	264	1421732	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	1802037	103.281	ng	0.00
7) Phenol-d6	6.951	99	2472934	107.666	ng	0.00
23) Nitrobenzene-d5	8.939	82	2579483	107.728	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	994114	111.437	ng	0.00
45) 2-Fluorobiphenyl	13.045	172	5738887	99.064	ng	0.00
79) Terphenyl-d14	19.786	244	7692376	100.800	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	371991	48.638	ng	100
3) Pyridine	3.681	79	1036652	52.340	ng	98
4) n-Nitrosodimethylamine	3.593	42	215821	52.696	ng	95
6) Aniline	7.116	93	1637124	53.917	ng	100
8) 2-Chlorophenol	7.351	128	1032007	53.773	ng	99
9) Benzaldehyde	6.928	77	709626	48.594	ng	99
10) Phenol	6.981	94	1296462	53.347	ng	99
11) bis(2-Chloroethyl)ether	7.216	93	1039221	53.164	ng	98
12) 1,3-Dichlorobenzene	7.681	146	1124523	52.412	ng	100
13) 1,4-Dichlorobenzene	7.822	146	1157927	51.870	ng	100
14) 1,2-Dichlorobenzene	8.140	146	1120717	52.663	ng	99
15) Benzyl Alcohol	8.022	79	921887	56.274	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.322	45	711769	52.417	ng	99
17) 2-Methylphenol	8.222	107	878183	54.767	ng	100
18) Hexachloroethane	8.869	117	431469	53.371	ng	99
19) n-Nitroso-di-n-propyla...	8.598	70	797673	56.751	ng	100
20) 3+4-Methylphenols	8.551	107	1208370	56.326	ng	98
22) Acetophenone	8.604	105	1616196	51.949	ng	99
24) Nitrobenzene	8.981	77	1151898	52.895	ng	99
25) Isophorone	9.510	82	2280770	55.187	ng	100
26) 2-Nitrophenol	9.692	139	552994	58.828	ng	98
27) 2,4-Dimethylphenol	9.751	122	1034501	53.511	ng	100
28) bis(2-Chloroethoxy)met...	9.992	93	1463553	54.208	ng	99
29) 2,4-Dichlorophenol	10.222	162	991973	55.499	ng	99
30) 1,2,4-Trichlorobenzene	10.439	180	1048864	52.734	ng	98
31) Naphthalene	10.628	128	3303529	51.309	ng	100
32) Benzoic acid	9.892	122	725550	59.774	ng	99
33) 4-Chloroaniline	10.728	127	1484906	54.095	ng	99
34) Hexachlorobutadiene	10.928	225	629209	53.185	ng	99
35) Caprolactam	11.510	113	348835	61.561	ng	98
36) 4-Chloro-3-methylphenol	11.857	107	1081009	56.680	ng	100
37) 2-Methylnaphthalene	12.239	142	2112994	54.400	ng	98
38) 1-Methylnaphthalene	12.463	142	2223932	53.945	ng	100
40) 1,2,4,5-Tetrachloroben...	12.610	216	1155693	51.030	ng	100
41) Hexachlorocyclopentadiene	12.598	237	749110	54.466	ng	98
43) 2,4,6-Trichlorophenol	12.845	196	810794	54.440	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050199.D
 Acq On : 05 Jun 2025 12:36
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC050

Quant Time: Jun 05 14:59:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.916	196	896452	54.822	ng	100
46) 1,1'-Biphenyl	13.257	154	2992234	50.951	ng	99
47) 2-Chloronaphthalene	13.292	162	2296762	50.305	ng	99
48) 2-Nitroaniline	13.486	65	591382	58.859	ng	99
49) Acenaphthylene	14.139	152	3878499	52.521	ng	100
50) Dimethylphthalate	13.880	163	2890771	54.500	ng	100
51) 2,6-Dinitrotoluene	13.986	165	627879	58.698	ng	97
52) Acenaphthene	14.480	154	2406891	52.109	ng	99
53) 3-Nitroaniline	14.316	138	706323	59.247	ng	98
54) 2,4-Dinitrophenol	14.516	184	350975	54.263	ng	91
55) Dibenzofuran	14.816	168	3522761	52.131	ng	100
56) 4-Nitrophenol	14.616	139	601906	59.319	ng	98
57) 2,4-Dinitrotoluene	14.774	165	876137	62.091	ng	97
58) Fluorene	15.463	166	2643113	51.068	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.039	232	786725	55.608	ng	92
60) Diethylphthalate	15.251	149	2921964	55.521	ng	99
61) 4-Chlorophenyl-phenyle...	15.463	204	1293938	51.728	ng	99
62) 4-Nitroaniline	15.480	138	669832	59.947	ng	99
63) Azobenzene	15.757	77	2829707	53.794	ng	99
65) 4,6-Dinitro-2-methylph...	15.533	198	487410	58.562	ng	98
66) n-Nitrosodiphenylamine	15.674	169	2459825	51.375	ng	100
67) 4-Bromophenyl-phenylether	16.357	248	868132	53.423	ng	97
68) Hexachlorobenzene	16.468	284	993147	52.310	ng	96
69) Atrazine	16.627	200	878900	57.610	ng	100
70) Pentachlorophenol	16.804	266	711985	57.680	ng	100
71) Phenanthrene	17.198	178	4421661	50.793	ng	100
72) Anthracene	17.286	178	4499098	51.663	ng	99
73) Carbazole	17.551	167	4203693	52.985	ng	100
74) Di-n-butylphthalate	18.139	149	4951541	57.359	ng	100
75) Fluoranthene	19.209	202	4957680	54.004	ng	99
77) Benzidine	19.392	184	2489372	61.214	ng	100
78) Pyrene	19.568	202	5143501	52.769	ng	99
80) Butylbenzylphthalate	20.486	149	2036123	62.625	ng	98
81) Benzo(a)anthracene	21.368	228	4836514	52.838	ng	99
82) 3,3'-Dichlorobenzidine	21.292	252	1721840	61.684	ng	99
83) Chrysene	21.433	228	4540308	52.145	ng	100
84) Bis(2-ethylhexyl)phtha...	21.321	149	3074950	61.400	ng	100
85) Di-n-octyl phthalate	22.462	149	4912405	66.103	ng	100
87) Indeno(1,2,3-cd)pyrene	27.762	276	4685372	51.183	ng	100
88) Benzo(b)fluoranthene	23.439	252	4510196	54.563	ng	99
89) Benzo(k)fluoranthene	23.503	252	4604565	54.567	ng	99
90) Benzo(a)pyrene	24.238	252	4255236	54.752	ng	100
91) Dibenzo(a,h)anthracene	27.821	278	3820009	51.410	ng	100
92) Benzo(g,h,i)perylene	28.803	276	3684864	49.548	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050199.D
 Acq On : 05 Jun 2025 12:36
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC050


Quant Time: Jun 05 14:59:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050199.D
Acq On : 05 Jun 2025 12:36
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Quant Time: Jun 05 14:59:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.787 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:152 Resp: 291057
Ion Ratio Lower Upper
152 100
150 155.1 122.2 183.4
115 59.5 46.3 69.5



#2
1,4-Dioxane
Concen: 48.638 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion: 88 Resp: 371991
Ion Ratio Lower Upper
88 100
58 53.5 42.9 64.3
43 14.0 11.0 16.6









#3
Pyridine
Concen: 52.340 ng
RT: 3.681 min Scan# 118
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 79 Resp: 1036652
Ion Ratio Lower Upper
79 100
52 38.8 32.1 48.1
51 18.6 15.4 23.0



#4
n-Nitrosodimethylamine
Concen: 52.696 ng
RT: 3.593 min Scan# 103
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion: 42 Resp: 215821
Ion Ratio Lower Upper
42 100
74 267.3 222.0 333.0
44 7.3 7.0 10.6





#5

2-Fluorophenol

Concen: 103.281 ng

RT: 5.369 min Scan# 405

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:112 Resp: 1802037

Ion Ratio Lower Upper

112 100

64 49.0 39.8 59.6

63 24.9 19.8 29.8



#6

Aniline

Concen: 53.917 ng

RT: 7.116 min Scan# 702

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion: 93 Resp: 1637124


Ion Ratio Lower Upper

93 100

66 34.1 27.4 41.2

65 17.0 13.6 20.4









#7
Phenol-d6
Concen: 107.666 ng
RT: 6.951 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 99 Resp: 2472934
Ion Ratio Lower Upper
99 100
42 8.8 6.9 10.3
71 32.0 25.3 37.9



#8
2-Chlorophenol
Concen: 53.773 ng
RT: 7.351 min Scan# 742
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion: 128 Resp: 1032007
Ion Ratio Lower Upper
128 100
130 32.3 12.6 52.6
64 38.8 19.8 59.8









#9
Benzaldehyde
Concen: 48.594 ng
RT: 6.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 77 Resp: 709626
Ion Ratio Lower Upper
77 100
105 100.3 80.9 120.9
106 96.5 77.1 117.1



#10
Phenol
Concen: 53.347 ng
RT: 6.981 min Scan# 679
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion: 94 Resp: 1296462
Ion Ratio Lower Upper
94 100
65 23.6 4.3 44.3
66 32.1 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 53.164 ng
RT: 7.216 min Scan# 718
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 93 Resp: 1039221
Ion Ratio Lower Upper
93 100
63 62.4 43.9 83.9
95 32.0 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 52.412 ng
RT: 7.681 min Scan# 798
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:146 Resp: 1124523
Ion Ratio Lower Upper
146 100
148 64.5 51.9 77.9
75 29.0 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 51.870 ng

RT: 7.822 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:146 Resp: 1157927

Ion Ratio Lower Upper

146 100

148 63.9 51.4 77.0

111 40.5 32.3 48.5



#14

1,2-Dichlorobenzene

Concen: 52.663 ng

RT: 8.140 min Scan# 876

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:146 Resp: 1120717


Ion Ratio Lower Upper

146 100

148 64.1 50.4 75.6

111 42.7 34.3 51.5





#15

Benzyl Alcohol

Concen: 56.274 ng

RT: 8.022 min Scan# 856

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :





BNA_M

ClientSampleId :

SSTDICC050

Tgt Ion: 79 Resp: 921887

Ion	Ratio	Lower	Upper
79	100		
108	78.0	64.0	96.0
77	63.0	52.2	78.2



#16

2,2'-oxybis(1-chloropropane)

Concen: 52.417 ng

RT: 8.322 min Scan# 907




Delta R.T. -0.000 min


Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion: 45 Resp: 711769

Ion	Ratio	Lower	Upper
45	100		
77	29.1	8.7	48.7
79	22.4	2.7	42.7





#17

2-Methylphenol

Concen: 54.767 ng

RT: 8.222 min Scan# 890

Delta R.T. -0.000 min

Lab File: BM050199.D


Acq: 05 Jun 2025 12:36

Instrument :

BNA_M




ClientSampleId :

SSTDICC050



Tgt Ion:107 Resp: 878183

Ion	Ratio	Lower	Upper
107	100		
108	114.0	90.8	136.2
77	44.6	35.7	53.5
79	44.3	35.0	52.4



#18

Hexachloroethane

Concen: 53.371 ng

RT: 8.869 min Scan# 1000




Delta R.T. -0.000 min


Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion:117 Resp: 431469

Ion	Ratio	Lower	Upper
117	100		
119	94.7	77.0	115.6
201	94.2	74.6	111.8





#19

n-Nitroso-di-n-propylamine

Concen: 56.751 ng

RT: 8.598 min Scan# 954

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050

Tgt Ion: 70 Resp: 797673





Ion Ratio Lower Upper

70 100

42 28.2 22.6 33.8

101 10.2 8.3 12.5

130 21.7 17.4 26.2



#20

3+4-Methylphenols

Concen: 56.326 ng

RT: 8.551 min Scan# 946

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion: 107 Resp: 1208370




Ion Ratio Lower Upper


107 100

108 88.2 70.4 110.4

77 31.7 12.2 52.2

79 26.6 7.7 47.7





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.581 min Scan# 1291

Delta R.T. 0.005 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDIC050

Tgt Ion:136 Resp: 1245308





Ion Ratio Lower Upper

136 100

137 10.6 8.6 13.0

54 4.7 3.8 5.8

68 5.8 4.9 7.3



#22

Acetophenone

Concen: 51.949 ng

RT: 8.604 min Scan# 955

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion:105 Resp: 1616196




Ion Ratio Lower Upper


105 100

71 7.1 5.0 7.4

51 14.7 11.4 17.2

120 22.7 18.2 27.4









#23
Nitrobenzene-d5
Concen: 107.728 ng
RT: 8.939 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 82 Resp: 2579483
Ion Ratio Lower Upper
82 100
128 44.2 35.2 52.8
54 30.5 24.5 36.7



#24
Nitrobenzene
Concen: 52.895 ng
RT: 8.981 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion: 77 Resp: 1151898
Ion Ratio Lower Upper
77 100
123 46.7 38.2 57.2
65 12.2 9.5 14.3





#25

Isophorone

Concen: 55.187 ng

RT: 9.510 min Scan# 1109

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :





BNA_M

ClientSampleId :

SSTDICC050

Tgt Ion: 82 Resp: 2280770

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.7	8.5
138	17.8	14.2	21.2



#26

2-Nitrophenol

Concen: 58.828 ng

RT: 9.692 min Scan# 1140




Delta R.T. 0.005 min


Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion:139 Resp: 552994

Ion	Ratio	Lower	Upper
139	100		
109	29.5	23.3	34.9
65	42.3	35.2	52.8





#27

2,4-Dimethylphenol

Concen: 53.511 ng

RT: 9.751 min Scan# 1150

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:122 Resp: 1034501

Ion Ratio Lower Upper

122 100

107 113.6 91.0 136.6

121 57.7 45.8 68.6



#28

bis(2-Chloroethoxy)methane

Concen: 54.208 ng

RT: 9.992 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion: 93 Resp: 1463553


Ion Ratio Lower Upper

93 100

95 32.4 26.2 39.4

123 11.5 8.9 13.3





#29

2,4-Dichlorophenol

Concen: 55.499 ng

RT: 10.222 min Scan# 1230

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:162 Resp: 991973

Ion Ratio Lower Upper

162 100

164 63.7 45.0 85.0

98 35.8 16.3 56.3



#30

1,2,4-Trichlorobenzene

Concen: 52.734 ng

RT: 10.439 min Scan# 1267

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:180 Resp: 1048864


Ion Ratio Lower Upper

180 100

182 93.8 76.6 114.8

145 30.5 24.2 36.4









#31
Naphthalene
Concen: 51.309 ng
RT: 10.628 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:128 Resp: 3303529
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.9 10.2 15.4



#32
Benzoic acid
Concen: 59.774 ng
RT: 9.892 min Scan# 1174
Delta R.T. 0.017 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36


Tgt Ion:122 Resp: 725550
Ion Ratio Lower Upper
122 100
105 122.3 103.4 143.4
77 92.1 70.3 110.3








#33
4-Chloroaniline
Concen: 54.095 ng
RT: 10.728 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Instrument :
BNA_M
ClientSampleId :
SSTDICC050




Tgt Ion:127 Resp: 1484906

Ion	Ratio	Lower	Upper
127	100		
129	32.2	25.4	38.0
65	26.6	21.0	31.4
92	18.7	14.8	22.2






#34
Hexachlorobutadiene
Concen: 53.185 ng
RT: 10.928 min Scan# 1350
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36



Tgt Ion:225 Resp: 629209

Ion	Ratio	Lower	Upper
225	100		
223	63.4	50.2	75.2
227	64.8	51.0	76.4









#35
Caprolactam
Concen: 61.561 ng
RT: 11.510 min Scan# 1449
Delta R.T. 0.011 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:113 Resp: 348835
Ion Ratio Lower Upper
113 100
55 100.7 83.5 123.5
56 81.8 60.6 100.6



#36
4-Chloro-3-methylphenol
Concen: 56.680 ng
RT: 11.857 min Scan# 1508
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:107 Resp: 1081009
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.6
142 81.6 64.8 97.2









#37
2-Methylnaphthalene
Concen: 54.400 ng
RT: 12.239 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:142 Resp: 2112994
Ion Ratio Lower Upper
142 100
141 86.6 71.5 107.3
115 34.7 27.8 41.8



#38
1-Methylnaphthalene
Concen: 53.945 ng
RT: 12.463 min Scan# 1611
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:142 Resp: 2223932
Ion Ratio Lower Upper
142 100
141 90.9 72.9 109.3
116 3.6 3.0 4.6





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.416 min Scan# 1943

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:164 Resp: 784298

Ion	Ratio	Lower	Upper

164	100		

162	101.1	80.2	120.4

160	44.0	35.7	53.5



#40

1,2,4,5-Tetrachlorobenzene

Concen: 51.030 ng

RT: 12.610 min Scan# 1636

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion:216 Resp: 1155693




Ion	Ratio	Lower	Upper


216	100		

214	78.9	63.0	94.4

179	20.8	16.6	25.0


108	19.7	15.9	23.9








#41
Hexachlorocyclopentadiene
Concen: 54.466 ng
RT: 12.598 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36


Instrument :
BNA_M
ClientSampleId :
SSTDICC050





Tgt Ion:237 Resp: 749110
Ion Ratio Lower Upper
237 100
235 62.8 44.6 84.6
272 12.7 0.0 33.1




#42
2,4,6-Tribromophenol
Concen: 111.437 ng
RT: 15.904 min Scan# 2196
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36



Tgt Ion:330 Resp: 994114
Ion Ratio Lower Upper
330 100
332 96.2 77.5 116.3
141 36.0 28.8 43.2





#43

2,4,6-Trichlorophenol

Concen: 54.440 ng

RT: 12.845 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM050199.D


Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050






Tgt Ion:196 Resp: 810794

Ion Ratio Lower Upper

196 100

198 95.9 76.6 115.0

200 30.9 24.5 36.7



#44

2,4,5-Trichlorophenol

Concen: 54.822 ng

RT: 12.916 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Tgt Ion:196 Resp: 896452




Ion Ratio Lower Upper


196 100

198 95.5 76.1 114.1

97 49.2 39.3 58.9

132 26.7 21.4 32.0









#45
2-Fluorobiphenyl
Concen: 99.064 ng
RT: 13.045 min Scan# 1710
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:172 Resp: 5738887
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.6
170 23.7 18.6 27.8



#46
1,1'-Biphenyl
Concen: 50.951 ng
RT: 13.257 min Scan# 1746
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:154 Resp: 2992234
Ion Ratio Lower Upper
154 100
153 39.6 19.2 59.2
76 12.1 0.0 32.0









#47
2-Chloronaphthalene
Concen: 50.305 ng
RT: 13.292 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:162 Resp: 2296762
Ion Ratio Lower Upper
162 100
127 39.4 31.1 46.7
164 32.7 26.5 39.7



#48
2-Nitroaniline
Concen: 58.859 ng
RT: 13.486 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion: 65 Resp: 591382
Ion Ratio Lower Upper
65 100
92 80.4 64.2 96.4
138 123.5 101.0 151.4





#49

Acenaphthylene

Concen: 52.521 ng

RT: 14.139 min Scan# 1896

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:152 Resp: 3878499

Ion Ratio Lower Upper

152 100

151 19.7 15.8 23.6

153 13.0 10.4 15.6



#50

Dimethylphthalate

Concen: 54.500 ng

RT: 13.880 min Scan# 1852

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:163 Resp: 2890771


Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9

164 10.1 7.9 11.9





#51

2,6-Dinitrotoluene

Concen: 58.698 ng

RT: 13.986 min Scan# 1870

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:165 Resp: 627879

Ion Ratio Lower Upper

165 100

63 47.2 35.8 53.6

89 58.2 44.8 67.2



#52

Acenaphthene

Concen: 52.109 ng

RT: 14.480 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:154 Resp: 2406891


Ion Ratio Lower Upper

154 100

153 112.5 88.7 133.1

152 53.1 42.6 63.8









#53
3-Nitroaniline
Concen: 59.247 ng
RT: 14.316 min Scan# 1925
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:138 Resp: 706323
Ion Ratio Lower Upper
138 100
108 10.4 8.8 13.2
92 113.9 92.9 139.3



#54
2,4-Dinitrophenol
Concen: 54.263 ng
RT: 14.516 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:184 Resp: 350975
Ion Ratio Lower Upper
184 100
63 50.4 45.3 67.9
154 60.9 54.3 81.5





#55

Dibenzenofuran

Concen: 52.131 ng

RT: 14.816 min Scan# 20

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:168 Resp: 3522761

Ion	Ratio	Lower	Upper

168	100		

139	37.2	29.9	44.9

169	13.3	10.6	16.0



#56

4-Nitrophenol

Concen: 59.319 ng

RT: 14.616 min Scan# 1977

Delta R.T. 0.006 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:139 Resp: 601906


Ion	Ratio	Lower	Upper

139	100		

109	74.2	52.6	92.6

65	81.7	64.3	104.3









#57
2,4-Dinitrotoluene
Concen: 62.091 ng
RT: 14.774 min Scan# 2003
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:165 Resp: 876137
Ion Ratio Lower Upper
165 100
63 33.8 28.4 42.6
89 69.4 58.0 87.0
182 2.6 2.2 3.2



#58
Fluorene
Concen: 51.068 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:166 Resp: 2643113
Ion Ratio Lower Upper
166 100
165 97.3 77.0 115.4
167 13.8 10.7 16.1









#59
2,3,4,6-Tetrachlorophenol
Concen: 55.608 ng
RT: 15.039 min Scan# 2049
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:232 Resp: 786725
Ion Ratio Lower Upper
232 100
131 45.2 32.9 49.3
130 2.4 1.8 2.8
166 31.2 20.8 31.2



#60
Diethylphthalate
Concen: 55.521 ng
RT: 15.251 min Scan# 2085
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:149 Resp: 2921964
Ion Ratio Lower Upper
149 100
177 22.1 17.2 25.8
150 12.2 10.0 15.0









#61
4-Chlorophenyl-phenylether
Concen: 51.728 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:204 Resp: 1293938
Ion Ratio Lower Upper
204 100
206 32.9 25.4 38.0
141 60.1 47.9 71.9



#62
4-Nitroaniline
Concen: 59.947 ng
RT: 15.480 min Scan# 2124
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36


Tgt Ion:138 Resp: 669832
Ion Ratio Lower Upper
138 100
92 49.0 28.8 68.8
108 77.4 58.1 98.1








#63
Azobenzene
Concen: 53.794 ng
RT: 15.757 min Scan# 2171
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36


Instrument :
BNA_M
ClientSampleId :
SSTDICC050





Tgt Ion: 77 Resp: 2829707
Ion Ratio Lower Upper
77 100
182 26.5 6.8 46.8
105 22.1 1.7 41.7
51 16.3 0.0 36.2




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36



Tgt Ion: 188 Resp: 1523841
Ion Ratio Lower Upper
188 100
94 9.3 8.1 12.1
80 9.9 8.6 13.0









#65
4,6-Dinitro-2-methylphenol
Concen: 58.562 ng
RT: 15.533 min Scan# 2133
Delta R.T. -0.001 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:198 Resp: 487410
Ion Ratio Lower Upper
198 100
51 25.3 4.5 44.5
105 46.7 24.9 64.9



#66
n-Nitrosodiphenylamine
Concen: 51.375 ng
RT: 15.674 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:169 Resp: 2459825
Ion Ratio Lower Upper
169 100
168 66.8 53.1 79.7
167 36.3 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 53.423 ng
RT: 16.357 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:248 Resp: 868132
Ion Ratio Lower Upper
248 100
250 97.0 78.0 117.0
141 69.9 60.1 90.1



#68
Hexachlorobenzene
Concen: 52.310 ng
RT: 16.468 min Scan# 2292
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:284 Resp: 993147
Ion Ratio Lower Upper
284 100
142 32.8 28.8 43.2
249 30.8 25.8 38.6









#69
Atrazine
Concen: 57.610 ng
RT: 16.627 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDIC050


Tgt Ion:200 Resp: 878900
Ion Ratio Lower Upper
200 100
173 26.2 6.0 46.0
215 48.4 28.1 68.1



#70
Pentachlorophenol
Concen: 57.680 ng
RT: 16.804 min Scan# 2349
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:266 Resp: 711985
Ion Ratio Lower Upper
266 100
268 63.9 51.0 76.6
264 62.9 50.8 76.2









#71
Phenanthrene
Concen: 50.793 ng
RT: 17.198 min Scan# 2416
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:178 Resp: 4421661
Ion Ratio Lower Upper
178 100
176 19.4 15.6 23.4
179 15.2 12.3 18.5



#72
Anthracene
Concen: 51.663 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:178 Resp: 4499098
Ion Ratio Lower Upper
178 100
176 18.9 14.8 22.2
179 15.6 12.3 18.5









#73
Carbazole
Concen: 52.985 ng
RT: 17.551 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:167 Resp: 4203693
Ion Ratio Lower Upper
167 100
166 21.1 17.0 25.6
139 12.8 10.2 15.4



#74
Di-n-butylphthalate
Concen: 57.359 ng
RT: 18.139 min Scan# 2576
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:149 Resp: 4951541
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.0
104 5.0 4.0 6.0









#75
Fluoranthene
Concen: 54.004 ng
RT: 19.209 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:202 Resp: 4957680
Ion Ratio Lower Upper
202 100
101 11.4 0.0 31.1
203 17.5 0.0 37.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.392 min Scan# 3129
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:240 Resp: 1446151
Ion Ratio Lower Upper
240 100
120 10.5 9.0 13.4
236 25.2 20.7 31.1









#77
Benzidine
Concen: 61.214 ng
RT: 19.392 min Scan# 2789
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:184 Resp: 2489372
Ion Ratio Lower Upper
184 100
185 14.5 11.5 17.3
183 11.1 9.0 13.4



#78
Pyrene
Concen: 52.769 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:202 Resp: 5143501
Ion Ratio Lower Upper
202 100
200 21.3 16.7 25.1
203 17.7 14.1 21.1









#79
Terphenyl-d14
Concen: 100.800 ng
RT: 19.786 min Scan# 2855
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:244 Resp: 7692376
Ion Ratio Lower Upper
244 100
212 7.6 6.0 9.0
122 10.2 8.6 12.8



#80
Butylbenzylphthalate
Concen: 62.625 ng
RT: 20.486 min Scan# 2975
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:149 Resp: 2036123
Ion Ratio Lower Upper
149 100
91 69.6 56.9 85.3
206 21.5 16.1 24.1









#81
Benzo(a)anthracene
Concen: 52.838 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:228 Resp: 4836514
Ion Ratio Lower Upper
228 100
226 27.8 21.7 32.5
229 19.5 15.7 23.5



#82
3,3'-Dichlorobenzidine
Concen: 61.684 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:252 Resp: 1721840
Ion Ratio Lower Upper
252 100
254 63.7 51.4 77.0
126 11.4 8.6 12.8





#83

Chrysene

Concen: 52.145 ng

RT: 21.433 min Scan# 3135

Delta R.T. 0.005 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:228 Resp: 4540308

Ion Ratio Lower Upper

228 100

226 30.5 24.3 36.5

229 19.8 15.8 23.8



#84

Bis(2-ethylhexyl)phthalate

Concen: 61.400 ng

RT: 21.321 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:149 Resp: 3074950


Ion Ratio Lower Upper

149 100

167 27.7 22.3 33.5

279 4.4 3.5 5.3









#85
Di-n-octyl phthalate
Concen: 66.103 ng
RT: 22.462 min Scan# 3311
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:149 Resp: 4912405
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.7 3.8 5.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:264 Resp: 1421732
Ion Ratio Lower Upper
264 100
260 24.1 18.6 28.0
265 21.5 16.8 25.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 51.183 ng
RT: 27.762 min Scan# 4211
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:276 Resp: 4685372
Ion Ratio Lower Upper
276 100
138 25.7 20.6 30.8
277 24.8 19.8 29.6



#88
Benzo(b)fluoranthene
Concen: 54.563 ng
RT: 23.439 min Scan# 3477
Delta R.T. 0.005 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:252 Resp: 4510196
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 10.2 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 54.567 ng

RT: 23.503 min Scan# 34

Delta R.T. 0.006 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:252 Resp: 4604565

Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.2

125 10.0 8.2 12.4



#90

Benzo(a)pyrene

Concen: 54.752 ng

RT: 24.238 min Scan# 3613

Delta R.T. -0.000 min

Lab File: BM050199.D

Acq: 05 Jun 2025 12:36




Tgt Ion:252 Resp: 4255236


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 11.4 9.0 13.4









#91
Dibenzo(a,h)anthracene
Concen: 51.410 ng
RT: 27.821 min Scan# 4389
Delta R.T. -0.000 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36




Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 3820009
Ion Ratio Lower Upper
278 100
139 15.9 12.8 19.2
279 23.5 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 49.548 ng
RT: 28.803 min Scan# 4389
Delta R.T. 0.006 min
Lab File: BM050199.D
Acq: 05 Jun 2025 12:36

Tgt Ion:276 Resp: 3684864
Ion Ratio Lower Upper
276 100
277 23.7 18.6 27.8
138 21.7 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050200.D
 Acq On : 05 Jun 2025 13:16
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060

Quant Time: Jun 05 14:59:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	258988	20.000	ng	0.00
21) Naphthalene-d8	10.580	136	988400	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	551586	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	997337	20.000	ng	0.00
76) Chrysene-d12	21.386	240	1056136	20.000	ng	0.00
86) Perylene-d12	24.374	264	1194662	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	1901282	122.462	ng	0.00
7) Phenol-d6	6.951	99	2424424	118.624	ng	0.00
23) Nitrobenzene-d5	8.939	82	2406327	126.617	ng	0.00
42) 2,4,6-Tribromophenol	15.898	330	772080	123.062	ng	0.00
45) 2-Fluorobiphenyl	13.045	172	4898196	120.225	ng	0.00
79) Terphenyl-d14	19.780	244	6053045	108.610	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	416291	61.170	ng	99
3) Pyridine	3.681	79	1079098	61.229	ng	100
4) n-Nitrosodimethylamine	3.599	42	227764	62.498	ng	99
6) Aniline	7.116	93	1603778	59.359	ng	100
8) 2-Chlorophenol	7.357	128	1039425	60.866	ng	98
9) Benzaldehyde	6.928	77	624366	48.050	ng	99
10) Phenol	6.981	94	1271828	58.814	ng	99
11) bis(2-Chloroethyl)ether	7.216	93	1006488	57.865	ng	98
12) 1,3-Dichlorobenzene	7.681	146	1154067	60.449	ng	99
13) 1,4-Dichlorobenzene	7.828	146	1184887	59.650	ng	99
14) 1,2-Dichlorobenzene	8.139	146	1126847	59.508	ng	100
15) Benzyl Alcohol	8.022	79	856127	58.731	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.316	45	670877	55.523	ng	98
17) 2-Methylphenol	8.222	107	836890	58.654	ng	100
18) Hexachloroethane	8.875	117	434277	60.370	ng	97
19) n-Nitroso-di-n-propyla...	8.598	70	703599	56.257	ng	99
20) 3+4-Methylphenols	8.551	107	1116126	58.468	ng	99
22) Acetophenone	8.604	105	1467978	59.449	ng	99
24) Nitrobenzene	8.981	77	1077660	62.349	ng	99
25) Isophorone	9.510	82	1934777	58.984	ng	100
26) 2-Nitrophenol	9.692	139	502043	67.289	ng	98
27) 2,4-Dimethylphenol	9.751	122	927674	60.457	ng	98
28) bis(2-Chloroethoxy)met...	9.992	93	1280678	59.764	ng	99
29) 2,4-Dichlorophenol	10.222	162	889620	62.710	ng	100
30) 1,2,4-Trichlorobenzene	10.439	180	977963	61.950	ng	99
31) Naphthalene	10.628	128	3038415	59.457	ng	100
32) Benzoic acid	9.881	122	624735	64.846	ng	98
33) 4-Chloroaniline	10.728	127	1304717	59.885	ng	100
34) Hexachlorobutadiene	10.928	225	584323	62.229	ng	99
35) Caprolactam	11.504	113	271534	60.374	ng	99
36) 4-Chloro-3-methylphenol	11.851	107	896859	59.248	ng	99
37) 2-Methylnaphthalene	12.239	142	1834221	59.497	ng	98
38) 1-Methylnaphthalene	12.457	142	1922042	58.741	ng	99
40) 1,2,4,5-Tetrachloroben...	12.610	216	1003997	63.035	ng	99
41) Hexachlorocyclopentadiene	12.598	237	663873	68.632	ng	97
43) 2,4,6-Trichlorophenol	12.845	196	665409	63.528	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050200.D
 Acq On : 05 Jun 2025 13:16
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060

Quant Time: Jun 05 14:59:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.916	196	730617	63.531	ng	98
46) 1,1'-Biphenyl	13.257	154	2494729	60.402	ng	100
47) 2-Chloronaphthalene	13.292	162	1947861	60.663	ng	100
48) 2-Nitroaniline	13.486	65	479127	67.806	ng	98
49) Acenaphthylene	14.139	152	3183780	61.302	ng	100
50) Dimethylphthalate	13.880	163	2224786	59.640	ng	100
51) 2,6-Dinitrotoluene	13.986	165	490781	65.238	ng	99
52) Acenaphthene	14.480	154	1953486	60.136	ng	99
53) 3-Nitroaniline	14.310	138	555944	66.308	ng	100
54) 2,4-Dinitrophenol	14.516	184	265588	58.459	ng	92
55) Dibenzofuran	14.816	168	2844670	59.857	ng	100
56) 4-Nitrophenol	14.610	139	477024	66.846	ng	98
57) 2,4-Dinitrotoluene	14.768	165	659402	66.447	ng	99
58) Fluorene	15.463	166	2133522	58.613	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.039	232	614431	61.752	ng	93
60) Diethylphthalate	15.251	149	2180022	58.899	ng	99
61) 4-Chlorophenyl-phenyle...	15.463	204	1031087	58.610	ng	99
62) 4-Nitroaniline	15.474	138	528186	67.214	ng	99
63) Azobenzene	15.757	77	2177933	58.871	ng	99
65) 4,6-Dinitro-2-methylph...	15.533	198	359804	66.052	ng	97
66) n-Nitrosodiphenylamine	15.674	169	1911910	61.011	ng	99
67) 4-Bromophenyl-phenylether	16.351	248	649254	61.046	ng	99
68) Hexachlorobenzene	16.462	284	761552	61.287	ng	99
69) Atrazine	16.621	200	634737	63.570	ng	98
70) Pentachlorophenol	16.804	266	528095	65.368	ng	99
71) Phenanthrene	17.198	178	3421380	60.051	ng	100
72) Anthracene	17.286	178	3485282	61.148	ng	100
73) Carbazole	17.551	167	3246806	62.528	ng	100
74) Di-n-butylphthalate	18.133	149	3550989	62.851	ng	100
75) Fluoranthene	19.209	202	3841414	63.934	ng	100
77) Benzidine	19.392	184	1902050	64.043	ng	100
78) Pyrene	19.568	202	4035214	56.687	ng	100
80) Butylbenzylphthalate	20.480	149	1522384	64.115	ng	97
81) Benzo(a)anthracene	21.368	228	4055461	60.666	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	1424421	69.874	ng	100
83) Chrysene	21.433	228	3854906	60.623	ng	99
84) Bis(2-ethylhexyl)phtha...	21.321	149	2349225	64.231	ng	100
85) Di-n-octyl phthalate	22.462	149	3946065	72.708	ng	100
87) Indeno(1,2,3-cd)pyrene	27.762	276	5275171	68.579	ng	# 92
88) Benzo(b)fluoranthene	23.433	252	4222358	60.790	ng	99
89) Benzo(k)fluoranthene	23.503	252	4265446	60.156	ng	100
90) Benzo(a)pyrene	24.238	252	4132525	63.280	ng	100
91) Dibenzo(a,h)anthracene	27.821	278	4258292	68.201	ng	99
92) Benzo(g,h,i)perylene	28.809	276	4279836	68.486	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050200.D
 Acq On : 05 Jun 2025 13:16
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060


Quant Time: Jun 05 14:59:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050200.D
Acq On : 05 Jun 2025 13:16
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Quant Time: Jun 05 14:59:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.787 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:152 Resp: 258988
Ion Ratio Lower Upper
152 100
150 155.9 122.2 183.4
115 60.6 46.3 69.5



#2
1,4-Dioxane
Concen: 61.170 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 88 Resp: 416291
Ion Ratio Lower Upper
88 100
58 52.4 42.9 64.3
43 13.5 11.0 16.6





#3

Pyridine

Concen: 61.229 ng

RT: 3.681 min Scan# 118

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :





BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion: 79 Resp: 1079098

Ion	Ratio	Lower	Upper
79	100		
52	39.8	32.1	48.1
51	19.4	15.4	23.0



#4

n-Nitrosodimethylamine

Concen: 62.498 ng

RT: 3.599 min Scan# 104




Delta R.T. 0.005 min


Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Tgt Ion: 42 Resp: 227764

Ion	Ratio	Lower	Upper
42	100		
74	276.4	222.0	333.0
44	7.4	7.0	10.6









#5
2-Fluorophenol
Concen: 122.462 ng
RT: 5.375 min Scan# 405
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:112 Resp: 1901282
Ion Ratio Lower Upper
112 100
64 47.1 39.8 59.6
63 24.2 19.8 29.8



#6
Aniline
Concen: 59.359 ng
RT: 7.116 min Scan# 702
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 93 Resp: 1603778
Ion Ratio Lower Upper
93 100
66 34.2 27.4 41.2
65 17.3 13.6 20.4









#7
Phenol-d6
Concen: 118.624 ng
RT: 6.951 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 99 Resp: 2424424
Ion Ratio Lower Upper
99 100
42 9.0 6.9 10.3
71 32.3 25.3 37.9



#8
2-Chlorophenol
Concen: 60.866 ng
RT: 7.357 min Scan# 743
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 128 Resp: 1039425
Ion Ratio Lower Upper
128 100
130 32.3 12.6 52.6
64 38.3 19.8 59.8









#9
Benzaldehyde
Concen: 48.050 ng
RT: 6.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 77 Resp: 624366
Ion Ratio Lower Upper
77 100
105 100.4 80.9 120.9
106 95.0 77.1 117.1



#10
Phenol
Concen: 58.814 ng
RT: 6.981 min Scan# 679
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 94 Resp: 1271828
Ion Ratio Lower Upper
94 100
65 23.8 4.3 44.3
66 32.7 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 57.865 ng
RT: 7.216 min Scan# 718
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 93 Resp: 1006488
Ion Ratio Lower Upper
93 100
63 62.2 43.9 83.9
95 32.1 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 60.449 ng
RT: 7.681 min Scan# 798
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 146 Resp: 1154067
Ion Ratio Lower Upper
146 100
148 63.8 51.9 77.9
75 29.4 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 59.650 ng

RT: 7.828 min Scan# 81

Delta R.T. 0.005 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:146 Resp: 1184887

Ion Ratio Lower Upper

146 100

148 63.6 51.4 77.0

111 39.4 32.3 48.5



#14

1,2-Dichlorobenzene

Concen: 59.508 ng

RT: 8.139 min Scan# 876

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:146 Resp: 1126847


Ion Ratio Lower Upper

146 100

148 63.1 50.4 75.6

111 42.6 34.3 51.5









#15
Benzyl Alcohol
Concen: 58.731 ng
RT: 8.022 min Scan# 856
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 79 Resp: 856127
Ion Ratio Lower Upper
79 100
108 79.7 64.0 96.0
77 64.2 52.2 78.2



#16
2,2'-oxybis(1-chloropropane)
Concen: 55.523 ng
RT: 8.316 min Scan# 906
Delta R.T. -0.006 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 45 Resp: 670877
Ion Ratio Lower Upper
45 100
77 30.4 8.7 48.7
79 22.1 2.7 42.7





#17

2-Methylphenol

Concen: 58.654 ng

RT: 8.222 min Scan# 890

Delta R.T. -0.000 min

Lab File: BM050200.D


Acq: 05 Jun 2025 13:16

Instrument :

BNA_M




ClientSampleId :

SSTDICC060



Tgt Ion:107 Resp: 836890

Ion	Ratio	Lower	Upper
107	100		
108	114.3	90.8	136.2
77	44.6	35.7	53.5
79	43.8	35.0	52.4



#18

Hexachloroethane

Concen: 60.370 ng

RT: 8.875 min Scan# 1001




Delta R.T. 0.005 min


Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Tgt Ion:117 Resp: 434277

Ion	Ratio	Lower	Upper
117	100		
119	94.4	77.0	115.6
201	97.5	74.6	111.8









#19
n-Nitroso-di-n-propylamine
Concen: 56.257 ng
RT: 8.598 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 70 Resp: 703599
Ion Ratio Lower Upper
70 100
42 29.3 22.6 33.8
101 10.4 8.3 12.5
130 22.5 17.4 26.2



#20
3+4-Methylphenols
Concen: 58.468 ng
RT: 8.551 min Scan# 946
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 107 Resp: 1116126
Ion Ratio Lower Upper
107 100
108 89.6 70.4 110.4
77 31.7 12.2 52.2
79 27.4 7.7 47.7





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.580 min Scan# 1291

Delta R.T. 0.005 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion:136 Resp: 988400





Ion Ratio Lower Upper

136 100

137 10.8 8.6 13.0

54 4.8 3.8 5.8

68 5.9 4.9 7.3



#22

Acetophenone

Concen: 59.449 ng

RT: 8.604 min Scan# 955

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Tgt Ion:105 Resp: 1467978




Ion Ratio Lower Upper


105 100

71 6.8 5.0 7.4

51 14.5 11.4 17.2

120 22.5 18.2 27.4









#23
Nitrobenzene-d5
Concen: 126.617 ng
RT: 8.939 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 82 Resp: 2406327
Ion Ratio Lower Upper
82 100
128 44.2 35.2 52.8
54 30.7 24.5 36.7



#24
Nitrobenzene
Concen: 62.349 ng
RT: 8.981 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 77 Resp: 1077660
Ion Ratio Lower Upper
77 100
123 48.1 38.2 57.2
65 12.0 9.5 14.3









#25
Isophorone
Concen: 58.984 ng
RT: 9.510 min Scan# 1109
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 82 Resp: 1934777
Ion Ratio Lower Upper
82 100
95 6.8 5.7 8.5
138 17.8 14.2 21.2



#26
2-Nitrophenol
Concen: 67.289 ng
RT: 9.692 min Scan# 1140
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 139 Resp: 502043
Ion Ratio Lower Upper
139 100
109 29.9 23.3 34.9
65 42.7 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 60.457 ng

RT: 9.751 min Scan# 1150

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:122 Resp: 927674

Ion Ratio Lower Upper

122 100

107 115.9 91.0 136.6

121 58.0 45.8 68.6



#28

bis(2-Chloroethoxy)methane

Concen: 59.764 ng

RT: 9.992 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion: 93 Resp: 1280678


Ion Ratio Lower Upper

93 100

95 32.4 26.2 39.4

123 11.5 8.9 13.3





#29

2,4-Dichlorophenol

Concen: 62.710 ng

RT: 10.222 min Scan# 1230

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:162 Resp: 889620

Ion Ratio Lower Upper

162 100

164 64.5 45.0 85.0

98 36.4 16.3 56.3



#30

1,2,4-Trichlorobenzene

Concen: 61.950 ng

RT: 10.439 min Scan# 1267

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:180 Resp: 977963


Ion Ratio Lower Upper

180 100

182 95.2 76.6 114.8

145 30.5 24.2 36.4









#31
Naphthalene
Concen: 59.457 ng
RT: 10.628 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:128 Resp: 3038415
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 12.9 10.2 15.4



#32
Benzoic acid
Concen: 64.846 ng
RT: 9.881 min Scan# 1172
Delta R.T. 0.006 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:122 Resp: 624735
Ion Ratio Lower Upper
122 100
105 123.7 103.4 143.4
77 93.3 70.3 110.3









#33
4-Chloroaniline
Concen: 59.885 ng
RT: 10.728 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 1304717




Ion	Ratio	Lower	Upper
127	100		
129	31.7	25.4	38.0
65	26.0	21.0	31.4
92	18.8	14.8	22.2




#34
Hexachlorobutadiene
Concen: 62.229 ng
RT: 10.928 min Scan# 1350
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:225 Resp: 584323

Ion	Ratio	Lower	Upper
225	100		
223	62.5	50.2	75.2
227	64.3	51.0	76.4









#35
Caprolactam
Concen: 60.374 ng
RT: 11.504 min Scan# 1448
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:113 Resp: 271534
Ion Ratio Lower Upper
113 100
55 101.8 83.5 123.5
56 80.4 60.6 100.6



#36
4-Chloro-3-methylphenol
Concen: 59.248 ng
RT: 11.851 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:107 Resp: 896859
Ion Ratio Lower Upper
107 100
144 26.6 21.0 31.6
142 81.9 64.8 97.2









#37
2-Methylnaphthalene
Concen: 59.497 ng
RT: 12.239 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:142 Resp: 1834221
Ion Ratio Lower Upper
142 100
141 87.1 71.5 107.3
115 34.3 27.8 41.8



#38
1-Methylnaphthalene
Concen: 58.741 ng
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:142 Resp: 1922042
Ion Ratio Lower Upper
142 100
141 90.5 72.9 109.3
116 3.9 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:164 Resp: 551586
Ion Ratio Lower Upper
164 100
162 97.6 80.2 120.4
160 44.8 35.7 53.5



#40
1,2,4,5-Tetrachlorobenzene
Concen: 63.035 ng
RT: 12.610 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16


Tgt Ion:216 Resp: 1003997
Ion Ratio Lower Upper
216 100
214 79.0 63.0 94.4
179 20.7 16.6 25.0
108 19.3 15.9 23.9








#41
Hexachlorocyclopentadiene
Concen: 68.632 ng
RT: 12.598 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16


Instrument :
BNA_M
ClientSampleId :
SSTDICC060





Tgt Ion:237 Resp: 663873
Ion Ratio Lower Upper
237 100
235 61.9 44.6 84.6
272 12.6 0.0 33.1




#42
2,4,6-Tribromophenol
Concen: 123.062 ng
RT: 15.898 min Scan# 2195
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16



Tgt Ion:330 Resp: 772080
Ion Ratio Lower Upper
330 100
332 96.5 77.5 116.3
141 35.9 28.8 43.2





#43

2,4,6-Trichlorophenol

Concen: 63.528 ng

RT: 12.845 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:196 Resp: 665409

Ion Ratio Lower Upper

196 100

198 96.5 76.6 115.0

200 31.6 24.5 36.7



#44

2,4,5-Trichlorophenol

Concen: 63.531 ng

RT: 12.916 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Tgt Ion:196 Resp: 730617




Ion Ratio Lower Upper


196 100

198 96.2 76.1 114.1

97 50.5 39.3 58.9

132 27.9 21.4 32.0









#45
2-Fluorobiphenyl
Concen: 120.225 ng
RT: 13.045 min Scan# 1710
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:172 Resp: 4898196
Ion Ratio Lower Upper
172 100
171 34.9 27.8 41.6
170 23.4 18.6 27.8



#46
1,1'-Biphenyl
Concen: 60.402 ng
RT: 13.257 min Scan# 1746
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:154 Resp: 2494729
Ion Ratio Lower Upper
154 100
153 39.2 19.2 59.2
76 12.3 0.0 32.0





#47

2-Chloronaphthalene

Concen: 60.663 ng

RT: 13.292 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:162 Resp: 1947861

Ion Ratio Lower Upper

162 100

127 38.9 31.1 46.7

164 33.3 26.5 39.7



#48

2-Nitroaniline

Concen: 67.806 ng

RT: 13.486 min Scan# 1785

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion: 65 Resp: 479127


Ion Ratio Lower Upper

65 100

92 78.3 64.2 96.4

138 123.5 101.0 151.4









#49
Acenaphthylene
Concen: 61.302 ng
RT: 14.139 min Scan# 1896
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:152 Resp: 3183780
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.2 10.4 15.6



#50
Dimethylphthalate
Concen: 59.640 ng
RT: 13.880 min Scan# 1852
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:163 Resp: 2224786
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.0 7.9 11.9





#51

2,6-Dinitrotoluene

Concen: 65.238 ng

RT: 13.986 min Scan# 1870

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:165 Resp: 490781

Ion Ratio Lower Upper

165 100

63 45.3 35.8 53.6

89 55.3 44.8 67.2



#52

Acenaphthene

Concen: 60.136 ng

RT: 14.480 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:154 Resp: 1953486


Ion Ratio Lower Upper

154 100

153 112.2 88.7 133.1

152 54.3 42.6 63.8









#53
3-Nitroaniline
Concen: 66.308 ng
RT: 14.310 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:138 Resp: 555944
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 115.5 92.9 139.3



#54
2,4-Dinitrophenol
Concen: 58.459 ng
RT: 14.516 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:184 Resp: 265588
Ion Ratio Lower Upper
184 100
63 51.3 45.3 67.9
154 61.2 54.3 81.5





#55

Dibenzofuran

Concen: 59.857 ng

RT: 14.816 min Scan# 2011

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:168 Resp: 2844670

Ion	Ratio	Lower	Upper

168	100		

139	37.2	29.9	44.9

169	13.3	10.6	16.0



#56

4-Nitrophenol

Concen: 66.846 ng

RT: 14.610 min Scan# 1976

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:139 Resp: 477024


Ion	Ratio	Lower	Upper

139	100		

109	74.4	52.6	92.6

65	85.7	64.3	104.3





#57

2,4-Dinitrotoluene

Concen: 66.447 ng

RT: 14.768 min Scan# 2003

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion:165 Resp: 659402





Ion Ratio Lower Upper

165 100

63 36.5 28.4 42.6

89 72.3 58.0 87.0

182 2.5 2.2 3.2



#58

Fluorene

Concen: 58.613 ng

RT: 15.463 min Scan# 2121

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:166 Resp: 2133522


Ion Ratio Lower Upper

166 100

165 97.2 77.0 115.4

167 13.8 10.7 16.1





#59

2,3,4,6-Tetrachlorophenol

Concen: 61.752 ng

RT: 15.039 min Scan# 2049

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion:232 Resp: 614431





Ion Ratio Lower Upper

232 100

131 45.0 32.9 49.3

130 2.4 1.8 2.8

166 31.1 20.8 31.2



#60

Diethylphthalate

Concen: 58.899 ng

RT: 15.251 min Scan# 2085

Delta R.T. 0.005 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:149 Resp: 2180022


Ion Ratio Lower Upper

149 100

177 22.2 17.2 25.8

150 12.0 10.0 15.0









#61
4-Chlorophenyl-phenylether
Concen: 58.610 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:204 Resp: 1031087
Ion Ratio Lower Upper
204 100
206 32.7 25.4 38.0
141 59.6 47.9 71.9



#62
4-Nitroaniline
Concen: 67.214 ng
RT: 15.474 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:138 Resp: 528186
Ion Ratio Lower Upper
138 100
92 49.7 28.8 68.8
108 77.9 58.1 98.1









#63
Azobenzene
Concen: 58.871 ng
RT: 15.757 min Scan# 2171
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 2177933




Ion	Ratio	Lower	Upper
77	100		
182	27.2	6.8	46.8
105	22.0	1.7	41.7
51	16.1	0.0	36.2




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion: 188 Resp: 997337

Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.1	12.1
80	10.6	8.6	13.0









#65
4,6-Dinitro-2-methylphenol
Concen: 66.052 ng
RT: 15.533 min Scan# 2133
Delta R.T. -0.001 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:198 Resp: 359804
Ion Ratio Lower Upper
198 100
51 23.3 4.5 44.5
105 43.0 24.9 64.9



#66
n-Nitrosodiphenylamine
Concen: 61.011 ng
RT: 15.674 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:169 Resp: 1911910
Ion Ratio Lower Upper
169 100
168 65.7 53.1 79.7
167 36.4 29.0 43.6





#67

4-Bromophenyl-phenylether

Concen: 61.046 ng

RT: 16.351 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:248 Resp: 649254

Ion Ratio Lower Upper

248 100

250 97.2 78.0 117.0

141 77.3 60.1 90.1



#68

Hexachlorobenzene

Concen: 61.287 ng

RT: 16.462 min Scan# 2291

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:284 Resp: 761552


Ion Ratio Lower Upper

284 100

142 35.8 28.8 43.2

249 31.1 25.8 38.6





#69

Atrazine

Concen: 63.570 ng

RT: 16.621 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:200 Resp: 634737

Ion Ratio Lower Upper

200 100

173 27.6 6.0 46.0

215 49.1 28.1 68.1



#70

Pentachlorophenol

Concen: 65.368 ng

RT: 16.804 min Scan# 2349

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:266 Resp: 528095


Ion Ratio Lower Upper

266 100

268 64.4 51.0 76.6

264 64.0 50.8 76.2









#71
Phenanthrene
Concen: 60.051 ng
RT: 17.198 min Scan# 2416
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:178 Resp: 3421380
Ion Ratio Lower Upper
178 100
176 19.4 15.6 23.4
179 15.2 12.3 18.5



#72
Anthracene
Concen: 61.148 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16


Tgt Ion:178 Resp: 3485282
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.3 12.3 18.5








#73
Carbazole
Concen: 62.528 ng
RT: 17.551 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Instrument :
BNA_M
ClientSampleId :
SSTDICC060







Tgt Ion:167 Resp: 3246806
Ion Ratio Lower Upper
167 100
166 21.1 17.0 25.6
139 12.8 10.2 15.4



#74
Di-n-butylphthalate
Concen: 62.851 ng
RT: 18.133 min Scan# 2575
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:149 Resp: 3550989
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.0
104 5.0 4.0 6.0









#75
Fluoranthene
Concen: 63.934 ng
RT: 19.209 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:202 Resp: 3841414
Ion Ratio Lower Upper
202 100
101 11.2 0.0 31.1
203 17.1 0.0 37.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:240 Resp: 1056136
Ion Ratio Lower Upper
240 100
120 11.5 9.0 13.4
236 25.5 20.7 31.1









#77
Benzidine
Concen: 64.043 ng
RT: 19.392 min Scan# 2789
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:184 Resp: 1902050
Ion Ratio Lower Upper
184 100
185 14.5 11.5 17.3
183 11.1 9.0 13.4



#78
Pyrene
Concen: 56.687 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:202 Resp: 4035214
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.4 14.1 21.1









#79
Terphenyl-d14
Concen: 108.610 ng
RT: 19.780 min Scan# 2855
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:244 Resp: 6053045
Ion Ratio Lower Upper
244 100
212 7.5 6.0 9.0
122 11.1 8.6 12.8



#80
Butylbenzylphthalate
Concen: 64.115 ng
RT: 20.480 min Scan# 2974
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:149 Resp: 1522384
Ion Ratio Lower Upper
149 100
91 74.6 56.9 85.3
206 20.5 16.1 24.1









#81
Benzo(a)anthracene
Concen: 60.666 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:228 Resp: 4055461
Ion Ratio Lower Upper
228 100
226 27.1 21.7 32.5
229 19.5 15.7 23.5



#82
3,3'-Dichlorobenzidine
Concen: 69.874 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:252 Resp: 1424421
Ion Ratio Lower Upper
252 100
254 64.0 51.4 77.0
126 11.2 8.6 12.8





#83

Chrysene

Concen: 60.623 ng

RT: 21.433 min Scan# 31

Delta R.T. 0.005 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:228 Resp: 3854906

Ion Ratio Lower Upper

228 100

226 29.6 24.3 36.5

229 19.3 15.8 23.8



#84

Bis(2-ethylhexyl)phthalate

Concen: 64.231 ng

RT: 21.321 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:149 Resp: 2349225


Ion Ratio Lower Upper

149 100

167 27.7 22.3 33.5

279 4.4 3.5 5.3









#85
Di-n-octyl phthalate
Concen: 72.708 ng
RT: 22.462 min Scan# 3311
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:149 Resp: 3946065
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.8 3.8 5.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:264 Resp: 1194662
Ion Ratio Lower Upper
264 100
260 24.1 18.6 28.0
265 22.1 16.8 25.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 68.579 ng
RT: 27.762 min Scan# 4211
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:276 Resp: 5275171
Ion Ratio Lower Upper
276 100
138 18.1 20.6 30.8#
277 24.8 19.8 29.6



#88
Benzo(b)fluoranthene
Concen: 60.790 ng
RT: 23.433 min Scan# 3476
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:252 Resp: 4222358
Ion Ratio Lower Upper
252 100
253 22.3 17.2 25.8
125 10.2 8.2 12.4









#89
Benzo(k)fluoranthene
Concen: 60.156 ng
RT: 23.503 min Scan# 3488
Delta R.T. 0.005 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:252 Resp: 4265446
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 10.0 8.2 12.4



#90
Benzo(a)pyrene
Concen: 63.280 ng
RT: 24.238 min Scan# 3613
Delta R.T. -0.000 min
Lab File: BM050200.D
Acq: 05 Jun 2025 13:16

Tgt Ion:252 Resp: 4132525
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 11.5 9.0 13.4





#91

Dibenzo(a,h)anthracene

Concen: 68.201 ng

RT: 27.821 min Scan# 4390

Delta R.T. -0.000 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:278 Resp: 4258292

Ion Ratio Lower Upper

278 100

139 16.0 12.8 19.2

279 23.7 18.6 28.0



#92

Benzo(g,h,i)perylene

Concen: 68.486 ng

RT: 28.809 min Scan# 4390

Delta R.T. 0.011 min

Lab File: BM050200.D

Acq: 05 Jun 2025 13:16




Tgt Ion:276 Resp: 4279836

Ion Ratio Lower Upper

276 100

277 23.4 18.6 27.8

138 22.3 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050201.D
 Acq On : 05 Jun 2025 13:56
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC080

Quant Time: Jun 05 14:59:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	291644	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	1166744	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	653677	20.000	ng	0.00
64) Phenanthrene-d10	17.157	188	1121619	20.000	ng	0.00
76) Chrysene-d12	21.386	240	978835	20.000	ng	0.00
86) Perylene-d12	24.374	264	1062363	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	2682374	153.426	ng	0.00
7) Phenol-d6	6.957	99	3537202	153.692	ng	0.00
23) Nitrobenzene-d5	8.945	82	3567235	159.011	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	1121188	150.796	ng	0.00
45) 2-Fluorobiphenyl	13.051	172	7130886	147.690	ng	0.00
79) Terphenyl-d14	19.786	244	7496313	145.128	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	568907	74.235	ng	99
3) Pyridine	3.681	79	1547303	77.965	ng	99
4) n-Nitrosodimethylamine	3.599	42	329339	80.251	ng	97
6) Aniline	7.122	93	2369250	77.871	ng	100
8) 2-Chlorophenol	7.357	128	1502113	78.110	ng	98
10) Phenol	6.981	94	1866613	76.653	ng	100
11) bis(2-Chloroethyl)ether	7.216	93	1480567	75.590	ng	98
12) 1,3-Dichlorobenzene	7.681	146	1643168	76.431	ng	99
13) 1,4-Dichlorobenzene	7.828	146	1685110	75.334	ng	100
14) 1,2-Dichlorobenzene	8.140	146	1624137	76.165	ng	99
15) Benzyl Alcohol	8.028	79	1284953	78.279	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.322	45	986129	72.476	ng	98
17) 2-Methylphenol	8.228	107	1245706	77.530	ng	100
18) Hexachloroethane	8.869	117	633589	78.215	ng	99
19) n-Nitroso-di-n-propyla...	8.604	70	1065330	75.641	ng	98
20) 3+4-Methylphenols	8.557	107	1677940	78.056	ng	100
22) Acetophenone	8.610	105	2192477	75.217	ng	100
24) Nitrobenzene	8.987	77	1617938	79.299	ng	99
25) Isophorone	9.516	82	3013669	77.831	ng	100
26) 2-Nitrophenol	9.692	139	788577	89.538	ng	98
27) 2,4-Dimethylphenol	9.757	122	1396376	77.093	ng	98
28) bis(2-Chloroethoxy)met...	9.998	93	1960708	77.513	ng	100
29) 2,4-Dichlorophenol	10.228	162	1350619	80.653	ng	99
30) 1,2,4-Trichlorobenzene	10.445	180	1456329	78.151	ng	100
31) Naphthalene	10.628	128	4521397	74.953	ng	100
32) Benzoic acid	9.916	122	1021552	89.827	ng	99
33) 4-Chloroaniline	10.734	127	1968786	76.553	ng	99
34) Hexachlorobutadiene	10.928	225	876137	79.043	ng	100
35) Caprolactam	11.522	113	417728	78.683	ng	97
36) 4-Chloro-3-methylphenol	11.857	107	1393503	77.985	ng	99
37) 2-Methylnaphthalene	12.239	142	2773230	76.205	ng	99
38) 1-Methylnaphthalene	12.463	142	2903763	75.179	ng	99
40) 1,2,4,5-Tetrachloroben...	12.610	216	1499750	79.455	ng	100
41) Hexachlorocyclopentadiene	12.598	237	1024557	89.378	ng	99
43) 2,4,6-Trichlorophenol	12.845	196	1035769	83.443	ng	100
44) 2,4,5-Trichlorophenol	12.916	196	1122919	82.394	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050201.D
 Acq On : 05 Jun 2025 13:56
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC080

Quant Time: Jun 05 14:59:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.257	154	3756723	76.752	ng	99
47) 2-Chloronaphthalene	13.292	162	2936913	77.180	ng	99
48) 2-Nitroaniline	13.486	65	725090	86.588	ng	98
49) Acenaphthylene	14.139	152	4768813	77.481	ng	100
50) Dimethylphthalate	13.886	163	3356061	75.915	ng	100
51) 2,6-Dinitrotoluene	13.986	165	743566	83.403	ng	98
52) Acenaphthene	14.486	154	2945563	76.514	ng	99
53) 3-Nitroaniline	14.316	138	820084	82.536	ng	98
54) 2,4-Dinitrophenol	14.516	184	419557	79.818	ng	95
55) Dibenzofuran	14.816	168	4229270	75.092	ng	99
56) 4-Nitrophenol	14.616	139	684831	80.978	ng	97
57) 2,4-Dinitrotoluene	14.774	165	994115	84.530	ng	97
58) Fluorene	15.463	166	3057322	70.874	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.039	232	915334	77.626	ng	93
60) Diethylphthalate	15.251	149	3236518	73.786	ng	99
61) 4-Chlorophenyl-phenyle...	15.463	204	1497395	71.823	ng	98
62) 4-Nitroaniline	15.480	138	746446	80.153	ng	99
63) Azobenzene	15.757	77	3216935	73.375	ng	99
65) 4,6-Dinitro-2-methylph...	15.533	198	540790	88.276	ng	98
66) n-Nitrosodiphenylamine	15.674	169	2805368	79.603	ng	100
67) 4-Bromophenyl-phenylether	16.357	248	977940	81.762	ng	97
68) Hexachlorobenzene	16.468	284	1117623	79.976	ng	97
69) Atrazine	16.627	200	913622	81.362	ng	99
70) Pentachlorophenol	16.804	266	766580	84.374	ng	99
71) Phenanthrene	17.198	178	4793533	74.812	ng	100
72) Anthracene	17.286	178	4900912	76.458	ng	99
73) Carbazole	17.551	167	4417127	75.641	ng	100
74) Di-n-butylphthalate	18.133	149	4904225	77.184	ng	100
75) Fluoranthene	19.209	202	5027290	74.400	ng	99
77) Benzidine	19.392	184	2034794	73.924	ng	100
78) Pyrene	19.568	202	5224283	79.187	ng	99
80) Butylbenzylphthalate	20.480	149	1897236	86.212	ng	98
81) Benzo(a)anthracene	21.368	228	4797580	77.435	ng	99
82) 3,3'-Dichlorobenzidine	21.292	252	1627542	86.143	ng	100
83) Chrysene	21.433	228	4547461	77.162	ng	100
84) Bis(2-ethylhexyl)phtha...	21.321	149	2796211	82.490	ng	99
85) Di-n-octyl phthalate	22.462	149	4575834	90.971	ng	100
87) Indeno(1,2,3-cd)pyrene	27.762	276	5980793	87.435	ng	100
88) Benzo(b)fluoranthene	23.439	252	4811379	77.897	ng	99
89) Benzo(k)fluoranthene	23.503	252	4862610	77.118	ng	100
90) Benzo(a)pyrene	24.239	252	4719647	81.270	ng	100
91) Dibenzo(a,h)anthracene	27.821	278	4849919	87.349	ng	99
92) Benzo(g,h,i)perylene	28.815	276	4870648	87.646	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050201.D
Acq On : 05 Jun 2025 13:56
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Quant Time: Jun 05 14:59:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050201.D
Acq On : 05 Jun 2025 13:56
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Quant Time: Jun 05 14:59:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.793 min Scan# 817
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:152 Resp: 291644
Ion Ratio Lower Upper
152 100
150 150.4 122.2 183.4
115 57.2 46.3 69.5



#2
1,4-Dioxane
Concen: 74.235 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 88 Resp: 568907
Ion Ratio Lower Upper
88 100
58 52.9 42.9 64.3
43 13.9 11.0 16.6









#3
Pyridine
Concen: 77.965 ng
RT: 3.681 min Scan# 118
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 79 Resp: 1547303
Ion Ratio Lower Upper
79 100
52 39.3 32.1 48.1
51 19.0 15.4 23.0



#4
n-Nitrosodimethylamine
Concen: 80.251 ng
RT: 3.599 min Scan# 104
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 42 Resp: 329339
Ion Ratio Lower Upper
42 100
74 271.2 222.0 333.0
44 7.0 7.0 10.6





#5

2-Fluorophenol

Concen: 153.426 ng

RT: 5.375 min Scan# 405

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :





BNA_M

ClientSampleId :

SSTDICC080

Tgt Ion:112 Resp: 2682374

Ion	Ratio	Lower	Upper
112	100		
64	47.2	39.8	59.6
63	24.3	19.8	29.8



#6

Aniline

Concen: 77.871 ng

RT: 7.122 min Scan# 703




Delta R.T. 0.006 min


Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Tgt Ion: 93 Resp: 2369250

Ion	Ratio	Lower	Upper
93	100		
66	34.0	27.4	41.2
65	17.0	13.6	20.4









#7
Phenol-d6
Concen: 153.692 ng
RT: 6.957 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 99 Resp: 3537202
Ion Ratio Lower Upper
99 100
42 8.9 6.9 10.3
71 32.0 25.3 37.9



#8
2-Chlorophenol
Concen: 78.110 ng
RT: 7.357 min Scan# 743
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 128 Resp: 1502113
Ion Ratio Lower Upper
128 100
130 32.8 12.6 52.6
64 38.0 19.8 59.8





#10

Phenol

Concen: 76.653 ng

RT: 6.981 min Scan# 61

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :





BNA_M

ClientSampleId :

SSTDICC080

Tgt Ion: 94 Resp: 1866613

Ion	Ratio	Lower	Upper
94	100		
65	24.4	4.3	44.3
66	33.0	12.9	52.9



#11

bis(2-Chloroethyl)ether

Concen: 75.590 ng

RT: 7.216 min Scan# 719




Delta R.T. 0.006 min


Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Tgt Ion: 93 Resp: 1480567

Ion	Ratio	Lower	Upper
93	100		
63	62.1	43.9	83.9
95	32.3	12.7	52.7





#12

1,3-Dichlorobenzene

Concen: 76.431 ng

RT: 7.681 min Scan# 798

Delta R.T. -0.000 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:146 Resp: 1643168

Ion Ratio Lower Upper

146 100

148 64.4 51.9 77.9

75 30.0 23.4 35.0



#13

1,4-Dichlorobenzene

Concen: 75.334 ng

RT: 7.828 min Scan# 823

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56




Tgt Ion:146 Resp: 1685110


Ion Ratio Lower Upper

146 100

148 64.6 51.4 77.0

111 40.6 32.3 48.5





#14

1,2-Dichlorobenzene

Concen: 76.165 ng

RT: 8.140 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:146 Resp: 1624137

Ion Ratio Lower Upper

146 100

148 63.4 50.4 75.6

111 43.3 34.3 51.5



#15

Benzyl Alcohol

Concen: 78.279 ng

RT: 8.028 min Scan# 857

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56




Tgt Ion: 79 Resp: 1284953


Ion Ratio Lower Upper

79 100

108 80.3 64.0 96.0

77 64.0 52.2 78.2









#16
2,2'-oxybis(1-Chloropropane)
Concen: 72.476 ng
RT: 8.322 min Scan# 907
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 45 Resp: 986129
Ion Ratio Lower Upper
45 100
77 30.1 8.7 48.7
79 22.6 2.7 42.7



#17
2-Methylphenol
Concen: 77.530 ng
RT: 8.228 min Scan# 891
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 107 Resp: 1245706
Ion Ratio Lower Upper
107 100
108 113.1 90.8 136.2
77 44.4 35.7 53.5
79 43.2 35.0 52.4





#18

Hexachloroethane

Concen: 78.215 ng

RT: 8.869 min Scan# 1000

Delta R.T. -0.000 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:117 Resp: 633589

Ion Ratio Lower Upper

117 100

119 94.2 77.0 115.6

201 92.8 74.6 111.8



#19

n-Nitroso-di-n-propylamine

Concen: 75.641 ng

RT: 8.604 min Scan# 955

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Tgt Ion: 70 Resp: 1065330




Ion Ratio Lower Upper


70 100

42 29.1 22.6 33.8

101 10.1 8.3 12.5

130 22.6 17.4 26.2









#20
3+4-Methylphenols
Concen: 78.056 ng
RT: 8.557 min Scan# 946
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:107 Resp: 1677940
Ion Ratio Lower Upper
107 100
108 90.3 70.4 110.4
77 32.3 12.2 52.2
79 27.3 7.7 47.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.581 min Scan# 1291
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:136 Resp: 1166744
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 4.8 3.8 5.8
68 6.0 4.9 7.3





#22

Acetophenone

Concen: 75.217 ng

RT: 8.610 min Scan# 91

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080

Tgt Ion:105 Resp: 2192477





Ion Ratio Lower Upper

105 100

71 6.6 5.0 7.4

51 14.2 11.4 17.2

120 22.7 18.2 27.4



#23

Nitrobenzene-d5

Concen: 159.011 ng

RT: 8.945 min Scan# 1013

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56




Tgt Ion: 82 Resp: 3567235


Ion Ratio Lower Upper

82 100

128 44.6 35.2 52.8

54 31.1 24.5 36.7









#24
Nitrobenzene
Concen: 79.299 ng
RT: 8.987 min Scan# 1019
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 77 Resp: 1617938
Ion Ratio Lower Upper
77 100
123 48.3 38.2 57.2
65 11.9 9.5 14.3



#25
Isophorone
Concen: 77.831 ng
RT: 9.516 min Scan# 1110
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 82 Resp: 3013669
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 17.8 14.2 21.2









#26
2-Nitrophenol
Concen: 89.538 ng
RT: 9.692 min Scan# 1139
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:139 Resp: 788577
Ion Ratio Lower Upper
139 100
109 29.6 23.3 34.9
65 42.2 35.2 52.8



#27
2,4-Dimethylphenol
Concen: 77.093 ng
RT: 9.757 min Scan# 1151
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:122 Resp: 1396376
Ion Ratio Lower Upper
122 100
107 116.1 91.0 136.6
121 58.6 45.8 68.6









#28
bis(2-Chloroethoxy)methane
Concen: 77.513 ng
RT: 9.998 min Scan# 1192
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 93 Resp: 1960708
Ion Ratio Lower Upper
93 100
95 32.6 26.2 39.4
123 11.5 8.9 13.3



#29
2,4-Dichlorophenol
Concen: 80.653 ng
RT: 10.228 min Scan# 1231
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:162 Resp: 1350619
Ion Ratio Lower Upper
162 100
164 64.0 45.0 85.0
98 35.8 16.3 56.3









#30
1,2,4-Trichlorobenzene
Concen: 78.151 ng
RT: 10.445 min Scan# 1267
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:180 Resp: 1456329
Ion Ratio Lower Upper
180 100
182 95.8 76.6 114.8
145 29.9 24.2 36.4



#31
Naphthalene
Concen: 74.953 ng
RT: 10.628 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:128 Resp: 4521397
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.8 10.2 15.4









#32
Benzoic acid
Concen: 89.827 ng
RT: 9.916 min Scan# 1178
Delta R.T. 0.041 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:122 Resp: 1021552
Ion Ratio Lower Upper
122 100
105 123.5 103.4 143.4
77 89.2 70.3 110.3



#33
4-Chloroaniline
Concen: 76.553 ng
RT: 10.734 min Scan# 1317
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:127 Resp: 1968786
Ion Ratio Lower Upper
127 100
129 32.1 25.4 38.0
65 25.7 21.0 31.4
92 18.4 14.8 22.2









#34
Hexachlorobutadiene
Concen: 79.043 ng
RT: 10.928 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:225 Resp: 876137
Ion Ratio Lower Upper
225 100
223 62.6 50.2 75.2
227 63.7 51.0 76.4



#35
Caprolactam
Concen: 78.683 ng
RT: 11.522 min Scan# 1451
Delta R.T. 0.023 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:113 Resp: 417728
Ion Ratio Lower Upper
113 100
55 100.6 83.5 123.5
56 77.8 60.6 100.6









#36
4-Chloro-3-methylphenol
Concen: 77.985 ng
RT: 11.857 min Scan# 1507
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:107 Resp: 1393503
Ion Ratio Lower Upper
107 100
144 26.0 21.0 31.6
142 81.5 64.8 97.2



#37
2-Methylnaphthalene
Concen: 76.205 ng
RT: 12.239 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:142 Resp: 2773230
Ion Ratio Lower Upper
142 100
141 87.8 71.5 107.3
115 35.1 27.8 41.8









#38
1-Methylnaphthalene
Concen: 75.179 ng
RT: 12.463 min Scan# 1611
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:142 Resp: 2903763
Ion Ratio Lower Upper
142 100
141 90.3 72.9 109.3
116 3.8 3.0 4.6



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:164 Resp: 653677
Ion Ratio Lower Upper
164 100
162 103.2 80.2 120.4
160 45.8 35.7 53.5









#40
1,2,4,5-Tetrachlorobenzene
Concen: 79.455 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:216 Resp: 1499750
Ion Ratio Lower Upper
216 100
214 78.5 63.0 94.4
179 20.7 16.6 25.0
108 19.5 15.9 23.9



#41
Hexachlorocyclopentadiene
Concen: 89.378 ng
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:237 Resp: 1024557
Ion Ratio Lower Upper
237 100
235 63.7 44.6 84.6
272 12.8 0.0 33.1





#42

2,4,6-Tribromophenol

Concen: 150.796 ng

RT: 15.904 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:330 Resp: 1121188

Ion Ratio Lower Upper

330 100

332 96.3 77.5 116.3

141 35.7 28.8 43.2



#43

2,4,6-Trichlorophenol

Concen: 83.443 ng

RT: 12.845 min Scan# 1676

Delta R.T. -0.000 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56




Tgt Ion:196 Resp: 1035769


Ion Ratio Lower Upper

196 100

198 95.6 76.6 115.0


200 30.6 24.5 36.7








#44
2,4,5-Trichlorophenol
Concen: 82.394 ng
RT: 12.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56


Instrument :
BNA_M
ClientSampleId :
SSTDICC080





Tgt Ion:196 Resp: 1122919
Ion Ratio Lower Upper
196 100
198 96.2 76.1 114.1
97 49.8 39.3 58.9
132 27.3 21.4 32.0




#45
2-Fluorobiphenyl
Concen: 147.690 ng
RT: 13.051 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56



Tgt Ion:172 Resp: 7130886
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.6
170 23.4 18.6 27.8





#46

1,1'-Biphenyl

Concen: 76.752 ng

RT: 13.257 min Scan# 1746

Delta R.T. -0.000 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:154 Resp: 3756723

Ion Ratio Lower Upper

154 100

153 39.7 19.2 59.2

76 12.2 0.0 32.0



#47

2-Chloronaphthalene

Concen: 77.180 ng

RT: 13.292 min Scan# 1752

Delta R.T. -0.000 min

Lab File: BM050201.D

Acq: 05 Jun 2025 13:56




Tgt Ion:162 Resp: 2936913


Ion Ratio Lower Upper

162 100

127 39.4 31.1 46.7

164 32.6 26.5 39.7









#48
2-Nitroaniline
Concen: 86.588 ng
RT: 13.486 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 65 Resp: 725090
Ion Ratio Lower Upper
65 100
92 80.2 64.2 96.4
138 123.1 101.0 151.4



#49
Acenaphthylene
Concen: 77.481 ng
RT: 14.139 min Scan# 1896
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 152 Resp: 4768813
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.0 10.4 15.6









#50
Dimethylphthalate
Concen: 75.915 ng
RT: 13.886 min Scan# 1852
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:163 Resp: 3356061
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 9.7 7.9 11.9



#51
2,6-Dinitrotoluene
Concen: 83.403 ng
RT: 13.986 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:165 Resp: 743566
Ion Ratio Lower Upper
165 100
63 46.9 35.8 53.6
89 57.1 44.8 67.2









#52
Acenaphthene
Concen: 76.514 ng
RT: 14.486 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:154 Resp: 2945563
Ion Ratio Lower Upper
154 100
153 112.1 88.7 133.1
152 53.4 42.6 63.8



#53
3-Nitroaniline
Concen: 82.536 ng
RT: 14.316 min Scan# 1926
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:138 Resp: 820084
Ion Ratio Lower Upper
138 100
108 10.7 8.8 13.2
92 113.5 92.9 139.3









#54
2,4-Dinitrophenol
Concen: 79.818 ng
RT: 14.516 min Scan# 1959
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:184 Resp: 419557
Ion Ratio Lower Upper
184 100
63 54.1 45.3 67.9
154 62.3 54.3 81.5



#55
Dibenzofuran
Concen: 75.092 ng
RT: 14.816 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:168 Resp: 4229270
Ion Ratio Lower Upper
168 100
139 37.8 29.9 44.9
169 13.5 10.6 16.0









#56
4-Nitrophenol
Concen: 80.978 ng
RT: 14.616 min Scan# 1976
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:139 Resp: 684831
Ion Ratio Lower Upper
139 100
109 74.9 52.6 92.6
65 81.6 64.3 104.3



#57
2,4-Dinitrotoluene
Concen: 84.530 ng
RT: 14.774 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:165 Resp: 994115
Ion Ratio Lower Upper
165 100
63 35.3 28.4 42.6
89 69.3 58.0 87.0
182 2.7 2.2 3.2









#58
Fluorene
Concen: 70.874 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:166 Resp: 3057322
Ion Ratio Lower Upper
166 100
165 97.4 77.0 115.4
167 13.8 10.7 16.1



#59
2,3,4,6-Tetrachlorophenol
Concen: 77.626 ng
RT: 15.039 min Scan# 2049
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:232 Resp: 915334
Ion Ratio Lower Upper
232 100
131 44.8 32.9 49.3
130 2.4 1.8 2.8
166 31.0 20.8 31.2









#60
Diethylphthalate
Concen: 73.786 ng
RT: 15.251 min Scan# 2085
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 3236518
Ion Ratio Lower Upper
149 100
177 22.4 17.2 25.8
150 12.3 10.0 15.0



#61
4-Chlorophenyl-phenylether
Concen: 71.823 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:204 Resp: 1497395
Ion Ratio Lower Upper
204 100
206 32.4 25.4 38.0
141 61.3 47.9 71.9









#62
4-Nitroaniline
Concen: 80.153 ng
RT: 15.480 min Scan# 2124
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:138 Resp: 746446
Ion Ratio Lower Upper
138 100
92 48.8 28.8 68.8
108 77.1 58.1 98.1



#63
Azobenzene
Concen: 73.375 ng
RT: 15.757 min Scan# 2171
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion: 77 Resp: 3216935
Ion Ratio Lower Upper
77 100
182 26.5 6.8 46.8
105 22.3 1.7 41.7
51 16.7 0.0 36.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2408
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:188 Resp: 1121619
Ion Ratio Lower Upper
188 100
94 9.4 8.1 12.1
80 10.2 8.6 13.0



#65
4,6-Dinitro-2-methylphenol
Concen: 88.276 ng
RT: 15.533 min Scan# 2133
Delta R.T. -0.001 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:198 Resp: 540790
Ion Ratio Lower Upper
198 100
51 25.7 4.5 44.5
105 46.3 24.9 64.9









#66
n-Nitrosodiphenylamine
Concen: 79.603 ng
RT: 15.674 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:169 Resp: 2805368
Ion Ratio Lower Upper
169 100
168 66.5 53.1 79.7
167 36.6 29.0 43.6



#67
4-Bromophenyl-phenylether
Concen: 81.762 ng
RT: 16.357 min Scan# 2273
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:248 Resp: 977940
Ion Ratio Lower Upper
248 100
250 96.2 78.0 117.0
141 71.2 60.1 90.1









#68
Hexachlorobenzene
Concen: 79.976 ng
RT: 16.468 min Scan# 2119
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:284 Resp: 1117623
Ion Ratio Lower Upper
284 100
142 33.9 28.8 43.2
249 30.7 25.8 38.6



#69
Atrazine
Concen: 81.362 ng
RT: 16.627 min Scan# 2319
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:200 Resp: 913622
Ion Ratio Lower Upper
200 100
173 26.1 6.0 46.0
215 48.9 28.1 68.1









#70
Pentachlorophenol
Concen: 84.374 ng
RT: 16.804 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:266 Resp: 766580
Ion Ratio Lower Upper
266 100
268 62.7 51.0 76.6
264 62.7 50.8 76.2



#71
Phenanthrene
Concen: 74.812 ng
RT: 17.198 min Scan# 2416
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:178 Resp: 4793533
Ion Ratio Lower Upper
178 100
176 19.4 15.6 23.4
179 15.5 12.3 18.5









#72
Anthracene
Concen: 76.458 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:178 Resp: 4900912
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.6 12.3 18.5



#73
Carbazole
Concen: 75.641 ng
RT: 17.551 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:167 Resp: 4417127
Ion Ratio Lower Upper
167 100
166 21.3 17.0 25.6
139 13.2 10.2 15.4









#74
Di-n-butylphthalate
Concen: 77.184 ng
RT: 18.133 min Scan# 2575
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 4904225
Ion Ratio Lower Upper
149 100
150 9.2 7.4 11.0
104 5.2 4.0 6.0



#75
Fluoranthene
Concen: 74.400 ng
RT: 19.209 min Scan# 2758
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:202 Resp: 5027290
Ion Ratio Lower Upper
202 100
101 11.2 0.0 31.1
203 17.5 0.0 37.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:240 Resp: 978835
Ion Ratio Lower Upper
240 100
120 11.6 9.0 13.4
236 25.5 20.7 31.1



#77
Benzidine
Concen: 73.924 ng
RT: 19.392 min Scan# 2789
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:184 Resp: 2034794
Ion Ratio Lower Upper
184 100
185 14.5 11.5 17.3
183 11.3 9.0 13.4









#78
Pyrene
Concen: 79.187 ng
RT: 19.568 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:202 Resp: 5224283
Ion Ratio Lower Upper
202 100
200 21.2 16.7 25.1
203 17.4 14.1 21.1



#79
Terphenyl-d14
Concen: 145.128 ng
RT: 19.786 min Scan# 2856
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:244 Resp: 7496313
Ion Ratio Lower Upper
244 100
212 7.5 6.0 9.0
122 10.0 8.6 12.8









#80
Butylbenzylphthalate
Concen: 86.212 ng
RT: 20.480 min Scan# 2974
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 1897236
Ion Ratio Lower Upper
149 100
91 72.9 56.9 85.3
206 20.7 16.1 24.1



#81
Benzo(a)anthracene
Concen: 77.435 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:228 Resp: 4797580
Ion Ratio Lower Upper
228 100
226 27.7 21.7 32.5
229 19.9 15.7 23.5









#82
3,3'-Dichlorobenzidine
Concen: 86.143 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 1627542
Ion Ratio Lower Upper
252 100
254 64.3 51.4 77.0
126 11.1 8.6 12.8



#83
Chrysene
Concen: 77.162 ng
RT: 21.433 min Scan# 3136
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:228 Resp: 4547461
Ion Ratio Lower Upper
228 100
226 30.5 24.3 36.5
229 19.6 15.8 23.8









#84
Bis(2-ethylhexyl)phthalate
Concen: 82.490 ng
RT: 21.321 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 2796211
Ion Ratio Lower Upper
149 100
167 27.5 22.3 33.5
279 4.4 3.5 5.3



#85
Di-n-octyl phthalate
Concen: 90.971 ng
RT: 22.462 min Scan# 3311
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:149 Resp: 4575834
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.7 3.8 5.8









#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 30
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:264 Resp: 1062363
Ion Ratio Lower Upper
264 100
260 23.1 18.6 28.0
265 22.3 16.8 25.2



#87
Indeno(1,2,3-cd)pyrene
Concen: 87.435 ng
RT: 27.762 min Scan# 4212
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:276 Resp: 5980793
Ion Ratio Lower Upper
276 100
138 25.6 20.6 30.8
277 24.9 19.8 29.6









#88
Benzo(b)fluoranthene
Concen: 77.897 ng
RT: 23.439 min Scan# 3476
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 4811379
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 9.9 8.2 12.4



#89
Benzo(k)fluoranthene
Concen: 77.118 ng
RT: 23.503 min Scan# 3488
Delta R.T. 0.006 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:252 Resp: 4862610
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 9.9 8.2 12.4









#90
Benzo(a)pyrene
Concen: 81.270 ng
RT: 24.239 min Scan# 3613
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 4719647
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 11.2 9.0 13.4



#91
Dibenzo(a,h)anthracene
Concen: 87.349 ng
RT: 27.821 min Scan# 4222
Delta R.T. -0.000 min
Lab File: BM050201.D
Acq: 05 Jun 2025 13:56

Tgt Ion:278 Resp: 4849919
Ion Ratio Lower Upper
278 100
139 16.2 12.8 19.2
279 23.8 18.6 28.0





#92

Benzo(g,h,i)perylene

Concen: 87.646 ng

RT: 28.815 min Scan# 41

Delta R.T. 0.017 min

Lab File: BM050201.D


Acq: 05 Jun 2025 13:56

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:276 Resp: 4870648

Ion Ratio Lower Upper

276 100

277 23.3 18.6 27.8

138 21.5 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/05/2025
 Supervised By :Jagrut Upadhyay 06/06/2025

Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	318987	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	1345797	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	824451	20.000	ng	0.00
64) Phenanthrene-d10	17.151	188	1474581	20.000	ng	0.00
76) Chrysene-d12	21.386	240	1199833	20.000	ng	0.00
86) Perylene-d12	24.374	264	1188981	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	1363817	71.321	ng	0.00
7) Phenol-d6	6.951	99	1828950	72.656	ng	0.00
23) Nitrobenzene-d5	8.940	82	1707276	65.977	ng	0.00
42) 2,4,6-Tribromophenol	15.898	330	671004	71.554	ng	0.00
45) 2-Fluorobiphenyl	13.045	172	3838077	63.026	ng	0.00
79) Terphenyl-d14	19.780	244	4364574	68.934	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	285619	34.075	ng	98
3) Pyridine	3.681	79	807832	37.216	ng	99
4) n-Nitrosodimethylamine	3.593	42	171804	38.276	ng	95
6) Aniline	7.116	93	1270681	38.184	ng	99
8) 2-Chlorophenol	7.351	128	812307	38.620	ng	99
9) Benzaldehyde	6.928	77	662318m	41.383	ng	
10) Phenol	6.975	94	1020433	38.313	ng	99
11) bis(2-Chloroethyl)ether	7.210	93	819158	38.237	ng	99
12) 1,3-Dichlorobenzene	7.681	146	890155	37.856	ng	99
13) 1,4-Dichlorobenzene	7.822	146	914542	37.381	ng	99
14) 1,2-Dichlorobenzene	8.140	146	881362	37.789	ng	99
15) Benzyl Alcohol	8.022	79	702528	39.129	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.316	45	556648	37.404	ng	98
17) 2-Methylphenol	8.222	107	686759	39.079	ng	99
18) Hexachloroethane	8.869	117	336898	38.024	ng	100
19) n-Nitroso-di-n-propyla...	8.598	70	613129	39.802	ng	99
20) 3+4-Methylphenols	8.551	107	925708	39.372	ng	100
22) Acetophenone	8.604	105	1264911	37.622	ng	99
24) Nitrobenzene	8.981	77	897378	38.131	ng	99
25) Isophorone	9.510	82	1719407	38.498	ng	99
26) 2-Nitrophenol	9.687	139	415162	40.867	ng	99
27) 2,4-Dimethylphenol	9.751	122	791547	37.886	ng	99
28) bis(2-Chloroethoxy)met...	9.992	93	1110519	38.061	ng	99
29) 2,4-Dichlorophenol	10.222	162	753084	38.988	ng	99
30) 1,2,4-Trichlorobenzene	10.439	180	807629	37.573	ng	100
31) Naphthalene	10.628	128	2548917	36.633	ng	100
32) Benzoic acid	9.875	122	519441	39.598	ng	98
33) 4-Chloroaniline	10.728	127	1127809	38.018	ng	99
34) Hexachlorobutadiene	10.922	225	479592	37.511	ng	99
35) Caprolactam	11.498	113	237644	38.807	ng	99
36) 4-Chloro-3-methylphenol	11.851	107	794929	38.568	ng	100
37) 2-Methylnaphthalene	12.239	142	1594515	37.986	ng	98
38) 1-Methylnaphthalene	12.457	142	1680613	37.722	ng	100
40) 1,2,4,5-Tetrachloroben...	12.610	216	900972	37.845	ng	99
41) Hexachlorocyclopentadiene	12.598	237	569438	39.386	ng	99
43) 2,4,6-Trichlorophenol	12.845	196	598909	38.255	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 06/05/2025
 Supervised By :Jagrut Upadhyay 06/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.916	196	658288	38.297	ng	100
46) 1,1'-Biphenyl	13.251	154	2269641	36.765	ng	99
47) 2-Chloronaphthalene	13.292	162	1740708	36.269	ng	100
48) 2-Nitroaniline	13.486	65	425146	40.253	ng	98
49) Acenaphthylene	14.139	152	2862688	36.877	ng	99
50) Dimethylphthalate	13.880	163	2084186	37.379	ng	100
51) 2,6-Dinitrotoluene	13.986	165	446609	39.718	ng	99
52) Acenaphthene	14.480	154	1769507	36.444	ng	99
53) 3-Nitroaniline	14.310	138	493099	39.347	ng	96
54) 2,4-Dinitrophenol	14.510	184	228596	34.596	ng	99
55) Dibenzofuran	14.816	168	2585454	36.397	ng	99
56) 4-Nitrophenol	14.610	139	411593	38.588	ng	98
57) 2,4-Dinitrotoluene	14.769	165	592521	39.946	ng	99
58) Fluorene	15.463	166	1965747	36.130	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.039	232	557824	37.508	ng	93
60) Diethylphthalate	15.245	149	2031289	36.717	ng	99
61) 4-Chlorophenyl-phenyle...	15.463	204	950092	36.132	ng	99
62) 4-Nitroaniline	15.469	138	449658	38.283	ng	99
63) Azobenzene	15.751	77	1998483	36.142	ng	98
65) 4,6-Dinitro-2-methylph...	15.533	198	312558	38.808	ng	99
66) n-Nitrosodiphenylamine	15.669	169	1747981	37.727	ng	100
67) 4-Bromophenyl-phenylether	16.351	248	604660	38.453	ng	99
68) Hexachlorobenzene	16.463	284	696502	37.911	ng	100
69) Atrazine	16.621	200	578491	39.186	ng	99
70) Pentachlorophenol	16.804	266	464946	38.925	ng	99
71) Phenanthrene	17.198	178	3046863	36.170	ng	99
72) Anthracene	17.286	178	3101683	36.806	ng	100
73) Carbazole	17.551	167	2778780	36.195	ng	99
74) Di-n-butylphthalate	18.133	149	3122305	37.377	ng	99
75) Fluoranthene	19.209	202	3117387	35.092	ng	100
77) Benzidine	19.392	184	1646454	48.798	ng	100
78) Pyrene	19.568	202	3262895	40.348	ng	99
80) Butylbenzylphthalate	20.480	149	1095153	40.599	ng	98
81) Benzo(a)anthracene	21.368	228	2824621	37.193	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	933068	40.289	ng	100
83) Chrysene	21.433	228	2642484	36.579	ng	99
84) Bis(2-ethylhexyl)phtha...	21.321	149	1606973	38.675	ng	100
85) Di-n-octyl phthalate	22.462	149	2382511	38.642	ng	100
87) Indeno(1,2,3-cd)pyrene	27.750	276	2970902	38.807	ng	100
88) Benzo(b)fluoranthene	23.433	252	2608359	37.732	ng	100
89) Benzo(k)fluoranthene	23.497	252	2566064	36.362	ng	100
90) Benzo(a)pyrene	24.239	252	2459411	37.840	ng	99
91) Dibenzo(a,h)anthracene	27.815	278	2412391	38.821	ng	100
92) Benzo(g,h,i)perylene	28.791	276	2411378	38.771	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 06/05/2025
 Supervised By :Jagrut Upadhyay 06/06/2025


Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration




Instrument :
BNA_M
ClientSampleId :
ICVBM060525

Manual Integrations


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.787 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:152 Resp: 31898
Ion Ratio Lower Upper
152 100
150 159.9 122.2 183.4
115 61.4 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#2
1,4-Dioxane
Concen: 34.075 ng
RT: 3.293 min Scan# 52
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion: 88 Resp: 285619
Ion Ratio Lower Upper
88 100
58 52.3 42.9 64.3
43 13.5 11.0 16.6





#3
Pyridine
Concen: 37.216 ng
RT: 3.681 min Scan# 118
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion: 79 Resp: 807832
Ion Ratio Lower Upper
79 100
52 39.8 32.1 48.1
51 18.6 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#4
n-Nitrosodimethylamine
Concen: 38.276 ng
RT: 3.593 min Scan# 103
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion: 42 Resp: 171804
Ion Ratio Lower Upper
42 100
74 267.3 222.0 333.0
44 7.5 7.0 10.6





#5
2-Fluorophenol
Concen: 71.321 ng
RT: 5.369 min Scan# 405
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:112 Resp: 136381
Ion Ratio Lower Upper
112 100
64 49.0 39.8 59.6
63 25.4 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#6
Aniline
Concen: 38.184 ng
RT: 7.116 min Scan# 702
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion: 93 Resp: 1270681
Ion Ratio Lower Upper
93 100
66 34.9 27.4 41.2
65 16.9 13.6 20.4





#7
Phenol-d6
Concen: 72.656 ng
RT: 6.951 min Scan# 674
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion: 99 Resp: 1828950
Ion Ratio Lower Upper
99 100
42 8.8 6.9 10.3
71 31.5 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#8
2-Chlorophenol
Concen: 38.620 ng
RT: 7.351 min Scan# 742
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:128 Resp: 812307
Ion Ratio Lower Upper
128 100
130 31.9 12.6 52.6
64 39.3 19.8 59.8





#9

Benzaldehyde

Concen: 41.383 ng m

RT: 6.928 min Scan# 6

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion: 77 Resp: 662318

Ion Ratio Lower Upper

77 100

105 101.1 80.9 120.9





106 97.0 77.1 117.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#10

Phenol

Concen: 38.313 ng

RT: 6.975 min Scan# 678

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion: 94 Resp: 1020433


Ion Ratio Lower Upper

94 100

65 24.4 4.3 44.3

66 32.1 12.9 52.9





#11

bis(2-Chloroethyl)ether

Concen: 38.237 ng

RT: 7.210 min Scan# 718

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion: 93 Resp: 81915

Ion Ratio Lower Upper

93 100

63 64.4 43.9 83.9





95 33.0 12.7 52.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#12

1,3-Dichlorobenzene

Concen: 37.856 ng

RT: 7.681 min Scan# 798

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion:146 Resp: 890155


Ion Ratio Lower Upper

146 100

148 64.6 51.9 77.9

75 30.0 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 37.381 ng

RT: 7.822 min Scan# 822

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion:146 Resp: 91454

Ion Ratio Lower Upper

146 100

148 63.9 51.4 77.0





111 41.1 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#14

1,2-Dichlorobenzene

Concen: 37.789 ng

RT: 8.140 min Scan# 876

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion:146 Resp: 881362


Ion Ratio Lower Upper

146 100

148 63.4 50.4 75.6


111 42.4 34.3 51.5





#15
Benzyl Alcohol
Concen: 39.129 ng
RT: 8.022 min Scan# 856
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion: 79 Resp: 702528
Ion Ratio Lower Upper
79 100
108 81.3 64.0 96.0
77 65.1 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.404 ng
RT: 8.316 min Scan# 906
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion: 45 Resp: 556648
Ion Ratio Lower Upper
45 100
77 29.9 8.7 48.7
79 22.4 2.7 42.7





#17

2-Methylphenol

Concen: 39.079 ng

RT: 8.222 min Scan# 890

Delta R.T. -0.000 min

Lab File: BM050202.D


Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525






Tgt Ion:	107	Resp:	686759
Ion	Ratio	Lower	Upper
107	100		
108	112.2	90.8	136.2
77	45.5	35.7	53.5
79	44.9	35.0	52.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#18

Hexachloroethane


Concen: 38.024 ng

RT: 8.869 min Scan# 1000



Delta R.T. -0.000 min


Lab File: BM050202.D

Acq: 05 Jun 2025 14:36



Tgt Ion:	117	Resp:	336898
Ion	Ratio	Lower	Upper
117	100		
119	96.2	77.0	115.6
201	93.2	74.6	111.8





#19

n-Nitroso-di-n-propylamine

Concen: 39.802 ng

RT: 8.598 min Scan# 954

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion: 70 Resp: 613129

Ion Ratio Lower Upper

70 100

42 29.4 22.6 33.8

101 10.5 8.3 12.5





130 21.9 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#20

3+4-Methylphenols

Concen: 39.372 ng

RT: 8.551 min Scan# 946

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Tgt Ion:107 Resp: 925708




Ion Ratio Lower Upper


107 100

108 90.3 70.4 110.4

77 31.8 12.2 52.2


79 27.7 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.581 min Scan# 1291
Delta R.T. 0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:136 Resp: 134579
Ion Ratio Lower Upper
136 100
137 10.7 8.6 13.0
54 4.6 3.8 5.8
68 5.8 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#22
Acetophenone
Concen: 37.622 ng
RT: 8.604 min Scan# 955
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:105 Resp: 1264911
Ion Ratio Lower Upper
105 100
71 5.8 5.0 7.4
51 14.0 11.4 17.2
120 22.7 18.2 27.4





#23
Nitrobenzene-d5
Concen: 65.977 ng
RT: 8.940 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion: 82 Resp: 1707270
Ion Ratio Lower Upper
82 100
128 43.4 35.2 52.8
54 30.6 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#24
Nitrobenzene
Concen: 38.131 ng
RT: 8.981 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion: 77 Resp: 897378
Ion Ratio Lower Upper
77 100
123 48.1 38.2 57.2
65 12.5 9.5 14.3





#25
Isophorone
Concen: 38.498 ng
RT: 9.510 min Scan# 1109
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion: 82 Resp: 171940
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 18.1 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#26
2-Nitrophenol
Concen: 40.867 ng
RT: 9.687 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:139 Resp: 415162
Ion Ratio Lower Upper
139 100
109 30.6 23.3 34.9
65 43.9 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 37.886 ng

RT: 9.751 min Scan# 1150

Delta R.T. -0.000 min

Lab File: BM050202.D


Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525



Tgt Ion:122 Resp: 79154

Ion Ratio Lower Upper

122 100

107 114.0 91.0 136.6




121 58.2 45.8 68.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#28

bis(2-Chloroethoxy)methane

Concen: 38.061 ng

RT: 9.992 min Scan# 1191

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion: 93 Resp: 1110519


Ion Ratio Lower Upper

93 100

95 32.3 26.2 39.4

123 11.3 8.9 13.3





#29

2,4-Dichlorophenol

Concen: 38.988 ng

RT: 10.222 min Scan# 1230

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion:162 Resp: 753084

Ion Ratio Lower Upper

162 100

164 64.2 45.0 85.0





98 37.6 16.3 56.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#30

1,2,4-Trichlorobenzene

Concen: 37.573 ng

RT: 10.439 min Scan# 1267

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion:180 Resp: 807629


Ion Ratio Lower Upper

180 100

182 95.6 76.6 114.8


145 30.2 24.2 36.4





#33
4-Chloroaniline
Concen: 38.018 ng
RT: 10.728 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:127 Resp: 1127809
Ion Ratio Lower Upper
127 100
129 32.4 25.4 38.0
65 26.0 21.0 31.4
92 18.6 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#34
Hexachlorobutadiene
Concen: 37.511 ng
RT: 10.922 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:225 Resp: 479592
Ion Ratio Lower Upper
225 100
223 63.7 50.2 75.2
227 62.9 51.0 76.4





#35
Caprolactam
Concen: 38.807 ng
RT: 11.498 min Scan# 1447
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:113 Resp: 237644
Ion Ratio Lower Upper
113 100
55 101.9 83.5 123.5
56 80.6 60.6 100.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#36
4-Chloro-3-methylphenol
Concen: 38.568 ng
RT: 11.851 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:107 Resp: 794929
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.6
142 80.8 64.8 97.2





#37
2-Methylnaphthalene
Concen: 37.986 ng
RT: 12.239 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525




Tgt Ion:142 Resp: 1594515
Ion Ratio Lower Upper
142 100
141 87.8 71.5 107.3
115 33.8 27.8 41.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#38
1-Methylnaphthalene
Concen: 37.722 ng
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Tgt Ion:142 Resp: 1680613
Ion Ratio Lower Upper
142 100
141 91.5 72.9 109.3
116 3.8 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:164 Resp: 82445
Ion Ratio Lower Upper
164 100
162 99.7 80.2 120.4
160 44.2 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.845 ng
RT: 12.610 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:216 Resp: 900972
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.4
179 20.7 16.6 25.0
108 19.4 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 39.386 ng
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:237 Resp: 569438
Ion Ratio Lower Upper
237 100
235 64.1 44.6 84.6
272 13.1 0.0 33.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#42
2,4,6-Tribromophenol
Concen: 71.554 ng
RT: 15.898 min Scan# 2195
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:330 Resp: 671004
Ion Ratio Lower Upper
330 100
332 96.6 77.5 116.3
141 35.6 28.8 43.2





#43

2,4,6-Trichlorophenol

Concen: 38.255 ng

RT: 12.845 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM050202.D


Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525



Tgt Ion:196 Resp: 598909

Ion Ratio Lower Upper

196 100

198 98.3 76.6 115.0




200 32.3 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#44

2,4,5-Trichlorophenol

Concen: 38.297 ng

RT: 12.916 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Tgt Ion:196 Resp: 658288




Ion Ratio Lower Upper


196 100

198 94.7 76.1 114.1

97 49.0 39.3 58.9

132 27.2 21.4 32.0





#45

2-Fluorobiphenyl

Concen: 63.026 ng

RT: 13.045 min Scan# 1710

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion:172 Resp: 383807

Ion Ratio Lower Upper

172 100

171 34.5 27.8 41.6





170 22.9 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#46

1,1'-Biphenyl

Concen: 36.765 ng

RT: 13.251 min Scan# 1745

Delta R.T. -0.006 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion:154 Resp: 2269641


Ion Ratio Lower Upper

154 100

153 39.5 19.2 59.2

76 12.7 0.0 32.0





#47

2-Chloronaphthalene

Concen: 36.269 ng

RT: 13.292 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM050202.D


Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525



Tgt Ion:162 Resp: 1740708

Ion Ratio Lower Upper

162 100

127 38.5 31.1 46.7




164 33.0 26.5 39.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#48

2-Nitroaniline

Concen: 40.253 ng

RT: 13.486 min Scan# 1785

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion: 65 Resp: 425146


Ion Ratio Lower Upper

65 100

92 79.8 64.2 96.4


138 123.8 101.0 151.4





#49
Acenaphthylene
Concen: 36.877 ng
RT: 14.139 min Scan# 1896
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:152 Resp: 2862688
Ion Ratio Lower Upper
152 100
151 19.3 15.8 23.6
153 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#50
Dimethylphthalate
Concen: 37.379 ng
RT: 13.880 min Scan# 1852
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:163 Resp: 2084186
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 9.9 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 39.718 ng
RT: 13.986 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:165 Resp: 446609
Ion Ratio Lower Upper
165 100
63 45.4 35.8 53.6
89 55.3 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#52
Acenaphthene
Concen: 36.444 ng
RT: 14.480 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:154 Resp: 1769507
Ion Ratio Lower Upper
154 100
153 112.5 88.7 133.1
152 54.0 42.6 63.8





#53
3-Nitroaniline
Concen: 39.347 ng
RT: 14.310 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:138 Resp: 493099
Ion Ratio Lower Upper
138 100
108 10.9 8.8 13.2
92 111.4 92.9 139.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#54
2,4-Dinitrophenol
Concen: 34.596 ng
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:184 Resp: 228596
Ion Ratio Lower Upper
184 100
63 56.0 45.3 67.9
154 67.7 54.3 81.5





#55
Dibenzenofuran
Concen: 36.397 ng
RT: 14.816 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:168 Resp: 2585454
Ion Ratio Lower Upper
168 100
139 36.4 29.9 44.9
169 12.9 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#56
4-Nitrophenol
Concen: 38.588 ng
RT: 14.610 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:139 Resp: 411593
Ion Ratio Lower Upper
139 100
109 72.8 52.6 92.6
65 81.6 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 39.946 ng
RT: 14.769 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:165 Resp: 59252
Ion Ratio Lower Upper
165 100
63 35.6 28.4 42.6
89 71.3 58.0 87.0
182 2.7 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#58
Fluorene
Concen: 36.130 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:166 Resp: 1965747
Ion Ratio Lower Upper
166 100
165 96.4 77.0 115.4
167 13.3 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 37.508 ng
RT: 15.039 min Scan# 2049
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525




Tgt Ion:232 Resp: 557824
Ion Ratio Lower Upper
232 100
131 44.8 32.9 49.3
130 2.4 1.8 2.8
166 30.8 20.8 31.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025




#60
Diethylphthalate
Concen: 36.717 ng
RT: 15.245 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Tgt Ion:149 Resp: 2031289
Ion Ratio Lower Upper
149 100
177 22.1 17.2 25.8
150 12.4 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 36.132 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:204 Resp: 950092
Ion Ratio Lower Upper
204 100
206 32.6 25.4 38.0
141 60.1 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#62
4-Nitroaniline
Concen: 38.283 ng
RT: 15.469 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:138 Resp: 449658
Ion Ratio Lower Upper
138 100
92 49.8 28.8 68.8
108 77.3 58.1 98.1





#63
Azobenzene
Concen: 36.142 ng
RT: 15.751 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion: 77 Resp: 199848
Ion Ratio Lower Upper
77 100
182 25.2 6.8 46.8
105 21.7 1.7 41.7
51 16.7 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.151 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:188 Resp: 1474581
Ion Ratio Lower Upper
188 100
94 10.0 8.1 12.1
80 11.1 8.6 13.0









#65
4,6-Dinitro-2-methylphenol
Concen: 38.808 ng
RT: 15.533 min Scan# 2133
Delta R.T. -0.001 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Instrument :
BNA_M
ClientSampleId :
ICVBM060525

Tgt Ion:198 Resp: 31255
Ion Ratio Lower Upper
198 100
51 23.8 4.5 44.5
105 44.1 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#66
n-Nitrosodiphenylamine
Concen: 37.727 ng
RT: 15.669 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:169 Resp: 1747981
Ion Ratio Lower Upper
169 100
168 66.9 53.1 79.7
167 36.2 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 38.453 ng
RT: 16.351 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:248 Resp: 604660
Ion Ratio Lower Upper
248 100
250 96.6 78.0 117.0
141 73.9 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#68
Hexachlorobenzene
Concen: 37.911 ng
RT: 16.463 min Scan# 2291
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:284 Resp: 696502
Ion Ratio Lower Upper
284 100
142 35.8 28.8 43.2
249 32.1 25.8 38.6





#69
Atrazine
Concen: 39.186 ng
RT: 16.621 min Scan# 2318
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:200 Resp: 57849
Ion Ratio Lower Upper
200 100
173 26.7 6.0 46.0
215 48.0 28.1 68.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#70
Pentachlorophenol
Concen: 38.925 ng
RT: 16.804 min Scan# 2349
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:266 Resp: 464946
Ion Ratio Lower Upper
266 100
268 63.5 51.0 76.6
264 62.5 50.8 76.2





#71
Phenanthrene
Concen: 36.170 ng
RT: 17.198 min Scan# 2416
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:178 Resp: 304686
Ion Ratio Lower Upper
178 100
176 18.9 15.6 23.4
179 15.2 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#72
Anthracene
Concen: 36.806 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:178 Resp: 3101683
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.4 12.3 18.5





#73
Carbazole
Concen: 36.195 ng
RT: 17.551 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:167 Resp: 2778780
Ion Ratio Lower Upper
167 100
166 20.9 17.0 25.6
139 12.6 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#74
Di-n-butylphthalate
Concen: 37.377 ng
RT: 18.133 min Scan# 2575
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:149 Resp: 3122305
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.0
104 5.1 4.0 6.0





#75
Fluoranthene
Concen: 35.092 ng
RT: 19.209 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:202 Resp: 311738
Ion Ratio Lower Upper
202 100
101 10.9 0.0 31.1
203 17.2 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:240 Resp: 1199833
Ion Ratio Lower Upper
240 100
120 11.0 9.0 13.4
236 25.2 20.7 31.1





#77
Benzidine
Concen: 48.798 ng
RT: 19.392 min Scan# 2789
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:184 Resp: 1646454
Ion Ratio Lower Upper
184 100
185 14.6 11.5 17.3
183 11.2 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#78
Pyrene
Concen: 40.348 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:202 Resp: 3262895
Ion Ratio Lower Upper
202 100
200 20.7 16.7 25.1
203 17.3 14.1 21.1





#79
Terphenyl-d14
Concen: 68.934 ng
RT: 19.780 min Scan# 2855
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:244 Resp: 4364574
Ion Ratio Lower Upper
244 100
212 7.3 6.0 9.0
122 11.0 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#80
Butylbenzylphthalate
Concen: 40.599 ng
RT: 20.480 min Scan# 2974
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:149 Resp: 1095153
Ion Ratio Lower Upper
149 100
91 72.7 56.9 85.3
206 19.8 16.1 24.1





#81
Benzo(a)anthracene
Concen: 37.193 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:228 Resp: 282462
Ion Ratio Lower Upper
228 100
226 26.8 21.7 32.5
229 19.8 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#82
3,3'-Dichlorobenzidine
Concen: 40.289 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:252 Resp: 933068
Ion Ratio Lower Upper
252 100
254 64.3 51.4 77.0
126 11.5 8.6 12.8





#83

Chrysene

Concen: 36.579 ng

RT: 21.433 min Scan# 3135

Delta R.T. 0.006 min

Lab File: BM050202.D


Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525



Tgt Ion:228 Resp: 2642484

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5




229 19.2 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 38.675 ng

RT: 21.321 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion:149 Resp: 1606973


Ion Ratio Lower Upper

149 100

167 27.6 22.3 33.5


279 4.4 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 38.642 ng
RT: 22.462 min Scan# 3311
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:149 Resp: 238251
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.8 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.374 min Scan# 3636
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36


Tgt Ion:264 Resp: 1188981
Ion Ratio Lower Upper
264 100
260 22.9 18.6 28.0
265 21.8 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 38.807 ng
RT: 27.750 min Scan# 4211
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:276 Resp: 297090
Ion Ratio Lower Upper
276 100
138 25.7 20.6 30.8
277 24.9 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#88
Benzo(b)fluoranthene
Concen: 37.732 ng
RT: 23.433 min Scan# 3476
Delta R.T. -0.000 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:252 Resp: 2608359
Ion Ratio Lower Upper
252 100
253 21.6 17.2 25.8
125 10.3 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 36.362 ng

RT: 23.497 min Scan# 34

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36

Instrument :

BNA_M

ClientSampleId :

ICVBM060525

Tgt Ion:252 Resp: 2566064

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2





125 10.2 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/05/2025

Supervised By :Jagrut Upadhyay 06/06/2025



#90

Benzo(a)pyrene

Concen: 37.840 ng

RT: 24.239 min Scan# 3613

Delta R.T. -0.000 min

Lab File: BM050202.D

Acq: 05 Jun 2025 14:36




Tgt Ion:252 Resp: 2459411


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4


125 10.6 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 38.821 ng
RT: 27.815 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36




Instrument :
BNA_M
ClientSampleId :
ICVBM060525



Tgt Ion:278 Resp: 241239
Ion Ratio Lower Upper
278 100
139 15.9 12.8 19.2
279 23.6 18.6 28.0




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/05/2025
Supervised By :Jagrut Upadhyay 06/06/2025



#92
Benzo(g,h,i)perylene
Concen: 38.771 ng
RT: 28.791 min Scan# 4387
Delta R.T. -0.006 min
Lab File: BM050202.D
Acq: 05 Jun 2025 14:36

Tgt Ion:276 Resp: 2411378
Ion Ratio Lower Upper
276 100
277 23.5 18.6 27.8
138 21.9 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.526	0.448	14.8	96	0.00
3	Pyridine	1.361	1.266	7.0	100	0.00
4	n-Nitrosodimethylamine	0.281	0.269	4.3	102	0.00
5 S	2-Fluorophenol	1.199	1.069	10.8	96	0.00
6	Aniline	2.086	1.992	4.5	100	0.00
7 S	Phenol-d6	1.578	1.433	9.2	95	0.00
8	2-Chlorophenol	1.319	1.273	3.5	102	0.00
9	Benzaldehyde	1.003	1.038	-3.5	122	0.00
10 C	Phenol	1.670	1.599	4.3	100	0.00
11	bis(2-Chloroethyl)ether	1.343	1.284	4.4	101	0.00
12	1,3-Dichlorobenzene	1.474	1.395	5.4	102	0.00
13 C	1,4-Dichlorobenzene	1.534	1.434	6.5	102	0.00
14	1,2-Dichlorobenzene	1.462	1.382	5.5	102	0.00
15	Benzyl Alcohol	1.126	1.101	2.2	100	0.00
16	2,2'-oxybis(1-Chloropropane	0.933	0.873	6.4	99	0.00
17	2-Methylphenol	1.102	1.076	2.4	100	0.00
18	Hexachloroethane	0.556	0.528	5.0	102	0.00
19 P	n-Nitroso-di-n-propylamine	0.966	0.961	0.5	99	0.00
20	3+4-Methylphenols	1.474	1.451	1.6	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.500	0.470	6.0	101	0.00
23 S	Nitrobenzene-d5	0.385	0.317	17.7	86	0.00
24	Nitrobenzene	0.350	0.333	4.9	100	0.00
25	Isophorone	0.664	0.639	3.8	99	0.00
26 C	2-Nitrophenol	0.151	0.154	-2.0	102	0.00
27	2,4-Dimethylphenol	0.310	0.294	5.2	99	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.413	4.8	99	0.00
29 C	2,4-Dichlorophenol	0.287	0.280	2.4	100	0.00
30	1,2,4-Trichlorobenzene	0.319	0.300	6.0	100	0.00
31	Naphthalene	1.034	0.947	8.4	99	0.00
32	Benzoic acid	0.195	0.193	1.0	97	-0.04
33	4-Chloroaniline	0.441	0.419	5.0	99	0.00
34 C	Hexachlorobutadiene	0.190	0.178	6.3	100	0.00
35	Caprolactam	0.091	0.088	3.3	93	-0.02
36 C	4-Chloro-3-methylphenol	0.306	0.295	3.6	96	0.00
37	2-Methylnaphthalene	0.624	0.592	5.1	98	0.00
38	1-Methylnaphthalene	0.662	0.624	5.7	98	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.578	0.546	5.5	102	0.00
41 P	Hexachlorocyclopentadiene	0.351	0.345	1.7	102	0.00
42 S	2,4,6-Tribromophenol	0.227	0.203	10.6	89	0.00
43 C	2,4,6-Trichlorophenol	0.380	0.363	4.5	98	0.00
44	2,4,5-Trichlorophenol	0.417	0.399	4.3	97	0.00
45 S	2-Fluorobiphenyl	1.477	1.164	21.2	85	0.00
46	1,1'-Biphenyl	1.498	1.376	8.1	99	0.00
47	2-Chloronaphthalene	1.164	1.056	9.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.256	0.258	-0.8	97	0.00
49	Acenaphthylene	1.883	1.736	7.8	96	0.00
50	Dimethylphthalate	1.353	1.264	6.6	95	0.00
51	2,6-Dinitrotoluene	0.273	0.271	0.7	94	0.00
52 C	Acenaphthene	1.178	1.073	8.9	95	0.00
53	3-Nitroaniline	0.304	0.299	1.6	93	0.00
54 P	2,4-Dinitrophenol	0.146	0.139	4.8	92	0.00
55	Dibenzofuran	1.723	1.568	9.0	95	0.00
56 P	4-Nitrophenol	0.259	0.250	3.5	92	0.00
57	2,4-Dinitrotoluene	0.360	0.359	0.3	91	0.00
58	Fluorene	1.320	1.192	9.7	95	0.00
59	2,3,4,6-Tetrachlorophenol	0.361	0.338	6.4	93	0.00
60	Diethylphthalate	1.342	1.232	8.2	92	0.00
61	4-Chlorophenyl-phenylether	0.638	0.576	9.7	95	0.00
62	4-Nitroaniline	0.285	0.273	4.2	89	-0.01
63	Azobenzene	1.341	1.212	9.6	93	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.109	0.106	2.8	88	0.00
66 c	n-Nitrosodiphenylamine	0.628	0.593	5.6	93	0.00
67	4-Bromophenyl-phenylether	0.213	0.205	3.8	93	0.00
68	Hexachlorobenzene	0.249	0.236	5.2	92	0.00
69	Atrazine	0.200	0.196	2.0	88	0.00
70 C	Pentachlorophenol	0.162	0.158	2.5	89	0.00
71	Phenanthrene	1.143	1.033	9.6	89	0.00
72	Anthracene	1.143	1.052	8.0	90	0.00
73	Carbazole	1.041	0.942	9.5	87	0.00
74	Di-n-butylphthalate	1.133	1.059	6.5	85	0.00
75 C	Fluoranthene	1.205	1.057	12.3	83	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
77	Benzidine	0.562	0.686	-22.1	90	0.00
78	Pyrene	1.348	1.360	-0.9	83	0.00
79 S	Terphenyl-d14	1.055	0.909	13.8	73	0.00
80	Butylbenzylphthalate	0.450	0.456	-1.3	75	0.00
81	Benzo(a)anthracene	1.266	1.177	7.0	78	0.00
82	3,3'-Dichlorobenzidine	0.386	0.389	-0.8	76	0.00
83	Chrysene	1.204	1.101	8.6	77	0.00
84	Bis(2-ethylhexyl)phthalate	0.693	0.670	3.3	73	0.00
85 c	Di-n-octyl phthalate	1.028	0.993	3.4	72	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	1.288	1.249	3.0	88	-0.01
88	Benzo(b)fluoranthene	1.163	1.097	5.7	79	0.00
89	Benzo(k)fluoranthene	1.187	1.079	9.1	78	0.00
90 C	Benzo(a)pyrene	1.093	1.034	5.4	81	0.00
91	Dibenzo(a,h)anthracene	1.045	1.014	3.0	88	0.00
92	Benzo(g,h,i)perylene	1.046	1.014	3.1	90	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050202.D
Acq On : 05 Jun 2025 14:36
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM060525

Quant Time: Jun 05 15:10:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	34.075	14.8	96	0.00
3	Pyridine	40.000	37.216	7.0	100	0.00
4	n-Nitrosodimethylamine	40.000	38.276	4.3	102	0.00
5 S	2-Fluorophenol	80.000	71.321	10.8	96	0.00
6	Aniline	40.000	38.184	4.5	100	0.00
7 S	Phenol-d6	80.000	72.656	9.2	95	0.00
8	2-Chlorophenol	40.000	38.620	3.5	102	0.00
9	Benzaldehyde	40.000	41.383	-3.5	122	0.00
10 C	Phenol	40.000	38.313	4.2	100	0.00
11	bis(2-Chloroethyl)ether	40.000	38.237	4.4	101	0.00
12	1,3-Dichlorobenzene	40.000	37.856	5.4	102	0.00
13 C	1,4-Dichlorobenzene	40.000	37.381	6.5	102	0.00
14	1,2-Dichlorobenzene	40.000	37.789	5.5	102	0.00
15	Benzyl Alcohol	40.000	39.129	2.2	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.404	6.5	99	0.00
17	2-Methylphenol	40.000	39.079	2.3	100	0.00
18	Hexachloroethane	40.000	38.024	4.9	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.802	0.5	99	0.00
20	3+4-Methylphenols	40.000	39.372	1.6	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	37.622	5.9	101	0.00
23 S	Nitrobenzene-d5	80.000	65.977	17.5	86	0.00
24	Nitrobenzene	40.000	38.131	4.7	100	0.00
25	Isophorone	40.000	38.498	3.8	99	0.00
26 C	2-Nitrophenol	40.000	40.867	-2.2	102	0.00
27	2,4-Dimethylphenol	40.000	37.886	5.3	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.061	4.8	99	0.00
29 C	2,4-Dichlorophenol	40.000	38.988	2.5	100	0.00
30	1,2,4-Trichlorobenzene	40.000	37.573	6.1	100	0.00
31	Naphthalene	40.000	36.633	8.4	99	0.00
32	Benzoic acid	40.000	39.598	1.0	97	-0.04
33	4-Chloroaniline	40.000	38.018	5.0	99	0.00
34 C	Hexachlorobutadiene	40.000	37.511	6.2	100	0.00
35	Caprolactam	40.000	38.807	3.0	93	-0.02
36 C	4-Chloro-3-methylphenol	40.000	38.568	3.6	96	0.00
37	2-Methylnaphthalene	40.000	37.986	5.0	98	0.00
38	1-Methylnaphthalene	40.000	37.722	5.7	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.845	5.4	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.386	1.5	102	0.00
42 S	2,4,6-Tribromophenol	80.000	71.554	10.6	89	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.255	4.4	98	0.00
44	2,4,5-Trichlorophenol	40.000	38.297	4.3	97	0.00
45 S	2-Fluorobiphenyl	80.000	63.026	21.2	85	0.00
46	1,1'-Biphenyl	40.000	36.765	8.1	99	0.00
47	2-Chloronaphthalene	40.000	36.269	9.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
 Data File : BM050202.D
 Acq On : 05 Jun 2025 14:36
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060525

Quant Time: Jun 05 15:10:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 05 14:55:18 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.253	-0.6	97	0.00
49	Acenaphthylene	40.000	36.877	7.8	96	0.00
50	Dimethylphthalate	40.000	37.379	6.6	95	0.00
51	2,6-Dinitrotoluene	40.000	39.718	0.7	94	0.00
52 C	Acenaphthene	40.000	36.444	8.9	95	0.00
53	3-Nitroaniline	40.000	39.347	1.6	93	0.00
54 P	2,4-Dinitrophenol	40.000	34.596	13.5	92	0.00
55	Dibenzofuran	40.000	36.397	9.0	95	0.00
56 P	4-Nitrophenol	40.000	38.588	3.5	92	0.00
57	2,4-Dinitrotoluene	40.000	39.946	0.1	91	0.00
58	Fluorene	40.000	36.130	9.7	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.508	6.2	93	0.00
60	Diethylphthalate	40.000	36.717	8.2	92	0.00
61	4-Chlorophenyl-phenylether	40.000	36.132	9.7	95	0.00
62	4-Nitroaniline	40.000	38.283	4.3	89	-0.01
63	Azobenzene	40.000	36.142	9.6	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.808	3.0	88	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.727	5.7	93	0.00
67	4-Bromophenyl-phenylether	40.000	38.453	3.9	93	0.00
68	Hexachlorobenzene	40.000	37.911	5.2	92	0.00
69	Atrazine	40.000	39.186	2.0	88	0.00
70 C	Pentachlorophenol	40.000	38.925	2.7	89	0.00
71	Phenanthrene	40.000	36.170	9.6	89	0.00
72	Anthracene	40.000	36.806	8.0	90	0.00
73	Carbazole	40.000	36.195	9.5	87	0.00
74	Di-n-butylphthalate	40.000	37.377	6.6	85	0.00
75 C	Fluoranthene	40.000	35.092	12.3	83	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	81	0.00
77	Benzydine	40.000	48.798	-22.0	90	0.00
78	Pyrene	40.000	40.348	-0.9	83	0.00
79 S	Terphenyl-d14	80.000	68.934	13.8	73	0.00
80	Butylbenzylphthalate	40.000	40.599	-1.5	75	0.00
81	Benzo(a)anthracene	40.000	37.193	7.0	78	0.00
82	3,3'-Dichlorobenzidine	40.000	40.289	-0.7	76	0.00
83	Chrysene	40.000	36.579	8.6	77	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.675	3.3	73	0.00
85 c	Di-n-octyl phthalate	40.000	38.642	3.4	72	0.00
86 I	Perylene-d12	20.000	20.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.807	3.0	88	-0.01
88	Benzo(b)fluoranthene	40.000	37.732	5.7	79	0.00
89	Benzo(k)fluoranthene	40.000	36.362	9.1	78	0.00
90 C	Benzo(a)pyrene	40.000	37.840	5.4	81	0.00
91	Dibenzo(a,h)anthracene	40.000	38.821	2.9	88	0.00
92	Benzo(g,h,i)perylene	40.000	38.771	3.1	90	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050202.D
Acq On : 05 Jun 2025 14:36
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM060525

Quant Time: Jun 05 15:10:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 14:55:18 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG No.: Q2236

Instrument ID: BNA_M Calibration Date/Time: 06/09/2025 17:03

Lab File ID: BM050245.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.361	1.396		2.6	
2-Fluorophenol	1.199	1.226		2.3	
Phenol-d6	1.578	1.604		1.6	
1,4-Dichlorobenzene	1.534	1.510		-1.6	20.0
2-Methylphenol	1.102	1.111		0.8	
3+4-Methylphenols	1.474	1.493		1.3	
Nitrobenzene-d5	0.385	0.402		4.4	
Hexachloroethane	0.555	0.568		2.3	
Nitrobenzene	0.350	0.361		3.1	
Hexachlorobutadiene	0.190	0.185		-2.6	20.0
2,4,6-Trichlorophenol	0.380	0.384		1.1	20.0
2-Fluorobiphenyl	1.477	1.434		-2.9	
2,4,5-Trichlorophenol	0.417	0.422		1.2	
2,4-Dinitrotoluene	0.360	0.399		10.8	
2,4,6-Tribromophenol	0.227	0.241		6.2	
Hexachlorobenzene	0.249	0.245		-1.6	
Pentachlorophenol	0.162	0.173		6.8	20.0
Terphenyl-d14	1.055	0.994		-5.8	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 01:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	360068	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	1478208	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	892439	20.000	ng	0.00
64) Phenanthrene-d10	17.139	188	1715618	20.000	ng	0.00
76) Chrysene-d12	21.374	240	1793890	20.000	ng	0.00
86) Perylene-d12	24.350	264	1989460	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	1766146	81.823	ng	0.00
7) Phenol-d6	6.934	99	2310860	81.327	ng	0.00
23) Nitrobenzene-d5	8.916	82	2377509	83.649	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	861064	84.827	ng	0.00
45) 2-Fluorobiphenyl	13.027	172	5120406	77.678	ng	0.00
79) Terphenyl-d14	19.768	244	7133435	75.356	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	378420	39.996	ng	97
3) Pyridine	3.669	79	1005551	41.039	ng	98
4) n-Nitrosodimethylamine	3.581	42	211301	41.704	ng	96
6) Aniline	7.092	93	1526241	40.631	ng	99
8) 2-Chlorophenol	7.334	128	968901	40.809	ng	99
9) Benzaldehyde	6.904	77	670731	37.127	ng	98
10) Phenol	6.957	94	1221466	40.628	ng	99
11) bis(2-Chloroethyl)ether	7.192	93	982078	40.612	ng	99
12) 1,3-Dichlorobenzene	7.657	146	1055731	39.775	ng	99
13) 1,4-Dichlorobenzene	7.804	146	1087164	39.366	ng	99
14) 1,2-Dichlorobenzene	8.122	146	1040935	39.539	ng	98
15) Benzyl Alcohol	8.004	79	830354	40.972	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.298	45	718218	42.755	ng	97
17) 2-Methylphenol	8.204	107	799802	40.319	ng	99
18) Hexachloroethane	8.851	117	409286	40.924	ng	99
19) n-Nitroso-di-n-propyla...	8.575	70	725113	41.701	ng	99
20) 3+4-Methylphenols	8.534	107	1075257	40.515	ng	98
22) Acetophenone	8.586	105	1461650	39.579	ng	# 99
24) Nitrobenzene	8.957	77	1065825	41.232	ng	98
25) Isophorone	9.486	82	2001054	40.790	ng	99
26) 2-Nitrophenol	9.669	139	482950	43.282	ng	99
27) 2,4-Dimethylphenol	9.733	122	922228	40.225	ng	100
28) bis(2-Chloroethoxy)met...	9.969	93	1315107	41.036	ng	99
29) 2,4-Dichlorophenol	10.198	162	865109	40.776	ng	98
30) 1,2,4-Trichlorobenzene	10.422	180	929903	39.387	ng	100
31) Naphthalene	10.604	128	2985277	39.061	ng	100
32) Benzoic acid	9.851	122	615676	42.730	ng	97
33) 4-Chloroaniline	10.704	127	1312714	40.288	ng	99
34) Hexachlorobutadiene	10.904	225	548050	39.026	ng	99
35) Caprolactam	11.486	113	291076	43.274	ng	97
36) 4-Chloro-3-methylphenol	11.833	107	927702	40.978	ng	99
37) 2-Methylnaphthalene	12.222	142	1847845	40.078	ng	98
38) 1-Methylnaphthalene	12.439	142	1954409	39.938	ng	99
40) 1,2,4,5-Tetrachloroben...	12.586	216	1006543	39.059	ng	100
41) Hexachlorocyclopentadiene	12.574	237	643926	41.145	ng	98
43) 2,4,6-Trichlorophenol	12.827	196	685864	40.472	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 01:24:52 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Jun 09 11:54:38 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.898	196	753509	40.497	ng	97
46) 1,1'-Biphenyl	13.233	154	2608348	39.033	ng	99
47) 2-Chloronaphthalene	13.274	162	2019609	38.875	ng	99
48) 2-Nitroaniline	13.469	65	529139	46.283	ng	95
49) Acenaphthylene	14.121	152	3355750	39.935	ng	100
50) Dimethylphthalate	13.863	163	2441625	40.454	ng	100
51) 2,6-Dinitrotoluene	13.968	165	531089	43.633	ng	98
52) Acenaphthene	14.463	154	2072586	39.434	ng	99
53) 3-Nitroaniline	14.292	138	602236	44.395	ng	99
54) 2,4-Dinitrophenol	14.498	184	284222	39.171	ng	96
55) Dibenzofuran	14.798	168	3045369	39.605	ng	99
56) 4-Nitrophenol	14.598	139	516019	44.693	ng	99
57) 2,4-Dinitrotoluene	14.751	165	711679	44.325	ng	96
58) Fluorene	15.445	166	2341722	39.762	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.021	232	662134m	41.131	ng	
60) Diethylphthalate	15.227	149	2473158	41.299	ng	100
61) 4-Chlorophenyl-phenyle...	15.445	204	1135846	39.906	ng	98
62) 4-Nitroaniline	15.457	138	592967	46.638	ng	99
63) Azobenzene	15.739	77	2429561	40.590	ng	99
65) 4,6-Dinitro-2-methylph...	15.515	198	396949	42.362	ng	98
66) n-Nitrosodiphenylamine	15.657	169	2105155	39.052	ng	100
67) 4-Bromophenyl-phenylether	16.333	248	717717	39.230	ng	98
68) Hexachlorobenzene	16.451	284	841548	39.370	ng	96
69) Atrazine	16.609	200	729517	42.473	ng	98
70) Pentachlorophenol	16.786	266	593145	42.681	ng	100
71) Phenanthrene	17.180	178	3789626	38.667	ng	99
72) Anthracene	17.268	178	3872739	39.499	ng	100
73) Carbazole	17.533	167	3630433	40.644	ng	100
74) Di-n-butylphthalate	18.115	149	4266878	43.903	ng	100
75) Fluoranthene	19.192	202	4285265	41.461	ng	99
77) Benzidine	19.374	184	2470920	48.982	ng	100
78) Pyrene	19.556	202	4533099	37.492	ng	100
80) Butylbenzylphthalate	20.468	149	1856362	46.028	ng	98
81) Benzo(a)anthracene	21.350	228	4491010	39.552	ng	100
82) 3,3'-Dichlorobenzidine	21.274	252	1643282	47.458	ng	99
83) Chrysene	21.415	228	4279795	39.625	ng	100
84) Bis(2-ethylhexyl)phtha...	21.297	149	2867346	46.156	ng	99
85) Di-n-octyl phthalate	22.433	149	4667340	50.631	ng	100
87) Indeno(1,2,3-cd)pyrene	27.709	276	5457981	42.608	ng	100
88) Benzo(b)fluoranthene	23.409	252	4576331	39.564	ng	99
89) Benzo(k)fluoranthene	23.474	252	4520328	38.282	ng	100
90) Benzo(a)pyrene	24.209	252	4395508	40.418	ng	99
91) Dibenzo(a,h)anthracene	27.773	278	4440353	42.705	ng	99
92) Benzo(g,h,i)perylene	28.756	276	4419492	42.468	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025


Quant Time: Jun 10 01:24:52 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Jun 09 11:54:38 2025

Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050245.D
Acq On : 09 Jun 2025 17:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025


Quant Time: Jun 10 01:24:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 36006
Ion Ratio Lower Upper
152 100
150 154.4 122.2 183.4
115 59.5 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#2
1,4-Dioxane
Concen: 39.996 ng
RT: 3.281 min Scan# 50
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion: 88 Resp: 378420
Ion Ratio Lower Upper
88 100
58 56.3 42.9 64.3
43 14.0 11.0 16.6





#3
Pyridine
Concen: 41.039 ng
RT: 3.669 min Scan# 116
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 100555
Ion Ratio Lower Upper
79 100
52 41.6 32.1 48.1
51 19.6 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#4
n-Nitrosodimethylamine
Concen: 41.704 ng
RT: 3.581 min Scan# 101
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion: 42 Resp: 211301
Ion Ratio Lower Upper
42 100
74 269.6 222.0 333.0
44 7.6 7.0 10.6





#5
2-Fluorophenol
Concen: 81.823 ng
RT: 5.357 min Scan# 405
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:112 Resp: 1766140
Ion Ratio Lower Upper
112 100
64 47.9 39.8 59.6
63 24.3 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#6
Aniline
Concen: 40.631 ng
RT: 7.092 min Scan# 698
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion: 93 Resp: 1526241
Ion Ratio Lower Upper
93 100
66 35.2 27.4 41.2
65 17.6 13.6 20.4





#7
Phenol-d6
Concen: 81.327 ng
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 99 Resp: 231086
Ion Ratio Lower Upper
99 100
42 8.7 6.9 10.3
71 32.0 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#8
2-Chlorophenol
Concen: 40.809 ng
RT: 7.334 min Scan# 739
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:128 Resp: 968901
Ion Ratio Lower Upper
128 100
130 32.9 12.6 52.6
64 40.4 19.8 59.8





#9
Benzaldehyde
Concen: 37.127 ng
RT: 6.904 min Scan# 60
Delta R.T. -0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 77 Resp: 67073
Ion Ratio Lower Upper
77 100
105 99.3 80.9 120.9
106 94.4 77.1 117.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#10
Phenol
Concen: 40.628 ng
RT: 6.957 min Scan# 675
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion: 94 Resp: 1221466
Ion Ratio Lower Upper
94 100
65 24.9 4.3 44.3
66 32.7 12.9 52.9





#11
bis(2-Chloroethyl)ether
Concen: 40.612 ng
RT: 7.192 min Scan# 718
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 93 Resp: 982078
Ion Ratio Lower Upper
93 100
63 64.0 43.9 83.9
95 31.5 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#12
1,3-Dichlorobenzene
Concen: 39.775 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03

Tgt Ion:146 Resp: 1055731
Ion Ratio Lower Upper
146 100
148 64.9 51.9 77.9
75 30.2 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 39.366 ng

RT: 7.804 min Scan# 81

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:146 Resp: 1087164

Ion Ratio Lower Upper

146 100

148 64.4 51.4 77.0





111 38.9 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#14

1,2-Dichlorobenzene

Concen: 39.539 ng

RT: 8.122 min Scan# 873

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03




Tgt Ion:146 Resp: 1040935


Ion Ratio Lower Upper

146 100

148 64.6 50.4 75.6


111 42.4 34.3 51.5





#15
Benzyl Alcohol
Concen: 40.972 ng
RT: 8.004 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 830354
Ion Ratio Lower Upper
79 100
108 81.5 64.0 96.0
77 63.8 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.755 ng
RT: 8.298 min Scan# 903
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion: 45 Resp: 718218
Ion Ratio Lower Upper
45 100
77 27.7 8.7 48.7
79 20.5 2.7 42.7





#17
2-Methylphenol
Concen: 40.319 ng
RT: 8.204 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:107 Resp: 799802
Ion Ratio Lower Upper
107 100
108 113.8 90.8 136.2
77 45.6 35.7 53.5
79 43.2 35.0 52.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#18
Hexachloroethane
Concen: 40.924 ng
RT: 8.851 min Scan# 997
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:117 Resp: 409286
Ion Ratio Lower Upper
117 100
119 97.7 77.0 115.6
201 93.2 74.6 111.8





#19
n-Nitroso-di-n-propylamine
Concen: 41.701 ng
RT: 8.575 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 70 Resp: 72511
Ion Ratio Lower Upper
70 100
42 27.5 22.6 33.8
101 10.0 8.3 12.5
130 21.2 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#20
3+4-Methylphenols
Concen: 40.515 ng
RT: 8.534 min Scan# 943
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03



Tgt Ion:107 Resp: 1075257
Ion Ratio Lower Upper
107 100
108 92.5 70.4 110.4
77 31.7 12.2 52.2
79 28.7 7.7 47.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.557 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:136 Resp: 1478208
Ion Ratio Lower Upper
136 100
137 10.9 8.6 13.0
54 4.8 3.8 5.8
68 5.9 4.9 7.3

Manual Integrations

APPROVED




Reviewed By :Rahul Chavli 06/10/2025


Supervised By :Jagrut Upadhyay 06/10/2025



#22
Acetophenone
Concen: 39.579 ng
RT: 8.586 min Scan# 952
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03

Tgt Ion:105 Resp: 1461650
Ion Ratio Lower Upper
105 100
71 4.3 5.0 7.4#
51 14.4 11.4 17.2
120 22.8 18.2 27.4





#23

Nitrobenzene-d5

Concen: 83.649 ng

RT: 8.916 min Scan# 1008

Delta R.T. -0.000 min

Lab File: BM050245.D


Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion: 82 Resp: 2377509

Ion Ratio Lower Upper

82 100

128 42.9 35.2 52.8




54 31.0 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#24

Nitrobenzene

Concen: 41.232 ng

RT: 8.957 min Scan# 1015

Delta R.T. -0.006 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03




Tgt Ion: 77 Resp: 1065825


Ion Ratio Lower Upper

77 100

123 45.9 38.2 57.2


65 12.1 9.5 14.3





#25
Isophorone
Concen: 40.790 ng
RT: 9.486 min Scan# 1105
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 2001054
Ion Ratio Lower Upper
82 100
95 6.7 5.7 8.5
138 17.5 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#26
2-Nitrophenol
Concen: 43.282 ng
RT: 9.669 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03

Tgt Ion:139 Resp: 482950
Ion Ratio Lower Upper
139 100
109 29.6 23.3 34.9
65 44.9 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 40.225 ng

RT: 9.733 min Scan# 1147

Delta R.T. -0.000 min

Lab File: BM050245.D


Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:122 Resp: 922228

Ion Ratio Lower Upper

122 100

107 113.7 91.0 136.6




121 57.0 45.8 68.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#28

bis(2-Chloroethoxy)methane


Concen: 41.036 ng

RT: 9.969 min Scan# 1187

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03





Tgt Ion: 93 Resp: 1315107


Ion Ratio Lower Upper

93 100

95 32.3 26.2 39.4

123 11.2 8.9 13.3





#29

2,4-Dichlorophenol

Concen: 40.776 ng

RT: 10.198 min Scan# 1226

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:162 Resp: 865109

Ion Ratio Lower Upper

162 100

164 63.7 45.0 85.0





98 37.5 16.3 56.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#30

1,2,4-Trichlorobenzene

Concen: 39.387 ng

RT: 10.422 min Scan# 1264

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03




Tgt Ion:180 Resp: 929903


Ion Ratio Lower Upper

180 100

182 95.6 76.6 114.8


145 29.7 24.2 36.4





#31
Naphthalene
Concen: 39.061 ng
RT: 10.604 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 298527
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 12.7 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#32
Benzoic acid
Concen: 42.730 ng
RT: 9.851 min Scan# 1167
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:122 Resp: 615676
Ion Ratio Lower Upper
122 100
105 125.9 103.4 143.4
77 94.0 70.3 110.3





#33
4-Chloroaniline
Concen: 40.288 ng
RT: 10.704 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:127 Resp: 1312714
Ion Ratio Lower Upper
127 100
129 32.1 25.4 38.0
65 27.2 21.0 31.4
92 19.1 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#34
Hexachlorobutadiene
Concen: 39.026 ng
RT: 10.904 min Scan# 1346
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:225 Resp: 548050
Ion Ratio Lower Upper
225 100
223 63.2 50.2 75.2
227 63.3 51.0 76.4





#35
Caprolactam
Concen: 43.274 ng
RT: 11.486 min Scan# 1447
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 291070
Ion Ratio Lower Upper
113 100
55 108.5 83.5 123.5
56 81.9 60.6 100.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#36
4-Chloro-3-methylphenol
Concen: 40.978 ng
RT: 11.833 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:107 Resp: 927702
Ion Ratio Lower Upper
107 100
144 26.2 21.0 31.6
142 80.2 64.8 97.2





#37
2-Methylnaphthalene
Concen: 40.078 ng
RT: 12.222 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:142 Resp: 1847849
Ion Ratio Lower Upper
142 100
141 87.0 71.5 107.3
115 34.1 27.8 41.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#38
1-Methylnaphthalene
Concen: 39.938 ng
RT: 12.439 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:142 Resp: 1954409
Ion Ratio Lower Upper
142 100
141 90.5 72.9 109.3
116 3.8 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:164 Resp: 892439
Ion Ratio Lower Upper
164 100
162 100.6 80.2 120.4
160 45.0 35.7 53.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.059 ng
RT: 12.586 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:216 Resp: 1006543
Ion Ratio Lower Upper
216 100
214 78.5 63.0 94.4
179 20.7 16.6 25.0
108 20.0 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 41.145 ng
RT: 12.574 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 237 Resp: 643920
Ion Ratio Lower Upper
237 100
235 63.2 44.6 84.6
272 12.4 0.0 33.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#42
2,4,6-Tribromophenol
Concen: 84.827 ng
RT: 15.880 min Scan# 2192
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion: 330 Resp: 861064
Ion Ratio Lower Upper
330 100
332 97.2 77.5 116.3
141 35.8 28.8 43.2





#43
2,4,6-Trichlorophenol
Concen: 40.472 ng
RT: 12.827 min Scan# 1673
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:196 Resp: 685864
Ion Ratio Lower Upper
196 100
198 96.0 76.6 115.0
200 32.1 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#44
2,4,5-Trichlorophenol
Concen: 40.497 ng
RT: 12.898 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:196 Resp: 753509
Ion Ratio Lower Upper
196 100
198 98.6 76.1 114.1
97 50.7 39.3 58.9
132 28.2 21.4 32.0





#45
2-Fluorobiphenyl
Concen: 77.678 ng
RT: 13.027 min Scan# 1710
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:172 Resp: 5120400
Ion Ratio Lower Upper
172 100
171 34.5 27.8 41.6
170 23.3 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#46
1,1'-Biphenyl
Concen: 39.033 ng
RT: 13.233 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:154 Resp: 2608348
Ion Ratio Lower Upper
154 100
153 39.3 19.2 59.2
76 12.7 0.0 32.0





#47
2-Chloronaphthalene
Concen: 38.875 ng
RT: 13.274 min Scan# 1749
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:162 Resp: 2019609
Ion Ratio Lower Upper
162 100
127 38.5 31.1 46.7
164 32.5 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#48
2-Nitroaniline
Concen: 46.283 ng
RT: 13.469 min Scan# 1782
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion: 65 Resp: 529139
Ion Ratio Lower Upper
65 100
92 78.5 64.2 96.4
138 118.9 101.0 151.4





#49
Acenaphthylene
Concen: 39.935 ng
RT: 14.121 min Scan# 1893
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 3355750
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#50
Dimethylphthalate
Concen: 40.454 ng
RT: 13.863 min Scan# 1849
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:163 Resp: 2441625
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 43.633 ng
RT: 13.968 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 531089
Ion Ratio Lower Upper
165 100
63 46.9 35.8 53.6
89 55.9 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#52
Acenaphthene
Concen: 39.434 ng
RT: 14.463 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:154 Resp: 2072586
Ion Ratio Lower Upper
154 100
153 112.5 88.7 133.1
152 53.1 42.6 63.8





#53
3-Nitroaniline
Concen: 44.395 ng
RT: 14.292 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 602230
Ion Ratio Lower Upper
138 100
108 10.8 8.8 13.2
92 117.6 92.9 139.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#54
2,4-Dinitrophenol
Concen: 39.171 ng
RT: 14.498 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:184 Resp: 284222
Ion Ratio Lower Upper
184 100
63 54.6 45.3 67.9
154 63.2 54.3 81.5





#55
Dibenzenofuran
Concen: 39.605 ng
RT: 14.798 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 3045369
Ion Ratio Lower Upper
168 100
139 37.9 29.9 44.9
169 13.4 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#56
4-Nitrophenol
Concen: 44.693 ng
RT: 14.598 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:139 Resp: 516019
Ion Ratio Lower Upper
139 100
109 71.1 52.6 92.6
65 84.3 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 44.325 ng
RT: 14.751 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 711679
Ion Ratio Lower Upper
165 100
63 38.8 28.4 42.6
89 75.4 58.0 87.0
182 2.6 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#58
Fluorene
Concen: 39.762 ng
RT: 15.445 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:166 Resp: 2341722
Ion Ratio Lower Upper
166 100
165 97.3 77.0 115.4
167 13.5 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 41.131 ng m
RT: 15.021 min Scan# 2046
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:232 Resp: 662134
Ion Ratio Lower Upper
232 100
131 41.5 32.9 49.3
130 2.4 1.8 2.8
166 25.9 20.8 31.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#60
Diethylphthalate
Concen: 41.299 ng
RT: 15.227 min Scan# 2081
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:149 Resp: 2473158
Ion Ratio Lower Upper
149 100
177 21.5 17.2 25.8
150 12.2 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 39.906 ng
RT: 15.445 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:204 Resp: 1135840
Ion Ratio Lower Upper
204 100
206 32.4 25.4 38.0
141 58.5 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#62
4-Nitroaniline
Concen: 46.638 ng
RT: 15.457 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:138 Resp: 592967
Ion Ratio Lower Upper
138 100
92 50.2 28.8 68.8
108 78.1 58.1 98.1





#63
Azobenzene
Concen: 40.590 ng
RT: 15.739 min Scan# 2171
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 242956
Ion Ratio Lower Upper
77 100
182 26.4 6.8 46.8
105 22.4 1.7 41.7
51 15.7 0.0 36.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2406
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:188 Resp: 1715618
Ion Ratio Lower Upper
188 100
94 9.8 8.1 12.1
80 10.5 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 42.362 ng
RT: 15.515 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:198 Resp: 396949
Ion Ratio Lower Upper
198 100
51 25.6 4.5 44.5
105 46.4 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#66
n-Nitrosodiphenylamine
Concen: 39.052 ng
RT: 15.657 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:169 Resp: 2105155
Ion Ratio Lower Upper
169 100
168 66.3 53.1 79.7
167 36.3 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 39.230 ng
RT: 16.333 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:248 Resp: 71771
Ion Ratio Lower Upper
248 100
250 98.7 78.0 117.0
141 77.6 60.1 90.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#68
Hexachlorobenzene
Concen: 39.370 ng
RT: 16.451 min Scan# 2289
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:284 Resp: 841548
Ion Ratio Lower Upper
284 100
142 33.1 28.8 43.2
249 30.5 25.8 38.6





#69
Atrazine
Concen: 42.473 ng
RT: 16.609 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:200 Resp: 72951
Ion Ratio Lower Upper
200 100
173 26.5 6.0 46.0
215 49.3 28.1 68.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#70
Pentachlorophenol
Concen: 42.681 ng
RT: 16.786 min Scan# 2346
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:266 Resp: 593145
Ion Ratio Lower Upper
266 100
268 63.8 51.0 76.6
264 63.0 50.8 76.2





#71
Phenanthrene
Concen: 38.667 ng
RT: 17.180 min Scan# 2416
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:178 Resp: 3789620
Ion Ratio Lower Upper
178 100
176 19.2 15.6 23.4
179 15.1 12.3 18.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#72
Anthracene
Concen: 39.499 ng
RT: 17.268 min Scan# 2428
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Tgt Ion:178 Resp: 3872739
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.4 12.3 18.5





#73
Carbazole
Concen: 40.644 ng
RT: 17.533 min Scan# 2476
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:167 Resp: 363043
Ion Ratio Lower Upper
167 100
166 21.2 17.0 25.6
139 13.0 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#74
Di-n-butylphthalate
Concen: 43.903 ng
RT: 18.115 min Scan# 2572
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:149 Resp: 4266878
Ion Ratio Lower Upper
149 100
150 9.0 7.4 11.0
104 5.1 4.0 6.0





#75
Fluoranthene
Concen: 41.461 ng
RT: 19.192 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 428526
Ion Ratio Lower Upper
202 100
101 11.7 0.0 31.1
203 16.9 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.374 min Scan# 3126
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:240 Resp: 1793890
Ion Ratio Lower Upper
240 100
120 11.0 9.0 13.4
236 25.3 20.7 31.1





#77
Benzidine
Concen: 48.982 ng
RT: 19.374 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 2470920
Ion Ratio Lower Upper
184 100
185 14.5 11.5 17.3
183 11.2 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#78
Pyrene
Concen: 37.492 ng
RT: 19.556 min Scan# 2817
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:202 Resp: 4533099
Ion Ratio Lower Upper
202 100
200 20.8 16.7 25.1
203 17.3 14.1 21.1





#79
Terphenyl-d14
Concen: 75.356 ng
RT: 19.768 min Scan# 2853
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:244 Resp: 713343
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 10.5 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#80
Butylbenzylphthalate
Concen: 46.028 ng
RT: 20.468 min Scan# 2972
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:149 Resp: 1856362
Ion Ratio Lower Upper
149 100
91 69.7 56.9 85.3
206 20.8 16.1 24.1





#81
Benzo(a)anthracene
Concen: 39.552 ng
RT: 21.350 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 4491010
Ion Ratio Lower Upper
228 100
226 27.2 21.7 32.5
229 20.0 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#82
3,3'-Dichlorobenzidine
Concen: 47.458 ng
RT: 21.274 min Scan# 3109
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:252 Resp: 1643282
Ion Ratio Lower Upper
252 100
254 64.5 51.4 77.0
126 11.8 8.6 12.8





#83
Chrysene
Concen: 39.625 ng
RT: 21.415 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 427979
Ion Ratio Lower Upper
228 100
226 30.6 24.3 36.5
229 19.6 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 46.156 ng
RT: 21.297 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:149 Resp: 2867346
Ion Ratio Lower Upper
149 100
167 27.3 22.3 33.5
279 4.0 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 50.631 ng
RT: 22.433 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 4667340
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.9 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.350 min Scan# 3632
Delta R.T. 0.006 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03


Tgt Ion:264 Resp: 1989460
Ion Ratio Lower Upper
264 100
260 23.7 18.6 28.0
265 21.8 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 42.608 ng
RT: 27.709 min Scan# 41
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:276 Resp: 545798
Ion Ratio Lower Upper
276 100
138 25.8 20.6 30.8
277 24.9 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#88
Benzo(b)fluoranthene
Concen: 39.564 ng
RT: 23.409 min Scan# 3472
Delta R.T. -0.000 min
Lab File: BM050245.D
Acq: 09 Jun 2025 17:03

Tgt Ion:252 Resp: 4576331
Ion Ratio Lower Upper
252 100
253 22.1 17.2 25.8
125 10.3 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 38.282 ng

RT: 23.474 min Scan# 34

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 4520328

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.2





125 10.2 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#90

Benzo(a)pyrene

Concen: 40.418 ng

RT: 24.209 min Scan# 3608

Delta R.T. -0.000 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03




Tgt Ion:252 Resp: 4395508


Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4

125 11.6 9.0 13.4





#91

Dibenzo(a,h)anthracene

Concen: 42.705 ng

RT: 27.773 min Scan# 41

Delta R.T. 0.006 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:278 Resp: 444035

Ion Ratio Lower Upper

278 100

139 16.6 12.8 19.2





279 23.9 18.6 28.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#92

Benzo(g,h,i)perylene

Concen: 42.468 ng

RT: 28.756 min Scan# 4381

Delta R.T. 0.006 min

Lab File: BM050245.D

Acq: 09 Jun 2025 17:03




Tgt Ion:276 Resp: 4419492

Ion Ratio Lower Upper

276 100

277 23.9 18.6 27.8

138 22.5 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 01:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.526	0.525	0.2	128	0.00
3	Pyridine	1.361	1.396	-2.6	124	0.00
4	n-Nitrosodimethylamine	0.281	0.293	-4.3	126	0.00
5 S	2-Fluorophenol	1.199	1.226	-2.3	125	0.00
6	Aniline	2.086	2.119	-1.6	120	0.00
7 S	Phenol-d6	1.578	1.604	-1.6	120	0.00
8	2-Chlorophenol	1.319	1.345	-2.0	121	0.00
9	Benzaldehyde	1.003	0.931	7.2	124	0.00
10 C	Phenol	1.670	1.696	-1.6	120	0.00
11	bis(2-Chloroethyl)ether	1.343	1.364	-1.6	121	0.00
12	1,3-Dichlorobenzene	1.474	1.466	0.5	121	0.00
13 C	1,4-Dichlorobenzene	1.534	1.510	1.6	121	0.00
14	1,2-Dichlorobenzene	1.462	1.445	1.2	120	0.00
15	Benzyl Alcohol	1.126	1.153	-2.4	118	0.00
16	2,2'-oxybis(1-Chloropropane	0.933	0.997	-6.9	128	0.00
17	2-Methylphenol	1.102	1.111	-0.8	117	0.00
18	Hexachloroethane	0.556	0.568	-2.2	124	0.00
19 P	n-Nitroso-di-n-propylamine	0.966	1.007	-4.2	117	0.00
20	3+4-Methylphenols	1.474	1.493	-1.3	115	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
22	Acetophenone	0.500	0.494	1.2	117	0.00
23 S	Nitrobenzene-d5	0.385	0.402	-4.4	120	0.00
24	Nitrobenzene	0.350	0.361	-3.1	119	0.00
25	Isophorone	0.664	0.677	-2.0	115	0.00
26 C	2-Nitrophenol	0.151	0.163	-7.9	118	0.00
27	2,4-Dimethylphenol	0.310	0.312	-0.6	116	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.445	-2.5	117	0.00
29 C	2,4-Dichlorophenol	0.287	0.293	-2.1	115	0.00
30	1,2,4-Trichlorobenzene	0.319	0.315	1.3	116	0.00
31	Naphthalene	1.034	1.010	2.3	116	0.00
32	Benzoic acid	0.195	0.208	-6.7	114	0.00
33	4-Chloroaniline	0.441	0.444	-0.7	115	0.00
34 C	Hexachlorobutadiene	0.190	0.185	2.6	114	0.00
35	Caprolactam	0.091	0.098	-7.7	114	0.00
36 C	4-Chloro-3-methylphenol	0.306	0.314	-2.6	112	0.00
37	2-Methylnaphthalene	0.624	0.625	-0.2	114	0.00
38	1-Methylnaphthalene	0.662	0.661	0.2	114	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.578	0.564	2.4	114	0.00
41 P	Hexachlorocyclopentadiene	0.351	0.361	-2.8	115	0.00
42 S	2,4,6-Tribromophenol	0.227	0.241	-6.2	114	0.00
43 C	2,4,6-Trichlorophenol	0.380	0.384	-1.1	112	0.00
44	2,4,5-Trichlorophenol	0.417	0.422	-1.2	111	0.00
45 S	2-Fluorobiphenyl	1.477	1.434	2.9	114	0.00
46	1,1'-Biphenyl	1.498	1.461	2.5	113	0.00
47	2-Chloronaphthalene	1.164	1.132	2.7	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 01:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.256	0.296	-15.6	120	0.00
49	Acenaphthylene	1.883	1.880	0.2	113	0.00
50	Dimethylphthalate	1.353	1.368	-1.1	111	0.00
51	2,6-Dinitrotoluene	0.273	0.298	-9.2	112	0.00
52 C	Acenaphthene	1.178	1.161	1.4	112	0.00
53	3-Nitroaniline	0.304	0.337	-10.9	114	0.00
54 P	2,4-Dinitrophenol	0.146	0.159	-8.9	114	0.00
55	Dibenzofuran	1.723	1.706	1.0	112	0.00
56 P	4-Nitrophenol	0.259	0.289	-11.6	115	0.00
57	2,4-Dinitrotoluene	0.360	0.399	-10.8	110	0.00
58	Fluorene	1.320	1.312	0.6	113	0.00
59	2,3,4,6-Tetrachlorophenol	0.361	0.371	-2.8	111	0.00
60	Diethylphthalate	1.342	1.386	-3.3	112	0.00
61	4-Chlorophenyl-phenylether	0.638	0.636	0.3	113	0.00
62	4-Nitroaniline	0.285	0.332	-16.5	118	0.00
63	Azobenzene	1.341	1.361	-1.5	113	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	0.109	0.116	-6.4	112	0.00
66 c	n-Nitrosodiphenylamine	0.628	0.614	2.2	111	0.00
67	4-Bromophenyl-phenylether	0.213	0.209	1.9	111	0.00
68	Hexachlorobenzene	0.249	0.245	1.6	111	0.00
69	Atrazine	0.200	0.213	-6.5	111	0.00
70 C	Pentachlorophenol	0.162	0.173	-6.8	114	0.00
71	Phenanthrene	1.143	1.104	3.4	111	0.00
72	Anthracene	1.143	1.129	1.2	112	0.00
73	Carbazole	1.041	1.058	-1.6	113	0.00
74	Di-n-butylphthalate	1.133	1.244	-9.8	116	0.00
75 C	Fluoranthene	1.205	1.249	-3.7	114	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
77	Benzidine	0.562	0.689	-22.6	135	0.00
78	Pyrene	1.348	1.263	6.3	115	0.00
79 S	Terphenyl-d14	1.055	0.994	5.8	119	0.00
80	Butylbenzylphthalate	0.450	0.517	-14.9	128	0.00
81	Benzo(a)anthracene	1.266	1.252	1.1	124	0.00
82	3,3'-Dichlorobenzidine	0.386	0.458	-18.7	133	0.00
83	Chrysene	1.204	1.193	0.9	125	0.00
84	Bis(2-ethylhexyl)phthalate	0.693	0.799	-15.3	130	0.00
85 c	Di-n-octyl phthalate	1.028	1.301	-26.6#	141	0.00
86 I	Perylene-d12	1.000	1.000	0.0	141	0.00
87	Indeno(1,2,3-cd)pyrene	1.288	1.372	-6.5	162#	0.00
88	Benzo(b)fluoranthene	1.163	1.150	1.1	139	0.00
89	Benzo(k)fluoranthene	1.187	1.136	4.3	137	0.00
90 C	Benzo(a)pyrene	1.093	1.105	-1.1	144	0.00
91	Dibenzo(a,h)anthracene	1.045	1.116	-6.8	162#	0.00
92	Benzo(g,h,i)perylene	1.046	1.111	-6.2	164#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050245.D
Acq On : 09 Jun 2025 17:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 01:24:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 01:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	117	0.00
2	1,4-Dioxane	40.000	39.996	0.0	128	0.00
3	Pyridine	40.000	41.039	-2.6	124	0.00
4	n-Nitrosodimethylamine	40.000	41.704	-4.3	126	0.00
5 S	2-Fluorophenol	80.000	81.823	-2.3	125	0.00
6	Aniline	40.000	40.631	-1.6	120	0.00
7 S	Phenol-d6	80.000	81.327	-1.7	120	0.00
8	2-Chlorophenol	40.000	40.809	-2.0	121	0.00
9	Benzaldehyde	40.000	37.127	7.2	124	0.00
10 C	Phenol	40.000	40.628	-1.6	120	0.00
11	bis(2-Chloroethyl)ether	40.000	40.612	-1.5	121	0.00
12	1,3-Dichlorobenzene	40.000	39.775	0.6	121	0.00
13 C	1,4-Dichlorobenzene	40.000	39.366	1.6	121	0.00
14	1,2-Dichlorobenzene	40.000	39.539	1.2	120	0.00
15	Benzyl Alcohol	40.000	40.972	-2.4	118	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.755	-6.9	128	0.00
17	2-Methylphenol	40.000	40.319	-0.8	117	0.00
18	Hexachloroethane	40.000	40.924	-2.3	124	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.701	-4.3	117	0.00
20	3+4-Methylphenols	40.000	40.515	-1.3	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	112	0.00
22	Acetophenone	40.000	39.579	1.1	117	0.00
23 S	Nitrobenzene-d5	80.000	83.649	-4.6	120	0.00
24	Nitrobenzene	40.000	41.232	-3.1	119	0.00
25	Isophorone	40.000	40.790	-2.0	115	0.00
26 C	2-Nitrophenol	40.000	43.282	-8.2	118	0.00
27	2,4-Dimethylphenol	40.000	40.225	-0.6	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.036	-2.6	117	0.00
29 C	2,4-Dichlorophenol	40.000	40.776	-1.9	115	0.00
30	1,2,4-Trichlorobenzene	40.000	39.387	1.5	116	0.00
31	Naphthalene	40.000	39.061	2.3	116	0.00
32	Benzoic acid	40.000	42.730	-6.8	114	0.00
33	4-Chloroaniline	40.000	40.288	-0.7	115	0.00
34 C	Hexachlorobutadiene	40.000	39.026	2.4	114	0.00
35	Caprolactam	40.000	43.274	-8.2	114	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.978	-2.4	112	0.00
37	2-Methylnaphthalene	40.000	40.078	-0.2	114	0.00
38	1-Methylnaphthalene	40.000	39.938	0.2	114	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.059	2.4	114	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.145	-2.9	115	0.00
42 S	2,4,6-Tribromophenol	80.000	84.827	-6.0	114	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.472	-1.2	112	0.00
44	2,4,5-Trichlorophenol	40.000	40.497	-1.2	111	0.00
45 S	2-Fluorobiphenyl	80.000	77.678	2.9	114	0.00
46	1,1'-Biphenyl	40.000	39.033	2.4	113	0.00
47	2-Chloronaphthalene	40.000	38.875	2.8	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050245.D
 Acq On : 09 Jun 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 01:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.283	-15.7	120	0.00
49	Acenaphthylene	40.000	39.935	0.2	113	0.00
50	Dimethylphthalate	40.000	40.454	-1.1	111	0.00
51	2,6-Dinitrotoluene	40.000	43.633	-9.1	112	0.00
52 C	Acenaphthene	40.000	39.434	1.4	112	0.00
53	3-Nitroaniline	40.000	44.395	-11.0	114	0.00
54 P	2,4-Dinitrophenol	40.000	39.171	2.1	114	0.00
55	Dibenzofuran	40.000	39.605	1.0	112	0.00
56 P	4-Nitrophenol	40.000	44.693	-11.7	115	0.00
57	2,4-Dinitrotoluene	40.000	44.325	-10.8	110	0.00
58	Fluorene	40.000	39.762	0.6	113	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.131	-2.8	111	0.00
60	Diethylphthalate	40.000	41.299	-3.2	112	0.00
61	4-Chlorophenyl-phenylether	40.000	39.906	0.2	113	0.00
62	4-Nitroaniline	40.000	46.638	-16.6	118	0.00
63	Azobenzene	40.000	40.590	-1.5	113	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.362	-5.9	112	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.052	2.4	111	0.00
67	4-Bromophenyl-phenylether	40.000	39.230	1.9	111	0.00
68	Hexachlorobenzene	40.000	39.370	1.6	111	0.00
69	Atrazine	40.000	42.473	-6.2	111	0.00
70 C	Pentachlorophenol	40.000	42.681	-6.7	114	0.00
71	Phenanthrene	40.000	38.667	3.3	111	0.00
72	Anthracene	40.000	39.499	1.3	112	0.00
73	Carbazole	40.000	40.644	-1.6	113	0.00
74	Di-n-butylphthalate	40.000	43.903	-9.8	116	0.00
75 C	Fluoranthene	40.000	41.461	-3.7	114	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	121	0.00
77	Benzydine	40.000	48.982	-22.5	135	0.00
78	Pyrene	40.000	37.492	6.3	115	0.00
79 S	Terphenyl-d14	80.000	75.356	5.8	119	0.00
80	Butylbenzylphthalate	40.000	46.028	-15.1	128	0.00
81	Benzo(a)anthracene	40.000	39.552	1.1	124	0.00
82	3,3'-Dichlorobenzidine	40.000	47.458	-18.6	133	0.00
83	Chrysene	40.000	39.625	0.9	125	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	46.156	-15.4	130	0.00
85 c	Di-n-octyl phthalate	40.000	50.631	-26.6#	141	0.00
86 I	Perylene-d12	20.000	20.000	0.0	141	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.608	-6.5	162	0.00
88	Benzo(b)fluoranthene	40.000	39.564	1.1	139	0.00
89	Benzo(k)fluoranthene	40.000	38.282	4.3	137	0.00
90 C	Benzo(a)pyrene	40.000	40.418	-1.0	144	0.00
91	Dibenzo(a,h)anthracene	40.000	42.705	-6.8	162	0.00
92	Benzo(g,h,i)perylene	40.000	42.468	-6.2	164	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050245.D
Acq On : 09 Jun 2025 17:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 01:24:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG No.: Q2236

Instrument ID: BNA_M Calibration Date/Time: 06/10/2025 04:48

Lab File ID: BM050262.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.361	1.448		6.4	
2-Fluorophenol	1.199	1.232		2.8	
Phenol-d6	1.578	1.600		1.4	
1,4-Dichlorobenzene	1.534	1.485		-3.2	20.0
2-Methylphenol	1.102	1.105		0.3	
3+4-Methylphenols	1.474	1.472		-0.1	
Nitrobenzene-d5	0.385	0.405		5.2	
Hexachloroethane	0.555	0.579		4.3	
Nitrobenzene	0.350	0.366		4.6	
Hexachlorobutadiene	0.190	0.181		-4.7	20.0
2,4,6-Trichlorophenol	0.380	0.384		1.1	20.0
2-Fluorobiphenyl	1.477	1.442		-2.4	
2,4,5-Trichlorophenol	0.417	0.418		0.2	
2,4-Dinitrotoluene	0.360	0.377		4.7	
2,4,6-Tribromophenol	0.227	0.228		0.4	
Hexachlorobenzene	0.249	0.248		-0.4	
Pentachlorophenol	0.162	0.224		38.3	20.0
Terphenyl-d14	1.055	1.019		-3.4	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 05:44:50 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Jun 09 11:54:38 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	460715	20.000	ng	0.00
21) Naphthalene-d8	10.551	136	1877125	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	1070949	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1957349	20.000	ng	0.00
76) Chrysene-d12	21.368	240	1968381	20.000	ng	0.00
86) Perylene-d12	24.338	264	2090330	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.351	112	2270046	82.193	ng	0.00
7) Phenol-d6	6.934	99	2948516	81.099	ng	0.00
23) Nitrobenzene-d5	8.916	82	3040809	84.250	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	977575	80.252	ng	0.00
45) 2-Fluorobiphenyl	13.021	172	6176210	78.077	ng	0.00
79) Terphenyl-d14	19.762	244	8024971	77.259	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.275	88	499652	41.272	ng	90
3) Pyridine	3.669	79	1334057	42.552	ng	93
4) n-Nitrosodimethylamine	3.581	42	281321	43.394	ng	99
6) Aniline	7.092	93	1927647	40.106	ng	98
8) 2-Chlorophenol	7.334	128	1232483	40.570	ng	97
9) Benzaldehyde	6.904	77	890670	38.531	ng	96
10) Phenol	6.957	94	1570209	40.818	ng	98
11) bis(2-Chloroethyl)ether	7.192	93	1265112	40.887	ng	96
12) 1,3-Dichlorobenzene	7.657	146	1325907	39.041	ng	98
13) 1,4-Dichlorobenzene	7.798	146	1367877	38.711	ng	98
14) 1,2-Dichlorobenzene	8.116	146	1317220	39.103	ng	99
15) Benzyl Alcohol	8.004	79	1053641	40.632	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.292	45	1029783	47.910	ng	93
17) 2-Methylphenol	8.204	107	1018406	40.123	ng	99
18) Hexachloroethane	8.845	117	533391	41.682	ng	96
19) n-Nitroso-di-n-propyla...	8.575	70	931068	41.848	ng	98
20) 3+4-Methylphenols	8.534	107	1356508	39.946	ng	99
22) Acetophenone	8.581	105	1869192	39.858	ng	98
24) Nitrobenzene	8.957	77	1374442	41.871	ng	97
25) Isophorone	9.486	82	2493995	40.035	ng	98
26) 2-Nitrophenol	9.669	139	584953	41.282	ng	98
27) 2,4-Dimethylphenol	9.733	122	1152793	39.596	ng	100
28) bis(2-Chloroethoxy)met...	9.969	93	1683958	41.378	ng	99
29) 2,4-Dichlorophenol	10.198	162	1065409	39.545	ng	98
30) 1,2,4-Trichlorobenzene	10.416	180	1169668	39.014	ng	99
31) Naphthalene	10.604	128	3777854	38.926	ng	100
32) Benzoic acid	9.851	122	565798	30.923	ng	97
33) 4-Chloroaniline	10.704	127	1625729	39.291	ng	98
34) Hexachlorobutadiene	10.898	225	680166	38.141	ng	99
35) Caprolactam	11.492	113	329763	38.607	ng	88
36) 4-Chloro-3-methylphenol	11.833	107	1135589	39.501	ng	99
37) 2-Methylnaphthalene	12.216	142	2270117	38.773	ng	99
38) 1-Methylnaphthalene	12.439	142	2417639	38.905	ng	100
40) 1,2,4,5-Tetrachloroben...	12.586	216	1221487	39.499	ng	99
41) Hexachlorocyclopentadiene	12.574	237	767530	40.868	ng	99
43) 2,4,6-Trichlorophenol	12.821	196	822241	40.432	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 05:44:50 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Jun 09 11:54:38 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.892	196	896098	40.133	ng	96
46) 1,1'-Biphenyl	13.233	154	3162089	39.432	ng	99
47) 2-Chloronaphthalene	13.269	162	2452782	39.343	ng	100
48) 2-Nitroaniline	13.468	65	634421	46.242	ng	92
49) Acenaphthylene	14.121	152	3990554	39.574	ng	100
50) Dimethylphthalate	13.863	163	2865663	39.566	ng	100
51) 2,6-Dinitrotoluene	13.968	165	614664	42.082	ng	94
52) Acenaphthene	14.463	154	2350065	37.260	ng	98
53) 3-Nitroaniline	14.292	138	690400	42.411	ng	99
54) 2,4-Dinitrophenol	14.492	184	124034	17.614	ng	91
55) Dibenzofuran	14.798	168	3611406	39.138	ng	100
56) 4-Nitrophenol	14.598	139	568795	41.052	ng	99
57) 2,4-Dinitrotoluene	14.751	165	808487	41.961	ng	92
58) Fluorene	15.445	166	2744828	38.838	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.021	232	758063m	39.241	ng	
60) Diethylphthalate	15.227	149	2879326	40.067	ng	100
61) 4-Chlorophenyl-phenyle...	15.445	204	1332679	39.017	ng	99
62) 4-Nitroaniline	15.457	138	649475	42.568	ng	97
63) Azobenzene	15.733	77	2958872	41.194	ng	97
65) 4,6-Dinitro-2-methylph...	15.515	198	264901	24.779	ng	95
66) n-Nitrosodiphenylamine	15.651	169	2442884	39.721	ng	100
67) 4-Bromophenyl-phenylether	16.333	248	828583	39.697	ng	98
68) Hexachlorobenzene	16.445	284	972175	39.865	ng	99
69) Atrazine	16.604	200	810041	41.337	ng	99
70) Pentachlorophenol	16.786	266	878713	55.421	ng	99
71) Phenanthrene	17.180	178	4374612	39.123	ng	99
72) Anthracene	17.268	178	4405600	39.385	ng	100
73) Carbazole	17.533	167	4064090	39.880	ng	100
74) Di-n-butylphthalate	18.115	149	4799970	43.289	ng	100
75) Fluoranthene	19.192	202	4786259	40.590	ng	100
77) Benzidine	19.374	184	2479851	44.801	ng	99
78) Pyrene	19.550	202	5069333	38.210	ng	100
80) Butylbenzylphthalate	20.462	149	1940495	43.849	ng	97
81) Benzo(a)anthracene	21.350	228	4905751	39.375	ng	100
82) 3,3'-Dichlorobenzidine	21.274	252	1678656	44.182	ng	99
83) Chrysene	21.409	228	4682061	39.507	ng	100
84) Bis(2-ethylhexyl)phtha...	21.292	149	3084575	45.251	ng	98
85) Di-n-octyl phthalate	22.427	149	4711085	46.575	ng	99
87) Indeno(1,2,3-cd)pyrene	27.697	276	5717703	42.482	ng	99
88) Benzo(b)fluoranthene	23.403	252	4894905	40.276	ng	99
89) Benzo(k)fluoranthene	23.468	252	4869113	39.246	ng	100
90) Benzo(a)pyrene	24.203	252	4660050	40.782	ng	99
91) Dibenzo(a,h)anthracene	27.762	278	4677689	42.817	ng	99
92) Benzo(g,h,i)perylene	28.738	276	4684760	42.844	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :


SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 05:44:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050262.D
Acq On : 10 Jun 2025 04:48
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 05:44:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.763 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 460715
Ion Ratio Lower Upper
152 100
150 153.5 122.2 183.4
115 61.2 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#2
1,4-Dioxane
Concen: 41.272 ng
RT: 3.275 min Scan# 49
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion: 88 Resp: 499652
Ion Ratio Lower Upper
88 100
58 61.9 42.9 64.3
43 15.1 11.0 16.6





#3
Pyridine
Concen: 42.552 ng
RT: 3.669 min Scan# 116
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 133405
Ion Ratio Lower Upper
79 100
52 45.4 32.1 48.1
51 21.3 15.4 23.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#4
n-Nitrosodimethylamine
Concen: 43.394 ng
RT: 3.581 min Scan# 101
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Tgt Ion: 42 Resp: 281321
Ion Ratio Lower Upper
42 100
74 275.6 222.0 333.0
44 9.2 7.0 10.6





#5
2-Fluorophenol
Concen: 82.193 ng
RT: 5.351 min Scan# 405
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 112 Resp: 2270040
Ion Ratio Lower Upper
112 100
64 53.5 39.8 59.6
63 27.2 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#6
Aniline
Concen: 40.106 ng
RT: 7.092 min Scan# 698
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion: 93 Resp: 1927647
Ion Ratio Lower Upper
93 100
66 35.7 27.4 41.2
65 17.8 13.6 20.4





#7
Phenol-d6
Concen: 81.099 ng
RT: 6.934 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 99 Resp: 2948510
Ion Ratio Lower Upper
99 100
42 8.8 6.9 10.3
71 32.5 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#8
2-Chlorophenol
Concen: 40.570 ng
RT: 7.334 min Scan# 739
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:128 Resp: 1232483
Ion Ratio Lower Upper
128 100
130 33.1 12.6 52.6
64 42.3 19.8 59.8





#9
Benzaldehyde
Concen: 38.531 ng
RT: 6.904 min Scan# 666
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 77 Resp: 890670
Ion Ratio Lower Upper
77 100
105 95.6 80.9 120.9
106 94.3 77.1 117.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#10
Phenol
Concen: 40.818 ng
RT: 6.957 min Scan# 675
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion: 94 Resp: 1570209
Ion Ratio Lower Upper
94 100
65 25.3 4.3 44.3
66 34.1 12.9 52.9





#11
bis(2-Chloroethyl)ether
Concen: 40.887 ng
RT: 7.192 min Scan# 715
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 93 Resp: 126511
Ion Ratio Lower Upper
93 100
63 68.4 43.9 83.9
95 32.1 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#12
1,3-Dichlorobenzene
Concen: 39.041 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:146 Resp: 1325907
Ion Ratio Lower Upper
146 100
148 65.3 51.9 77.9
75 32.1 23.4 35.0





#13
1,4-Dichlorobenzene
Concen: 38.711 ng
RT: 7.798 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:146 Resp: 136787
Ion Ratio Lower Upper
146 100
148 63.2 51.4 77.0
111 42.1 32.3 48.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#14
1,2-Dichlorobenzene
Concen: 39.103 ng
RT: 8.116 min Scan# 872
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Tgt Ion:146 Resp: 1317220
Ion Ratio Lower Upper
146 100
148 63.2 50.4 75.6
111 44.4 34.3 51.5





#15
Benzyl Alcohol
Concen: 40.632 ng
RT: 8.004 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 105364
Ion Ratio Lower Upper
79 100
108 79.0 64.0 96.0
77 64.1 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.910 ng
RT: 8.292 min Scan# 902
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion: 45 Resp: 1029783
Ion Ratio Lower Upper
45 100
77 25.4 8.7 48.7
79 19.1 2.7 42.7





#17
2-Methylphenol
Concen: 40.123 ng
RT: 8.204 min Scan# 890
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:107 Resp: 1018400
Ion Ratio Lower Upper
107 100
108 114.5 90.8 136.2
77 45.9 35.7 53.5
79 44.2 35.0 52.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#18
Hexachloroethane
Concen: 41.682 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:117 Resp: 533391
Ion Ratio Lower Upper
117 100
119 95.6 77.0 115.6
201 86.7 74.6 111.8





#19
n-Nitroso-di-n-propylamine
Concen: 41.848 ng
RT: 8.575 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 70 Resp: 931068
Ion Ratio Lower Upper
70 100
42 28.4 22.6 33.8
101 9.6 8.3 12.5
130 19.8 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#20
3+4-Methylphenols
Concen: 39.946 ng
RT: 8.534 min Scan# 943
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Tgt Ion:107 Resp: 1356508
Ion Ratio Lower Upper
107 100
108 89.6 70.4 110.4
77 32.9 12.2 52.2
79 28.5 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1290
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:136 Resp: 1877129
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 5.5 3.8 5.8
68 6.3 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#22
Acetophenone
Concen: 39.858 ng
RT: 8.581 min Scan# 951
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:105 Resp: 1869192
Ion Ratio Lower Upper
105 100
71 7.4 5.0 7.4
51 15.1 11.4 17.2
120 22.3 18.2 27.4





#23
Nitrobenzene-d5
Concen: 84.250 ng
RT: 8.916 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 3040809
Ion Ratio Lower Upper
82 100
128 41.5 35.2 52.8
54 31.9 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#24
Nitrobenzene
Concen: 41.871 ng
RT: 8.957 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion: 77 Resp: 1374442
Ion Ratio Lower Upper
77 100
123 45.2 38.2 57.2
65 12.5 9.5 14.3





#25
Isophorone
Concen: 40.035 ng
RT: 9.486 min Scan# 1105
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 249399
Ion Ratio Lower Upper
82 100
95 6.8 5.7 8.5
138 16.9 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#26
2-Nitrophenol
Concen: 41.282 ng
RT: 9.669 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:139 Resp: 584953
Ion Ratio Lower Upper
139 100
109 29.2 23.3 34.9
65 46.4 35.2 52.8





#27
2,4-Dimethylphenol
Concen: 39.596 ng
RT: 9.733 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:122 Resp: 115279
Ion Ratio Lower Upper
122 100
107 113.9 91.0 136.6
121 56.9 45.8 68.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#28
bis(2-Chloroethoxy)methane
Concen: 41.378 ng
RT: 9.969 min Scan# 1187
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48

Tgt Ion: 93 Resp: 1683958
Ion Ratio Lower Upper
93 100
95 32.4 26.2 39.4
123 10.8 8.9 13.3





#29

2,4-Dichlorophenol

Concen: 39.545 ng

RT: 10.198 min Scan# 1226

Delta R.T. -0.000 min

Lab File: BM050262.D

Acq: 10 Jun 2025 04:48

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:162 Resp: 1065409

Ion Ratio Lower Upper

162 100

164 64.5 45.0 85.0





98 39.1 16.3 56.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#30

1,2,4-Trichlorobenzene

Concen: 39.014 ng

RT: 10.416 min Scan# 1263

Delta R.T. -0.006 min

Lab File: BM050262.D

Acq: 10 Jun 2025 04:48




Tgt Ion:180 Resp: 1169668


Ion Ratio Lower Upper

180 100

182 95.0 76.6 114.8


145 30.3 24.2 36.4





#31
Naphthalene
Concen: 38.926 ng
RT: 10.604 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 3777854
Ion Ratio Lower Upper
128 100
129 11.0 8.6 12.8
127 12.7 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#32
Benzoic acid
Concen: 30.923 ng
RT: 9.851 min Scan# 1167
Delta R.T. 0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:122 Resp: 565798
Ion Ratio Lower Upper
122 100
105 125.4 103.4 143.4
77 95.0 70.3 110.3





#33
4-Chloroaniline
Concen: 39.291 ng
RT: 10.704 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:127 Resp: 1625729
Ion Ratio Lower Upper
127 100
129 32.0 25.4 38.0
65 28.3 21.0 31.4
92 18.9 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#34
Hexachlorobutadiene
Concen: 38.141 ng
RT: 10.898 min Scan# 1345
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:225 Resp: 680166
Ion Ratio Lower Upper
225 100
223 62.8 50.2 75.2
227 64.8 51.0 76.4





#35
Caprolactam
Concen: 38.607 ng
RT: 11.492 min Scan# 1447
Delta R.T. 0.012 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 32976
Ion Ratio Lower Upper
113 100
55 116.1 83.5 123.5
56 90.7 60.6 100.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#36
4-Chloro-3-methylphenol
Concen: 39.501 ng
RT: 11.833 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:107 Resp: 1135589
Ion Ratio Lower Upper
107 100
144 26.0 21.0 31.6
142 80.2 64.8 97.2





#37
2-Methylnaphthalene
Concen: 38.773 ng
RT: 12.216 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 227011
Ion Ratio Lower Upper
142 100
141 88.8 71.5 107.3
115 34.9 27.8 41.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#38
1-Methylnaphthalene
Concen: 38.905 ng
RT: 12.439 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:142 Resp: 2417639
Ion Ratio Lower Upper
142 100
141 91.1 72.9 109.3
116 3.8 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:164 Resp: 1070949
Ion Ratio Lower Upper
164 100
162 100.8 80.2 120.4
160 44.8 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.499 ng
RT: 12.586 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:216 Resp: 1221487
Ion Ratio Lower Upper
216 100
214 79.1 63.0 94.4
179 21.1 16.6 25.0
108 20.5 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 40.868 ng
RT: 12.574 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:237 Resp: 767530
Ion Ratio Lower Upper
237 100
235 63.3 44.6 84.6
272 13.0 0.0 33.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#42
2,4,6-Tribromophenol
Concen: 80.252 ng
RT: 15.880 min Scan# 2192
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Tgt Ion:330 Resp: 977575
Ion Ratio Lower Upper
330 100
332 96.0 77.5 116.3
141 37.1 28.8 43.2





#43
2,4,6-Trichlorophenol
Concen: 40.432 ng
RT: 12.821 min Scan# 1676
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:196 Resp: 82224
Ion Ratio Lower Upper
196 100
198 95.1 76.6 115.0
200 30.1 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#44
2,4,5-Trichlorophenol
Concen: 40.133 ng
RT: 12.892 min Scan# 1684
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:196 Resp: 896098
Ion Ratio Lower Upper
196 100
198 92.9 76.1 114.1
97 53.2 39.3 58.9
132 29.1 21.4 32.0





#45
2-Fluorobiphenyl
Concen: 78.077 ng
RT: 13.021 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:172 Resp: 6176210
Ion Ratio Lower Upper
172 100
171 34.8 27.8 41.6
170 23.0 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#46
1,1'-Biphenyl
Concen: 39.432 ng
RT: 13.233 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Tgt Ion:154 Resp: 3162089
Ion Ratio Lower Upper
154 100
153 39.5 19.2 59.2
76 13.3 0.0 32.0





#47
2-Chloronaphthalene
Concen: 39.343 ng
RT: 13.269 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:162 Resp: 245278
Ion Ratio Lower Upper
162 100
127 39.2 31.1 46.7
164 32.9 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#48
2-Nitroaniline
Concen: 46.242 ng
RT: 13.468 min Scan# 1782
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48

Tgt Ion: 65 Resp: 634421
Ion Ratio Lower Upper
65 100
92 75.8 64.2 96.4
138 115.1 101.0 151.4





#49

Acenaphthylene

Concen: 39.574 ng

RT: 14.121 min Scan# 1893

Delta R.T. -0.000 min

Lab File: BM050262.D


Acq: 10 Jun 2025 04:48

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:152 Resp: 3990554

Ion Ratio Lower Upper

152 100

151 19.6 15.8 23.6




153 12.9 10.4 15.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#50

Dimethylphthalate


Concen: 39.566 ng

RT: 13.863 min Scan# 1849

Delta R.T. -0.000 min

Lab File: BM050262.D

Acq: 10 Jun 2025 04:48





Tgt Ion:163 Resp: 2865663


Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9


164 9.8 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 42.082 ng
RT: 13.968 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 614664
Ion Ratio Lower Upper
165 100
63 49.9 35.8 53.6
89 59.3 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#52
Acenaphthene
Concen: 37.260 ng
RT: 14.463 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:154 Resp: 2350065
Ion Ratio Lower Upper
154 100
153 113.1 88.7 133.1
152 53.9 42.6 63.8





#53
3-Nitroaniline
Concen: 42.411 ng
RT: 14.292 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:138 Resp: 690400
Ion Ratio Lower Upper
138 100
108 10.9 8.8 13.2
92 116.9 92.9 139.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#54
2,4-Dinitrophenol
Concen: 17.614 ng
RT: 14.492 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Tgt Ion:184 Resp: 124034
Ion Ratio Lower Upper
184 100
63 63.4 45.3 67.9
154 74.8 54.3 81.5





#55
Dibenzenofuran
Concen: 39.138 ng
RT: 14.798 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 3611400
Ion Ratio Lower Upper
168 100
139 37.4 29.9 44.9
169 13.3 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#56
4-Nitrophenol
Concen: 41.052 ng
RT: 14.598 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:139 Resp: 568795
Ion Ratio Lower Upper
139 100
109 73.1 52.6 92.6
65 86.1 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 41.961 ng
RT: 14.751 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 80848
Ion Ratio Lower Upper
165 100
63 40.7 28.4 42.6
89 79.2 58.0 87.0
182 2.5 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#58
Fluorene
Concen: 38.838 ng
RT: 15.445 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:166 Resp: 2744828
Ion Ratio Lower Upper
166 100
165 96.5 77.0 115.4
167 13.5 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 39.241 ng m
RT: 15.021 min Scan# 2046
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:232 Resp: 758061
Ion Ratio Lower Upper
232 100
131 43.6 32.9 49.3
130 2.4 1.8 2.8
166 26.6 20.8 31.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#60
Diethylphthalate
Concen: 40.067 ng
RT: 15.227 min Scan# 2081
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:149 Resp: 2879326
Ion Ratio Lower Upper
149 100
177 21.6 17.2 25.8
150 12.1 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 39.017 ng
RT: 15.445 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:204 Resp: 1332679
Ion Ratio Lower Upper
204 100
206 32.8 25.4 38.0
141 59.2 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#62
4-Nitroaniline
Concen: 42.568 ng
RT: 15.457 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:138 Resp: 649475
Ion Ratio Lower Upper
138 100
92 51.5 28.8 68.8
108 80.0 58.1 98.1





#63
Azobenzene
Concen: 41.194 ng
RT: 15.733 min Scan# 2171
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 77 Resp: 295887
Ion Ratio Lower Upper
77 100
182 23.6 6.8 46.8
105 21.2 1.7 41.7
51 16.0 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:188 Resp: 1957349
Ion Ratio Lower Upper
188 100
94 10.4 8.1 12.1
80 11.1 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 24.779 ng
RT: 15.515 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:198 Resp: 26490
Ion Ratio Lower Upper
198 100
51 26.9 4.5 44.5
105 48.6 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#66
n-Nitrosodiphenylamine
Concen: 39.721 ng
RT: 15.651 min Scan# 2153
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:169 Resp: 2442884
Ion Ratio Lower Upper
169 100
168 66.3 53.1 79.7
167 36.3 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 39.697 ng
RT: 16.333 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:248 Resp: 82858
Ion Ratio Lower Upper
248 100
250 95.2 78.0 117.0
141 76.5 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#68
Hexachlorobenzene
Concen: 39.865 ng
RT: 16.445 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48

Tgt Ion:284 Resp: 972175
Ion Ratio Lower Upper
284 100
142 36.3 28.8 43.2
249 31.6 25.8 38.6





#69

Atrazine

Concen: 41.337 ng

RT: 16.604 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM050262.D


Acq: 10 Jun 2025 04:48

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:200 Resp: 81004

Ion Ratio Lower Upper

200 100

173 26.8 6.0 46.0




215 47.8 28.1 68.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#70

Pentachlorophenol

Concen: 55.421 ng

RT: 16.786 min Scan# 2346

Delta R.T. -0.000 min

Lab File: BM050262.D

Acq: 10 Jun 2025 04:48




Tgt Ion:266 Resp: 878713


Ion Ratio Lower Upper

266 100

268 64.5 51.0 76.6


264 63.1 50.8 76.2





#71
Phenanthrene
Concen: 39.123 ng
RT: 17.180 min Scan# 2416
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 437461
Ion Ratio Lower Upper
178 100
176 19.0 15.6 23.4
179 15.3 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#72
Anthracene
Concen: 39.385 ng
RT: 17.268 min Scan# 2428
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:178 Resp: 4405600
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.3 12.3 18.5





#73
Carbazole
Concen: 39.880 ng
RT: 17.533 min Scan# 2473
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:167 Resp: 4064090
Ion Ratio Lower Upper
167 100
166 21.3 17.0 25.6
139 12.8 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#74
Di-n-butylphthalate
Concen: 43.289 ng
RT: 18.115 min Scan# 2572
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:149 Resp: 4799970
Ion Ratio Lower Upper
149 100
150 9.1 7.4 11.0
104 5.1 4.0 6.0





#75
Fluoranthene
Concen: 40.590 ng
RT: 19.192 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 478625
Ion Ratio Lower Upper
202 100
101 11.2 0.0 31.1
203 17.1 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:240 Resp: 1968381
Ion Ratio Lower Upper
240 100
120 11.3 9.0 13.4
236 25.6 20.7 31.1





#77
Benzidine
Concen: 44.801 ng
RT: 19.374 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 247985
Ion Ratio Lower Upper
184 100
185 14.8 11.5 17.3
183 11.0 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#78
Pyrene
Concen: 38.210 ng
RT: 19.550 min Scan# 2816
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:202 Resp: 5069333
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.6 14.1 21.1





#79
Terphenyl-d14
Concen: 77.259 ng
RT: 19.762 min Scan# 2852
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:244 Resp: 802497
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 11.0 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#80
Butylbenzylphthalate
Concen: 43.849 ng
RT: 20.462 min Scan# 2971
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:149 Resp: 1940495
Ion Ratio Lower Upper
149 100
91 74.0 56.9 85.3
206 20.1 16.1 24.1





#81
Benzo(a)anthracene
Concen: 39.375 ng
RT: 21.350 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 490575
Ion Ratio Lower Upper
228 100
226 27.4 21.7 32.5
229 19.5 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#82
3,3'-Dichlorobenzidine
Concen: 44.182 ng
RT: 21.274 min Scan# 3109
Delta R.T. -0.000 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:252 Resp: 1678656
Ion Ratio Lower Upper
252 100
254 64.8 51.4 77.0
126 11.1 8.6 12.8





#83
Chrysene
Concen: 39.507 ng
RT: 21.409 min Scan# 3112
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 468206
Ion Ratio Lower Upper
228 100
226 30.3 24.3 36.5
229 19.6 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 45.251 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.012 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:149 Resp: 3084575
Ion Ratio Lower Upper
149 100
167 26.8 22.3 33.5
279 3.8 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 46.575 ng
RT: 22.427 min Scan# 311
Delta R.T. -0.012 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 4711085
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 5.0 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.338 min Scan# 3630
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48


Tgt Ion:264 Resp: 2090330
Ion Ratio Lower Upper
264 100
260 23.2 18.6 28.0
265 21.2 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 42.482 ng
RT: 27.697 min Scan# 41
Delta R.T. -0.012 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:276 Resp: 571770
Ion Ratio Lower Upper
276 100
138 26.3 20.6 30.8
277 25.0 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#88
Benzo(b)fluoranthene
Concen: 40.276 ng
RT: 23.403 min Scan# 3471
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48

Tgt Ion:252 Resp: 4894905
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 10.5 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 39.246 ng

RT: 23.468 min Scan# 34

Delta R.T. -0.006 min

Lab File: BM050262.D

Acq: 10 Jun 2025 04:48

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 486911

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2





125 10.4 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#90

Benzo(a)pyrene

Concen: 40.782 ng

RT: 24.203 min Scan# 3607

Delta R.T. -0.006 min

Lab File: BM050262.D

Acq: 10 Jun 2025 04:48




Tgt Ion:252 Resp: 4660050


Ion Ratio Lower Upper

252 100

253 21.6 17.6 26.4


125 11.3 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 42.817 ng
RT: 27.762 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:278 Resp: 4677689
Ion Ratio Lower Upper
278 100
139 16.7 12.8 19.2
279 23.7 18.6 28.0




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#92
Benzo(g,h,i)perylene
Concen: 42.844 ng
RT: 28.738 min Scan# 4378
Delta R.T. -0.012 min
Lab File: BM050262.D
Acq: 10 Jun 2025 04:48

Tgt Ion:276 Resp: 4684760
Ion Ratio Lower Upper
276 100
277 23.3 18.6 27.8
138 22.3 17.8 26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 05:44:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	149	0.00
2	1,4-Dioxane	0.526	0.542	-3.0	169#	0.00
3	Pyridine	1.361	1.448	-6.4	165#	0.00
4	n-Nitrosodimethylamine	0.281	0.305	-8.5	168#	0.00
5 S	2-Fluorophenol	1.199	1.232	-2.8	160#	0.00
6	Aniline	2.086	2.092	-0.3	151#	0.00
7 S	Phenol-d6	1.578	1.600	-1.4	153#	0.00
8	2-Chlorophenol	1.319	1.338	-1.4	154#	0.00
9	Benzaldehyde	1.003	0.967	3.6	164#	0.00
10 C	Phenol	1.670	1.704	-2.0	155#	0.00
11	bis(2-Chloroethyl)ether	1.343	1.373	-2.2	156#	0.00
12	1,3-Dichlorobenzene	1.474	1.439	2.4	152#	0.00
13 C	1,4-Dichlorobenzene	1.534	1.485	3.2	153#	0.00
14	1,2-Dichlorobenzene	1.462	1.430	2.2	152#	0.00
15	Benzyl Alcohol	1.126	1.143	-1.5	150	0.00
16	2,2'-oxybis(1-Chloropropane	0.933	1.118	-19.8	183#	0.00
17	2-Methylphenol	1.102	1.105	-0.3	149	0.00
18	Hexachloroethane	0.556	0.579	-4.1	162#	0.00
19 P	n-Nitroso-di-n-propylamine	0.966	1.010	-4.6	151#	0.00
20	3+4-Methylphenols	1.474	1.472	0.1	146	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	143	0.00
22	Acetophenone	0.500	0.498	0.4	149	0.00
23 S	Nitrobenzene-d5	0.385	0.405	-5.2	153#	0.00
24	Nitrobenzene	0.350	0.366	-4.6	154#	0.00
25	Isophorone	0.664	0.664	0.0	144	0.00
26 C	2-Nitrophenol	0.151	0.156	-3.3	143	0.00
27	2,4-Dimethylphenol	0.310	0.307	1.0	144	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.449	-3.5	150	0.00
29 C	2,4-Dichlorophenol	0.287	0.284	1.0	141	0.00
30	1,2,4-Trichlorobenzene	0.319	0.312	2.2	146	0.00
31	Naphthalene	1.034	1.006	2.7	147	0.00
32	Benzoic acid	0.195	0.151	22.6	105	0.00
33	4-Chloroaniline	0.441	0.433	1.8	143	0.00
34 C	Hexachlorobutadiene	0.190	0.181	4.7	142	0.00
35	Caprolactam	0.091	0.088	3.3	129	0.01
36 C	4-Chloro-3-methylphenol	0.306	0.302	1.3	138	0.00
37	2-Methylnaphthalene	0.624	0.605	3.0	140	0.00
38	1-Methylnaphthalene	0.662	0.644	2.7	141	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
40	1,2,4,5-Tetrachlorobenzene	0.578	0.570	1.4	138	0.00
41 P	Hexachlorocyclopentadiene	0.351	0.358	-2.0	137	0.00
42 S	2,4,6-Tribromophenol	0.227	0.228	-0.4	129	0.00
43 C	2,4,6-Trichlorophenol	0.380	0.384	-1.1	134	0.00
44	2,4,5-Trichlorophenol	0.417	0.418	-0.2	132	0.00
45 S	2-Fluorobiphenyl	1.477	1.442	2.4	137	0.00
46	1,1'-Biphenyl	1.498	1.476	1.5	138	0.00
47	2-Chloronaphthalene	1.164	1.145	1.6	138	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 05:44:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.256	0.296	-15.6	144	0.00
49	Acenaphthylene	1.883	1.863	1.1	134	0.00
50	Dimethylphthalate	1.353	1.338	1.1	130	0.00
51	2,6-Dinitrotoluene	0.273	0.287	-5.1	130	0.00
52 C	Acenaphthene	1.178	1.097	6.9	127	0.00
53	3-Nitroaniline	0.304	0.322	-5.9	130	0.00
54 P	2,4-Dinitrophenol	0.146	0.058	60.3#	50#	0.00
55	Dibenzofuran	1.723	1.686	2.1	133	0.00
56 P	4-Nitrophenol	0.259	0.266	-2.7	127	0.00
57	2,4-Dinitrotoluene	0.360	0.377	-4.7	125	0.00
58	Fluorene	1.320	1.281	3.0	132	0.00
59	2,3,4,6-Tetrachlorophenol	0.361	0.354	1.9	127	0.00
60	Diethylphthalate	1.342	1.344	-0.1	131	0.00
61	4-Chlorophenyl-phenylether	0.638	0.622	2.5	133	0.00
62	4-Nitroaniline	0.285	0.303	-6.3	129	0.00
63	Azobenzene	1.341	1.381	-3.0	138	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
65	4,6-Dinitro-2-methylphenol	0.109	0.068	37.6#	75	0.00
66 c	n-Nitrosodiphenylamine	0.628	0.624	0.6	129	0.00
67	4-Bromophenyl-phenylether	0.213	0.212	0.5	128	0.00
68	Hexachlorobenzene	0.249	0.248	0.4	129	0.00
69	Atrazine	0.200	0.207	-3.5	124	0.00
70 C	Pentachlorophenol	0.162	0.224	-38.3#	169#	0.00
71	Phenanthrene	1.143	1.117	2.3	128	0.00
72	Anthracene	1.143	1.125	1.6	128	0.00
73	Carbazole	1.041	1.038	0.3	127	0.00
74	Di-n-butylphthalate	1.133	1.226	-8.2	130	0.00
75 C	Fluoranthene	1.205	1.223	-1.5	128	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
77	Benzidine	0.562	0.630	-12.1	135	0.00
78	Pyrene	1.348	1.288	4.5	129	0.00
79 S	Terphenyl-d14	1.055	1.019	3.4	134	0.00
80	Butylbenzylphthalate	0.450	0.493	-9.6	133	0.00
81	Benzo(a)anthracene	1.266	1.246	1.6	136	0.00
82	3,3'-Dichlorobenzidine	0.386	0.426	-10.4	136	0.00
83	Chrysene	1.204	1.189	1.2	137	0.00
84	Bis(2-ethylhexyl)phthalate	0.693	0.784	-13.1	140	-0.01
85 c	Di-n-octyl phthalate	1.028	1.197	-16.4	142	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	148	0.00
87	Indeno(1,2,3-cd)pyrene	1.288	1.368	-6.2	170#	-0.01
88	Benzo(b)fluoranthene	1.163	1.171	-0.7	148	0.00
89	Benzo(k)fluoranthene	1.187	1.165	1.9	148	0.00
90 C	Benzo(a)pyrene	1.093	1.115	-2.0	153#	0.00
91	Dibenzo(a,h)anthracene	1.045	1.119	-7.1	171#	0.00
92	Benzo(g,h,i)perylene	1.046	1.121	-7.2	174#	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050262.D
Acq On : 10 Jun 2025 04:48
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 05:44:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 05:44:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	149	0.00
2	1,4-Dioxane	40.000	41.272	-3.2	169	0.00
3	Pyridine	40.000	42.552	-6.4	165	0.00
4	n-Nitrosodimethylamine	40.000	43.394	-8.5	168	0.00
5 S	2-Fluorophenol	80.000	82.193	-2.7	160	0.00
6	Aniline	40.000	40.106	-0.3	151	0.00
7 S	Phenol-d6	80.000	81.099	-1.4	153	0.00
8	2-Chlorophenol	40.000	40.570	-1.4	154	0.00
9	Benzaldehyde	40.000	38.531	3.7	164	0.00
10 C	Phenol	40.000	40.818	-2.0	155	0.00
11	bis(2-Chloroethyl)ether	40.000	40.887	-2.2	156	0.00
12	1,3-Dichlorobenzene	40.000	39.041	2.4	152	0.00
13 C	1,4-Dichlorobenzene	40.000	38.711	3.2	153	0.00
14	1,2-Dichlorobenzene	40.000	39.103	2.2	152	0.00
15	Benzyl Alcohol	40.000	40.632	-1.6	150	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	47.910	-19.8	183	0.00
17	2-Methylphenol	40.000	40.123	-0.3	149	0.00
18	Hexachloroethane	40.000	41.682	-4.2	162	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.848	-4.6	151	0.00
20	3+4-Methylphenols	40.000	39.946	0.1	146	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	143	0.00
22	Acetophenone	40.000	39.858	0.4	149	0.00
23 S	Nitrobenzene-d5	80.000	84.250	-5.3	153	0.00
24	Nitrobenzene	40.000	41.871	-4.7	154	0.00
25	Isophorone	40.000	40.035	-0.1	144	0.00
26 C	2-Nitrophenol	40.000	41.282	-3.2	143	0.00
27	2,4-Dimethylphenol	40.000	39.596	1.0	144	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.378	-3.4	150	0.00
29 C	2,4-Dichlorophenol	40.000	39.545	1.1	141	0.00
30	1,2,4-Trichlorobenzene	40.000	39.014	2.5	146	0.00
31	Naphthalene	40.000	38.926	2.7	147	0.00
32	Benzoic acid	40.000	30.923	22.7	105	0.00
33	4-Chloroaniline	40.000	39.291	1.8	143	0.00
34 C	Hexachlorobutadiene	40.000	38.141	4.6	142	0.00
35	Caprolactam	40.000	38.607	3.5	129	0.01
36 C	4-Chloro-3-methylphenol	40.000	39.501	1.2	138	0.00
37	2-Methylnaphthalene	40.000	38.773	3.1	140	0.00
38	1-Methylnaphthalene	40.000	38.905	2.7	141	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	131	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.499	1.3	138	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.868	-2.2	137	0.00
42 S	2,4,6-Tribromophenol	80.000	80.252	-0.3	129	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.432	-1.1	134	0.00
44	2,4,5-Trichlorophenol	40.000	40.133	-0.3	132	0.00
45 S	2-Fluorobiphenyl	80.000	78.077	2.4	137	0.00
46	1,1'-Biphenyl	40.000	39.432	1.4	138	0.00
47	2-Chloronaphthalene	40.000	39.343	1.6	138	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050262.D
 Acq On : 10 Jun 2025 04:48
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 05:44:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.242	-15.6	144	0.00
49	Acenaphthylene	40.000	39.574	1.1	134	0.00
50	Dimethylphthalate	40.000	39.566	1.1	130	0.00
51	2,6-Dinitrotoluene	40.000	42.082	-5.2	130	0.00
52 C	Acenaphthene	40.000	37.260	6.9	127	0.00
53	3-Nitroaniline	40.000	42.411	-6.0	130	0.00
54 P	2,4-Dinitrophenol	40.000	17.614	56.0#	50	0.00
55	Dibenzofuran	40.000	39.138	2.2	133	0.00
56 P	4-Nitrophenol	40.000	41.052	-2.6	127	0.00
57	2,4-Dinitrotoluene	40.000	41.961	-4.9	125	0.00
58	Fluorene	40.000	38.838	2.9	132	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.241	1.9	127	0.00
60	Diethylphthalate	40.000	40.067	-0.2	131	0.00
61	4-Chlorophenyl-phenylether	40.000	39.017	2.5	133	0.00
62	4-Nitroaniline	40.000	42.568	-6.4	129	0.00
63	Azobenzene	40.000	41.194	-3.0	138	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	123	0.00
65	4,6-Dinitro-2-methylphenol	40.000	24.779	38.1#	75	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.721	0.7	129	0.00
67	4-Bromophenyl-phenylether	40.000	39.697	0.8	128	0.00
68	Hexachlorobenzene	40.000	39.865	0.3	129	0.00
69	Atrazine	40.000	41.337	-3.3	124	0.00
70 C	Pentachlorophenol	40.000	55.421	-38.6#	169	0.00
71	Phenanthrene	40.000	39.123	2.2	128	0.00
72	Anthracene	40.000	39.385	1.5	128	0.00
73	Carbazole	40.000	39.880	0.3	127	0.00
74	Di-n-butylphthalate	40.000	43.289	-8.2	130	0.00
75 C	Fluoranthene	40.000	40.590	-1.5	128	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	133	0.00
77	Benzydine	40.000	44.801	-12.0	135	0.00
78	Pyrene	40.000	38.210	4.5	129	0.00
79 S	Terphenyl-d14	80.000	77.259	3.4	134	0.00
80	Butylbenzylphthalate	40.000	43.849	-9.6	133	0.00
81	Benzo(a)anthracene	40.000	39.375	1.6	136	0.00
82	3,3'-Dichlorobenzidine	40.000	44.182	-10.5	136	0.00
83	Chrysene	40.000	39.507	1.2	137	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.251	-13.1	140	-0.01
85 c	Di-n-octyl phthalate	40.000	46.575	-16.4	142	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	148	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.482	-6.2	170	-0.01
88	Benzo(b)fluoranthene	40.000	40.276	-0.7	148	0.00
89	Benzo(k)fluoranthene	40.000	39.246	1.9	148	0.00
90 C	Benzo(a)pyrene	40.000	40.782	-2.0	153	0.00
91	Dibenzo(a,h)anthracene	40.000	42.817	-7.0	171	0.00
92	Benzo(g,h,i)perylene	40.000	42.844	-7.1	174	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050262.D
Acq On : 10 Jun 2025 04:48
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 05:44:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 1




QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050193.D
Acq On : 05 Jun 2025 08:40
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jun 05 16:20:25 2025






AutoFind: Scans 2486, 2487, 2488; Background Corrected with Scan 2479

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.3	101444	PASS
68	69	0.00	2	1.6	2964	PASS
69	198	0.00	100	37.2	185664	PASS
70	69	0.00	2	0.6	1032	PASS
127	198	10	80	47.5	237035	PASS
197	198	0.00	2	0.4	1811	PASS
198	198	100	100	100.0	499264	PASS
199	198	5	9	7.0	34904	PASS
275	198	10	60	24.6	122771	PASS
365	198	1	100	3.2	16128	PASS
441	198	0.01	100	11.6	57920	PASS
442	442	50	100	100.0	371584	PASS
443	442	15	24	19.9	74069	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050193.D
Acq On : 05 Jun 2025 08:40
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 05 16:27:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 16:20:25 2025
Response via : Initial Calibration



TIC: BM050193.D\data.ms

(70) Pentachlorophenol (C)

16.804min (+ 0.000) 1708888.69 ng




response 747502

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.80
264.00	63.50	62.45
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060525\
Data File : BM050193.D
Acq On : 05 Jun 2025 08:40
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 05 16:27:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 05 16:20:25 2025
Response via : Initial Calibration



TIC: BM050193.D\data.ms

(77) Benzidine

19.398min (+ 0.006) 33233.28 ng

response 3215775

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.84
183.00	11.20	11.16
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/5/2025	BNA_M	<u>BM050193.D</u>
Compound Name	Response	Retention Time
DDT	1539212	20.645
DDD	20154	20.257
DDE	1832	19.692
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
21986	1561198	1.41


Instrument :
BNA_M
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050244.D
Acq On : 09 Jun 2025 16:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Jun 09 11:54:38 2025






AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.2	88096	PASS
68	69	0.00	2	1.4	2134	PASS
69	198	0.00	100	36.0	157327	PASS
70	69	0.00	2	0.3	522	PASS
127	198	10	80	46.5	203328	PASS
197	198	0.00	2	0.4	1683	PASS
198	198	100	100	100.0	437035	PASS
199	198	5	9	6.8	29595	PASS
275	198	10	60	25.3	110448	PASS
365	198	1	100	3.4	14806	PASS
441	198	0.01	100	12.7	55720	PASS
442	442	50	100	100.0	360981	PASS
443	442	15	24	19.1	68901	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050244.D
Acq On : 09 Jun 2025 16:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 09 17:38:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration



TIC: BM050244.D\data.ms

(70) Pentachlorophenol (C)

16.786min (-0.000) 282541.19 ng




response 572172

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	62.74
264.00	63.50	63.02
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050244.D
Acq On : 09 Jun 2025 16:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 09 17:38:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration



TIC: BM050244.D\data.ms

(77) Benzidine

19.374min (-0.000) 185831.97 ng

response 2889837

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.40
183.00	11.20	11.15
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/9/2025	BNA_M	<u>BM050244.D</u>
Compound Name	Response	Retention Time
DDT	1441801	20.621
DDD	16704	20.233
DDE	1198	19.674
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
17902	1459703	1.23


Instrument :
BNA_M
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050261.D
Acq On : 10 Jun 2025 04:09
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Jun 09 11:54:38 2025






AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.8	99808	PASS
68	69	0.00	2	1.5	2792	PASS
69	198	0.00	100	45.4	189995	PASS
70	69	0.00	2	0.5	984	PASS
127	198	10	80	52.3	219093	PASS
197	198	0.00	2	0.3	1212	PASS
198	198	100	100	100.0	418709	PASS
199	198	5	9	6.6	27819	PASS
275	198	10	60	24.1	100909	PASS
365	198	1	100	3.1	12955	PASS
441	198	0.01	100	11.3	47504	PASS
442	442	50	100	100.0	294400	PASS
443	442	15	24	20.2	59373	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050261.D
Acq On : 10 Jun 2025 04:09
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 10 05:34:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration



TIC: BM050261.D\data.ms

(70) Pentachlorophenol (C)

16.786min (-0.000) 158695.78 ng




response 460207

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.15
264.00	63.50	62.70
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050261.D
Acq On : 10 Jun 2025 04:09
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jun 10 05:34:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration



TIC: BM050261.D\data.ms

(77) Benzidine**19.374min (-0.000) 99825.69 ng****response 2680856**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.26
183.00	11.20	11.06
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/9/2025	BNA_M	<u>BM050261.D</u>
Compound Name	Response	Retention Time
DDT	1335236	20.615
DDD	14959	20.233
DDE	1895	19.668
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
16854	1352090	1.25

Instrument :
BNA_M
ClientSampleId :
DFTPP

Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	PB168352BL	SDG No.:	Q2236
Lab Sample ID:	PB168352BL	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050263.D	1	06/09/25 10:45	06/10/25 05:28	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.30	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.53	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	0.65	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	0.76	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	0.54	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.20	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	0.52	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	151		15 (23) - 110 (138)	101%	SPK: 150
13127-88-3	Phenol-d6	142		15 (10) - 110 (134)	94%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.5		30 (67) - 130 (132)	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	91.6		30 (52) - 130 (132)	92%	SPK: 100
118-79-6	2,4,6-Tribromophenol	156		15 (44) - 110 (137)	104%	SPK: 150
1718-51-0	Terphenyl-d14	88.3		30 (42) - 130 (152)	88%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	335000	7.763			
1146-65-2	Naphthalene-d8	1240000	10.551			
15067-26-2	Acenaphthene-d10	699000	14.392			
1517-22-2	Phenanthrene-d10	1380000	17.133			
1719-03-5	Chrysene-d12	1500000	21.362			
1520-96-3	Perylene-d12	1570000	24.338			

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:		
Client Sample ID:	PB168352BL		SDG No.:	Q2236	
Lab Sample ID:	PB168352BL		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050263.D	1	06/09/25 10:45	06/10/25 05:28	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050263.D
Acq On : 10 Jun 2025 05:28
Operator : RC/JU
Sample : PB168352BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168352BL

Quant Time: Jun 10 06:52:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	334540	20.000	ng	0.00
21) Naphthalene-d8	10.551	136	1240838	20.000	ng	0.00
39) Acenaphthene-d10	14.392	164	698531	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1381505	20.000	ng	0.00
76) Chrysene-d12	21.362	240	1501560	20.000	ng	-0.01
86) Perylene-d12	24.338	264	1574973	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	3032395	151.207	ng	0.00
7) Phenol-d6	6.934	99	3739604	141.651	ng	0.00
23) Nitrobenzene-d5	8.916	82	2206555	92.485	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1240049	156.073	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	4725754	91.592	ng	0.00
79) Terphenyl-d14	19.762	244	6998812	88.328	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050263.D
Acq On : 10 Jun 2025 05:28
Operator : RC/JU
Sample : PB168352BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168352BL


Quant Time: Jun 10 06:52:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050263.D
Acq On : 10 Jun 2025 05:28
Operator : RC/JU
Sample : PB168352BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168352BL

Quant Time: Jun 10 06:52:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.763 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28




Instrument :
BNA_M
ClientSampleId :
PB168352BL


Tgt Ion:152 Resp: 334540
Ion Ratio Lower Upper
152 100
150 156.9 122.2 183.4
115 61.2 46.3 69.5



#5
2-Fluorophenol
Concen: 151.207 ng
RT: 5.357 min Scan# 403
Delta R.T. 0.000 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28

Tgt Ion:112 Resp: 3032395
Ion Ratio Lower Upper
112 100
64 48.5 39.8 59.6
63 24.7 19.8 29.8









#7
Phenol-d6
Concen: 141.651 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28




Instrument :
BNA_M
ClientSampleId :
PB168352BL


Tgt Ion: 99 Resp: 3739604
Ion Ratio Lower Upper
99 100
42 8.8 6.9 10.3
71 31.9 25.3 37.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1286
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28

Tgt Ion: 136 Resp: 1240838
Ion Ratio Lower Upper
136 100
137 10.8 8.6 13.0
54 5.0 3.8 5.8
68 6.1 4.9 7.3









#23
Nitrobenzene-d5
Concen: 92.485 ng
RT: 8.916 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28




Instrument :
BNA_M
ClientSampleId :
PB168352BL


Tgt Ion: 82 Resp: 2206555
Ion Ratio Lower Upper
82 100
128 43.4 35.2 52.8
54 30.4 24.5 36.7



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.392 min Scan# 1939
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28


Tgt Ion:164 Resp: 698531
Ion Ratio Lower Upper
164 100
162 99.7 80.2 120.4
160 46.0 35.7 53.5








#42
2,4,6-Tribromophenol
Concen: 156.073 ng
RT: 15.880 min Scan# 2192
Delta R.T. 0.000 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28

Instrument :
BNA_M
ClientSampleId :
PB168352BL







Tgt Ion:330 Resp: 1240049
Ion Ratio Lower Upper
330 100
332 96.8 77.5 116.3
141 36.2 28.8 43.2



#45
2-Fluorobiphenyl
Concen: 91.592 ng
RT: 13.022 min Scan# 1706
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28

Tgt Ion:172 Resp: 4725754
Ion Ratio Lower Upper
172 100
171 35.0 27.8 41.6
170 23.1 18.6 27.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28




Instrument :
BNA_M
ClientSampleId :
PB168352BL


Tgt Ion:188 Resp: 1381505
Ion Ratio Lower Upper
188 100
94 10.0 8.1 12.1
80 10.6 8.6 13.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.362 min Scan# 3124
Delta R.T. -0.012 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28

Tgt Ion:240 Resp: 1501560
Ion Ratio Lower Upper
240 100
120 11.8 9.0 13.4
236 25.3 20.7 31.1









#79
Terphenyl-d14
Concen: 88.328 ng
RT: 19.762 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28




Instrument :
BNA_M
ClientSampleId :
PB168352BL

Tgt Ion:244 Resp: 6998812
Ion Ratio Lower Upper
244 100
212 7.3 6.0 9.0
122 10.9 8.6 12.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.338 min Scan# 3630
Delta R.T. -0.006 min
Lab File: BM050263.D
Acq: 10 Jun 2025 05:28

Tgt Ion:264 Resp: 1574973
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 21.9 16.8 25.2



Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	PB168352BS	SDG No.:	Q2236
Lab Sample ID:	PB168352BS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050264.D	1	06/09/25 10:45	06/10/25 06:07	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	43.6		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	46.9		0.53	5.00	ug/L
95-48-7	2-Methylphenol	49.2		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	49.6		1.10	10.0	ug/L
67-72-1	Hexachloroethane	49.5		0.65	5.00	ug/L
98-95-3	Nitrobenzene	51.3		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	48.2		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	52.5		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	52.9		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	58.9		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	52.0		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	110	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	146		15 (23) - 110 (138)	97%	SPK: 150
13127-88-3	Phenol-d6	140		15 (10) - 110 (134)	93%	SPK: 150
4165-60-0	Nitrobenzene-d5	88.6		30 (67) - 130 (132)	89%	SPK: 100
321-60-8	2-Fluorobiphenyl	86.1		30 (52) - 130 (132)	86%	SPK: 100
118-79-6	2,4,6-Tribromophenol	153		15 (44) - 110 (137)	102%	SPK: 150
1718-51-0	Terphenyl-d14	86.8		30 (42) - 130 (152)	87%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	334000		7.763		
1146-65-2	Naphthalene-d8	1310000		10.551		
15067-26-2	Acenaphthene-d10	755000		14.398		
1517-22-2	Phenanthrene-d10	1420000		17.133		
1719-03-5	Chrysene-d12	1520000		21.368		
1520-96-3	Perylene-d12	1660000		24.339		

Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	PB168352BS	SDG No.:	Q2236
Lab Sample ID:	PB168352BS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050264.D	1	06/09/25 10:45	06/10/25 06:07	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050264.D
 Acq On : 10 Jun 2025 06:07
 Operator : RC/JU
 Sample : PB168352BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB168352BS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 06:53:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	333914	20.000	ng	0.00
21) Naphthalene-d8	10.551	136	1311547	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	754658	20.000	ng	0.00
64) Phenanthrene-d10	17.133	188	1423727	20.000	ng	0.00
76) Chrysene-d12	21.368	240	1520709	20.000	ng	0.00
86) Perylene-d12	24.339	264	1658221	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	2923158	146.033	ng	0.00
7) Phenol-d6	6.934	99	3690740	140.063	ng	0.00
23) Nitrobenzene-d5	8.916	82	2233083	88.551	ng	0.00
42) 2,4,6-Tribromophenol	15.880	330	1315918	153.304	ng	0.00
45) 2-Fluorobiphenyl	13.022	172	4797135	86.060	ng	0.00
79) Terphenyl-d14	19.762	244	6968788	86.841	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.281	88	340181	38.770	ng	95
3) Pyridine	3.669	79	991163	43.620	ng	98
4) n-Nitrosodimethylamine	3.581	42	225565	48.006	ng	97
6) Aniline	7.093	93	1261529	36.214	ng	97
8) 2-Chlorophenol	7.334	128	1112221	50.515	ng	98
9) Benzaldehyde	6.904	77	568218	33.917	ng	100
10) Phenol	6.963	94	1412735	50.671	ng	99
11) bis(2-Chloroethyl)ether	7.193	93	1096770	48.907	ng	98
12) 1,3-Dichlorobenzene	7.657	146	1171169	47.580	ng	98
13) 1,4-Dichlorobenzene	7.798	146	1200936	46.892	ng	99
14) 1,2-Dichlorobenzene	8.116	146	1150887	47.139	ng	98
15) Benzyl Alcohol	8.004	79	923134	49.118	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.293	45	818561	52.545	ng	97
17) 2-Methylphenol	8.204	107	904893	49.189	ng	97
18) Hexachloroethane	8.845	117	458802	49.468	ng	99
19) n-Nitroso-di-n-propyla...	8.575	70	786038	48.746	ng	99
20) 3+4-Methylphenols	8.534	107	1220427	49.586	ng	99
22) Acetophenone	8.581	105	1613029	49.228	ng	99
24) Nitrobenzene	8.957	77	1176763	51.308	ng	98
25) Isophorone	9.487	82	2138641	49.135	ng	100
26) 2-Nitrophenol	9.663	139	500374	50.542	ng	96
27) 2,4-Dimethylphenol	9.734	122	1027023	50.488	ng	100
28) bis(2-Chloroethoxy)met...	9.969	93	1439847	50.637	ng	100
29) 2,4-Dichlorophenol	10.198	162	968227	51.435	ng	98
30) 1,2,4-Trichlorobenzene	10.416	180	1026445	49.001	ng	99
31) Naphthalene	10.604	128	3253648	47.982	ng	100
32) Benzoic acid	9.857	122	633129	49.525	ng	99
33) 4-Chloroaniline	10.704	127	761701	26.347	ng	99
34) Hexachlorobutadiene	10.898	225	600603	48.203	ng	99
35) Caprolactam	11.481	113	299123m	50.122	ng	
36) 4-Chloro-3-methylphenol	11.833	107	1025222	51.040	ng	100
37) 2-Methylnaphthalene	12.216	142	2016294	49.289	ng	99
38) 1-Methylnaphthalene	12.439	142	2115216	48.717	ng	99
40) 1,2,4,5-Tetrachloroben...	12.586	216	1090962	50.064	ng	99
41) Hexachlorocyclopentadiene	12.575	237	1406851	106.306	ng	99
43) 2,4,6-Trichlorophenol	12.828	196	751627	52.450	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050264.D
 Acq On : 10 Jun 2025 06:07
 Operator : RC/JU
 Sample : PB168352BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB168352BS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 10 06:53:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.892	196	831872	52.871	ng	98
46) 1,1'-Biphenyl	13.233	154	2818663	49.881	ng	99
47) 2-Chloronaphthalene	13.269	162	2178306	49.584	ng	99
48) 2-Nitroaniline	13.469	65	565550	58.499	ng	95
49) Acenaphthylene	14.116	152	3592939	50.565	ng	100
50) Dimethylphthalate	13.857	163	2653207	51.986	ng	99
51) 2,6-Dinitrotoluene	13.969	165	566954	55.084	ng	96
52) Acenaphthene	14.463	154	2142880	48.215	ng	100
53) 3-Nitroaniline	14.292	138	393184	34.276	ng	100
54) 2,4-Dinitrophenol	14.498	184	533234	89.355	ng	96
55) Dibenzofuran	14.798	168	3297110	50.708	ng	99
56) 4-Nitrophenol	14.604	139	1171665	120.006	ng	98
57) 2,4-Dinitrotoluene	14.751	165	799521	58.887	ng	96
58) Fluorene	15.445	166	2563768	51.480	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.021	232	729601m	53.596	ng	
60) Diethylphthalate	15.227	149	2700060	53.319	ng	99
61) 4-Chlorophenyl-phenyle...	15.445	204	1240228	51.528	ng	97
62) 4-Nitroaniline	15.457	138	602665	56.055	ng	99
63) Azobenzene	15.733	77	2693437	53.214	ng	98
65) 4,6-Dinitro-2-methylph...	15.516	198	396144	50.943	ng	98
66) n-Nitrosodiphenylamine	15.651	169	2304314	51.511	ng	99
67) 4-Bromophenyl-phenylether	16.333	248	787638	51.878	ng	98
68) Hexachlorobenzene	16.445	284	922265	51.992	ng	99
69) Atrazine	16.604	200	814129	57.117	ng	100
70) Pentachlorophenol	16.786	266	1318637	114.339	ng	99
71) Phenanthrene	17.180	178	4203765	51.686	ng	99
72) Anthracene	17.268	178	4262245	52.384	ng	99
73) Carbazole	17.533	167	4035831	54.446	ng	99
74) Di-n-butylphthalate	18.115	149	4624100	57.333	ng	99
75) Fluoranthene	19.192	202	4747982	55.356	ng	100
77) Benzidine	19.374	184	1720352	40.229	ng	100
78) Pyrene	19.551	202	5054422	49.313	ng	100
80) Butylbenzylphthalate	20.462	149	2005247	58.652	ng	97
81) Benzo(a)anthracene	21.350	228	5042893	52.391	ng	100
82) 3,3'-Dichlorobenzidine	21.274	252	1108524	37.765	ng	100
83) Chrysene	21.409	228	4818888	52.631	ng	100
84) Bis(2-ethylhexyl)phtha...	21.298	149	3140586	59.636	ng	99
85) Di-n-octyl phthalate	22.427	149	5083521	65.052	ng	100
87) Indeno(1,2,3-cd)pyrene	27.703	276	6192744	58.001	ng	100
88) Benzo(b)fluoranthene	23.403	252	5189054	53.823	ng	100
89) Benzo(k)fluoranthene	23.468	252	5163179	52.460	ng	100
90) Benzo(a)pyrene	24.203	252	4948650	54.593	ng	99
91) Dibenzo(a,h)anthracene	27.762	278	5037803	58.130	ng	98
92) Benzo(g,h,i)perylene	28.744	276	5052425	58.247	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050264.D
 Acq On : 10 Jun 2025 06:07
 Operator : RC/JU
 Sample : PB168352BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

PB168352BS

Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025


Quant Time: Jun 10 06:53:09 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Jun 09 11:54:38 2025


Response via : Initial Calibration




Instrument :
BNA_M
ClientSampleId :
PB168352BS

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.763 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:152 Resp: 333914
Ion Ratio Lower Upper
152 100
150 159.0 122.2 183.4
115 60.6 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#2
1,4-Dioxane
Concen: 38.770 ng
RT: 3.281 min Scan# 50
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 88 Resp: 340181
Ion Ratio Lower Upper
88 100
58 57.2 42.9 64.3
43 15.2 11.0 16.6





#3
Pyridine
Concen: 43.620 ng
RT: 3.669 min Scan# 116
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 79 Resp: 99116
Ion Ratio Lower Upper
79 100
52 41.8 32.1 48.1
51 19.7 15.4 23.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#4
n-Nitrosodimethylamine
Concen: 48.006 ng
RT: 3.581 min Scan# 101
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 42 Resp: 225565
Ion Ratio Lower Upper
42 100
74 283.3 222.0 333.0
44 10.1 7.0 10.6





#5
2-Fluorophenol
Concen: 146.033 ng
RT: 5.357 min Scan# 405
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 112 Resp: 292315
Ion Ratio Lower Upper
112 100
64 50.6 39.8 59.6
63 25.1 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#6
Aniline
Concen: 36.214 ng
RT: 7.093 min Scan# 698
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 93 Resp: 1261529
Ion Ratio Lower Upper
93 100
66 35.9 27.4 41.2
65 18.0 13.6 20.4





#7
Phenol-d6
Concen: 140.063 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 99 Resp: 3690740
Ion Ratio Lower Upper
99 100
42 8.6 6.9 10.3
71 32.9 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#8
2-Chlorophenol
Concen: 50.515 ng
RT: 7.334 min Scan# 739
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:128 Resp: 1112221
Ion Ratio Lower Upper
128 100
130 32.8 12.6 52.6
64 41.5 19.8 59.8





#9
Benzaldehyde
Concen: 33.917 ng
RT: 6.904 min Scan# 60
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 77 Resp: 568218
Ion Ratio Lower Upper
77 100
105 100.1 80.9 120.9
106 97.0 77.1 117.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#10
Phenol
Concen: 50.671 ng
RT: 6.963 min Scan# 676
Delta R.T. 0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 94 Resp: 1412735
Ion Ratio Lower Upper
94 100
65 24.5 4.3 44.3
66 33.3 12.9 52.9





#11
bis(2-Chloroethyl)ether
Concen: 48.907 ng
RT: 7.193 min Scan# 715
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 93 Resp: 1096770
Ion Ratio Lower Upper
93 100
63 66.1 43.9 83.9
95 32.4 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#12
1,3-Dichlorobenzene
Concen: 47.580 ng
RT: 7.657 min Scan# 794
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07

Tgt Ion:146 Resp: 1171169
Ion Ratio Lower Upper
146 100
148 64.5 51.9 77.9
75 31.6 23.4 35.0









#13
1,4-Dichlorobenzene
Concen: 46.892 ng
RT: 7.798 min Scan# 811
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07

Instrument :
BNA_M
ClientSampleId :
PB168352BS

Tgt Ion:146 Resp: 1200930
Ion Ratio Lower Upper
146 100
148 63.4 51.4 77.0
111 41.4 32.3 48.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#14
1,2-Dichlorobenzene
Concen: 47.139 ng
RT: 8.116 min Scan# 872
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:146 Resp: 1150887
Ion Ratio Lower Upper
146 100
148 64.9 50.4 75.6
111 43.1 34.3 51.5





#15
Benzyl Alcohol
Concen: 49.118 ng
RT: 8.004 min Scan# 856
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 79 Resp: 923134
Ion Ratio Lower Upper
79 100
108 78.9 64.0 96.0
77 63.9 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 52.545 ng
RT: 8.293 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 45 Resp: 818561
Ion Ratio Lower Upper
45 100
77 27.5 8.7 48.7
79 20.5 2.7 42.7





#17
2-Methylphenol
Concen: 49.189 ng
RT: 8.204 min Scan# 890
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:107 Resp: 90489
Ion Ratio Lower Upper
107 100
108 116.1 90.8 136.2
77 46.8 35.7 53.5
79 46.2 35.0 52.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#18
Hexachloroethane
Concen: 49.468 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:117 Resp: 458802
Ion Ratio Lower Upper
117 100
119 95.7 77.0 115.6
201 90.9 74.6 111.8





#19
n-Nitroso-di-n-propylamine
Concen: 48.746 ng
RT: 8.575 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion: 70 Resp: 786038
Ion Ratio Lower Upper
70 100
42 28.4 22.6 33.8
101 9.6 8.3 12.5
130 20.6 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#20
3+4-Methylphenols
Concen: 49.586 ng
RT: 8.534 min Scan# 943
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Tgt Ion:107 Resp: 1220427
Ion Ratio Lower Upper
107 100
108 91.3 70.4 110.4
77 32.8 12.2 52.2
79 29.1 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.551 min Scan# 1286
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:136 Resp: 131154
Ion Ratio Lower Upper
136 100
137 11.2 8.6 13.0
54 5.0 3.8 5.8
68 6.0 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#22
Acetophenone
Concen: 49.228 ng
RT: 8.581 min Scan# 951
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:105 Resp: 1613029
Ion Ratio Lower Upper
105 100
71 6.3 5.0 7.4
51 14.7 11.4 17.2
120 23.1 18.2 27.4





#23
Nitrobenzene-d5
Concen: 88.551 ng
RT: 8.916 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 82 Resp: 223308
Ion Ratio Lower Upper
82 100
128 43.0 35.2 52.8
54 30.9 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#24
Nitrobenzene
Concen: 51.308 ng
RT: 8.957 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 77 Resp: 1176763
Ion Ratio Lower Upper
77 100
123 45.7 38.2 57.2
65 11.9 9.5 14.3





#25
Isophorone
Concen: 49.135 ng
RT: 9.487 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 82 Resp: 213864
Ion Ratio Lower Upper
82 100
95 7.0 5.7 8.5
138 17.6 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#26
2-Nitrophenol
Concen: 50.542 ng
RT: 9.663 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:139 Resp: 500374
Ion Ratio Lower Upper
139 100
109 29.6 23.3 34.9
65 48.0 35.2 52.8





#27
2,4-Dimethylphenol
Concen: 50.488 ng
RT: 9.734 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:122 Resp: 102702
Ion Ratio Lower Upper
122 100
107 114.4 91.0 136.6
121 57.3 45.8 68.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#28
bis(2-Chloroethoxy)methane
Concen: 50.637 ng
RT: 9.969 min Scan# 1187
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion: 93 Resp: 1439847
Ion Ratio Lower Upper
93 100
95 32.6 26.2 39.4
123 11.0 8.9 13.3





#29
2,4-Dichlorophenol
Concen: 51.435 ng
RT: 10.198 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:162 Resp: 96822
Ion Ratio Lower Upper
162 100
164 63.3 45.0 85.0
98 37.8 16.3 56.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#30
1,2,4-Trichlorobenzene
Concen: 49.001 ng
RT: 10.416 min Scan# 1263
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:180 Resp: 1026445
Ion Ratio Lower Upper
180 100
182 94.6 76.6 114.8
145 30.8 24.2 36.4





#31
Naphthalene
Concen: 47.982 ng
RT: 10.604 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:128 Resp: 3253648
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.6 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#32
Benzoic acid
Concen: 49.525 ng
RT: 9.857 min Scan# 1168
Delta R.T. 0.012 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:122 Resp: 633129
Ion Ratio Lower Upper
122 100
105 122.7 103.4 143.4
77 91.2 70.3 110.3





#33
4-Chloroaniline
Concen: 26.347 ng
RT: 10.704 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:127 Resp: 76170
Ion Ratio Lower Upper
127 100
129 31.9 25.4 38.0
65 27.9 21.0 31.4
92 18.5 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#34
Hexachlorobutadiene
Concen: 48.203 ng
RT: 10.898 min Scan# 1345
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:225 Resp: 600603
Ion Ratio Lower Upper
225 100
223 63.9 50.2 75.2
227 64.0 51.0 76.4





#35
Caprolactam
Concen: 50.122 ng m
RT: 11.481 min Scan# 1447
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion:113 Resp: 29912
Ion Ratio Lower Upper
113 100
55 117.9 83.5 123.5
56 91.3 60.6 100.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#36
4-Chloro-3-methylphenol
Concen: 51.040 ng
RT: 11.833 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Tgt Ion:107 Resp: 1025222
Ion Ratio Lower Upper
107 100
144 26.0 21.0 31.6
142 80.7 64.8 97.2





#37
2-Methylnaphthalene
Concen: 49.289 ng
RT: 12.216 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion:142 Resp: 2016294
Ion Ratio Lower Upper
142 100
141 87.8 71.5 107.3
115 34.8 27.8 41.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#38
1-Methylnaphthalene
Concen: 48.717 ng
RT: 12.439 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Tgt Ion:142 Resp: 2115216
Ion Ratio Lower Upper
142 100
141 90.0 72.9 109.3
116 3.8 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:164 Resp: 754658
Ion Ratio Lower Upper
164 100
162 100.4 80.2 120.4
160 45.2 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.064 ng
RT: 12.586 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:216 Resp: 1090962
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.4
179 20.4 16.6 25.0
108 22.3 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 106.306 ng
RT: 12.575 min Scan# 1630
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion:237 Resp: 140685
Ion Ratio Lower Upper
237 100
235 63.5 44.6 84.6
272 12.7 0.0 33.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#42
2,4,6-Tribromophenol
Concen: 153.304 ng
RT: 15.880 min Scan# 2192
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Tgt Ion:330 Resp: 1315918
Ion Ratio Lower Upper
330 100
332 96.3 77.5 116.3
141 35.8 28.8 43.2





#43
2,4,6-Trichlorophenol
Concen: 52.450 ng
RT: 12.828 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion:196 Resp: 75162
Ion Ratio Lower Upper
196 100
198 96.9 76.6 115.0
200 30.7 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#44
2,4,5-Trichlorophenol
Concen: 52.871 ng
RT: 12.892 min Scan# 1684
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Tgt Ion:196 Resp: 831872
Ion Ratio Lower Upper
196 100
198 95.9 76.1 114.1
97 52.6 39.3 58.9
132 28.3 21.4 32.0





#45
2-Fluorobiphenyl
Concen: 86.060 ng
RT: 13.022 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion:172 Resp: 4797135
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.6
170 23.3 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#46
1,1'-Biphenyl
Concen: 49.881 ng
RT: 13.233 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Tgt Ion:154 Resp: 2818663
Ion Ratio Lower Upper
154 100
153 39.6 19.2 59.2
76 12.9 0.0 32.0





#47
2-Chloronaphthalene
Concen: 49.584 ng
RT: 13.269 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:162 Resp: 2178300
Ion Ratio Lower Upper
162 100
127 38.6 31.1 46.7
164 32.2 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#48
2-Nitroaniline
Concen: 58.499 ng
RT: 13.469 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07

Tgt Ion: 65 Resp: 565550
Ion Ratio Lower Upper
65 100
92 76.3 64.2 96.4
138 119.6 101.0 151.4





#49
Acenaphthylene
Concen: 50.565 ng
RT: 14.116 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07

Instrument :

BNA_M

ClientSampleId :





PB168352BS

Tgt Ion:152 Resp: 3592939
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.1 10.4 15.6

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#50
Dimethylphthalate
Concen: 51.986 ng
RT: 13.857 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:163 Resp: 2653207
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.2 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 55.084 ng
RT: 13.969 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:165 Resp: 566954
Ion Ratio Lower Upper
165 100
63 48.0 35.8 53.6
89 58.1 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#52
Acenaphthene
Concen: 48.215 ng
RT: 14.463 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:154 Resp: 2142880
Ion Ratio Lower Upper
154 100
153 110.9 88.7 133.1
152 53.4 42.6 63.8





#53
3-Nitroaniline
Concen: 34.276 ng
RT: 14.292 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:138 Resp: 393184
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 115.7 92.9 139.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#54
2,4-Dinitrophenol
Concen: 89.355 ng
RT: 14.498 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:184 Resp: 533234
Ion Ratio Lower Upper
184 100
63 55.7 45.3 67.9
154 63.2 54.3 81.5





#55
Dibenzofuran
Concen: 50.708 ng
RT: 14.798 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:168 Resp: 3297110
Ion Ratio Lower Upper
168 100
139 36.6 29.9 44.9
169 13.0 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#56
4-Nitrophenol
Concen: 120.006 ng
RT: 14.604 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:139 Resp: 1171665
Ion Ratio Lower Upper
139 100
109 73.8 52.6 92.6
65 82.8 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 58.887 ng
RT: 14.751 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:165 Resp: 79952
Ion Ratio Lower Upper
165 100
63 38.5 28.4 42.6
89 75.1 58.0 87.0
182 2.6 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#58
Fluorene
Concen: 51.480 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:166 Resp: 2563768
Ion Ratio Lower Upper
166 100
165 96.9 77.0 115.4
167 13.2 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 53.596 ng m
RT: 15.021 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:232 Resp: 72960
Ion Ratio Lower Upper
232 100
131 41.9 32.9 49.3
130 2.4 1.8 2.8
166 25.9 20.8 31.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#60
Diethylphthalate
Concen: 53.319 ng
RT: 15.227 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:149 Resp: 2700060
Ion Ratio Lower Upper
149 100
177 21.7 17.2 25.8
150 12.1 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 51.528 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:204 Resp: 1240228
Ion Ratio Lower Upper
204 100
206 32.7 25.4 38.0
141 56.7 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#62
4-Nitroaniline
Concen: 56.055 ng
RT: 15.457 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:138 Resp: 602665
Ion Ratio Lower Upper
138 100
92 51.0 28.8 68.8
108 78.0 58.1 98.1





#63
Azobenzene
Concen: 53.214 ng
RT: 15.733 min Scan# 2171
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion: 77 Resp: 269343
Ion Ratio Lower Upper
77 100
182 24.5 6.8 46.8
105 21.6 1.7 41.7
51 16.2 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:188 Resp: 1423727
Ion Ratio Lower Upper
188 100
94 9.8 8.1 12.1
80 11.1 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 50.943 ng
RT: 15.516 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:198 Resp: 396144
Ion Ratio Lower Upper
198 100
51 24.1 4.5 44.5
105 46.4 24.9 64.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#66
n-Nitrosodiphenylamine
Concen: 51.511 ng
RT: 15.651 min Scan# 2153
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:169 Resp: 2304314
Ion Ratio Lower Upper
169 100
168 67.3 53.1 79.7
167 35.8 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 51.878 ng
RT: 16.333 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:248 Resp: 787638
Ion Ratio Lower Upper
248 100
250 99.0 78.0 117.0
141 73.4 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#68
Hexachlorobenzene
Concen: 51.992 ng
RT: 16.445 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:284 Resp: 922265
Ion Ratio Lower Upper
284 100
142 36.3 28.8 43.2
249 31.5 25.8 38.6





#69
Atrazine
Concen: 57.117 ng
RT: 16.604 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:200 Resp: 814129
Ion Ratio Lower Upper
200 100
173 26.1 6.0 46.0
215 47.7 28.1 68.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#70
Pentachlorophenol
Concen: 114.339 ng
RT: 16.786 min Scan# 2346
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:266 Resp: 1318637
Ion Ratio Lower Upper
266 100
268 64.4 51.0 76.6
264 63.1 50.8 76.2





#71
Phenanthrene
Concen: 51.686 ng
RT: 17.180 min Scan# 2416
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:178 Resp: 420376
Ion Ratio Lower Upper
178 100
176 19.1 15.6 23.4
179 15.2 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#72
Anthracene
Concen: 52.384 ng
RT: 17.268 min Scan# 2428
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:178 Resp: 4262245
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.5 12.3 18.5





#73
Carbazole
Concen: 54.446 ng
RT: 17.533 min Scan# 2476
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:167 Resp: 403583
Ion Ratio Lower Upper
167 100
166 21.0 17.0 25.6
139 12.6 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#74
Di-n-butylphthalate
Concen: 57.333 ng
RT: 18.115 min Scan# 2572
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:149 Resp: 4624100
Ion Ratio Lower Upper
149 100
150 8.9 7.4 11.0
104 5.0 4.0 6.0





#75
Fluoranthene
Concen: 55.356 ng
RT: 19.192 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:202 Resp: 474798
Ion Ratio Lower Upper
202 100
101 11.2 0.0 31.1
203 17.2 0.0 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:240 Resp: 1520709
Ion Ratio Lower Upper
240 100
120 11.2 9.0 13.4
236 25.0 20.7 31.1





#77
Benzidine
Concen: 40.229 ng
RT: 19.374 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:184 Resp: 1720352
Ion Ratio Lower Upper
184 100
185 14.7 11.5 17.3
183 11.2 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#78
Pyrene
Concen: 49.313 ng
RT: 19.551 min Scan# 2816
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:202 Resp: 5054422
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.4 14.1 21.1





#79
Terphenyl-d14
Concen: 86.841 ng
RT: 19.762 min Scan# 2852
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:244 Resp: 6968788
Ion Ratio Lower Upper
244 100
212 7.2 6.0 9.0
122 10.8 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#80
Butylbenzylphthalate
Concen: 58.652 ng
RT: 20.462 min Scan# 2971
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:149 Resp: 2005247
Ion Ratio Lower Upper
149 100
91 73.7 56.9 85.3
206 20.1 16.1 24.1





#81
Benzo(a)anthracene
Concen: 52.391 ng
RT: 21.350 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:228 Resp: 504289
Ion Ratio Lower Upper
228 100
226 27.3 21.7 32.5
229 19.8 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#82
3,3'-Dichlorobenzidine
Concen: 37.765 ng
RT: 21.274 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:252 Resp: 1108524
Ion Ratio Lower Upper
252 100
254 64.1 51.4 77.0
126 10.7 8.6 12.8





#83
Chrysene
Concen: 52.631 ng
RT: 21.409 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:228 Resp: 481888
Ion Ratio Lower Upper
228 100
226 30.5 24.3 36.5
229 19.8 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 59.636 ng
RT: 21.298 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:149 Resp: 3140586
Ion Ratio Lower Upper
149 100
167 27.5 22.3 33.5
279 4.4 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 65.052 ng
RT: 22.427 min Scan# 31
Delta R.T. -0.012 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:149 Resp: 508352
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.9 3.8 5.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.339 min Scan# 3630
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07


Tgt Ion:264 Resp: 1658221
Ion Ratio Lower Upper
264 100
260 23.2 18.6 28.0
265 21.1 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 58.001 ng
RT: 27.703 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS



Tgt Ion:276 Resp: 6192744
Ion Ratio Lower Upper
276 100
138 25.8 20.6 30.8
277 24.9 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#88
Benzo(b)fluoranthene
Concen: 53.823 ng
RT: 23.403 min Scan# 3471
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07

Tgt Ion:252 Resp: 5189054
Ion Ratio Lower Upper
252 100
253 21.5 17.2 25.8
125 10.3 8.2 12.4





#89

Benzo(k)fluoranthene

Concen: 52.460 ng

RT: 23.468 min Scan# 34

Delta R.T. -0.006 min

Lab File: BM050264.D

Acq: 10 Jun 2025 06:07

Instrument :

BNA_M

ClientSampleId :

PB168352BS

Tgt Ion:252 Resp: 5163179

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.2





125 10.3 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#90

Benzo(a)pyrene

Concen: 54.593 ng

RT: 24.203 min Scan# 3607

Delta R.T. -0.006 min

Lab File: BM050264.D

Acq: 10 Jun 2025 06:07




Tgt Ion:252 Resp: 4948650


Ion Ratio Lower Upper

252 100

253 22.3 17.6 26.4


125 11.7 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 58.130 ng
RT: 27.762 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07




Instrument :
BNA_M
ClientSampleId :
PB168352BS




Tgt Ion:278 Resp: 503780
Ion Ratio Lower Upper
278 100
139 16.4 12.8 19.2
279 24.3 18.6 28.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#92
Benzo(g,h,i)perylene
Concen: 58.247 ng
RT: 28.744 min Scan# 4379
Delta R.T. -0.006 min
Lab File: BM050264.D
Acq: 10 Jun 2025 06:07



Tgt Ion:276 Resp: 5052425
Ion Ratio Lower Upper
276 100
277 24.4 18.6 27.8
138 22.1 17.8 26.6



Report of Analysis

Client:	ENTACT	Date Collected:	06/02/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	TP06-MHI-WCMS	SDG No.:	Q2236
Lab Sample ID:	Q2226-04MS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050250.D	1	06/09/25 10:45	06/09/25 20:19	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	360		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	400		5.30	50.0	ug/L
95-48-7	2-Methylphenol	430		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	430		11.0	100	ug/L
67-72-1	Hexachloroethane	420		6.50	50.0	ug/L
98-95-3	Nitrobenzene	470		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	430		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	480		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	520		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	490		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	990	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	125		15 (23) - 110 (138)	83%	SPK: 150
13127-88-3	Phenol-d6	111		15 (10) - 110 (134)	74%	SPK: 150
4165-60-0	Nitrobenzene-d5	83.6		30 (67) - 130 (132)	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.0		30 (52) - 130 (132)	80%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		15 (44) - 110 (137)	88%	SPK: 150
1718-51-0	Terphenyl-d14	83.3		30 (42) - 130 (152)	83%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	393000	7.769			
1146-65-2	Naphthalene-d8	1520000	10.558			
15067-26-2	Acenaphthene-d10	827000	14.398			
1517-22-2	Phenanthrene-d10	1460000	17.139			
1719-03-5	Chrysene-d12	1330000	21.369			
1520-96-3	Perylene-d12	1380000	24.345			

Report of Analysis

Client:	ENTACT		Date Collected:	06/02/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	TP06-MHI-WCMS		SDG No.:	Q2236	
Lab Sample ID:	Q2226-04MS		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050250.D	1	06/09/25 10:45	06/09/25 20:19	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050250.D
 Acq On : 09 Jun 2025 20:19
 Operator : RC/JU
 Sample : Q2226-04MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP06-MHI-WCMS

Manual Integrations
 APPROVED

Reviewed By :Rahul
 Chavli

06/10/2025

Supervised By :Jagrut

Upadhyay

06/10/2025

Quant Time: Jun 10 01:26:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	392950	20.000	ng	0.00
21) Naphthalene-d8	10.558	136	1521763	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	827375	20.000	ng	0.00
64) Phenanthrene-d10	17.139	188	1462387	20.000	ng	0.00
76) Chrysene-d12	21.369	240	1332936	20.000	ng	0.00
86) Perylene-d12	24.345	264	1379954	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	2937446	124.700	ng	0.00
7) Phenol-d6	6.934	99	3430486	110.627	ng	0.00
23) Nitrobenzene-d5	8.916	82	2445053	83.563	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	1247878	132.601	ng	0.00
45) 2-Fluorobiphenyl	13.028	172	4888299	79.988	ng	0.00
79) Terphenyl-d14	19.763	244	5857343	83.273	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	388538	37.629	ng	# 73
3) Pyridine	3.682	79	966583	36.148	ng	96
4) n-Nitrosodimethylamine	3.587	42	245837	44.460	ng	97
6) Aniline	7.099	93	721210	17.593	ng	98
8) 2-Chlorophenol	7.334	128	1168140	45.083	ng	98
9) Benzaldehyde	6.905	77	453430	22.999	ng	96
10) Phenol	6.964	94	1296102	39.503	ng	99
11) bis(2-Chloroethyl)ether	7.193	93	1187468	44.996	ng	98
12) 1,3-Dichlorobenzene	7.658	146	1166230	40.261	ng	98
13) 1,4-Dichlorobenzene	7.805	146	1214259	40.289	ng	99
14) 1,2-Dichlorobenzene	8.122	146	1179906	41.067	ng	99
15) Benzyl Alcohol	8.005	79	994445	44.963	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.299	45	905020	49.367	ng	95
17) 2-Methylphenol	8.205	107	938045	43.331	ng	99
18) Hexachloroethane	8.852	117	461080	42.245	ng	99
19) n-Nitroso-di-n-propyla...	8.575	70	842877	44.418	ng	97
20) 3+4-Methylphenols	8.534	107	1243110	42.920	ng	99
22) Acetophenone	8.587	105	1692937	44.530	ng	# 99
24) Nitrobenzene	8.958	77	1256342	47.211	ng	97
25) Isophorone	9.487	82	2303628	45.614	ng	99
26) 2-Nitrophenol	9.669	139	547550	47.667	ng	97
27) 2,4-Dimethylphenol	9.734	122	1078632	45.700	ng	100
28) bis(2-Chloroethoxy)met...	9.969	93	1539341	46.658	ng	99
29) 2,4-Dichlorophenol	10.205	162	1002287	45.889	ng	99
30) 1,2,4-Trichlorobenzene	10.416	180	1063647	43.762	ng	100
31) Naphthalene	10.605	128	3430345	43.600	ng	100
32) Benzoic acid	9.858	122	630781	42.526	ng	97
33) 4-Chloroaniline	10.710	127	306819	9.147	ng	98
34) Hexachlorobutadiene	10.899	225	619791	42.871	ng	100
35) Caprolactam	11.481	113	238482	34.440	ng	90
36) 4-Chloro-3-methylphenol	11.834	107	1028377	44.125	ng	99
37) 2-Methylnaphthalene	12.216	142	2119896	44.663	ng	98
38) 1-Methylnaphthalene	12.440	142	2209544	43.860	ng	99
40) 1,2,4,5-Tetrachloroben...	12.587	216	1122982	47.004	ng	99
41) Hexachlorocyclopentadiene	12.581	237	1415330	97.547	ng	97
43) 2,4,6-Trichlorophenol	12.828	196	760650	48.414	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050250.D
 Acq On : 09 Jun 2025 20:19
 Operator : RC/JU
 Sample : Q2226-04MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

TP06-MHI-WCMS

Manual Integrations**APPROVED**

Reviewed By :Rahul

Chavli

06/10/2025

Supervised By :Jagrut

Upadhyay

06/10/2025

Quant Time: Jun 10 01:26:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.899	196	832691	48.272	ng	98
46) 1,1'-Biphenyl	13.234	154	2928102	47.263	ng	99
47) 2-Chloronaphthalene	13.275	162	2248213	46.678	ng	99
48) 2-Nitroaniline	13.469	65	562921	53.110	ng	94
49) Acenaphthylene	14.116	152	3621022	46.481	ng	100
50) Dimethylphthalate	13.863	163	2619856	46.821	ng	100
51) 2,6-Dinitrotoluene	13.969	165	557207	49.379	ng	96
52) Acenaphthene	14.463	154	2154317	44.212	ng	99
53) 3-Nitroaniline	14.293	138	339356	26.984	ng	100
54) 2,4-Dinitrophenol	14.498	184	609950	94.097	ng	95
55) Dibenzofuran	14.798	168	3304639	46.357	ng	100
56) 4-Nitrophenol	14.598	139	978143	91.379	ng	97
57) 2,4-Dinitrotoluene	14.751	165	767682	51.572	ng	97
58) Fluorene	15.445	166	2523640	46.221	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.022	232	700206m	46.916	ng	
60) Diethylphthalate	15.228	149	2611641	47.041	ng	99
61) 4-Chlorophenyl-phenyle...	15.445	204	1219957	46.231	ng	98
62) 4-Nitroaniline	15.457	138	537036	45.560	ng	99
63) Azobenzene	15.734	77	2612861	47.085	ng	98
65) 4,6-Dinitro-2-methylph...	15.516	198	385706	48.290	ng	99
66) n-Nitrosodiphenylamine	15.657	169	2213402	48.171	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	768128	49.256	ng	100
68) Hexachlorobenzene	16.445	284	884845	48.564	ng	99
69) Atrazine	16.610	200	734379	50.160	ng	100
70) Pentachlorophenol	16.787	266	1177160	99.373	ng	99
71) Phenanthrene	17.181	178	3861939	46.228	ng	99
72) Anthracene	17.269	178	3937922	47.119	ng	100
73) Carbazole	17.534	167	3576865	46.979	ng	100
74) Di-n-butylphthalate	18.116	149	4215930	50.890	ng	100
75) Fluoranthene	19.192	202	4091775	46.445	ng	99
77) Benzidine	19.375	184	358876	9.574	ng	99
78) Pyrene	19.557	202	4304284	47.910	ng	99
80) Butylbenzylphthalate	20.469	149	1632985	54.492	ng	99
81) Benzo(a)anthracene	21.351	228	3990464	47.298	ng	100
82) 3,3'-Dichlorobenzidine	21.275	252	824418	32.043	ng	99
83) Chrysene	21.416	228	3797384	47.317	ng	99
84) Bis(2-ethylhexyl)phtha...	21.298	149	2525715	54.716	ng	99
85) Di-n-octyl phthalate	22.433	149	3901373	56.957	ng	100
87) Indeno(1,2,3-cd)pyrene	27.703	276	4771886	53.706	ng	# 94
88) Benzo(b)fluoranthene	23.410	252	3868153	48.213	ng	99
89) Benzo(k)fluoranthene	23.474	252	3920735	47.870	ng	100
90) Benzo(a)pyrene	24.204	252	3723908	49.366	ng	100
91) Dibenzo(a,h)anthracene	27.768	278	3892102	53.966	ng	99
92) Benzo(g,h,i)perylene	28.745	276	3923152	54.349	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050250.D
 Acq On : 09 Jun 2025 20:19
 Operator : RC/JU
 Sample : Q2226-04MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

TP06-MHI-WCMS

Manual Integrations**APPROVED**

Reviewed By :Rahul

Chavli


06/10/2025

Supervised By :Jagrut

Upadhyay

06/10/2025

Quant Time: Jun 10 01:26:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050250.D
Acq On : 09 Jun 2025 20:19
Operator : RC/JU
Sample : Q2226-04MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli


06/10/2025


Supervised By :Jagrut

Upadhyay

06/10/2025


Quant Time: Jun 10 01:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. 0.000 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:152 Resp: 392950
Ion Ratio Lower Upper
152 100
150 156.6 122.2 183.4
115 59.5 46.3 69.5

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#2
1,4-Dioxane
Concen: 37.629 ng
RT: 3.293 min Scan# 52
Delta R.T. 0.012 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion: 88 Resp: 388538
Ion Ratio Lower Upper
88 100
58 58.7 42.9 64.3
43 56.7 11.0 16.6





#3
Pyridine
Concen: 36.148 ng
RT: 3.682 min Scan# 118
Delta R.T. 0.012 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 79 Resp: 96658
Ion Ratio Lower Upper
79 100
52 43.6 32.1 48.1
51 20.0 15.4 23.0

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#4
n-Nitrosodimethylamine
Concen: 44.460 ng
RT: 3.587 min Scan# 102
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Tgt Ion: 42 Resp: 245837
Ion Ratio Lower Upper
42 100
74 282.9 222.0 333.0
44 10.4 7.0 10.6





#5
2-Fluorophenol
Concen: 124.700 ng
RT: 5.358 min Scan# 405
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19





Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS

Tgt Ion:112 Resp: 2937440
Ion Ratio Lower Upper
112 100
64 50.7 39.8 59.6
63 25.6 19.8 29.8

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#6
Aniline
Concen: 17.593 ng
RT: 7.099 min Scan# 699
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion: 93 Resp: 721210
Ion Ratio Lower Upper
93 100
66 35.5 27.4 41.2
65 17.8 13.6 20.4





#7
Phenol-d6
Concen: 110.627 ng
RT: 6.934 min Scan# 61
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 99 Resp: 343048
Ion Ratio Lower Upper
99 100
42 8.6 6.9 10.3
71 32.5 25.3 37.9

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#8
2-Chlorophenol
Concen: 45.083 ng
RT: 7.334 min Scan# 739
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:128 Resp: 1168140
Ion Ratio Lower Upper
128 100
130 32.4 12.6 52.6
64 41.5 19.8 59.8





#9
Benzaldehyde
Concen: 22.999 ng
RT: 6.905 min Scan# 606
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 77 Resp: 453430
Ion Ratio Lower Upper
77 100
105 96.4 80.9 120.9
106 94.4 77.1 117.1

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#10
Phenol
Concen: 39.503 ng
RT: 6.964 min Scan# 676
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion: 94 Resp: 1296102
Ion Ratio Lower Upper
94 100
65 25.0 4.3 44.3
66 32.4 12.9 52.9





#11
bis(2-Chloroethyl)ether
Concen: 44.996 ng
RT: 7.193 min Scan# 718
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 93 Resp: 118746
Ion Ratio Lower Upper
93 100
63 66.6 43.9 83.9
95 32.4 12.7 52.7

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#12
1,3-Dichlorobenzene
Concen: 40.261 ng
RT: 7.658 min Scan# 794
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Tgt Ion:146 Resp: 1166230
Ion Ratio Lower Upper
146 100
148 63.3 51.9 77.9
75 31.1 23.4 35.0





#13

1,4-Dichlorobenzene

Concen: 40.289 ng

RT: 7.805 min Scan# 81

Delta R.T. 0.001 min

Lab File: BM050250.D

Acq: 09 Jun 2025 20:19

Instrument :

BNA_M

ClientSampleId :

TP06-MHI-WCMS

Tgt Ion:146 Resp: 1214259

Ion Ratio Lower Upper

146 100

148 64.0 51.4 77.0

111 39.8 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :Rahul





Chavli

06/10/2025

Supervised By :Jagrut

Upadhyay

06/10/2025



#14

1,2-Dichlorobenzene

Concen: 41.067 ng

RT: 8.122 min Scan# 873

Delta R.T. 0.001 min

Lab File: BM050250.D

Acq: 09 Jun 2025 20:19




Tgt Ion:146 Resp: 1179906


Ion Ratio Lower Upper

146 100

148 63.8 50.4 75.6


111 42.1 34.3 51.5





#15
Benzyl Alcohol
Concen: 44.963 ng
RT: 8.005 min Scan# 801
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 79 Resp: 994445
Ion Ratio Lower Upper
79 100
108 77.6 64.0 96.0
77 63.2 52.2 78.2

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 49.367 ng
RT: 8.299 min Scan# 903
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Tgt Ion: 45 Resp: 905020
Ion Ratio Lower Upper
45 100
77 26.4 8.7 48.7
79 20.0 2.7 42.7





#17

2-Methylphenol

Concen: 43.331 ng

RT: 8.205 min Scan# 81

Delta R.T. 0.001 min

Lab File: BM050250.D

Acq: 09 Jun 2025 20:19

Instrument :

BNA_M

ClientSampleId :

TP06-MHI-WCMS

Tgt Ion:107 Resp: 93804

Ion Ratio Lower Upper

107 100

108 114.1 90.8 136.2

77 45.7 35.7 53.5

79 43.7 35.0 52.4

Manual Integrations

APPROVED

Reviewed By :Rahul





Chavli

06/10/2025

Supervised By :Jagrut

Upadhyay

06/10/2025



#18

Hexachloroethane

Concen: 42.245 ng

RT: 8.852 min Scan# 997

Delta R.T. 0.001 min

Lab File: BM050250.D

Acq: 09 Jun 2025 20:19




Tgt Ion:117 Resp: 461080


Ion Ratio Lower Upper

117 100

119 95.6 77.0 115.6


201 91.5 74.6 111.8





#19
n-Nitroso-di-n-propylamine
Concen: 44.418 ng
RT: 8.575 min Scan# 91
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS



Tgt Ion: 70 Resp: 84287
Ion Ratio Lower Upper
70 100
42 27.3 22.6 33.8
101 9.0 8.3 12.5
130 20.1 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
06/10/2025
Supervised By :Jagrut
Upadhyay
06/10/2025



#20
3+4-Methylphenols
Concen: 42.920 ng
RT: 8.534 min Scan# 943
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:107 Resp: 1243110
Ion Ratio Lower Upper
107 100
108 90.7 70.4 110.4
77 32.7 12.2 52.2
79 28.9 7.7 47.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.558 min Scan# 1287
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS



Tgt Ion:136 Resp: 152176
Ion Ratio Lower Upper
136 100
137 10.6 8.6 13.0
54 5.0 3.8 5.8
68 6.2 4.9 7.3




Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
06/10/2025
Supervised By :Jagrut
Upadhyay
06/10/2025



#22
Acetophenone
Concen: 44.530 ng
RT: 8.587 min Scan# 952
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:105 Resp: 1692937
Ion Ratio Lower Upper
105 100
71 4.5 5.0 7.4#
51 14.6 11.4 17.2
120 22.9 18.2 27.4





#23
Nitrobenzene-d5
Concen: 83.563 ng
RT: 8.916 min Scan# 1008
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 82 Resp: 244505
Ion Ratio Lower Upper
82 100
128 42.9 35.2 52.8
54 32.0 24.5 36.7

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#24
Nitrobenzene
Concen: 47.211 ng
RT: 8.958 min Scan# 1015
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion: 77 Resp: 1256342
Ion Ratio Lower Upper
77 100
123 45.4 38.2 57.2
65 12.3 9.5 14.3





#25
Isophorone
Concen: 45.614 ng
RT: 9.487 min Scan# 1105
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion: 82 Resp: 2303628
Ion Ratio Lower Upper
82 100
95 6.9 5.7 8.5
138 17.2 14.2 21.2

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#26
2-Nitrophenol
Concen: 47.667 ng
RT: 9.669 min Scan# 1136
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:139 Resp: 547550
Ion Ratio Lower Upper
139 100
109 29.3 23.3 34.9
65 46.8 35.2 52.8





#27
2,4-Dimethylphenol
Concen: 45.700 ng
RT: 9.734 min Scan# 1150
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:122 Resp: 107863
Ion Ratio Lower Upper
122 100
107 113.5 91.0 136.6
121 57.7 45.8 68.6

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#28
bis(2-Chloroethoxy)methane
Concen: 46.658 ng
RT: 9.969 min Scan# 1187
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion: 93 Resp: 1539341
Ion Ratio Lower Upper
93 100
95 32.2 26.2 39.4
123 11.0 8.9 13.3





#29
2,4-Dichlorophenol
Concen: 45.889 ng
RT: 10.205 min Scan# 1227
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:162 Resp: 100228
Ion Ratio Lower Upper
162 100
164 64.6 45.0 85.0
98 37.4 16.3 56.3

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#30
1,2,4-Trichlorobenzene
Concen: 43.762 ng
RT: 10.416 min Scan# 1263
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:180 Resp: 1063647
Ion Ratio Lower Upper
180 100
182 95.3 76.6 114.8
145 30.6 24.2 36.4





#31
Naphthalene
Concen: 43.600 ng
RT: 10.605 min Scan# 1168
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:128 Resp: 3430349
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.9 10.2 15.4

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#32
Benzoic acid
Concen: 42.526 ng
RT: 9.858 min Scan# 1168
Delta R.T. 0.012 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:122 Resp: 630781
Ion Ratio Lower Upper
122 100
105 125.8 103.4 143.4
77 94.5 70.3 110.3





#33
4-Chloroaniline
Concen: 9.147 ng
RT: 10.710 min Scan# 1313
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS



Tgt Ion:127 Resp: 306819
Ion Ratio Lower Upper
127 100
129 29.9 25.4 38.0
65 26.7 21.0 31.4
92 17.7 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
06/10/2025
Supervised By :Jagrut
Upadhyay
06/10/2025



#34
Hexachlorobutadiene
Concen: 42.871 ng
RT: 10.899 min Scan# 1345
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:225 Resp: 619791
Ion Ratio Lower Upper
225 100
223 62.5 50.2 75.2
227 63.2 51.0 76.4





#35
Caprolactam
Concen: 34.440 ng
RT: 11.481 min Scan# 1447
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:113 Resp: 23848
Ion Ratio Lower Upper
113 100
55 113.7 83.5 123.5
56 89.2 60.6 100.6

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#36
4-Chloro-3-methylphenol
Concen: 44.125 ng
RT: 11.834 min Scan# 1504
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:107 Resp: 1028377
Ion Ratio Lower Upper
107 100
144 25.6 21.0 31.6
142 80.0 64.8 97.2





#37
2-Methylnaphthalene
Concen: 44.663 ng
RT: 12.216 min Scan# 1569
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:142 Resp: 2119890
Ion Ratio Lower Upper
142 100
141 87.0 71.5 107.3
115 34.7 27.8 41.8

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#38
1-Methylnaphthalene
Concen: 43.860 ng
RT: 12.440 min Scan# 1607
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:142 Resp: 2209544
Ion Ratio Lower Upper
142 100
141 92.5 72.9 109.3
116 4.1 3.0 4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1943
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:164 Resp: 827375
Ion Ratio Lower Upper
164 100
162 101.7 80.2 120.4
160 45.4 35.7 53.5

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.004 ng
RT: 12.587 min Scan# 1632
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:216 Resp: 1122982
Ion Ratio Lower Upper
216 100
214 78.4 63.0 94.4
179 20.7 16.6 25.0
108 22.4 15.9 23.9





#41
Hexachlorocyclopentadiene
Concen: 97.547 ng
RT: 12.581 min Scan# 1631
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:237 Resp: 1415330
Ion Ratio Lower Upper
237 100
235 62.2 44.6 84.6
272 12.9 0.0 33.1


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
06/10/2025



Supervised By :Jagrit
Upadhyay
06/10/2025




#42
2,4,6-Tribromophenol
Concen: 132.601 ng
RT: 15.881 min Scan# 2192
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Tgt Ion:330 Resp: 1247878
Ion Ratio Lower Upper
330 100
332 96.3 77.5 116.3
141 36.5 28.8 43.2





#43
2,4,6-Trichlorophenol
Concen: 48.414 ng
RT: 12.828 min Scan# 1673
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:196 Resp: 760650
Ion Ratio Lower Upper
196 100
198 96.6 76.6 115.0
200 31.2 24.5 36.7


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
06/10/2025



Supervised By :Jagrut
Upadhyay
06/10/2025




#44
2,4,5-Trichlorophenol
Concen: 48.272 ng
RT: 12.899 min Scan# 1685
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Tgt Ion:196 Resp: 832691
Ion Ratio Lower Upper
196 100
198 96.8 76.1 114.1
97 51.7 39.3 58.9
132 27.6 21.4 32.0





#45
2-Fluorobiphenyl
Concen: 79.988 ng
RT: 13.028 min Scan# 1710
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:172 Resp: 4888295
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.6
170 23.1 18.6 27.8


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
06/10/2025



Supervised By :Jagrut
Upadhyay
06/10/2025




#46
1,1'-Biphenyl
Concen: 47.263 ng
RT: 13.234 min Scan# 1742
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Tgt Ion:154 Resp: 2928102
Ion Ratio Lower Upper
154 100
153 39.4 19.2 59.2
76 13.0 0.0 32.0





#47
2-Chloronaphthalene
Concen: 46.678 ng
RT: 13.275 min Scan# 1749
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:162 Resp: 224821
Ion Ratio Lower Upper
162 100
127 38.1 31.1 46.7
164 32.7 26.5 39.7

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#48
2-Nitroaniline
Concen: 53.110 ng
RT: 13.469 min Scan# 1782
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion: 65 Resp: 562921
Ion Ratio Lower Upper
65 100
92 76.8 64.2 96.4
138 118.2 101.0 151.4





#49
Acenaphthylene
Concen: 46.481 ng
RT: 14.116 min Scan# 1892
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:152 Resp: 362102
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.2 10.4 15.6

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#50
Dimethylphthalate
Concen: 46.821 ng
RT: 13.863 min Scan# 1849
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:163 Resp: 2619856
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 7.9 11.9





#51
2,6-Dinitrotoluene
Concen: 49.379 ng
RT: 13.969 min Scan# 1867
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:165 Resp: 55720
Ion Ratio Lower Upper
165 100
63 47.9 35.8 53.6
89 58.3 44.8 67.2

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#52
Acenaphthene
Concen: 44.212 ng
RT: 14.463 min Scan# 1951
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:154 Resp: 2154317
Ion Ratio Lower Upper
154 100
153 113.0 88.7 133.1
152 53.2 42.6 63.8





#53
3-Nitroaniline
Concen: 26.984 ng
RT: 14.293 min Scan# 1922
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:138 Resp: 339350
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 115.7 92.9 139.3

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#54
2,4-Dinitrophenol
Concen: 94.097 ng
RT: 14.498 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:184 Resp: 609950
Ion Ratio Lower Upper
184 100
63 55.5 45.3 67.9
154 62.0 54.3 81.5





#55
Dibenzofuran
Concen: 46.357 ng
RT: 14.798 min Scan# 2011
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:168 Resp: 3304639
Ion Ratio Lower Upper
168 100
139 37.2 29.9 44.9
169 13.4 10.6 16.0

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#56
4-Nitrophenol
Concen: 91.379 ng
RT: 14.598 min Scan# 1974
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:139 Resp: 978143
Ion Ratio Lower Upper
139 100
109 74.1 52.6 92.6
65 87.6 64.3 104.3





#57
2,4-Dinitrotoluene
Concen: 51.572 ng
RT: 14.751 min Scan# 2000
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS



Tgt Ion:165 Resp: 76768
Ion Ratio Lower Upper
165 100
63 38.7 28.4 42.6
89 74.2 58.0 87.0
182 2.5 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
06/10/2025
Supervised By :Jagrut
Upadhyay
06/10/2025



#58
Fluorene
Concen: 46.221 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:166 Resp: 2523640
Ion Ratio Lower Upper
166 100
165 96.8 77.0 115.4
167 13.6 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 46.916 ng m
RT: 15.022 min Scan# 2046
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:232 Resp: 700200
Ion Ratio Lower Upper
232 100
131 41.7 32.9 49.3
130 2.4 1.8 2.8
166 26.0 20.8 31.2

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#60
Diethylphthalate
Concen: 47.041 ng
RT: 15.228 min Scan# 2081
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:149 Resp: 2611641
Ion Ratio Lower Upper
149 100
177 21.9 17.2 25.8
150 12.4 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 46.231 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:204 Resp: 121995
Ion Ratio Lower Upper
204 100
206 32.8 25.4 38.0
141 58.9 47.9 71.9


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
06/10/2025



Supervised By :Jagrut
Upadhyay
06/10/2025




#62
4-Nitroaniline
Concen: 45.560 ng
RT: 15.457 min Scan# 2120
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Tgt Ion:138 Resp: 537036
Ion Ratio Lower Upper
138 100
92 49.7 28.8 68.8
108 76.9 58.1 98.1





#63
Azobenzene
Concen: 47.085 ng
RT: 15.734 min Scan# 2171
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS



Tgt Ion: 77 Resp: 261286
Ion Ratio Lower Upper
77 100
182 24.2 6.8 46.8
105 21.8 1.7 41.7
51 16.0 0.0 36.2




Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
06/10/2025
Supervised By :Jagrut
Upadhyay
06/10/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2406
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:188 Resp: 1462387
Ion Ratio Lower Upper
188 100
94 9.2 8.1 12.1
80 10.1 8.6 13.0





#65
4,6-Dinitro-2-methylphenol
Concen: 48.290 ng
RT: 15.516 min Scan# 2130
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:198 Resp: 385700
Ion Ratio Lower Upper
198 100
51 24.1 4.5 44.5
105 45.6 24.9 64.9

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#66
n-Nitrosodiphenylamine
Concen: 48.171 ng
RT: 15.657 min Scan# 2154
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:169 Resp: 2213402
Ion Ratio Lower Upper
169 100
168 66.2 53.1 79.7
167 35.7 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 49.256 ng
RT: 16.334 min Scan# 21
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:248 Resp: 768128
Ion Ratio Lower Upper
248 100
250 97.4 78.0 117.0
141 75.6 60.1 90.1


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
06/10/2025



Supervised By :Jagrut
Upadhyay
06/10/2025




#68
Hexachlorobenzene
Concen: 48.564 ng
RT: 16.445 min Scan# 2288
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Tgt Ion:284 Resp: 884845
Ion Ratio Lower Upper
284 100
142 36.3 28.8 43.2
249 31.2 25.8 38.6





#69
Atrazine
Concen: 50.160 ng
RT: 16.610 min Scan# 21
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:200 Resp: 734379
Ion Ratio Lower Upper
200 100
173 26.2 6.0 46.0
215 48.3 28.1 68.1

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#70
Pentachlorophenol
Concen: 99.373 ng
RT: 16.787 min Scan# 2346
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:266 Resp: 1177160
Ion Ratio Lower Upper
266 100
268 63.8 51.0 76.6
264 62.5 50.8 76.2





#71
Phenanthrene
Concen: 46.228 ng
RT: 17.181 min Scan# 2416
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:178 Resp: 3861939
Ion Ratio Lower Upper
178 100
176 19.1 15.6 23.4
179 15.3 12.3 18.5

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#72
Anthracene
Concen: 47.119 ng
RT: 17.269 min Scan# 2428
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:178 Resp: 3937922
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.3 12.3 18.5





#73
Carbazole
Concen: 46.979 ng
RT: 17.534 min Scan# 2476
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:167 Resp: 357686
Ion Ratio Lower Upper
167 100
166 21.1 17.0 25.6
139 12.6 10.2 15.4

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#74
Di-n-butylphthalate
Concen: 50.890 ng
RT: 18.116 min Scan# 2572
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:149 Resp: 4215930
Ion Ratio Lower Upper
149 100
150 9.1 7.4 11.0
104 5.1 4.0 6.0





#75
Fluoranthene
Concen: 46.445 ng
RT: 19.192 min Scan# 21
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:202 Resp: 409177
Ion Ratio Lower Upper
202 100
101 11.8 0.0 31.1
203 17.5 0.0 37.0

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.369 min Scan# 3125
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:240 Resp: 1332936
Ion Ratio Lower Upper
240 100
120 11.1 9.0 13.4
236 24.9 20.7 31.1





#77
Benzidine
Concen: 9.574 ng
RT: 19.375 min Scan# 21
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:184 Resp: 358870
Ion Ratio Lower Upper
184 100
185 14.3 11.5 17.3
183 10.8 9.0 13.4

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#78
Pyrene
Concen: 47.910 ng
RT: 19.557 min Scan# 2817
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:202 Resp: 4304284
Ion Ratio Lower Upper
202 100
200 20.5 16.7 25.1
203 17.3 14.1 21.1





#79
Terphenyl-d14
Concen: 83.273 ng
RT: 19.763 min Scan# 2852
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:244 Resp: 585734
Ion Ratio Lower Upper
244 100
212 7.2 6.0 9.0
122 11.3 8.6 12.8

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#80
Butylbenzylphthalate
Concen: 54.492 ng
RT: 20.469 min Scan# 2972
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:149 Resp: 1632985
Ion Ratio Lower Upper
149 100
91 70.4 56.9 85.3
206 20.9 16.1 24.1





#81
Benzo(a)anthracene
Concen: 47.298 ng
RT: 21.351 min Scan# 3109
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS



Tgt Ion:228 Resp: 3990464
Ion Ratio Lower Upper
228 100
226 27.1 21.7 32.5
229 19.7 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
06/10/2025
Supervised By :Jagrut
Upadhyay
06/10/2025



#82
3,3'-Dichlorobenzidine
Concen: 32.043 ng
RT: 21.275 min Scan# 3109
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:252 Resp: 824418
Ion Ratio Lower Upper
252 100
254 64.7 51.4 77.0
126 11.2 8.6 12.8





#83
Chrysene
Concen: 47.317 ng
RT: 21.416 min Scan# 3135
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:228 Resp: 3797384
Ion Ratio Lower Upper
228 100
226 29.9 24.3 36.5
229 19.5 15.8 23.8

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 54.716 ng
RT: 21.298 min Scan# 3113
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:149 Resp: 2525715
Ion Ratio Lower Upper
149 100
167 27.4 22.3 33.5
279 4.1 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 56.957 ng
RT: 22.433 min Scan# 31
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:149 Resp: 390137
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 5.0 3.8 5.8

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.345 min Scan# 3631
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:264 Resp: 1379954
Ion Ratio Lower Upper
264 100
260 23.0 18.6 28.0
265 21.4 16.8 25.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 53.706 ng
RT: 27.703 min Scan# 41
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS





Tgt Ion:276 Resp: 477188
Ion Ratio Lower Upper
276 100
138 19.8 20.6 30.8
277 25.0 19.8 29.6

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#88
Benzo(b)fluoranthene
Concen: 48.213 ng
RT: 23.410 min Scan# 3472
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:252 Resp: 3868153
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 10.3 8.2 12.4





#89
Benzo(k)fluoranthene
Concen: 47.870 ng
RT: 23.474 min Scan# 3483
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS





Tgt Ion:252 Resp: 3920735
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 10.2 8.2 12.4

Manual Integrations
APPROVED




Reviewed By :Rahul
Chavli
06/10/2025


Supervised By :Jagrut
Upadhyay
06/10/2025



#90
Benzo(a)pyrene
Concen: 49.366 ng
RT: 24.204 min Scan# 3607
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19


Tgt Ion:252 Resp: 3723908
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 11.3 9.0 13.4





#91
Dibenzo(a,h)anthracene
Concen: 53.966 ng
RT: 27.768 min Scan# 41
Delta R.T. 0.001 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMS






Tgt Ion:278 Resp: 389210
Ion Ratio Lower Upper
278 100
139 16.9 12.8 19.2
279 23.7 18.6 28.0


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
06/10/2025



Supervised By :Jagrut
Upadhyay
06/10/2025



#92
Benzo(g,h,i)perylene
Concen: 54.349 ng
RT: 28.745 min Scan# 4379
Delta R.T. -0.005 min
Lab File: BM050250.D
Acq: 09 Jun 2025 20:19



Tgt Ion:276 Resp: 3923152
Ion Ratio Lower Upper
276 100
277 23.3 18.6 27.8
138 22.4 17.8 26.6



Report of Analysis

Client:	ENTACT	Date Collected:	06/02/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	TP06-MHI-WCMSD	SDG No.:	Q2236
Lab Sample ID:	Q2226-04MSD	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050251.D	1	06/09/25 10:45	06/09/25 20:58	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	380		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	440		5.30	50.0	ug/L
95-48-7	2-Methylphenol	460		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	460		11.0	100	ug/L
67-72-1	Hexachloroethane	450		6.50	50.0	ug/L
98-95-3	Nitrobenzene	500		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	490		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	530		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	530		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	550		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	530		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	1100	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	132		15 (23) - 110 (138)	88%	SPK: 150
13127-88-3	Phenol-d6	118		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.0		30 (67) - 130 (132)	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	86.3		30 (52) - 130 (132)	86%	SPK: 100
118-79-6	2,4,6-Tribromophenol	143		15 (44) - 110 (137)	95%	SPK: 150
1718-51-0	Terphenyl-d14	90.5		30 (42) - 130 (152)	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	396000	7.769			
1146-65-2	Naphthalene-d8	1510000	10.557			
15067-26-2	Acenaphthene-d10	836000	14.398			
1517-22-2	Phenanthrene-d10	1450000	17.139			
1719-03-5	Chrysene-d12	1290000	21.368			
1520-96-3	Perylene-d12	1340000	24.344			

Report of Analysis

Client:	ENTACT		Date Collected:	06/02/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	TP06-MHI-WCMSD		SDG No.:	Q2236	
Lab Sample ID:	Q2226-04MSD		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050251.D	1	06/09/25 10:45	06/09/25 20:58	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050251.D
 Acq On : 09 Jun 2025 20:58
 Operator : RC/JU
 Sample : Q2226-04MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP06-MHI-WCMSD

Quant Time: Jun 10 01:27:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	395507	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	1511863	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	835521	20.000	ng	0.00
64) Phenanthrene-d10	17.139	188	1452446	20.000	ng	0.00
76) Chrysene-d12	21.368	240	1291557	20.000	ng	0.00
86) Perylene-d12	24.344	264	1337173	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.363	112	3126533	131.869	ng	0.00
7) Phenol-d6	6.940	99	3679931	117.904	ng	0.00
23) Nitrobenzene-d5	8.922	82	2614873	89.952	ng	0.00
42) 2,4,6-Tribromophenol	15.886	330	1359589	143.063	ng	0.00
45) 2-Fluorobiphenyl	13.027	172	5328381	86.339	ng	0.00
79) Terphenyl-d14	19.762	244	6171435	90.550	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	410805	39.528	ng	# 78
3) Pyridine	3.681	79	1012714	37.628	ng	99
4) n-Nitrosodimethylamine	3.587	42	254911	45.803	ng	96
6) Aniline	7.098	93	692920	16.794	ng	100
8) 2-Chlorophenol	7.334	128	1261327	48.365	ng	99
9) Benzaldehyde	6.910	77	470499	23.710	ng	99
10) Phenol	6.963	94	1374952	41.635	ng	99
11) bis(2-Chloroethyl)ether	7.193	93	1256020	47.286	ng	99
12) 1,3-Dichlorobenzene	7.657	146	1284181	44.046	ng	99
13) 1,4-Dichlorobenzene	7.804	146	1332501	43.927	ng	99
14) 1,2-Dichlorobenzene	8.122	146	1289747	44.600	ng	99
15) Benzyl Alcohol	8.004	79	1057318	47.497	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.298	45	867084	46.991	ng	99
17) 2-Methylphenol	8.210	107	1009984	46.352	ng	99
18) Hexachloroethane	8.851	117	492829	44.862	ng	98
19) n-Nitroso-di-n-propyla...	8.575	70	889624	46.578	ng	99
20) 3+4-Methylphenols	8.534	107	1333740	45.751	ng	100
22) Acetophenone	8.587	105	1836141	48.613	ng	# 99
24) Nitrobenzene	8.963	77	1324519	50.098	ng	100
25) Isophorone	9.492	82	2452735	48.885	ng	99
26) 2-Nitrophenol	9.669	139	608472	53.317	ng	99
27) 2,4-Dimethylphenol	9.734	122	1160854	49.506	ng	99
28) bis(2-Chloroethoxy)met...	9.969	93	1638720	49.995	ng	100
29) 2,4-Dichlorophenol	10.204	162	1109642	51.137	ng	99
30) 1,2,4-Trichlorobenzene	10.422	180	1167906	48.367	ng	98
31) Naphthalene	10.604	128	3699253	47.325	ng	100
32) Benzoic acid	9.863	122	703609	47.746	ng	99
33) 4-Chloroaniline	10.710	127	262265	7.870	ng	99
34) Hexachlorobutadiene	10.904	225	699158	48.678	ng	99
35) Caprolactam	11.486	113	262971	38.226	ng	99
36) 4-Chloro-3-methylphenol	11.833	107	1110198	47.948	ng	99
37) 2-Methylnaphthalene	12.216	142	2288847	48.538	ng	99
38) 1-Methylnaphthalene	12.439	142	2392878	47.810	ng	100
40) 1,2,4,5-Tetrachloroben...	12.592	216	1254185	51.984	ng	99
41) Hexachlorocyclopentadiene	12.580	237	1602258	109.354	ng	98
43) 2,4,6-Trichlorophenol	12.827	196	843116	53.140	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
 Data File : BM050251.D
 Acq On : 09 Jun 2025 20:58
 Operator : RC/JU
 Sample : Q2226-04MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP06-MHI-WCMSD

Quant Time: Jun 10 01:27:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 09 11:54:38 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.898	196	917115	52.647	ng	98
46) 1,1'-Biphenyl	13.233	154	3172114	50.703	ng	99
47) 2-Chloronaphthalene	13.275	162	2421623	49.788	ng	100
48) 2-Nitroaniline	13.469	65	595329	55.620	ng	96
49) Acenaphthylene	14.122	152	3967313	50.430	ng	99
50) Dimethylphthalate	13.863	163	2849131	50.422	ng	100
51) 2,6-Dinitrotoluene	13.969	165	612698	53.767	ng	98
52) Acenaphthene	14.463	154	2335159	47.456	ng	99
53) 3-Nitroaniline	14.292	138	344695	27.141	ng	100
54) 2,4-Dinitrophenol	14.498	184	691966	109.274	ng	95
55) Dibenzofuran	14.798	168	3575384	49.666	ng	100
56) 4-Nitrophenol	14.604	139	1047823	96.935	ng	97
57) 2,4-Dinitrotoluene	14.757	165	830530	55.251	ng	98
58) Fluorene	15.445	166	2702504	49.014	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.021	232	769866	51.081	ng	93
60) Diethylphthalate	15.227	149	2797659	49.900	ng	99
61) 4-Chlorophenyl-phenyle...	15.445	204	1333102	50.026	ng	97
62) 4-Nitroaniline	15.457	138	582061	48.899	ng	99
63) Azobenzene	15.739	77	2743738	48.962	ng	99
65) 4,6-Dinitro-2-methylph...	15.516	198	428217	53.979	ng	99
66) n-Nitrosodiphenylamine	15.657	169	2388396	52.335	ng	100
67) 4-Bromophenyl-phenylether	16.333	248	838429	54.132	ng	99
68) Hexachlorobenzene	16.451	284	957173	52.893	ng	97
69) Atrazine	16.610	200	795271	54.691	ng	98
70) Pentachlorophenol	16.786	266	1281055	108.884	ng	99
71) Phenanthrene	17.180	178	4147484	49.986	ng	99
72) Anthracene	17.268	178	4211491	50.737	ng	100
73) Carbazole	17.533	167	3792742	50.155	ng	100
74) Di-n-butylphthalate	18.115	149	4481884	54.471	ng	100
75) Fluoranthene	19.192	202	4355639	49.778	ng	99
77) Benzidine	19.374	184	441789	12.164	ng	99
78) Pyrene	19.556	202	4546225	52.225	ng	99
80) Butylbenzylphthalate	20.468	149	1735253	59.759	ng	98
81) Benzo(a)anthracene	21.350	228	4167289	50.976	ng	100
82) 3,3'-Dichlorobenzidine	21.274	252	826974	33.172	ng	99
83) Chrysene	21.409	228	3981460	51.200	ng	100
84) Bis(2-ethylhexyl)phtha...	21.298	149	2662705	59.532	ng	99
85) Di-n-octyl phthalate	22.433	149	4131361	62.247	ng	100
87) Indeno(1,2,3-cd)pyrene	27.703	276	5098501	59.218	ng	100
88) Benzo(b)fluoranthene	23.409	252	4049157	52.083	ng	99
89) Benzo(k)fluoranthene	23.474	252	4106648	51.744	ng	99
90) Benzo(a)pyrene	24.203	252	3915530	53.567	ng	100
91) Dibenzo(a,h)anthracene	27.768	278	4153403	59.431	ng	99
92) Benzo(g,h,i)perylene	28.744	276	4187458	59.866	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050251.D
Acq On : 09 Jun 2025 20:58
Operator : RC/JU
Sample : Q2226-04MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Quant Time: Jun 10 01:27:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050251.D
Acq On : 09 Jun 2025 20:58
Operator : RC/JU
Sample : Q2226-04MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD

Quant Time: Jun 10 01:27:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 817
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:152 Resp: 395507
Ion Ratio Lower Upper
152 100
150 157.9 122.2 183.4
115 58.2 46.3 69.5



#2
1,4-Dioxane
Concen: 39.528 ng
RT: 3.293 min Scan# 52
Delta R.T. 0.012 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 88 Resp: 410805
Ion Ratio Lower Upper
88 100
58 54.3 42.9 64.3
43 55.5 11.0 16.6









#3
Pyridine
Concen: 37.628 ng
RT: 3.681 min Scan# 118
Delta R.T. 0.012 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion: 79 Resp: 1012714
Ion Ratio Lower Upper
79 100
52 40.6 32.1 48.1
51 19.6 15.4 23.0



#4
n-Nitrosodimethylamine
Concen: 45.803 ng
RT: 3.587 min Scan# 102
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 42 Resp: 254911
Ion Ratio Lower Upper
42 100
74 270.2 222.0 333.0
44 9.4 7.0 10.6









#5
2-Fluorophenol
Concen: 131.869 ng
RT: 5.363 min Scan# 405
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:112 Resp: 3126533
Ion Ratio Lower Upper
112 100
64 47.4 39.8 59.6
63 23.9 19.8 29.8



#6
Aniline
Concen: 16.794 ng
RT: 7.098 min Scan# 699
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 93 Resp: 692920
Ion Ratio Lower Upper
93 100
66 34.0 27.4 41.2
65 16.9 13.6 20.4









#7
Phenol-d6
Concen: 117.904 ng
RT: 6.940 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion: 99 Resp: 3679931
Ion Ratio Lower Upper
99 100
42 8.9 6.9 10.3
71 31.5 25.3 37.9



#8
2-Chlorophenol
Concen: 48.365 ng
RT: 7.334 min Scan# 739
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 128 Resp: 1261327
Ion Ratio Lower Upper
128 100
130 32.8 12.6 52.6
64 40.9 19.8 59.8









#9
Benzaldehyde
Concen: 23.710 ng
RT: 6.910 min Scan# 60
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCM5D


Tgt Ion: 77 Resp: 470499
Ion Ratio Lower Upper
77 100
105 103.6 80.9 120.9
106 97.3 77.1 117.1



#10
Phenol
Concen: 41.635 ng
RT: 6.963 min Scan# 676
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 94 Resp: 1374952
Ion Ratio Lower Upper
94 100
65 23.9 4.3 44.3
66 32.5 12.9 52.9









#11
bis(2-Chloroethyl)ether
Concen: 47.286 ng
RT: 7.193 min Scan# 715
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion: 93 Resp: 1256020
Ion Ratio Lower Upper
93 100
63 63.1 43.9 83.9
95 31.8 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 44.046 ng
RT: 7.657 min Scan# 794
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:146 Resp: 1284181
Ion Ratio Lower Upper
146 100
148 64.2 51.9 77.9
75 29.8 23.4 35.0









#13
1,4-Dichlorobenzene
Concen: 43.927 ng
RT: 7.804 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:146 Resp: 1332501
Ion Ratio Lower Upper
146 100
148 62.8 51.4 77.0
111 40.3 32.3 48.5



#14
1,2-Dichlorobenzene
Concen: 44.600 ng
RT: 8.122 min Scan# 873
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:146 Resp: 1289747
Ion Ratio Lower Upper
146 100
148 63.9 50.4 75.6
111 42.2 34.3 51.5





#15

Benzyl Alcohol

Concen: 47.497 ng

RT: 8.004 min Scan# 80

Delta R.T. 0.000 min

Lab File: BM050251.D

Acq: 09 Jun 2025 20:58

Instrument :





BNA_M

ClientSampleId :

TP06-MHI-WCMSD

Tgt Ion: 79 Resp: 1057318

Ion	Ratio	Lower	Upper
79	100		
108	80.9	64.0	96.0
77	64.8	52.2	78.2



#16

2,2'-oxybis(1-Chloropropane)

Concen: 46.991 ng

RT: 8.298 min Scan# 903




Delta R.T. 0.006 min


Lab File: BM050251.D

Acq: 09 Jun 2025 20:58

Tgt Ion: 45 Resp: 867084

Ion	Ratio	Lower	Upper
45	100		
77	28.5	8.7	48.7
79	21.7	2.7	42.7





#17

2-Methylphenol

Concen: 46.352 ng

RT: 8.210 min Scan# 81

Delta R.T. 0.006 min

Lab File: BM050251.D


Acq: 09 Jun 2025 20:58

Instrument :

BNA_M



ClientSampleId :

TP06-MHI-WCMSD



Tgt Ion:107 Resp: 1009984

Ion	Ratio	Lower	Upper
107	100		
108	115.4	90.8	136.2
77	44.3	35.7	53.5
79	43.7	35.0	52.4



#18

Hexachloroethane


Concen: 44.862 ng

RT: 8.851 min Scan# 997

Delta R.T. 0.000 min



Lab File: BM050251.D


Acq: 09 Jun 2025 20:58



Tgt Ion:117 Resp: 492829

Ion	Ratio	Lower	Upper
117	100		
119	96.5	77.0	115.6
201	96.1	74.6	111.8









#19
n-Nitroso-di-n-propylamine
Concen: 46.578 ng
RT: 8.575 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion: 70 Resp: 889624
Ion Ratio Lower Upper
70 100
42 28.7 22.6 33.8
101 9.6 8.3 12.5
130 22.0 17.4 26.2



#20
3+4-Methylphenols
Concen: 45.751 ng
RT: 8.534 min Scan# 943
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 107 Resp: 1333740
Ion Ratio Lower Upper
107 100
108 90.2 70.4 110.4
77 31.8 12.2 52.2
79 27.3 7.7 47.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.557 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:136 Resp: 1511863
Ion Ratio Lower Upper
136 100
137 10.7 8.6 13.0
54 4.9 3.8 5.8
68 5.6 4.9 7.3



#22
Acetophenone
Concen: 48.613 ng
RT: 8.587 min Scan# 952
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:105 Resp: 1836141
Ion Ratio Lower Upper
105 100
71 4.4 5.0 7.4#
51 14.0 11.4 17.2
120 22.6 18.2 27.4









#23
Nitrobenzene-d5
Concen: 89.952 ng
RT: 8.922 min Scan# 1009
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion: 82 Resp: 2614873
Ion Ratio Lower Upper
82 100
128 44.2 35.2 52.8
54 30.4 24.5 36.7



#24
Nitrobenzene
Concen: 50.098 ng
RT: 8.963 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 77 Resp: 1324519
Ion Ratio Lower Upper
77 100
123 48.0 38.2 57.2
65 11.9 9.5 14.3









#25
Isophorone
Concen: 48.885 ng
RT: 9.492 min Scan# 1106
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion: 82 Resp: 2452735
Ion Ratio Lower Upper
82 100
95 6.7 5.7 8.5
138 17.9 14.2 21.2



#26
2-Nitrophenol
Concen: 53.317 ng
RT: 9.669 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:139 Resp: 608472
Ion Ratio Lower Upper
139 100
109 28.9 23.3 34.9
65 44.4 35.2 52.8





#27

2,4-Dimethylphenol

Concen: 49.506 ng

RT: 9.734 min Scan# 1150

Delta R.T. 0.000 min

Lab File: BM050251.D

Acq: 09 Jun 2025 20:58

Instrument :





BNA_M

ClientSampleId :

TP06-MHI-WCMSD

Tgt Ion:122 Resp: 1160854

Ion	Ratio	Lower	Upper
122	100		
107	114.9	91.0	136.6
121	57.7	45.8	68.6



#28

bis(2-Chloroethoxy)methane

Concen: 49.995 ng

RT: 9.969 min Scan# 1187




Delta R.T. 0.000 min


Lab File: BM050251.D

Acq: 09 Jun 2025 20:58

Tgt Ion: 93 Resp: 1638720

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.2	39.4
123	10.9	8.9	13.3





#29

2,4-Dichlorophenol

Concen: 51.137 ng

RT: 10.204 min Scan# 11

Delta R.T. 0.006 min

Lab File: BM050251.D

Acq: 09 Jun 2025 20:58

Instrument :





BNA_M

ClientSampleId :

TP06-MHI-WCMSD

Tgt Ion:162 Resp: 1109642

Ion	Ratio	Lower	Upper
162	100		
164	65.6	45.0	85.0
98	37.2	16.3	56.3



#30

1,2,4-Trichlorobenzene

Concen: 48.367 ng

RT: 10.422 min Scan# 1264




Delta R.T. 0.000 min


Lab File: BM050251.D

Acq: 09 Jun 2025 20:58

Tgt Ion:180 Resp: 1167906


Ion	Ratio	Lower	Upper
180	100		
182	97.8	76.6	114.8
145	30.0	24.2	36.4








#31
Naphthalene
Concen: 47.325 ng
RT: 10.604 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD







Tgt Ion:128 Resp: 3699253
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.9 10.2 15.4



#32
Benzoic acid
Concen: 47.746 ng
RT: 9.863 min Scan# 1169
Delta R.T. 0.018 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:122 Resp: 703609
Ion Ratio Lower Upper
122 100
105 123.5 103.4 143.4
77 91.4 70.3 110.3









#33
4-Chloroaniline
Concen: 7.870 ng
RT: 10.710 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:	127	Resp:	262265
Ion	Ratio	Lower	Upper
127	100		
129	31.3	25.4	38.0
65	26.4	21.0	31.4
92	18.3	14.8	22.2



#34
Hexachlorobutadiene
Concen: 48.678 ng
RT: 10.904 min Scan# 1346
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:	225	Resp:	699158
Ion	Ratio	Lower	Upper
225	100		
223	62.1	50.2	75.2
227	64.8	51.0	76.4









#35
Caprolactam
Concen: 38.226 ng
RT: 11.486 min Scan# 1445
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:113 Resp: 262971
Ion Ratio Lower Upper
113 100
55 102.4 83.5 123.5
56 79.1 60.6 100.6



#36
4-Chloro-3-methylphenol
Concen: 47.948 ng
RT: 11.833 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:107 Resp: 1110198
Ion Ratio Lower Upper
107 100
144 25.6 21.0 31.6
142 80.0 64.8 97.2









#37
2-Methylnaphthalene
Concen: 48.538 ng
RT: 12.216 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:142 Resp: 2288847
Ion Ratio Lower Upper
142 100
141 87.9 71.5 107.3
115 34.1 27.8 41.8



#38
1-Methylnaphthalene
Concen: 47.810 ng
RT: 12.439 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:142 Resp: 2392878
Ion Ratio Lower Upper
142 100
141 90.6 72.9 109.3
116 3.7 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.398 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:164 Resp: 835521
Ion Ratio Lower Upper
164 100
162 100.9 80.2 120.4
160 45.9 35.7 53.5



#40
1,2,4,5-Tetrachlorobenzene
Concen: 51.984 ng
RT: 12.592 min Scan# 1633
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:216 Resp: 1254185
Ion Ratio Lower Upper
216 100
214 78.1 63.0 94.4
179 20.7 16.6 25.0
108 21.5 15.9 23.9









#41
Hexachlorocyclopentadiene
Concen: 109.354 ng
RT: 12.580 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:237 Resp: 1602258
Ion Ratio Lower Upper
237 100
235 62.5 44.6 84.6
272 13.0 0.0 33.1



#42
2,4,6-Tribromophenol
Concen: 143.063 ng
RT: 15.886 min Scan# 2193
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Tgt Ion:330 Resp: 1359589
Ion Ratio Lower Upper
330 100
332 96.3 77.5 116.3
141 34.8 28.8 43.2








#43
2,4,6-Trichlorophenol
Concen: 53.140 ng
RT: 12.827 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD





Tgt Ion:196 Resp: 843116
Ion Ratio Lower Upper
196 100
198 94.9 76.6 115.0
200 30.9 24.5 36.7




#44
2,4,5-Trichlorophenol
Concen: 52.647 ng
RT: 12.898 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58



Tgt Ion:196 Resp: 917115
Ion Ratio Lower Upper
196 100
198 96.9 76.1 114.1
97 49.7 39.3 58.9
132 27.6 21.4 32.0









#45
2-Fluorobiphenyl
Concen: 86.339 ng
RT: 13.027 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:172 Resp: 5328381
Ion Ratio Lower Upper
172 100
171 34.8 27.8 41.6
170 23.4 18.6 27.8



#46
1,1'-Biphenyl
Concen: 50.703 ng
RT: 13.233 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:154 Resp: 3172114
Ion Ratio Lower Upper
154 100
153 39.1 19.2 59.2
76 12.7 0.0 32.0









#47
2-Chloronaphthalene
Concen: 49.788 ng
RT: 13.275 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:162 Resp: 2421623
Ion Ratio Lower Upper
162 100
127 39.0 31.1 46.7
164 33.0 26.5 39.7



#48
2-Nitroaniline
Concen: 55.620 ng
RT: 13.469 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Tgt Ion: 65 Resp: 595329
Ion Ratio Lower Upper
65 100
92 78.2 64.2 96.4
138 120.0 101.0 151.4








#49
Acenaphthylene
Concen: 50.430 ng
RT: 14.122 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD





Tgt Ion:152 Resp: 3967313
Ion Ratio Lower Upper
152 100
151 19.3 15.8 23.6
153 12.8 10.4 15.6




#50
Dimethylphthalate
Concen: 50.422 ng
RT: 13.863 min Scan# 1849
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58



Tgt Ion:163 Resp: 2849131
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.0 7.9 11.9









#51
2,6-Dinitrotoluene
Concen: 53.767 ng
RT: 13.969 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:165 Resp: 612698
Ion Ratio Lower Upper
165 100
63 46.7 35.8 53.6
89 56.9 44.8 67.2



#52
Acenaphthene
Concen: 47.456 ng
RT: 14.463 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:154 Resp: 2335159
Ion Ratio Lower Upper
154 100
153 111.9 88.7 133.1
152 53.3 42.6 63.8









#53
3-Nitroaniline
Concen: 27.141 ng
RT: 14.292 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:138 Resp: 344695
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 116.5 92.9 139.3



#54
2,4-Dinitrophenol
Concen: 109.274 ng
RT: 14.498 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:184 Resp: 691966
Ion Ratio Lower Upper
184 100
63 55.2 45.3 67.9
154 61.6 54.3 81.5









#55
Dibenzenofuran
Concen: 49.666 ng
RT: 14.798 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:168 Resp: 3575384
Ion Ratio Lower Upper
168 100
139 37.1 29.9 44.9
169 13.5 10.6 16.0



#56
4-Nitrophenol
Concen: 96.935 ng
RT: 14.604 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:139 Resp: 1047823
Ion Ratio Lower Upper
139 100
109 71.7 52.6 92.6
65 80.5 64.3 104.3









#57
2,4-Dinitrotoluene
Concen: 55.251 ng
RT: 14.757 min Scan# 2003
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:165 Resp: 830530
Ion Ratio Lower Upper
165 100
63 34.8 28.4 42.6
89 70.6 58.0 87.0
182 2.6 2.2 3.2



#58
Fluorene
Concen: 49.014 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Tgt Ion:166 Resp: 2702504
Ion Ratio Lower Upper
166 100
165 97.5 77.0 115.4
167 13.6 10.7 16.1








#59
2,3,4,6-Tetrachlorophenol
Concen: 51.081 ng
RT: 15.021 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD





Tgt Ion:232 Resp: 769866
Ion Ratio Lower Upper
232 100
131 44.9 32.9 49.3
130 2.3 1.8 2.8
166 30.7 20.8 31.2




#60
Diethylphthalate
Concen: 49.900 ng
RT: 15.227 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58



Tgt Ion:149 Resp: 2797659
Ion Ratio Lower Upper
149 100
177 21.8 17.2 25.8
150 12.2 10.0 15.0









#61
4-Chlorophenyl-phenylether
Concen: 50.026 ng
RT: 15.445 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:204 Resp: 1333102
Ion Ratio Lower Upper
204 100
206 32.7 25.4 38.0
141 57.7 47.9 71.9



#62
4-Nitroaniline
Concen: 48.899 ng
RT: 15.457 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58


Tgt Ion:138 Resp: 582061
Ion Ratio Lower Upper
138 100
92 50.1 28.8 68.8
108 77.9 58.1 98.1








#63
Azobenzene
Concen: 48.962 ng
RT: 15.739 min Scan# 2171
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD







Tgt Ion: 77 Resp: 2743738
Ion Ratio Lower Upper
77 100
182 26.2 6.8 46.8
105 22.2 1.7 41.7
51 16.2 0.0 36.2



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2406
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion: 188 Resp: 1452446
Ion Ratio Lower Upper
188 100
94 9.4 8.1 12.1
80 10.1 8.6 13.0









#65
4,6-Dinitro-2-methylphenol
Concen: 53.979 ng
RT: 15.516 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:198 Resp: 428217
Ion Ratio Lower Upper
198 100
51 25.3 4.5 44.5
105 45.5 24.9 64.9



#66
n-Nitrosodiphenylamine
Concen: 52.335 ng
RT: 15.657 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:169 Resp: 2388396
Ion Ratio Lower Upper
169 100
168 66.6 53.1 79.7
167 36.5 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 54.132 ng
RT: 16.333 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:248 Resp: 838429
Ion Ratio Lower Upper
248 100
250 97.3 78.0 117.0
141 76.8 60.1 90.1



#68
Hexachlorobenzene
Concen: 52.893 ng
RT: 16.451 min Scan# 2289
Delta R.T. 0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:284 Resp: 957173
Ion Ratio Lower Upper
284 100
142 33.7 28.8 43.2
249 31.1 25.8 38.6









#69
Atrazine
Concen: 54.691 ng
RT: 16.610 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.9	6.0	46.0
215	49.7	28.1	68.1



#70
Pentachlorophenol
Concen: 108.884 ng
RT: 16.786 min Scan# 2346
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion	Ratio	Lower	Upper
266	100		
268	64.3	51.0	76.6
264	63.0	50.8	76.2









#71
Phenanthrene
Concen: 49.986 ng
RT: 17.180 min Scan# 2416
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:178 Resp: 4147484
Ion Ratio Lower Upper
178 100
176 19.1 15.6 23.4
179 15.4 12.3 18.5



#72
Anthracene
Concen: 50.737 ng
RT: 17.268 min Scan# 2428
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:178 Resp: 4211491
Ion Ratio Lower Upper
178 100
176 18.7 14.8 22.2
179 15.3 12.3 18.5









#73
Carbazole
Concen: 50.155 ng
RT: 17.533 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCM5D


Tgt Ion:167 Resp: 3792742
Ion Ratio Lower Upper
167 100
166 21.0 17.0 25.6
139 12.8 10.2 15.4



#74
Di-n-butylphthalate
Concen: 54.471 ng
RT: 18.115 min Scan# 2572
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:149 Resp: 4481884
Ion Ratio Lower Upper
149 100
150 9.1 7.4 11.0
104 5.1 4.0 6.0









#75
Fluoranthene
Concen: 49.778 ng
RT: 19.192 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:202 Resp: 4355639
Ion Ratio Lower Upper
202 100
101 11.6 0.0 31.1
203 17.5 0.0 37.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.368 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:240 Resp: 1291557
Ion Ratio Lower Upper
240 100
120 11.4 9.0 13.4
236 25.1 20.7 31.1









#77
Benzidine
Concen: 12.164 ng
RT: 19.374 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCM5D


Tgt Ion:184 Resp: 441789
Ion Ratio Lower Upper
184 100
185 15.1 11.5 17.3
183 11.3 9.0 13.4



#78
Pyrene
Concen: 52.225 ng
RT: 19.556 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:202 Resp: 4546225
Ion Ratio Lower Upper
202 100
200 20.7 16.7 25.1
203 17.2 14.1 21.1









#79
Terphenyl-d14
Concen: 90.550 ng
RT: 19.762 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:244 Resp: 6171435
Ion Ratio Lower Upper
244 100
212 7.1 6.0 9.0
122 11.3 8.6 12.8



#80
Butylbenzylphthalate
Concen: 59.759 ng
RT: 20.468 min Scan# 2972
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:149 Resp: 1735253
Ion Ratio Lower Upper
149 100
91 69.3 56.9 85.3
206 21.6 16.1 24.1









#81
Benzo(a)anthracene
Concen: 50.976 ng
RT: 21.350 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:228 Resp: 4167289
Ion Ratio Lower Upper
228 100
226 27.3 21.7 32.5
229 19.6 15.7 23.5



#82
3,3'-Dichlorobenzidine
Concen: 33.172 ng
RT: 21.274 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:252 Resp: 826974
Ion Ratio Lower Upper
252 100
254 64.5 51.4 77.0
126 11.5 8.6 12.8









#83
Chrysene
Concen: 51.200 ng
RT: 21.409 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:228 Resp: 3981460
Ion Ratio Lower Upper
228 100
226 30.3 24.3 36.5
229 19.8 15.8 23.8



#84
Bis(2-ethylhexyl)phthalate
Concen: 59.532 ng
RT: 21.298 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:149 Resp: 2662705
Ion Ratio Lower Upper
149 100
167 27.6 22.3 33.5
279 4.0 3.5 5.3









#85
Di-n-octyl phthalate
Concen: 62.247 ng
RT: 22.433 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:149 Resp: 4131361
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 4.8 3.8 5.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.344 min Scan# 3631
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:264 Resp: 1337173
Ion Ratio Lower Upper
264 100
260 23.9 18.6 28.0
265 21.8 16.8 25.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 59.218 ng
RT: 27.703 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:276 Resp: 5098501
Ion Ratio Lower Upper
276 100
138 25.7 20.6 30.8
277 24.7 19.8 29.6



#88
Benzo(b)fluoranthene
Concen: 52.083 ng
RT: 23.409 min Scan# 3472
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:252 Resp: 4049157
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 9.8 8.2 12.4









#89
Benzo(k)fluoranthene
Concen: 51.744 ng
RT: 23.474 min Scan# 34
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD


Tgt Ion:252 Resp: 4106648
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.2
125 10.2 8.2 12.4



#90
Benzo(a)pyrene
Concen: 53.567 ng
RT: 24.203 min Scan# 3607
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:252 Resp: 3915530
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.4 9.0 13.4









#91
Dibenzo(a,h)anthracene
Concen: 59.431 ng
RT: 27.768 min Scan# 41
Delta R.T. 0.000 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58




Instrument :
BNA_M
ClientSampleId :
TP06-MHI-WCMSD

Tgt Ion:278 Resp: 4153403
Ion Ratio Lower Upper
278 100
139 16.2 12.8 19.2
279 23.9 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 59.866 ng
RT: 28.744 min Scan# 4379
Delta R.T. -0.006 min
Lab File: BM050251.D
Acq: 09 Jun 2025 20:58

Tgt Ion:276 Resp: 4187458
Ion Ratio Lower Upper
276 100
277 23.4 18.6 27.8
138 21.9 17.8 26.6



Manual Integration Report

Sequence:	BM060525	Instrument	BNA_m

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BM050195.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/6/2025 12:19:17 PM	Jagrut	6/6/2025 1:16:08 PM	Peak Integrated by Software
SSTDICC005	BM050195.D	Benzaldehyde	Rahul	6/6/2025 12:19:17 PM	Jagrut	6/6/2025 1:16:08 PM	Peak Integrated by Software
SSTDICC010	BM050196.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/5/2025 4:48:25 PM	Jagrut	6/6/2025 1:16:10 PM	Peak Integrated by Software
SSTDICC020	BM050197.D	Benzaldehyde	Rahul	6/6/2025 12:19:19 PM	Jagrut	6/6/2025 1:16:13 PM	Peak Integrated by Software
SSTDICCC040	BM050198.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/5/2025 4:48:28 PM	Jagrut	6/6/2025 1:16:15 PM	Peak Integrated by Software
SSTDICV040	BM050202.D	Benzaldehyde	Rahul	6/5/2025 4:48:30 PM	Jagrut	6/6/2025 1:16:18 PM	Peak Integrated by Software

Manual Integration Report

Sequence:

bm060925

Instrument

BNA_m

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BM050236.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/10/2025 5:09:05 PM	Jagrut	6/10/2025 5:10:15 PM	Peak Integrated by Software
SSTDCCC040	BM050245.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/10/2025 5:09:14 PM	Jagrut	6/10/2025 5:10:26 PM	Peak Integrated by Software
Q2226-04MS	BM050250.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/10/2025 5:09:16 PM	Jagrut	6/10/2025 5:10:28 PM	Peak Integrated by Software
Q2236-19	BM050256.D	Naphthalene	Rahul	6/10/2025 5:09:34 PM	Jagrut	6/10/2025 5:10:31 PM	Peak Integrated by Software
SSTDCCC040	BM050262.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/10/2025 5:09:18 PM	Jagrut	6/10/2025 5:10:33 PM	Peak Integrated by Software
PB168352BS	BM050264.D	2,3,4,6-Tetrachlorophen ol	Rahul	6/10/2025 5:09:25 PM	Jagrut	6/10/2025 5:10:37 PM	Peak Integrated by Software
PB168352BS	BM050264.D	Caprolactam	Rahul	6/10/2025 5:09:25 PM	Jagrut	6/10/2025 5:10:37 PM	Peak Integrated by Software

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM060525

Review By	Rahul	Review On	6/6/2025 12:27:19 PM		
Supervise By	Jagrut	Supervise On	6/6/2025 1:16:37 PM		
SubDirectory	BM060525	HP Acquire Method	BNA_M	HP Processing Method	bm060525
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791			
CCC		SP6787			
Internal Standard/PEM		S12667,10ul/1000ul sample			
ICV/I.BLK		SP6796			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM050193.D	05 Jun 2025 08:40	RC/JU	Ok
2	SSTDICC2.5	BM050194.D	05 Jun 2025 09:20	RC/JU	Ok
3	SSTDICC005	BM050195.D	05 Jun 2025 09:59	RC/JU	Ok,M
4	SSTDICC010	BM050196.D	05 Jun 2025 10:38	RC/JU	Ok,M
5	SSTDICC020	BM050197.D	05 Jun 2025 11:17	RC/JU	Ok,M
6	SSTDICCC040	BM050198.D	05 Jun 2025 11:57	RC/JU	Ok,M
7	SSTDICC050	BM050199.D	05 Jun 2025 12:36	RC/JU	Ok
8	SSTDICC060	BM050200.D	05 Jun 2025 13:16	RC/JU	Ok
9	SSTDICC080	BM050201.D	05 Jun 2025 13:56	RC/JU	Ok
10	SSTDICV040	BM050202.D	05 Jun 2025 14:36	RC/JU	Ok,M
11	PB168224TB	BM050203.D	05 Jun 2025 16:35	RC/JU	Ok
12	SP6794	BM050204.D	05 Jun 2025 17:15	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM060925

Review By	Rahul	Review On	6/10/2025 5:10:05 PM
Supervise By	Jagrut	Supervise On	6/10/2025 5:10:49 PM
SubDirectory	BM060925	HP Acquire Method	BNA_M
		HP Processing Method	BM060625
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791		
CCC	SP6787		
Internal Standard/PEM	S12667,10ul/1000ul sample		
ICV/I.BLK	SP6796		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM050235.D	09 Jun 2025 10:07	RC/JU	Ok
2	SSTDCCC040	BM050236.D	09 Jun 2025 10:47	RC/JU	Ok,M
3	PB168340BL	BM050237.D	09 Jun 2025 11:26	RC/JU	Ok
4	PB168340BS	BM050238.D	09 Jun 2025 12:05	RC/JU	Ok,M
5	PB168340TB	BM050239.D	09 Jun 2025 12:44	RC/JU	Ok
6	Q2125-08	BM050240.D	09 Jun 2025 13:29	RC/JU	Ok
7	Q2125-08MS	BM050241.D	09 Jun 2025 14:08	RC/JU	Ok,M
8	Q2125-08MSD	BM050242.D	09 Jun 2025 14:47	RC/JU	Ok,M
9	Q2226-04	BM050243.D	09 Jun 2025 15:39	RC/JU	Ok
10	DFTPP	BM050244.D	09 Jun 2025 16:24	RC/JU	Ok
11	SSTDCCC040	BM050245.D	09 Jun 2025 17:03	RC/JU	Ok,M
12	PB168311TB	BM050246.D	09 Jun 2025 17:43	RC/JU	Ok
13	Q2227-04	BM050247.D	09 Jun 2025 18:22	RC/JU	Ok
14	Q2236-07	BM050248.D	09 Jun 2025 19:01	RC/JU	Ok
15	Q2228-04	BM050249.D	09 Jun 2025 19:40	RC/JU	Ok
16	Q2226-04MS	BM050250.D	09 Jun 2025 20:19	RC/JU	Ok,M
17	Q2226-04MSD	BM050251.D	09 Jun 2025 20:58	RC/JU	Ok
18	Q2235-03	BM050252.D	09 Jun 2025 21:38	RC/JU	Ok
19	Q2236-03	BM050253.D	09 Jun 2025 22:17	RC/JU	Ok
20	Q2236-11	BM050254.D	09 Jun 2025 22:56	RC/JU	Ok
21	Q2236-15	BM050255.D	09 Jun 2025 23:35	RC/JU	Ok

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM060925

Review By	Rahul	Review On	6/10/2025 5:10:05 PM
Supervise By	Jagrut	Supervise On	6/10/2025 5:10:49 PM
SubDirectory	BM060925	HP Acquire Method	BNA_M
		HP Processing Method	BM060625
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791		
CCC	SP6787		
Internal Standard/PEM	S12667,10ul/1000ul sample		
ICV/I.BLK	SP6796		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2236-19	BM050256.D	10 Jun 2025 00:15	RC/JU	Ok,M
23	Q2241-04	BM050257.D	10 Jun 2025 00:54	RC/JU	Ok
24	Q2241-08	BM050258.D	10 Jun 2025 01:33	RC/JU	Ok
25	Q2260-04	BM050259.D	10 Jun 2025 02:12	RC/JU	Ok
26	Q2265-04	BM050260.D	10 Jun 2025 02:51	RC/JU	Ok
27	DFTPP	BM050261.D	10 Jun 2025 04:09	RC/JU	Ok
28	SSTDCCC040	BM050262.D	10 Jun 2025 04:48	RC/JU	Ok,M
29	PB168352BL	BM050263.D	10 Jun 2025 05:28	RC/JU	Ok
30	PB168352BS	BM050264.D	10 Jun 2025 06:07	RC/JU	Ok,M
31	PB168333TB	BM050265.D	10 Jun 2025 06:46	RC/JU	Ok
32	Q2242-04	BM050266.D	10 Jun 2025 07:25	RC/JU	Ok
33	Q2244-04	BM050267.D	10 Jun 2025 08:04	RC/JU	Ok
34	Q2266-04	BM050268.D	10 Jun 2025 08:43	RC/JU	Ok
35	Q2266-08	BM050269.D	10 Jun 2025 09:22	RC/JU	ReRun
36	Q2240-04	BM050270.D	10 Jun 2025 10:01	RC/JU	Ok
37	Q2240-08	BM050271.D	10 Jun 2025 10:41	RC/JU	Ok
38	Q2240-12	BM050272.D	10 Jun 2025 11:20	RC/JU	Ok
39	Q2231-02DL	BM050273.D	10 Jun 2025 11:59	RC/JU	Dilution
40	Q2202-03DL	BM050274.D	10 Jun 2025 12:38	RC/JU	Ok
41	Q2231-02DL2	BM050275.D	10 Jun 2025 13:28	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM060525

Review By	Rahul	Review On	6/6/2025 12:27:19 PM
Supervise By	Jagrut	Supervise On	6/6/2025 1:16:37 PM
SubDirectory	BM060525	HP Acquire Method	BNA_M
		HP Processing Method	bm060525

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050193.D	05 Jun 2025 08:40		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050194.D	05 Jun 2025 09:20		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050195.D	05 Jun 2025 09:59	Compound #32,54,65,77 removed from 5PPM	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM050196.D	05 Jun 2025 10:38	Compound #54 kept on QR	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050197.D	05 Jun 2025 11:17		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050198.D	05 Jun 2025 11:57	The compound # 85 failed in the Calibration.	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM050199.D	05 Jun 2025 12:36		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BM050200.D	05 Jun 2025 13:16		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050201.D	05 Jun 2025 13:56	Compound #9 removed from 80PPM	RC/JU	Ok
10	SSTDICV040	ICVBM060525	BM050202.D	05 Jun 2025 14:36		RC/JU	Ok,M
11	PB168224TB	PB168224TB	BM050203.D	05 Jun 2025 16:35		RC/JU	Ok
12	SP6794	SP6794	BM050204.D	05 Jun 2025 17:15		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM060925

Review By	Rahul	Review On	6/10/2025 5:10:05 PM
Supervise By	Jagrut	Supervise On	6/10/2025 5:10:49 PM
SubDirectory	BM060925	HP Acquire Method	BNA_M
		HP Processing Method	BM060625

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050235.D	09 Jun 2025 10:07		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050236.D	09 Jun 2025 10:47		RC/JU	Ok,M
3	PB168340BL	PB168340BL	BM050237.D	09 Jun 2025 11:26		RC/JU	Ok
4	PB168340BS	PB168340BS	BM050238.D	09 Jun 2025 12:05		RC/JU	Ok,M
5	PB168340TB	PB168340TB	BM050239.D	09 Jun 2025 12:44		RC/JU	Ok
6	Q2125-08	GSB3	BM050240.D	09 Jun 2025 13:29		RC/JU	Ok
7	Q2125-08MS	GSB3MS	BM050241.D	09 Jun 2025 14:08		RC/JU	Ok,M
8	Q2125-08MSD	GSB3MSD	BM050242.D	09 Jun 2025 14:47		RC/JU	Ok,M
9	Q2226-04	TP06-MHI-WC	BM050243.D	09 Jun 2025 15:39		RC/JU	Ok
10	DFTPP	DFTPP	BM050244.D	09 Jun 2025 16:24		RC/JU	Ok
11	SSTDCCC040	SSTDCCC040	BM050245.D	09 Jun 2025 17:03		RC/JU	Ok,M
12	PB168311TB	PB168311TB	BM050246.D	09 Jun 2025 17:43		RC/JU	Ok
13	Q2227-04	TP07-MHH-WC	BM050247.D	09 Jun 2025 18:22		RC/JU	Ok
14	Q2236-07	WC-A2-04-C	BM050248.D	09 Jun 2025 19:01		RC/JU	Ok
15	Q2228-04	TP08-MHI-WC	BM050249.D	09 Jun 2025 19:40		RC/JU	Ok
16	Q2226-04MS	TP06-MHI-WCMS	BM050250.D	09 Jun 2025 20:19		RC/JU	Ok,M
17	Q2226-04MSD	TP06-MHI-WCMSD	BM050251.D	09 Jun 2025 20:58		RC/JU	Ok
18	Q2235-03	WC-A2-08-C	BM050252.D	09 Jun 2025 21:38		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM060925

Review By	Rahul	Review On	6/10/2025 5:10:05 PM		
Supervise By	Jagrut	Supervise On	6/10/2025 5:10:49 PM		
SubDirectory	BM060925	HP Acquire Method	BNA_M	HP Processing Method	BM060625
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791			
CCC		SP6787			
Internal Standard/PEM		S12667,10ul/1000ul sample			
ICV/I.BLK		SP6796			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2236-03	WC-A4-05A-C	BM050253.D	09 Jun 2025 22:17		RC/JU	Ok
20	Q2236-11	WC-A2-05-C	BM050254.D	09 Jun 2025 22:56		RC/JU	Ok
21	Q2236-15	WC-A2-06-C	BM050255.D	09 Jun 2025 23:35		RC/JU	Ok
22	Q2236-19	WC-A2-07-C	BM050256.D	10 Jun 2025 00:15		RC/JU	Ok,M
23	Q2241-04	TP-N	BM050257.D	10 Jun 2025 00:54		RC/JU	Ok
24	Q2241-08	TP-S	BM050258.D	10 Jun 2025 01:33		RC/JU	Ok
25	Q2260-04	TP10-MHG-WC	BM050259.D	10 Jun 2025 02:12		RC/JU	Ok
26	Q2265-04	TP11-MHL-WC	BM050260.D	10 Jun 2025 02:51		RC/JU	Ok
27	DFTPP	DFTPP	BM050261.D	10 Jun 2025 04:09		RC/JU	Ok
28	SSTDCCC040	SSTDCCC040	BM050262.D	10 Jun 2025 04:48		RC/JU	Ok,M
29	PB168352BL	PB168352BL	BM050263.D	10 Jun 2025 05:28		RC/JU	Ok
30	PB168352BS	PB168352BS	BM050264.D	10 Jun 2025 06:07		RC/JU	Ok,M
31	PB168333TB	PB168333TB	BM050265.D	10 Jun 2025 06:46		RC/JU	Ok
32	Q2242-04	TP09-MHJ	BM050266.D	10 Jun 2025 07:25		RC/JU	Ok
33	Q2244-04	TP03-MHC	BM050267.D	10 Jun 2025 08:04		RC/JU	Ok
34	Q2266-04	WC-3	BM050268.D	10 Jun 2025 08:43		RC/JU	Ok
35	Q2266-08	WC-5	BM050269.D	10 Jun 2025 09:22	Internal Standard Failed & Surrogates failed for low IS recoveries.	RC/JU	ReRun
36	Q2240-04	TP-3	BM050270.D	10 Jun 2025 10:01		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM060925

Review By	Rahul	Review On	6/10/2025 5:10:05 PM		
Supervise By	Jagrut	Supervise On	6/10/2025 5:10:49 PM		
SubDirectory	BM060925	HP Acquire Method	BNA_M	HP Processing Method	BM060625
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791			
CCC		SP6787			
Internal Standard/PEM		S12667,10ul/1000ul sample			
ICV/I.BLK		SP6796			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

37	Q2240-08	TP-2	BM050271.D	10 Jun 2025 10:41		RC/JU	Ok
38	Q2240-12	TP-1	BM050272.D	10 Jun 2025 11:20		RC/JU	Ok
39	Q2231-02DL	MW-14-20250604DL	BM050273.D	10 Jun 2025 11:59	Need further 5X dilution	RC/JU	Dilution
40	Q2202-03DL	MW-12-20250603DL	BM050274.D	10 Jun 2025 12:38		RC/JU	Ok
41	Q2231-02DL2	MW-14-20250604DL2	BM050275.D	10 Jun 2025 13:28		RC/JU	Ok

M : Manual Integration

SOP ID : M1311-TCLP-16

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-7

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pipette ID : WC

Tumbler ID : T-1 / T-2

TCLP Filter ID : 115525

Start Prep Date : 06/05/2025 Time : 15:00

End Prep Date : 06/06/2025 Time : 09:20

Combination Ratio : 20

ZHE Cleaning Batch : N/A

Initial Room Temperature: 24 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : JP

Supervisor By : 12

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
HCL-TCLP,1N	N/A	WP112797
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	N/A	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120 ml Plastic	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 /T-2 checked,30 rpm. Q2242-04 IS USED FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/06/25 11:30	<u>JP</u> / <u>ICB Room</u>	<u>RJ / Ext</u>
	Preparation Group	Analysis Group <u>JP / Noting</u>

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168311TB	LEB311	16	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q2226-04	TP06-MHI-WC	01	100.02	2000	N/A	N/A	N/A	4.0	1.0	T-1
Q2227-04	TP07-MHH-WC	02	100.03	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q2228-04	TP08-MHI-WC	03	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q2235-03	WC-A2-08-C	04	100.01	2000	N/A	N/A	N/A	12.0	1.5	T-1
Q2236-03	WC-A4-05A-C	05	100.02	2000	N/A	N/A	N/A	12.0	1.0	T-1
Q2236-07	WC-A2-04-C	06	100.03	2000	N/A	N/A	N/A	12.0	1.5	T-1
Q2236-11	WC-A2-05-C	07	100.00	2000	N/A	N/A	N/A	12.0	1.0	T-1
Q2236-15	WC-A2-06-C	08	100.01	2000	N/A	N/A	N/A	11.5	1.5	T-1
Q2236-19	WC-A2-07-C	09	100.02	2000	N/A	N/A	N/A	11.5	1.0	T-1
Q2240-04	TP-3	10	100.03	2000	N/A	N/A	N/A	7.2	1.5	T-1
Q2240-08	TP-2	11	100.02	2000	N/A	N/A	N/A	7.0	1.0	T-2
Q2240-12	TP-1	12	100.02	2000	N/A	N/A	N/A	7.0	1.0	T-2
Q2241-04	TP-N	13	100.04	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q2241-08	TP-S	14	100.02	2000	N/A	N/A	N/A	7.0	1.0	T-2
Q2242-04	TP09-MHJ	15	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB168311TB	LEB311	N/A	N/A	N/A	N/A	N/A	N/A
Q2226-04	TP06-MHI-WC	N/A	N/A	N/A	N/A	100	N/A
Q2227-04	TP07-MHH-WC	N/A	N/A	N/A	N/A	100	N/A
Q2228-04	TP08-MHI-WC	N/A	N/A	N/A	N/A	100	N/A
Q2235-03	WC-A2-08-C	N/A	N/A	N/A	N/A	100	N/A
Q2236-03	WC-A4-05A-C	N/A	N/A	N/A	N/A	100	N/A
Q2236-07	WC-A2-04-C	N/A	N/A	N/A	N/A	100	N/A
Q2236-11	WC-A2-05-C	N/A	N/A	N/A	N/A	100	N/A
Q2236-15	WC-A2-06-C	N/A	N/A	N/A	N/A	100	N/A
Q2236-19	WC-A2-07-C	N/A	N/A	N/A	N/A	100	N/A
Q2240-04	TP-3	N/A	N/A	N/A	N/A	100	N/A
Q2240-08	TP-2	N/A	N/A	N/A	N/A	100	N/A
Q2240-12	TP-1	N/A	N/A	N/A	N/A	100	N/A
Q2241-04	TP-N	N/A	N/A	N/A	N/A	100	N/A
Q2241-08	TP-S	N/A	N/A	N/A	N/A	100	N/A
Q2242-04	TP09-MHJ	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB168311TB	LEB311	N/A	N/A	N/A	N/A	#1	4.93
Q2226-04	TP06-MHI-WC	5.02	96.5	6.6	2.0	#1	4.93
Q2227-04	TP07-MHH-WC	5.03	96.5	5.6	1.5	#1	4.93
Q2228-04	TP08-MHI-WC	5.02	96.5	5.5	1.5	#1	4.93
Q2235-03	WC-A2-08-C	5.01	96.5	12.5	4.5	#1	4.93
Q2236-03	WC-A4-05A-C	5.02	96.5	12.5	4.5	#1	4.93
Q2236-07	WC-A2-04-C	5.03	96.5	12.5	4.5	#1	4.93
Q2236-11	WC-A2-05-C	5.02	96.5	12.5	4.5	#1	4.93
Q2236-15	WC-A2-06-C	5.01	96.5	12.0	4.0	#1	4.93
Q2236-19	WC-A2-07-C	5.02	96.5	12.0	4.0	#1	4.93
Q2240-04	TP-3	5.03	96.5	9.1	3.5	#1	4.93
Q2240-08	TP-2	5.02	96.5	8.6	3.0	#1	4.93
Q2240-12	TP-1	5.03	96.5	8.6	3.0	#1	4.93
Q2241-04	TP-N	5.02	96.5	8.4	3.0	#1	4.93
Q2241-08	TP-S	5.01	96.5	8.6	3.5	#1	4.93
Q2242-04	TP09-MHJ	5.02	96.5	5.5	2.0	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tclp q2235

WorkList ID : 189978

Department : TCLP Extraction

Date : 06-05-2025 13:53:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2226-04	TP06-MHI-WC	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N42	06/04/2025	1311
Q2227-04	TP07-MHH-WC	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N41	06/03/2025	1311
Q2228-04	TP08-MHI-WC	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N41	06/04/2025	1311
Q2235-03	WC-A2-08-C	Solid	TCLP Extraction	Cool 4 deg C	ENTA05	N41	06/04/2025	1311
Q2236-03	WC-A4-05A-C	Solid	TCLP Extraction	Cool 4 deg C	ENTA05	N31	06/04/2025	1311
Q2236-07	WC-A2-04-C	Solid	TCLP Extraction	Cool 4 deg C	ENTA05	N31	06/04/2025	1311
Q2236-11	WC-A2-05-C	Solid	TCLP Extraction	Cool 4 deg C	ENTA05	N31	06/04/2025	1311
Q2236-15	WC-A2-06-C	Solid	TCLP Extraction	Cool 4 deg C	ENTA05	N31	06/04/2025	1311
Q2236-19	WC-A2-07-C	Solid	TCLP Extraction	Cool 4 deg C	ENTA05	N31	06/04/2025	1311
Q2240-04	TP-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N31	06/04/2025	1311
Q2240-08	TP-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N31	06/04/2025	1311
Q2240-12	TP-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N31	06/04/2025	1311
Q2241-04	TP-N	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N41	06/05/2025	1311
Q2241-08	TP-S	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	N41	06/05/2025	1311
Q2242-04	TP09-MHJ	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L21	06/05/2025	1311

Date/Time 06/05/25 14:10

Raw Sample Received by: SP WOC

Raw Sample Relinquished by: CP SM

Date/Time 06/05/25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: SP WOC

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: N/A

Extraction Start Date : 06/09/2025

Matrix : Water

Extraction Start Time : 10:45

Weigh By: EH

Extraction By: RJ

Extraction End Date : 06/09/2025

Balance check: N/A

Filter By: RJ

Extraction End Time : 15:25

Balance ID: N/A

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: E3880

Hood ID: 4,5,6,7

Supervisor By : rajesh

Extraction Method: ☒ Separatory Funnel

☐ Continuous Liquid/Liquid

☐ Sonication

☐ Waste Dilution

☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6794
Surrogate	1.0ML	100/150 PPM	SP6754
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	EP2612
Baked Na2SO4	N/A	EP2620
10N NaOH	N/A	E2865
H2SO4 1:1	N/A	E3939
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5ML Vail Lot # 2210443.p H Adjusted<2 with 1:1 H2SO4 & 10 N NaOH.

KD Bath ID: WATER BATH-1,2

Envap ID: NEVAP-02

KD Bath Temperature: 60 °C

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
06/09/25	RP (Eff. Lab)	Relsvor
15:30	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 06/09/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168311TB	PB168311TB	TCLP BNA	100	6	RUPESH	rajesh	1			SEP-1
PB168333TB	PB168333TB	TCLP BNA	100	6	RUPESH	rajesh	1			2
PB168352BL	PB168352BL	TCLP BNA	1000	6	RUPESH	rajesh	1			3
PB168352BS	PB168352BS	TCLP BNA	1000	6	RUPESH	rajesh	1			4
Q2226-04	TP06-MHI-WC	TCLP BNA	100	6	RUPESH	rajesh	1	A		5
Q2226-04MS	TP06-MHI-WCMS	TCLP BNA	100	6	RUPESH	rajesh	1	A		6
Q2226-04MS D	TP06-MHI-WCMSD	TCLP BNA	100	6	RUPESH	rajesh	1	A		7
Q2227-04	TP07-MHH-WC	TCLP BNA	100	6	RUPESH	rajesh	1	A		8
Q2228-04	TP08-MHI-WC	TCLP BNA	100	6	RUPESH	rajesh	1	A		9
Q2235-03	WC-A2-08-C	TCLP BNA	100	6	RUPESH	rajesh	1	A		10
Q2236-03	WC-A4-05A-C	TCLP BNA	100	6	RUPESH	rajesh	1	A		11
Q2236-07	WC-A2-04-C	TCLP BNA	100	6	RUPESH	rajesh	1	A		12
Q2236-11	WC-A2-05-C	TCLP BNA	100	6	RUPESH	rajesh	1	A		13
Q2236-15	WC-A2-06-C	TCLP BNA	100	6	RUPESH	rajesh	1	A		14
Q2236-19	WC-A2-07-C	TCLP BNA	100	6	RUPESH	rajesh	1	A		15
Q2240-04	TP-3	TCLP BNA	100	6	RUPESH	rajesh	1	A		16
Q2240-08	TP-2	TCLP BNA	100	6	RUPESH	rajesh	1	A		SEP-01
Q2240-12	TP-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		2
Q2241-04	TP-N	TCLP BNA	100	6	RUPESH	rajesh	1	A		3
Q2241-08	TP-S	TCLP BNA	100	6	RUPESH	rajesh	1	A		4
Q2242-04	TP09-MHJ	TCLP BNA	100	6	RUPESH	rajesh	1	A		5
Q2244-04	TP03-MHC	TCLP BNA	100	6	RUPESH	rajesh	1	A		6
Q2260-04	TP10-MHG-WC	TCLP BNA	100	6	RUPESH	rajesh	1	A		7
Q2265-04	TP-MHL-WC	TCLP BNA	100	6	RUPESH	rajesh	1	A		8
Q2266-04	WC-3	TCLP BNA	100	6	RUPESH	rajesh	1	A		9
Q2266-08	WC-4	TCLP BNA	100	6	RUPESH	rajesh	1	A		10

* Extracts relinquished on the same date as received.

8/5/25

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168311TB	LEB311	16	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q2226-04	TP06-MHI-WC	01	100.02	2000	N/A	N/A	N/A	4.0	1.0	T-1
Q2227-04	TP07-MHH-WC	02	100.03	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q2228-04	TP08-MHI-WC	03	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q2235-03	WC-A2-08-C	04	100.01	2000	N/A	N/A	N/A	12.0	1.5	T-1
Q2236-03	WC-A4-05A-C	05	100.02	2000	N/A	N/A	N/A	12.0	1.0	T-1
Q2236-07	WC-A2-04-C	06	100.03	2000	N/A	N/A	N/A	12.0	1.5	T-1
Q2236-11	WC-A2-05-C	07	100.00	2000	N/A	N/A	N/A	12.0	1.0	T-1
Q2236-15	WC-A2-06-C	08	100.01	2000	N/A	N/A	N/A	11.5	1.5	T-1
Q2236-19	WC-A2-07-C	09	100.02	2000	N/A	N/A	N/A	11.5	1.0	T-1
Q2240-04	TP-3	10	100.03	2000	N/A	N/A	N/A	7.2	1.5	T-1
Q2240-08	TP-2	11	100.02	2000	N/A	N/A	N/A	7.0	1.0	T-2
Q2240-12	TP-1	12	100.02	2000	N/A	N/A	N/A	7.0	1.0	T-2
Q2241-04	TP-N	13	100.04	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q2241-08	TP-S	14	100.02	2000	N/A	N/A	N/A	7.0	1.0	T-2
Q2242-04	TP09-MHJ	15	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-2

06/06/25
11/30

TCLP EXTRACTION LOGPAGE

PB168333

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168333TB	LEB333	09	N/A	2000	N/A	N/A	N/A	4.94	1.0	T-1
Q2137-05	MOO-25-0151	01	100.02	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q2244-04	TP03-MHC	02	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q2260-04	TP10-MHG-WC	03	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-1
Q2262-02	ARS20-0032	04	100.03	2000	N/A	N/A	N/A	9.1	1.0	T-1
Q2262-04	ARS20-0001	05	100.02	2000	N/A	N/A	N/A	10.0	1.5	T-1
Q2265-04	TP-MHL-WC	06	100.01	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q2266-04	WC-3	07	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-1
Q2266-08	WC-4	08	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 788-9222
www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2236

COC Number: 2042113

Page 1 of 2

CLIENT INFORMATION

COMPANY: ENTACT, LLC
ADDRESS: 150 Bay Street, Suite 806
CITY: Jersey City STATE: NJ ZIP: 07302
ATTENTION: Austin Farmerie
PHONE: 412-716-1366 FAX:

PROJECT INFORMATION

PROJECT NAME: 540 Degraw St Brooklyn, NY
PROJECT #: E9309 LOCATION: Brooklyn, NY
PROJECT MANAGER: Austin Farmerie
E-MAIL: afarmerie@entact.com
PHONE: 412-716-1366 FAX:

BILLING INFORMATION

BILL TO: ENTACT, LLC PO# E9309
ADDRESS: 999 Oakmont Plaza Drive, Suite 300
CITY: Westmont STATE: IL ZIP: 60559
ATTENTION: Wendy Murray PHONE: 800-936-8228

DATA TURNAROUND INFORMATION

FAX: 3 DAYS*
HARD COPY: DAYS*
EDD 3 DAYS*
* TO BE APPROVED BY ALLIANCE
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESULTS ONLY ☐ USEPA CLP
☐ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☐ EDD Format

ANALYSIS

TCLP VOCs	TCLP ICP Metals + Cu, Ni, Zn	TCLP Herb	TCLP Pest	TCLP SVOCs	TCLP pH	I/C/R	PCBs	Oil & Grease
1	2	3	4	5	6	7	8	9

PRESERVATIVES

COMMENTS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		E	E	E	E	E	E	E	E	E	
1.	WC-A4-05A-G	Soil		X	6/4	12:00	1	X									
2.	WC-A4-05A-C	Soil	X		6/4	12:00	11		X	X	X	X	X	X	X	X	
3.	WC-A2-04-G	Soil		X	6/4	12:00	1	X									
4.	WC-A2-04-C	Soil	X		6/4	12:00	11		X	X	X	X	X	X	X	X	
5.	WC-A2-05-G	Soil		X	6/4	12:00	1	X									
6.	WC-A2-05-C	Soil	X		6/4	12:00	11		X	X	X	X	X	X	X	X	
7.	WC-A2-06-G	Soil		X	6/4	12:00	1	X									
8.	WC-A2-06-C	Soil	X		6/4	12:00	11		X	X	X	X	X	X	X	X	
9.	WC-A2-07-G	Soil		X	6/4	12:00	1	X									
10.	WC-A2-07-C	Soil	X		6/4	12:00	11		X	X	X	X	X	X	X	X	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER 1. Austin Farmerie	DATE/TIME 6/4 11:00	RECEIVED BY 1. [Signature] 64.25	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 4.4 ☐ Ice in Cooler? _____
RELINQUISHED BY	DATE/TIME	RECEIVED BY	Comments:
2.		2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	
3.	6.4.25	3.	

Page _____ of _____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 788-9222

www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2236

COC Number: 2042113

Page 2 of 2

CLIENT INFORMATION

COMPANY: ENTACT, LLC

ADDRESS: 150 Bay Street, Suite 806

CITY: Jersey City STATE: NJ ZIP: 07302

ATTENTION: Austin Farmerie

PHONE: 412-716-1366

FAX:

PROJECT INFORMATION

PROJECT NAME: 540 Degraw St Brooklyn, NY

PROJECT #: E9309 LOCATION: Brooklyn, NY

PROJECT MANAGER: Austin Farmerie

E-MAIL: afarmerie@entact.com

PHONE: 412-716-1366

FAX:

BILLING INFORMATION

BILL TO: ENTACT, LLC

PO# E9309

ADDRESS: 999 Oakmont Plaza Drive, Suite 300

CITY: Westmont

STATE: IL ZIP: 60559

ATTENTION: Wendy Murray

PHONE: 800-936-8228

ANALYSIS

DATA TURNAROUND INFORMATION

FAX: 3 DAYS*

HARD COPY: 3 DAYS*

EDD 3 DAYS*

* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESEULTS ONLY☐ USEPA CLP☐ RESULTS + QC☐ New York State ASP "B"☐ New Jersey REDUCED☐ New York State ASP "A"☐ New Jersey CLP☐ Other _____☐ EDD Format _____

ASTM COD	ASTM Ammonia-Nitrogen	ASTM O&G	ASTM TS	TS, TVS	pH	Paint Filter		
10	11	12	13	14	15	16		

PRESERVATIVES

COMMENTS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	E	E	E	E	E	E	E	E	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9
1.	WC-A4-05A-G	Soil		X	5/6	12:00	1									
2.	WC-A4-05A-C	Soil	X		5/6	12:00	11	X	X	X	X	X	X	X		
3.	WC-A2-04-G	Soil		X	5/7	12:00	1									
4.	WC-A2-04-C	Soil	X		5/7	12:00	11	X	X	X	X	X	X	X		
5.	WC-A2-05-G	Soil		X	5/8	12:00	1									
6.	WC-A2-05-C	Soil	X		5/8	12:00	11	X	X	X	X	X	X	X		
7.	WC-A2-06-G	Soil		X	5/8	12:00	1									
8.	WC-A2-06-C	Soil	X		5/8	12:00	11	X	X	X	X	X	X	X		
9.	WC-A2-07-G	Soil		X	5/8	12:00	1									
10.	WC-A2-07-C	Soil	X		5/8	12:00	11	X	X	X	X	X	X	X		

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER 1. Austin Farmerie	DATE/TIME 6-4-25 1655	RECEIVED BY 	DATE/TIME 6-4-25 1655
RELINQUISHED BY 2.	DATE/TIME 	RECEIVED BY 	DATE/TIME
RELINQUISHED BY 3.	DATE/TIME 6-4-25 1830	RECEIVED FOR LAB BY 	DATE/TIME

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 4.4 ☐ Ice in Cooler? _____

Comments:

Page _____ of _____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488