

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/09/2025

Lab File ID: BM050245.D Time Analyzed: 17:03

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	360068	7.769	1478210	10.56	892439	14.40
UPPER LIMIT	720136	8.269	2956420	11.057	1784880	14.898
LOWER LIMIT	180034	7.269	739105	10.057	446220	13.898
EPA SAMPLE NO.						
01 TP06-MHI-WCMS	392950	7.77	1521760	10.56	827375	14.40
02 TP06-MHI-WCMSD	395507	7.77	1511860	10.56	835521	14.40
03 WC-A2-04-C	334436	7.77	1312450	10.56	761229	14.40
04 PB168311TB	348706	7.77	1311820	10.55	755301	14.40
05 WC-A4-05A-C	339236	7.77	1291740	10.56	734983	14.40
06 WC-A2-05-C	372657	7.77	1419600	10.56	799341	14.40
07 WC-A2-06-C	313628	7.77	1214920	10.55	691337	14.40
08 WC-A2-07-C	372220	7.77	1448450	10.56	829081	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.