

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG No.: Q2236

Instrument ID: MSVOA_X Calibration Date/Time: 06/07/2025 00:34

Lab File ID: VX046546.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:42 12:57

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.630	0.611		-3.02	20
1,1-Dichloroethene	0.594	0.589		-0.84	20
2-Butanone	0.495	0.584		17.98	20
Carbon Tetrachloride	0.588	0.564		-4.08	20
Chloroform	1.342	1.497		11.55	20
Benzene	1.479	1.500		1.42	20
1,2-Dichloroethane	0.615	0.672		9.27	20
Trichloroethene	0.376	0.349		-7.18	20
Tetrachloroethene	0.334	0.303		-9.28	20
Chlorobenzene	1.176	1.165	0.3	-0.94	20
1,2-Dichloroethane-d4	0.890	0.964		8.31	20
Dibromofluoromethane	0.368	0.381		3.53	20
Toluene-d8	1.212	1.194		-1.49	20
4-Bromofluorobenzene	0.502	0.498		-0.8	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.