

Surrogate Summary

SDG No.: **Q2236**

Client: **ENTACT**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2236-01	WC-A4-05A-G	1,2-Dichloroethane-d4	50	43.3	87	70 (74)	130 (125)
		Dibromofluoromethane	50	42.8	86	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
Q2236-05	WC-A2-04-G	1,2-Dichloroethane-d4	50	49.4	99	70 (74)	130 (125)
		Dibromofluoromethane	50	33.8	68 *	70 (75)	130 (124)
		Toluene-d8	50	53.0	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	57.6	115	70 (77)	130 (121)
Q2236-05RE	WC-A2-04-GRE	1,2-Dichloroethane-d4	50	43.2	86	70 (74)	130 (125)
		Dibromofluoromethane	50	32.6	65 *	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.6	97	70 (77)	130 (121)
Q2236-09	WC-A2-05-G	1,2-Dichloroethane-d4	50	41.5	83	70 (74)	130 (125)
		Dibromofluoromethane	50	40.2	80	70 (75)	130 (124)
		Toluene-d8	50	50.6	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	98	70 (77)	130 (121)
Q2236-13	WC-A2-06-G	1,2-Dichloroethane-d4	50	46.2	92	70 (74)	130 (125)
		Dibromofluoromethane	50	41.6	83	70 (75)	130 (124)
		Toluene-d8	50	52.6	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	59.0	118	70 (77)	130 (121)
Q2236-17	WC-A2-07-G	1,2-Dichloroethane-d4	50	47.5	95	70 (74)	130 (125)
		Dibromofluoromethane	50	34.9	70	70 (75)	130 (124)
		Toluene-d8	50	52.8	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	58.1	116	70 (77)	130 (121)
Q2236-17RE	WC-A2-07-GRE	1,2-Dichloroethane-d4	50	42.8	86	70 (74)	130 (125)
		Dibromofluoromethane	50	31.8	64 *	70 (75)	130 (124)
		Toluene-d8	50	50.6	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.8	96	70 (77)	130 (121)
VX0606WBL02	VX0606WBL02	1,2-Dichloroethane-d4	50	49.3	99	70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98	70 (75)	130 (124)
		Toluene-d8	50	48.8	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.2	108	70 (77)	130 (121)
VX0606WBS02	VX0606WBS02	1,2-Dichloroethane-d4	50	49.1	98	70 (74)	130 (125)
		Dibromofluoromethane	50	48.7	97	70 (75)	130 (124)
		Toluene-d8	50	49.3	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.6	101	70 (77)	130 (121)

Surrogate Summary

SDG No.: Q2236

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0609WBL01	VN0609WBL01	1,2-Dichloroethane-d4	50	42.8	86	70 (74)	130 (125)
		Dibromofluoromethane	50	48.1	96	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
VN0609WBS01	VN0609WBS01	1,2-Dichloroethane-d4	50	56.6	113	70 (74)	130 (125)
		Dibromofluoromethane	50	58.6	117	70 (75)	130 (124)
		Toluene-d8	50	55.2	110	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.7	111	70 (77)	130 (121)
VN0609WBSD0	VN0609WBSD01	1,2-Dichloroethane-d4	50	44.5	89	70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101	70 (75)	130 (124)
		Toluene-d8	50	48.2	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.7	95	70 (77)	130 (121)

