

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2236

**Client:** ENTACT

**Analytical Method:** SW8260D

**Datafile :** VX046548.D

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD |
|---------------|----------------------|-------|--------|------|-----|-----|------|---------|----------------|-----|
| VX0606WBS02   | Vinyl chloride       | 20    | 18.5   | ug/L | 93  |     |      | 70 (65) | 130 (117)      |     |
|               | 1,1-Dichloroethene   | 20    | 19.1   | ug/L | 96  |     |      | 70 (74) | 130 (110)      |     |
|               | 2-Butanone           | 100   | 98.5   | ug/L | 99  |     |      | 40 (65) | 160 (122)      |     |
|               | Carbon Tetrachloride | 20    | 19.1   | ug/L | 96  |     |      | 70 (77) | 130 (113)      |     |
|               | Chloroform           | 20    | 20.7   | ug/L | 104 |     |      | 70 (79) | 130 (113)      |     |
|               | Benzene              | 20    | 19.9   | ug/L | 100 |     |      | 70 (82) | 130 (109)      |     |
|               | 1,2-Dichloroethane   | 20    | 19.9   | ug/L | 100 |     |      | 70 (80) | 130 (115)      |     |
|               | Trichloroethene      | 20    | 18.3   | ug/L | 92  |     |      | 70 (77) | 130 (113)      |     |
|               | Tetrachloroethene    | 20    | 18.7   | ug/L | 94  |     |      | 70 (67) | 130 (123)      |     |
|               | Chlorobenzene        | 20    | 19.5   | ug/L | 98  |     |      | 70 (82) | 130 (109)      |     |