

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236
 Lab File ID: VX046546.D Date Analyzed: 06/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 00:34
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	74224	5.56	135585	6.76	121168	10.06
UPPER LIMIT	148448	6.062	271170	7.263	242336	10.555
LOWER LIMIT	37112	5.062	67792.5	6.263	60584	9.555
EPA SAMPLE NO.						
WC-A2-04-G	117002	5.57	224778	6.78	236932	10.06
WC-A2-06-G	110472	5.57	211587	6.78	221449	10.06
WC-A2-07-G	119993	5.57	228870	6.78	241518	10.06
VX0606WBL02	95381	5.56	186443	6.76	185656	10.06
VX0606WBS02	86561	5.57	153201	6.77	137797	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.