

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236
 Lab File ID: VX046546.D Date Analyzed: 06/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 00:34
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	59451	12.018				
UPPER LIMIT	118902	12.518				
LOWER LIMIT	29725.5	11.518				
EPA SAMPLE NO.						
WC-A2-04-G	118237	12.02				
WC-A2-06-G	117146	12.02				
WC-A2-07-G	121477 *	12.02				
VX0606WBL02	90343	12.02				
VX0606WBS02	70682	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.