

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136026

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	06/05/2025	14:30
2	Q2218-03		1	13	100.00	>212.0	>212.0	06/05/2025	15:00
3	Q2218-03DUP		1	13	100.00	>212.0	>212.0	06/05/2025	15:30
4	Q2237-02		1	15	100.00	>212.0	>212.0	06/05/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136026

WorkList Name : fp-06-05 WorkList ID : 189982 Department : Wet-Chemistry Date : 06-05-2025 09:58:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2218-03	3309	Water	Flash Point	Cool 4 deg C	PSEG03	N31	06/04/2025	1010B
Q2237-02	TW-WTS-10	Water	Flash Point	Cool 4 deg C	ENTA05	N31	06/04/2025	1010B

Date/Time 06/05/2025 14:20
Raw Sample Received by: EM (WC)
Raw Sample Relinquished by: JH (WC)

Date/Time 06/05/2025 16:46
Raw Sample Received by: JH (WC)
Raw Sample Relinquished by: EM (WC)