



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q2237
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2237 **SAS No.:** Q2237
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.92	6.0	99	65 - 135	P	06/06/2025	13:37	LB136052
	Copper	20.8	20.0	104	65 - 135	P	06/06/2025	13:37	LB136052
	Lead	13.8	12.0	115	65 - 135	P	06/06/2025	13:37	LB136052
	Nickel	39.5	40.0	99	65 - 135	P	06/06/2025	13:37	LB136052
	Zinc	45.4	40.0	113	65 - 135	P	06/06/2025	13:37	LB136052
CRA	Mercury	0.20	0.2	102	70 - 130	CV	06/10/2025	10:56	LB136079
CRI01	Cadmium	6.07	6.0	101	65 - 135	P	06/10/2025	14:50	LB136097
	Copper	22.7	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Lead	13.4	12.0	112	65 - 135	P	06/10/2025	14:50	LB136097
	Nickel	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Zinc	44.1	40.0	110	65 - 135	P	06/10/2025	14:50	LB136097