



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q2237				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q2237	SAS No.:	Q2237		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB31	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	10:41	LB136079	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2237

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2237

SAS No.: Q2237

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB17	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	10:48	LB136079
CCB18	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	11:37	LB136079
CCB19	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	12:14	LB136079
CCB20	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	12:59	LB136079
CCB21	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	13:32	LB136079
CCB22	Mercury	0.076	+/-0.2	U	0.20	CV	06/10/2025	14:30	LB136079

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2237

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2237

SAS No.: Q2237

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	13:33	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	13:33	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	13:33	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	13:33	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	13:33	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT SDG No.: Q2237
Contract: ENTA05 Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	14:09	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	14:09	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	14:09	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
CCB02	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	15:00	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	15:00	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	15:00	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
CCB03	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	15:57	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	15:57	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	15:57	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
CCB04	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	16:51	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	16:51	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	16:51	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
CCB05	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	17:46	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	17:46	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	17:46	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
CCB06	Cadmium	0.50	+/-3	U	6.00	P	06/07/2025	03:07	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/07/2025	03:07	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/07/2025	03:07	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
CCB07	Cadmium	0.50	+/-3	U	6.00	P	06/07/2025	03:44	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/07/2025	03:44	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/07/2025	03:44	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/07/2025	03:44	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/07/2025	03:44	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2237

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2237

SAS No.: Q2237

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	14:45	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	14:45	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	14:45	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q2237							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q2237	SAS No.: Q2237						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	15:35	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	15:35	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
CCB02	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	16:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	16:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
CCB03	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	17:40	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	17:40	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
CCB04	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	18:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	18:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
CCB05	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	19:22	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	19:22	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
CCB06	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	20:14	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	20:14	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
CCB07	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	21:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	21:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
CCB08	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	22:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	22:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	22:08	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2237</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2237</u>	SAS No.: <u>Q2237</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
CCB09	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	23:01	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	23:01	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
CCB10	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	00:02	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	00:02	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
CCB11	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	01:39	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	01:39	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
CCB12	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	02:25	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	02:25	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	02:25	LB136097