

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

Continuing Calib Date: 06/06/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 09:47 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.65       | 5.66      | 5.56      | 5.76  | 0.01       |
| Aroclor-1248-2 (2)   | 5.92       | 5.93      | 5.83      | 6.03  | 0.01       |
| Aroclor-1248-3 (3)   | 6.13       | 6.13      | 6.03      | 6.23  | 0.00       |
| Aroclor-1248-4 (4)   | 6.52       | 6.53      | 6.43      | 6.63  | 0.01       |
| Aroclor-1248-5 (5)   | 6.56       | 6.57      | 6.47      | 6.67  | 0.01       |
| Tetrachloro-m-xylene | 4.50       | 4.50      | 4.40      | 4.60  | 0.00       |
| Decachlorobiphenyl   | 10.20      | 10.22     | 10.12     | 10.32 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

Continuing Calib Date: 06/06/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 09:47 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.88       | 4.88      | 4.78      | 4.98 | 0.00       |
| Aroclor-1248-2       | (2) | 5.11       | 5.12      | 5.02      | 5.22 | 0.01       |
| Aroclor-1248-3       | (3) | 5.15       | 5.16      | 5.06      | 5.26 | 0.01       |
| Aroclor-1248-4       | (4) | 5.33       | 5.33      | 5.23      | 5.43 | 0.01       |
| Aroclor-1248-5       | (5) | 5.72       | 5.73      | 5.63      | 5.83 | 0.01       |
| Tetrachloro-m-xylene |     | 3.79       | 3.80      | 3.70      | 3.90 | 0.01       |
| Decachlorobiphenyl   |     | 8.81       | 8.83      | 8.73      | 8.93 | 0.02       |