

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP072698.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.652	5.552	5.752	417.960	500.000	-16.4
Aroclor-1016-2	5.674	5.574	5.774	456.000	500.000	-8.8
Aroclor-1016-3	5.736	5.636	5.836	456.070	500.000	-8.8
Aroclor-1016-4	5.834	5.734	5.934	479.790	500.000	-4.0
Aroclor-1016-5	6.126	6.027	6.227	447.650	500.000	-10.5
Aroclor-1260-1	7.243	7.147	7.347	401.130	500.000	-19.8
Aroclor-1260-2	7.496	7.401	7.601	400.520	500.000	-19.9
Aroclor-1260-3	7.855	7.760	7.960	404.370	500.000	-19.1
Aroclor-1260-4	8.077	7.985	8.185	404.120	500.000	-19.2
Aroclor-1260-5	8.397	8.303	8.503	405.630	500.000	-18.9
Decachlorobiphenyl	10.200	10.113	10.313	42.560	50.000	-14.9
Tetrachloro-m-xylene	4.500	4.396	4.596	47.080	50.000	-5.8

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP072698.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.877	4.777	4.977	493.320	500.000	-1.3
Aroclor-1016-2	4.896	4.795	4.995	506.140	500.000	1.2
Aroclor-1016-3	5.072	4.973	5.173	517.690	500.000	3.5
Aroclor-1016-4	5.114	5.015	5.215	525.930	500.000	5.2
Aroclor-1016-5	5.328	5.229	5.429	520.380	500.000	4.1
Aroclor-1260-1	6.360	6.264	6.464	493.380	500.000	-1.3
Aroclor-1260-2	6.549	6.453	6.653	454.650	500.000	-9.1
Aroclor-1260-3	6.701	6.606	6.806	476.990	500.000	-4.6
Aroclor-1260-4	7.171	7.077	7.277	458.520	500.000	-8.3
Aroclor-1260-5	7.414	7.319	7.519	441.700	500.000	-11.7
Decachlorobiphenyl	8.813	8.723	8.923	49.420	50.000	-1.2
Tetrachloro-m-xylene	3.796	3.691	3.891	51.110	50.000	2.2