

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP072713.D Time Analyzed: 21:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.655	5.552	5.752	423.620	500.000	-15.3
Aroclor-1016-2	5.676	5.574	5.774	461.250	500.000	-7.8
Aroclor-1016-3	5.738	5.636	5.836	461.460	500.000	-7.7
Aroclor-1016-4	5.835	5.734	5.934	477.850	500.000	-4.4
Aroclor-1016-5	6.128	6.027	6.227	471.130	500.000	-5.8
Aroclor-1260-1	7.245	7.147	7.347	464.990	500.000	-7.0
Aroclor-1260-2	7.499	7.401	7.601	452.750	500.000	-9.5
Aroclor-1260-3	7.857	7.760	7.960	448.410	500.000	-10.3
Aroclor-1260-4	8.081	7.985	8.185	439.380	500.000	-12.1
Aroclor-1260-5	8.399	8.303	8.503	450.860	500.000	-9.8
Decachlorobiphenyl	10.203	10.113	10.313	46.120	50.000	-7.8
Tetrachloro-m-xylene	4.503	4.396	4.596	47.490	50.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP072713.D Time Analyzed: 21:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.877	4.777	4.977	512.610	500.000	2.5
Aroclor-1016-2	4.895	4.795	4.995	527.060	500.000	5.4
Aroclor-1016-3	5.071	4.973	5.173	530.630	500.000	6.1
Aroclor-1016-4	5.113	5.015	5.215	532.950	500.000	6.6
Aroclor-1016-5	5.327	5.229	5.429	563.590	500.000	12.7
Aroclor-1260-1	6.358	6.264	6.464	526.630	500.000	5.3
Aroclor-1260-2	6.547	6.453	6.653	501.840	500.000	0.4
Aroclor-1260-3	6.699	6.606	6.806	544.320	500.000	8.9
Aroclor-1260-4	7.169	7.077	7.277	543.610	500.000	8.7
Aroclor-1260-5	7.412	7.319	7.519	535.710	500.000	7.1
Decachlorobiphenyl	8.810	8.723	8.923	54.260	50.000	8.5
Tetrachloro-m-xylene	3.796	3.691	3.891	54.070	50.000	8.1