

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP072726.D Time Analyzed: 02:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.653	5.552	5.752	424.100	500.000	-15.2
Aroclor-1016-2	5.675	5.574	5.774	455.180	500.000	-9.0
Aroclor-1016-3	5.737	5.636	5.836	456.060	500.000	-8.8
Aroclor-1016-4	5.834	5.734	5.934	468.650	500.000	-6.3
Aroclor-1016-5	6.126	6.027	6.227	465.410	500.000	-6.9
Aroclor-1260-1	7.244	7.147	7.347	468.970	500.000	-6.2
Aroclor-1260-2	7.497	7.401	7.601	446.370	500.000	-10.7
Aroclor-1260-3	7.855	7.760	7.960	460.440	500.000	-7.9
Aroclor-1260-4	8.079	7.985	8.185	467.630	500.000	-6.5
Aroclor-1260-5	8.397	8.303	8.503	467.580	500.000	-6.5
Decachlorobiphenyl	10.198	10.113	10.313	47.120	50.000	-5.8
Tetrachloro-m-xylene	4.501	4.396	4.596	47.490	50.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/07/2025

Lab Sample No.: AR1660CCC500 Data File : PP072726.D Time Analyzed: 02:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.876	4.777	4.977	525.000	500.000	5.0
Aroclor-1016-2	4.894	4.795	4.995	529.980	500.000	6.0
Aroclor-1016-3	5.071	4.973	5.173	530.470	500.000	6.1
Aroclor-1016-4	5.113	5.015	5.215	525.130	500.000	5.0
Aroclor-1016-5	5.326	5.229	5.429	585.290	500.000	17.1
Aroclor-1260-1	6.358	6.264	6.464	533.740	500.000	6.7
Aroclor-1260-2	6.547	6.453	6.653	497.670	500.000	-0.5
Aroclor-1260-3	6.698	6.606	6.806	528.580	500.000	5.7
Aroclor-1260-4	7.169	7.077	7.277	527.330	500.000	5.5
Aroclor-1260-5	7.411	7.319	7.519	525.850	500.000	5.2
Decachlorobiphenyl	8.809	8.723	8.923	56.500	50.000	13.0
Tetrachloro-m-xylene	3.796	3.691	3.891	53.410	50.000	6.8