

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 06/09/2025

Lab Sample No.: AR1660CCC500 Data File : PP072748.D Time Analyzed: 15:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.652	5.552	5.752	452.290	500.000	-9.5
Aroclor-1016-2	5.674	5.574	5.774	498.820	500.000	-0.2
Aroclor-1016-3	5.736	5.636	5.836	493.520	500.000	-1.3
Aroclor-1016-4	5.833	5.734	5.934	510.500	500.000	2.1
Aroclor-1016-5	6.126	6.027	6.227	501.700	500.000	0.3
Aroclor-1260-1	7.243	7.147	7.347	479.450	500.000	-4.1
Aroclor-1260-2	7.497	7.401	7.601	470.330	500.000	-5.9
Aroclor-1260-3	7.855	7.760	7.960	495.010	500.000	-1.0
Aroclor-1260-4	8.079	7.985	8.185	484.380	500.000	-3.1
Aroclor-1260-5	8.397	8.303	8.503	505.090	500.000	1.0
Decachlorobiphenyl	10.198	10.113	10.313	51.700	50.000	3.4
Tetrachloro-m-xylene	4.500	4.396	4.596	52.050	50.000	4.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 06/09/2025

Lab Sample No.: AR1660CCC500 Data File : PP072748.D Time Analyzed: 15:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.777	4.977	531.190	500.000	6.2
Aroclor-1016-2	4.892	4.795	4.995	544.200	500.000	8.8
Aroclor-1016-3	5.069	4.973	5.173	547.570	500.000	9.5
Aroclor-1016-4	5.110	5.015	5.215	552.240	500.000	10.4
Aroclor-1016-5	5.324	5.229	5.429	544.880	500.000	9.0
Aroclor-1260-1	6.356	6.264	6.464	546.980	500.000	9.4
Aroclor-1260-2	6.546	6.453	6.653	514.030	500.000	2.8
Aroclor-1260-3	6.697	6.606	6.806	538.170	500.000	7.6
Aroclor-1260-4	7.167	7.077	7.277	541.700	500.000	8.3
Aroclor-1260-5	7.410	7.319	7.519	548.420	500.000	9.7
Decachlorobiphenyl	8.807	8.723	8.923	55.900	50.000	11.8
Tetrachloro-m-xylene	3.793	3.691	3.891	54.700	50.000	9.4