

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2237

**Analytical Method:** 8082A

**Client:** ENTACT

**Datafile :** PP072739.D

| Lab Sample ID             | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD     |
|---------------------------|-----------|-------|--------|-------|-----|-----|------|------|---------|-----------|---------|
|                           |           |       |        |       |     |     |      | Qual | Low     | High      |         |
| PB168325BSD<br>(Column 1) | AR1016    | 5     | 4.50   | ug/L  | 90  | 7   |      |      | 40 (77) | 140 (107) | 20 (20) |
|                           | AR1260    | 5     | 4.40   | ug/L  | 88  | 12  |      |      | 40 (66) | 140 (113) | 20 (20) |
| PB168325BSD<br>(Column 2) | AR1016    | 5     | 4.80   | ug/L  | 96  | 0   |      |      | 40 (77) | 140 (107) | 20 (20) |
|                           | AR1260    | 5     | 4.80   | ug/L  | 96  | 0   |      |      | 40 (66) | 140 (113) | 20 (20) |