

Surrogate Summary

SW-846

SDG No.: Q2237

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168323BL	PB168323BL	2-Fluorophenol	150	146	97		15 (23)	110 (138)
		Phenol-d6	150	137	92		15 (10)	110 (134)
		Nitrobenzene-d5	100	85.2	85		30 (67)	130 (132)
		2-Fluorobiphenyl	100	84.3	84		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	143	95		15 (44)	110 (137)
		Terphenyl-d14	100	90.1	90		30 (42)	130 (152)
PB168323BS	PB168323BS	2-Fluorophenol	150	143	95		15 (23)	110 (138)
		Phenol-d6	150	136	91		15 (10)	110 (134)
		Nitrobenzene-d5	100	79.4	79		30 (67)	130 (132)
		2-Fluorobiphenyl	100	77.9	78		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	88		15 (44)	110 (137)
		Terphenyl-d14	100	79.2	79		30 (42)	130 (152)
Q2230-03MS	GW-MW01-060425MS	2-Fluorophenol	150	79.2	53		15 (23)	110 (138)
		Phenol-d6	150	50.7	34		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.5	91		30 (67)	130 (132)
		2-Fluorobiphenyl	100	85.6	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	162	108		15 (44)	110 (137)
		Terphenyl-d14	100	80.4	80		30 (42)	130 (152)
Q2230-04MSD	GW-MW01-060425MSD	2-Fluorophenol	150	74.2	49		15 (23)	110 (138)
		Phenol-d6	150	46.3	31		15 (10)	110 (134)
		Nitrobenzene-d5	100	88.0	88		30 (67)	130 (132)
		2-Fluorobiphenyl	100	85.9	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	159	106		15 (44)	110 (137)
		Terphenyl-d14	100	89.3	89		30 (42)	130 (152)
Q2237-02	TW-WTS-10	2-Fluorophenol	150	76.8	51		15 (23)	110 (138)
		Phenol-d6	150	50.6	34		15 (10)	110 (134)
		Nitrobenzene-d5	100	83.7	84		30 (67)	130 (132)
		2-Fluorobiphenyl	100	83.2	83		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	157	105		15 (44)	110 (137)
		Terphenyl-d14	100	80.0	80		30 (42)	130 (152)