

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ENTA05

Lab Code: CHEM

SAS No.: Q2237 SDG NO.: Q2237

Lab File ID: BP024859.D

DFTPP Injection Date: 06/06/2025

Instrument ID: BNA_P

DFTPP Injection Time: 09:49

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.2
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	36.9
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	31.2
365	Greater than 1% of mass 198	4.6
441	Present, but less than mass 443	13.1
442	Greater than 50% of mass 198	84
443	15.0 - 24.0% of mass 442	16.1 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024860.D	06/06/2025	10:30
SSTDICC005	SSTDICC005	BP024861.D	06/06/2025	11:11
SSTDICC010	SSTDICC010	BP024862.D	06/06/2025	11:52
SSTDICC020	SSTDICC020	BP024863.D	06/06/2025	12:33
SSTDICCC040	SSTDICCC040	BP024864.D	06/06/2025	13:14
SSTDICC050	SSTDICC050	BP024865.D	06/06/2025	13:56
SSTDICC060	SSTDICC060	BP024866.D	06/06/2025	14:37
SSTDICC080	SSTDICC080	BP024867.D	06/06/2025	15:18