

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG No.: Q2237

Instrument ID: MSVOA_N Calibration Date/Time: 06/06/2025 14:03

Lab File ID: VN086865.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.015	1.933		-4.07	
Carbon Tetrachloride	0.433	0.409		-5.54	
Chloroform	1.118	1.061		-5.1	
1,1,1-Trichloroethane	0.951	0.895		-5.89	
Benzene	1.444	1.345		-6.86	
Toluene	0.882	0.835		-5.33	
Tetrachloroethene	0.316	0.294		-6.96	
Ethyl Benzene	1.900	1.796		-5.47	
m/p-Xylenes	0.727	0.701		-3.58	
o-Xylene	0.696	0.674		-3.16	
1,2-Dichloroethane-d4	0.669	0.500		-25.26	
Dibromofluoromethane	0.296	0.219		-26.01	
Toluene-d8	1.173	0.861		-26.6	
4-Bromofluorobenzene	0.436	0.325		-25.46	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.