

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	TW-WTS-10		SDG No.:	Q2237	
Lab Sample ID:	Q2237-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group4	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086875.D	1		06/06/25 18:42	VN060625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.44	J	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
100-41-4	Ethyl Benzene	0.43	J	0.13	1.00	ug/L
1330-20-7	Total Xylenes	0.26	U	0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.2		70 (74) - 130 (125)	94%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		70 (75) - 130 (124)	98%	SPK: 50
2037-26-5	Toluene-d8	51.5		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.1		70 (77) - 130 (121)	96%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	336000	8.235			
540-36-3	1,4-Difluorobenzene	595000	9.106			
3114-55-4	Chlorobenzene-d5	514000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	237000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products