

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN060625**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 6/9/2025 8:08:23 AM               |
| Supervise By | Maresh Dadoda | Supervise On      | 6/9/2025 1:13:51 PM               |
| SubDirectory | VN060625      | HP Acquire Method | HP Processing Method 82N060625W.M |

| STD. NAME                                                                                          | STD REF.# |
|----------------------------------------------------------------------------------------------------|-----------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134155  |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134156  |

| Sr# | SampleID     | ClientID            | Data File Name | Date-Time         | Comment                        | Operator | Status |
|-----|--------------|---------------------|----------------|-------------------|--------------------------------|----------|--------|
| 1   | BFB          | BFB                 | VN086861.D     | 06 Jun 2025 07:59 |                                | JC\MD    | Ok     |
| 2   | VSTDIC001    | VSTDIC001           | VN086862.D     | 06 Jun 2025 12:44 | Method failed for com.#13      | JC\MD    | Ok,M   |
| 3   | VSTDIC005    | VSTDIC005           | VN086863.D     | 06 Jun 2025 13:17 |                                | JC\MD    | Ok,M   |
| 4   | VSTDIC020    | VSTDIC020           | VN086864.D     | 06 Jun 2025 13:40 |                                | JC\MD    | Ok,M   |
| 5   | VSTDIC050    | VSTDIC050           | VN086865.D     | 06 Jun 2025 14:03 |                                | JC\MD    | Ok,M   |
| 6   | VSTDIC100    | VSTDIC100           | VN086866.D     | 06 Jun 2025 14:26 |                                | JC\MD    | Ok,M   |
| 7   | VSTDIC150    | VSTDIC150           | VN086867.D     | 06 Jun 2025 14:49 |                                | JC\MD    | Ok,M   |
| 8   | IBLK         | IBLK                | VN086868.D     | 06 Jun 2025 15:12 |                                | JC\MD    | Ok     |
| 9   | VSTDICV050   | ICVVN060625         | VN086869.D     | 06 Jun 2025 15:54 |                                | JC\MD    | Ok,M   |
| 10  | VN0606WBL01  | VN0606WBL01         | VN086870.D     | 06 Jun 2025 16:47 |                                | JC\MD    | Ok     |
| 11  | VN0606WBL02  | VN0606WBL02         | VN086871.D     | 06 Jun 2025 17:10 |                                | JC\MD    | Ok     |
| 12  | VN0606WBS01  | VN0606WBS01         | VN086872.D     | 06 Jun 2025 17:33 |                                | JC\MD    | Ok,M   |
| 13  | VN0606WBSD01 | VN0606WBSD01        | VN086873.D     | 06 Jun 2025 17:56 |                                | JC\MD    | Ok,M   |
| 14  | Q2254-01     | BP-VPB-182-GW-810-8 | VN086874.D     | 06 Jun 2025 18:19 | vial A pH<2 endccc out of tune | JC\MD    | Not Ok |
| 15  | Q2237-02     | TW-WTS-10           | VN086875.D     | 06 Jun 2025 18:42 | vial A pH<2                    | JC\MD    | Ok     |
| 16  | Q2216-02     | 3887                | VN086876.D     | 06 Jun 2025 19:05 | vial A pH<2                    | JC\MD    | Ok     |
| 17  | Q2216-03     | 3888                | VN086877.D     | 06 Jun 2025 19:28 | vial A pH<2                    | JC\MD    | Ok     |
| 18  | Q2216-04     | 3864                | VN086878.D     | 06 Jun 2025 19:51 | vial A pH<2                    | JC\MD    | Ok     |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN060625**

|                                                                                                    |                  |                   |                                   |
|----------------------------------------------------------------------------------------------------|------------------|-------------------|-----------------------------------|
| Review By                                                                                          | John Carlone     | Review On         | 6/9/2025 8:08:23 AM               |
| Supervise By                                                                                       | Maresh Dadoda    | Supervise On      | 6/9/2025 1:13:51 PM               |
| SubDirectory                                                                                       | VN060625         | HP Acquire Method | HP Processing Method 82N060625W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b> |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134155         |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134156         |                   |                                   |

|    |            |              |            |                   |                         |       |        |
|----|------------|--------------|------------|-------------------|-------------------------|-------|--------|
| 19 | Q2216-05   | 3865         | VN086879.D | 06 Jun 2025 20:13 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 20 | Q2216-06   | 3851         | VN086880.D | 06 Jun 2025 20:36 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 21 | Q2206-04   | TP-1         | VN086881.D | 06 Jun 2025 20:59 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 22 | Q2242-04   | TP09-MHJ     | VN086882.D | 06 Jun 2025 21:21 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 23 | Q2192-01   | SB-1         | VN086883.D | 06 Jun 2025 21:44 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 24 | Q2198-02   | B-202-SB02   | VN086884.D | 06 Jun 2025 22:07 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 25 | Q2198-04   | B-207-SB02   | VN086885.D | 06 Jun 2025 22:29 | vial A pH<2 Out of Tune | JC\MD | Not Ok |
| 26 | VSTDCCC050 | VSTDCCC050EC | VN086886.D | 06 Jun 2025 22:52 | Out of Tune             | JC\MD | Not Ok |

M : Manual Integration