

Surrogate Summary

SDG No.: Q2237

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2237-02	TW-WTS-10	1,2-Dichloroethane-d4	50	47.2	94	70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98	70 (75)	130 (124)
		Toluene-d8	50	51.5	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.1	96	70 (77)	130 (121)
VN0606WBL02	VN0606WBL02	1,2-Dichloroethane-d4	50	47.3	95	70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.6	95	70 (77)	130 (121)
VN0606WBS01	VN0606WBS01	1,2-Dichloroethane-d4	50	50.4	101	70 (74)	130 (125)
		Dibromofluoromethane	50	52.8	106	70 (75)	130 (124)
		Toluene-d8	50	51.7	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.3	105	70 (77)	130 (121)
VN0606WBSD0	VN0606WBSD01	1,2-Dichloroethane-d4	50	50.6	101	70 (74)	130 (125)
		Dibromofluoromethane	50	52.0	104	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.2	104	70 (77)	130 (121)

