

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237
Lab File ID: VN086865.D Date Analyzed: 06/06/2025
Instrument ID: MSVOA_N Time Analyzed: 14:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	207354	8.24	378587	9.11	332766	11.87
UPPER LIMIT	414708	8.736	757174	9.606	665532	12.365
LOWER LIMIT	103677	7.736	189294	8.606	166383	11.365
EPA SAMPLE NO.						
TW-WTS-10	335826	8.24	595271	9.11	514382	11.87
TW-WTS-10	335826	8.24	595271	9.11	514382	11.87
VN0606WBL02	344048	8.24	613851	9.11	529685	11.87
VN0606WBL02	344048	8.24	613851	9.11	529685	11.87
VN0606WBS01	225012	8.23	402533	9.11	357932	11.87
VN0606WBS01	225012	8.23	402533	9.11	357932	11.87
VN0606WBSD01	219730	8.24	395186	9.11	346762	11.87
VN0606WBSD01	219730	8.24	395186	9.11	346762	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.