

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2237 SAS No.: Q2237 SDG NO.: Q2237
 Lab File ID: VN086865.D Date Analyzed: 06/06/2025
 Instrument ID: MSVOA_N Time Analyzed: 14:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	167532	13.794				
UPPER LIMIT	335064	14.294				
LOWER LIMIT	83766	13.294				
EPA SAMPLE NO.						
TW-WTS-10	237388	13.79				
TW-WTS-10	237388	13.79				
VN0606WBL02	242229	13.79				
VN0606WBL02	242229	13.79				
VN0606WBS01	176650	13.79				
VN0606WBS01	176650	13.79				
VN0606WBSD01	171633	13.79				
VN0606WBSD01	171633	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.