



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB136022

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.424	0.4	106	90-110	06/05/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.373	0.4	93	90-110	06/05/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.383	0.4	96	90-110	06/05/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB136039

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	06/06/2025
Sample ID: CCV2 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV4 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV5 Ammonia as N	mg/L	1	1	100	90-110	06/06/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB136063

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.095	0.099	96	90-110	06/09/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	06/09/2025
Sample ID: Cyanide	CCV2	mg/L	0.23	0.25	92	90-110	06/09/2025
Sample ID: Cyanide	CCV3	mg/L	0.24	0.25	96	90-110	06/09/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB136064

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
COD		mg/L	50.962	50	102	95-105	05/28/2025
Sample ID:	CCV1						
COD		mg/L	47.915	50	96	95-105	06/09/2025
Sample ID:	CCV2						
COD		mg/L	48.931	50	98	95-105	06/09/2025
Sample ID:	CCV3						
COD		mg/L	49.946	50	100	95-105	06/09/2025