

Preparation Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Project: Rotor Clip NJ WTD - 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136022BL Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	06/05/2025
Sample ID: LB136028BL BOD5	mg/L	< 0.2000	0.2000	U	0.20	2.0	06/05/2025
Sample ID: LB136054BL TSS	mg/L	1	2.0000	J	1	4	06/09/2025
Sample ID: LB136056BL Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	06/09/2025
Sample ID: LB136064BL COD	mg/L	< 5.0000	5.0000	U	1.5	10.0	06/09/2025
Sample ID: PB168326BL Ammonia as N	mg/L	0.03	0.0500	J	0.03	0.1	06/06/2025
Sample ID: PB168347BL Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/09/2025