

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2191-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	90.8		90.9		1	0.11		06/09/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2197-01
Client ID:	DSN002DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
COD	mg/L	+/-20	36.7		34.7		1	5.6		06/09/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2197-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	79.4		80.4		1	1.25		06/09/2025

Duplicate Sample Summary

Client: VERINA CONSULTING GROUP, LLC	SDG No.: Q2243
Project: Rotor Clip NJ WTD - 2025	Sample ID: Q2203-01
Client ID: 001-WILLETS-PT-BLVD(JUNE)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	5.50	OR	5.40	OR	1	2		06/06/2025
Ammonia as N	mg/L	+/-20	5.30	D	5.30	D	5	0		06/06/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2203-01
Client ID:	001-WILLETS-PT-BLVD(JUNE)MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	6.70	OR	6.50	OR	1	3		06/06/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2229-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	444		424		1	4.61		06/05/2025
TSS	mg/L	+/-5	121		121		1	0		06/09/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2243-01
Client ID:	WATER-TREATMENT-DISCHARGEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.078	J	0.088	J	1	12.05		06/05/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2243-01
Client ID:	WATER-TREATMENT-DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.46		0.47		1	2.13		06/05/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2243-02
Client ID:	WATER-TREATMENT-DISCHARGEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0046	J	0.0044	J	1	4		06/09/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2243-02
Client ID:	WATER-TREATMENT-DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.040		0.039		1	3		06/09/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2243-02
Client ID:	WATER-TREATMENT-DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	23.9		24.3		1	1.66		06/09/2025