

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136022

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136022BS							
Residual Chlorine	mg/L	0.4	0.41		103	1	90-110	06/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136028

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136028BS							
BOD5	mg/L	198	210		106	1	84.6-115.4	06/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136054

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136054BS							
TSS	mg/L	550	533		97	1	90-110	06/09/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136056

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136056BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	06/09/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136064

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136064BS							
COD	mg/L	50	52.0		104	1	90-110	06/09/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136039

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168326BS							
Ammonia as N	mg/L	1	1.10		110	1	90-110	06/06/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2243
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136063

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168347BS							
Cyanide	mg/L	0.1	0.096		96	1	85-115	06/09/2025