

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	502	510	98	90 - 110	P	06/06/2025	13:16	LB136052
	Chromium	522	520	100	90 - 110	P	06/06/2025	13:16	LB136052
	Copper	527	510	103	90 - 110	P	06/06/2025	13:16	LB136052
	Lead	980	1000	98	90 - 110	P	06/06/2025	13:16	LB136052
	Nickel	533	530	100	90 - 110	P	06/06/2025	13:16	LB136052
	Silver	243	250	97	90 - 110	P	06/06/2025	13:16	LB136052
	Zinc	991	1000	99	90 - 110	P	06/06/2025	13:16	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	6.02	6.0	100	80 - 120	P	06/06/2025	13:29	LB136052
	Chromium	9.45	10.0	94	80 - 120	P	06/06/2025	13:29	LB136052
	Copper	21.0	20.0	105	80 - 120	P	06/06/2025	13:29	LB136052
	Lead	13.9	12.0	116	80 - 120	P	06/06/2025	13:29	LB136052
	Nickel	39.7	40.0	99	80 - 120	P	06/06/2025	13:29	LB136052
	Silver	9.79	10.0	98	80 - 120	P	06/06/2025	13:29	LB136052
	Zinc	43.4	40.0	108	80 - 120	P	06/06/2025	13:29	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Chromium	995	1000	100	90 - 110	P	06/06/2025	14:05	LB136052
	Copper	1250	1250	100	90 - 110	P	06/06/2025	14:05	LB136052
	Lead	4910	5000	98	90 - 110	P	06/06/2025	14:05	LB136052
	Nickel	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Silver	1240	1250	99	90 - 110	P	06/06/2025	14:05	LB136052
	Zinc	2480	2500	99	90 - 110	P	06/06/2025	14:05	LB136052
CCV02	Cadmium	2470	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Chromium	993	1000	99	90 - 110	P	06/06/2025	14:56	LB136052
	Copper	1280	1250	102	90 - 110	P	06/06/2025	14:56	LB136052
	Lead	4960	5000	99	90 - 110	P	06/06/2025	14:56	LB136052
	Nickel	2480	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Silver	1250	1250	100	90 - 110	P	06/06/2025	14:56	LB136052
	Zinc	2490	2500	100	90 - 110	P	06/06/2025	14:56	LB136052
CCV03	Cadmium	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Chromium	998	1000	100	90 - 110	P	06/06/2025	15:50	LB136052
	Copper	1280	1250	103	90 - 110	P	06/06/2025	15:50	LB136052
	Lead	5000	5000	100	90 - 110	P	06/06/2025	15:50	LB136052
	Nickel	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Silver	1240	1250	99	90 - 110	P	06/06/2025	15:50	LB136052
	Zinc	2470	2500	99	90 - 110	P	06/06/2025	15:50	LB136052
CCV04	Cadmium	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Chromium	993	1000	99	90 - 110	P	06/06/2025	16:46	LB136052
	Copper	1240	1250	99	90 - 110	P	06/06/2025	16:46	LB136052
	Lead	4890	5000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Nickel	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Silver	1230	1250	98	90 - 110	P	06/06/2025	16:46	LB136052
	Zinc	2460	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
CCV05	Cadmium	2390	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Chromium	977	1000	98	90 - 110	P	06/06/2025	17:40	LB136052
	Copper	1250	1250	100	90 - 110	P	06/06/2025	17:40	LB136052
	Lead	4810	5000	96	90 - 110	P	06/06/2025	17:40	LB136052
	Nickel	2410	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Silver	1230	1250	99	90 - 110	P	06/06/2025	17:40	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC SDG No.: Q2243
Contract: VERI01 Lab Code: CHEM Case No.: Q2243 SAS No.: Q2243
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Zinc	2450	2500	98	90 - 110	P	06/06/2025	17:40	LB136052
CCV06	Cadmium	2300	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Chromium	907	1000	91	90 - 110	P	06/07/2025	03:03	LB136052
	Copper	1170	1250	94	90 - 110	P	06/07/2025	03:03	LB136052
	Lead	4580	5000	92	90 - 110	P	06/07/2025	03:03	LB136052
	Nickel	2300	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Silver	1120	1250	90	90 - 110	P	06/07/2025	03:03	LB136052
	Zinc	2190	2500	88	90 - 110	P	06/07/2025	03:03	LB136052
CCV07	Cadmium	2330	2500	93	90 - 110	P	06/07/2025	03:40	LB136052
	Chromium	951	1000	95	90 - 110	P	06/07/2025	03:40	LB136052
	Copper	1180	1250	94	90 - 110	P	06/07/2025	03:40	LB136052
	Lead	4630	5000	92	90 - 110	P	06/07/2025	03:40	LB136052
	Nickel	2330	2500	93	90 - 110	P	06/07/2025	03:40	LB136052
	Silver	1180	1250	95	90 - 110	P	06/07/2025	03:40	LB136052
	Zinc	2300	2500	92	90 - 110	P	06/07/2025	03:40	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC SDG No.: Q2243

Contract: VERI01 Lab Code: CHEM Case No.: Q2243 SAS No.: Q2243

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	1920	2000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Chromium	794	800	99	90 - 110	P	06/10/2025	14:37	LB136097
	Copper	994	1000	99	90 - 110	P	06/10/2025	14:37	LB136097
	Lead	3800	4000	95	90 - 110	P	06/10/2025	14:37	LB136097
	Nickel	1930	2000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Silver	978	1000	98	90 - 110	P	06/10/2025	14:37	LB136097
	Zinc	1980	2000	99	90 - 110	P	06/10/2025	14:37	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	6.31	6.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Chromium	10.8	10.0	108	80 - 120	P	06/10/2025	14:41	LB136097
	Copper	22.6	20.0	113	80 - 120	P	06/10/2025	14:41	LB136097
	Lead	12.8	12.0	107	80 - 120	P	06/10/2025	14:41	LB136097
	Nickel	41.6	40.0	104	80 - 120	P	06/10/2025	14:41	LB136097
	Silver	10.5	10.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Zinc	43.7	40.0	109	80 - 120	P	06/10/2025	14:41	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	15:30	LB136097
	Copper	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Lead	5030	5000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Nickel	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Silver	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Zinc	2560	2500	102	90 - 110	P	06/10/2025	15:30	LB136097
CCV02	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	16:28	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	16:28	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Lead	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Nickel	2540	2500	101	90 - 110	P	06/10/2025	16:28	LB136097
	Silver	1270	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	16:28	LB136097
CCV03	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	17:36	LB136097
	Lead	5080	5000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Nickel	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Silver	1280	1250	103	90 - 110	P	06/10/2025	17:36	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	17:36	LB136097
CCV04	Cadmium	2580	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	18:28	LB136097
	Copper	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Lead	5160	5000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Nickel	2570	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Zinc	2540	2500	101	90 - 110	P	06/10/2025	18:28	LB136097
CCV05	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	19:18	LB136097
	Lead	5090	5000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Nickel	2540	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Silver	1260	1250	101	90 - 110	P	06/10/2025	19:18	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Zinc	2510	2500	100	90 - 110	P	06/10/2025	19:18	LB136097
CCV06	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	20:09	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	20:09	LB136097
	Lead	5110	5000	102	90 - 110	P	06/10/2025	20:09	LB136097
	Nickel	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	20:09	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	20:09	LB136097
CCV07	Cadmium	2600	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	21:04	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Lead	5210	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Zinc	2540	2500	102	90 - 110	P	06/10/2025	21:04	LB136097
CCV08	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Copper	1280	1250	103	90 - 110	P	06/10/2025	22:04	LB136097
	Lead	5130	5000	102	90 - 110	P	06/10/2025	22:04	LB136097
	Nickel	2560	2500	103	90 - 110	P	06/10/2025	22:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	22:04	LB136097
	Zinc	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
CCV09	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:56	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	22:56	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	22:56	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	22:56	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	22:56	LB136097
CCV10	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Chromium	1060	1000	106	90 - 110	P	06/10/2025	23:57	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	23:57	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	23:57	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q2243

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Silver	1310	1250	104	90 - 110	P	06/10/2025	23:57	LB136097
	Zinc	2630	2500	105	90 - 110	P	06/10/2025	23:57	LB136097
CCV11	Cadmium	2570	2500	103	90 - 110	P	06/11/2025	01:35	LB136097
	Chromium	1040	1000	104	90 - 110	P	06/11/2025	01:35	LB136097
	Copper	1290	1250	103	90 - 110	P	06/11/2025	01:35	LB136097
	Lead	5160	5000	103	90 - 110	P	06/11/2025	01:35	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
	Silver	1280	1250	103	90 - 110	P	06/11/2025	01:35	LB136097
	Zinc	2600	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
CCV12	Cadmium	2400	2500	96	90 - 110	P	06/11/2025	02:21	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/11/2025	02:21	LB136097
	Copper	1200	1250	96	90 - 110	P	06/11/2025	02:21	LB136097
	Lead	4810	5000	96	90 - 110	P	06/11/2025	02:21	LB136097
	Nickel	2420	2500	97	90 - 110	P	06/11/2025	02:21	LB136097
	Silver	1280	1250	102	90 - 110	P	06/11/2025	02:21	LB136097
	Zinc	2590	2500	104	90 - 110	P	06/11/2025	02:21	LB136097