



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	13:33	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	13:33	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	13:33	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	13:33	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	13:33	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	13:33	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	13:33	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	14:09	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	14:09	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	14:09	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	14:09	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	14:09	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
CCB02	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	15:00	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	15:00	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	15:00	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	15:00	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	15:00	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
CCB03	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	15:57	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	15:57	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	15:57	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	15:57	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	15:57	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
CCB04	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	16:51	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	16:51	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	16:51	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	16:51	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	16:51	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
CCB05	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	17:46	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	17:46	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	17:46	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	17:46	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	17:46	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
CCB06	Cadmium	0.50	+/-3	U	6.00	P	06/07/2025	03:07	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/07/2025	03:07	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/07/2025	03:07	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/07/2025	03:07	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	3.06	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/07/2025	03:07	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
CCB07	Cadmium	0.50	+/-3	U	6.00	P	06/07/2025	03:44	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/07/2025	03:44	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/07/2025	03:44	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/07/2025	03:44	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/07/2025	03:44	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/07/2025	03:44	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/07/2025	03:44	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	14:45	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	14:45	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	14:45	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	14:45	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	14:45	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC		SDG No.: Q2243							
Contract: VERI01	Lab Code: CHEM	Case No.: Q2243	SAS No.: Q2243						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	15:35	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	15:35	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	15:35	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	15:35	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
CCB02	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	16:32	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	16:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	16:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	16:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
CCB03	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	17:40	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	17:40	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	17:40	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	17:40	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
CCB04	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	18:32	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	18:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	18:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	18:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
CCB05	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	19:22	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	19:22	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	19:22	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	19:22	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
CCB06	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	20:14	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	20:14	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	20:14	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	20:14	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
CCB07	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	21:08	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	21:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
CCB08	Lead	2.30	+/-6	U	12.0	P	06/10/2025	21:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	21:08	LB136097
CCB09	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	22:08	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	22:08	LB136097
CCB10	Copper	4.60	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	22:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
CCB11	Silver	1.62	+/-5	U	10.0	P	06/10/2025	22:08	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	23:01	LB136097
CCB12	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	23:01	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	23:01	LB136097
CCB13	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	23:01	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
CCB14	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	00:02	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	00:02	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
CCB15	Lead	2.30	+/-6	U	12.0	P	06/11/2025	00:02	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	00:02	LB136097
CCB16	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	01:39	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	01:39	LB136097
CCB17	Copper	4.60	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	01:39	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
CCB18	Silver	1.62	+/-5	U	10.0	P	06/11/2025	01:39	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	02:25	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	02:25	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	02:25	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	02:25	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	02:25	LB136097