

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168284BL	WATER			Batch Number:	PB168284		Prep Date:	06/05/2025	
	Cadmium	0.25	<1.5	U	3.00	P	06/10/2025	18:44	LB136097
	Chromium	1.06	<2.5	U	5.00	P	06/06/2025	14:22	LB136052
	Copper	2.30	<5	U	10.0	P	06/10/2025	18:44	LB136097
	Lead	1.15	<3	U	6.00	P	06/06/2025	14:22	LB136052
	Nickel	1.53	<10	U	20.0	P	06/10/2025	18:44	LB136097
	Silver	0.81	<2.5	U	5.00	P	06/06/2025	14:22	LB136052
	Zinc	8.33	<10	U	20.0	P	06/10/2025	18:44	LB136097