

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q2243
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243
Matrix: Water **Sample ID:** Q2197-05 **Client ID:** DSN003DUP
Percent Solids for Sample: NA **Duplicate ID** Q2197-05DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	2.00	J	2.01	J	0		P
Copper	ug/L	20	5.15	J	5.24	J	2		P
Lead	ug/L	20	6.00	U	6.00	U			P
Nickel	ug/L	20	10.1	J	10.9	J	7		P
Silver	ug/L	20	5.00	U	5.00	U			P
Zinc	ug/L	20	73.5		74.5		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q2243
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q2243 **SAS No.:** Q2243
Matrix: Water **Sample ID:** Q2197-05MS **Client ID:** DSN003MSD
Percent Solids for Sample: NA **Duplicate ID** Q2197-05MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	102		105		3		P
Chromium	ug/L	20	208		216		4		P
Copper	ug/L	20	159		164		3		P
Lead	ug/L	20	480		493		3		P
Nickel	ug/L	20	266		273		3		P
Silver	ug/L	20	34.5		36.9		7		P
Zinc	ug/L	20	182		192		5		P