

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2243

Contract: VERI01

Lab Code: CHEM

Case No.: Q2243

SAS No.: Q2243

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168284BS							
Cadmium	ug/L	100	111		111	80 - 120	P
Chromium	ug/L	200	204		102	80 - 120	P
Copper	ug/L	150	170		113	80 - 120	P
Lead	ug/L	500	490		98	80 - 120	P
Nickel	ug/L	250	277		111	80 - 120	P
Silver	ug/L	37.5	41.6		111	80 - 120	P
Zinc	ug/L	100	114		114	80 - 120	P