

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

DSN003L

Lab Name: Chemtech Consulting Group

Contract: VERI01

Lab Code: CHEM **Lb No.:** lb136097 **Lab Sample ID :** Q2197-05L **SDG No.:** Q2243

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Cadmium	3.00	U	15.0	U			P
Chromium	2.00	J	25.0	U	100.0		P
Copper	5.15	J	13.1	J	154		P
Lead	6.00	U	30.0	U			P
Nickel	10.1	J	12.9	J	28		P
Silver	5.00	U	25.0	U			P
Zinc	73.5		72.7	J	1		P