ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

MATRIX: Water

ORDER ID: Q2249

METHOD: 8260D NA NO YES 1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) 2. GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series. 4. GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series. 5. GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration File ID VX046657.D met the requirements except for Vinyl Chloride is failing marginally low while 1,2-Dichloroethane-d4 which is not our target compound, therefore no corrective action taken. 6. Blank Contamination - If yes, list compounds and concentrations in each blank: 7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. 8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. 9. Internal Standard Area/Retention Time Shift Meet Criteria Comments: The Internal Standards Areas met the acceptable requirements except for MW-06-6.5-060525, MW-06-6.5-060525RE, sample was reanalyzed to confirm the failure and reported. 10. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

S:
۱

Sample MW-06-6.5-060525 was diluted at straight dilution after checking past history of this sample.

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW	Date	