

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2249 SAS No.: Q2249 SDG No.: Q2249

Instrument ID: MSVOA_X Calibration Date/Time: 06/11/2025 14:46

Lab File ID: VX046636.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:42 12:57

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.630	0.569		-9.68	20
1,1-Dichloroethene	0.594	0.596		0.34	20
1,1-Dichloroethane	1.281	1.325	0.1	3.43	20
cis-1,2-Dichloroethene	0.773	0.799		3.36	20
1,1,1-Trichloroethane	1.148	1.186		3.31	20
Benzene	1.479	1.534		3.72	20
1,2-Dichloroethane	0.615	0.621		0.98	20
Trichloroethene	0.376	0.379		0.8	20
1,1,2-Trichloroethane	0.365	0.403		10.41	20
Tetrachloroethene	0.334	0.330		-1.2	20
1,2-Dichloroethane-d4	0.890	0.830		-6.74	20
Dibromofluoromethane	0.368	0.382		3.8	20
Toluene-d8	1.212	1.219		0.58	20
4-Bromofluorobenzene	0.502	0.503		0.2	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.