

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2249 SAS No.: Q2249 SDG NO.: Q2249
Lab File ID: VX046636.D Date Analyzed: 06/11/2025
Instrument ID: MSVOA_X Time Analyzed: 14:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	77468	5.56	133266	6.76	119186	10.06
UPPER LIMIT	154936	6.062	266532	7.263	238372	10.555
LOWER LIMIT	38734	5.062	66633	6.263	59593	9.555
EPA SAMPLE NO.						
MW-06-6.5-060525	113972	5.57	227958	6.78	237036	10.06
VX0611WBL02	73470	5.56	147517	6.77	153084	10.06
VX0611WBS02	79398	5.57	136732	6.78	125565	10.06
VX0611WBSD02	75823	5.56	130739	6.76	119823	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.