

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2249 SAS No.: Q2249 SDG NO.: Q2249
 Lab File ID: VX046657.D Date Analyzed: 06/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:11
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	82603	5.56	125406	6.77	118101	10.06
UPPER LIMIT	165206	6.056	250812	7.269	236202	10.555
LOWER LIMIT	41301.5	5.056	62703	6.269	59050.5	9.555
EPA SAMPLE NO.						
MW-06-6.5-060525RE	130843	5.56	256854 *	6.77	263328 *	10.06
VX0612WBL01	92224	5.57	180952	6.78	188641	10.06
VX0612WBS01	85578	5.56	144622	6.77	130715	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.