

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2249 SAS No.: Q2249 SDG NO.: Q2249
 Lab File ID: VX046636.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 14:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	60295	12.018				
UPPER LIMIT	120590	12.518				
LOWER LIMIT	30147.5	11.518				
EPA SAMPLE NO.						
MW-06-6.5-060525	124632 *	12.02				
VX0611WBL02	77047	12.02				
VX0611WBS02	66542	12.02				
VX0611WBSD02	62816	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.