

## **DATA PACKAGE VOLATILE ORGANICS**

**PROJECT NAME : FORMER SCHLUMBERGER STC PTC SITE D3868221**

**JACOBS ENGINEERING GROUP, INC.**

**412 Mt. Kemble Ave**

**Downtown Building**

**Morristown, NJ - 07960**

**Phone No: 9732670555**

**ORDER ID : Q2250**

**ATTENTION : John Ynfante**



**Laboratory Certification ID # 20012**

Q2250-VOCMS Group3



**1 of 902**

<b>1) VOLATILES DATA</b>	<b>2</b>
<b>2) Signature Page</b>	<b>4</b>
<b>3) Case Narrative</b>	<b>5</b>
<b>4) Qualifier Page</b>	<b>7</b>
<b>5) Conformance/Non Conformance</b>	<b>8</b>
<b>6) QA Checklist</b>	<b>10</b>
<b>7) Chronicle</b>	<b>11</b>
<b>8) Hit Summary</b>	<b>12</b>
<b>9) QC Data Summary For VOCMS Group3</b>	<b>13</b>
<b>9.1) Deuterated Monitoring Compound Summary</b>	<b>14</b>
<b>9.2) MS/MSD Summary</b>	<b>15</b>
<b>9.3) LCS/LCSD Summary</b>	<b>17</b>
<b>9.4) Method Blank Summary</b>	<b>19</b>
<b>9.5) GS/MS Tune Summary</b>	<b>21</b>
<b>9.6) Internal Standard Area and RT Summary</b>	<b>24</b>
<b>10) Sample Data</b>	<b>28</b>
<b>10.1) MW-11A-13.5-060525</b>	<b>29</b>
<b>10.2) MW-11A-13.5-060525DL</b>	<b>42</b>
<b>10.3) EB02-060525</b>	<b>52</b>
<b>10.4) TB-01-060525</b>	<b>60</b>
<b>11) Calibration Data Summary</b>	<b>67</b>
<b>11.1) Initial Calibration Data</b>	<b>68</b>
<b>11.1.1) VN060625</b>	<b>68</b>
<b>11.2) Continued Calibration Data</b>	<b>437</b>
<b>11.2.1) VN086913.D</b>	<b>437</b>
<b>11.2.2) VN086940.D</b>	<b>496</b>
<b>12) QC Sample Data</b>	<b>555</b>
<b>12.1) Tune Raw Data</b>	<b>556</b>
<b>12.2) Method Blank Data</b>	<b>559</b>
<b>12.3) LCS Data</b>	<b>573</b>
<b>12.4) MS Data</b>	<b>679</b>
<b>12.5) MSD Data</b>	<b>732</b>
<b>13) Manual Integration</b>	<b>785</b>
<b>14) Analytical Runlogs</b>	<b>789</b>
<b>15) Standard Prep Logs</b>	<b>801</b>

## Table Of Contents for Q2250

<b>16) Shipping Document</b>	<b>898</b>	<b>1</b>
<b>16.1) Chain Of Custody</b>	<b>899</b>	<b>2</b>
<b>16.2) Lab Certificate</b>	<b>900</b>	<b>3</b>
<b>16.3) Internal COC</b>	<b>901</b>	<b>4</b>
		<b>5</b>
		<b>6</b>
		<b>7</b>
		<b>8</b>
		<b>9</b>
		<b>10</b>
		<b>11</b>
		<b>12</b>
		<b>13</b>
		<b>14</b>
		<b>15</b>
		<b>16</b>

## Cover Page

**Order ID :** Q2250

**Project ID :** Former Schlumberger STC PTC Site D3868221

**Client :** JACOBS Engineering Group, Inc.

### Lab Sample Number

Q2250-01  
Q2250-02  
Q2250-03  
Q2250-04  
Q2250-05  
Q2250-06  
Q2250-07

### Client Sample Number

MW-11A-13.5-060525  
MW-11A-13.5-060525MS  
MW-11A-13.5-060525MSD  
MW-06-6.5-060525  
EB02-060525  
TB-01-060525  
EB02-060525

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 6/18/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## CASE NARRATIVE

**JACOBS Engineering Group, Inc.**

**Project Name:** Former Schlumberger STC PTC Site D3868221

**Project # N/A**

**Order ID # Q2250**

**Test Name:** VOCMS Group3

### **A. Number of Samples and Date of Receipt:**

7 Water samples were received on 06/05/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Alkalinity, Anions Group1, Dissolved ICP-Group2, Dissolved Metals Group3, Mercury, Metals ICP-TAL, METALS-TAL, SVOC-SIMGroup1, TDS and VOCMS Group3. This data package contains results for VOCMS Group3.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_N were done using GC column Rx-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group3 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {Q2250-02MS} with File ID: VN086922.D recoveries met the requirements for all compounds except for cis-1,2-Dichloroethene[2100%] and Trichloroethene[3000%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The MSD {Q2250-03MSD} with File ID: VN086923.D recoveries met the acceptable requirements except for cis-1,2-Dichloroethene[2500%] and Trichloroethene[3800%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The RPD for {Q2250-03MSD} with File ID: VN086923.D met criteria except for Trichloroethene[24%] due to difference in results of MS and MSD.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Initial Calibration met the requirements.  
The Continuous Calibration met the requirements.  
The Tuning criteria met requirements.

Sample MW-11A-13.5-060525 was diluted due to high concentration.

**E. Additional Comments:**

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

**DATA REPORTING QUALIFIERS- ORGANIC**

For reporting results, the following "Results Qualifiers" are used:

- Value** If the result is a value greater than or equal to the detection limit, report the value
- U** Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
- ND** Indicates the analyte was analyzed for, but not detected
- J** Indicates an estimated value. This flag is used:  
(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)  
(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
- B** Indicates the analyte was found in the blank as well as the sample report as "12 B".
- E** Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.
- D** This flag identifies all compounds identified in an analysis at a secondary dilution factor.
- P** This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
- N** This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
- A** This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
- Q** Indicates the LCS did not meet the control limits requirements

# **ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092**

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## **GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

ORDER ID: Q2250

MATRIX: Water

METHOD: 8260D

- |  | NA | NO | YES |
|--|----|----|-----|
| 1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)   |    |    | ✓   |
| 2. GC/MS Tuning Specifications<br>BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY<br>ASP CLP, CLP AND NJ)   |    |    | ✓   |
| 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for<br>8000 Series.   |    |    | ✓   |
| 4. GC/MS Calibration - Initial Calibration performed before sample analysis and<br>continuing calibration performed within 24 hours of sample analysis for 600 series and<br>12 hours for 8000 series.   |    |    | ✓   |
| 5. GC/MS Calibration Requirements.<br><br>The Initial Calibration met the requirements.<br>The Continuous Calibration met the requirements.  |    |    | ✓   |
| 6. Blank Contamination - If yes, list compounds and concentrations in each blank:  |    |    | ✓   |
| 7. Surrogate Recoveries Meet Criteria<br><br>If not met, list those compounds and their recoveries which fall outside the acceptable<br>ranges.  |    |    | ✓   |
| 8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria<br><br>If not met, list those compounds and their recoveries which fall outside the acceptable<br>range.<br><br>The MS {Q2250-02MS} with File ID: VN086922.D recoveries met the requirements<br>for all compounds except for cis-1,2-Dichloroethene[2100%] and<br>Trichloroethene[3000%] these compounds did not meet the NJDKQP criteria and in-<br>house criteria due to matrix interference.<br><br>The MSD {Q2250-03MSD} with File ID: VN086923.D recoveries met the acceptable<br>requirements except for cis-1,2-Dichloroethene[2500%] and Trichloroethene[3800%]<br>these compounds did not meet the NJDKQP criteria and in-house criteria due to<br>matrix interference.<br><br>The RPD for {Q2250-03MSD} with File ID: VN086923.D met criteria except for<br>Trichloroethene[24%] due to difference in results of MS and MSD. |    |    | ✓   |
| 9. Internal Standard Area/Retention Time Shift Meet Criteria   |    |    | ✓   |

Comments:

**ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092**

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)**

NA      NO      YES

10. Analysis Holding Time Met ✓

If not met, list number of days exceeded for each sample:

**ADDITIONAL COMMENTS:**

Sample MW-11A-13.5-060525 was diluted due to high concentration.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

---

QA REVIEW

---

Date

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16

## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q2250

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication, airbill, sample management lab chronicle, login page) ✓

Check chain-of-custody for proper relinquish/return of samples ✓

Is the chain of custody signed and complete ✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts ✓

Collect information for each project id from server. Were all requirements followed ✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page ✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody ✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results ✓

Do requested analyses on Chain of Custody agree with the log-in page ✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody ✓

Were the samples received within hold time ✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle ✓

#### ANALYTICAL:

Was method requirement followed? ✓

Was client requirement followed? ✓

Does the case narrative summarize all QC failure? ✓

All runlogs and manual integration are reviewed for requirements ✓

All manual calculations and /or hand notations verified ✓

QA Review Signature: SOHIL JODHANI

Date: 06/18/2025

## LAB CHRONICLE

<b>OrderID:</b>	Q2250	<b>OrderDate:</b>	6/5/2025 4:22:00 PM					
<b>Client:</b>	JACOBS Engineering Group, Inc.	<b>Project:</b>	Former Schlumberger STC PTC Site D3868221					
<b>Contact:</b>	John Ynfante	<b>Location:</b>	D22,N31,VOA Ref. #3 Water					
<hr/>								
LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2250-01	<b>MW-11A-13.5-060525</b>	<b>Water</b>	VOCMS Group3	8260-Low	<b>06/05/25</b>		<b>06/05/25</b>	
Q2250-01DL	<b>MW-11A-13.5-060525 DL</b>	<b>Water</b>	VOCMS Group3	8260-Low	<b>06/05/25</b>		<b>06/05/25</b>	
Q2250-05	<b>EB02-060525</b>	<b>Water</b>	VOCMS Group3	8260-Low	<b>06/05/25</b>		<b>06/05/25</b>	
Q2250-06	<b>TB-01-060525</b>	<b>Water</b>	VOCMS Group3	8260-Low	<b>06/05/25</b>		<b>06/05/25</b>	

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16

### Hit Summary Sheet SW-846

**SDG No.:** Q2250  
**Client:** JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
<b>Client ID:</b>	<b>MW-11A-13.5-060525</b>							
Q2250-01	MW-11A-13.5-0605 Water	Vinyl Chloride		5.00		0.26	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	1,1-Dichloroethene		4.20		0.23	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	1,1-Dichloroethane		2.80		0.23	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	cis-1,2-Dichloroethene		850	E	0.19	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	Benzene		0.42	J	0.15	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	Trichloroethene		1400	E	0.090	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	1,1,2-Trichloroethane		0.50	J	0.21	1.00	ug/L
Q2250-01	MW-11A-13.5-0605 Water	Tetrachloroethene		9.40		0.23	1.00	ug/L
<b>Total Voc :</b>				2270				
<b>Total Concentration:</b>				2270				
<b>Client ID:</b>	<b>MW-11A-13.5-060525DL</b>							
Q2250-01DL	MW-11A-13.5-0605 Water	Vinyl Chloride		6.60	JD	5.20	20.0	ug/L
Q2250-01DL	MW-11A-13.5-0605 Water	cis-1,2-Dichloroethene		1000	D	3.80	20.0	ug/L
Q2250-01DL	MW-11A-13.5-0605 Water	Trichloroethene		1700	D	1.90	20.0	ug/L
Q2250-01DL	MW-11A-13.5-0605 Water	Tetrachloroethene		12.2	JD	4.60	20.0	ug/L
<b>Total Voc :</b>				2720				
<b>Total Concentration:</b>				2720				



# QC SUMMARY

### Surrogate Summary

**SDG No.:** Q2250

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2250-01	MW-11A-13.5-060525	1,2-Dichloroethane-d4	50	47.8	96	70 (74)	130 (125)
		Dibromofluoromethane	50	49.4	99	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
Q2250-01DL	MW-11A-13.5-060525DL	4-Bromofluorobenzene	50	50.0	100	70 (77)	130 (121)
		1,2-Dichloroethane-d4	50	46.3	93	70 (74)	130 (125)
		Dibromofluoromethane	50	48.4	97	70 (75)	130 (124)
Q2250-02MS	MW-11A-13.5-060525MS	Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98	70 (77)	130 (121)
		1,2-Dichloroethane-d4	50	49.9	100	70 (74)	130 (125)
Q2250-03MSD	MW-11A-13.5-060525MSD	Dibromofluoromethane	50	52.4	105	70 (75)	130 (124)
		Toluene-d8	50	51.2	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.9	102	70 (77)	130 (121)
Q2250-05	EB02-060525	1,2-Dichloroethane-d4	50	52.7	105	70 (74)	130 (125)
		Dibromofluoromethane	50	56.6	113	70 (75)	130 (124)
		Toluene-d8	50	54.8	110	70 (86)	130 (113)
Q2250-06	TB-01-060525	4-Bromofluorobenzene	50	55.3	111	70 (77)	130 (121)
		1,2-Dichloroethane-d4	50	47.9	96	70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98	70 (75)	130 (124)
VN0610WBL01	VN0610WBL01	Toluene-d8	50	51.5	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
		1,2-Dichloroethane-d4	50	47.2	94	70 (74)	130 (125)
VN0610WBS01	VN0610WBS01	Dibromofluoromethane	50	48.8	98	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98	70 (77)	130 (121)
VN0611WBL01	VN0611WBL01	1,2-Dichloroethane-d4	50	46.7	93	70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98	70 (75)	130 (124)
		Toluene-d8	50	51.5	103	70 (86)	130 (113)
VN0611WBS02	VN0611WBS02	4-Bromofluorobenzene	50	49.0	98	70 (77)	130 (121)
		1,2-Dichloroethane-d4	50	45.4	91	70 (74)	130 (125)
		Dibromofluoromethane	50	49.5	99	70 (75)	130 (124)
VN0611WBL01	VN0611WBL01	Toluene-d8	50	48.6	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.3	97	70 (77)	130 (121)
		1,2-Dichloroethane-d4	50	48.3	97	70 (74)	130 (125)
VN0611WBS02	VN0611WBS02	Dibromofluoromethane	50	49.3	99	70 (75)	130 (124)
		Toluene-d8	50	51.7	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.4	99	70 (77)	130 (121)
VN0611WBL01	VN0611WBL01	1,2-Dichloroethane-d4	50	47.1	94	70 (74)	130 (125)
		Dibromofluoromethane	50	51.1	102	70 (75)	130 (124)
		Toluene-d8	50	48.7	97	70 (86)	130 (113)
VN0611WBS02	VN0611WBS02	4-Bromofluorobenzene	50	50.0	100	70 (77)	130 (121)

( ) = LABORATORY INHOUSE LIMIT

### Matrix Spike/Matrix Spike Duplicate Summary

**SW-846**

**SDG No.:** Q2250

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD Qual	Limits		RPD
					Rec	Qual		Low	High	
<b>Lab Sample ID :</b>	<b>Q2250-02MS</b>	<b>Client Sample ID :</b>	<b>MW-11A-13.5-060525MS</b>		<b>Datafile :</b>		<b>VN086922.D</b>			
Vinyl chloride	50	5.00	50.4	ug/L	91			70 (69)	130 (125)	
1,1-Dichloroethene	50	4.20	47.9	ug/L	87			70 (53)	130 (162)	
1,1-Dichloroethane	50	2.80	45.2	ug/L	85			70 (78)	130 (122)	
cis-1,2-Dichloroethene	50	850	1900	ug/L	2100	*		70 (82)	130 (124)	
1,1,1-Trichloroethane	50	0	37.2	ug/L	74			70 (83)	130 (117)	
Benzene	50	0.42	39.1	ug/L	77			70 (81)	130 (128)	
1,2-Dichloroethane	50	0	39.3	ug/L	79			70 (76)	130 (120)	
Trichloroethene	50	1400	2900	ug/L	3000	*		70 (28)	130 (175)	
1,1,2-Trichloroethane	50	0.50	40.0	ug/L	79			70 (27)	130 (175)	
Tetrachloroethene	50	9.40	57.0	ug/L	95			70 (48)	130 (153)	

( ) = LABORATORY INHOUSE LIMIT

### Matrix Spike/Matrix Spike Duplicate Summary

**SW-846**

**SDG No.:** Q2250

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD Qual	Limits			RPD	
					Rec	Qual		Low	High			
<b>Lab Sample ID :</b>	<b>Q2250-03MSD</b>	<b>Client Sample ID :</b>	<b>MW-11A-13.5-060525MSD</b>					<b>Datafile :</b>	<b>VN086923.D</b>			
Vinyl chloride	50	5.00	57.2	ug/L	104		13	70 (69)	130 (125)	20 (20)		
1,1-Dichloroethene	50	4.20	53.1	ug/L	98		12	70 (53)	130 (162)	20 (20)		
1,1-Dichloroethane	50	2.80	49.5	ug/L	93		9	70 (78)	130 (122)	20 (20)		
cis-1,2-Dichloroethene	50	850	2100	ug/L	2500	*	17	70 (82)	130 (124)	20 (20)		
1,1,1-Trichloroethane	50	0	41.4	ug/L	83		11	70 (83)	130 (117)	20 (20)		
Benzene	50	0.42	43.0	ug/L	85		10	70 (81)	130 (128)	20 (20)		
1,2-Dichloroethane	50	0	43.1	ug/L	86		9	70 (76)	130 (120)	20 (20)		
Trichloroethene	50	1400	3300	ug/L	3800	*	24	*	70 (28)	130 (175)	20 (20)	
1,1,2-Trichloroethane	50	0.50	44.5	ug/L	88		11	70 (27)	130 (175)	20 (20)		
Tetrachloroethene	50	9.40	66.0	ug/L	113		17	70 (48)	130 (153)	20 (20)		

( ) = LABORATORY INHOUSE LIMIT

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2250

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW8260-Low

**Datafile :** VN086916.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits		RPD
								Low	High	
VN0610WBS01	Vinyl chloride	20	18.0	ug/L	90			70 (65)	130 (117)	
	1,1-Dichloroethene	20	18.7	ug/L	94			70 (74)	130 (110)	
	1,1-Dichloroethane	20	18.2	ug/L	91			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	18.7	ug/L	94			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	18.1	ug/L	91			70 (80)	130 (108)	
	Benzene	20	18.5	ug/L	93			70 (82)	130 (109)	
	1,2-Dichloroethane	20	18.7	ug/L	94			70 (80)	130 (115)	
	Trichloroethene	20	19.3	ug/L	97			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	19.5	ug/L	98			70 (83)	130 (112)	
	Tetrachloroethene	20	17.9	ug/L	90			70 (67)	130 (123)	

( ) = LABORATORY INHOUSE LIMIT

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2250

**Client:** JACOBS Engineering Group, Inc.

**Analytical Method:** SW8260-Low

**Datafile :** VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits		RPD
								Low	High	
VN0611WBS02	Vinyl chloride	20	18.9	ug/L	95			70 (65)	130 (117)	
	1,1-Dichloroethene	20	19.2	ug/L	96			70 (74)	130 (110)	
	1,1-Dichloroethane	20	19.4	ug/L	97			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			70 (80)	130 (108)	
	Benzene	20	19.3	ug/L	97			70 (82)	130 (109)	
	1,2-Dichloroethane	20	19.5	ug/L	98			70 (80)	130 (115)	
	Trichloroethene	20	19.9	ug/L	100			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	20.4	ug/L	102			70 (83)	130 (112)	
	Tetrachloroethene	20	19.2	ug/L	96			70 (67)	130 (123)	

( ) = LABORATORY INHOUSE LIMIT



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0610WBL01

Lab Name: CHEMTECH

Contract: JACO05

Lab Code: CHEM Case No.: Q2250

SAS No.: Q2250 SDG NO.: Q2250

Lab File ID: VN086915.D

Lab Sample ID: VN0610WBL01

Date Analyzed: 06/10/2025

Time Analyzed: 10:09

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0610WBS01	VN0610WBS01	VN086916.D	06/10/2025
EB02-060525	Q2250-05	VN086919.D	06/10/2025
TB-01-060525	Q2250-06	VN086920.D	06/10/2025
MW-11A-13.5-060525	Q2250-01	VN086921.D	06/10/2025
MW-11A-13.5-060525MS	Q2250-02MS	VN086922.D	06/10/2025
MW-11A-13.5-060525MSD	Q2250-03MSD	VN086923.D	06/10/2025

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

**VN0611WBL01**

Lab Name: CHEMTECH

Contract: JACO05

Lab Code: CHEM Case No.: Q2250

SAS No.: Q2250 SDG NO.: Q2250

Lab File ID: VN086942.D

Lab Sample ID: VN0611WBL01

Date Analyzed: 06/11/2025

Time Analyzed: 12:28

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
<b>VN0611WBS02</b>	<b>VN0611WBS02</b>	<b>VN086944.D</b>	<b>06/11/2025</b>
<b>MW-11A-13.5-060525DL</b>	<b>Q2250-01DL</b>	<b>VN086948.D</b>	<b>06/11/2025</b>

COMMENTS:

---

---



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name:	CHEMTECH	Contract:	JAC005
Lab Code:	CHEM	Case No.:	Q2250
Lab File ID:	VN086861.D	SAS No.:	Q2250
Instrument ID:	MSVOA_N	SDG NO.:	Q2250
GC Column:	RXI-624 ID: 0.25 (mm)	BFB Injection Date:	06/06/2025
		BFB Injection Time:	07:59
		Heated Purge:	Y/N
			N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.3
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.7 ( 1 ) 1
174	50.0 - 100.0% of mass 95	66.6
175	5.0 - 9.0% of mass 174	4.7 ( 7.1 ) 1
176	95.0 - 101.0% of mass 174	65.3 ( 98.1 ) 1
177	5.0 - 9.0% of mass 176	4.4 ( 6.8 ) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC001	VSTDICC001	VN086862.D	06/06/2025	12:44
VSTDICC005	VSTDICC005	VN086863.D	06/06/2025	13:17
VSTDICC020	VSTDICC020	VN086864.D	06/06/2025	13:40
VSTDICCC050	VSTDICCC050	VN086865.D	06/06/2025	14:03
VSTDICC100	VSTDICC100	VN086866.D	06/06/2025	14:26
VSTDICC150	VSTDICC150	VN086867.D	06/06/2025	14:49



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name:	CHEMTECH	Contract:	JAC005
Lab Code:	CHEM	Case No.:	Q2250
Lab File ID:	VN086912.D	SAS No.:	Q2250
Instrument ID:	MSVOA_N	SDG NO.:	Q2250
GC Column:	RXI-624	Heated Purge: Y/N	N
		BFB Injection Date:	06/10/2025
		BFB Injection Time:	08:39

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.1
75	30.0 - 60.0% of mass 95	47.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.6 ( 0.8 ) 1
174	50.0 - 100.0% of mass 95	73.1
175	5.0 - 9.0% of mass 174	5.3 ( 7.2 ) 1
176	95.0 - 101.0% of mass 174	71.3 ( 97.5 ) 1
177	5.0 - 9.0% of mass 176	4.9 ( 6.9 ) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086913.D	06/10/2025	09:13
VN0610WBL01	VN0610WBL01	VN086915.D	06/10/2025	10:09
VN0610WBS01	VN0610WBS01	VN086916.D	06/10/2025	10:30
EB02-060525	Q2250-05	VN086919.D	06/10/2025	11:47
TB-01-060525	Q2250-06	VN086920.D	06/10/2025	12:08
MW-11A-13.5-060525	Q2250-01	VN086921.D	06/10/2025	12:30
MW-11A-13.5-060525MS	Q2250-02MS	VN086922.D	06/10/2025	12:51
MW-11A-13.5-060525MSD	Q2250-03MSD	VN086923.D	06/10/2025	13:12



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name:	CHEMTECH	Contract:	JAC005
Lab Code:	CHEM	Case No.:	Q2250
Lab File ID:	VN086939.D	SAS No.:	Q2250
Instrument ID:	MSVOA_N	BFB Injection Date:	06/11/2025
GC Column:	RXI-624 ID: 0.25 (mm)	BFB Injection Time:	10:22
		Heated Purge:	Y/N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.7
75	30.0 - 60.0% of mass 95	46.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.7 ( 1 ) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.4 ( 7.5 ) 1
176	95.0 - 101.0% of mass 174	68.8 ( 96.8 ) 1
177	5.0 - 9.0% of mass 176	4.8 ( 7 ) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086940.D	06/11/2025	11:32
VN0611WBL01	VN0611WBL01	VN086942.D	06/11/2025	12:28
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025	13:27
MW-11A-13.5-060525DL	Q2250-01DL	VN086948.D	06/11/2025	15:06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JAC005  
 Lab Code: CHEM Case No.: Q2250 SAS No.: Q2250 SDG No.: Q2250  
 Lab File ID: VN086913.D Date Analyzed: 06/10/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 09:13  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	239452	8.23	422119	9.11	362285	11.87
	478904	8.73	844238	9.606	724570	12.365
	119726	7.73	211060	8.606	181143	11.365
EPA SAMPLE NO.						
MW-11A-13.5-060525	361832	8.23	672514	9.11	593044	11.87
MW-11A-13.5-060525MS	174378	8.23	310932	9.11	270945	11.87
MW-11A-13.5-060525MSD	172955	8.23	309901	9.11	266637	11.87
EB02-060525	232754	8.23	433385	9.11	381245	11.87
TB-01-060525	350151	8.23	648560	9.11	563762	11.87
VN0610WBL01	241251	8.23	440901	9.11	386425	11.87
VN0610WBS01	249843	8.23	443174	9.11	389297	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	CHEMTECH	Contract:	JAC005		
Lab Code:	<u>CHEM</u>	SAS No.:	<u>Q2250</u>	SDG NO.:	<u>Q2250</u>
Lab File ID:	<u>VN086913.D</u>	Date Analyzed:	<u>06/10/2025</u>		
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>09:13</u>		
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)	Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	170987	13.788				
	341974	14.288				
	85493.5	13.288				
EPA SAMPLE NO.						
MW-11A-13.5-060525	281220	13.79				
MW-11A-13.5-060525MS	132628	13.79				
MW-11A-13.5-060525MSD	129873	13.79				
EB02-060525	180063	13.79				
TB-01-060525	266280	13.79				
VN0610WBL01	183380	13.79				
VN0610WBS01	189620	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	CHEMTECH	Contract:	JAC005				
Lab Code:	CHEM	Case No.:	Q2250	SAS No.:	Q2250	SDG NO.:	Q2250
Lab File ID:	VN086940.D	Date Analyzed:	06/11/2025				
Instrument ID:	MSVOA_N	Time Analyzed:	11:32				
GC Column:	RXI-624	ID:	0.25 (mm)	Heated Purge:	(Y/N) N		

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	205098	8.23	357643	9.11	312173	11.87
UPPER LIMIT	410196	8.73	715286	9.606	624346	12.365
LOWER LIMIT	102549	7.73	178822	8.606	156087	11.365
EPA SAMPLE NO.						
MW-11A-13.5-060525DL	272350	8.24	500042	9.11	434591	11.87
VN0611WBL01	328120	8.23	618651	9.11	543433	11.87
VN0611WBS02	206799	8.24	370524	9.11	321188	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	CHEMTECH	Contract:	JAC005		
Lab Code:	<u>CHEM</u>	SAS No.:	<u>Q2250</u>	SDG NO.:	<u>Q2250</u>
Lab File ID:	<u>VN086940.D</u>	Date Analyzed:	<u>06/11/2025</u>		
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>11:32</u>		
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)	Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	152136	13.788				
UPPER LIMIT	304272	14.288				
LOWER LIMIT	76068	13.288				
EPA SAMPLE NO.						
MW-11A-13.5-060525DL	208257	13.79				
VN0611WBL01	257257	13.79				
VN0611WBS02	157403	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.



# SAMPLE

# DATA

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:	06/05/25
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:	06/05/25
Client Sample ID:	MW-11A-13.5-060525			SDG No.:	Q2250
Lab Sample ID:	Q2250-01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL			Test:	VOCMS Group3
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086921.D	1		06/10/25 12:30	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	5.00		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	4.20		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	2.80		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	850	E	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.42	J	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	1400	E	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	J	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	9.40		0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	47.8		70 (74) - 130 (125)	96%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		70 (75) - 130 (124)	99%	SPK: 50
2037-26-5	Toluene-d8	51.3		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		70 (77) - 130 (121)	100%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	362000	8.23			
540-36-3	1,4-Difluorobenzene	673000	9.106			
3114-55-4	Chlorobenzene-d5	593000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	281000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086921.D  
 Acq On : 10 Jun 2025 12:30  
 Operator : JC\MD  
 Sample : Q2250-01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**MW-11A-13.5-060525**

Quant Time: Jun 11 01:52:07 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	361832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	672514	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	593044	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	281220	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.582	65	231529	47.792	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	95.580%	
35) Dibromofluoromethane	8.177	113	196952	49.417	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	98.840%	
50) Toluene-d8	10.565	98	809307	51.295	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	102.600%	
62) 4-Bromofluorobenzene	12.847	95	293403	50.054	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	100.100%	
<b>Target Compounds</b>						
				Qvalue		
4) Vinyl Chloride	2.559	62	24226	5.041	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.400	101	302235	76.617	ug/l	99
12) 1,1-Dichloroethene	4.371	96	16986	4.218	ug/l	100
21) trans-1,2-Dichloroethene	5.812	96	16472	3.676	ug/l	88
24) 1,1-Dichloroethane	6.594	63	22723	2.805	ug/l	96
27) cis-1,2-Dichloroethene	7.500	96	4534458	846.190	ug/l	90
29) Tetrahydrofuran	7.853	42	9084	3.339	ug/l	95
40) Benzene	8.612	78	8119	0.418	ug/l	95
44) Trichloroethene	9.359	130	6309047	1370.098	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	2269	0.497	ug/l	95
64) Tetrachloroethene	11.106	164	35442	9.443	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086921.D  
 Acq On : 10 Jun 2025 12:30  
 Operator : JC\MD  
 Sample : Q2250-01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-11A-13.5-060525


Quant Time: Jun 11 01:52:07 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30


Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525





Tgt Ion:168 Resp: 361832  
Ion Ratio Lower Upper  
168 100  
99 58.0 49.1 73.7




#4  
Vinyl Chloride  
Concen: 5.041 ug/l  
RT: 2.559 min Scan# 103  
Delta R.T. 0.006 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30




Tgt Ion: 62 Resp: 24226  
Ion Ratio Lower Upper  
62 100  
64 30.0 26.0 39.0









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 76.617 ug/l  
RT: 4.400 min Scan# 4  
Instrument: MSVOA\_N  
Delta R.T. 0.006 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30 ClientSampleId : MW-11A-13.5-060525





Tgt Ion:101 Resp: 302235  
Ion Ratio Lower Upper  
101 100  
85 41.6 33.8 50.8  
151 74.1 58.7 88.1






#12  
1,1-Dichloroethene  
Concen: 4.218 ug/l  
RT: 4.371 min Scan# 411  
Delta R.T. 0.012 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30



Tgt Ion: 96 Resp: 16986  
Ion Ratio Lower Upper  
96 100  
61 157.0 125.8 188.6  
98 65.3 52.5 78.7








#21  
trans-1,2-Dichloroethene  
Concen: 3.676 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. 0.006 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30



Instrument: MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525




Tgt Ion: 96 Resp: 16472  
Ion Ratio Lower Upper  
96 100  
61 118.7 109.2 163.8  
98 66.1 49.8 74.8



#24  
1,1-Dichloroethane  
Concen: 2.805 ug/l  
RT: 6.594 min Scan# 789  
Delta R.T. 0.012 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30





Tgt Ion: 63 Resp: 22723  
Ion Ratio Lower Upper  
63 100  
98 5.8 3.6 10.8  
100 6.6 2.6 7.8






#27  
cis-1,2-Dichloroethene  
Concen: 846.190 ug/l  
RT: 7.500 min Scan# 9  
Instrument: MSVOA\_N  
Delta R.T. -0.000 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30  
ClientSampleId : MW-11A-13.5-060525




Tgt Ion: 96 Resp: 4534458  
Ion Ratio Lower Upper  
96 100  
61 121.3 0.0 278.0  
98 64.3 0.0 128.2



#29  
Tetrahydrofuran  
Concen: 3.339 ug/l  
RT: 7.853 min Scan# 1003  
Delta R.T. -0.000 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30

Tgt Ion: 42 Resp: 9084  
Ion Ratio Lower Upper  
42 100  
72 45.2 37.8 56.6  
71 40.1 35.6 53.4





#33

1,2-Dichloroethane-d4

Concen: 47.792 ug/l

RT: 8.582 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086921.D

Acq: 10 Jun 2025 12:30

Instrument:

MSVOA\_N

ClientSampleId :




MW-11A-13.5-060525

Tgt Ion: 65 Resp: 231529

Ion Ratio Lower Upper

65 100

67 54.8 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086921.D

Acq: 10 Jun 2025 12:30



Tgt Ion:114 Resp: 672514







Ion Ratio Lower Upper

114 100

63 19.4 0.0 39.6

88 15.7 0.0 30.2





#35

Dibromofluoromethane

Concen: 49.417 ug/l

RT: 8.177 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086921.D

Acq: 10 Jun 2025 12:30

Instrument:

MSVOA\_N

ClientSampleId :

MW-11A-13.5-060525


Tgt Ion:113 Resp: 196952

Ion Ratio Lower Upper

113 100

111 101.7 84.2 126.2

192 17.8 14.2 21.4



#40

Benzene

Concen: 0.418 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086921.D


Acq: 10 Jun 2025 12:30







Tgt Ion: 78 Resp: 8119

Ion Ratio Lower Upper

78 100

77 21.1 18.7 28.1





#44

Trichloroethene

Concen: 1370.098 ug/l

RT: 9.359 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086921.D ClientSampleId :

Acq: 10 Jun 2025 12:30 MW-11A-13.5-060525

Instrument :


MSVOA\_N

Tgt Ion:130 Resp: 6309047

Ion Ratio Lower Upper

130 100

95 100.4 0.0 196.2



#50

Toluene-d8

Concen: 51.295 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.006 min

Lab File: VN086921.D


Acq: 10 Jun 2025 12:30




Tgt Ion: 98 Resp: 809307

Ion Ratio Lower Upper

98 100

100 66.5 53.4 80.0





#55

1,1,2-Trichloroethane  
Concen: 0.497 ug/l  
RT: 11.018 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525

Tgt Ion: 97 Resp: 2269

Ion	Ratio	Lower	Upper
97	100		
83	88.0	69.7	104.5
85	47.5	45.1	67.7
99	60.4	50.6	75.8

Abundance




#62






4-Bromofluorobenzene  
Concen: 50.054 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30

Tgt Ion: 95 Resp: 293403

Ion	Ratio	Lower	Upper
95	100		
174	70.2	0.0	141.8
176	68.2	0.0	132.6

Abundance






#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30


Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525


Tgt Ion:117 Resp: 593044  
Ion Ratio Lower Upper  
117 100  
82 55.4 44.6 67.0  
119 32.3 25.5 38.3



#64  
Tetrachloroethene  
Concen: 9.443 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. -0.000 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30

Tgt Ion:164 Resp: 35442  
Ion Ratio Lower Upper  
164 100  
166 131.5 101.8 152.8  
129 98.4 83.0 124.4  
131 95.1 78.2 117.2






#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2  
Delta R.T. -0.006 min  
Lab File: VN086921.D  
Acq: 10 Jun 2025 12:30

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525

Tgt Ion:152 Resp: 281220  
Ion Ratio Lower Upper  
152 100  
115 60.9 30.1 90.5  
150 158.4 0.0 345.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:	06/05/25	
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:	06/05/25	
Client Sample ID:	MW-11A-13.5-060525DL			SDG No.:	Q2250	
Lab Sample ID:	Q2250-01DL			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:	uL			Test:	VOCMS Group3	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086948.D	20		06/11/25 15:06	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	6.60	JD	5.20	20.0	ug/L
75-35-4	1,1-Dichloroethene	4.60	UD	4.60	20.0	ug/L
75-34-3	1,1-Dichloroethane	4.60	UD	4.60	20.0	ug/L
156-59-2	cis-1,2-Dichloroethene	1000	D	3.80	20.0	ug/L
71-55-6	1,1,1-Trichloroethane	4.00	UD	4.00	20.0	ug/L
71-43-2	Benzene	3.00	UD	3.00	20.0	ug/L
107-06-2	1,2-Dichloroethane	4.40	UD	4.40	20.0	ug/L
79-01-6	Trichloroethene	1700	D	1.90	20.0	ug/L
79-00-5	1,1,2-Trichloroethane	4.20	UD	4.20	20.0	ug/L
127-18-4	Tetrachloroethene	12.2	JD	4.60	20.0	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	46.3		70 (74) - 130 (125)	93%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		70 (75) - 130 (124)	97%	SPK: 50
2037-26-5	Toluene-d8	51.1		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.1		70 (77) - 130 (121)	98%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	272000	8.236			
540-36-3	1,4-Difluorobenzene	500000	9.106			
3114-55-4	Chlorobenzene-d5	435000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	208000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086948.D  
 Acq On : 11 Jun 2025 15:06  
 Operator : JC\MD  
 Sample : Q2250-01DL 20X  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**MW-11A-13.5-060525DL**

Quant Time: Jun 12 01:33:54 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.236	168	272350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	500042	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	434591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	208257	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	168991	46.344	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	92.680%	
35) Dibromofluoromethane	8.177	113	143354	48.375	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	96.760%	
50) Toluene-d8	10.571	98	599641	51.115	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	102.240%	
62) 4-Bromofluorobenzene	12.847	95	214176	49.140	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	98.280%	
<b>Target Compounds</b>						
				Qvalue		
4) Vinyl Chloride	2.559	62	1187	0.328	ug/l	90
9) 1,1,2-Trichlorotrifluo...	4.395	101	13861	4.668	ug/l	97
27) cis-1,2-Dichloroethene	7.500	96	208706	51.744	ug/l	89
44) Trichloroethene	9.359	130	290594	84.873	ug/l	98
64) Tetrachloroethene	11.106	164	1673	0.608	ug/l #	78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
Data File : VN086948.D  
Acq On : 11 Jun 2025 15:06  
Operator : JC\MD  
Sample : Q2250-01DL 20X  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-11A-13.5-060525DL


Quant Time: Jun 12 01:33:54 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.236 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06


Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525DL





Tgt Ion:168 Resp: 272350  
Ion Ratio Lower Upper  
168 100  
99 59.2 49.1 73.7






#4  
Vinyl Chloride  
Concen: 0.328 ug/l  
RT: 2.559 min Scan# 103  
Delta R.T. 0.006 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06



Tgt Ion: 62 Resp: 1187  
Ion Ratio Lower Upper  
62 100  
64 38.0 26.0 39.0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 4.668 ug/l

RT: 4.395 min Scan# 4

Delta R.T. -0.000 min

Lab File: VN086948.D

Acq: 11 Jun 2025 15:06

Instrument:

MSVOA\_N

ClientSampleId :

MW-11A-13.5-060525DL





Tgt Ion:101 Resp: 13861

Ion Ratio Lower Upper

101 100

85 39.8 33.8 50.8

151 75.5 58.7 88.1



#27

cis-1,2-Dichloroethene

Concen: 51.744 ug/l

RT: 7.500 min Scan# 943

Delta R.T. 0.000 min

Lab File: VN086948.D

Acq: 11 Jun 2025 15:06


Tgt Ion: 96 Resp: 208706


Ion Ratio Lower Upper

96 100


61 120.1 0.0 278.0

98 63.6 0.0 128.2







Abundance Scan 1128 (8.588 min): VN086948.D\data.ms




Abundance Scan 1128 (8.588 min): VN086948.D\data.ms (-1)



Abundance Scan 1216 (9.106 min): VN086865.D\data.ms (-1)



Abundance Scan 1216 (9.106 min): VN086948.D\data.ms (-1)



#33

1,2-Dichloroethane-d4

Concen: 46.344 ug/l

RT: 8.588 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086948.D

Acq: 11 Jun 2025 15:06

Instrument:

MSVOA\_N

ClientSampleId :


MW-11A-13.5-060525DL

Tgt Ion: 65 Resp: 168991

Ion Ratio Lower Upper

65 100

67 54.6 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN086948.D

Acq: 11 Jun 2025 15:06


Tgt Ion: 114 Resp: 500042


Ion Ratio Lower Upper

114 100

63 19.2 0.0 39.6

88 15.4 0.0 30.2









#35

Dibromofluoromethane  
Concen: 48.375 ug/l  
RT: 8.177 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06


Instrument: MSVOA\_N  
ClientSampleId: MW-11A-13.5-060525DL





Tgt Ion:113 Resp: 143354  
Ion Ratio Lower Upper  
113 100  
111 103.8 84.2 126.2  
192 18.3 14.2 21.4






#44  
Trichloroethene  
Concen: 84.873 ug/l  
RT: 9.359 min Scan# 1259  
Delta R.T. 0.000 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06







Tgt Ion:130 Resp: 290594  
Ion Ratio Lower Upper  
130 100  
95 96.5 0.0 196.2






#50  
Toluene-d8  
Concen: 51.115 ug/l  
RT: 10.571 min Scan# 1  
Instrument: MSVOA\_N  
Delta R.T. 0.000 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06 ClientSampleId : MW-11A-13.5-060525DL


Tgt Ion: 98 Resp: 599641  
Ion Ratio Lower Upper  
98 100  
100 67.1 53.4 80.0



#62  
4-Bromofluorobenzene  
Concen: 49.140 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. 0.000 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06



Tgt Ion: 95 Resp: 214176  
Ion Ratio Lower Upper  
95 100  
174 70.6 0.0 141.8  
176 67.6 0.0 132.6





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Instrument: MSVOA\_N  
Delta R.T. 0.000 min  
Lab File: VN086948.D  
ClientSampleId : MW-11A-13.5-060525DL  
Acq: 11 Jun 2025 15:06


Tgt Ion:117 Resp: 434591  
Ion Ratio Lower Upper  
117 100  
82 54.9 44.6 67.0  
119 32.0 25.5 38.3



#64  
Tetrachloroethene  
Concen: 0.608 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. 0.000 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06


Tgt Ion:164 Resp: 1673  
Ion Ratio Lower Upper  
164 100  
166 108.4 101.8 152.8  
129 85.6 83.0 124.4  
131 64.4 78.2 117.2#

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2  
Instrument: MSVOA\_N  
Delta R.T. -0.006 min  
Lab File: VN086948.D  
Acq: 11 Jun 2025 15:06 ClientSampleId : MW-11A-13.5-060525DL

Tgt Ion:152 Resp: 208257  
Ion Ratio Lower Upper  
152 100  
115 60.5 30.1 90.5  
150 157.5 0.0 345.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:	06/05/25	
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:	06/05/25	
Client Sample ID:	EB02-060525			SDG No.:	Q2250	
Lab Sample ID:	Q2250-05			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:	uL			Test:	VOCMS Group3	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086919.D	1		06/10/25 11:47	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	47.9		70 (74) - 130 (125)	96%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		70 (75) - 130 (124)	98%	SPK: 50
2037-26-5	Toluene-d8	51.5		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.9		70 (77) - 130 (121)	98%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	233000	8.23			
540-36-3	1,4-Difluorobenzene	433000	9.106			
3114-55-4	Chlorobenzene-d5	381000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	180000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

( ) = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086919.D  
 Acq On : 10 Jun 2025 11:47  
 Operator : JC\MD  
 Sample : Q2250-05  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 EB02-060525

Quant Time: Jun 11 01:51:21 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	232754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	433385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	381245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	180063	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.583	65	149250	47.893	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.780%		
35) Dibromofluoromethane	8.177	113	125229	48.759	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.520%		
50) Toluene-d8	10.571	98	523905	51.528	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.060%		
62) 4-Bromofluorobenzene	12.847	95	184714	48.899	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.800%		
<b>Target Compounds</b>						
37) Ethyl Acetate	7.571	43	12844	2.636	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086919.D  
Acq On : 10 Jun 2025 11:47  
Operator : JC\MD  
Sample : Q2250-05  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
EB02-060525


Quant Time: Jun 11 01:51:21 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086919.D  
Acq: 10 Jun 2025 11:47


Instrument : MSVOA\_N  
ClientSampleId : EB02-060525





Tgt Ion:168 Resp: 232754  
Ion Ratio Lower Upper  
168 100  
99 58.4 49.1 73.7






#33  
1,2-Dichloroethane-d4  
Concen: 47.893 ug/l  
RT: 8.583 min Scan# 1127  
Delta R.T. -0.006 min  
Lab File: VN086919.D  
Acq: 10 Jun 2025 11:47



Tgt Ion: 65 Resp: 149250  
Ion Ratio Lower Upper  
65 100  
67 55.1 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086919.D

Acq: 10 Jun 2025 11:47

Instrument:

MSVOA\_N

ClientSampleId :

EB02-060525

Tgt Ion:114 Resp: 433385

Ion Ratio Lower Upper

114 100

63 20.3

88 15.6

0.0 39.6

0.0 30.2

Abundance

200000





150000

100000

50000

0

Time--&gt;



#35

Dibromofluoromethane

Concen: 48.759 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086919.D

Acq: 10 Jun 2025 11:47

Tgt Ion:113 Resp: 125229

Ion Ratio Lower Upper

113 100

111 103.7

192 18.4

84.2 126.2

14.2 21.4

Abundance

50000

40000


30000




20000

10000

0

Time--&gt;





#37

Ethyl Acetate

Concen: 2.636 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086919.D

Acq: 10 Jun 2025 11:47

Instrument:

MSVOA\_N

ClientSampleId :

EB02-060525

Tgt Ion: 43 Resp: 12844

Ion Ratio Lower Upper





43 100

61 15.8 12.1 18.1

70 13.4 9.8 14.6

Abundance

7.571



#50

Toluene-d8

Concen: 51.528 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086919.D

Acq: 10 Jun 2025 11:47

Tgt Ion: 98 Resp: 523905


Ion Ratio Lower Upper


98 100

100 66.5 53.4 80.0

Abundance



10.571






#62  
4-Bromofluorobenzene  
Concen: 48.899 ug/l  
RT: 12.847 min Scan# 1  
Instrument: MSVOA\_N  
Delta R.T. -0.000 min  
Lab File: VN086919.D  
Acq: 10 Jun 2025 11:47  
ClientSampleId : EB02-060525


Tgt Ion: 95 Resp: 184714  
Ion Ratio Lower Upper  
95 100  
174 73.1 0.0 141.8  
176 70.0 0.0 132.6



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1685  
Delta R.T. -0.000 min  
Lab File: VN086919.D  
Acq: 10 Jun 2025 11:47

Tgt Ion:117 Resp: 381245  
Ion Ratio Lower Upper  
117 100  
82 55.0 44.6 67.0  
119 32.0 25.5 38.3






#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2  
Delta R.T. -0.006 min  
Lab File: VN086919.D  
Acq: 10 Jun 2025 11:47

Instrument : MSVOA\_N  
ClientSampleId : EB02-060525

Tgt Ion:152 Resp: 180063  
Ion Ratio Lower Upper  
152 100  
115 60.2 30.1 90.5  
150 157.3 0.0 345.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:	06/05/25	
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:	06/05/25	
Client Sample ID:	TB-01-060525			SDG No.:	Q2250	
Lab Sample ID:	Q2250-06			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:	uL			Test:	VOCMS Group3	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086920.D	1		06/10/25 12:08	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	47.2		70 (74) - 130 (125)	94%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		70 (75) - 130 (124)	98%	SPK: 50
2037-26-5	Toluene-d8	51.1		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.1		70 (77) - 130 (121)	98%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	350000	8.23			
540-36-3	1,4-Difluorobenzene	649000	9.106			
3114-55-4	Chlorobenzene-d5	564000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	266000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

( ) = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086920.D  
 Acq On : 10 Jun 2025 12:08  
 Operator : JC\MD  
 Sample : Q2250-06  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**TB-01-060525**

Quant Time: Jun 11 01:51:43 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	350151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	648560	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	563762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	266280	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.583	65	221297	47.204	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.400%		
35) Dibromofluoromethane	8.177	113	187512	48.787	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.580%		
50) Toluene-d8	10.565	98	777698	51.112	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.220%		
62) 4-Bromofluorobenzene	12.847	95	277459	49.082	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.160%		


Target Compounds	Qvalue
(#)= qualifier out of range (m)= manual integration (+)= signals summed	

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086920.D  
Acq On : 10 Jun 2025 12:08  
Operator : JC\MD  
Sample : Q2250-06  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
TB-01-060525


Quant Time: Jun 11 01:51:43 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086920.D  
Acq: 10 Jun 2025 12:08


Instrument : MSVOA\_N  
ClientSampleId : TB-01-060525





Tgt Ion:168 Resp: 350151  
Ion Ratio Lower Upper  
168 100  
99 57.3 49.1 73.7









#33  
1,2-Dichloroethane-d4  
Concen: 47.204 ug/l  
RT: 8.583 min Scan# 1127  
Delta R.T. -0.006 min  
Lab File: VN086920.D  
Acq: 10 Jun 2025 12:08



Tgt Ion: 65 Resp: 221297  
Ion Ratio Lower Upper  
65 100  
67 54.6 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086920.D

Acq: 10 Jun 2025 12:08

Instrument:

MSVOA\_N

ClientSampleId :

TB-01-060525

Tgt Ion:114 Resp: 648560

Ion Ratio Lower Upper


114 100

63 20.4

88 16.2

0.0 39.6

0.0 30.2



#35

Dibromofluoromethane

Concen: 48.787 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086920.D

Acq: 10 Jun 2025 12:08

Tgt Ion:113 Resp: 187512

Ion Ratio Lower Upper


113 100




111 102.2

192 18.2

84.2 126.2

14.2 21.4









#50  
Toluene-d8  
Concen: 51.112 ug/l  
RT: 10.565 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086920.D  
Acq: 10 Jun 2025 12:08


Instrument : MSVOA\_N  
ClientSampleId : TB-01-060525


Tgt Ion: 98 Resp: 777698  
Ion Ratio Lower Upper  
98 100  
100 66.9 53.4 80.0



#62  
4-Bromofluorobenzene  
Concen: 49.082 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086920.D  
Acq: 10 Jun 2025 12:08

Tgt Ion: 95 Resp: 277459  
Ion Ratio Lower Upper  
95 100  
174 70.9 0.0 141.8  
176 67.9 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086920.D  
Acq: 10 Jun 2025 12:08


Instrument : MSVOA\_N  
ClientSampleId : TB-01-060525

Tgt Ion:117 Resp: 563762  
Ion Ratio Lower Upper  
117 100  
82 54.9 44.6 67.0  
119 32.1 25.5 38.3



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2012  
Delta R.T. -0.006 min  
Lab File: VN086920.D  
Acq: 10 Jun 2025 12:08

Tgt Ion:152 Resp: 266280  
Ion Ratio Lower Upper  
152 100  
115 61.7 30.1 90.5  
150 160.6 0.0 345.0





# CALIBRATION

# SUMMARY

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	CHEMTECH	Contract:	JAC005	
Lab Code:	CHEM	Case No.:	Q2250	
Instrument ID:	MSVOA_N	Calibration Date(s):	06/06/2025	
Heated Purge:	(Y/N) N	Calibration Time(s):	12:44	14:49
GC Column:	RXI-624	ID:	0.25 (mm)	

LAB FILE ID:	RRF001 = VN086862.D	RRF005 = VN086863.D	RRF020 = VN086864.D					
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664	2.5
1,1-Dichloroethene	0.573	0.593	0.563	0.533	0.550	0.527	0.557	4.4
1,1-Dichloroethane	1.192	1.153	1.156	1.063	1.110	1.043	1.120	5.2
cis-1,2-Dichloroethene	0.786	0.766	0.762	0.699	0.729	0.701	0.740	4.9
1,1,1-Trichloroethane	1.029	0.995	0.969	0.895	0.925	0.893	0.951	5.9
Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444	6.2
1,2-Dichloroethane	0.473	0.456	0.444	0.411	0.430	0.413	0.438	5.6
Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.328	0.342	4.2
1,1,2-Trichloroethane	0.359	0.355	0.342	0.322	0.335	0.323	0.340	4.7
Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.5
1,2-Dichloroethane-d4		0.732	0.707	0.500	0.656	0.751	0.669	15.1
Dibromofluoromethane		0.303	0.310	0.219	0.298	0.351	0.296	16.2
Toluene-d8		1.245	1.203	0.861	1.178	1.377	1.173	16.2
4-Bromofluorobenzene		0.441	0.446	0.325	0.446	0.521	0.436	16.2

\* Compounds with required minimum RRF and maximum %RSD values.

All other compounds must meet a minimum RRF of 0.010.

RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\

Method File : 82N060625W.M

Title : SW846 8260

Last Update : Sat Jun 07 02:12:50 2025

Response Via : Initial Calibration

## Calibration Files

1 =VN086862.D 5 =VN086863.D 20 =VN086864.D 50 =VN086865.D 100 =VN086866.D 150 =VN086867.D

Compound	1	5	20	50	100	150	Avg	%RSD
----------	---	---	----	----	-----	-----	-----	------

1) I	Pentafluorobenzene	-----	ISTD-----					
2) T	Dichlorodifluo...	0.467	0.444	0.539	0.501	0.535	0.506	0.499
3) P	Chloromethane	0.762	0.654	0.645	0.597	0.617	0.587	0.644
4) C	Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664
5) T	Bromomethane	0.375	0.380	0.357	0.379	0.368	0.372	0.372
6) T	Chloroethane	0.460	0.444	0.442	0.408	0.418	0.402	0.429
7) T	Trichlorofluor...	0.882	0.903	0.904	0.834	0.858	0.825	0.868
8) T	Diethyl Ether	0.372	0.397	0.383	0.363	0.384	0.369	0.378
9) T	1,1,2-Trichlor...	0.554	0.567	0.563	0.519	0.546	0.520	0.545
10) T	Methyl Iodide	0.720	0.729	0.683	0.722	0.679	0.707	0.707
11) T	Tert butyl alc...	0.192	0.194	0.176	0.180	0.166	0.182	0.182
12) CM	1,1-Dichloroet...	0.573	0.593	0.563	0.533	0.550	0.527	0.557
13) T	Acrolein	0.073	0.051	0.045	0.052	0.066	0.057	0.057
14) T	Allyl chloride	0.985	0.911	0.925	0.865	0.905	0.947	0.923
15) T	Acrylonitrile	0.434	0.434	0.445	0.407	0.423	0.404	0.425
16) T	Acetone	0.426	0.366	0.366	0.322	0.334	0.316	0.355
17) T	Carbon Disulfide	1.718	1.621	1.542	1.426	1.496	1.433	1.539
18) T	Methyl Acetate	1.035	1.049	1.078	0.986	1.049	1.011	1.035
19) T	Methyl tert-bu...	2.120	2.038	2.051	1.933	2.021	1.926	2.015
20) T	Methylene Chlo...	0.822	0.688	0.643	0.605	0.629	0.601	0.665
21) T	trans-1,2-Dich...	0.700	0.674	0.621	0.567	0.591	0.561	0.619
22) T	Diisopropyl ether	2.018	2.020	2.036	1.856	1.915	1.828	1.945
23) T	Vinyl Acetate	1.743	1.692	1.715	1.591	1.604	1.517	1.644
24) P	1,1-Dichloroet...	1.192	1.152	1.156	1.063	1.110	1.043	1.120
25) T	2-Butanone	0.604	0.598	0.604	0.551	0.573	0.533	0.577
26) T	2,2-Dichloropr...	0.967	0.796	0.778	0.905	0.912	0.868	0.871
27) T	cis-1,2-Dichlo...	0.786	0.766	0.762	0.699	0.729	0.701	0.740
28) T	Bromochloromet...	0.579	0.564	0.616	0.466	0.517	0.560	0.550
29) T	Tetrahydrofuran	0.390	0.390	0.399	0.358	0.372	0.346	0.376
30) C	Chloroform	1.235	1.151	1.145	1.061	1.085	1.030	1.118
31) T	Cyclohexane	1.303	1.116	1.004	1.030	0.976	1.086	12.21
32) T	1,1,1-Trichlor...	1.029	0.995	0.969	0.895	0.925	0.893	0.951
33) S	1,2-Dichloroet...	0.732	0.707	0.500	0.656	0.751	0.669	15.08
34) I	1,4-Difluorobenzene	-----	ISTD-----					
35) S	Dibromofluorom...	0.303	0.310	0.219	0.298	0.351	0.296	16.15
36) T	1,1-Dichloropr...	0.467	0.458	0.438	0.418	0.442	0.426	0.442
37) T	Ethyl Acetate	0.544	0.615	0.577	0.536	0.565	0.535	0.562
38) T	Carbon Tetrach...	0.453	0.449	0.434	0.409	0.435	0.421	0.433
39) T	Methylcyclohexane	0.633	0.645	0.588	0.570	0.603	0.589	0.605
40) TM	Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444
41) T	Methacrylonitrile	0.356	0.319	0.318	0.303	0.299	0.299	0.316
42) TM	1,2-Dichloroet...	0.473	0.456	0.444	0.411	0.430	0.413	0.438
43) T	Isopropyl Acetate	0.975	0.950	0.906	0.852	0.884	0.845	0.902
44) TM	Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.327	0.342
45) C	1,2-Dichloropr...	0.366	0.369	0.354	0.332	0.352	0.335	0.351
46) T	Dibromomethane	0.233	0.242	0.241	0.221	0.237	0.226	0.233
47) T	Bromodichlorom...	0.510	0.484	0.480	0.457	0.483	0.465	0.480
48) T	Methyl methacr...	0.444	0.415	0.421	0.394	0.421	0.397	0.415
49) T	1,4-Dioxane	0.006	0.007	0.008	0.008	0.008	0.007	0.008
50) S	Toluene-d8	1.245	1.203	0.861	1.178	1.377	1.173	16.23
51) T	4-Methyl-2-Pen...	0.505	0.549	0.576	0.538	0.562	0.528	0.543
52) CM	Toluene	0.918	0.914	0.885	0.835	0.883	0.859	0.882
53) T	t-1,3-Dichloro...	0.571	0.526	0.524	0.522	0.548	0.530	0.537
54) T	cis-1,3-Dichlo...	0.609	0.577	0.564	0.551	0.584	0.561	0.574
55) T	1,1,2-Trichlor...	0.359	0.355	0.342	0.322	0.335	0.323	0.340
56) T	Ethyl methacry...	0.433	0.516	0.567	0.556	0.595	0.578	0.541

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\  
 Method File : 82N060625W.M

57) T	1,3-Dichloropr...	0.643	0.600	0.587	0.561	0.583	0.559	0.589	5.27
58) T	2-Chloroethyl ...	0.278	0.305	0.334	0.274	0.340	0.405	0.323	15.07
59) T	2-Hexanone	0.312	0.282	0.363	0.368	0.397	0.376	0.350	12.36
60) T	Dibromochlorom...	0.361	0.351	0.358	0.340	0.363	0.349	0.354	2.49
61) T	1,2-Dibromoethane	0.353	0.358	0.354	0.329	0.354	0.341	0.348	3.18
62) S	4-Bromofluorob...	0.441	0.446	0.325	0.446	0.521	0.436		16.16
63) I	Chlorobenzene-d5	-----ISTD-----							
64) T	Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.51
65) PM	Chlorobenzene	1.233	1.135	1.107	1.023	1.089	1.030	1.103	7.01
66) T	1,1,1,2-Tetra...	0.362	0.374	0.358	0.338	0.356	0.338	0.354	3.99
67) C	Ethyl Benzene	1.989	1.975	1.907	1.796	1.913	1.816	1.899	4.19#
68) T	m/p-Xylenes	0.730	0.751	0.741	0.701	0.736	0.703	0.727	2.82
69) T	o-Xylene	0.714	0.699	0.702	0.674	0.711	0.678	0.696	2.42
70) T	Styrene	1.177	1.193	1.226	1.162	1.229	1.164	1.192	2.51
71) P	Bromoform	0.217	0.266	0.276	0.265	0.286	0.267	0.263	9.09
72) I	1,4-Dichlorobenzen...	-----ISTD-----							
73) T	Isopropylbenzene	3.864	3.749	3.649	3.426	3.621	3.546	3.643	4.21
74) T	N-amyl acetate	1.376	1.110	1.187	1.266	1.372	1.327	1.273	8.42
75) P	1,1,2,2-Tetra...	1.292	1.299	1.273	1.178	1.205	1.157	1.234	4.99
76) T	1,2,3-Trichlor...	1.297	1.189	1.129	1.173	1.201	1.142	1.189	5.03
77) T	Bromobenzene	0.870	0.855	0.840	0.790	0.838	0.819	0.835	3.36
78) T	n-propylbenzene	4.530	4.627	4.449	4.211	4.424	4.317	4.427	3.35
79) T	2-Chlorotoluene	2.852	2.723	2.674	2.507	2.618	2.554	2.655	4.69
80) T	1,3,5-Trimethyl...	3.004	3.091	3.035	2.898	3.064	2.953	3.007	2.39
81) T	trans-1,4-Dich...		0.510	0.522	0.476	0.546	0.528	0.516	5.06
82) T	4-Chlorotoluene	2.899	2.757	2.671	2.531	2.664	2.595	2.686	4.81
83) T	tert-Butylbenzene	2.942	2.810	2.736	2.617	2.745	2.668	2.753	4.14
84) T	1,2,4-Trimethyl...	2.952	3.122	3.063	2.914	3.075	2.968	3.016	2.73
85) T	sec-Butylbenzene	4.114	4.126	4.037	3.829	4.018	3.861	3.998	3.15
86) T	p-Isopropyltol...	3.305	3.425	3.321	3.184	3.366	3.227	3.305	2.68
87) T	1,3-Dichlorobe...	1.763	1.709	1.657	1.554	1.612	1.566	1.644	5.01
88) T	1,4-Dichlorobe...	1.820	1.786	1.657	1.572	1.642	1.576	1.676	6.27
89) T	n-Butylbenzene	3.473	3.292	3.140	3.059	3.188	3.054	3.201	4.99
90) T	Hexachloroethane	0.580	0.537	0.569	0.537	0.577	0.562	0.560	3.43
91) T	1,2-Dichlorobe...	1.675	1.651	1.596	1.500	1.557	1.496	1.579	4.77
92) T	1,2-Dibromo-3...	0.339	0.317	0.290	0.272	0.283	0.270	0.295	9.29
93) T	1,2,4-Trichlor...	1.042	1.037	0.991	0.969	1.016	0.994	1.008	2.82
94) T	Hexachlorobuta...	0.364	0.391	0.385	0.363	0.382	0.369	0.376	3.11
95) T	Naphthalene	3.820	3.727	3.761	3.600	3.839	3.772	3.753	2.27
96) T	1,2,3-Trichlor...	1.073	1.009	0.993	0.950	1.000	0.987	1.002	4.05

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086862.D  
 Acq On : 06 Jun 2025 12:44  
 Operator : JC\MD  
 Sample : VSTDICC001  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC001

Quant Time: Jun 07 01:54:26 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.235	168	212682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	393457	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167399	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount 50.000	Range 74 - 125		Recovery	=	0.000%#	
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount 50.000	Range 75 - 124		Recovery	=	0.000%#	
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount 50.000	Range 86 - 113		Recovery	=	0.000%#	
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount 50.000	Range 77 - 121		Recovery	=	0.000%#	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.153	85	1985	0.936	ug/l	72
3) Chloromethane	2.400	50	3241	1.184	ug/l	98
4) Vinyl Chloride	2.559	62	2848	1.008	ug/l #	88
6) Chloroethane	3.159	64	1957	1.073	ug/l #	79
7) Trichlorofluoromethane	3.530	101	3752	1.017	ug/l	97
8) Diethyl Ether	3.994	74	1583	0.984	ug/l	76
9) 1,1,2-Trichlorotrifluo...	4.377	101	2358m	1.017	ug/l	
12) 1,1-Dichloroethene	4.365	96	2436	1.029	ug/l	85
14) Allyl chloride	5.041	41	4189	1.067	ug/l #	75
15) Acrylonitrile	5.735	53	9222	5.106	ug/l	96
16) Acetone	4.459	43	9069	6.006	ug/l	95
17) Carbon Disulfide	4.730	76	7306	1.116	ug/l	97
18) Methyl Acetate	5.053	43	4402	1.000	ug/l #	66
19) Methyl tert-butyl Ether	5.824	73	9019	1.052	ug/l	92
20) Methylene Chloride	5.294	84	3496	1.237	ug/l	93
21) trans-1,2-Dichloroethene	5.800	96	2979	1.131	ug/l	87
22) Diisopropyl ether	6.682	45	8584	1.037	ug/l #	96
23) Vinyl Acetate	6.618	43	37075	5.303	ug/l	95
24) 1,1-Dichloroethane	6.588	63	5071	1.065	ug/l #	81
25) 2-Butanone	7.500	43	12841	5.231	ug/l	93
26) 2,2-Dichloropropane	7.500	77	4112	1.110	ug/l	94
27) cis-1,2-Dichloroethene	7.494	96	3345	1.062	ug/l	95
28) Bromochloromethane	7.824	49	2464	1.052	ug/l #	88
29) Tetrahydrofuran	7.853	42	8301	5.191	ug/l	100
30) Chloroform	7.971	83	5255	1.105	ug/l	97
32) 1,1,1-Trichloroethane	8.171	97	4378	1.082	ug/l #	50
36) 1,1-Dichloropropene	8.376	75	3678	1.059	ug/l	96
37) Ethyl Acetate	7.577	43	4282	0.968	ug/l #	78
38) Carbon Tetrachloride	8.376	117	3563	1.045	ug/l #	93
39) Methylcyclohexane	9.606	83	4981	1.046	ug/l #	90
40) Benzene	8.612	78	12498	1.100	ug/l	100
41) Methacrylonitrile	7.782	41	2799	1.127	ug/l	89
42) 1,2-Dichloroethane	8.682	62	3725	1.081	ug/l	82
43) Isopropyl Acetate	8.694	43	7673	1.081	ug/l #	96
44) Trichloroethene	9.365	130	2822	1.047	ug/l	71
45) 1,2-Dichloropropane	9.629	63	2883	1.043	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086862.D  
 Acq On : 06 Jun 2025 12:44  
 Operator : JC\MD  
 Sample : VSTDICC001  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC001

Quant Time: Jun 07 01:54:26 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.712	93	1835	0.999	ug/1	94
47) Bromodichloromethane	9.888	83	4017	1.063	ug/1 #	96
48) Methyl methacrylate	9.688	41	3492	1.069	ug/1	88
49) 1,4-Dioxane	9.706	88	1004	16.891	ug/1 #	74
51) 4-Methyl-2-Pentanone	10.453	43	19886	4.653	ug/1	99
52) Toluene	10.635	92	7227	1.041	ug/1	95
53) t-1,3-Dichloropropene	10.841	75	4494	1.064	ug/1	97
54) cis-1,3-Dichloropropene	10.312	75	4792	1.060	ug/1	97
55) 1,1,2-Trichloroethane	11.023	97	2828	1.058	ug/1	89
56) Ethyl methacrylate	10.882	69	3406	0.800	ug/1 #	85
57) 1,3-Dichloropropane	11.170	76	5062	1.092	ug/1	98
58) 2-Chloroethyl Vinyl ether	10.165	63	10924	4.304	ug/1	97
59) 2-Hexanone	11.235	43	12295m	4.466	ug/1	
60) Dibromochloromethane	11.359	129	2843	1.021	ug/1	96
61) 1,2-Dibromoethane	11.470	107	2777	1.014	ug/1	96
64) Tetrachloroethene	11.106	164	2473	1.123	ug/1	94
65) Chlorobenzene	11.894	112	8581	1.118	ug/1 #	87
66) 1,1,1,2-Tetrachloroethane	11.959	131	2523	1.023	ug/1 #	63
67) Ethyl Benzene	11.970	91	13845	1.047	ug/1	94
68) m/p-Xylenes	12.070	106	10163	2.008	ug/1	95
69) o-Xylene	12.400	106	4970	1.025	ug/1	93
70) Styrene	12.417	104	8189	0.987	ug/1	98
71) Bromoform	12.582	173	1508	0.825	ug/1 #	78
73) Isopropylbenzene	12.694	105	12938	1.061	ug/1	98
74) N-amyl acetate	12.647	43	4608m	1.138	ug/1	
75) 1,1,2,2-Tetrachloroethane	12.935	83	4326	1.047	ug/1 #	93
76) 1,2,3-Trichloropropane	13.000	75	4343m	1.088	ug/1	
77) Bromobenzene	12.982	156	2913	1.042	ug/1	96
78) n-propylbenzene	13.035	91	15168	1.023	ug/1	98
79) 2-Chlorotoluene	13.123	91	9550	1.074	ug/1	97
80) 1,3,5-Trimethylbenzene	13.170	105	10058	0.999	ug/1	96
82) 4-Chlorotoluene	13.223	91	9706	1.079	ug/1	96
83) tert-Butylbenzene	13.435	119	9849	1.069	ug/1	98
84) 1,2,4-Trimethylbenzene	13.482	105	9882	0.979	ug/1	99
85) sec-Butylbenzene	13.617	105	13775	1.029	ug/1	100
86) p-Isopropyltoluene	13.729	119	11065	1.000	ug/1	97
87) 1,3-Dichlorobenzene	13.735	146	5903	1.073	ug/1	94
88) 1,4-Dichlorobenzene	13.811	146	6094m	1.086	ug/1	
89) n-Butylbenzene	14.058	91	11627	1.085	ug/1	94
90) Hexachloroethane	14.329	117	1942	1.036	ug/1	93
91) 1,2-Dichlorobenzene	14.106	146	5607	1.061	ug/1	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	1136	1.150	ug/1	78
93) 1,2,4-Trichlorobenzene	15.388	180	3489	1.034	ug/1	98
94) Hexachlorobutadiene	15.500	225	1219	0.969	ug/1	88
95) Naphthalene	15.635	128	12789	1.018	ug/1	98
96) 1,2,3-Trichlorobenzene	15.835	180	3594	1.072	ug/1	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
Data File : VN086862.D  
Acq On : 06 Jun 2025 12:44  
Operator : JC\MD  
Sample : VSTDICC001  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Jun 07 01:54:26 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 01:51:02 2025  
Response via : Initial Calibration

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC001

## Manual Integrations APPROVED


Reviewed By :John Caralone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.235 min Scan# 1  
Delta R.T. -0.001 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001




Tgt Ion:168 Resp: 21268  
Ion Ratio Lower Upper  
168 100  
99 58.1 49.1 73.7

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025









#2  
Dichlorodifluoromethane  
Concen: 0.936 ug/l  
RT: 2.153 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44



Tgt Ion: 85 Resp: 1985  
Ion Ratio Lower Upper  
85 100  
87 48.3 16.2 48.6





#3

Chloromethane

Concen: 1.184 ug/l

RT: 2.400 min Scan# 7

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#4

Vinyl Chloride

Concen: 1.008 ug/l

RT: 2.559 min Scan# 103

Delta R.T. 0.006 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44


Tgt Ion: 62 Resp: 2848

Ion Ratio Lower Upper

62 100


64 25.9 26.0 39.0#





#6  
Chloroethane  
Concen: 1.073 ug/l  
RT: 3.159 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001




Tgt Ion: 64 Resp: 1951  
Ion Ratio Lower Upper  
64 100  
66 42.3 24.6 36.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#7  
Trichlorofluoromethane  
Concen: 1.017 ug/l  
RT: 3.530 min Scan# 268  
Delta R.T. 0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44



Tgt Ion:101 Resp: 3752  
Ion Ratio Lower Upper  
101 100  
103 67.2 52.2 78.2





#8

Diethyl Ether

Concen: 0.984 ug/l

RT: 3.994 min Scan# 3

Instrument:

Delta R.T. 0.011 min


MSVOA\_N

Lab File: VN086862.D

ClientSampleId :

Acq: 06 Jun 2025 12:44

VSTDICC001



Tgt Ion: 74 Resp: 1583

Ion Ratio Lower Upper







74 100




45 113.8 45.5 136.5

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 1.017 ug/l m  
RT: 4.377 min Scan# 412  
Delta R.T. -0.018 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44Tgt Ion:101 Resp: 2358  
Ion Ratio Lower Upper  
101 100  
85 27.8 33.8 50.8#  
151 72.1 58.7 88.1



#12

1,1-Dichloroethene

Concen: 1.029 ug/l

RT: 4.365 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 96 Resp: 2430

Ion Ratio Lower Upper

96 100




61 137.1 125.8 188.6

98 55.8 52.5 78.7

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 1.067 ug/l

RT: 5.041 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion: 41 Resp: 4189







Ion Ratio Lower Upper

41 100

39 35.3 51.3 76.9#

76 36.4 31.4 47.0





#15

Acrylonitrile

Concen: 5.106 ug/l

RT: 5.735 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 53 Resp: 922

Ion Ratio Lower Upper

53 100

52 85.4 65.0 97.6


51 32.7 27.2 40.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 6.006 ug/l

RT: 4.459 min Scan# 426

Delta R.T. 0.011 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44




Tgt Ion: 43 Resp: 9069

Ion Ratio Lower Upper

43 100

58 38.0 28.0 42.0





#17

Carbon Disulfide

Concen: 1.116 ug/l

RT: 4.730 min Scan# 4

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 1.000 ug/l

RT: 5.053 min Scan# 527

Delta R.T. 0.011 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44




Tgt Ion: 43 Resp: 4402

Ion Ratio Lower Upper

43 100

74 8.7 20.7 31.1#





#19

Methyl tert-butyl Ether

Concen: 1.052 ug/l

RT: 5.824 min Scan# 6

Delta R.T. 0.011 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 1.237 ug/l


RT: 5.294 min Scan# 568




Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt	Ion	Resp:	3496
	84	100	
	49	125.7	90.5 135.7
	51	35.9	28.5 42.7
	86	60.5	48.1 72.1









#21  
trans-1,2-Dichloroethene  
Concen: 1.131 ug/l  
RT: 5.800 min Scan# 6  
Delta R.T. -0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#22  
Diisopropyl ether  
Concen: 1.037 ug/l  
RT: 6.682 min Scan# 804  
Delta R.T. -0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Tgt Ion: 45 Resp: 8584  
Ion Ratio Lower Upper  
45 100  
43 51.5 41.9 62.9  
87 24.5 22.4 33.6  
59 8.4 9.8 14.6#





#23

**Vinyl Acetate**

Concen: 5.303 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

**1,1-Dichloroethane**

Concen: 1.065 ug/l

RT: 6.588 min Scan# 788


Delta R.T. 0.006 min




Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 63 Resp: 5071

Ion	Ratio	Lower	Upper
63	100		
98	0.0	3.6	10.8#
100	0.0	2.6	7.8#





#25

2-Butanone

Concen: 5.231 ug/l

RT: 7.500 min Scan# 9

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 1.110 ug/l


RT: 7.500 min Scan# 943







Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 77 Resp: 4112  
 Ion Ratio Lower Upper  
 77 100  
 97 20.1 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 1.062 ug/l

RT: 7.494 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA\_N


ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 1.052 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion: 49 Resp: 2464







Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 67.5 62.2 93.2





#29

Tetrahydrofuran

Concen: 5.191 ug/l

RT: 7.853 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 1.105 ug/l


RT: 7.971 min Scan# 1023


Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 83 Resp: 5255  
 Ion Ratio Lower Upper  
 83 100  
 85 68.6 52.9 79.3







#32

1,1,1-Trichloroethane  
Concen: 1.082 ug/l  
RT: 8.171 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#33


1,2-Dichloroethane-d4  
Concen: 0.000 ug/l  
Expected RT: 8.59 min

Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44


Tgt Ion: 65  
Sig Exp Ratio  
65 100  
67 52.8



Abundance Scan 1216 (9.106 min): VN086862.D\data.ms



Abundance Scan 1216 (9.106 min): VN086862.D\data.ms (-1)



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 8.17 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 113

Sig Exp Ratio

113 100


111 105.2


192 17.8

Abundance Ion 112.80 (112.50 to 113.50): VN086862.D\data.ms

Ion 110.80 (110.50 to 111.50): VN086862.D\data.ms

Ion 191.70 (191.40 to 192.40): VN086862.D\data.ms





#36

1,1-Dichloropropene

Concen: 1.059 ug/l

RT: 8.376 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 75 Resp: 3673

Ion Ratio Lower Upper

75 100

110 31.8 17.6 52.9



77 30.2 25.3 37.9


Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025





#38

Carbon Tetrachloride

Concen: 1.045 ug/l

RT: 8.376 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion:117 Resp: 3561

Ion Ratio Lower Upper

117 100

119 89.0 75.0 112.6



121 23.8 24.9 37.3




Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025





#40

Benzene

Concen: 1.100 ug/l

RT: 8.612 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#41

Methacrylonitrile

Concen: 1.127 ug/l

RT: 7.782 min Scan# 991


Delta R.T. -0.006 min


Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 41 Resp: 2799

Ion	Ratio	Lower	Upper
41	100		
39	50.3	41.2	61.8
67	62.4	61.4	92.2
52	22.9	22.7	34.1





#42

1,2-Dichloroethane

Concen: 1.081 ug/l


RT: 8.682 min Scan# 1143

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
 ClientSampleId : VSTDICC001



Tgt Ion: 62 Resp: 3725




Ion Ratio Lower Upper

62 100

98 3.9 0.0 21.2

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#43  
 Isopropyl Acetate  
 Concen: 1.081 ug/l  
 RT: 8.694 min Scan# 1146  
 Delta R.T. -0.000 min  
 Lab File: VN086862.D  
 Acq: 06 Jun 2025 12:44




Tgt Ion: 43 Resp: 7673


Ion Ratio Lower Upper

43 100

61 26.2 22.1 33.1

87 11.7 11.8 17.6#





#44

Trichloroethene

Concen: 1.047 ug/l

RT: 9.365 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086862.D



Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001



Tgt Ion: 130 Resp: 2822

Ion Ratio Lower Upper



130 100

95 69.1 0.0 196.2

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#45

1,2-Dichloropropane


Concen: 1.043 ug/l

RT: 9.629 min Scan# 1305

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44






Tgt Ion: 63 Resp: 2883

Ion Ratio Lower Upper

63 100

65 29.0 25.5 38.3







#46  
Dibromomethane  
Concen: 0.999 ug/l  
RT: 9.712 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001


Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#47  
Bromodichloromethane  
Concen: 1.063 ug/l  
RT: 9.888 min Scan# 1349  
Delta R.T. -0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Tgt Ion: 83 Resp: 4017  
Ion Ratio Lower Upper  
83 100  
85 62.7 50.9 76.3  
127 0.0 6.6 10.0#





#48

Methyl methacrylate

Concen: 1.069 ug/l

RT: 9.688 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 41 Resp: 349

Ion Ratio Lower Upper

41 100

69 79.4 74.6 112.0



39 44.9 40.2 60.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#49

1,4-Dioxane

Concen: 16.891 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion: 88 Resp: 1004


Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 94.0 57.7 86.5#







#50  
Toluene-d8  
Concen: 0.000 ug/l  
Expected RT: 10.57 min  
  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44  
  
Tgt Ion: 98  
Sig Exp Ratio  
98 100  
100 66.7


Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICC001

**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#51  
4-Methyl-2-Pentanone  
Concen: 4.653 ug/l  
RT: 10.453 min Scan# 1445  
Delta R.T. 0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44





Tgt Ion: 43 Resp: 19886  
Ion Ratio Lower Upper  
43 100  
58 42.3 34.2 51.4





#52  
Toluene  
Concen: 1.041 ug/l  
RT: 10.635 min Scan# 1476  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44



Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001




Tgt Ion: 92 Resp: 7222  
Ion Ratio Lower Upper  
92 100  
91 178.1 136.5 204.7

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#53  
t-1,3-Dichloropropene  
Concen: 1.064 ug/l  
RT: 10.841 min Scan# 1511  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44





Tgt Ion: 75 Resp: 4494  
Ion Ratio Lower Upper  
75 100  
77 31.4 26.4 39.6






Abundance Scan 1421 (10.312 min): VN086862.D\data.ms (-)



Abundance Scan 1542 (11.023 min): VN086862.D\data.ms (-)



Abundance Scan 1542 (11.023 min): VN086862.D\data.ms (-)

#54

cis-1,3-Dichloropropene

Concen: 1.060 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument :

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 75 Resp: 4792

Ion Ratio Lower Upper

75 100

77 35.9 26.2 39.4


39 41.1 32.3 48.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#55

1,1,2-Trichloroethane

Concen: 1.058 ug/l

RT: 11.023 min Scan# 1542

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 97 Resp: 2828


Ion Ratio Lower Upper




97 100

83 80.6 69.7 104.5

85 46.5 45.1 67.7

99 52.5 50.6 75.8





#56

Ethyl methacrylate

Concen: 0.800 ug/l

RT: 10.882 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#57

1,3-Dichloropropane

Concen: 1.092 ug/l


RT: 11.170 min Scan# 1567




Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion: 76 Resp: 5062  
 Ion Ratio Lower Upper  
 76 100  
 78 32.8 25.3 37.9





#58

2-Chloroethyl Vinyl ether

Concen: 4.304 ug/l

RT: 10.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#59

2-Hexanone

Concen: 4.466 ug/l

RT: 11.235 min Scan# 1578

Delta R.T. 0.035 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44




Tgt Ion: 43 Resp: 12295

Ion Ratio Lower Upper

43 100

58 4.7 30.4 91.3#





#60

Dibromochloromethane

Concen: 1.021 ug/l

RT: 11.359 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#61

1,2-Dibromoethane

Concen: 1.014 ug/l

RT: 11.470 min Scan# 1618

Delta R.T. -0.000 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44


Tgt Ion:107 Resp: 2777

Ion Ratio Lower Upper

107 100

109 91.2 76.3 114.5













#62  
4-Bromofluorobenzene  
Concen: 0.000 ug/l  
Expected RT: 12.85 min  
  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44  
  
Tgt Ion: 95  
Sig Exp Ratio  
95 100  
174 70.9  
176 66.3

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICC001

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025





#64

Tetrachloroethene

Concen: 1.123 ug/l

RT: 11.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

Abundance

1500

1100

700

300




100

0

Time--&gt;

11.05 11.10 11.15

11.106



#65

Chlorobenzene

Concen: 1.118 ug/l

RT: 11.894 min Scan# 1690

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:112 Resp: 8581

Ion Ratio Lower Upper

112 100

114 25.7 26.2 39.2#

Abundance

4000

3000

2000


1000

0


Time--&gt;

11.894

11.90



Abundance Scan 1701 (11.959 min): VN086862.D\data.ms (-)



#66

1,1,1,2-Tetrachloroethane

Concen: 1.023 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#67

Ethyl Benzene

Concen: 1.047 ug/l

RT: 11.970 min Scan# 1703

Delta R.T. 0.006 min

Lab File: VN086862.D



Acq: 06 Jun 2025 12:44




Tgt Ion: 91 Resp: 13845

Ion Ratio Lower Upper

91 100

106 28.7 25.4 38.2









#68  
m/p-Xylenes  
Concen: 2.008 ug/l  
RT: 12.070 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001

Tgt Ion:106 Resp: 1016  
Ion Ratio Lower Upper  
106 100  
91 207.3 159.4 239.0


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#69  
o-Xylene  
Concen: 1.025 ug/l  
RT: 12.400 min Scan# 1776  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Tgt Ion:106 Resp: 4970  
Ion Ratio Lower Upper  
106 100  
91 223.6 106.1 318.1









#70  
Styrene  
Concen: 0.987 ug/l  
RT: 12.417 min Scan# 1  
Instrument : MSVOA\_N  
Delta R.T. 0.006 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44  
ClientSampleId : VSTDICC001

Tgt Ion:104 Resp: 8189  
Ion Ratio Lower Upper  
104 100  
78 49.6 38.5 57.7  
103 52.5 43.1 64.7


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#71  
Bromoform  
Concen: 0.825 ug/l  
RT: 12.582 min Scan# 1807  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Tgt Ion:173 Resp: 1508  
Ion Ratio Lower Upper  
173 100  
175 63.1 24.1 72.4  
254 0.0 0.0 0.0






1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.794 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



Isopropylbenzene  
Concen: 1.061 ug/l  
RT: 12.694 min Scan# 1826  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Tgt Ion:105 Resp: 12938  
Ion Ratio Lower Upper  
105 100  
120 28.1 13.5 40.4





#74

N-amyl acetate

Concen: 1.138 ug/l m

RT: 12.647 min Scan# 1

Delta R.T. 0.135 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 43 Resp: 460#

Ion Ratio Lower Upper

43 100

70 0.0 37.6 56.4#

55 0.0 19.0 28.6#





61 0.0 20.8 31.2#

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 1.047 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion: 83 Resp: 4326


Ion Ratio Lower Upper

83 100


131 4.1 4.9 14.7#

85 68.6 32.0 96.0






Abundance Scan 1878 (13.000 min): VN086862.D\data.ms (-)



Abundance Scan 1878 (13.000 min): VN086862.D\data.ms (-)



#76

1,2,3-Trichloropropane

Concen: 1.088 ug/l m

RT: 13.000 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC001


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 1875 (12.982 min): VN086862.D\data.ms (-)



Abundance Scan 1875 (12.982 min): VN086862.D\data.ms (-)



#77

Bromobenzene

Concen: 1.042 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44




Tgt Ion:156 Resp: 2913

Ion Ratio Lower Upper

156 100

77 230.2 117.7 353.1

158 102.5 48.1 144.4



#77

Bromobenzene

Concen: 1.042 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion:156 Resp: 2913




Ion Ratio Lower Upper

156 100

77 230.2 117.7 353.1

158 102.5 48.1 144.4





#78

n-propylbenzene

Concen: 1.023 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC001

Tgt Ion: 91 Resp: 15163

Ion Ratio Lower Upper

91 100





120 21.9 11.4 34.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#79

2-Chlorotoluene

Concen: 1.074 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN086862.D


Acq: 06 Jun 2025 12:44




Tgt Ion: 91 Resp: 9550

Ion Ratio Lower Upper

91 100

126 30.4 16.2 48.5





#80

1,3,5-Trimethylbenzene

Concen: 0.999 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#82

4-Chlorotoluene

Concen: 1.079 ug/l


RT: 13.223 min Scan# 1916




Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt	Ion:	91	Resp:	9706
Ion	Ratio	Lower	Upper	
91	100			
126	31.1	16.7	50.1	





#83

tert-Butylbenzene

Concen: 1.069 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.979 ug/l


RT: 13.482 min Scan# 1960




Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt	Ion	105	100	9882
	Ion Ratio			
	Lower			
105	100			
120	47.2	23.2	69.6	









#85  
sec-Butylbenzene  
Concen: 1.029 ug/l  
RT: 13.617 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC001


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#86  
p-Isopropyltoluene  
Concen: 1.000 ug/l  
RT: 13.729 min Scan# 2002  
Delta R.T. -0.000 min  
Lab File: VN086862.D  
Acq: 06 Jun 2025 12:44

Tgt Ion:119 Resp: 11065  
Ion Ratio Lower Upper  
119 100  
134 25.4 13.5 40.4  
91 24.1 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 1.073 ug/l

RT: 13.735 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 1.086 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

2

Tgt Ion:146 Resp: 6094

3

Ion Ratio Lower Upper

4

146 100

5

111 44.4 19.9 59.7

6

148 57.8 31.8 95.4

7

8

9

10

11


12




13

14

15

16





#89

n-Butylbenzene

Concen: 1.085 ug/l

RT: 14.058 min Scan# 2

Delta R.T. 0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 1.036 ug/l

RT: 14.329 min Scan# 2104


Delta R.T. -0.006 min


Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:117 Resp: 1942

Ion	Ratio	Lower	Upper
117	100		
201	60.6	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 1.061 ug/l

RT: 14.106 min Scan# 210

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 1.150 ug/l


RT: 14.717 min Scan# 2170







Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt	Ion:	Resp:	1136
Ion	Ratio	Lower	Upper
75	100		
155	63.6	41.1	123.4
157	81.0	52.7	158.1





#93

1,2,4-Trichlorobenzene

Concen: 1.034 ug/l

RT: 15.388 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 0.969 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44


Tgt Ion:225 Resp: 1219




Ion Ratio Lower Upper

225 100

223 53.6 33.5 100.4

227 57.5 31.1 93.5





#95

Naphthalene

Concen: 1.018 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.006 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC001

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 1.072 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN086862.D

Acq: 06 Jun 2025 12:44

Tgt Ion:180 Resp: 3594


Ion Ratio Lower Upper

180 100

182 91.6 47.0 141.0

145 29.0 16.6 49.8

Time--&gt;



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086863.D  
 Acq On : 06 Jun 2025 13:17  
 Operator : JC\MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC005

Quant Time: Jun 07 01:55:22 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	229701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	423980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181065	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	16818	5.468	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	10.940%#	
35) Dibromofluoromethane	8.177	113	12867	5.121	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	10.240%#	
50) Toluene-d8	10.571	98	52794	5.308	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	10.620%#	
62) 4-Bromofluorobenzene	12.847	95	18697	5.059	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	10.120%#	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	10196	4.452	ug/l	93
3) Chloromethane	2.395	50	15025	5.080	ug/l	98
4) Vinyl Chloride	2.554	62	15384	5.043	ug/l	97
5) Bromomethane	2.989	94	8606	5.041	ug/l	97
6) Chloroethane	3.154	64	10190	5.171	ug/l	96
7) Trichlorofluoromethane	3.518	101	20740	5.203	ug/l	99
8) Diethyl Ether	3.971	74	9121	5.252	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.401	101	13034	5.205	ug/l	97
10) Methyl Iodide	4.606	142	16530	5.092	ug/l	98
11) Tert butyl alcohol	5.542	59	22000	26.363	ug/l	99
12) 1,1-Dichloroethene	4.359	96	13610	5.323	ug/l	94
13) Acrolein	4.195	56	8401	31.804	ug/l	99
14) Allyl chloride	5.042	41	20925	4.935	ug/l	99
15) Acrylonitrile	5.736	53	49826	25.545	ug/l	99
16) Acetone	4.448	43	41981	25.741	ug/l	99
17) Carbon Disulfide	4.736	76	37246	5.267	ug/l	97
18) Methyl Acetate	5.048	43	24096	5.070	ug/l	97
19) Methyl tert-butyl Ether	5.812	73	46823	5.058	ug/l	95
20) Methylene Chloride	5.300	84	15809	5.178	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	15480	5.442	ug/l #	81
22) Diisopropyl ether	6.689	45	46396	5.191	ug/l	99
23) Vinyl Acetate	6.618	43	194334	25.735	ug/l	100
24) 1,1-Dichloroethane	6.583	63	26473	5.147	ug/l	99
25) 2-Butanone	7.494	43	68629	25.888	ug/l	99
26) 2,2-Dichloropropane	7.500	77	18288	4.571	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	17594	5.172	ug/l	98
28) Bromochloromethane	7.830	49	12958	5.124	ug/l #	97
29) Tetrahydrofuran	7.853	42	44749	25.912	ug/l	98
30) Chloroform	7.977	83	26450	5.149	ug/l	95
31) Cyclohexane	8.271	56	29938	6.002	ug/l	95
32) 1,1,1-Trichloroethane	8.177	97	22857	5.232	ug/l #	50
36) 1,1-Dichloropropene	8.377	75	19410	5.184	ug/l	98
37) Ethyl Acetate	7.571	43	26073	5.470	ug/l	97
38) Carbon Tetrachloride	8.371	117	19056	5.184	ug/l	99
39) Methylcyclohexane	9.606	83	27353	5.333	ug/l	95
40) Benzene	8.612	78	63646	5.199	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086863.D  
 Acq On : 06 Jun 2025 13:17  
 Operator : JC\MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC005

Quant Time: Jun 07 01:55:22 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	13518	5.050	ug/1	99
42) 1,2-Dichloroethane	8.683	62	19329	5.205	ug/1	98
43) Isopropyl Acetate	8.700	43	40268	5.264	ug/1	98
44) Trichloroethene	9.359	130	15245	5.251	ug/1	96
45) 1,2-Dichloropropane	9.624	63	15632	5.248	ug/1	99
46) Dibromomethane	9.718	93	10241	5.176	ug/1	99
47) Bromodichloromethane	9.888	83	20512	5.039	ug/1	97
48) Methyl methacrylate	9.688	41	17591	4.996	ug/1	94
49) 1,4-Dioxane	9.700	88	6141	95.874	ug/1 #	93
51) 4-Methyl-2-Pentanone	10.447	43	116434	25.283	ug/1	97
52) Toluene	10.630	92	38744	5.178	ug/1	99
53) t-1,3-Dichloropropene	10.841	75	22285	4.896	ug/1	98
54) cis-1,3-Dichloropropene	10.318	75	24477	5.026	ug/1	93
55) 1,1,2-Trichloroethane	11.018	97	15068	5.234	ug/1	97
56) Ethyl methacrylate	10.888	69	21893	4.774	ug/1	95
57) 1,3-Dichloropropane	11.165	76	25452	5.096	ug/1	98
58) 2-Chloroethyl Vinyl ether	10.165	63	64722	23.664	ug/1	98
59) 2-Hexanone	11.212	43	59884	20.188	ug/1	99
60) Dibromochloromethane	11.359	129	14884	4.962	ug/1	98
61) 1,2-Dibromoethane	11.471	107	15172	5.141	ug/1	99
64) Tetrachloroethene	11.106	164	12322	5.236	ug/1	92
65) Chlorobenzene	11.894	112	42195	5.145	ug/1	98
66) 1,1,1,2-Tetrachloroethane	11.965	131	13897	5.271	ug/1	97
67) Ethyl Benzene	11.965	91	73443	5.199	ug/1	97
68) m/p-Xylenes	12.071	106	55859	10.330	ug/1	100
69) o-Xylene	12.400	106	26004	5.021	ug/1	96
70) Styrene	12.412	104	44380	5.008	ug/1	98
71) Bromoform	12.582	173	9890	5.063	ug/1 #	96
73) Isopropylbenzene	12.694	105	67878	5.146	ug/1	100
74) N-amyl acetate	12.559	43	20096m	4.590	ug/1	
75) 1,1,2,2-Tetrachloroethane	12.935	83	23516	5.262	ug/1	98
76) 1,2,3-Trichloropropane	12.994	75	21527m	4.986	ug/1	
77) Bromobenzene	12.976	156	15485	5.119	ug/1	92
78) n-propylbenzene	13.035	91	83781	5.226	ug/1	99
79) 2-Chlorotoluene	13.124	91	49297	5.128	ug/1	100
80) 1,3,5-Trimethylbenzene	13.171	105	55961	5.139	ug/1	100
81) trans-1,4-Dichloro-2-b...	12.741	75	9234	4.939	ug/1	94
82) 4-Chlorotoluene	13.223	91	49924	5.132	ug/1	100
83) tert-Butylbenzene	13.435	119	50874	5.103	ug/1	98
84) 1,2,4-Trimethylbenzene	13.482	105	56525	5.176	ug/1	100
85) sec-Butylbenzene	13.618	105	74711	5.161	ug/1	99
86) p-Isopropyltoluene	13.729	119	62016	5.182	ug/1	99
87) 1,3-Dichlorobenzene	13.735	146	30949	5.200	ug/1	98
88) 1,4-Dichlorobenzene	13.812	146	32342	5.330	ug/1	96
89) n-Butylbenzene	14.053	91	59601	5.142	ug/1	99
90) Hexachloroethane	14.335	117	9725	4.794	ug/1	95
91) 1,2-Dichlorobenzene	14.106	146	29897	5.228	ug/1	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5731	5.363	ug/1	93
93) 1,2,4-Trichlorobenzene	15.394	180	18778	5.143	ug/1	99
94) Hexachlorobutadiene	15.500	225	7072	5.199	ug/1	99
95) Naphthalene	15.641	128	67481	4.965	ug/1	99
96) 1,2,3-Trichlorobenzene	15.841	180	18262	5.034	ug/1	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086863.D  
 Acq On : 06 Jun 2025 13:17  
 Operator : JC\MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC005

Quant Time: Jun 07 01:55:22 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086863.D  
 Acq On : 06 Jun 2025 13:17  
 Operator : JC\MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1


Quant Time: Jun 07 01:55:22 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC005

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005




Tgt Ion:168 Resp: 22970  
Ion Ratio Lower Upper  
168 100  
99 59.2 49.1 73.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#2  
Dichlorodifluoromethane  
Concen: 4.452 ug/l  
RT: 2.154 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17



Tgt Ion: 85 Resp: 10196  
Ion Ratio Lower Upper  
85 100  
87 28.6 16.2 48.6









#3  
Chloromethane  
Concen: 5.080 ug/l  
RT: 2.395 min Scan# 7  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#4  
Vinyl Chloride  
Concen: 5.043 ug/l  
RT: 2.554 min Scan# 102  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 62 Resp: 15384  
Ion Ratio Lower Upper  
62 100  
64 34.3 26.0 39.0





#5

Bromomethane

Concen: 5.041 ug/l

RT: 2.989 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005



Tgt Ion: 94 Resp: 8600

Ion Ratio Lower Upper







94 100




96 100.6 77.9 116.9

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#6  
Chloroethane  
Concen: 5.171 ug/l  
RT: 3.154 min Scan# 204  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17Tgt Ion: 64 Resp: 10190  
Ion Ratio Lower Upper  
64 100  
66 28.8 24.6 36.8






#7  
Trichlorofluoromethane  
Concen: 5.203 ug/l  
RT: 3.518 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005



### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#8  
Diethyl Ether  
Concen: 5.252 ug/l  
RT: 3.971 min Scan# 343  
Delta R.T. -0.012 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 74 Resp: 9121  
Ion Ratio Lower Upper  
74 100  
45 95.8 45.5 136.5








#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 5.205 ug/l  
RT: 4.401 min Scan# 415  
Delta R.T. 0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17



Tgt Ion:101 Resp: 13034  
Ion Ratio Lower Upper  
101 100  
85 43.3 33.8 50.8  
151 70.8 58.7 88.1

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#10  
Methyl Iodide  
Concen: 5.092 ug/l  
RT: 4.606 min Scan# 451  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17



Tgt Ion:142 Resp: 16530  
Ion Ratio Lower Upper  
142 100  
127 39.8 30.9 46.3  
141 14.5 10.4 15.6






#11

Tert butyl alcohol  
Concen: 26.363 ug/l  
RT: 5.542 min Scan# 6  
Delta R.T. -0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#12


1,1-Dichloroethene  
Concen: 5.323 ug/l  
RT: 4.359 min Scan# 409  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 96 Resp: 13610  
Ion Ratio Lower Upper  
96 100  
61 148.1 125.8 188.6  
98 67.5 52.5 78.7






Abundance Scan 381 (4.195 min): VN086863.D\data.ms



Abundance Scan 381 (4.195 min): VN086863.D\data.ms (-33)



#13

Acrolein

Concen: 31.804 ug/l

RT: 4.195 min Scan# 3

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 840

Ion Ratio Lower Upper

56 100


55 70.2 55.6 83.4

Manual Integrations


APPROVED

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 525 (5.042 min): VN086863.D\data.ms



Abundance Scan 525 (5.042 min): VN086863.D\data.ms (-47)



#14

Allyl chloride

Concen: 4.935 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion: 41 Resp: 20925


Ion Ratio Lower Upper

41 100


39 63.2 51.3 76.9

76 39.0 31.4 47.0






Abundance Scan 643 (5.736 min): VN086863.D\data.ms



Abundance Scan 643 (5.736 min): VN086863.D\data.ms (-59)



#15

Acrylonitrile

Concen: 25.545 ug/l

RT: 5.736 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005

Tgt Ion: 53 Resp: 49820

Ion Ratio Lower Upper

53 100


52 82.4 65.0 97.6

51 32.9 27.2 40.8


**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 424 (4.448 min): VN086863.D\data.ms



Abundance Scan 424 (4.448 min): VN086863.D\data.ms (-37)



#16

Acetone

Concen: 25.741 ug/l

RT: 4.448 min Scan# 424

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17




Tgt Ion: 43 Resp: 41981

Ion Ratio Lower Upper

43 100

58 34.1 28.0 42.0





#17

Carbon Disulfide

Concen: 5.267 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 5.070 ug/l


RT: 5.048 min Scan# 526


Delta R.T. 0.006 min

Lab File: VN086863.D





Acq: 06 Jun 2025 13:17

Tgt Ion: 43 Resp: 24096  
 Ion Ratio Lower Upper  
 43 100  
 74 24.4 20.7 31.1





Abundance Scan 656 (5.812 min): VN086863.D\data.ms



#19


Methyl tert-butyl Ether  
Concen: 5.058 ug/l

RT: 5.812 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#20


Methylene Chloride  
Concen: 5.178 ug/l  
RT: 5.300 min Scan# 569  
Delta R.T. 0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 84 Resp: 15809  
Ion Ratio Lower Upper  
84 100  
49 106.8 90.5 135.7  
51 36.6 28.5 42.7  
86 60.4 48.1 72.1






Abundance Scan 655 (5.806 min): VN086863.D\data.ms



Abundance Scan 655 (5.806 min): VN086863.D\data.ms (-60)



#21

trans-1,2-Dichloroethene

Concen: 5.442 ug/l

RT: 5.806 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC005


**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 805 (6.689 min): VN086863.D\data.ms



Abundance Scan 805 (6.689 min): VN086863.D\data.ms (-75)



#22

Diisopropyl ether

Concen: 5.191 ug/l

RT: 6.689 min Scan# 805

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion: 45 Resp: 46396


Ion Ratio Lower Upper




45 100

43 53.2 41.9 62.9

87 27.9 22.4 33.6

59 12.6 9.8 14.6





#23

**Vinyl Acetate**

Concen: 25.735 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

**1,1-Dichloroethane**

Concen: 5.147 ug/l

RT: 6.583 min Scan# 787

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion: 63 Resp: 26473




Ion Ratio Lower Upper

63 100

98 6.4 3.6 10.8

100 5.1 2.6 7.8





#25

2-Butanone

Concen: 25.888 ug/l

RT: 7.494 min Scan# 9

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 4.571 ug/l

RT: 7.500 min Scan# 943

Delta R.T. 0.000 min

Lab File: VN086863.D



Acq: 06 Jun 2025 13:17




Tgt Ion: 77 Resp: 18288

Ion Ratio Lower Upper

77 100

97 23.3 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 5.172 ug/l

RT: 7.500 min Scan# 9

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 5.124 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion: 49 Resp: 12958


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 75.2 62.2 93.2





#29

Tetrahydrofuran

Concen: 25.912 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005



Tgt Ion: 42 Resp: 44745

Ion Ratio Lower Upper

42 100

72 48.6 37.8 56.6




71 43.5 35.6 53.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 5.149 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086863.D




Acq: 06 Jun 2025 13:17


Tgt Ion: 83 Resp: 26450

Ion Ratio Lower Upper

83 100

85 69.9 52.9 79.3





#31

Cyclohexane

Concen: 6.002 ug/l

RT: 8.271 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005



Tgt Ion: 56 Resp: 2993

Ion Ratio Lower Upper

56 100







69 33.2 25.4 38.0


84 81.1 68.8 103.2

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025

#32  
1,1,1-Trichloroethane  
Concen: 5.232 ug/l  
RT: 8.177 min Scan# 1058  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17Tgt Ion: 97 Resp: 22857  
Ion Ratio Lower Upper  
97 100  
99 0.0 52.0 78.0#  
61 42.9 35.9 53.9



#33  
1,2-Dichloroethane-d4  
Concen: 5.468 ug/l  
RT: 8.588 min Scan# 1128  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005




Tgt Ion: 65 Resp: 1681  
Ion Ratio Lower Upper  
65 100  
67 54.4 0.0 105.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17



Tgt Ion:114 Resp: 423980  
Ion Ratio Lower Upper  
114 100  
63 20.8 0.0 39.6  
88 15.9 0.0 30.2





#35

Dibromofluoromethane

Concen: 5.121 ug/l


RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17




Instrument : MSVOA\_N  
 ClientSampleId : VSTDICC005




Tgt	Ion:	113	Resp:	1286
Ion	Ratio	Lower	Upper	
113	100			
111	106.5	84.2	126.2	
192	18.0	14.2	21.4	

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025









#36  
 1,1-Dichloropropene  
 Concen: 5.184 ug/l  
 RT: 8.377 min Scan# 1092  
 Delta R.T. 0.000 min  
 Lab File: VN086863.D  
 Acq: 06 Jun 2025 13:17



Tgt	Ion:	75	Resp:	19410
Ion	Ratio	Lower	Upper	
75	100			
110	37.4	17.6	52.9	
77	31.7	25.3	37.9	





#37

**Ethyl Acetate**

Concen: 5.470 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 2607

Ion Ratio Lower Upper

43 100


61 14.1 12.1 18.1

70 10.9 9.8 14.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

**Carbon Tetrachloride**

Concen: 5.184 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion:117 Resp: 19056


Ion Ratio Lower Upper

117 100

119 93.8 75.0 112.6

121 28.5 24.9 37.3





#39

Methylcyclohexane

Concen: 5.333 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005



Tgt Ion: 83 Resp: 2735

Ion Ratio Lower Upper

83 100



55 73.6 61.0 91.6

98 42.6 38.9 58.3

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 5.199 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17




Tgt Ion: 78 Resp: 63646

Ion Ratio Lower Upper

78 100

77 22.6 18.7 28.1





#41

Methacrylonitrile

Concen: 5.050 ug/l

RT: 7.783 min Scan# 991

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#42

1,2-Dichloroethane

Concen: 5.205 ug/l

RT: 8.683 min Scan# 1144

Delta R.T. 0.006 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17


Tgt Ion: 62 Resp: 19329

Ion Ratio Lower Upper

62 100

98 9.9 0.0 21.2





#43

Isopropyl Acetate


Concen: 5.264 ug/l

RT: 8.700 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005

Tgt Ion: 43 Resp: 40263

Ion Ratio Lower Upper

43 100

61 26.2 22.1 33.1




87 14.1 11.8 17.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene


Concen: 5.251 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17






Tgt Ion:130 Resp: 15245

Ion Ratio Lower Upper

130 100

95 101.6 0.0 196.2





#45

1,2-Dichloropropane

Concen: 5.248 ug/l


RT: 9.624 min Scan# 1

Delta R.T. -0.006 min


Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Abundance Scan 1304 (9.624 min): VN086863.D\data.ms



Abundance Scan 1304 (9.624 min): VN086863.D\data.ms (-1)



Tgt Ion: 63 Resp: 1563

Ion Ratio Lower Upper





63 100




65 32.7 25.5 38.3

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#46  
Dibromomethane  
Concen: 5.176 ug/l  
RT: 9.718 min Scan# 1320  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17Tgt Ion: 93 Resp: 10241  
Ion Ratio Lower Upper  
93 100  
95 84.9 67.1 100.7  
174 93.4 74.2 111.4



#47

Bromodichloromethane

Concen: 5.039 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 4.996 ug/l


RT: 9.688 min Scan# 1315




Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion: 41 Resp: 17591  
 Ion Ratio Lower Upper  
 41 100  
 69 85.2 74.6 112.0  
 39 52.3 40.2 60.4





#49

1,4-Dioxane

Concen: 95.874 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 5.308 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17


Tgt Ion: 98 Resp: 52794

Ion Ratio Lower Upper


98 100

100 68.1 53.4 80.0






Abundance Scan 1444 (10.447 min): VN086863.D\data.ms



Abundance Scan 1444 (10.447 min): VN086863.D\data.ms (-)



#51

4-Methyl-2-Pentanone

Concen: 25.283 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC005


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 1475 (10.630 min): VN086863.D\data.ms



Abundance Scan 1475 (10.630 min): VN086863.D\data.ms (-)



#52

Toluene

Concen: 5.178 ug/l

RT: 10.630 min Scan# 1475

Delta R.T. -0.006 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17




Tgt Ion: 92 Resp: 38744

Ion Ratio Lower Upper

92 100

91 169.5 136.5 204.7





#53

t-1,3-Dichloropropene

Concen: 4.896 ug/l

RT: 10.841 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#54

cis-1,3-Dichloropropene

Concen: 5.026 ug/l

RT: 10.318 min Scan# 1422

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion: 75 Resp: 24477




Ion Ratio Lower Upper

75 100

77 34.5 26.2 39.4

39 46.5 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 5.234 ug/l

RT: 11.018 min Scan# 15063

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#56

Ethyl methacrylate

Concen: 4.774 ug/l


RT: 10.888 min Scan# 1519




Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt	Ion	97	15063
Ion	Ratio	Lower	Upper
97	100		
83	82.5	69.7	104.5
85	53.5	45.1	67.7
99	62.6	50.6	75.8





#57

1,3-Dichloropropane

Concen: 5.096 ug/l

RT: 11.165 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 23.664 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17


Tgt Ion: 63 Resp: 64722

Ion Ratio Lower Upper


63 100

106 27.9 23.0 34.6









Abundance Scan 1574 (11.212 min): VN086863.D\data.ms



Abundance Scan 1574 (11.212 min): VN086863.D\data.ms (-)



Abundance Scan 1599 (11.359 min): VN086863.D\data.ms (-)



#59

2-Hexanone

Concen: 20.188 ug/l

RT: 11.212 min Scan# 1

Delta R.T. 0.012 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 59884

Ion Ratio Lower Upper


43 100

58 60.5 30.4 91.3

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#60

Dibromochloromethane

Concen: 4.962 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17




Tgt Ion:129 Resp: 14884

Ion Ratio Lower Upper

129 100

127 75.9 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 5.141 ug/l  
RT: 11.471 min Scan# 1517  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#62  
4-Bromofluorobenzene  
Concen: 5.059 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 95 Resp: 18697  
Ion Ratio Lower Upper  
95 100  
174 75.5 0.0 141.8  
176 70.2 0.0 132.6







Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument: MSVOA\_N  
ClientSampleId: VSTDICC005


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carbone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



Tetrachloroethene  
Concen: 5.236 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion:164 Resp: 12322  
Ion Ratio Lower Upper  
164 100  
166 117.9 101.8 152.8  
129 96.6 83.0 124.4  
131 88.0 78.2 117.2







#65  
Chlorobenzene  
Concen: 5.145 ug/l  
RT: 11.894 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 5.271 ug/l  
RT: 11.965 min Scan# 1702  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion:131 Resp: 13897  
Ion Ratio Lower Upper  
131 100  
133 91.7 48.3 144.8  
119 69.1 34.0 101.8





#67

Ethyl Benzene

Concen: 5.199 ug/l

RT: 11.965 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#68

m/p-Xylenes

Concen: 10.330 ug/l


RT: 12.071 min Scan# 1720


Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:106 Resp: 55859  
 Ion Ratio Lower Upper  
 106 100  
 91 199.0 159.4 239.0







#69  
o-Xylene  
Concen: 5.021 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#70  
Styrene  
Concen: 5.008 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion:104 Resp: 44380  
Ion Ratio Lower Upper  
104 100  
78 50.4 38.5 57.7  
103 54.4 43.1 64.7





#71

Bromoform

Concen: 5.063 ug/l

RT: 12.582 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion:152 Resp: 181065




Ion Ratio Lower Upper

152 100

115 61.2 30.1 90.5

150 157.8 0.0 345.0





#73

Isopropylbenzene

Concen: 5.146 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#74

N-amyl acetate

Concen: 4.590 ug/l

RT: 12.559 min Scan# 1803

Delta R.T. 0.047 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion: 43 Resp: 20096


Ion Ratio Lower Upper




43 100

70 0.0 37.6 56.4#

55 0.0 19.0 28.6#

61 10.0 20.8 31.2#









#75  
1,1,2,2-Tetrachloroethane  
Concen: 5.262 ug/l  
RT: 12.935 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#76  
1,2,3-Trichloropropane  
Concen: 4.986 ug/l m  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 75 Resp: 21527  
Ion Ratio Lower Upper  
75 100  
77 179.7 79.3 237.9





#77

Bromobenzene

Concen: 5.119 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#78

n-propylbenzene

Concen: 5.226 ug/l


RT: 13.035 min Scan# 1884




Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt	Ion:	91	Resp:	83781
Ion	Ratio	Lower	Upper	
91	100			
120	22.1	11.4	34.2	





#79

2-Chlorotoluene

Concen: 5.128 ug/l

RT: 13.124 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 5.139 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17




Tgt Ion:105 Resp: 55961

Ion Ratio Lower Upper

105 100

120 49.8 24.9 74.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 4.939 ug/l  
RT: 12.741 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC005


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#82  
4-Chlorotoluene  
Concen: 5.132 ug/l  
RT: 13.223 min Scan# 1916  
Delta R.T. -0.000 min  
Lab File: VN086863.D  
Acq: 06 Jun 2025 13:17

Tgt Ion: 91 Resp: 49924  
Ion Ratio Lower Upper  
91 100  
126 33.3 16.7 50.1





#83

tert-Butylbenzene

Concen: 5.103 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 5.176 ug/l


RT: 13.482 min Scan# 1960


Delta R.T. 0.000 min

Lab File: VN086863.D


Acq: 06 Jun 2025 13:17

Tgt	Ion:105	Resp:	56525
Ion Ratio	Lower	Upper	
105	100		
120	46.6	23.2	69.6






Abundance Scan 1983 (13.618 min): VN086863.D\data.ms (-)



Abundance Scan 1983 (13.618 min): VN086863.D\data.ms (-)



#85

sec-Butylbenzene

Concen: 5.161 ug/l

RT: 13.618 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC005


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 2002 (13.729 min): VN086863.D\data.ms (-)



Abundance Scan 2002 (13.729 min): VN086863.D\data.ms (-)



#86

p-Isopropyltoluene

Concen: 5.182 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion:119 Resp: 62016


Ion Ratio Lower Upper

119 100

134 26.4 13.5 40.4

91 22.9 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 5.200 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 5.330 ug/l

RT: 13.812 min Scan# 2016


Delta R.T. -0.000 min




Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:146 Resp: 32342

Ion	Ratio	Lower	Upper
146	100		
111	46.1	19.9	59.7
148	63.3	31.8	95.4





#89

n-Butylbenzene

Concen: 5.142 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 4.794 ug/l


RT: 14.335 min Scan# 2105




Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt	Ion:117	Resp:	9725
Ion	Ratio	Lower	Upper
117	100		
201	70.6	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 5.228 ug/l

RT: 14.106 min Scan# 2106

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 5.363 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion: 75 Resp: 5731







Ion Ratio Lower Upper

75 100

155 76.0 41.1 123.4

157 98.8 52.7 158.1





#93

1,2,4-Trichlorobenzene

Concen: 5.143 ug/l

RT: 15.394 min Scan# 2

Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument :

MSVOA\_N


ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 5.199 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17


Tgt Ion:225 Resp: 7072




Ion Ratio Lower Upper

225 100

223 66.4 33.5 100.4

227 61.7 31.1 93.5





#95

Naphthalene

Concen: 4.965 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.000 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC005

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 5.034 ug/l

RT: 15.841 min Scan# 2361


Delta R.T. 0.006 min

Lab File: VN086863.D

Acq: 06 Jun 2025 13:17

Tgt Ion:180 Resp: 18262

Ion	Ratio	Lower	Upper
180	100		
182	90.9	47.0	141.0
145	35.3	16.6	49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086864.D  
 Acq On : 06 Jun 2025 13:40  
 Operator : JC\MD  
 Sample : VSTDICC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC020

Quant Time: Jun 07 01:56:20 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.235	168	224006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	418154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	363172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	181178	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	63390	21.136	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 42.280%	#	
35) Dibromofluoromethane	8.177	113	51797	20.902	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 41.800%	#	
50) Toluene-d8	10.571	98	201268	20.516	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 41.040%	#	
62) 4-Bromofluorobenzene	12.853	95	74622	20.474	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 40.940%	#	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.159	85	48254	21.605	ug/l	92
3) Chloromethane	2.401	50	57778	20.033	ug/l	93
4) Vinyl Chloride	2.559	62	61254	20.589	ug/l	99
5) Bromomethane	3.006	94	34022	20.435	ug/l	95
6) Chloroethane	3.159	64	39625	20.619	ug/l	99
7) Trichlorofluoromethane	3.524	101	80978	20.832	ug/l	94
8) Diethyl Ether	3.983	74	34288	20.245	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.395	101	50458	20.661	ug/l	98
10) Methyl Iodide	4.612	142	65360	20.646	ug/l	98
11) Tert butyl alcohol	5.547	59	87102	107.031	ug/l	99
12) 1,1-Dichloroethene	4.359	96	50485	20.249	ug/l	96
13) Acrolein	4.200	56	22866	88.767	ug/l	97
14) Allyl chloride	5.047	41	82924	20.055	ug/l	98
15) Acrylonitrile	5.742	53	199498	104.881	ug/l	99
16) Acetone	4.447	43	163979	103.099	ug/l	98
17) Carbon Disulfide	4.736	76	138195	20.038	ug/l	97
18) Methyl Acetate	5.047	43	96547	20.830	ug/l	99
19) Methyl tert-butyl Ether	5.818	73	183780	20.358	ug/l	98
20) Methylene Chloride	5.294	84	57611	19.348	ug/l	98
21) trans-1,2-Dichloroethene	5.806	96	55643	20.059	ug/l	94
22) Diisopropyl ether	6.689	45	182472	20.935	ug/l	98
23) Vinyl Acetate	6.624	43	768556	104.365	ug/l	99
24) 1,1-Dichloroethane	6.583	63	103578	20.650	ug/l	99
25) 2-Butanone	7.494	43	270477	104.621	ug/l	100
26) 2,2-Dichloropropane	7.506	77	69692	17.862	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	68292	20.585	ug/l	98
28) Bromochloromethane	7.824	49	55196	22.381	ug/l	99
29) Tetrahydrofuran	7.853	42	178571	106.031	ug/l	99
30) Chloroform	7.977	83	102615	20.485	ug/l	93
31) Cyclohexane	8.271	56	100026	20.563	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	86810	20.376	ug/l	98
36) 1,1-Dichloropropene	8.383	75	73252	19.838	ug/l	100
37) Ethyl Acetate	7.571	43	96467	20.522	ug/l	98
38) Carbon Tetrachloride	8.377	117	72628	20.034	ug/l	97
39) Methylcyclohexane	9.606	83	98425	19.455	ug/l	99
40) Benzene	8.612	78	241444	19.996	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086864.D  
 Acq On : 06 Jun 2025 13:40  
 Operator : JC\MD  
 Sample : VSTDICC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC020

Quant Time: Jun 07 01:56:20 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	53244	20.169	ug/1	98
42) 1,2-Dichloroethane	8.677	62	74238	20.270	ug/1	100
43) Isopropyl Acetate	8.694	43	151601	20.094	ug/1	99
44) Trichloroethene	9.359	130	57110	19.946	ug/1	96
45) 1,2-Dichloropropane	9.630	63	59293	20.184	ug/1	97
46) Dibromomethane	9.712	93	40290	20.648	ug/1	98
47) Bromodichloromethane	9.894	83	80353	20.015	ug/1	98
48) Methyl methacrylate	9.682	41	70457	20.290	ug/1	97
49) 1,4-Dioxane	9.706	88	28231	446.886	ug/1 #	98
51) 4-Methyl-2-Pentanone	10.447	43	481741	106.066	ug/1	99
52) Toluene	10.635	92	148000	20.056	ug/1	98
53) t-1,3-Dichloropropene	10.841	75	87683	19.532	ug/1	99
54) cis-1,3-Dichloropropene	10.318	75	94315	19.636	ug/1	99
55) 1,1,2-Trichloroethane	11.018	97	57195	20.143	ug/1	99
56) Ethyl methacrylate	10.882	69	94829	20.968	ug/1	98
57) 1,3-Dichloropropane	11.165	76	98205	19.938	ug/1	98
58) 2-Chloroethyl Vinyl ether	10.165	63	278986	103.426	ug/1	99
59) 2-Hexanone	11.206	43	303686	103.803	ug/1	95
60) Dibromochloromethane	11.359	129	59961	20.268	ug/1	99
61) 1,2-Dibromoethane	11.471	107	59139	20.320	ug/1	97
64) Tetrachloroethene	11.106	164	45374	19.742	ug/1	96
65) Chlorobenzene	11.894	112	160803	20.076	ug/1	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	52069	20.223	ug/1	99
67) Ethyl Benzene	11.965	91	277087	20.083	ug/1	98
68) m/p-Xylenes	12.071	106	215378	40.780	ug/1	100
69) o-Xylene	12.400	106	102050	20.175	ug/1	98
70) Styrene	12.412	104	178061	20.572	ug/1	99
71) Bromoform	12.576	173	40070	21.004	ug/1 #	99
73) Isopropylbenzene	12.694	105	264442	20.035	ug/1	100
74) N-amyl acetate	12.523	43	86003	19.633	ug/1	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	92264	20.634	ug/1	99
76) 1,2,3-Trichloropropane	12.994	75	81846m	18.945	ug/1	
77) Bromobenzene	12.982	156	60877	20.111	ug/1	98
78) n-propylbenzene	13.035	91	322436	20.102	ug/1	100
79) 2-Chlorotoluene	13.123	91	193774	20.144	ug/1	100
80) 1,3,5-Trimethylbenzene	13.170	105	219937	20.183	ug/1	100
81) trans-1,4-Dichloro-2-b...	12.735	75	37808	20.209	ug/1	92
82) 4-Chlorotoluene	13.223	91	193544	19.885	ug/1	99
83) tert-Butylbenzene	13.435	119	198303	19.880	ug/1	99
84) 1,2,4-Trimethylbenzene	13.482	105	221990	20.315	ug/1	100
85) sec-Butylbenzene	13.617	105	292563	20.196	ug/1	100
86) p-Isopropyltoluene	13.729	119	240682	20.099	ug/1	100
87) 1,3-Dichlorobenzene	13.735	146	120099	20.166	ug/1	99
88) 1,4-Dichlorobenzene	13.812	146	120102	19.780	ug/1	97
89) n-Butylbenzene	14.053	91	227536	19.617	ug/1	100
90) Hexachloroethane	14.335	117	41202	20.299	ug/1	100
91) 1,2-Dichlorobenzene	14.106	146	115645	20.212	ug/1	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	21025	19.661	ug/1	98
93) 1,2,4-Trichlorobenzene	15.388	180	71826	19.659	ug/1	98
94) Hexachlorobutadiene	15.500	225	27887	20.486	ug/1	97
95) Naphthalene	15.641	128	272584	20.043	ug/1	100
96) 1,2,3-Trichlorobenzene	15.841	180	71929	19.815	ug/1	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086864.D  
 Acq On : 06 Jun 2025 13:40  
 Operator : JC\MD  
 Sample : VSTDICC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC020

Quant Time: Jun 07 01:56:20 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086864.D  
 Acq On : 06 Jun 2025 13:40  
 Operator : JC\MD  
 Sample : VSTDICC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1




Quant Time: Jun 07 01:56:20 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**VSTDICC020**

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.235 min Scan# 1

Delta R.T. -0.001 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#2

Dichlorodifluoromethane


Concen: 21.605 ug/l

RT: 2.159 min Scan# 35

Delta R.T. 0.006 min



Lab File: VN086864.D







Acq: 06 Jun 2025 13:40



Tgt Ion: 85 Resp: 48254

Ion	Ratio	Lower	Upper
85	100		
87	37.1	16.2	48.6





#3

Chloromethane

Concen: 20.033 ug/l

RT: 2.401 min Scan# 7

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#4

Vinyl Chloride

Concen: 20.589 ug/l


RT: 2.559 min Scan# 103


Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion: 62 Resp: 61254  
 Ion Ratio Lower Upper  
 62 100  
 64 31.7 26.0 39.0





#5  
Bromomethane  
Concen: 20.435 ug/l  
RT: 3.006 min Scan# 1  
Delta R.T. 0.012 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40



Instrument : MSVOA\_N  
ClientSampleId : VSTDICC020




Tgt Ion: 94 Resp: 3402  
Ion Ratio Lower Upper  
94 100  
96 92.9 77.9 116.9

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#6  
Chloroethane  
Concen: 20.619 ug/l  
RT: 3.159 min Scan# 205  
Delta R.T. 0.006 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40




Tgt Ion: 64 Resp: 39625  
Ion Ratio Lower Upper  
64 100  
66 30.4 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 20.832 ug/l  
RT: 3.524 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC020




Tgt Ion: 101 Resp: 80973  
Ion Ratio Lower Upper  
101 100  
103 60.4 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#8  
Diethyl Ether  
Concen: 20.245 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40




Tgt Ion: 74 Resp: 34288  
Ion Ratio Lower Upper  
74 100  
45 92.9 45.5 136.5









#9  
 1,1,2-Trichlorotrifluoroethane  
 Concen: 20.661 ug/l  
 RT: 4.395 min Scan# 4  
 Delta R.T. -0.000 min  
 Lab File: VN086864.D  
 Acq: 06 Jun 2025 13:40  
 Instrument: MSVOA\_N  
 ClientSampleId: VSTDICC020





Tgt Ion:101 Resp: 50453  
 Ion Ratio Lower Upper  
 101 100  
 85 41.7 33.8 50.8  
 151 71.3 58.7 88.1  
**Manual Integrations APPROVED**  
 Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025






#10  
 Methyl Iodide  
 Concen: 20.646 ug/l  
 RT: 4.612 min Scan# 452  
 Delta R.T. 0.006 min  
 Lab File: VN086864.D  
 Acq: 06 Jun 2025 13:40



Tgt Ion:142 Resp: 65360  
 Ion Ratio Lower Upper  
 142 100  
 127 39.2 30.9 46.3  
 141 14.3 10.4 15.6





#11

Tert butyl alcohol  
Concen: 107.031 ug/l  
RT: 5.547 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025




#12


1,1-Dichloroethene  
Concen: 20.249 ug/l  
RT: 4.359 min Scan# 409  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Tgt Ion: 96 Resp: 50485  
Ion Ratio Lower Upper  
96 100  
61 153.1 125.8 188.6  
98 61.0 52.5 78.7





Abundance Scan 382 (4.200 min): VN086864.D\data.ms



#13

Acrolein

Concen: 88.767 ug/l

RT: 4.200 min Scan# 3

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 22860

Ion Ratio Lower Upper

56 100


55 71.6 55.6 83.4

Manual Integrations



APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 526 (5.047 min): VN086864.D\data.ms



#14

Allyl chloride

Concen: 20.055 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion: 41 Resp: 82924




Ion Ratio Lower Upper

41 100

39 62.6 51.3 76.9

76 38.1 31.4 47.0





#15

Acrylonitrile

Concen: 104.881 ug/l

RT: 5.742 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 103.099 ug/l


RT: 4.447 min Scan# 424




Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion: 43 Resp: 163979  
 Ion Ratio Lower Upper  
 43 100  
 58 35.9 28.0 42.0





#17

Carbon Disulfide

Concen: 20.038 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 20.830 ug/l


RT: 5.047 min Scan# 526


Delta R.T. 0.006 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40

Tgt Ion: 43 Resp: 96547  
 Ion Ratio Lower Upper  
 43 100  
 74 26.2 20.7 31.1







Abundance Scan 657 (5.818 min): VN086864.D\data.ms




Abundance Scan 657 (5.818 min): VN086864.D\data.ms (-60)



Abundance Scan 568 (5.294 min): VN086864.D\data.ms



Abundance Scan 568 (5.294 min): VN086864.D\data.ms (-51)



#19

Methyl tert-butyl Ether

Concen: 20.358 ug/l

RT: 5.818 min Scan# 6

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC020

Tgt Ion: 73 Resp: 183780

Ion Ratio Lower Upper

73 100


57 22.9 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 19.348 ug/l

RT: 5.294 min Scan# 568

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion: 84 Resp: 57611


Ion Ratio Lower Upper


84 100

49 114.5 90.5 135.7

51 35.9 28.5 42.7

86 63.0 48.1 72.1







#21  
 trans-1,2-Dichloroethene  
 Concen: 20.059 ug/l  
 RT: 5.806 min Scan# 6  
 Delta R.T. -0.000 min  
 Lab File: VN086864.D  
 Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N  
 ClientSampleId : VSTDICC020


**Manual Integrations**  
**APPROVED**







Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#22  
 Diisopropyl ether  
 Concen: 20.935 ug/l  
 RT: 6.689 min Scan# 805  
 Delta R.T. -0.000 min  
 Lab File: VN086864.D  
 Acq: 06 Jun 2025 13:40

Tgt Ion: 45 Resp: 182472  
 Ion Ratio Lower Upper  
 45 100  
 43 50.7 41.9 62.9  
 87 27.0 22.4 33.6  
 59 13.1 9.8 14.6





#23

Vinyl Acetate

Concen: 104.365 ug/l

RT: 6.624 min Scan# 7

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument :

MSVOA\_N


ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

1,1-Dichloroethane

Concen: 20.650 ug/l

RT: 6.583 min Scan# 787


Delta R.T. -0.000 min




Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion: 63 Resp: 103578

Ion	Ratio	Lower	Upper
63	100		
98	7.5	3.6	10.8
100	4.7	2.6	7.8





#25

2-Butanone

Concen: 104.621 ug/l

RT: 7.494 min Scan# 9

Instrument :

Delta R.T. -0.000 min

MSVOA\_N

Lab File: VN086864.D




ClientSampleId : VSTDICC020

Acq: 06 Jun 2025 13:40

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 17.862 ug/l

RT: 7.506 min Scan# 944

Delta R.T. 0.006 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40




Tgt Ion: 77 Resp: 69692

Ion Ratio Lower Upper

77 100

97 23.7 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 20.585 ug/l

RT: 7.500 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 22.381 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion: 49 Resp: 55196




Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 77.1 62.2 93.2





#29

Tetrahydrofuran

Concen: 106.031 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 20.485 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40


Tgt Ion: 83 Resp: 102615

Ion Ratio Lower Upper

83 100

85 60.7 52.9 79.3





#31

Cyclohexane

Concen: 20.563 ug/l

RT: 8.271 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VN086864.D



Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC020



Tgt Ion: 56 Resp: 100020

Ion Ratio Lower Upper

56 100






69 29.0 25.4 38.0


84 85.6 68.8 103.2

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#32  
1,1,1-Trichloroethane  
Concen: 20.376 ug/l  
RT: 8.177 min Scan# 1058  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40Tgt Ion: 97 Resp: 86810  
Ion Ratio Lower Upper  
97 100  
99 63.8 52.0 78.0  
61 43.3 35.9 53.9



#33

1,2-Dichloroethane-d4



Concen: 21.136 ug/l

RT: 8.588 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40



Tgt Ion: 65 Resp: 63390

Ion Ratio Lower Upper






65 100


67 55.3 0.0 105.6

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40Tgt Ion:114 Resp: 418154  
Ion Ratio Lower Upper  
114 100  
63 20.6 0.0 39.6  
88 15.8 0.0 30.2



#35

Dibromofluoromethane


Concen: 20.902 ug/l

RT: 8.177 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40



Tgt Ion: 113 Resp: 51791

Ion Ratio Lower Upper

113 100

111 105.1 84.2 126.2

192 17.7 14.2 21.4

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#36

1,1-Dichloropropene





Concen: 19.838 ug/l







RT: 8.383 min Scan# 1093

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40





#37

**Ethyl Acetate**

Concen: 20.522 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

**Carbon Tetrachloride**

Concen: 20.034 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion:117 Resp: 72628


Ion Ratio Lower Upper

117 100

119 91.3 75.0 112.6

121 28.1 24.9 37.3





#39

Methylcyclohexane

Concen: 19.455 ug/l

RT: 9.606 min Scan# 1


Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N

ClientSampleId : VSTDICC020



Tgt Ion: 83 Resp: 9842

Ion Ratio Lower Upper

83 100



55 75.7 61.0 91.6

98 47.4 38.9 58.3

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 19.996 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086864.D



Acq: 06 Jun 2025 13:40

Tgt Ion: 78 Resp: 241444

Ion Ratio Lower Upper

78 100

77 23.8 18.7 28.1



Tgt Ion: 78 Resp: 241444

Ion Ratio Lower Upper

78 100







77 23.8 18.7 28.1

Tgt Ion: 78 Resp: 241444

Ion Ratio Lower Upper

78 100

77 23.8 18.7 28.1



#41

Methacrylonitrile

Concen: 20.169 ug/l

RT: 7.788 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC020

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

Abundance


30000

20000

10000

0

7.788



#42

1,2-Dichloroethane

Concen: 20.270 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion: 62 Resp: 74238

Ion Ratio Lower Upper

62 100

98 10.6 0.0 21.2

Abundance


30000




20000

10000

0

8.677





#43

Isopropyl Acetate

Concen: 20.094 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#44

Trichloroethene

Concen: 19.946 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40


Tgt Ion:130 Resp: 57110

Ion Ratio Lower Upper

130 100

95 102.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 20.184 ug/l

RT: 9.630 min Scan# 1305

Delta R.T. -0.000 min

Lab File: VN086864.D



Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC020



Tgt Ion: 63 Resp: 5929

Ion Ratio Lower Upper



63 100

65 30.4 25.5 38.3

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane


Concen: 20.648 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40





Tgt Ion: 93 Resp: 40290




Ion Ratio Lower Upper

93 100

95 84.1 67.1 100.7

174 90.1 74.2 111.4





#47

Bromodichloromethane

Concen: 20.015 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40




Instrument: MSVOA\_N

ClientSampleId: VSTDICC020

**Manual Integrations  
APPROVED**

 Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Tgt Ion:	Ion Ratio	Lower	Upper
83	100		
85	62.3	50.9	76.3
127	7.6	6.6	10.0



#48

Methyl methacrylate

Concen: 20.290 ug/l


RT: 9.682 min Scan# 1314

Delta R.T. -0.000 min




Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	Ion Ratio	Lower	Upper
41	100		
69	89.4	74.6	112.0
39	50.3	40.2	60.4



VN086864.D 82N060625W.M



#49

1,4-Dioxane

Concen: 446.886 ug/l

RT: 9.706 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC020


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025



Supervised By :Mahesh Dadoda 06/09/2025




#50  
Toluene-d8  
Concen: 20.516 ug/l  
RT: 10.571 min Scan# 1465  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40




Tgt Ion: 98 Resp: 201268  
Ion Ratio Lower Upper  
98 100  
100 67.8 53.4 80.0






Abundance Scan 1444 (10.447 min): VN086864.D\data.ms



Abundance Scan 1444 (10.447 min): VN086864.D\data.ms (-)



#51

4-Methyl-2-Pentanone

Concen: 106.066 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC020


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 1476 (10.635 min): VN086864.D\data.ms



Abundance Scan 1476 (10.635 min): VN086864.D\data.ms (-)



#52

Toluene

Concen: 20.056 ug/l

RT: 10.635 min Scan# 1476

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40




Tgt Ion: 92 Resp: 148000

Ion Ratio Lower Upper

92 100

91 172.9 136.5 204.7





Abundance Scan 1422 (10.318 min): VN086864.D\data.ms (-)

#53

t-1,3-Dichloropropene

Concen: 19.532 ug/l

RT: 10.841 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 1422 (10.318 min): VN086864.D\data.ms (-)

#54

cis-1,3-Dichloropropene

Concen: 19.636 ug/l


RT: 10.318 min Scan# 1422


Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt	Ion	75	100	Resp:	94315
Ion	Ratio	77	32.5	Lower	26.2
39	41.3	39	32.3	Upper	39.4






#55  
 1,1,2-Trichloroethane  
 Concen: 20.143 ug/l  
 RT: 11.018 min Scan# 1  
 Delta R.T. -0.000 min  
 Lab File: VN086864.D  
 Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N  
 ClientSampleId : VSTDICC020


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#56  
 Ethyl methacrylate  
 Concen: 20.968 ug/l  
 RT: 10.882 min Scan# 1518  
 Delta R.T. -0.000 min  
 Lab File: VN086864.D  
 Acq: 06 Jun 2025 13:40

Tgt Ion: 69 Resp: 94829  
 Ion Ratio Lower Upper  
 69 100  
 41 58.9 47.2 70.8  
 39 25.9 23.0 34.6





#57

1,3-Dichloropropane

Concen: 19.938 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC020

**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025




#58  
2-Chloroethyl Vinyl ether  
Concen: 103.426 ug/l  
RT: 10.165 min Scan# 1396  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Tgt Ion: 63 Resp: 278986  
Ion Ratio Lower Upper  
63 100  
106 28.1 23.0 34.6








Abundance Scan 1573 (11.206 min): VN086864.D\data.ms




Abundance Scan 1573 (11.206 min): VN086864.D\data.ms (-)



Abundance Scan 1599 (11.359 min): VN086864.D\data.ms



Abundance Scan 1599 (11.359 min): VN086864.D\data.ms (-)



#59

2-Hexanone

Concen: 103.803 ug/l

RT: 11.206 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#60

Dibromochloromethane

Concen: 20.268 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40




Tgt Ion:129 Resp: 59961

Ion Ratio Lower Upper

129 100

127 76.3 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 20.320 ug/l  
RT: 11.471 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC020


1 Manual Integrations  
2 APPROVED




3 Reviewed By :John Carlone 06/09/2025  
4 Supervised By :Mahesh Dadoda 06/09/2025



#62  
4-Bromofluorobenzene  
Concen: 20.474 ug/l  
RT: 12.853 min Scan# 1853  
Delta R.T. 0.006 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Tgt Ion: 95 Resp: 74622  
Ion Ratio Lower Upper  
95 100  
174 71.7 0.0 141.8  
176 68.0 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#64

Tetrachloroethene


Concen: 19.742 ug/l


RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40







#65  
Chlorobenzene  
Concen: 20.076 ug/l  
RT: 11.894 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC020


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 20.223 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Tgt Ion:131 Resp: 52069  
Ion Ratio Lower Upper  
131 100  
133 96.3 48.3 144.8  
119 69.6 34.0 101.8





#67

Ethyl Benzene

Concen: 20.083 ug/l

RT: 11.965 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#68

m/p-Xylenes

Concen: 40.780 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40


Tgt Ion:106 Resp: 215378

Ion Ratio Lower Upper

106 100

91 199.8 159.4 239.0







#69  
o-Xylene  
Concen: 20.175 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC020


**Manual Integrations**  
**APPROVED**







Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#70  
Styrene  
Concen: 20.572 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Tgt Ion:104 Resp: 178061  
Ion Ratio Lower Upper  
104 100  
78 48.3 38.5 57.7  
103 54.4 43.1 64.7









#71

Bromoform  
Concen: 21.004 ug/lRT: 12.576 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICC020**Manual Integrations  
APPROVED**Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025

#72

1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2012  
Delta R.T. -0.006 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40Tgt Ion:152 Resp: 181178  
Ion Ratio Lower Upper  
152 100  
115 61.4 30.1 90.5  
150 164.5 0.0 345.0



#73

Isopropylbenzene

Concen: 20.035 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#74

N-amyl acetate

Concen: 19.633 ug/l

RT: 12.523 min Scan# 1797


Delta R.T. 0.012 min


Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion: 43 Resp: 86003

Ion	Ratio	Lower	Upper
43	100		
70	47.7	37.6	56.4
55	26.6	19.0	28.6
61	25.9	20.8	31.2







#75  
1,1,2,2-Tetrachloroethane  
Concen: 20.634 ug/l  
RT: 12.941 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC020


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#76  
1,2,3-Trichloropropane  
Concen: 18.945 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086864.D  
Acq: 06 Jun 2025 13:40

Tgt Ion: 75 Resp: 81846  
Ion Ratio Lower Upper  
75 100  
77 176.5 79.3 237.9





#77

Bromobenzene

Concen: 20.111 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#78

n-propylbenzene

Concen: 20.102 ug/l


RT: 13.035 min Scan# 1884




Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt	Ion:	91	Resp:	322436
Ion	Ratio	Lower	Upper	
91	100			
120	22.9	11.4	34.2	





#79

2-Chlorotoluene

Concen: 20.144 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 20.183 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40




Tgt Ion:105 Resp: 219937

Ion Ratio Lower Upper

105 100

120 49.6 24.9 74.6





#81

trans-1,4-Dichloro-2-butene

Concen: 20.209 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40





Instrument: MSVOA\_N

ClientSampleId: VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#82

4-Chlorotoluene

Concen: 19.885 ug/l


RT: 13.223 min Scan# 1916




Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
	91	100			
	126	33.7	193544	16.7	50.1





#83

tert-Butylbenzene

Concen: 19.880 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 20.315 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086864.D


Acq: 06 Jun 2025 13:40


Tgt Ion:105 Resp: 221990

Ion Ratio Lower Upper



105 100

120 46.1 23.2 69.6






Abundance Scan 1983 (13.617 min): VN086864.D\data.ms (-)



Abundance Scan 2002 (13.729 min): VN086864.D\data.ms (-)



#85

sec-Butylbenzene

Concen: 20.196 ug/l

RT: 13.617 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#86

p-Isopropyltoluene

Concen: 20.099 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion:119 Resp: 240682


Ion Ratio Lower Upper

119 100

134 27.1 13.5 40.4

91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 20.166 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 19.780 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion:146 Resp: 120102




Ion Ratio Lower Upper

146 100

111 42.5 19.9 59.7

148 64.7 31.8 95.4





#89

n-Butylbenzene

Concen: 19.617 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 20.299 ug/l



RT: 14.335 min Scan# 2105




Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt	Ion:117	Resp:	41202
Ion	Ratio	Lower	Upper
117	100		
201	66.2	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 20.212 ug/l

RT: 14.106 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.661 ug/l


RT: 14.717 min Scan# 2170







Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:	Ion Ratio	Resp:	Lower	Upper
75	100	21025		
155	81.8	41.1	123.4	
157	109.0	52.7	158.1	





#93

1,2,4-Trichlorobenzene

Concen: 19.659 ug/l

RT: 15.388 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 20.486 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40


Tgt Ion:225 Resp: 27887




Ion Ratio Lower Upper

225 100

223 63.0 33.5 100.4

227 63.3 31.1 93.5





#95

Naphthalene

Concen: 20.043 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC020

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 19.815 ug/l

RT: 15.841 min Scan# 2361


Delta R.T. 0.006 min

Lab File: VN086864.D

Acq: 06 Jun 2025 13:40

Tgt Ion:180 Resp: 71929

Ion	Ratio	Lower	Upper
180	100		
182	95.3	47.0	141.0
145	33.1	16.6	49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086865.D  
 Acq On : 06 Jun 2025 14:03  
 Operator : JC\MD  
 Sample : VSTDICCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICCC050

Quant Time: Jun 07 01:57:17 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.236	168	207354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	378587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	332766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167532	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	103736	37.366	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	74.740%	
35) Dibromofluoromethane	8.171	113	83058	37.020	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	74.040%#	
50) Toluene-d8	10.571	98	326105	36.716	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	73.440%#	
62) 4-Bromofluorobenzene	12.847	95	122965	37.264	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	74.520%#	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	103958	50.283	ug/l	100
3) Chloromethane	2.395	50	123866	46.397	ug/l	100
4) Vinyl Chloride	2.554	62	132783	48.215	ug/l	100
5) Bromomethane	2.995	94	73974	47.999	ug/l	100
6) Chloroethane	3.154	64	84505	47.504	ug/l	100
7) Trichlorofluoromethane	3.524	101	172860	48.041	ug/l	100
8) Diethyl Ether	3.983	74	75193	47.963	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.395	101	107695	47.640	ug/l	100
10) Methyl Iodide	4.606	142	141678	48.346	ug/l	100
11) Tert butyl alcohol	5.548	59	182315	242.020	ug/l	100
12) 1,1-Dichloroethene	4.359	96	110549	47.900	ug/l	100
13) Acrolein	4.201	56	47026	197.217	ug/l	100
14) Allyl chloride	5.042	41	179313	46.848	ug/l	100
15) Acrylonitrile	5.742	53	421941	239.638	ug/l	100
16) Acetone	4.448	43	333857	226.765	ug/l	100
17) Carbon Disulfide	4.730	76	295753	46.328	ug/l	100
18) Methyl Acetate	5.042	43	204538	47.673	ug/l	100
19) Methyl tert-butyl Ether	5.812	73	400775	47.960	ug/l	100
20) Methylene Chloride	5.295	84	125498	45.531	ug/l	100
21) trans-1,2-Dichloroethene	5.806	96	117660	45.822	ug/l	100
22) Diisopropyl ether	6.689	45	384939	47.711	ug/l	100
23) Vinyl Acetate	6.618	43	1649023	241.910	ug/l	100
24) 1,1-Dichloroethane	6.583	63	220477	47.485	ug/l	100
25) 2-Butanone	7.494	43	571221	238.693	ug/l	100
26) 2,2-Dichloropropane	7.500	77	187567	51.933	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	144911	47.189	ug/l	100
28) Bromochloromethane	7.824	49	96707	42.362	ug/l	100
29) Tetrahydrofuran	7.853	42	371357	238.210	ug/l	100
30) Chloroform	7.977	83	220093	47.467	ug/l	100
31) Cyclohexane	8.271	56	208105	46.217	ug/l	100
32) 1,1,1-Trichloroethane	8.177	97	185502	47.037	ug/l	100
36) 1,1-Dichloropropene	8.377	75	158341	47.362	ug/l	100
37) Ethyl Acetate	7.577	43	203091	47.720	ug/l	100
38) Carbon Tetrachloride	8.371	117	154677	47.126	ug/l	100
39) Methylcyclohexane	9.606	83	215849	47.126	ug/l	100
40) Benzene	8.612	78	509334	46.590	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086865.D  
 Acq On : 06 Jun 2025 14:03  
 Operator : JC\MD  
 Sample : VSTDICCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICCC050

Quant Time: Jun 07 01:57:17 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	114727	48.000	ug/l	100
42) 1,2-Dichloroethane	8.677	62	155713	46.960	ug/l	100
43) Isopropyl Acetate	8.694	43	322741	47.248	ug/l	100
44) Trichloroethene	9.359	130	123847	47.776	ug/l	100
45) 1,2-Dichloropropane	9.630	63	125673	47.251	ug/l	100
46) Dibromomethane	9.718	93	83811	47.442	ug/l	100
47) Bromodichloromethane	9.894	83	173200	47.651	ug/l	100
48) Methyl methacrylate	9.683	41	149019	47.399	ug/l	100
49) 1,4-Dioxane	9.700	88	59060	1032.607	ug/l	#
51) 4-Methyl-2-Pentanone	10.447	43	1017688	247.484	ug/l	100
52) Toluene	10.635	92	316219	47.331	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	197475	48.587	ug/l	100
54) cis-1,3-Dichloropropene	10.318	75	208523	47.950	ug/l	100
55) 1,1,2-Trichloroethane	11.018	97	121969	47.445	ug/l	100
56) Ethyl methacrylate	10.882	69	210467	51.401	ug/l	100
57) 1,3-Dichloropropane	11.165	76	212243	47.593	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	519139	212.570	ug/l	100
59) 2-Hexanone	11.200	43	696567	262.977	ug/l	100
60) Dibromochloromethane	11.359	129	128715	48.056	ug/l	100
61) 1,2-Dibromoethane	11.471	107	124522	47.258	ug/l	100
64) Tetrachloroethene	11.106	164	97679	46.382	ug/l	100
65) Chlorobenzene	11.894	112	340439	46.386	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.965	131	112333	47.615	ug/l	100
67) Ethyl Benzene	11.965	91	597593	47.271	ug/l	100
68) m/p-Xylenes	12.071	106	466620	96.424	ug/l	100
69) o-Xylene	12.400	106	224299	48.395	ug/l	100
70) Styrene	12.412	104	386596	48.745	ug/l	100
71) Bromoform	12.582	173	88127	50.415	ug/l	#
73) Isopropylbenzene	12.694	105	573905	47.023	ug/l	100
74) N-amyl acetate	12.512	43	212064	52.353	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.941	83	197360	47.732	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	196473m	49.183	ug/l	
77) Bromobenzene	12.982	156	132381	47.296	ug/l	100
78) n-propylbenzene	13.035	91	705524	47.567	ug/l	100
79) 2-Chlorotoluene	13.124	91	419986	47.216	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	485499	48.181	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	79732	46.089	ug/l	100
82) 4-Chlorotoluene	13.224	91	424066	47.117	ug/l	100
83) tert-Butylbenzene	13.435	119	438417	47.531	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	488189	48.314	ug/l	100
85) sec-Butylbenzene	13.618	105	641523	47.893	ug/l	100
86) p-Isopropyltoluene	13.729	119	533366	48.169	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	260310	47.269	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	263358	46.907	ug/l	100
89) n-Butylbenzene	14.053	91	512519	47.787	ug/l	100
90) Hexachloroethane	14.335	117	89896	47.896	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	251235	47.486	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.718	75	45535	46.049	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	162374	48.061	ug/l	100
94) Hexachlorobutadiene	15.500	225	60866	48.356	ug/l	100
95) Naphthalene	15.641	128	603088	47.958	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	159113	47.402	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086865.D  
 Acq On : 06 Jun 2025 14:03  
 Operator : JC\MD  
 Sample : VSTDICCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICCC050

Quant Time: Jun 07 01:57:17 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086865.D  
 Acq On : 06 Jun 2025 14:03  
 Operator : JC\MD  
 Sample : VSTDICCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1


Quant Time: Jun 07 01:57:17 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**VSTDICCC050**

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.236 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03




Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050




Tgt Ion:168 Resp: 207354  
Ion Ratio Lower Upper  
168 100  
99 61.4 49.1 73.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#2  
Dichlorodifluoromethane  
Concen: 50.283 ug/l  
RT: 2.154 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03



Tgt Ion: 85 Resp: 103958  
Ion Ratio Lower Upper  
85 100  
87 32.4 16.2 48.6









#3  
 Chloromethane  
 Concen: 46.397 ug/l  
 RT: 2.395 min Scan# 7  
 Delta R.T. 0.000 min  
 Lab File: VN086865.D  
 Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
 ClientSampleId : VSTDICCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#4  
 Vinyl Chloride  
 Concen: 48.215 ug/l  
 RT: 2.554 min Scan# 102  
 Delta R.T. 0.000 min  
 Lab File: VN086865.D  
 Acq: 06 Jun 2025 14:03

Tgt Ion: 62 Resp: 132783  
 Ion Ratio Lower Upper  
 62 100  
 64 32.5 26.0 39.0





#5

Bromomethane

Concen: 47.999 ug/l

RT: 2.995 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#6

Chloroethane

Concen: 47.504 ug/l


RT: 3.154 min Scan# 204


Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03


Tgt Ion: 64 Resp: 84505  
 Ion Ratio Lower Upper  
 64 100  
 66 30.7 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 48.041 ug/l  
RT: 3.524 min Scan# 2  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03




Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050




Tgt Ion:101 Resp: 172860  
Ion Ratio Lower Upper  
101 100  
103 65.2 52.2 78.2

Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#8  
Diethyl Ether  
Concen: 47.963 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03




Tgt Ion: 74 Resp: 75193  
Ion Ratio Lower Upper  
74 100  
45 91.0 45.5 136.5





#9  
 1,1,2-Trichlorotrifluoroethane  
 Concen: 47.640 ug/l  
 RT: 4.395 min Scan# 4  
 Delta R.T. 0.000 min  
 Lab File: VN086865.D  
 Acq: 06 Jun 2025 14:03




Instrument : MSVOA\_N  
 ClientSampleId : VSTDICCC050




Tgt Ion:101 Resp: 10769  
 Ion Ratio Lower Upper  
 101 100  
 85 42.3 33.8 50.8  
 151 73.4 58.7 88.1

Manual Integrations  
**APPROVED**



Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025






#10  
 Methyl Iodide  
 Concen: 48.346 ug/l  
 RT: 4.606 min Scan# 451  
 Delta R.T. 0.000 min  
 Lab File: VN086865.D  
 Acq: 06 Jun 2025 14:03



Tgt Ion:142 Resp: 141678  
 Ion Ratio Lower Upper  
 142 100  
 127 38.6 30.9 46.3  
 141 13.0 10.4 15.6









#11

Tert butyl alcohol  
Concen: 242.020 ug/l  
RT: 5.548 min Scan# 6  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#12


1,1-Dichloroethene  
Concen: 47.900 ug/l  
RT: 4.359 min Scan# 409  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion: 96 Resp: 110549  
Ion Ratio Lower Upper  
96 100  
61 157.2 125.8 188.6  
98 65.6 52.5 78.7






Abundance Scan 382 (4.201 min): VN086865.D\data.ms



Abundance Scan 382 (4.201 min): VN086865.D\data.ms (-33)



#13

Acrolein

Concen: 197.217 ug/l

RT: 4.201 min Scan# 3

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICCC050


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 525 (5.042 min): VN086865.D\data.ms



Abundance Scan 525 (5.042 min): VN086865.D\data.ms (-47)



#14

Allyl chloride

Concen: 46.848 ug/l

RT: 5.042 min Scan# 525

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03


Tgt Ion: 41 Resp: 179313




Ion Ratio Lower Upper

41 100

39 64.1 51.3 76.9

76 39.2 31.4 47.0





#15

Acrylonitrile

Concen: 239.638 ug/l

RT: 5.742 min Scan# 6

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 53 Resp: 42194

Ion Ratio Lower Upper

53 100

52 81.3 65.0 97.6




51 34.0 27.2 40.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 226.765 ug/l

RT: 4.448 min Scan# 424

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion: 43 Resp: 333857

Ion Ratio Lower Upper

43 100

58 35.0 28.0 42.0





#17

Carbon Disulfide

Concen: 46.328 ug/l

RT: 4.730 min Scan# 4

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 47.673 ug/l

RT: 5.042 min Scan# 525

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion: 43 Resp: 204538

Ion Ratio Lower Upper

43 100

74 25.9 20.7 31.1









#19

Methyl tert-butyl Ether  
Concen: 47.960 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#20



Methylene Chloride  
Concen: 45.531 ug/l  
RT: 5.295 min Scan# 568  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion: 84 Resp: 125498  
Ion Ratio Lower Upper  
84 100  
49 113.1 90.5 135.7  
51 35.6 28.5 42.7  
86 60.1 48.1 72.1






Abundance Scan 655 (5.806 min): VN086865.D\data.ms



Abundance Scan 805 (6.689 min): VN086865.D\data.ms



#21

trans-1,2-Dichloroethene

Concen: 45.822 ug/l

RT: 5.806 min Scan# 6

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#22

Diisopropyl ether

Concen: 47.711 ug/l

RT: 6.689 min Scan# 805

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion: 45 Resp: 384939


Ion Ratio Lower Upper




45 100

43 52.4 41.9 62.9

87 28.0 22.4 33.6

59 12.2 9.8 14.6





#23

**Vinyl Acetate**

Concen: 241.910 ug/l

RT: 6.618 min Scan# 7

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

**1,1-Dichloroethane**

Concen: 47.485 ug/l

RT: 6.583 min Scan# 787

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03


Tgt Ion: 63 Resp: 220477




Ion Ratio Lower Upper

63 100

98 7.2 3.6 10.8

100 5.2 2.6 7.8





#25

2-Butanone

Concen: 238.693 ug/l

RT: 7.494 min Scan# 9

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 51.933 ug/l

RT: 7.500 min Scan# 943

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion: 77 Resp: 187567

Ion Ratio Lower Upper

77 100

97 23.1 11.6 34.7









#27

cis-1,2-Dichloroethene  
Concen: 47.189 ug/l  
RT: 7.500 min Scan# 9

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#28

Bromochloromethane  
Concen: 42.362 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion: 49 Resp: 96707  
Ion Ratio Lower Upper  
49 100  
129 1.9 0.0 3.8  
130 77.7 62.2 93.2





#29

Tetrahydrofuran

Concen: 238.210 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 47.467 ug/l


RT: 7.977 min Scan# 1024


Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt	Ion	Resp:	220093
Ion	Ratio	Lower	Upper
83	100		
85	66.1	52.9	79.3





#31

Cyclohexane

Concen: 46.217 ug/l

RT: 8.271 min Scan# 1074

Delta R.T. 0.000 min

Lab File: VN086865.D



Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICCC050



Tgt Ion: 56 Resp: 208100

Ion Ratio Lower Upper

56 100



69 31.7 25.4 38.0

84 86.0 68.8 103.2

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane



Concen: 47.037 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03




Tgt Ion: 97 Resp: 185502


Ion Ratio Lower Upper

97 100

99 65.0 52.0 78.0

61 44.9 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 37.366 ug/l


RT: 8.588 min Scan# 1128

Delta R.T. 0.000 min


Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Abundance Scan 1128 (8.588 min): VN086865.D\data.ms



Abundance Scan 1128 (8.588 min): VN086865.D\data.ms (-1)



Tgt Ion: 65 Resp: 103730

Ion Ratio Lower Upper

65 100



67 52.8 0.0 105.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l


RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Abundance Scan 1216 (9.106 min): VN086865.D\data.ms




Tgt Ion:114 Resp: 378587

Ion Ratio Lower Upper


114 100


63 19.8 0.0 39.6

88 15.1 0.0 30.2



Abundance Scan 1216 (9.106 min): VN086865.D\data.ms (-1)





#35

Dibromofluoromethane

Concen: 37.020 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN086865.D





Acq: 06 Jun 2025 14:03

Instrument :





MSVOA\_N




ClientSampleId :

VSTDICCC050



#36  
1,1-Dichloropropene  
Concen: 47.362 ug/l  
RT: 8.377 min Scan# 1092  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03





#37

**Ethyl Acetate**

Concen: 47.720 ug/l

RT: 7.577 min Scan# 9

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

**Carbon Tetrachloride**

Concen: 47.126 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03



Tgt Ion:117 Resp: 154677


Ion Ratio Lower Upper

117 100

119 93.8 75.0 112.6

121 31.1 24.9 37.3





#39

Methylcyclohexane

Concen: 47.126 ug/l

RT: 9.606 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 215849

Ion Ratio Lower Upper

83 100

55 76.3 61.0 91.6



98 48.6 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 46.590 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion: 78 Resp: 509334

Ion Ratio Lower Upper

78 100

77 23.4 18.7 28.1





#41

Methacrylonitrile

Concen: 48.000 ug/l

RT: 7.789 min Scan# 992

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#42

1,2-Dichloroethane

Concen: 46.960 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03


Tgt Ion: 62 Resp: 155713

Ion Ratio Lower Upper

62 100

98 10.6 0.0 21.2





#43

Isopropyl Acetate



Concen: 47.248 ug/l

RT: 8.694 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03



Tgt Ion: 43 Resp: 32274

Ion Ratio Lower Upper

43 100





61 27.6 22.1 33.1

87 14.7 11.8 17.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025


#44  
Trichloroethene  
Concen: 47.776 ug/l  
RT: 9.359 min Scan# 1259  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03


Tgt Ion:130 Resp: 123847

Ion Ratio Lower Upper

130 100

95 98.1 0.0 196.2





#45

1,2-Dichloropropane

Concen: 47.251 ug/l

RT: 9.630 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#46

Dibromomethane

Concen: 47.442 ug/l

RT: 9.718 min Scan# 1320


Delta R.T. 0.000 min




Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion: 93 Resp: 83811

Ion	Ratio	Lower	Upper
93	100		
95	83.9	67.1	100.7
174	92.8	74.2	111.4





#47

Bromodichloromethane

Concen: 47.651 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 47.399 ug/l

RT: 9.683 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03


Tgt Ion: 41 Resp: 149019


Ion Ratio Lower Upper

41 100

69 93.3 74.6 112.0

39 50.3 40.2 60.4





#49

1,4-Dioxane

Concen: 1032.607 ug/l

RT: 9.700 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D



Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N

ClientSampleId :

VSTDICCC050



Tgt Ion: 88 Resp: 59060

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0



58 72.1 57.7 86.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8


Concen: 36.716 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03






Tgt Ion: 98 Resp: 326105

Ion Ratio Lower Upper


98 100

100 66.7 53.4 80.0








Abundance Scan 1444 (10.447 min): VN086865.D\data.ms




Abundance Scan 1444 (10.447 min): VN086865.D\data.ms (-)



Abundance Scan 1476 (10.635 min): VN086865.D\data.ms



Abundance Scan 1476 (10.635 min): VN086865.D\data.ms (-)



Abundance Scan 1476 (10.635 min): VN086865.D\data.ms (-)

#51

4-Methyl-2-Pentanone

Concen: 247.484 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N


ClientSampleId :

VSTDICCC050

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#52

Toluene

Concen: 47.331 ug/l

RT: 10.635 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion: 92 Resp: 316219

Ion Ratio Lower Upper

92 100

91 170.6 136.5 204.7





Abundance Scan 1422 (10.318 min): VN086865.D\data.ms (-)

m/z-->

#53

t-1,3-Dichloropropene

Concen: 48.587 ug/l

RT: 10.841 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N


ClientSampleId :

VSTDICCC050

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#54

cis-1,3-Dichloropropene

Concen: 47.950 ug/l

RT: 10.318 min Scan# 1422

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03


Tgt Ion: 75 Resp: 208523


Ion Ratio Lower Upper

75 100

77 32.8 26.2 39.4

39 40.4 32.3 48.5







#55  
 1,1,2-Trichloroethane  
 Concen: 47.445 ug/l  
 RT: 11.018 min Scan# 1  
 Delta R.T. 0.000 min  
 Lab File: VN086865.D  
 Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
 ClientSampleId : VSTDICCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#56  
 Ethyl methacrylate  
 Concen: 51.401 ug/l  
 RT: 10.882 min Scan# 1518  
 Delta R.T. 0.000 min  
 Lab File: VN086865.D  
 Acq: 06 Jun 2025 14:03

Tgt Ion: 69 Resp: 210467  
 Ion Ratio Lower Upper  
 69 100  
 41 59.0 47.2 70.8  
 39 28.8 23.0 34.6





#57

1,3-Dichloropropane

Concen: 47.593 ug/l

RT: 11.165 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 212.570 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion: 63 Resp: 519139

Ion Ratio Lower Upper

63 100

106 28.8 23.0 34.6









#59  
2-Hexanone  
Concen: 262.977 ug/l  
RT: 11.200 min Scan# 1  
Instrument : MSVOA\_N  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03  
ClientSampleId : VSTDICCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#60  
Dibromochloromethane  
Concen: 48.056 ug/l  
RT: 11.359 min Scan# 1599  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion:129 Resp: 128715  
Ion Ratio Lower Upper  
129 100  
127 77.2 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 47.258 ug/l  
RT: 11.471 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICCC050


**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#62  
4-Bromofluorobenzene  
Concen: 37.264 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion: 95 Resp: 122965  
Ion Ratio Lower Upper  
95 100  
174 70.9 0.0 141.8  
176 66.3 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03


Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050

Tgt Ion:117 Resp: 332760  
Ion Ratio Lower Upper  
117 100  
82 55.8 44.6 67.0  
119 31.9 25.5 38.3

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025







#65  
Chlorobenzene  
Concen: 46.386 ug/l  
RT: 11.894 min Scan# 1  
Instrument : MSVOA\_N  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03  
ClientSampleId : VSTDICCC050


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#66  
1,1,1,2-Tetrachloroethane  
Concen: 47.615 ug/l  
RT: 11.965 min Scan# 1702  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion:131 Resp: 112333  
Ion Ratio Lower Upper  
131 100  
133 96.5 48.3 144.8  
119 67.9 34.0 101.8






Abundance Scan 1702 (11.965 min): VN086865.D\data.ms (-)



Abundance Scan 1702 (11.965 min): VN086865.D\data.ms (-)






Abundance Scan 1720 (12.071 min): VN086865.D\data.ms (-)



Abundance Scan 1720 (12.071 min): VN086865.D\data.ms (-)









#69  
o-Xylene  
Concen: 48.395 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050


Manual Integrations  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#70  
Styrene  
Concen: 48.745 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion:104 Resp: 386596  
Ion Ratio Lower Upper  
104 100  
78 48.1 38.5 57.7  
103 53.9 43.1 64.7





#71

Bromoform

Concen: 50.415 ug/l

RT: 12.582 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l


RT: 13.794 min Scan# 2013




Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt	Ion:152	Resp:	167532
Ion	Ratio	Lower	Upper
152	100		
115	60.3	30.1	90.5
150	172.5	0.0	345.0





#73

Isopropylbenzene

Concen: 47.023 ug/l

RT: 12.694 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#74

N-amyl acetate

Concen: 52.353 ug/l

RT: 12.512 min Scan# 1795

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion: 43 Resp: 212064


Ion Ratio Lower Upper




43 100

70 47.0 37.6 56.4

55 23.8 19.0 28.6

61 26.0 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 47.732 ug/l  
RT: 12.941 min Scan# 1868  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#76  
1,2,3-Trichloropropane  
Concen: 49.183 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion: 75 Resp: 196473  
Ion Ratio Lower Upper  
75 100  
77 158.6 79.3 237.9





#77

Bromobenzene

Concen: 47.296 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#78

n-propylbenzene

Concen: 47.567 ug/l


RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min




Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt	Ion:	91	Resp:	705524
Ion	Ratio	Lower	Upper	
91	100			
120	22.8	11.4	34.2	



VN086865.D 82N060625W.M



#79

2-Chlorotoluene

Concen: 47.216 ug/l

RT: 13.124 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 48.181 ug/l


RT: 13.171 min Scan# 1907




Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion:105 Resp: 485499  
 Ion Ratio Lower Upper  
 105 100  
 120 49.7 24.9 74.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 46.089 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#82  
4-Chlorotoluene  
Concen: 47.117 ug/l  
RT: 13.224 min Scan# 1916  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion: 91 Resp: 424066  
Ion Ratio Lower Upper  
91 100  
126 33.4 16.7 50.1





#83

tert-Butylbenzene

Concen: 47.531 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N



ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 48.314 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN086865.D


Acq: 06 Jun 2025 14:03




Tgt Ion:105 Resp: 488189

Ion Ratio Lower Upper

105 100

120 46.4 23.2 69.6





#85

sec-Butylbenzene

Concen: 47.893 ug/l

RT: 13.618 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#86

p-Isopropyltoluene

Concen: 48.169 ug/l


RT: 13.729 min Scan# 2002




Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt Ion:119 Resp: 533366  
 Ion Ratio Lower Upper  
 119 100  
 134 26.9 13.5 40.4  
 91 22.9 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 47.269 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 46.907 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03


Tgt Ion:146 Resp: 263358




Ion Ratio Lower Upper

146 100

111 39.8 19.9 59.7

148 63.6 31.8 95.4





#89

n-Butylbenzene

Concen: 47.787 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 47.896 ug/l


RT: 14.335 min Scan# 2105




Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
	117	100			
	201	66.4	89896	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 47.486 ug/l

RT: 14.106 min Scan# 2170

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 46.049 ug/l


RT: 14.718 min Scan# 2170

Delta R.T. 0.000 min




Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Tgt	Ion:	Ion Ratio	Resp:	Lower	Upper
	75	100			
	155	82.3	45535	41.1	123.4
	157	105.4		52.7	158.1



VN086865.D 82N060625W.M



#93

1,2,4-Trichlorobenzene

Concen: 48.061 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN086865.D

Acq: 06 Jun 2025 14:03

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICCC050


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025





#94  
Hexachlorobutadiene  
Concen: 48.356 ug/l  
RT: 15.500 min Scan# 2303  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03



Tgt Ion:225 Resp: 60866  
Ion Ratio Lower Upper  
225 100  
223 66.9 33.5 100.4  
227 62.3 31.1 93.5






#95  
Naphthalene  
Concen: 47.958 ug/l  
RT: 15.641 min Scan# 2  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Instrument : MSVOA\_N  
ClientSampleId : VSTDICCC050


**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#96  
1,2,3-Trichlorobenzene  
Concen: 47.402 ug/l  
RT: 15.835 min Scan# 2360  
Delta R.T. 0.000 min  
Lab File: VN086865.D  
Acq: 06 Jun 2025 14:03

Tgt Ion:180 Resp: 159113  
Ion Ratio Lower Upper  
180 100  
182 94.0 47.0 141.0  
145 33.2 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086866.D  
 Acq On : 06 Jun 2025 14:26  
 Operator : JC\MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC100

Quant Time: Jun 07 01:58:16 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.235	168	181996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	327782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	290313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147098	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	238957	98.065	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125			Recovery =	196.120%#	
35) Dibromofluoromethane	8.177	113	195212	100.495	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124			Recovery =	200.980%#	
50) Toluene-d8	10.570	98	772210	100.419	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113			Recovery =	200.840%#	
62) 4-Bromofluorobenzene	12.847	95	292385	102.339	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121			Recovery =	204.680%#	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.153	85	194799	107.350	ug/l	100
3) Chloromethane	2.395	50	224558	95.833	ug/l	98
4) Vinyl Chloride	2.553	62	245135	101.414	ug/l	98
5) Bromomethane	2.977	94	137967	101.996	ug/l	99
6) Chloroethane	3.147	64	152150	97.447	ug/l	97
7) Trichlorofluoromethane	3.524	101	312394	98.917	ug/l	99
8) Diethyl Ether	3.983	74	139844	101.631	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.389	101	198739	100.163	ug/l	99
10) Methyl Iodide	4.612	142	262771	102.162	ug/l	99
11) Tert butyl alcohol	5.547	59	327987	496.062	ug/l	99
12) 1,1-Dichloroethene	4.359	96	200212	98.838	ug/l	99
13) Acrolein	4.200	56	94171	449.961	ug/l	98
14) Allyl chloride	5.047	41	329412	98.055	ug/l	99
15) Acrylonitrile	5.735	53	770470	498.552	ug/l	99
16) Acetone	4.447	43	608178	470.648	ug/l	98
17) Carbon Disulfide	4.736	76	544511	97.179	ug/l	100
18) Methyl Acetate	5.047	43	381808	101.390	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	735694	100.306	ug/l	100
20) Methylene Chloride	5.294	84	228800	94.575	ug/l	97
21) trans-1,2-Dichloroethene	5.806	96	215177	95.476	ug/l	99
22) Diisopropyl ether	6.688	45	696881	98.409	ug/l	98
23) Vinyl Acetate	6.618	43	2919350	487.936	ug/l	100
24) 1,1-Dichloroethane	6.588	63	404207	99.186	ug/l	99
25) 2-Butanone	7.494	43	1043385	496.743	ug/l	100
26) 2,2-Dichloropropane	7.500	77	331995	104.729	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	265248	98.410	ug/l	99
28) Bromochloromethane	7.824	49	188050	93.851	ug/l	99
29) Tetrahydrofuran	7.847	42	677636	495.240	ug/l	99
30) Chloroform	7.977	83	394927	97.040	ug/l	99
31) Cyclohexane	8.265	56	374933	94.869	ug/l	99
32) 1,1,1-Trichloroethane	8.177	97	336819	97.306	ug/l	99
36) 1,1-Dichloropropene	8.382	75	289907	100.156	ug/l	99
37) Ethyl Acetate	7.571	43	370652	100.591	ug/l	99
38) Carbon Tetrachloride	8.371	117	285325	100.404	ug/l	98
39) Methylcyclohexane	9.606	83	395603	99.758	ug/l	99
40) Benzene	8.612	78	927083	97.946	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086866.D  
 Acq On : 06 Jun 2025 14:26  
 Operator : JC\MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC100

Quant Time: Jun 07 01:58:16 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.794	41	195946	94.688	ug/l	95
42) 1,2-Dichloroethane	8.676	62	282015	98.232	ug/l	100
43) Isopropyl Acetate	8.694	43	579464	97.981	ug/l	99
44) Trichloroethene	9.359	130	222822	99.280	ug/l	97
45) 1,2-Dichloropropane	9.629	63	230454	100.077	ug/l	98
46) Dibromomethane	9.712	93	155530	101.685	ug/l	99
47) Bromodichloromethane	9.894	83	316375	100.533	ug/l	100
48) Methyl methacrylate	9.682	41	275710	101.288	ug/l	99
49) 1,4-Dioxane	9.706	88	104771	2115.744	ug/l #	99
51) 4-Methyl-2-Pentanone	10.447	43	1843274	517.729	ug/l	99
52) Toluene	10.635	92	578978	100.092	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	359200	102.077	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	383030	101.730	ug/l	99
55) 1,1,2-Trichloroethane	11.017	97	219922	98.808	ug/l	98
56) Ethyl methacrylate	10.882	69	389802	109.955	ug/l	99
57) 1,3-Dichloropropane	11.165	76	382510	99.069	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	1113523	526.621	ug/l	100
59) 2-Hexanone	11.200	43	1301539	567.535	ug/l	99
60) Dibromochloromethane	11.359	129	237965	102.616	ug/l	100
61) 1,2-Dibromoethane	11.470	107	232214	101.787	ug/l	98
64) Tetrachloroethene	11.106	164	181568	98.824	ug/l	94
65) Chlorobenzene	11.894	112	632563	98.793	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	206895	100.521	ug/l	100
67) Ethyl Benzene	11.964	91	1110803	100.717	ug/l	99
68) m/p-Xylenes	12.070	106	854557	202.411	ug/l	99
69) o-Xylene	12.400	106	412719	102.070	ug/l	100
70) Styrene	12.412	104	713548	103.127	ug/l	100
71) Bromoform	12.576	173	165781	108.707	ug/l #	99
73) Isopropylbenzene	12.694	105	1065297	99.410	ug/l	100
74) N-amyl acetate	12.506	43	403587	113.475	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	354386	97.616	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	353334m	100.738	ug/l	
77) Bromobenzene	12.982	156	246413	100.266	ug/l	99
78) n-propylbenzene	13.035	91	1301631	99.948	ug/l	100
79) 2-Chlorotoluene	13.123	91	770308	98.630	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	901317	101.873	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	160736	105.820	ug/l	92
82) 4-Chlorotoluene	13.223	91	783657	99.166	ug/l	100
83) tert-Butylbenzene	13.435	119	807528	99.710	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	904794	101.983	ug/l	100
85) sec-Butylbenzene	13.617	105	1182155	100.513	ug/l	99
86) p-Isopropyltoluene	13.729	119	990273	101.857	ug/l	100
87) 1,3-Dichlorobenzene	13.735	146	474162	98.063	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	483003	97.978	ug/l	99
89) n-Butylbenzene	14.053	91	937913	99.598	ug/l	100
90) Hexachloroethane	14.335	117	169772	103.019	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	458049	98.602	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	83243	95.877	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	298870	100.751	ug/l	99
94) Hexachlorobutadiene	15.500	225	112496	101.789	ug/l	98
95) Naphthalene	15.641	128	1129434	102.290	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	294127	99.796	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086866.D  
 Acq On : 06 Jun 2025 14:26  
 Operator : JC\MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC100

Quant Time: Jun 07 01:58:16 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
(#)						

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086866.D  
 Acq On : 06 Jun 2025 14:26  
 Operator : JC\MD  
 Sample : VSTDIICC100  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1




Quant Time: Jun 07 01:58:16 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDIICC100

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.235 min Scan# 1

Delta R.T. -0.001 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#2

Dichlorodifluoromethane

Concen: 107.350 ug/l


RT: 2.153 min Scan# 34




Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 85 Resp: 194799  
 Ion Ratio Lower Upper  
 85 100  
 87 32.3 16.2 48.6








#3  
Chloromethane  
Concen: 95.833 ug/l  
RT: 2.395 min Scan# 7  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100

Tgt Ion: 50 Resp: 224555  
Ion Ratio Lower Upper  
50 100  
52 33.4 25.7 38.5


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#4  
Vinyl Chloride  
Concen: 101.414 ug/l  
RT: 2.553 min Scan# 102  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion: 62 Resp: 245135  
Ion Ratio Lower Upper  
62 100  
64 31.5 26.0 39.0





#5

Bromomethane

Concen: 101.996 ug/l

RT: 2.977 min Scan# 1

Delta R.T. -0.018 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#6

Chloroethane

Concen: 97.447 ug/l

RT: 3.147 min Scan# 203

Delta R.T. -0.006 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26


Tgt Ion: 64 Resp: 152150

Ion Ratio Lower Upper

64 100


66 32.5 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 98.917 ug/l  
RT: 3.524 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100




Tgt Ion:101 Resp: 31239  
Ion Ratio Lower Upper  
101 100  
103 64.7 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#8  
Diethyl Ether  
Concen: 101.631 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26




Tgt Ion: 74 Resp: 139844  
Ion Ratio Lower Upper  
74 100  
45 89.5 45.5 136.5









#9  
 1,1,2-Trichlorotrifluoroethane  
 Concen: 100.163 ug/l  
 RT: 4.389 min Scan# 415  
 Delta R.T. -0.006 min  
 Lab File: VN086866.D  
 Acq: 06 Jun 2025 14:26  
**Instrument:** MSVOA\_N  
**ClientSampleId:** VSTDICC100





Tgt Ion:101 Resp: 198739  
 Ion Ratio Lower Upper  
 101 100  
 85 42.2 33.8 50.8  
 151 74.2 58.7 88.1  
**Manual Integrations APPROVED**  
 Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025






#10  
 Methyl Iodide  
 Concen: 102.162 ug/l  
 RT: 4.612 min Scan# 452  
 Delta R.T. 0.006 min  
 Lab File: VN086866.D  
 Acq: 06 Jun 2025 14:26



Tgt Ion:142 Resp: 262771  
 Ion Ratio Lower Upper  
 142 100  
 127 37.9 30.9 46.3  
 141 13.6 10.4 15.6









#11

Tert butyl alcohol  
Concen: 496.062 ug/l  
RT: 5.547 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#12

1,1-Dichloroethene  
Concen: 98.838 ug/l  
RT: 4.359 min Scan# 409  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion: 96 Resp: 200212  
Ion Ratio Lower Upper  
96 100  
61 156.9 125.8 188.6  
98 63.7 52.5 78.7





#13

Acrolein

Concen: 449.961 ug/l

RT: 4.200 min Scan# 3

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025





Supervised By :Mahesh Dadoda 06/09/2025

Tgt Ion: 56 Resp: 9417

Ion Ratio Lower Upper

56 100

55 70.9 55.6 83.4



#14

Allyl chloride

Concen: 98.055 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion: 41 Resp: 329412




Ion Ratio Lower Upper

41 100

39 63.7 51.3 76.9

76 39.7 31.4 47.0





#15

Acrylonitrile

Concen: 498.552 ug/l

RT: 5.735 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 470.648 ug/l

RT: 4.447 min Scan# 424


Delta R.T. -0.000 min




Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 608178

Ion	Ratio	Lower	Upper
43	100		
58	36.2	28.0	42.0





#17

Carbon Disulfide

Concen: 97.179 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 101.390 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26


Tgt Ion: 43 Resp: 381808

Ion Ratio Lower Upper



43 100

74 26.1 20.7 31.1






Abundance Scan 657 (5.818 min): VN086866.D\data.ms



Abundance Scan 568 (5.294 min): VN086866.D\data.ms



#19

Methyl tert-butyl Ether

Concen: 100.306 ug/l

RT: 5.818 min Scan# 6

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#20

Methylene Chloride

Concen: 94.575 ug/l

RT: 5.294 min Scan# 568

Delta R.T. -0.000 min


Lab File: VN086866.D


Acq: 06 Jun 2025 14:26

Tgt Ion: 84 Resp: 228800



Ion Ratio Lower Upper

84	100		
49	111.4	90.5	135.7
51	34.2	28.5	42.7
86	63.5	48.1	72.1






Abundance Scan 655 (5.806 min): VN086866.D\data.ms



Abundance Scan 805 (6.688 min): VN086866.D\data.ms



Abundance Scan 805 (6.688 min): VN086866.D\data.ms (-75)

Sub

m/z-->

#21

trans-1,2-Dichloroethene

Concen: 95.476 ug/l

RT: 5.806 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC100

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#22

Diisopropyl ether

Concen: 98.409 ug/l

RT: 6.688 min Scan# 805

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 45 Resp: 696881


Ion Ratio Lower Upper




45 100

43 51.1 41.9 62.9

87 29.6 22.4 33.6

59 12.2 9.8 14.6





#23

**Vinyl Acetate**

Concen: 487.936 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

**1,1-Dichloroethane**

Concen: 99.186 ug/l

RT: 6.588 min Scan# 788

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion: 63 Resp: 404207




Ion Ratio Lower Upper

63 100

98 7.3 3.6 10.8

100 4.7 2.6 7.8





#25

2-Butanone

Concen: 496.743 ug/l

RT: 7.494 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 104.729 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26




Tgt Ion: 77 Resp: 331995

Ion Ratio Lower Upper

77 100

97 23.4 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 98.410 ug/l

RT: 7.500 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 93.851 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion: 49 Resp: 188050




Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 78.9 62.2 93.2





#29

Tetrahydrofuran

Concen: 495.240 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 97.040 ug/l

RT: 7.977 min Scan# 1024


Delta R.T. -0.000 min




Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 83 Resp: 394927

Ion	Ratio	Lower	Upper
83	100		
85	65.1	52.9	79.3





#31

Cyclohexane

Concen: 94.869 ug/l

RT: 8.265 min Scan# 1074

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane

Concen: 97.306 ug/l


RT: 8.177 min Scan# 1058


Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 97 Resp: 336819  
 Ion Ratio Lower Upper  
 97 100  
 99 65.7 52.0 78.0  
 61 44.2 35.9 53.9







#33

1,2-Dichloroethane-d4  
Concen: 98.065 ug/l  
RT: 8.588 min Scan# 1128  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26



Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100




Tgt Ion: 65 Resp: 23895  
Ion Ratio Lower Upper  
65 100  
67 54.8 0.0 105.6

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26



Tgt Ion:114 Resp: 327782  
Ion Ratio Lower Upper  
114 100  
63 19.7 0.0 39.6  
88 15.4 0.0 30.2





#35

Dibromofluoromethane

Concen: 100.495 ug/l

RT: 8.177 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC100



Tgt Ion:113 Resp: 19521

Ion Ratio Lower Upper

113 100

111 102.9 84.2 126.2







192 18.2 14.2 21.4







Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#36  
1,1-Dichloropropene  
Concen: 100.156 ug/l  
RT: 8.382 min Scan# 1093  
Delta R.T. 0.006 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26Tgt Ion: 75 Resp: 289907  
Ion Ratio Lower Upper  
75 100  
110 34.9 17.6 52.9  
77 31.1 25.3 37.9



#37

**Ethyl Acetate**

Concen: 100.591 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 37065

Ion Ratio Lower Upper

43 100


61 15.3 12.1 18.1

70 12.6 9.8 14.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

**Carbon Tetrachloride**

Concen: 100.404 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion:117 Resp: 285325


Ion Ratio Lower Upper

117 100

119 95.4 75.0 112.6

121 28.6 24.9 37.3





#39

Methylcyclohexane

Concen: 99.758 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC100



Tgt Ion: 83 Resp: 39560

Ion Ratio Lower Upper

83 100




55 76.8 61.0 91.6

98 47.6 38.9 58.3

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 97.946 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086866.D



Acq: 06 Jun 2025 14:26







Tgt Ion: 78 Resp: 927083

Ion Ratio Lower Upper

78 100

77 23.6 18.7 28.1





#41

Methacrylonitrile

Concen: 94.688 ug/l

RT: 7.794 min Scan# 993

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument :

MSVOA\_N


ClientSampleId :

VSTDICC100

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#42

1,2-Dichloroethane

Concen: 98.232 ug/l

RT: 8.676 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26




Tgt Ion: 62 Resp: 282015

Ion Ratio Lower Upper

62 100

98 10.6 0.0 21.2





#43

Isopropyl Acetate

Concen: 97.981 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N

ClientSampleId :





VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

Abundance



#44

Trichloroethene

Concen: 99.280 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion:130 Resp: 222822


Ion Ratio Lower Upper

130 100

95 101.4 0.0 196.2

Abundance





#45

1,2-Dichloropropane

Concen: 100.077 ug/l

RT: 9.629 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC100

Abundance Scan 1305 (9.629 min): VN086866.D\data.ms



Tgt Ion: 63 Resp: 230454

Ion Ratio Lower Upper






63 100




65 30.6 25.5 38.3

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025

#46  
Dibromomethane  
Concen: 101.685 ug/l  
RT: 9.712 min Scan# 1319  
Delta R.T. -0.006 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26Tgt Ion: 93 Resp: 155530  
Ion Ratio Lower Upper  
93 100  
95 82.9 67.1 100.7  
174 91.6 74.2 111.4



#47

Bromodichloromethane

Concen: 100.533 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 101.288 ug/l


RT: 9.682 min Scan# 1314




Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt	Ion	Resp:	275710
Ion	Ratio	Lower	Upper
41	100		
69	92.4	74.6	112.0
39	50.1	40.2	60.4





#49

1,4-Dioxane

Concen: 2115.744 ug/l

RT: 9.706 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 100.419 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26




Tgt Ion: 98 Resp: 772210

Ion Ratio Lower Upper

98 100

100 67.6 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 517.729 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#52

Toluene

Concen: 100.092 ug/l

RT: 10.635 min Scan# 1476

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26




Tgt Ion: 92 Resp: 578978

Ion Ratio Lower Upper

92 100

91 170.6 136.5 204.7






#53

t-1,3-Dichloropropene  
Concen: 102.077 ug/l  
RT: 10.841 min Scan# 1422  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#54

cis-1,3-Dichloropropene  
Concen: 101.730 ug/l  
RT: 10.318 min Scan# 1422  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 383030  
Ion Ratio Lower Upper  
75 100  
77 32.3 26.2 39.4  
39 40.1 32.3 48.5











#55  
1,1,2-Trichloroethane  
Concen: 98.808 ug/l  
RT: 11.017 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025





#57

1,3-Dichloropropane

Concen: 99.069 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 526.621 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26


Tgt Ion: 63 Resp: 1113523

Ion Ratio Lower Upper


63 100

106 29.1 23.0 34.6








Abundance Scan 1572 (11.200 min): VN086866.D\data.ms (-)




Abundance Scan 1572 (11.200 min): VN086866.D\data.ms (-)



Abundance Scan 1599 (11.359 min): VN086866.D\data.ms (-)



Abundance Scan 1599 (11.359 min): VN086866.D\data.ms (-)



#59

2-Hexanone

Concen: 567.535 ug/l

RT: 11.200 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC100

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#60

Dibromochloromethane

Concen: 102.616 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26




Tgt Ion:129 Resp: 237965

Ion Ratio Lower Upper

129 100

127 77.1 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 101.787 ug/l  
RT: 11.470 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICC100


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#62  
4-Bromofluorobenzene  
Concen: 102.339 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion: 95 Resp: 292385  
Ion Ratio Lower Upper  
95 100  
174 71.6 0.0 141.8  
176 68.6 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.870 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#64  
Tetrachloroethene  
Concen: 98.824 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion:164 Resp: 181568  
Ion Ratio Lower Upper  
164 100  
166 121.8 101.8 152.8  
129 96.8 83.0 124.4  
131 91.5 78.2 117.2





#65

Chlorobenzene

Concen: 98.793 ug/l

RT: 11.894 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 100.521 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion:131 Resp: 206895




Ion Ratio Lower Upper

131 100

133 96.4 48.3 144.8

119 68.2 34.0 101.8





#67

Ethyl Benzene

Concen: 100.717 ug/l

RT: 11.964 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#68

m/p-Xylenes

Concen: 202.411 ug/l


RT: 12.070 min Scan# 1720


Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion:106 Resp: 854557  
 Ion Ratio Lower Upper  
 106 100  
 91 200.4 159.4 239.0







#69  
o-Xylene  
Concen: 102.070 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#70  
Styrene  
Concen: 103.127 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion:104 Resp: 713548  
Ion Ratio Lower Upper  
104 100  
78 48.4 38.5 57.7  
103 54.0 43.1 64.7





#71

Bromoform

Concen: 108.707 ug/l

RT: 12.576 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26



Tgt Ion:152 Resp: 147098




Ion Ratio Lower Upper

152 100

115 60.1 30.1 90.5

150 190.4 0.0 345.0





#73

Isopropylbenzene

Concen: 99.410 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#74

N-amyl acetate

Concen: 113.475 ug/l

RT: 12.506 min Scan# 1794

Delta R.T. -0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion: 43 Resp: 403587


Ion Ratio Lower Upper




43 100

70 47.7 37.6 56.4

55 25.0 19.0 28.6

61 27.3 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 97.616 ug/l  
RT: 12.941 min Scan# 1868  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#76  
1,2,3-Trichloropropane  
Concen: 100.738 ug/l m  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion: 75 Resp: 353334  
Ion Ratio Lower Upper  
75 100  
77 162.2 79.3 237.9





#77

Bromobenzene

Concen: 100.266 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#78

n-propylbenzene

Concen: 99.948 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26




Tgt Ion: 91 Resp: 1301631

Ion Ratio Lower Upper

91 100

120 22.9 11.4 34.2





#79

2-Chlorotoluene

Concen: 98.630 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 101.873 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086866.D


Acq: 06 Jun 2025 14:26


Tgt Ion:105 Resp: 901317

Ion Ratio Lower Upper

105 100

120 49.7 24.9 74.6







#81  
trans-1,4-Dichloro-2-butene  
Concen: 105.820 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC100


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#82  
4-Chlorotoluene  
Concen: 99.166 ug/l  
RT: 13.223 min Scan# 1916  
Delta R.T. -0.000 min  
Lab File: VN086866.D  
Acq: 06 Jun 2025 14:26

Tgt Ion: 91 Resp: 783657  
Ion Ratio Lower Upper  
91 100  
126 33.2 16.7 50.1





#83

tert-Butylbenzene

Concen: 99.710 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 101.983 ug/l



RT: 13.482 min Scan# 1960




Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt	Ion:105	Resp:	904794
Ion	Ratio	Lower	Upper
105	100		
120	46.7	23.2	69.6





#85

sec-Butylbenzene

Concen: 100.513 ug/l

RT: 13.617 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#86

p-Isopropyltoluene

Concen: 101.857 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion:119 Resp: 990273




Ion Ratio Lower Upper

119 100

134 26.7 13.5 40.4

91 22.6 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 98.063 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 97.978 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion:146 Resp: 483003


Ion Ratio Lower Upper

146 100

111 39.3 19.9 59.7

148 63.0 31.8 95.4





#89

n-Butylbenzene

Concen: 99.598 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC100

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 103.019 ug/l

RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min




Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion:117 Resp: 169772

Ion	Ratio	Lower	Upper
117	100		
201	65.8	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 98.602 ug/l

RT: 14.106 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 95.877 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26


Tgt Ion: 75 Resp: 83243

Ion Ratio Lower Upper

75 100




155 86.1 41.1 123.4

157 109.8 52.7 158.1



Abundance

14.717



#93

1,2,4-Trichlorobenzene

Concen: 100.751 ug/l

RT: 15.394 min Scan# 2303

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC100

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 101.789 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt Ion:225 Resp: 112496




Ion Ratio Lower Upper

225 100

223 64.2 33.5 100.4

227 62.1 31.1 93.5





#95

Naphthalene

Concen: 102.290 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC100

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 99.796 ug/l


RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086866.D

Acq: 06 Jun 2025 14:26

Tgt	Ion:180	Resp:	294127
Ion	Ratio	Lower	Upper
180	100		
182	95.3	47.0	141.0
145	33.5	16.6	49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086867.D  
 Acq On : 06 Jun 2025 14:49  
 Operator : JC\MD  
 Sample : VSTDICC150  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC150

Quant Time: Jun 07 01:59:15 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.229	168	172832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	307937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	278552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136630	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	389297	168.233	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 336.460%	#	
35) Dibromofluoromethane	8.176	113	324474	177.803	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 355.600%	#	
50) Toluene-d8	10.570	98	1272348	176.120	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 352.240%	#	
62) 4-Bromofluorobenzene	12.847	95	481421	179.364	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 358.720%	#	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.153	85	262143	152.122	ug/l	100
3) Chloromethane	2.394	50	304558	136.865	ug/l	100
4) Vinyl Chloride	2.553	62	335834	146.303	ug/l	100
5) Bromomethane	2.965	94	190786	148.522	ug/l	98
6) Chloroethane	3.136	64	208568	140.664	ug/l	96
7) Trichlorofluoromethane	3.518	101	427871	142.666	ug/l	100
8) Diethyl Ether	3.983	74	191555	146.593	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.388	101	269810	143.192	ug/l	100
10) Methyl Iodide	4.606	142	352020	144.118	ug/l	99
11) Tert butyl alcohol	5.553	59	430866	686.213	ug/l	100
12) 1,1-Dichloroethene	4.353	96	273388	142.118	ug/l	98
13) Acrolein	4.194	56	171625	863.526	ug/l	99
14) Allyl chloride	5.035	41	490852	153.858	ug/l	92
15) Acrylonitrile	5.735	53	1048372	714.344	ug/l	99
16) Acetone	4.447	43	819041	667.435	ug/l	100
17) Carbon Disulfide	4.730	76	742788	139.594	ug/l	100
18) Methyl Acetate	5.041	43	523996	146.527	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	998781	143.396	ug/l	99
20) Methylene Chloride	5.300	84	311597	135.628	ug/l	99
21) trans-1,2-Dichloroethene	5.806	96	290960	135.947	ug/l	97
22) Diisopropyl ether	6.688	45	947638	140.915	ug/l	98
23) Vinyl Acetate	6.618	43	3932872	692.189	ug/l	99
24) 1,1-Dichloroethane	6.582	63	540881	139.761	ug/l	99
25) 2-Butanone	7.494	43	1381940	692.809	ug/l	99
26) 2,2-Dichloropropane	7.500	77	450112	149.519	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	363398	141.974	ug/l	99
28) Bromochloromethane	7.824	49	290594	152.718	ug/l	98
29) Tetrahydrofuran	7.847	42	898159	691.211	ug/l	98
30) Chloroform	7.976	83	534038	138.179	ug/l	97
31) Cyclohexane	8.265	56	505795	134.766	ug/l	96
32) 1,1,1-Trichloroethane	8.176	97	462854	140.807	ug/l	99
36) 1,1-Dichloropropene	8.376	75	393166	144.584	ug/l	100
37) Ethyl Acetate	7.571	43	494000	142.706	ug/l	99
38) Carbon Tetrachloride	8.371	117	388595	145.557	ug/l	98
39) Methylcyclohexane	9.606	83	544437	146.136	ug/l	97
40) Benzene	8.612	78	1266139	142.388	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086867.D  
 Acq On : 06 Jun 2025 14:49  
 Operator : JC\MD  
 Sample : VSTDICC150  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC150

Quant Time: Jun 07 01:59:15 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	276400	142.174	ug/l	98
42) 1,2-Dichloroethane	8.676	62	381516	141.455	ug/l	100
43) Isopropyl Acetate	8.694	43	780791	140.531	ug/l	99
44) Trichloroethene	9.359	130	302548	143.490	ug/l	96
45) 1,2-Dichloropropane	9.623	63	309047	142.856	ug/l	100
46) Dibromomethane	9.712	93	208461	145.074	ug/l	100
47) Bromodichloromethane	9.894	83	430011	145.449	ug/l	100
48) Methyl methacrylate	9.688	41	366967	143.502	ug/l	99
49) 1,4-Dioxane	9.712	88	138035	2967.116	ug/l #	100
51) 4-Methyl-2-Pentanone	10.447	43	2438311	728.995	ug/l	98
52) Toluene	10.635	92	793270	145.976	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	489809	148.163	ug/l	97
54) cis-1,3-Dichloropropene	10.318	75	518092	146.469	ug/l	99
55) 1,1,2-Trichloroethane	11.017	97	298151	142.588	ug/l	99
56) Ethyl methacrylate	10.876	69	533912	160.311	ug/l	98
57) 1,3-Dichloropropane	11.165	76	516391	142.363	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	1869543	941.147	ug/l	99
59) 2-Hexanone	11.200	43	1735777	805.662	ug/l	99
60) Dibromochloromethane	11.365	129	322064	147.832	ug/l	98
61) 1,2-Dibromoethane	11.470	107	314606	146.790	ug/l	99
64) Tetrachloroethene	11.106	164	245122	139.048	ug/l	97
65) Chlorobenzene	11.894	112	860345	140.041	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	282739	143.171	ug/l	99
67) Ethyl Benzene	11.964	91	1517883	143.438	ug/l	98
68) m/p-Xylenes	12.070	106	1175378	290.155	ug/l	99
69) o-Xylene	12.400	106	566301	145.966	ug/l	100
70) Styrene	12.411	104	972318	146.459	ug/l	100
71) Bromoform	12.582	173	223200	152.538	ug/l #	99
73) Isopropylbenzene	12.694	105	1453651	146.042	ug/l	100
74) N-amyl acetate	12.500	43	543747	164.596	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.941	83	474426	140.693	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	468041m	143.665	ug/l	
77) Bromobenzene	12.982	156	335734	147.077	ug/l	98
78) n-propylbenzene	13.035	91	1769694	146.301	ug/l	99
79) 2-Chlorotoluene	13.123	91	1046916	144.317	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	1210325	147.279	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	216253	153.277	ug/l	95
82) 4-Chlorotoluene	13.223	91	1063581	144.900	ug/l	100
83) tert-Butylbenzene	13.435	119	1093406	145.352	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	1216529	147.626	ug/l	100
85) sec-Butylbenzene	13.617	105	1582762	144.885	ug/l	99
86) p-Isopropyltoluene	13.729	119	1322780	146.482	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	641966	142.939	ug/l	100
88) 1,4-Dichlorobenzene	13.811	146	646188	141.124	ug/l	99
89) n-Butylbenzene	14.053	91	1251819	143.117	ug/l	100
90) Hexachloroethane	14.335	117	230195	150.387	ug/l	98
91) 1,2-Dichlorobenzene	14.105	146	613177	142.108	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	110677	137.242	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	407621	147.940	ug/l	99
94) Hexachlorobutadiene	15.500	225	151172	147.264	ug/l	98
95) Naphthalene	15.641	128	1545981	150.743	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	404450	147.743	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086867.D  
 Acq On : 06 Jun 2025 14:49  
 Operator : JC\MD  
 Sample : VSTDICC150  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC150

Quant Time: Jun 07 01:59:15 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 01:51:02 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
(#)						

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
Data File : VN086867.D  
Acq On : 06 Jun 2025 14:49  
Operator : JC\MD  
Sample : VSTDIICC150  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Jun 07 01:59:15 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 01:51:02 2025  
Response via : Initial Calibration

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC150

## Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.229 min Scan# 1  
Delta R.T. -0.007 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC150




Tgt Ion:168 Resp: 17283  
Ion Ratio Lower Upper  
168 100  
99 61.8 49.1 73.7

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#2  
Dichlorodifluoromethane  
Concen: 152.122 ug/l  
RT: 2.153 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49



Tgt Ion: 85 Resp: 262143  
Ion Ratio Lower Upper  
85 100  
87 32.3 16.2 48.6





#3

Chloromethane

Concen: 136.865 ug/l

RT: 2.394 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#4

Vinyl Chloride

Concen: 146.303 ug/l


RT: 2.553 min Scan# 102


Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion: 62 Resp: 335834  
 Ion Ratio Lower Upper  
 62 100  
 64 32.6 26.0 39.0





#5  
Bromomethane  
Concen: 148.522 ug/l  
RT: 2.965 min Scan# 1  
Delta R.T. -0.030 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49




Instrument : MSVOA\_N  
ClientSampleId : VSTDICC150




Tgt Ion: 94 Resp: 190780  
Ion Ratio Lower Upper  
94 100  
96 95.5 77.9 116.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#6  
Chloroethane  
Concen: 140.664 ug/l  
RT: 3.136 min Scan# 201  
Delta R.T. -0.018 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49




Tgt Ion: 64 Resp: 208568  
Ion Ratio Lower Upper  
64 100  
66 32.7 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 142.666 ug/l  
RT: 3.518 min Scan# 2  
Delta R.T. -0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49




**Instrument:** MSVOA\_N  
**ClientSampleId:** VSTDICC150




Tgt Ion:101 Resp: 42787:  
Ion Ratio Lower Upper  
101 100  
103 65.5 52.2 78.2

**Manual Integrations APPROVED**



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#8  
Diethyl Ether  
Concen: 146.593 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49




Tgt Ion: 74 Resp: 191555  
Ion Ratio Lower Upper  
74 100  
45 90.3 45.5 136.5








#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 143.192 ug/l  
RT: 4.388 min Scan# 4  
Instrument : MSVOA\_N  
Delta R.T. -0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49




Tgt Ion:101 Resp: 269810  
Ion Ratio Lower Upper  
101 100  
85 42.3 33.8 50.8  
151 73.5 58.7 88.1

**Manual Integrations APPROVED**



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#10  
Methyl Iodide  
Concen: 144.118 ug/l  
RT: 4.606 min Scan# 451  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49



Tgt Ion:142 Resp: 352020  
Ion Ratio Lower Upper  
142 100  
127 39.1 30.9 46.3  
141 13.5 10.4 15.6









#11

Tert butyl alcohol  
Concen: 686.213 ug/l  
RT: 5.553 min Scan# 6  
Delta R.T. 0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC150

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#12


1,1-Dichloroethene  
Concen: 142.118 ug/l  
RT: 4.353 min Scan# 408  
Delta R.T. -0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 96 Resp: 273388  
Ion Ratio Lower Upper  
96 100  
61 154.5 125.8 188.6  
98 64.9 52.5 78.7





Abundance Scan 381 (4.194 min): VN086867.D\data.ms



#13

Acrolein

Concen: 863.526 ug/l

RT: 4.194 min Scan# 3

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC150


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 524 (5.035 min): VN086867.D\data.ms



#14

Allyl chloride

Concen: 153.858 ug/l


RT: 5.035 min Scan# 524




Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt	Ion	Resp:	
	41	490852	
Ion	Ratio	Lower	Upper
41	100		
39	56.4	51.3	76.9
76	35.7	31.4	47.0





#15

Acrylonitrile

Concen: 714.344 ug/l

RT: 5.735 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 667.435 ug/l

RT: 4.447 min Scan# 424


Delta R.T. -0.000 min




Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 819041

Ion	Ratio	Lower	Upper
43	100		
58	34.7	28.0	42.0





#17

Carbon Disulfide

Concen: 139.594 ug/l

RT: 4.730 min Scan# 4

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 146.527 ug/l


RT: 5.041 min Scan# 525


Delta R.T. -0.000 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 523996  
 Ion Ratio Lower Upper  
 43 100  
 74 26.1 20.7 31.1






Abundance Scan 657 (5.818 min): VN086867.D\data.ms



Abundance Scan 657 (5.818 min): VN086867.D\data.ms (-60)



#19

Methyl tert-butyl Ether

Concen: 143.396 ug/l

RT: 5.818 min Scan# 6

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC150


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 569 (5.300 min): VN086867.D\data.ms (-51)



#20

Methylene Chloride

Concen: 135.628 ug/l

RT: 5.300 min Scan# 569

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion: 84 Resp: 311597


Ion Ratio Lower Upper


84 100

49 113.0 90.5 135.7



51 36.1 28.5 42.7

86 62.3 48.1 72.1






Abundance Scan 655 (5.806 min): VN086867.D\data.ms



Abundance Scan 805 (6.688 min): VN086867.D\data.ms



#21

trans-1,2-Dichloroethene

Concen: 135.947 ug/l

RT: 5.806 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC150

### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



Time--&gt;

5.60 5.80 6.00



#23

**Vinyl Acetate**

Concen: 692.189 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

**1,1-Dichloroethane**

Concen: 139.761 ug/l

RT: 6.582 min Scan# 787


Delta R.T. -0.000 min




Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion: 63 Resp: 540881

Ion	Ratio	Lower	Upper
63	100		
98	7.3	3.6	10.8
100	4.4	2.6	7.8





#25

2-Butanone

Concen: 692.809 ug/l

RT: 7.494 min Scan# 9

Instrument :

Delta R.T. -0.000 min

MSVOA\_N

Lab File: VN086867.D

ClientSampleId :




Acq: 06 Jun 2025 14:49

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 149.519 ug/l

RT: 7.500 min Scan# 943

Delta R.T. -0.000 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49




Tgt Ion: 77 Resp: 450112

Ion Ratio Lower Upper

77 100

97 23.4 11.6 34.7









#27

cis-1,2-Dichloroethene  
Concen: 141.974 ug/l  
RT: 7.500 min Scan# 9

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICC150  
Acq: 06 Jun 2025 14:49

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#28

Bromochloromethane  
Concen: 152.718 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 49 Resp: 290594  
Ion Ratio Lower Upper  
49 100  
129 2.0 0.0 3.8  
130 79.8 62.2 93.2





#29

Tetrahydrofuran

Concen: 691.211 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 138.179 ug/l

RT: 7.976 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49

Tgt Ion: 83 Resp: 534038

Ion Ratio Lower Upper

83 100

85 63.4 52.9 79.3



#31

Cyclohexane

Concen: 134.766 ug/l

RT: 8.265 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC150



Tgt Ion: 56 Resp: 505795

Ion Ratio Lower Upper

56 100

69 32.0 25.4 38.0


84 90.6 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane


Concen: 140.807 ug/l

RT: 8.176 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49




Tgt Ion: 97 Resp: 462854

Ion Ratio Lower Upper


97 100


99 64.9 52.0 78.0

61 43.2 35.9 53.9





Abundance





#33

1,2-Dichloroethane-d4  
Concen: 168.233 ug/l  
RT: 8.588 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49










Tgt Ion: 65 Resp: 389297

Ion	Ratio	Lower	Upper
65	100		
67	54.9	0.0	105.6

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025





#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion:114 Resp: 307937

Ion	Ratio	Lower	Upper
114	100		
63	19.8	0.0	39.6
88	16.5	0.0	30.2





#35

Dibromofluoromethane

Concen: 177.803 ug/l

RT: 8.176 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49

Instrument :

MSVOA\_N

ClientSampleId :

VSTDICC150



Tgt Ion: 113 Resp: 324474

Ion Ratio Lower Upper

113 100



111 102.5 84.2 126.2

192 17.7 14.2 21.4

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#36

1,1-Dichloropropene

Concen: 144.584 ug/l

RT: 8.376 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49



Tgt Ion: 75 Resp: 393166




Ion Ratio Lower Upper

75 100

110 35.6 17.6 52.9

77 31.4 25.3 37.9





#37

**Ethyl Acetate**

Concen: 142.706 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 494000

Ion Ratio Lower Upper

43 100





61 15.4 12.1 18.1

70 12.7 9.8 14.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

**Carbon Tetrachloride**

Concen: 145.557 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:117 Resp: 388595




Ion Ratio Lower Upper

117 100

119 94.9 75.0 112.6

121 29.1 24.9 37.3





#39

Methylcyclohexane

Concen: 146.136 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#40

Benzene

Concen: 142.388 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49




Tgt Ion: 78 Resp: 1266139

Ion Ratio Lower Upper

78 100

77 23.6 18.7 28.1





#41

Methacrylonitrile

Concen: 142.174 ug/l

RT: 7.788 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#42

1,2-Dichloroethane

Concen: 141.455 ug/l

RT: 8.676 min Scan# 1143


Delta R.T. -0.000 min


Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion: 62 Resp: 381516

	Ion Ratio	Lower	Upper
62	100		
98	10.5	0.0	21.2






Abundance Scan 1146 (8.694 min): VN086867.D\data.ms



Abundance Scan 1146 (8.694 min): VN086867.D\data.ms (-1)



#43

Isopropyl Acetate

Concen: 140.531 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N

ClientSampleId :


VSTDICC150

**Manual Integrations  
APPROVED**


Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025


Abundance




Abundance Scan 1259 (9.359 min): VN086865.D\data.ms (-1)



Abundance Scan 1259 (9.359 min): VN086867.D\data.ms



Abundance Scan 1259 (9.359 min): VN086867.D\data.ms (-1)



#44

Trichloroethene

Concen: 143.490 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:130 Resp: 302548


Ion Ratio Lower Upper

130 100

95 101.6 0.0 196.2

Abundance





#45

1,2-Dichloropropane

Concen: 142.856 ug/l

RT: 9.623 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086867.D



Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC150



Tgt Ion: 63 Resp: 30904

Ion Ratio Lower Upper



63 100

65 31.9 25.5 38.3


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025



Supervised By :Mahesh Dadoda 06/09/2025






#46  
Dibromomethane  
Concen: 145.074 ug/l  
RT: 9.712 min Scan# 1319  
Delta R.T. -0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49



Tgt Ion: 93 Resp: 208461  
Ion Ratio Lower Upper  
93 100  
95 84.8 67.1 100.7  
174 92.8 74.2 111.4





#47

Bromodichloromethane

Concen: 145.449 ug/l

RT: 9.894 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49





Instrument:

MSVOA\_N

ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

 Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025


#48

Methyl methacrylate

Concen: 143.502 ug/l


RT: 9.688 min Scan# 1315




Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt	Ion	41	Resp:	366967
Ion	Ratio	Lower	Upper	
41	100			
69	94.2	74.6	112.0	
39	50.8	40.2	60.4	





#49

1,4-Dioxane

Concen: 2967.116 ug/l

RT: 9.712 min Scan# 1

Delta R.T. 0.011 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument :

MSVOA\_N



ClientSampleId :

VSTDICC150


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025



Supervised By :Mahesh Dadoda 06/09/2025






#50  
Toluene-d8  
Concen: 176.120 ug/l  
RT: 10.570 min Scan# 1465  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49



Tgt Ion: 98 Resp: 1272348  
Ion Ratio Lower Upper  
98 100  
100 67.5 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 728.995 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**



Reviewed By :John Carlone 06/09/2025




Supervised By :Mahesh Dadoda 06/09/2025



#52  
Toluene  
Concen: 145.976 ug/l  
RT: 10.635 min Scan# 1476  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49



Tgt Ion: 92 Resp: 793270  
Ion Ratio Lower Upper  
92 100  
91 170.2 136.5 204.7





#53  
 t-1,3-Dichloropropene  
 Concen: 148.163 ug/l  
 RT: 10.841 min Scan# 1422  
 Delta R.T. -0.000 min  
 Lab File: VN086867.D  
 Acq: 06 Jun 2025 14:49


Tgt Ion: 75 Resp: 489809  
 Ion Ratio Lower Upper  
 75 100  
 77 31.4 26.4 39.6



Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC150

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#55  
1,1,2-Trichloroethane  
Concen: 142.588 ug/l  
RT: 11.017 min Scan# 1  
Instrument : MSVOA\_N  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49  
ClientSampleId : VSTDICC150


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#56  
Ethyl methacrylate  
Concen: 160.311 ug/l  
RT: 10.876 min Scan# 1517  
Delta R.T. -0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 69 Resp: 533912  
Ion Ratio Lower Upper  
69 100  
41 57.5 47.2 70.8  
39 27.0 23.0 34.6





#57

1,3-Dichloropropane

Concen: 142.363 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**



Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025




#58  
2-Chloroethyl Vinyl ether  
Concen: 941.147 ug/l  
RT: 10.165 min Scan# 1396  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 63 Resp: 1869543  
Ion Ratio Lower Upper  
63 100  
106 29.2 23.0 34.6








Abundance Scan 1572 (11.200 min): VN086867.D\data.ms




Abundance Scan 1572 (11.200 min): VN086867.D\data.ms (-)



Abundance Scan 1600 (11.365 min): VN086867.D\data.ms



Abundance Scan 1600 (11.365 min): VN086867.D\data.ms (-)



#59

2-Hexanone

Concen: 805.662 ug/l

RT: 11.200 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#60

Dibromochloromethane

Concen: 147.832 ug/l

RT: 11.365 min Scan# 1600

Delta R.T. 0.006 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49




Tgt Ion:129 Resp: 322064

Ion Ratio Lower Upper

129 100

127 78.6 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 146.790 ug/l  
RT: 11.470 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDICC150


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#62  
4-Bromofluorobenzene  
Concen: 179.364 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 95 Resp: 481421  
Ion Ratio Lower Upper  
95 100  
174 70.7 0.0 141.8  
176 67.8 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.864 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#64

Tetrachloroethene


Concen: 139.048 ug/l


RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49







#65  
Chlorobenzene  
Concen: 140.041 ug/l  
RT: 11.894 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC150


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 143.171 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion:131 Resp: 282739  
Ion Ratio Lower Upper  
131 100  
133 95.2 48.3 144.8  
119 67.8 34.0 101.8





#67

Ethyl Benzene

Concen: 143.438 ug/l

RT: 11.964 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#68

m/p-Xylenes

Concen: 290.155 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49


Tgt Ion:106 Resp: 1175378

Ion Ratio Lower Upper

106 100

91 200.8 159.4 239.0







#69  
o-Xylene  
Concen: 145.966 ug/l  
RT: 12.400 min Scan# 1  
Instrument : MSVOA\_N  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

ClientSampleId : VSTDICC150


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#70  
Styrene  
Concen: 146.459 ug/l  
RT: 12.411 min Scan# 1778  
Instrument : MSVOA\_N  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion:104 Resp: 972318  
Ion Ratio Lower Upper  
104 100  
78 48.4 38.5 57.7  
103 54.0 43.1 64.7





#71

Bromoform

Concen: 152.538 ug/l

RT: 12.582 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:152 Resp: 136630




Ion Ratio Lower Upper

152 100

115 61.1 30.1 90.5

150 206.7 0.0 345.0





#73

Isopropylbenzene

Concen: 146.042 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#74

N-amyl acetate

Concen: 164.596 ug/l

RT: 12.500 min Scan# 1793

Delta R.T. -0.012 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion: 43 Resp: 543747


Ion Ratio Lower Upper




43 100

70 50.0 37.6 56.4

55 26.1 19.0 28.6

61 27.6 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 140.693 ug/l  
RT: 12.941 min Scan# 1868  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC150


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#76  
1,2,3-Trichloropropane  
Concen: 143.665 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 75 Resp: 468041  
Ion Ratio Lower Upper  
75 100  
77 166.0 79.3 237.9





#77

Bromobenzene

Concen: 147.077 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#78

n-propylbenzene

Concen: 146.301 ug/l


RT: 13.035 min Scan# 1884




Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion: 91 Resp: 1769694  
 Ion Ratio Lower Upper  
 91 100  
 120 23.2 11.4 34.2





#79

2-Chlorotoluene

Concen: 144.317 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 147.279 ug/l


RT: 13.170 min Scan# 1907




Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt Ion:105 Resp: 1210325  
 Ion Ratio Lower Upper  
 105 100  
 120 50.0 24.9 74.6








#81  
trans-1,4-Dichloro-2-butene  
Concen: 153.277 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Instrument : MSVOA\_N  
ClientSampleId : VSTDICC150



### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#82  
4-Chlorotoluene  
Concen: 144.900 ug/l  
RT: 13.223 min Scan# 1916  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion: 91 Resp: 1063581  
Ion Ratio Lower Upper  
91 100  
126 33.4 16.7 50.1





#83

tert-Butylbenzene

Concen: 145.352 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N




ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 147.626 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086867.D


Acq: 06 Jun 2025 14:49




Tgt Ion:105 Resp: 1216529

Ion Ratio Lower Upper

105 100

120 46.7 23.2 69.6





#85

sec-Butylbenzene

Concen: 144.885 ug/l

RT: 13.617 min Scan# 1983

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#86

p-Isopropyltoluene

Concen: 146.482 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:119 Resp: 1322780


Ion Ratio Lower Upper

119 100

134 26.4 13.5 40.4

91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 142.939 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N



ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 141.124 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:146 Resp: 646188




Ion Ratio Lower Upper

146 100

111 40.2 19.9 59.7

148 63.2 31.8 95.4





#89

n-Butylbenzene

Concen: 143.117 ug/l

RT: 14.053 min Scan# 2105

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#90

Hexachloroethane

Concen: 150.387 ug/l


RT: 14.335 min Scan# 2105




Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt	Ion:117	Resp:	230195
Ion	Ratio	Lower	Upper
117	100		
201	64.5	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 142.108 ug/l

RT: 14.105 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N





ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 137.242 ug/l


RT: 14.717 min Scan# 2170







Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Tgt	Ion	Resp:	110677
Ion	Ratio	Lower	Upper
75	100		
155	85.5	41.1	123.4
157	110.2	52.7	158.1





#93

1,2,4-Trichlorobenzene

Concen: 147.940 ug/l

RT: 15.394 min Scan# 2

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**


Reviewed By :John Carlone 06/09/2025







Supervised By :Mahesh Dadoda 06/09/2025



#94  
Hexachlorobutadiene  
Concen: 147.264 ug/l  
RT: 15.500 min Scan# 2303  
Delta R.T. -0.000 min  
Lab File: VN086867.D  
Acq: 06 Jun 2025 14:49

Tgt Ion:225 Resp: 151172  
Ion Ratio Lower Upper  
225 100  
223 64.8 33.5 100.4  
227 63.0 31.1 93.5





#95

Naphthalene

Concen: 150.743 ug/l

RT: 15.641 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49

Instrument:

MSVOA\_N


ClientSampleId :

VSTDICC150

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 147.743 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086867.D

Acq: 06 Jun 2025 14:49


Tgt Ion:180 Resp: 404450

Ion Ratio Lower Upper

180 100

182 94.7 47.0 141.0

145 33.6 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICVN060625

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	218122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	394140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	172710	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	152255	52.135	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 104.260%		
35) Dibromofluoromethane	8.177	113	127102	54.416	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 108.840%		
50) Toluene-d8	10.571	98	491265	53.129	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 106.260%		
62) 4-Bromofluorobenzene	12.847	95	185792	54.081	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 108.160%		
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	120237	55.286	ug/l	99
3) Chloromethane	2.401	50	139027	49.505	ug/l	99
4) Vinyl Chloride	2.554	62	151951	52.451	ug/l	99
5) Bromomethane	2.989	94	89638	55.292	ug/l	95
6) Chloroethane	3.154	64	95654	51.117	ug/l	96
7) Trichlorofluoromethane	3.524	101	196138	51.819	ug/l	97
8) Diethyl Ether	3.983	74	87405	53.001	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.395	101	127498	53.615	ug/l	99
10) Methyl Iodide	4.612	142	164080	53.227	ug/l	97
11) Tert butyl alcohol	5.542	59	207768	262.193	ug/l	99
12) 1,1-Dichloroethene	4.359	96	125461	51.678	ug/l	99
13) Acrolein	4.195	56	60336	240.545	ug/l	98
14) Allyl chloride	5.042	41	203344	50.504	ug/l	98
15) Acrylonitrile	5.736	53	479502	258.885	ug/l	100
16) Acetone	4.448	43	382146	246.750	ug/l	98
17) Carbon Disulfide	4.736	76	339466	50.550	ug/l	100
18) Methyl Acetate	5.042	43	234290	51.912	ug/l	100
19) Methyl tert-butyl Ether	5.818	73	454641	51.720	ug/l	100
20) Methylene Chloride	5.295	84	143862	49.617	ug/l	97
21) trans-1,2-Dichloroethene	5.806	96	133535	49.437	ug/l	99
22) Diisopropyl ether	6.689	45	440080	51.853	ug/l	99
23) Vinyl Acetate	6.618	43	1828919	255.055	ug/l	99
24) 1,1-Dichloroethane	6.589	63	251473	51.487	ug/l	98
25) 2-Butanone	7.494	43	644194	255.897	ug/l	99
26) 2,2-Dichloropropane	7.500	77	216678	57.031	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	164985	51.073	ug/l	99
28) Bromochloromethane	7.824	49	108810	45.310	ug/l	99
29) Tetrahydrofuran	7.847	42	420891	256.656	ug/l	100
30) Chloroform	7.977	83	248429	50.933	ug/l	97
31) Cyclohexane	8.265	56	237836	50.212	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	209927	50.603	ug/l	97
36) 1,1-Dichloropropene	8.377	75	181414	52.123	ug/l	99
37) Ethyl Acetate	7.571	43	222090	50.125	ug/l	99
38) Carbon Tetrachloride	8.371	117	179281	52.466	ug/l	98
39) Methylcyclohexane	9.606	83	251231	52.686	ug/l	98
40) Benzene	8.612	78	579667	50.931	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICVN060625

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	130903	52.607	ug/1	100
42) 1,2-Dichloroethane	8.677	62	177746	51.489	ug/1	100
43) Isopropyl Acetate	8.694	43	359378	50.536	ug/1	100
44) Trichloroethene	9.359	130	139763	51.788	ug/1	100
45) 1,2-Dichloropropane	9.624	63	142463	51.450	ug/1	99
46) Dibromomethane	9.712	93	97199	52.849	ug/1	98
47) Bromodichloromethane	9.894	83	196104	51.824	ug/1	100
48) Methyl methacrylate	9.688	41	168001	51.328	ug/1	98
49) 1,4-Dioxane	9.706	88	66102	1110.123	ug/1 #	99
51) 4-Methyl-2-Pentanone	10.447	43	1146622	267.835	ug/1	99
52) Toluene	10.635	92	361963	52.040	ug/1	99
53) t-1,3-Dichloropropene	10.835	75	224839	53.137	ug/1	99
54) cis-1,3-Dichloropropene	10.318	75	237774	52.519	ug/1	99
55) 1,1,2-Trichloroethane	11.018	97	137998	51.562	ug/1	99
56) Ethyl methacrylate	10.882	69	241649	56.688	ug/1	98
57) 1,3-Dichloropropane	11.165	76	238572	51.386	ug/1	99
58) 2-Chloroethyl Vinyl ether	10.165	63	625479	246.006	ug/1	100
59) 2-Hexanone	11.200	43	781717	283.478	ug/1	100
60) Dibromochloromethane	11.359	129	146749	52.627	ug/1	99
61) 1,2-Dibromoethane	11.471	107	142338	51.887	ug/1	99
64) Tetrachloroethene	11.106	164	111409	50.450	ug/1	98
65) Chlorobenzene	11.894	112	391661	50.892	ug/1	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	129045	52.164	ug/1	99
67) Ethyl Benzene	11.965	91	683879	51.590	ug/1	98
68) m/p-Xylenes	12.071	106	527293	103.912	ug/1	98
69) o-Xylene	12.400	106	254600	52.387	ug/1	100
70) Styrene	12.412	104	438751	52.758	ug/1	99
71) Bromoform	12.582	173	100357	54.751	ug/1 #	99
73) Isopropylbenzene	12.694	105	652045	51.823	ug/1	100
74) N-amyl acetate	12.506	43	238355	54.212	ug/1	97
75) 1,1,2,2-Tetrachloroethane	12.941	83	220643	51.763	ug/1	100
76) 1,2,3-Trichloropropane	12.994	75	220957m	53.821	ug/1	
77) Bromobenzene	12.982	156	150950	52.313	ug/1	100
78) n-propylbenzene	13.035	91	801180	52.397	ug/1	100
79) 2-Chlorotoluene	13.123	91	473629	51.650	ug/1	99
80) 1,3,5-Trimethylbenzene	13.171	105	548347	52.787	ug/1	99
81) trans-1,4-Dichloro-2-b...	12.735	75	99341	55.702	ug/1	97
82) 4-Chlorotoluene	13.223	91	481403	51.884	ug/1	99
83) tert-Butylbenzene	13.441	119	499519	52.532	ug/1	99
84) 1,2,4-Trimethylbenzene	13.482	105	552333	53.024	ug/1	100
85) sec-Butylbenzene	13.618	105	723986	52.428	ug/1	100
86) p-Isopropyltoluene	13.729	119	606979	53.174	ug/1	99
87) 1,3-Dichlorobenzene	13.735	146	290810	51.224	ug/1	100
88) 1,4-Dichlorobenzene	13.812	146	298412	51.557	ug/1	100
89) n-Butylbenzene	14.059	91	579636	52.424	ug/1	100
90) Hexachloroethane	14.335	117	102254	52.847	ug/1	100
91) 1,2-Dichlorobenzene	14.106	146	280025	51.340	ug/1	100
92) 1,2-Dibromo-3-Chloropr...	14.723	75	50115	49.161	ug/1	98
93) 1,2,4-Trichlorobenzene	15.394	180	181064	51.986	ug/1	99
94) Hexachlorobutadiene	15.500	225	68570	52.843	ug/1	99
95) Naphthalene	15.641	128	674187	52.005	ug/1	100
96) 1,2,3-Trichlorobenzene	15.841	180	174888	50.539	ug/1	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
ICVVN060625

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
Data File : VN086869.D  
Acq On : 06 Jun 2025 15:54  
Operator : JC\MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 9 Sample Multiplier: 1


Quant Time: Jun 07 02:16:19 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
ICVWN060625

## Manual Integrations APPROVED


Reviewed By :John Carbone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54




Instrument : MSVOA\_N  
ClientSampleId : ICVN060625




Tgt Ion:168 Resp: 21812  
Ion Ratio Lower Upper  
168 100  
99 58.6 49.1 73.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025









#2  
Dichlorodifluoromethane  
Concen: 55.286 ug/l  
RT: 2.154 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54



Tgt Ion: 85 Resp: 120237  
Ion Ratio Lower Upper  
85 100  
87 33.2 16.2 48.6





#3

Chloromethane

Concen: 49.505 ug/l

RT: 2.401 min Scan# 7

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N


ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#4

Vinyl Chloride

Concen: 52.451 ug/l


RT: 2.554 min Scan# 102




Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 62 Resp: 151951  
 Ion Ratio Lower Upper  
 62 100  
 64 32.0 26.0 39.0





#5

Bromomethane

Concen: 55.292 ug/l

RT: 2.989 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#6

Chloroethane

Concen: 51.117 ug/l


RT: 3.154 min Scan# 204


Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 64 Resp: 95654  
 Ion Ratio Lower Upper  
 64 100  
 66 32.6 24.6 36.8







#7

Trichlorofluoromethane  
Concen: 51.819 ug/l  
RT: 3.524 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54



Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625




Tgt Ion: 101 Resp: 196138  
Ion Ratio Lower Upper  
101 100  
103 62.9 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#8  
Diethyl Ether  
Concen: 53.001 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54




Tgt Ion: 74 Resp: 87405  
Ion Ratio Lower Upper  
74 100  
45 90.9 45.5 136.5





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 53.615 ug/l  
RT: 4.395 min Scan# 4  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54




Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625




Tgt Ion:101 Resp: 127493  
Ion Ratio Lower Upper  
101 100  
85 41.9 33.8 50.8  
151 72.8 58.7 88.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#10  
Methyl Iodide  
Concen: 53.227 ug/l  
RT: 4.612 min Scan# 452  
Delta R.T. 0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54



Tgt Ion:142 Resp: 164080  
Ion Ratio Lower Upper  
142 100  
127 36.6 30.9 46.3  
141 14.0 10.4 15.6









#11

Tert butyl alcohol  
Concen: 262.193 ug/l  
RT: 5.542 min Scan# 6  
Delta R.T. -0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#12


1,1-Dichloroethene  
Concen: 51.678 ug/l  
RT: 4.359 min Scan# 409  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion: 96 Resp: 125461  
Ion Ratio Lower Upper  
96 100  
61 158.0 125.8 188.6  
98 65.2 52.5 78.7





Abundance Scan 381 (4.195 min): VN086869.D\data.ms



#13

Acrolein

Concen: 240.545 ug/l

RT: 4.195 min Scan# 3

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#14

Allyl chloride

Concen: 50.504 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion: 41 Resp: 203344




Ion Ratio Lower Upper

41 100

39 62.3 51.3 76.9

76 40.4 31.4 47.0





#15

Acrylonitrile

Concen: 258.885 ug/l

RT: 5.736 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#16

Acetone

Concen: 246.750 ug/l

RT: 4.448 min Scan# 424


Delta R.T. -0.000 min




Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 43 Resp: 382146

Ion Ratio	Lower	Upper
43	100	
58	35.9	28.0
		42.0





#17

Carbon Disulfide

Concen: 50.550 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#18

Methyl Acetate

Concen: 51.912 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54


Tgt Ion: 43 Resp: 234290

Ion Ratio Lower Upper

43 100

74 25.9 20.7 31.1






#19

Methyl tert-butyl Ether  
Concen: 51.720 ug/l  
RT: 5.818 min Scan# 6  
Delta R.T. 0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025




#20

Methylene Chloride  
Concen: 49.617 ug/l  
RT: 5.295 min Scan# 568  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion: 84 Resp: 143862  
Ion Ratio Lower Upper  
84 100  
49 110.3 90.5 135.7  
51 31.8 28.5 42.7  
86 60.8 48.1 72.1



Abundance Scan 568 (5.295 min): VN086869.D\data.ms (-51)





#21  
trans-1,2-Dichloroethene  
Concen: 49.437 ug/l  
RT: 5.806 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625


**Manual Integrations**  
**APPROVED**







Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#22  
Diisopropyl ether  
Concen: 51.853 ug/l  
RT: 6.689 min Scan# 805  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion: 45 Resp: 440080  
Ion Ratio Lower Upper  
45 100  
43 52.4 41.9 62.9  
87 29.5 22.4 33.6  
59 12.4 9.8 14.6





#23

**Vinyl Acetate**

Concen: 255.055 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N


ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#24

**1,1-Dichloroethane**

Concen: 51.487 ug/l

RT: 6.589 min Scan# 788

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion: 63 Resp: 251473




Ion Ratio Lower Upper

63 100

98 7.7 3.6 10.8

100 4.6 2.6 7.8





#25

2-Butanone

Concen: 255.897 ug/l

RT: 7.494 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#26

2,2-Dichloropropane

Concen: 57.031 ug/l


RT: 7.500 min Scan# 943




Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 77 Resp: 216678  
 Ion Ratio Lower Upper  
 77 100  
 97 23.7 11.6 34.7





#27

cis-1,2-Dichloroethene

Concen: 51.073 ug/l

RT: 7.500 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#28

Bromochloromethane

Concen: 45.310 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion: 49 Resp: 108810




Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 78.5 62.2 93.2





#29

Tetrahydrofuran

Concen: 256.656 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54




Instrument : MSVOA\_N

ClientSampleId : ICVVN060625

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#30

Chloroform

Concen: 50.933 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54




Tgt Ion: 83 Resp: 248429

Ion Ratio Lower Upper

83 100

85 63.5 52.9 79.3





#31

Cyclohexane

Concen: 50.212 ug/l

RT: 8.265 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#32

1,1,1-Trichloroethane

Concen: 50.603 ug/l

RT: 8.177 min Scan# 1058


Delta R.T. -0.000 min


Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 97 Resp: 209927

	Ion Ratio	Lower	Upper
97	100		
99	61.7	52.0	78.0
61	45.4	35.9	53.9







#33  
 1,2-Dichloroethane-d4  
 Concen: 52.135 ug/l  
 RT: 8.588 min Scan# 1  
 Delta R.T. -0.000 min  
 Lab File: VN086869.D  
 Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
 ClientSampleId : ICVN060625


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 9.106 min Scan# 1216  
 Delta R.T. -0.000 min  
 Lab File: VN086869.D  
 Acq: 06 Jun 2025 15:54

Tgt Ion:114 Resp: 394140  
 Ion Ratio Lower Upper  
 114 100  
 63 20.2 0.0 39.6  
 88 16.4 0.0 30.2





#35

Dibromofluoromethane

Concen: 54.416 ug/l


RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54



Instrument : MSVOA\_N  
 ClientSampleId : ICVVN060625




Tgt Ion:113 Resp: 127100  
 Ion Ratio Lower Upper  
 113 100  
 111 103.4 84.2 126.2  
 192 18.0 14.2 21.4

Manual Integrations  
**APPROVED**



Reviewed By :John Carlone 06/09/2025  
 Supervised By :Mahesh Dadoda 06/09/2025









#36  
 1,1-Dichloropropene  
 Concen: 52.123 ug/l  
 RT: 8.377 min Scan# 1092  
 Delta R.T. -0.000 min  
 Lab File: VN086869.D  
 Acq: 06 Jun 2025 15:54



Tgt Ion: 75 Resp: 181414  
 Ion Ratio Lower Upper  
 75 100  
 110 35.0 17.6 52.9  
 77 31.0 25.3 37.9





#37

**Ethyl Acetate**

Concen: 50.125 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N


ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#38

**Carbon Tetrachloride**

Concen: 52.466 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion:117 Resp: 179281


Ion Ratio Lower Upper

117 100

119 94.8 75.0 112.6

121 28.5 24.9 37.3





#39

Methylcyclohexane

Concen: 52.686 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N

ClientSampleId :

ICVVN060625

Tgt Ion: 83 Resp: 25123

Ion Ratio Lower Upper

83 100

55 74.5 61.0 91.6



98 47.2 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025







Tgt Ion: 78 Resp: 579667

Ion Ratio Lower Upper

78 100

77 23.7 18.7 28.1





#41

Methacrylonitrile

Concen: 52.607 ug/l

RT: 7.789 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#42

1,2-Dichloroethane

Concen: 51.489 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 62 Resp: 177746

Ion Ratio Lower Upper

62 100

98 10.6 0.0 21.2

Time--&gt;

7.70

7.75

7.80

Time--&gt;

8.60

8.70

8.80


Time--&gt;

8.677

Time--&gt;

8.677

Time--&gt;



#43

Isopropyl Acetate

Concen: 50.536 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N

ClientSampleId :

ICVVN060625



Tgt Ion: 43 Resp: 359373

Ion Ratio Lower Upper

43 100

61 27.9 22.1 33.1






87 14.8 11.8 17.6


Manual Integrations

APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025


#44  
Trichloroethene  
Concen: 51.788 ug/l  
RT: 9.359 min Scan# 1259  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54Tgt Ion:130 Resp: 139763  
Ion Ratio Lower Upper  
130 100  
95 98.0 0.0 196.2



#45

1,2-Dichloropropane  
Concen: 51.450 ug/l  
RT: 9.624 min Scan# 1305  
Delta R.T. -0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54



Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625




Tgt Ion: 63 Resp: 142461  
Ion Ratio Lower Upper  
63 100  
65 32.7 25.5 38.3

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025






#46  
Dibromomethane  
Concen: 52.849 ug/l  
RT: 9.712 min Scan# 1319  
Delta R.T. -0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54



Tgt Ion: 93 Resp: 97199  
Ion Ratio Lower Upper  
93 100  
95 82.2 67.1 100.7  
174 90.9 74.2 111.4





#47

Bromodichloromethane

Concen: 51.824 ug/l

RT: 9.894 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#48

Methyl methacrylate

Concen: 51.328 ug/l


RT: 9.688 min Scan# 1315

Delta R.T. 0.006 min







Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt	Ion	Ion Ratio	Lower	Upper
	41	100		
	69	95.2	74.6	112.0
	39	51.0	40.2	60.4



VN086869.D 82N060625W.M



#49

1,4-Dioxane

Concen: 1110.123 ug/l

RT: 9.706 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA\_N


ClientSampleId :

ICVVN060625

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#50

Toluene-d8

Concen: 53.129 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54




Tgt Ion: 98 Resp: 491265

Ion Ratio Lower Upper

98 100

100 67.6 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 267.835 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#52

Toluene

Concen: 52.040 ug/l

RT: 10.635 min Scan# 1476

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54




Tgt Ion: 92 Resp: 361963

Ion Ratio Lower Upper

92 100

91 169.8 136.5 204.7








#53

t-1,3-Dichloropropene  
Concen: 53.137 ug/l  
RT: 10.835 min Scan# 1510  
Delta R.T. -0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025







#54

cis-1,3-Dichloropropene  
Concen: 52.519 ug/l  
RT: 10.318 min Scan# 1422  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 237774  
Ion Ratio Lower Upper  
75 100  
77 32.1 26.2 39.4  
39 41.3 32.3 48.5

1 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16





#55

1,1,2-Trichloroethane

Concen: 51.562 ug/l

RT: 11.018 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#56

Ethyl methacrylate

Concen: 56.688 ug/l

RT: 10.882 min Scan# 1518

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion: 69 Resp: 241649




Ion Ratio Lower Upper

69 100

41 57.7 47.2 70.8

39 27.4 23.0 34.6





#57

1,3-Dichloropropane

Concen: 51.386 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 246.006 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54


Tgt Ion: 63 Resp: 625479

Ion Ratio Lower Upper


63 100

106 28.9 23.0 34.6






Abundance Scan 1572 (11.200 min): VN086869.D\data.ms (-)



Abundance Scan 1572 (11.200 min): VN086869.D\data.ms (-)



#59

2-Hexanone

Concen: 283.478 ug/l

RT: 11.200 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N



ClientSampleId :

ICVVN060625


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 1599 (11.359 min): VN086869.D\data.ms (-)



Abundance Scan 1599 (11.359 min): VN086869.D\data.ms (-)



#60

Dibromochloromethane

Concen: 52.627 ug/l


RT: 11.359 min Scan# 1599




Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion:129 Resp: 146749  
Ion Ratio Lower Upper  
129 100  
127 78.1 38.6 115.8





#61

1,2-Dibromoethane

Concen: 51.887 ug/l

RT: 11.471 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#62

4-Bromofluorobenzene

Concen: 54.081 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion: 95 Resp: 185792


Ion Ratio Lower Upper

95 100

174 71.4 0.0 141.8

176 68.1 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#64  
Tetrachloroethene  
Concen: 50.450 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion:164 Resp: 111409  
Ion Ratio Lower Upper  
164 100  
166 128.5 101.8 152.8  
129 99.9 83.0 124.4  
131 97.9 78.2 117.2







#65  
Chlorobenzene  
Concen: 50.892 ug/l  
RT: 11.894 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625


**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 52.164 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion:131 Resp: 129045  
Ion Ratio Lower Upper  
131 100  
133 94.5 48.3 144.8  
119 67.4 34.0 101.8



#67

Ethyl Benzene

Concen: 51.590 ug/l

RT: 11.965 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D




Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N

ClientSampleId :





ICVVN060625



Tgt Ion: 91 Resp: 683879  
 Ion Ratio Lower Upper  
 91 100  
 106 30.9 25.4 38.2


Manual Integrations  
APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#68  
 m/p-Xylenes  
 Concen: 103.912 ug/l  
 RT: 12.071 min Scan# 1720  
 Delta R.T. -0.000 min  
 Lab File: VN086869.D  
 Acq: 06 Jun 2025 15:54

Tgt Ion: 106 Resp: 527293  
 Ion Ratio Lower Upper  
 106 100  
 91 202.0 159.4 239.0









#69  
o-Xylene  
Concen: 52.387 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument :  
MSVOA\_N  
ClientSampleId :  
ICVVN060625


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#70  
Styrene  
Concen: 52.758 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion:104 Resp: 438751  
Ion Ratio Lower Upper  
104 100  
78 48.7 38.5 57.7  
103 53.7 43.1 64.7





#71

Bromoform

Concen: 54.751 ug/l

RT: 12.582 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion:152 Resp: 172710


Ion Ratio Lower Upper

152 100


115 60.6 30.1 90.5

150 174.0 0.0 345.0






Abundance Scan 1826 (12.694 min): VN086869.D\data.ms



Abundance Scan 1826 (12.694 min): VN086869.D\data.ms (-)



#73

Isopropylbenzene

Concen: 51.823 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N


ClientSampleId :

ICVVN060625


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025


Supervised By :Mahesh Dadoda 06/09/2025



Abundance Scan 1794 (12.506 min): VN086869.D\data.ms



Abundance Scan 1794 (12.506 min): VN086869.D\data.ms (-)



#74

N-amyl acetate

Concen: 54.212 ug/l

RT: 12.506 min Scan# 1794

Delta R.T. -0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 43 Resp: 238355


Ion Ratio Lower Upper




43 100

70 49.3 37.6 56.4

55 25.3 19.0 28.6

61 25.2 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 51.763 ug/l  
RT: 12.941 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVN060625


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#76  
1,2,3-Trichloropropane  
Concen: 53.821 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 220957  
Ion Ratio Lower Upper  
75 100  
77 160.5 79.3 237.9





#77

Bromobenzene

Concen: 52.313 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#78

n-propylbenzene

Concen: 52.397 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54




Tgt Ion: 91 Resp: 801180

Ion Ratio Lower Upper

91 100

120 22.9 11.4 34.2





#79

2-Chlorotoluene

Concen: 51.650 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 52.787 ug/l



RT: 13.171 min Scan# 1907

Delta R.T. -0.000 min


Lab File: VN086869.D


Acq: 06 Jun 2025 15:54

Tgt Ion:105 Resp: 548347  
 Ion Ratio Lower Upper  
 105 100  
 120 50.6 24.9 74.6



Abundance







#81  
trans-1,4-Dichloro-2-butene  
Concen: 55.702 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Instrument : MSVOA\_N  
ClientSampleId : ICVVN060625


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/09/2025  
Supervised By :Mahesh Dadoda 06/09/2025



#82  
4-Chlorotoluene  
Concen: 51.884 ug/l  
RT: 13.223 min Scan# 1916  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion: 91 Resp: 481403  
Ion Ratio Lower Upper  
91 100  
126 33.1 16.7 50.1





#83

tert-Butylbenzene

Concen: 52.532 ug/l

RT: 13.441 min Scan# 1952

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 53.024 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086869.D


Acq: 06 Jun 2025 15:54




Tgt Ion:105 Resp: 552333

Ion Ratio Lower Upper

105 100

120 46.6 23.2 69.6





#85

sec-Butylbenzene

Concen: 52.428 ug/l

RT: 13.618 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N





ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#86

p-Isopropyltoluene

Concen: 53.174 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion:119 Resp: 606979


Ion Ratio Lower Upper

119 100

134 26.5 13.5 40.4

91 22.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 51.224 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N


ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#88

1,4-Dichlorobenzene

Concen: 51.557 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion:146 Resp: 298412

Ion Ratio Lower Upper

146 100

111 40.0 19.9 59.7

148 63.3 31.8 95.4



#89

n-Butylbenzene

Concen: 52.424 ug/l

RT: 14.059 min Scan# 2

Instrument :

Delta R.T. 0.006 min


MSVOA\_N

Lab File: VN086869.D


ClientSampleId :

Acq: 06 Jun 2025 15:54

ICVVN060625



Abundance Scan 2058 (14.059 min): VN086869.D\data.ms



Tgt Ion: 91 Resp: 579630




Ion Ratio Lower Upper

Tgt Ion	Ion Ratio	Lower	Upper
91	100		
92	55.0	27.5	82.4
134	25.5	13.0	39.0

**Manual Integrations****APPROVED**



Reviewed By :John Carlone 06/09/2025




Supervised By :Mahesh Dadoda 06/09/2025



#90  
Hexachloroethane  
Concen: 52.847 ug/l  
RT: 14.335 min Scan# 2105  
Delta R.T. -0.000 min  
Lab File: VN086869.D  
Acq: 06 Jun 2025 15:54

Tgt Ion:117 Resp: 102254  
Ion Ratio Lower Upper  
117 100  
201 66.4 33.2 99.6





#91

1,2-Dichlorobenzene

Concen: 51.340 ug/l

RT: 14.106 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N




ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 49.161 ug/l

RT: 14.723 min Scan# 2171

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt Ion: 75 Resp: 50115


Ion Ratio Lower Upper







75 100

155 83.5 41.1 123.4

157 108.6 52.7 158.1

Time--&gt; 14.00 14.10 14.20





#93

1,2,4-Trichlorobenzene

Concen: 51.986 ug/l

RT: 15.394 min Scan# 2

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument :

MSVOA\_N


ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#94

Hexachlorobutadiene

Concen: 52.843 ug/l


RT: 15.500 min Scan# 2303



Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Tgt	Ion:225	Resp:	68570
Ion	Ratio	Lower	Upper
225	100		
223	65.3	33.5	100.4
227	62.5	31.1	93.5





#95

Naphthalene

Concen: 52.005 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54

Instrument:

MSVOA\_N


ClientSampleId :

ICVVN060625

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/09/2025

Supervised By :Mahesh Dadoda 06/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 50.539 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086869.D

Acq: 06 Jun 2025 15:54


Tgt Ion:180 Resp: 174888

Ion Ratio Lower Upper

180 100

182 95.2 47.0 141.0

145 33.8 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
ICVVN060625

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.499	0.551	-10.4	116	0.00
3 P	Chloromethane	0.644	0.637	1.1	112	0.00
4 C	Vinyl Chloride	0.664	0.697	-5.0#	114	0.00
5 T	Bromomethane	0.372	0.411	-10.5	121	0.00
6 T	Chloroethane	0.429	0.439	-2.3	113	0.00
7 T	Trichlorofluoromethane	0.868	0.899	-3.6	113	0.00
8 T	Diethyl Ether	0.378	0.401	-6.1	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.585	-7.3	118	0.00
10 T	Methyl Iodide	0.707	0.752	-6.4	116	0.00
11 T	Tert butyl alcohol	0.182	0.191	-4.9	114	0.00
12 CM	1,1-Dichloroethene	0.557	0.575	-3.2#	113	0.00
13 T	Acrolein	0.057	0.055	3.5	128	0.00
14 T	Allyl chloride	0.923	0.932	-1.0	113	0.00
15 T	Acrylonitrile	0.425	0.440	-3.5	114	0.00
16 T	Acetone	0.355	0.350	1.4	114	0.00
17 T	Carbon Disulfide	1.539	1.556	-1.1	115	0.00
18 T	Methyl Acetate	1.035	1.074	-3.8	115	0.00
19 T	Methyl tert-butyl Ether	2.015	2.084	-3.4	113	0.00
20 T	Methylene Chloride	0.665	0.660	0.8	115	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.612	1.1	113	0.00
22 T	Diisopropyl ether	1.945	2.018	-3.8	114	0.00
23 T	Vinyl Acetate	1.644	1.677	-2.0	111	0.00
24 P	1,1-Dichloroethane	1.120	1.153	-2.9	114	0.00
25 T	2-Butanone	0.577	0.591	-2.4	113	0.00
26 T	2,2-Dichloropropane	0.871	0.993	-14.0	116	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.756	-2.2	114	0.00
28 T	Bromochloromethane	0.550	0.499	9.3	113	0.00
29 T	Tetrahydrofuran	0.376	0.386	-2.7	113	0.00
30 C	Chloroform	1.118	1.139	-1.9#	113	0.00
31 T	Cyclohexane	1.086	1.090	-0.4	114	0.00
32 T	1,1,1-Trichloroethane	0.951	0.962	-1.2	113	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.698	-4.3	147	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.296	0.322	-8.8	153#	0.00
36 T	1,1-Dichloropropene	0.442	0.460	-4.1	115	0.00
37 T	Ethyl Acetate	0.562	0.563	-0.2	109	0.00
38 T	Carbon Tetrachloride	0.433	0.455	-5.1	116	0.00
39 T	Methylcyclohexane	0.605	0.637	-5.3	116	0.00
40 TM	Benzene	1.444	1.471	-1.9	114	0.00
41 T	Methacrylonitrile	0.316	0.332	-5.1	114	0.00
42 TM	1,2-Dichloroethane	0.438	0.451	-3.0	114	0.00
43 T	Isopropyl Acetate	0.902	0.912	-1.1	111	0.00
44 TM	Trichloroethene	0.342	0.355	-3.8	113	0.00
45 C	1,2-Dichloropropane	0.351	0.361	-2.8#	113	0.00
46 T	Dibromomethane	0.233	0.247	-6.0	116	0.00
47 T	Bromodichloromethane	0.480	0.498	-3.8	113	0.00
48 T	Methyl methacrylate	0.415	0.426	-2.7	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**ICVVN060625**

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	112	0.00
50 S	Toluene-d8	1.173	1.246	-6.2	151#	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.582	-7.2	113	0.00
52 CM	Toluene	0.882	0.918	-4.1#	114	0.00
53 T	t-1,3-Dichloropropene	0.537	0.570	-6.1	114	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.603	-5.1	114	0.00
55 T	1,1,2-Trichloroethane	0.340	0.350	-2.9	113	0.00
56 T	Ethyl methacrylate	0.541	0.613	-13.3	115	0.00
57 T	1,3-Dichloropropane	0.589	0.605	-2.7	112	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.317	1.9	120	0.00
59 T	2-Hexanone	0.350	0.397	-13.4	112	0.00
60 T	Dibromochloromethane	0.354	0.372	-5.1	114	0.00
61 T	1,2-Dibromoethane	0.348	0.361	-3.7	114	0.00
62 S	4-Bromofluorobenzene	0.436	0.471	-8.0	151#	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.316	0.319	-0.9	114	0.00
65 PM	Chlorobenzene	1.103	1.122	-1.7	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.370	-4.5	115	0.00
67 C	Ethyl Benzene	1.899	1.960	-3.2#	114	0.00
68 T	m/p-Xylenes	0.727	0.756	-4.0	113	0.00
69 T	o-Xylene	0.696	0.730	-4.9	114	0.00
70 T	Styrene	1.192	1.257	-5.5	113	0.00
71 P	Bromoform	0.263	0.288	-9.5	114	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.643	3.775	-3.6	114	0.00
74 T	N-amyl acetate	1.273	1.380	-8.4	112	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.278	-3.6	112	0.00
76 T	1,2,3-Trichloropropane	1.189	1.279	-7.6	112	0.00
77 T	Bromobenzene	0.835	0.874	-4.7	114	0.00
78 T	n-propylbenzene	4.427	4.639	-4.8	114	0.00
79 T	2-Chlorotoluene	2.655	2.742	-3.3	113	0.00
80 T	1,3,5-Trimethylbenzene	3.007	3.175	-5.6	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.575	-11.4	125	0.00
82 T	4-Chlorotoluene	2.686	2.787	-3.8	114	0.00
83 T	tert-Butylbenzene	2.753	2.892	-5.0	114	0.00
84 T	1,2,4-Trimethylbenzene	3.016	3.198	-6.0	113	0.00
85 T	sec-Butylbenzene	3.998	4.192	-4.9	113	0.00
86 T	p-Isopropyltoluene	3.305	3.514	-6.3	114	0.00
87 T	1,3-Dichlorobenzene	1.644	1.684	-2.4	112	0.00
88 T	1,4-Dichlorobenzene	1.676	1.728	-3.1	113	0.00
89 T	n-Butylbenzene	3.201	3.356	-4.8	113	0.00
90 T	Hexachloroethane	0.560	0.592	-5.7	114	0.00
91 T	1,2-Dichlorobenzene	1.579	1.621	-2.7	111	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.290	1.7	110	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.048	-4.0	112	0.00
94 T	Hexachlorobutadiene	0.376	0.397	-5.6	113	0.00
95 T	Naphthalene	3.753	3.904	-4.0	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
Data File : VN086869.D  
Acq On : 06 Jun 2025 15:54  
Operator : JC\MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
ICVVN060625

Quant Time: Jun 07 02:16:19 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	1.002	1.013	-1.1	110	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
ICVVN060625

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	55.286	-10.6	116	0.00
3 P	Chloromethane	50.000	49.505	1.0	112	0.00
4 C	Vinyl Chloride	50.000	52.451	-4.9#	114	0.00
5 T	Bromomethane	50.000	55.292	-10.6	121	0.00
6 T	Chloroethane	50.000	51.117	-2.2	113	0.00
7 T	Trichlorofluoromethane	50.000	51.819	-3.6	113	0.00
8 T	Diethyl Ether	50.000	53.001	-6.0	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.615	-7.2	118	0.00
10 T	Methyl Iodide	50.000	53.227	-6.5	116	0.00
11 T	Tert butyl alcohol	250.000	262.193	-4.9	114	0.00
12 CM	1,1-Dichloroethene	50.000	51.678	-3.4#	113	0.00
13 T	Acrolein	250.000	240.545	3.8	128	0.00
14 T	Allyl chloride	50.000	50.504	-1.0	113	0.00
15 T	Acrylonitrile	250.000	258.885	-3.6	114	0.00
16 T	Acetone	250.000	246.750	1.3	114	0.00
17 T	Carbon Disulfide	50.000	50.550	-1.1	115	0.00
18 T	Methyl Acetate	50.000	51.912	-3.8	115	0.00
19 T	Methyl tert-butyl Ether	50.000	51.720	-3.4	113	0.00
20 T	Methylene Chloride	50.000	49.617	0.8	115	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.437	1.1	113	0.00
22 T	Diisopropyl ether	50.000	51.853	-3.7	114	0.00
23 T	Vinyl Acetate	250.000	255.055	-2.0	111	0.00
24 P	1,1-Dichloroethane	50.000	51.487	-3.0	114	0.00
25 T	2-Butanone	250.000	255.897	-2.4	113	0.00
26 T	2,2-Dichloropropane	50.000	57.031	-14.1	116	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.073	-2.1	114	0.00
28 T	Bromochloromethane	50.000	45.310	9.4	113	0.00
29 T	Tetrahydrofuran	250.000	256.656	-2.7	113	0.00
30 C	Chloroform	50.000	50.933	-1.9#	113	0.00
31 T	Cyclohexane	50.000	50.212	-0.4	114	0.00
32 T	1,1,1-Trichloroethane	50.000	50.603	-1.2	113	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.135	-4.3	147	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	54.416	-8.8	153	0.00
36 T	1,1-Dichloropropene	50.000	52.123	-4.2	115	0.00
37 T	Ethyl Acetate	50.000	50.125	-0.3	109	0.00
38 T	Carbon Tetrachloride	50.000	52.466	-4.9	116	0.00
39 T	Methylcyclohexane	50.000	52.686	-5.4	116	0.00
40 TM	Benzene	50.000	50.931	-1.9	114	0.00
41 T	Methacrylonitrile	50.000	52.607	-5.2	114	0.00
42 TM	1,2-Dichloroethane	50.000	51.489	-3.0	114	0.00
43 T	Isopropyl Acetate	50.000	50.536	-1.1	111	0.00
44 TM	Trichloroethene	50.000	51.788	-3.6	113	0.00
45 C	1,2-Dichloropropane	50.000	51.450	-2.9#	113	0.00
46 T	Dibromomethane	50.000	52.849	-5.7	116	0.00
47 T	Bromodichloromethane	50.000	51.824	-3.6	113	0.00
48 T	Methyl methacrylate	50.000	51.328	-2.7	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086869.D  
 Acq On : 06 Jun 2025 15:54  
 Operator : JC\MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
ICVVN060625

Quant Time: Jun 07 02:16:19 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1110.123	-11.0	112	0.00
50 S	Toluene-d8	50.000	53.129	-6.3	151	0.00
51 T	4-Methyl-2-Pentanone	250.000	267.835	-7.1	113	0.00
52 CM	Toluene	50.000	52.040	-4.1#	114	0.00
53 T	t-1,3-Dichloropropene	50.000	53.137	-6.3	114	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.519	-5.0	114	0.00
55 T	1,1,2-Trichloroethane	50.000	51.562	-3.1	113	0.00
56 T	Ethyl methacrylate	50.000	56.688	-13.4	115	0.00
57 T	1,3-Dichloropropane	50.000	51.386	-2.8	112	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.006	1.6	120	0.00
59 T	2-Hexanone	250.000	283.478	-13.4	112	0.00
60 T	Dibromochloromethane	50.000	52.627	-5.3	114	0.00
61 T	1,2-Dibromoethane	50.000	51.887	-3.8	114	0.00
62 S	4-Bromofluorobenzene	50.000	54.081	-8.2	151	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	50.450	-0.9	114	0.00
65 PM	Chlorobenzene	50.000	50.892	-1.8	115	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.164	-4.3	115	0.00
67 C	Ethyl Benzene	50.000	51.590	-3.2#	114	0.00
68 T	m/p-Xylenes	100.000	103.912	-3.9	113	0.00
69 T	o-Xylene	50.000	52.387	-4.8	114	0.00
70 T	Styrene	50.000	52.758	-5.5	113	0.00
71 P	Bromoform	50.000	54.751	-9.5	114	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	103	0.00
73 T	Isopropylbenzene	50.000	51.823	-3.6	114	0.00
74 T	N-amyl acetate	50.000	54.212	-8.4	112	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.763	-3.5	112	0.00
76 T	1,2,3-Trichloropropane	50.000	53.821	-7.6	112	0.00
77 T	Bromobenzene	50.000	52.313	-4.6	114	0.00
78 T	n-propylbenzene	50.000	52.397	-4.8	114	0.00
79 T	2-Chlorotoluene	50.000	51.650	-3.3	113	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.787	-5.6	113	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	55.702	-11.4	125	0.00
82 T	4-Chlorotoluene	50.000	51.884	-3.8	114	0.00
83 T	tert-Butylbenzene	50.000	52.532	-5.1	114	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.024	-6.0	113	0.00
85 T	sec-Butylbenzene	50.000	52.428	-4.9	113	0.00
86 T	p-Isopropyltoluene	50.000	53.174	-6.3	114	0.00
87 T	1,3-Dichlorobenzene	50.000	51.224	-2.4	112	0.00
88 T	1,4-Dichlorobenzene	50.000	51.557	-3.1	113	0.00
89 T	n-Butylbenzene	50.000	52.424	-4.8	113	0.00
90 T	Hexachloroethane	50.000	52.847	-5.7	114	0.00
91 T	1,2-Dichlorobenzene	50.000	51.340	-2.7	111	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.161	1.7	110	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.986	-4.0	112	0.00
94 T	Hexachlorobutadiene	50.000	52.843	-5.7	113	0.00
95 T	Naphthalene	50.000	52.005	-4.0	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
Data File : VN086869.D  
Acq On : 06 Jun 2025 15:54  
Operator : JC\MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
ICVVN060625

Quant Time: Jun 07 02:16:19 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	Dev(min)
96 T 1,2,3-Trichlorobenzene	50.000	50.539	-1.1	110	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>CHEMTECH</u>		Contract:	<u>JAC005</u>	
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2250</u>	SAS No.:	<u>Q2250</u>
Instrument ID:	<u>MSVOA_N</u>		Calibration Date/Time:	<u>06/10/2025</u>	<u>09:13</u>
Lab File ID:	<u>VN086913.D</u>		Init. Calib. Date(s):	<u>06/06/2025</u>	<u>06/06/2025</u>
Heated Purge:	(Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:44</u>	<u>14:49</u>
GC Column:	<u>RXI-624</u>	ID: <u>0.25</u> (mm)			

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.653		-1.66	20
1,1-Dichloroethene	0.557	0.548		-1.62	20
1,1-Dichloroethane	1.120	1.079	0.1	-3.66	20
cis-1,2-Dichloroethene	0.740	0.721		-2.57	20
1,1,1-Trichloroethane	0.951	0.894		-5.99	20
Benzene	1.444	1.400		-3.05	20
1,2-Dichloroethane	0.438	0.426		-2.74	20
Trichloroethene	0.342	0.345		0.88	20
1,1,2-Trichloroethane	0.340	0.338		-0.59	20
Tetrachloroethene	0.316	0.308		-2.53	20
1,2-Dichloroethane-d4	0.669	0.613		-8.37	20
Dibromofluoromethane	0.296	0.301		1.69	20
Toluene-d8	1.173	1.153		-1.71	20
4-Bromofluorobenzene	0.436	0.420		-3.67	20

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	239452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	422119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	362285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	170987	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.583	65	146800	45.789	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	91.580%	
35) Dibromofluoromethane	8.171	113	127253	50.869	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	101.740%	
50) Toluene-d8	10.565	98	486698	49.146	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	98.300%	
62) 4-Bromofluorobenzene	12.847	95	177087	48.131	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	96.260%	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	120199	50.345	ug/l	98
3) Chloromethane	2.395	50	132091	42.845	ug/l	99
4) Vinyl Chloride	2.554	62	156438	49.190	ug/l	98
5) Bromomethane	2.989	94	85611	48.104	ug/l	96
6) Chloroethane	3.154	64	100888	49.111	ug/l	96
7) Trichlorofluoromethane	3.530	101	203113	48.882	ug/l	99
8) Diethyl Ether	3.983	74	90847	50.181	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.395	101	126828	48.583	ug/l	98
10) Methyl Iodide	4.612	142	142941	42.239	ug/l	99
11) Tert butyl alcohol	5.536	59	177401	203.929	ug/l	99
12) 1,1-Dichloroethene	4.371	96	131127	49.200	ug/l	96
13) Acrolein	4.195	56	63063	229.021	ug/l	98
14) Allyl chloride	5.048	41	197016	44.573	ug/l	96
15) Acrylonitrile	5.736	53	467151	229.750	ug/l	100
16) Acetone	4.442	43	416888	245.204	ug/l	100
17) Carbon Disulfide	4.736	76	341962	46.386	ug/l	99
18) Methyl Acetate	5.042	43	225420	45.497	ug/l	98
19) Methyl tert-butyl Ether	5.812	73	478684	49.605	ug/l	99
20) Methylene Chloride	5.295	84	147723	46.410	ug/l	95
21) trans-1,2-Dichloroethene	5.806	96	140464	47.370	ug/l	97
22) Diisopropyl ether	6.683	45	439557	47.178	ug/l	98
23) Vinyl Acetate	6.618	43	1865051	236.925	ug/l	98
24) 1,1-Dichloroethane	6.589	63	258382	48.190	ug/l	100
25) 2-Butanone	7.489	43	590344	213.617	ug/l	96
26) 2,2-Dichloropropane	7.500	77	225847	54.150	ug/l	100
27) cis-1,2-Dichloroethene	7.500	96	172582	48.666	ug/l	99
28) Bromochloromethane	7.824	49	113231	42.951	ug/l	92
29) Tetrahydrofuran	7.847	42	389154	216.165	ug/l	96
30) Chloroform	7.977	83	252183	47.097	ug/l	99
31) Cyclohexane	8.265	56	217182	41.767	ug/l	94
32) 1,1,1-Trichloroethane	8.177	97	214121	47.016	ug/l	98
36) 1,1-Dichloropropene	8.377	75	184378	49.463	ug/l	99
37) Ethyl Acetate	7.571	43	220517	46.471	ug/l	98
38) Carbon Tetrachloride	8.371	117	181026	49.466	ug/l	99
39) Methylcyclohexane	9.606	83	220974	43.269	ug/l	93
40) Benzene	8.612	78	591102	48.493	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	120605	45.256	ug/l	93
42) 1,2-Dichloroethane	8.677	62	179943	48.671	ug/l	100
43) Isopropyl Acetate	8.694	43	350324	45.998	ug/l	97
44) Trichloroethene	9.359	130	145683	50.404	ug/l	96
45) 1,2-Dichloropropane	9.624	63	145320	49.004	ug/l	100
46) Dibromomethane	9.712	93	99996	50.766	ug/l	99
47) Bromodichloromethane	9.888	83	202301	49.918	ug/l	99
48) Methyl methacrylate	9.683	41	159853	45.601	ug/l	96
49) 1,4-Dioxane	9.700	88	56947	892.983	ug/l #	97
51) 4-Methyl-2-Pentanone	10.447	43	1081942	235.975	ug/l	97
52) Toluene	10.630	92	367587	49.345	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	234554	51.759	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	249744	51.506	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	142700	49.785	ug/l	98
56) Ethyl methacrylate	10.877	69	248574	54.447	ug/l	93
57) 1,3-Dichloropropane	11.165	76	244911	49.255	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	665707	244.474	ug/l	97
59) 2-Hexanone	11.200	43	715095	242.131	ug/l	93
60) Dibromochloromethane	11.359	129	153400	51.366	ug/l	99
61) 1,2-Dibromoethane	11.471	107	147800	50.307	ug/l	97
64) Tetrachloroethene	11.106	164	111601	48.675	ug/l	97
65) Chlorobenzene	11.888	112	404797	50.661	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	132787	51.699	ug/l	99
67) Ethyl Benzene	11.965	91	668179	48.548	ug/l	99
68) m/p-Xylenes	12.071	106	526722	99.975	ug/l	99
69) o-Xylene	12.394	106	258839	51.297	ug/l	95
70) Styrene	12.412	104	446437	51.704	ug/l	99
71) Bromoform	12.576	173	104086	54.693	ug/l #	100
73) Isopropylbenzene	12.694	105	621794	49.917	ug/l	99
74) N-amyl acetate	12.506	43	231590	53.204	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.935	83	216753	51.363	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	182840m	44.986	ug/l	
77) Bromobenzene	12.976	156	154418	54.054	ug/l	97
78) n-propylbenzene	13.035	91	736021	48.621	ug/l	98
79) 2-Chlorotoluene	13.124	91	443618	48.865	ug/l	97
80) 1,3,5-Trimethylbenzene	13.171	105	505593	49.161	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	91865m	52.029	ug/l	
82) 4-Chlorotoluene	13.218	91	463176	50.423	ug/l	99
83) tert-Butylbenzene	13.435	119	440956	46.840	ug/l	99
84) 1,2,4-Trimethylbenzene	13.476	105	509675	49.422	ug/l	98
85) sec-Butylbenzene	13.612	105	629702	46.060	ug/l	99
86) p-Isopropyltoluene	13.729	119	535293	47.367	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	286869	51.039	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	290947	50.774	ug/l	100
89) n-Butylbenzene	14.053	91	485793	44.380	ug/l	99
90) Hexachloroethane	14.329	117	92000	48.027	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	274399	50.816	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.718	75	47339	46.906	ug/l	93
93) 1,2,4-Trichlorobenzene	15.388	180	156993	45.529	ug/l	99
94) Hexachlorobutadiene	15.500	225	48622	37.848	ug/l	95
95) Naphthalene	15.635	128	625189	48.711	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	146962	42.897	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound R.T. QIon Response Conc Units Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1


Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC050

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13




Instrument : MSVOA\_N  
ClientSampleId : VSTDCCCC050




Tgt Ion:168 Resp: 23945  
Ion Ratio Lower Upper  
168 100  
99 57.3 49.1 73.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#2  
Dichlorodifluoromethane  
Concen: 50.345 ug/l  
RT: 2.154 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13



Tgt Ion: 85 Resp: 120199  
Ion Ratio Lower Upper  
85 100  
87 33.5 16.2 48.6









#3  
Chloromethane  
Concen: 42.845 ug/l  
RT: 2.395 min Scan# 7  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCCC050


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#4  
Vinyl Chloride  
Concen: 49.190 ug/l  
RT: 2.554 min Scan# 102  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13


Tgt Ion: 62 Resp: 156438  
Ion Ratio Lower Upper  
62 100  
64 31.2 26.0 39.0





#5  
Bromomethane  
Concen: 48.104 ug/l  
RT: 2.989 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13




Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050




Tgt Ion: 94 Resp: 8561  
Ion Ratio Lower Upper  
94 100  
96 93.0 77.9 116.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#6  
Chloroethane  
Concen: 49.111 ug/l  
RT: 3.154 min Scan# 204  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13




Tgt Ion: 64 Resp: 100888  
Ion Ratio Lower Upper  
64 100  
66 33.0 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 48.882 ug/l  
RT: 3.530 min Scan# 2  
Delta R.T. 0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13




Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050




Tgt Ion:101 Resp: 20311  
Ion Ratio Lower Upper  
101 100  
103 65.6 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#8  
Diethyl Ether  
Concen: 50.181 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13



Tgt Ion: 74 Resp: 90847  
Ion Ratio Lower Upper  
74 100  
45 88.0 45.5 136.5








#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 48.583 ug/l  
RT: 4.395 min Scan# 4  
Delta R.T. -0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#10  
Methyl Iodide  
Concen: 42.239 ug/l  
RT: 4.612 min Scan# 452  
Delta R.T. 0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:142 Resp: 142941  
Ion Ratio Lower Upper  
142 100  
127 38.6 30.9 46.3  
141 14.4 10.4 15.6









#11

Tert butyl alcohol  
Concen: 203.929 ug/l  
RT: 5.536 min Scan# 6

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050  
Acq: 10 Jun 2025 09:13

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#12



1,1-Dichloroethene  
Concen: 49.200 ug/l  
RT: 4.371 min Scan# 411  
Delta R.T. 0.012 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 96 Resp: 131127  
Ion Ratio Lower Upper  
96 100  
61 150.8 125.8 188.6  
98 63.8 52.5 78.7






Abundance Scan 381 (4.195 min): VN086913.D\data.ms



Abundance Scan 526 (5.048 min): VN086913.D\data.ms



Abundance Scan 526 (5.048 min): VN086913.D\data.ms (-47)

#13

Acrolein

Concen: 229.021 ug/l

RT: 4.195 min Scan# 3

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13



Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCCC050

**Manual Integrations  
APPROVED**

 Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025


#14

Allyl chloride


Concen: 44.573 ug/l




RT: 5.048 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

 Tgt Ion: 41 Resp: 197016  
 Ion Ratio Lower Upper  
 41 100  
 39 62.2 51.3 76.9  
 76 43.5 31.4 47.0




#15

Acrylonitrile

Concen: 229.750 ug/l

RT: 5.736 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 53 Resp: 46715

Ion Ratio Lower Upper

53 100

52 80.7 65.0 97.6




51 34.0 27.2 40.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#16

Acetone

Concen: 245.204 ug/l

RT: 4.442 min Scan# 423

Delta R.T. -0.006 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion: 43 Resp: 416888

Ion Ratio Lower Upper

43 100

58 35.0 28.0 42.0





#17

Carbon Disulfide

Concen: 46.386 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#18

Methyl Acetate

Concen: 45.497 ug/l

RT: 5.042 min Scan# 525

Delta R.T. -0.000 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13


Tgt Ion: 43 Resp: 225420

Ion Ratio Lower Upper


43 100

74 27.0 20.7 31.1







Abundance Scan 656 (5.812 min): VN086913.D\data.ms




Abundance Scan 656 (5.812 min): VN086913.D\data.ms (-60)



Abundance Scan 568 (5.295 min): VN086865.D\data.ms (-55)



Abundance Scan 568 (5.295 min): VN086913.D\data.ms (-51)




#19

Methyl tert-butyl Ether  
Concen: 49.605 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument: MSVOA\_N  
ClientSampleId: VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#20

Methylene Chloride  
Concen: 46.410 ug/l  
RT: 5.295 min Scan# 568  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 84 Resp: 147723  
Ion Ratio Lower Upper  
84 100  
49 107.5 90.5 135.7  
51 33.1 28.5 42.7  
86 63.4 48.1 72.1








#21  
trans-1,2-Dichloroethene  
Concen: 47.370 ug/l  
RT: 5.806 min Scan# 6  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#22  
Diisopropyl ether  
Concen: 47.178 ug/l  
RT: 6.683 min Scan# 804  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 45 Resp: 439557  
Ion Ratio Lower Upper  
45 100  
43 52.7 41.9 62.9  
87 30.9 22.4 33.6  
59 12.6 9.8 14.6





#23

**Vinyl Acetate**

Concen: 236.925 ug/l

RT: 6.618 min Scan# 7

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#24

**1,1-Dichloroethane**

Concen: 48.190 ug/l

RT: 6.589 min Scan# 788


Delta R.T. 0.006 min




Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt Ion: 63 Resp: 258382

Ion	Ratio	Lower	Upper
63	100		
98	7.2	3.6	10.8
100	4.9	2.6	7.8





#25

2-Butanone

Concen: 213.617 ug/l

RT: 7.489 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#26

2,2-Dichloropropane

Concen: 54.150 ug/l

RT: 7.500 min Scan# 943

Delta R.T. 0.000 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion: 77 Resp: 225847

Ion Ratio Lower Upper

77 100

97 23.2 11.6 34.7









#27

cis-1,2-Dichloroethene  
Concen: 48.666 ug/l  
RT: 7.500 min Scan# 943  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025









#28

Bromochloromethane  
Concen: 42.951 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 49 Resp: 113231  
Ion Ratio Lower Upper  
49 100  
129 1.9 0.0 3.8  
130 84.6 62.2 93.2





#29

Tetrahydrofuran

Concen: 216.165 ug/l

RT: 7.847 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#30

Chloroform

Concen: 47.097 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. 0.000 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion: 83 Resp: 252183

Ion Ratio Lower Upper

83 100

85 67.0 52.9 79.3





#31

Cyclohexane

Concen: 41.767 ug/l

RT: 8.265 min Scan# 1074

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#32

1,1,1-Trichloroethane

Concen: 47.016 ug/l


RT: 8.177 min Scan# 1058




Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt	Ion	Resp:	214121
Ion	Ratio	Lower	Upper
97	100		
99	66.9	52.0	78.0
61	43.4	35.9	53.9








#33

1,2-Dichloroethane-d4  
Concen: 45.789 ug/l  
RT: 8.583 min Scan# 1128  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument: MSVOA\_N  
ClientSampleId: VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:114 Resp: 422119  
Ion Ratio Lower Upper  
114 100  
63 18.6 0.0 39.6  
88 15.8 0.0 30.2





#35

Dibromofluoromethane

Concen: 50.869 ug/l


RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
 ClientSampleId : VSTDCCC050



Tgt Ion:113 Resp: 127251

Ion Ratio Lower Upper

113 100




111 103.6 84.2 126.2

192 17.6 14.2 21.4




Manual Integrations  
**APPROVED**




Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#36  
 1,1-Dichloropropene  
 Concen: 49.463 ug/l  
 RT: 8.377 min Scan# 1092  
 Delta R.T. 0.000 min  
 Lab File: VN086913.D  
 Acq: 10 Jun 2025 09:13





#37

**Ethyl Acetate**

Concen: 46.471 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCCC050

Tgt Ion: 43 Resp: 22051

Ion Ratio Lower Upper

43 100




61 15.0 12.1 18.1

70 13.4 9.8 14.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#38

**Carbon Tetrachloride**

Concen: 49.466 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13


Tgt Ion: 117 Resp: 181026


Ion Ratio Lower Upper

117 100

119 92.9 75.0 112.6

121 31.1 24.9 37.3





#39

Methylcyclohexane



Concen: 43.269 ug/l

RT: 9.606 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13



Tgt Ion: 83 Resp: 220974

Ion Ratio Lower Upper

83 100



55 69.9 61.0 91.6

98 45.2 38.9 58.3

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#40

Benzene


Concen: 48.493 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13








Tgt Ion: 78 Resp: 591102

Ion Ratio Lower Upper

78 100

77 23.4 18.7 28.1





#41

Methacrylonitrile

Concen: 45.256 ug/l

RT: 7.783 min Scan# 991

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N

ClientSampleId :





VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

Abundance



#42

1,2-Dichloroethane

Concen: 48.671 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13


Tgt Ion: 62 Resp: 179943


Ion Ratio Lower Upper

62 100

98 10.7 0.0 21.2

Abundance





#43

Isopropyl Acetate



Concen: 45.998 ug/l

RT: 8.694 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13



Tgt Ion: 43 Resp: 350324

Ion Ratio Lower Upper

43 100



61 29.4 22.1 33.1

87 15.9 11.8 17.6

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#44

Trichloroethene


Concen: 50.404 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13





Tgt Ion:130 Resp: 145683

Ion Ratio Lower Upper


130 100

95 93.7 0.0 196.2





Sub


m/z--&gt;



Abundance Scan 1304 (9.624 min): VN086913.D\data.ms






Abundance Scan 1319 (9.712 min): VN086913.D\data.ms



Abundance Scan 1319 (9.712 min): VN086913.D\data.ms (-1)

Sub



#47

Bromodichloromethane

Concen: 49.918 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#48

Methyl methacrylate

Concen: 45.601 ug/l


RT: 9.683 min Scan# 1314




Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt	Ion	Ion	Resp:	159853
		Ratio	Lower	Upper
	41	100		
	69	99.2	74.6	112.0
	39	50.8	40.2	60.4





#49

1,4-Dioxane

Concen: 892.983 ug/l

RT: 9.700 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N



ClientSampleId :

VSTDCCC050


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025



Supervised By :Mahesh Dadoda 06/11/2025






#50  
Toluene-d8  
Concen: 49.146 ug/l  
RT: 10.565 min Scan# 1464  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13



Tgt Ion: 98 Resp: 486698  
Ion Ratio Lower Upper  
98 100  
100 67.9 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 235.975 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#52

Toluene

Concen: 49.345 ug/l

RT: 10.630 min Scan# 1475

Delta R.T. -0.006 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion: 92 Resp: 367587

Ion Ratio Lower Upper

92 100

91 167.7 136.5 204.7









#53  
t-1,3-Dichloropropene  
Concen: 51.759 ug/l  
RT: 10.835 min Scan# 1510  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#54  
cis-1,3-Dichloropropene  
Concen: 51.506 ug/l  
RT: 10.312 min Scan# 1421  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 75 Resp: 249744  
Ion Ratio Lower Upper  
75 100  
77 32.3 26.2 39.4  
39 39.4 32.3 48.5







#55  
1,1,2-Trichloroethane  
Concen: 49.785 ug/l  
RT: 11.018 min Scan# 1541  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carbone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#56  
Ethyl methacrylate  
Concen: 54.447 ug/l  
RT: 10.877 min Scan# 1517  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 69 Resp: 248574  
Ion Ratio Lower Upper  
69 100  
41 54.6 47.2 70.8  
39 24.1 23.0 34.6









#57

1,3-Dichloropropane  
Concen: 49.255 ug/l  
RT: 11.165 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#58

2-Chloroethyl Vinyl ether  
Concen: 244.474 ug/l  
RT: 10.159 min Scan# 1395  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 63 Resp: 665707  
Ion Ratio Lower Upper  
63 100  
106 30.3 23.0 34.6









#59  
2-Hexanone  
Concen: 242.131 ug/l  
RT: 11.200 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#60  
Dibromochloromethane  
Concen: 51.366 ug/l  
RT: 11.359 min Scan# 1599  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:129 Resp: 153400  
Ion Ratio Lower Upper  
129 100  
127 78.2 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 50.307 ug/l  
RT: 11.471 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDCCC050


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#62  
4-Bromofluorobenzene  
Concen: 48.131 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 95 Resp: 177087  
Ion Ratio Lower Upper  
95 100  
174 72.9 0.0 141.8  
176 69.1 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations


**APPROVED**




Reviewed By :John Carbone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#64  
Tetrachloroethene  
Concen: 48.675 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:164 Resp: 111601  
Ion Ratio Lower Upper  
164 100  
166 127.6 101.8 152.8  
129 97.4 83.0 124.4  
131 94.3 78.2 117.2








#65  
Chlorobenzene  
Concen: 50.661 ug/l  
RT: 11.888 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050



1 Manual Integrations  
2 APPROVED


3 Reviewed By :John Carlone 06/11/2025  
4 Supervised By :Mahesh Dadoda 06/11/2025




#66  
1,1,1,2-Tetrachloroethane  
Concen: 51.699 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:131 Resp: 132787  
Ion Ratio Lower Upper  
131 100  
133 94.8 48.3 144.8  
119 67.2 34.0 101.8





Abundance Scan 1702 (11.965 min): VN086913.D\data.ms (-)



#67

Ethyl Benzene

Concen: 48.548 ug/l

RT: 11.965 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N



ClientSampleId :

VSTDCCC050


**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



Abundance Scan 1720 (12.071 min): VN086913.D\data.ms (-)



#68

m/p-Xylenes

Concen: 99.975 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion:106 Resp: 526722

Ion Ratio Lower Upper

106 100

91 198.0 159.4 239.0









#69  
o-Xylene  
Concen: 51.297 ug/l  
RT: 12.394 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDCCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#70  
Styrene  
Concen: 51.704 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:104 Resp: 446437  
Ion Ratio Lower Upper  
104 100  
78 46.9 38.5 57.7  
103 53.0 43.1 64.7





#71

Bromoform

Concen: 54.693 ug/l

RT: 12.576 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13


Tgt Ion:152 Resp: 170987




Ion Ratio Lower Upper

152 100

115 59.8 30.1 90.5

150 174.7 0.0 345.0





#73

Isopropylbenzene

Concen: 49.917 ug/l

RT: 12.694 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#74

N-amyl acetate

Concen: 53.204 ug/l

RT: 12.506 min Scan# 1794

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt Ion: 43 Resp: 231590


Ion Ratio Lower Upper




43 100

70 52.9 37.6 56.4

55 27.1 19.0 28.6

61 23.3 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 51.363 ug/l  
RT: 12.935 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#76  
1,2,3-Trichloropropane  
Concen: 44.986 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 75 Resp: 182840  
Ion Ratio Lower Upper  
75 100  
77 192.1 79.3 237.9





#77

Bromobenzene

Concen: 54.054 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carbone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#78

n-propylbenzene

Concen: 48.621 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion: 91 Resp: 736021

Ion Ratio Lower Upper

91 100

120 23.6 11.4 34.2





#79

2-Chlorotoluene

Concen: 48.865 ug/l

RT: 13.124 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carbone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#80

1,3,5-Trimethylbenzene

Concen: 49.161 ug/l

RT: 13.171 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN086913.D


Acq: 10 Jun 2025 09:13




Tgt Ion:105 Resp: 505593

Ion Ratio Lower Upper

105 100

120 51.4 24.9 74.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 52.029 ug/l m  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#82  
4-Chlorotoluene  
Concen: 50.423 ug/l  
RT: 13.218 min Scan# 1915  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion: 91 Resp: 463176  
Ion Ratio Lower Upper  
91 100  
126 33.7 16.7 50.1





#83

tert-Butylbenzene

Concen: 46.840 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**



Reviewed By :John Carlone 06/11/2025




Supervised By :Mahesh Dadoda 06/11/2025



#84  
1,2,4-Trimethylbenzene  
Concen: 49.422 ug/l  
RT: 13.476 min Scan# 1959  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:105 Resp: 509675  
Ion Ratio Lower Upper  
105 100  
120 47.5 23.2 69.6









#85  
sec-Butylbenzene  
Concen: 46.060 ug/l  
RT: 13.612 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#86  
p-Isopropyltoluene  
Concen: 47.367 ug/l  
RT: 13.729 min Scan# 2002  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:119 Resp: 535293  
Ion Ratio Lower Upper  
119 100  
134 26.2 13.5 40.4  
91 21.7 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 51.039 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#88

1,4-Dichlorobenzene

Concen: 50.774 ug/l


RT: 13.812 min Scan# 2016




Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt	Ion:146	Resp:	290947
Ion	Ratio	Lower	Upper
146	100		
111	40.0	19.9	59.7
148	63.4	31.8	95.4





#89

n-Butylbenzene

Concen: 44.380 ug/l

RT: 14.053 min Scan# 2

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#90

Hexachloroethane

Concen: 48.027 ug/l


RT: 14.329 min Scan# 2104




Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt Ion:117 Resp: 92000  
 Ion Ratio Lower Upper  
 117 100  
 201 65.8 33.2 99.6





#91

1,2-Dichlorobenzene

Concen: 50.816 ug/l

RT: 14.106 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 46.906 ug/l


RT: 14.718 min Scan# 2170




Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
	75	100			
	155	87.0	47339	41.1	123.4
	157	114.2		52.7	158.1








#93

1,2,4-Trichlorobenzene  
Concen: 45.529 ug/l  
RT: 15.388 min Scan# 2284  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025









#94

Hexachlorobutadiene  
Concen: 37.848 ug/l  
RT: 15.500 min Scan# 2303  
Delta R.T. 0.000 min  
Lab File: VN086913.D  
Acq: 10 Jun 2025 09:13

Tgt Ion:225 Resp: 48622  
Ion Ratio Lower Upper  
225 100  
223 62.1 33.5 100.4  
227 59.7 31.1 93.5





#95

Naphthalene

Concen: 48.711 ug/l

RT: 15.635 min Scan# 2360

Delta R.T. -0.006 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#96

1,2,3-Trichlorobenzene

Concen: 42.897 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN086913.D

Acq: 10 Jun 2025 09:13


Tgt Ion:180 Resp: 146962

Ion Ratio Lower Upper

180 100

182 93.7 47.0 141.0

145 33.9 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	115	0.00
2 T	Dichlorodifluoromethane	0.499	0.502	-0.6	116	0.00
3 P	Chloromethane	0.644	0.552	14.3	107	0.00
4 C	Vinyl Chloride	0.664	0.653	1.7#	118	0.00
5 T	Bromomethane	0.372	0.358	3.8	116	0.00
6 T	Chloroethane	0.429	0.421	1.9	119	0.00
7 T	Trichlorofluoromethane	0.868	0.848	2.3	118	0.00
8 T	Diethyl Ether	0.378	0.379	-0.3	121	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.530	2.8	118	0.00
10 T	Methyl Iodide	0.707	0.597	15.6	101	0.00
11 T	Tert butyl alcohol	0.182	0.148	18.7	97	-0.01
12 CM	1,1-Dichloroethene	0.557	0.548	1.6#	119	0.01
13 T	Acrolein	0.057	0.053	7.0	134	0.00
14 T	Allyl chloride	0.923	0.823	10.8	110	0.00
15 T	Acrylonitrile	0.425	0.390	8.2	111	0.00
16 T	Acetone	0.355	0.348	2.0	125	0.00
17 T	Carbon Disulfide	1.539	1.428	7.2	116	0.00
18 T	Methyl Acetate	1.035	0.941	9.1	110	0.00
19 T	Methyl tert-butyl Ether	2.015	1.999	0.8	119	0.00
20 T	Methylene Chloride	0.665	0.617	7.2	118	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.587	5.2	119	0.00
22 T	Diisopropyl ether	1.945	1.836	5.6	114	0.00
23 T	Vinyl Acetate	1.644	1.558	5.2	113	0.00
24 P	1,1-Dichloroethane	1.120	1.079	3.7	117	0.00
25 T	2-Butanone	0.577	0.493	14.6	103	0.00
26 T	2,2-Dichloropropane	0.871	0.943	-8.3	120	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.721	2.6	119	0.00
28 T	Bromochloromethane	0.550	0.473	14.0	117	0.00
29 T	Tetrahydrofuran	0.376	0.325	13.6	105	0.00
30 C	Chloroform	1.118	1.053	5.8#	115	0.00
31 T	Cyclohexane	1.086	0.907	16.5	104	0.00
32 T	1,1,1-Trichloroethane	0.951	0.894	6.0	115	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.613	8.4	142	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
35 S	Dibromofluoromethane	0.296	0.301	-1.7	153#	0.00
36 T	1,1-Dichloropropene	0.442	0.437	1.1	116	0.00
37 T	Ethyl Acetate	0.562	0.522	7.1	109	0.00
38 T	Carbon Tetrachloride	0.433	0.429	0.9	117	0.00
39 T	Methylcyclohexane	0.605	0.523	13.6	102	0.00
40 TM	Benzene	1.444	1.400	3.0	116	0.00
41 T	Methacrylonitrile	0.316	0.286	9.5	105	0.00
42 TM	1,2-Dichloroethane	0.438	0.426	2.7	116	0.00
43 T	Isopropyl Acetate	0.902	0.830	8.0	109	0.00
44 TM	Trichloroethene	0.342	0.345	-0.9	118	0.00
45 C	1,2-Dichloropropane	0.351	0.344	2.0#	116	0.00
46 T	Dibromomethane	0.233	0.237	-1.7	119	0.00
47 T	Bromodichloromethane	0.480	0.479	0.2	117	0.00
48 T	Methyl methacrylate	0.415	0.379	8.7	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	96	0.00
50 S	Toluene-d8	1.173	1.153	1.7	149	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.513	5.5	106	0.00
52 CM	Toluene	0.882	0.871	1.2#	116	0.00
53 T	t-1,3-Dichloropropene	0.537	0.556	-3.5	119	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.592	-3.1	120	0.00
55 T	1,1,2-Trichloroethane	0.340	0.338	0.6	117	0.00
56 T	Ethyl methacrylate	0.541	0.589	-8.9	118	0.00
57 T	1,3-Dichloropropane	0.589	0.580	1.5	115	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.315	2.5	128	0.00
59 T	2-Hexanone	0.350	0.339	3.1	103	0.00
60 T	Dibromochloromethane	0.354	0.363	-2.5	119	0.00
61 T	1,2-Dibromoethane	0.348	0.350	-0.6	119	0.00
62 S	4-Bromofluorobenzene	0.436	0.420	3.7	144	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.316	0.308	2.5	114	0.00
65 PM	Chlorobenzene	1.103	1.117	-1.3	119	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.367	-3.7	118	0.00
67 C	Ethyl Benzene	1.899	1.844	2.9#	112	0.00
68 T	m/p-Xylenes	0.727	0.727	0.0	113	0.00
69 T	o-Xylene	0.696	0.714	-2.6	115	0.00
70 T	Styrene	1.192	1.232	-3.4	115	0.00
71 P	Bromoform	0.263	0.287	-9.1	118	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.643	3.636	0.2	108	0.00
74 T	N-amyl acetate	1.273	1.354	-6.4	109	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.268	-2.8	110	0.00
76 T	1,2,3-Trichloropropane	1.189	1.069	10.1	93	0.00
77 T	Bromobenzene	0.835	0.903	-8.1	117	0.00
78 T	n-propylbenzene	4.427	4.305	2.8	104	0.00
79 T	2-Chlorotoluene	2.655	2.594	2.3	106	0.00
80 T	1,3,5-Trimethylbenzene	3.007	2.957	1.7	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.537	-4.1	115	0.00
82 T	4-Chlorotoluene	2.686	2.709	-0.9	109	0.00
83 T	tert-Butylbenzene	2.753	2.579	6.3	101	0.00
84 T	1,2,4-Trimethylbenzene	3.016	2.981	1.2	104	0.00
85 T	sec-Butylbenzene	3.998	3.683	7.9	98	0.00
86 T	p-Isopropyltoluene	3.305	3.131	5.3	100	0.00
87 T	1,3-Dichlorobenzene	1.644	1.678	-2.1	110	0.00
88 T	1,4-Dichlorobenzene	1.676	1.702	-1.6	110	0.00
89 T	n-Butylbenzene	3.201	2.841	11.2	95	0.00
90 T	Hexachloroethane	0.560	0.538	3.9	102	0.00
91 T	1,2-Dichlorobenzene	1.579	1.605	-1.6	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.277	6.1	104	0.00
93 T	1,2,4-Trichlorobenzene	1.008	0.918	8.9	97	0.00
94 T	Hexachlorobutadiene	0.376	0.284	24.5	80	0.00
95 T	Naphthalene	3.753	3.656	2.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086913.D  
Acq On : 10 Jun 2025 09:13  
Operator : JC\MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	1.002	0.859	14.3	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	115	0.00
2 T	Dichlorodifluoromethane	50.000	50.345	-0.7	116	0.00
3 P	Chloromethane	50.000	42.845	14.3	107	0.00
4 C	Vinyl Chloride	50.000	49.190	1.6#	118	0.00
5 T	Bromomethane	50.000	48.104	3.8	116	0.00
6 T	Chloroethane	50.000	49.111	1.8	119	0.00
7 T	Trichlorofluoromethane	50.000	48.882	2.2	118	0.00
8 T	Diethyl Ether	50.000	50.181	-0.4	121	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.583	2.8	118	0.00
10 T	Methyl Iodide	50.000	42.239	15.5	101	0.00
11 T	Tert butyl alcohol	250.000	203.929	18.4	97	-0.01
12 CM	1,1-Dichloroethene	50.000	49.200	1.6#	119	0.01
13 T	Acrolein	250.000	229.021	8.4	134	0.00
14 T	Allyl chloride	50.000	44.573	10.9	110	0.00
15 T	Acrylonitrile	250.000	229.750	8.1	111	0.00
16 T	Acetone	250.000	245.204	1.9	125	0.00
17 T	Carbon Disulfide	50.000	46.386	7.2	116	0.00
18 T	Methyl Acetate	50.000	45.497	9.0	110	0.00
19 T	Methyl tert-butyl Ether	50.000	49.605	0.8	119	0.00
20 T	Methylene Chloride	50.000	46.410	7.2	118	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.370	5.3	119	0.00
22 T	Diisopropyl ether	50.000	47.178	5.6	114	0.00
23 T	Vinyl Acetate	250.000	236.925	5.2	113	0.00
24 P	1,1-Dichloroethane	50.000	48.190	3.6	117	0.00
25 T	2-Butanone	250.000	213.617	14.6	103	0.00
26 T	2,2-Dichloropropane	50.000	54.150	-8.3	120	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.666	2.7	119	0.00
28 T	Bromochloromethane	50.000	42.951	14.1	117	0.00
29 T	Tetrahydrofuran	250.000	216.165	13.5	105	0.00
30 C	Chloroform	50.000	47.097	5.8#	115	0.00
31 T	Cyclohexane	50.000	41.767	16.5	104	0.00
32 T	1,1,1-Trichloroethane	50.000	47.016	6.0	115	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.789	8.4	142	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	111	0.00
35 S	Dibromofluoromethane	50.000	50.869	-1.7	153	0.00
36 T	1,1-Dichloropropene	50.000	49.463	1.1	116	0.00
37 T	Ethyl Acetate	50.000	46.471	7.1	109	0.00
38 T	Carbon Tetrachloride	50.000	49.466	1.1	117	0.00
39 T	Methylcyclohexane	50.000	43.269	13.5	102	0.00
40 TM	Benzene	50.000	48.493	3.0	116	0.00
41 T	Methacrylonitrile	50.000	45.256	9.5	105	0.00
42 TM	1,2-Dichloroethane	50.000	48.671	2.7	116	0.00
43 T	Isopropyl Acetate	50.000	45.998	8.0	109	0.00
44 TM	Trichloroethene	50.000	50.404	-0.8	118	0.00
45 C	1,2-Dichloropropane	50.000	49.004	2.0#	116	0.00
46 T	Dibromomethane	50.000	50.766	-1.5	119	0.00
47 T	Bromodichloromethane	50.000	49.918	0.2	117	0.00
48 T	Methyl methacrylate	50.000	45.601	8.8	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086913.D  
 Acq On : 10 Jun 2025 09:13  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	892.983	10.7	96	0.00
50 S	Toluene-d8	50.000	49.146	1.7	149	0.00
51 T	4-Methyl-2-Pentanone	250.000	235.975	5.6	106	0.00
52 CM	Toluene	50.000	49.345	1.3#	116	0.00
53 T	t-1,3-Dichloropropene	50.000	51.759	-3.5	119	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.506	-3.0	120	0.00
55 T	1,1,2-Trichloroethane	50.000	49.785	0.4	117	0.00
56 T	Ethyl methacrylate	50.000	54.447	-8.9	118	0.00
57 T	1,3-Dichloropropane	50.000	49.255	1.5	115	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.474	2.2	128	0.00
59 T	2-Hexanone	250.000	242.131	3.1	103	0.00
60 T	Dibromochloromethane	50.000	51.366	-2.7	119	0.00
61 T	1,2-Dibromoethane	50.000	50.307	-0.6	119	0.00
62 S	4-Bromofluorobenzene	50.000	48.131	3.7	144	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	109	0.00
64 T	Tetrachloroethene	50.000	48.675	2.7	114	0.00
65 PM	Chlorobenzene	50.000	50.661	-1.3	119	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.699	-3.4	118	0.00
67 C	Ethyl Benzene	50.000	48.548	2.9#	112	0.00
68 T	m/p-Xylenes	100.000	99.975	0.0	113	0.00
69 T	o-Xylene	50.000	51.297	-2.6	115	0.00
70 T	Styrene	50.000	51.704	-3.4	115	0.00
71 P	Bromoform	50.000	54.693	-9.4	118	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	49.917	0.2	108	0.00
74 T	N-amyl acetate	50.000	53.204	-6.4	109	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.363	-2.7	110	0.00
76 T	1,2,3-Trichloropropane	50.000	44.986	10.0	93	0.00
77 T	Bromobenzene	50.000	54.054	-8.1	117	0.00
78 T	n-propylbenzene	50.000	48.621	2.8	104	0.00
79 T	2-Chlorotoluene	50.000	48.865	2.3	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.161	1.7	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.029	-4.1	115	0.00
82 T	4-Chlorotoluene	50.000	50.423	-0.8	109	0.00
83 T	tert-Butylbenzene	50.000	46.840	6.3	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.422	1.2	104	0.00
85 T	sec-Butylbenzene	50.000	46.060	7.9	98	0.00
86 T	p-Isopropyltoluene	50.000	47.367	5.3	100	0.00
87 T	1,3-Dichlorobenzene	50.000	51.039	-2.1	110	0.00
88 T	1,4-Dichlorobenzene	50.000	50.774	-1.5	110	0.00
89 T	n-Butylbenzene	50.000	44.380	11.2	95	0.00
90 T	Hexachloroethane	50.000	48.027	3.9	102	0.00
91 T	1,2-Dichlorobenzene	50.000	50.816	-1.6	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.906	6.2	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	45.529	8.9	97	0.00
94 T	Hexachlorobutadiene	50.000	37.848	24.3	80	0.00
95 T	Naphthalene	50.000	48.711	2.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086913.D  
Acq On : 10 Jun 2025 09:13  
Operator : JC\MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Jun 11 01:47:05 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)
96 T 1,2,3-Trichlorobenzene	50.000	42.897	14.2	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>CHEMTECH</u>		Contract:	<u>JAC005</u>	
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2250</u>	SAS No.:	<u>Q2250</u>
Instrument ID:	<u>MSVOA_N</u>		Calibration Date/Time:	<u>06/11/2025</u>	<u>11:32</u>
Lab File ID:	<u>VN086940.D</u>		Init. Calib. Date(s):	<u>06/06/2025</u>	<u>06/06/2025</u>
Heated Purge:	(Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:44</u>	<u>14:49</u>
GC Column:	<u>RXI-624</u>	ID: <u>0.25</u> (mm)			

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.701		5.57	20
1,1-Dichloroethene	0.557	0.593		6.46	20
1,1-Dichloroethane	1.120	1.178	0.1	5.18	20
cis-1,2-Dichloroethene	0.740	0.793		7.16	20
1,1,1-Trichloroethane	0.951	0.976		2.63	20
Benzene	1.444	1.554		7.55	20
1,2-Dichloroethane	0.438	0.472		7.76	20
Trichloroethene	0.342	0.382		11.7	20
1,1,2-Trichloroethane	0.340	0.383		12.65	20
Tetrachloroethene	0.316	0.337		6.65	20
1,2-Dichloroethane-d4	0.669	0.622		-7.03	20
Dibromofluoromethane	0.296	0.309		4.39	20
Toluene-d8	1.173	1.181		0.68	20
4-Bromofluorobenzene	0.436	0.440		0.92	20

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	205098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	152136	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.583	65	127629	46.478	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	92.960%	
35) Dibromofluoromethane	8.177	113	110532	52.151	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	104.300%	
50) Toluene-d8	10.565	98	422391	50.342	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	100.680%	
62) 4-Bromofluorobenzene	12.847	95	157388	50.489	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	100.980%	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	110255	53.916	ug/l	100
3) Chloromethane	2.395	50	113731	43.069	ug/l	94
4) Vinyl Chloride	2.554	62	143755	52.773	ug/l	99
5) Bromomethane	2.989	94	63914	41.928	ug/l	96
6) Chloroethane	3.154	64	93702	53.253	ug/l	99
7) Trichlorofluoromethane	3.524	101	190657	53.570	ug/l	99
8) Diethyl Ether	3.983	74	85722	55.281	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.401	101	119322	53.364	ug/l	99
10) Methyl Iodide	4.618	142	115825	39.959	ug/l	97
11) Tert butyl alcohol	5.542	59	183049	245.667	ug/l	99
12) 1,1-Dichloroethene	4.365	96	121690	53.307	ug/l	94
13) Acrolein	4.201	56	68006	288.340	ug/l	99
14) Allyl chloride	5.048	41	187099	49.420	ug/l	96
15) Acrylonitrile	5.736	53	456903	262.349	ug/l	99
16) Acetone	4.442	43	393802	270.423	ug/l	97
17) Carbon Disulfide	4.736	76	308706	48.889	ug/l	100
18) Methyl Acetate	5.042	43	222452	52.419	ug/l	98
19) Methyl tert-butyl Ether	5.812	73	445459	53.894	ug/l	98
20) Methylene Chloride	5.295	84	138117	50.660	ug/l	95
21) trans-1,2-Dichloroethene	5.806	96	129921	51.154	ug/l	98
22) Diisopropyl ether	6.683	45	412097	51.639	ug/l	99
23) Vinyl Acetate	6.618	43	1766271	261.960	ug/l	98
24) 1,1-Dichloroethane	6.589	63	241575	52.602	ug/l	99
25) 2-Butanone	7.489	43	582343	246.018	ug/l	95
26) 2,2-Dichloropropane	7.506	77	215080	60.206	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	162658	53.551	ug/l	98
28) Bromochloromethane	7.824	49	118461	52.462	ug/l	95
29) Tetrahydrofuran	7.847	42	380103	246.502	ug/l	96
30) Chloroform	7.977	83	241191	52.589	ug/l	98
31) Cyclohexane	8.265	56	208644	46.846	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	200139	51.307	ug/l	97
36) 1,1-Dichloropropene	8.377	75	172642	54.664	ug/l	99
37) Ethyl Acetate	7.571	43	211755	52.670	ug/l	99
38) Carbon Tetrachloride	8.371	117	168940	54.485	ug/l	94
39) Methylcyclohexane	9.606	83	206018	47.613	ug/l	96
40) Benzene	8.612	78	555597	53.798	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	115191	51.017	ug/1	93
42) 1,2-Dichloroethane	8.677	62	168630	53.833	ug/1	99
43) Isopropyl Acetate	8.694	43	338399	52.442	ug/1	97
44) Trichloroethene	9.353	130	136748	55.842	ug/1	98
45) 1,2-Dichloropropane	9.624	63	137250	54.626	ug/1	100
46) Dibromomethane	9.712	93	94620	56.697	ug/1	99
47) Bromodichloromethane	9.888	83	190362	55.440	ug/1	100
48) Methyl methacrylate	9.683	41	154138	51.898	ug/1	96
49) 1,4-Dioxane	9.700	88	60969	1128.409	ug/1 #	96
51) 4-Methyl-2-Pentanone	10.447	43	1054797	271.529	ug/1	98
52) Toluene	10.630	92	349873	55.435	ug/1	98
53) t-1,3-Dichloropropene	10.835	75	217752	56.714	ug/1	100
54) cis-1,3-Dichloropropene	10.312	75	236296	57.518	ug/1	97
55) 1,1,2-Trichloroethane	11.018	97	137099	56.454	ug/1	99
56) Ethyl methacrylate	10.877	69	231226	59.778	ug/1	95
57) 1,3-Dichloropropane	11.165	76	232746	55.247	ug/1	98
58) 2-Chloroethyl Vinyl ether	10.159	63	665780	288.579	ug/1	99
59) 2-Hexanone	11.200	43	705660	282.011	ug/1	97
60) Dibromochloromethane	11.359	129	148069	58.520	ug/1	99
61) 1,2-Dibromoethane	11.471	107	137779	55.351	ug/1	99
64) Tetrachloroethene	11.106	164	105084	53.190	ug/1	95
65) Chlorobenzene	11.888	112	381598	55.424	ug/1	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	124964	56.463	ug/1	99
67) Ethyl Benzene	11.965	91	642857	54.206	ug/1	99
68) m/p-Xylenes	12.071	106	500895	110.334	ug/1	99
69) o-Xylene	12.400	106	244038	56.127	ug/1	98
70) Styrene	12.412	104	424060	56.996	ug/1	100
71) Bromoform	12.576	173	99616	60.747	ug/1 #	100
73) Isopropylbenzene	12.694	105	593080	53.511	ug/1	99
74) N-amyl acetate	12.512	43	208229	53.765	ug/1	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	213329	56.816	ug/1	99
76) 1,2,3-Trichloropropane	12.994	75	179987m	49.771	ug/1	
77) Bromobenzene	12.976	156	146181	57.512	ug/1	98
78) n-propylbenzene	13.035	91	717012	53.234	ug/1	99
79) 2-Chlorotoluene	13.123	91	439522	54.413	ug/1	98
80) 1,3,5-Trimethylbenzene	13.171	105	490727	53.628	ug/1	98
81) trans-1,4-Dichloro-2-b...	12.735	75	93790	59.701	ug/1 #	87
82) 4-Chlorotoluene	13.218	91	446237	54.598	ug/1	99
83) tert-Butylbenzene	13.435	119	448106	53.498	ug/1	97
84) 1,2,4-Trimethylbenzene	13.482	105	495936	54.048	ug/1	99
85) sec-Butylbenzene	13.612	105	617792	50.788	ug/1	99
86) p-Isopropyltoluene	13.729	119	524440	52.156	ug/1	98
87) 1,3-Dichlorobenzene	13.729	146	280357	56.062	ug/1	99
88) 1,4-Dichlorobenzene	13.812	146	287122	56.315	ug/1	99
89) n-Butylbenzene	14.053	91	477306	49.007	ug/1	100
90) Hexachloroethane	14.329	117	89886	52.738	ug/1	99
91) 1,2-Dichlorobenzene	14.106	146	267932	55.766	ug/1	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	47326	52.704	ug/1	92
93) 1,2,4-Trichlorobenzene	15.388	180	160030	52.161	ug/1	99
94) Hexachlorobutadiene	15.500	225	51073	44.682	ug/1	99
95) Naphthalene	15.635	128	624421	54.679	ug/1	100
96) 1,2,3-Trichlorobenzene	15.835	180	151411	49.672	ug/1	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
(#)						

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1




Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC050

### Manual Integrations APPROVED

Reviewed By :John Carlane 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025









#1

Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#2

Dichlorodifluoromethane  
Concen: 53.916 ug/l  
RT: 2.154 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 85 Resp: 110255  
Ion Ratio Lower Upper  
85 100  
87 32.4 16.2 48.6









#3  
Chloromethane  
Concen: 43.069 ug/l  
RT: 2.395 min Scan# 7  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#4  
Vinyl Chloride  
Concen: 52.773 ug/l  
RT: 2.554 min Scan# 102  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32


Tgt Ion: 62 Resp: 143755  
Ion Ratio Lower Upper  
62 100  
64 31.8 26.0 39.0





#5  
Bromomethane  
Concen: 41.928 ug/l  
RT: 2.989 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32




Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050




Tgt Ion: 94 Resp: 63914  
Ion Ratio Lower Upper  
94 100  
96 93.9 77.9 116.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025




#6  
Chloroethane  
Concen: 53.253 ug/l  
RT: 3.154 min Scan# 204  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32




Tgt Ion: 64 Resp: 93702  
Ion Ratio Lower Upper  
64 100  
66 31.2 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 53.570 ug/l  
RT: 3.524 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32




Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050




Tgt Ion:101 Resp: 19065  
Ion Ratio Lower Upper  
101 100  
103 64.5 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025









#8  
Diethyl Ether  
Concen: 55.281 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32



Tgt Ion: 74 Resp: 85722  
Ion Ratio Lower Upper  
74 100  
45 85.9 45.5 136.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 53.364 ug/l

RT: 4.401 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025

Abundance

30000

20000

10000

0

Time--> 4.30 4.40 4.50

#10

Methyl Iodide

Concen: 39.959 ug/l

RT: 4.618 min Scan# 453

Delta R.T. 0.012 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt Ion:142 Resp: 115825

Ion Ratio Lower Upper

142 100

127 36.0 30.9 46.3

141 13.1 10.4 15.6

Abundance




30000

20000

10000

0

Time--> 4.50 4.60 4.70







#11

Tert butyl alcohol  
Concen: 245.667 ug/l  
RT: 5.542 min Scan# 6  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#12

1,1-Dichloroethene  
Concen: 53.307 ug/l  
RT: 4.365 min Scan# 410  
Delta R.T. 0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 96 Resp: 121690  
Ion Ratio Lower Upper  
96 100  
61 148.3 125.8 188.6  
98 62.3 52.5 78.7





#13

Acrolein

Concen: 288.340 ug/l

RT: 4.201 min Scan# 3

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#14

Allyl chloride

Concen: 49.420 ug/l

RT: 5.048 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion: 41 Resp: 187099







Ion Ratio Lower Upper

41 100

39 62.0 51.3 76.9

76 42.2 31.4 47.0





#15

Acrylonitrile

Concen: 262.349 ug/l

RT: 5.736 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 53 Resp: 45690

Ion Ratio Lower Upper

53 100

52 80.2 65.0 97.6


51 33.6 27.2 40.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#16

Acetone

Concen: 270.423 ug/l

RT: 4.442 min Scan# 423

Delta R.T. -0.006 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32




Tgt Ion: 43 Resp: 393802

Ion Ratio Lower Upper

43 100

58 36.9 28.0 42.0





#17

Carbon Disulfide

Concen: 48.889 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#18

Methyl Acetate

Concen: 52.419 ug/l

RT: 5.042 min Scan# 525


Delta R.T. -0.000 min


Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt Ion: 43 Resp: 222452

Ion	Ratio	Lower	Upper
43	100		
74	26.7	20.7	31.1






#19

Methyl tert-butyl Ether  
Concen: 53.894 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32




Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050




Tgt Ion: 73 Resp: 445459  
Ion Ratio Lower Upper  
73 100  
57 22.7 17.4 26.2

Manual Integrations  
APPROVED



Reviewed By :John Carbone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#20  
Methylene Chloride  
Concen: 50.660 ug/l  
RT: 5.295 min Scan# 568  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32



Tgt Ion: 84 Resp: 138117  
Ion Ratio Lower Upper  
84 100  
49 106.1 90.5 135.7  
51 34.5 28.5 42.7  
86 62.4 48.1 72.1








#21  
trans-1,2-Dichloroethene  
Concen: 51.154 ug/l  
RT: 5.806 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#22  
Diisopropyl ether  
Concen: 51.639 ug/l  
RT: 6.683 min Scan# 804  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 45 Resp: 412097  
Ion Ratio Lower Upper  
45 100  
43 52.4 41.9 62.9  
87 30.3 22.4 33.6  
59 12.3 9.8 14.6





#23

**Vinyl Acetate**

Concen: 261.960 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#24

**1,1-Dichloroethane**

Concen: 52.602 ug/l

RT: 6.589 min Scan# 788

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion: 63 Resp: 241575




Ion Ratio Lower Upper

63 100

98 7.4 3.6 10.8

100 4.5 2.6 7.8









#25  
2-Butanone  
Concen: 246.018 ug/l  
RT: 7.489 min Scan# 9  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#26  
2,2-Dichloropropane  
Concen: 60.206 ug/l  
RT: 7.506 min Scan# 944  
Delta R.T. 0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 77 Resp: 215080  
Ion Ratio Lower Upper  
77 100  
97 23.4 11.6 34.7









#27  
cis-1,2-Dichloroethene  
Concen: 53.551 ug/l  
RT: 7.500 min Scan# 943  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
Client SampleId : VSTDCCC050


### Manual Integrations APPROVED







Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#28  
Bromochloromethane  
Concen: 52.462 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 49 Resp: 118461  
Ion Ratio Lower Upper  
49 100  
129 2.2 0.0 3.8  
130 82.4 62.2 93.2





#29

Tetrahydrofuran

Concen: 246.502 ug/l

RT: 7.847 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#30

Chloroform

Concen: 52.589 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32




Tgt Ion: 83 Resp: 241191

Ion Ratio Lower Upper

83 100

85 64.5 52.9 79.3





#31

Cyclohexane

Concen: 46.846 ug/l

RT: 8.265 min Scan# 1074

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#32

1,1,1-Trichloroethane

Concen: 51.307 ug/l

RT: 8.177 min Scan# 1058


Delta R.T. -0.000 min




Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt Ion: 97 Resp: 200139

Ion	Ratio	Lower	Upper
97	100		
99	62.4	52.0	78.0
61	43.5	35.9	53.9








#33

1,2-Dichloroethane-d4  
Concen: 46.478 ug/l  
RT: 8.583 min Scan# 1128  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument: MSVOA\_N  
ClientSampleId: VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025









#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:114 Resp: 357643  
Ion Ratio Lower Upper  
114 100  
63 20.1 0.0 39.6  
88 16.2 0.0 30.2





#35

Dibromofluoromethane

Concen: 52.151 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCC050

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#36

1,1-Dichloropropene

Concen: 54.664 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion: 75 Resp: 172642




Ion Ratio Lower Upper

75 100

110 35.6 17.6 52.9

77 31.1 25.3 37.9





#37

Ethyl Acetate

Concen: 52.670 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 21175

Ion Ratio Lower Upper

43 100

61 15.2 12.1 18.1




70 12.8 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#38

Carbon Tetrachloride

Concen: 54.485 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion:117 Resp: 168940







Ion Ratio Lower Upper

117 100

119 100.6 75.0 112.6

121 30.4 24.9 37.3





#39

Methylcyclohexane

Concen: 47.613 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#40

Benzene

Concen: 53.798 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32




Tgt Ion: 78 Resp: 555597

Ion Ratio Lower Upper

78 100

77 24.3 18.7 28.1





#41

Methacrylonitrile

Concen: 51.017 ug/l

RT: 7.789 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :





VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carbone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025

Abundance



#42

1,2-Dichloroethane

Concen: 53.833 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion: 62 Resp: 168630


Ion Ratio Lower Upper

62 100

98 11.0 0.0 21.2

Abundance





#43

Isopropyl Acetate

Concen: 52.442 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 43 Resp: 338395

Ion Ratio Lower Upper

43 100




61 29.2 22.1 33.1

87 15.7 11.8 17.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#44

Trichloroethene

Concen: 55.842 ug/l

RT: 9.353 min Scan# 1258

Delta R.T. -0.006 min

Lab File: VN086940.D



Acq: 11 Jun 2025 11:32

Tgt Ion: 130 Resp: 136748


Ion Ratio Lower Upper


130 100

95 96.1 0.0 196.2




Sub








Abundance Scan 1304 (9.624 min): VN086940.D\data.ms




Abundance Scan 1304 (9.624 min): VN086940.D\data.ms (-1)



Abundance Scan 1319 (9.712 min): VN086940.D\data.ms



Abundance Scan 1319 (9.712 min): VN086940.D\data.ms (-1)



#45

1,2-Dichloropropane

Concen: 54.626 ug/l

RT: 9.624 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#46

Dibromomethane

Concen: 56.697 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion: 93 Resp: 94620


Ion Ratio Lower Upper

93 100

95 82.9 67.1 100.7

174 94.1 74.2 111.4





#47

Bromodichloromethane




Concen: 55.440 ug/l

RT: 9.888 min Scan# 1350

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32



#48

Methyl methacrylate






Concen: 51.898 ug/l




RT: 9.683 min Scan# 1314

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32





#49

1,4-Dioxane

Concen: 1128.409 ug/l

RT: 9.700 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N



ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**




Reviewed By :John Carlone 06/12/2025




Supervised By :Semsettin Yesilyurt 06/12/2025



#50  
Toluene-d8  
Concen: 50.342 ug/l  
RT: 10.565 min Scan# 1464  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 98 Resp: 422391  
Ion Ratio Lower Upper  
98 100  
100 68.2 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 271.529 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N




ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#52

Toluene

Concen: 55.435 ug/l

RT: 10.630 min Scan# 1475

Delta R.T. -0.006 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32


Tgt Ion: 92 Resp: 349873

Ion Ratio Lower Upper


92 100

91 167.2 136.5 204.7






Abundance Scan 1510 (10.835 min): VN086940.D\data.ms (-)



Abundance Scan 1510 (10.835 min): VN086940.D\data.ms (-)



#53

t-1,3-Dichloropropene

Concen: 56.714 ug/l

RT: 10.835 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N



ClientSampleId :

VSTDCCC050


**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025


Supervised By :Semsettin Yesilyurt 06/12/2025



Abundance Scan 1421 (10.312 min): VN086940.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086940.D\data.ms (-)



#54

cis-1,3-Dichloropropene

Concen: 57.518 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion: 75 Resp: 236296


Ion Ratio Lower Upper

75 100

77 31.6 26.2 39.4

39 38.5 32.3 48.5






#55  
1,1,2-Trichloroethane  
Concen: 56.454 ug/l  
RT: 11.018 min Scan# 1541  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050



### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#56  
Ethyl methacrylate  
Concen: 59.778 ug/l  
RT: 10.877 min Scan# 1517  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 69 Resp: 231226  
Ion Ratio Lower Upper  
69 100  
41 55.6 47.2 70.8  
39 26.0 23.0 34.6









#57

1,3-Dichloropropane  
Concen: 55.247 ug/l  
RT: 11.165 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#58

2-Chloroethyl Vinyl ether  
Concen: 288.579 ug/l  
RT: 10.159 min Scan# 1395  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 63 Resp: 665780  
Ion Ratio Lower Upper  
63 100  
106 29.5 23.0 34.6









#59  
2-Hexanone  
Concen: 282.011 ug/l  
RT: 11.200 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument: MSVOA\_N  
ClientSampleId: VSTDCCC050


1 Manual Integrations  
2 APPROVED




3 Reviewed By :John Carlone 06/12/2025  
4 Supervised By :Semsettin Yesilyurt 06/12/2025



#60  
Dibromochloromethane  
Concen: 58.520 ug/l  
RT: 11.359 min Scan# 1599  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:129 Resp: 148069  
Ion Ratio Lower Upper  
129 100  
127 76.4 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 55.351 ug/l  
RT: 11.471 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


Manual Integrations  
**APPROVED**


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#62  
4-Bromofluorobenzene  
Concen: 50.489 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 95 Resp: 157388  
Ion Ratio Lower Upper  
95 100  
174 71.7 0.0 141.8  
176 68.2 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#64  
Tetrachloroethene  
Concen: 53.190 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:164 Resp: 105084  
Ion Ratio Lower Upper  
164 100  
166 124.8 101.8 152.8  
129 95.6 83.0 124.4  
131 93.8 78.2 117.2








#65  
Chlorobenzene  
Concen: 55.424 ug/l  
RT: 11.888 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

Manual Integrations  
**APPROVED**

Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#67

Ethyl Benzene

Concen: 54.206 ug/l

RT: 11.965 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :




VSTDCCC050



Tgt Ion: 91 Resp: 642851  
 Ion Ratio Lower Upper  
 91 100  
 106 32.5 25.4 38.2



### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025



#68  
 m/p-Xylenes  
 Concen: 110.334 ug/l  
 RT: 12.071 min Scan# 1720  
 Delta R.T. -0.000 min  
 Lab File: VN086940.D  
 Acq: 11 Jun 2025 11:32

Tgt Ion:106 Resp: 500895  
 Ion Ratio Lower Upper  
 106 100  
 91 198.0 159.4 239.0









#69  
o-Xylene  
Concen: 56.127 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument :  
MSVOA\_N  
ClientSampleId :  
VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#70  
Styrene  
Concen: 56.996 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:104 Resp: 424060  
Ion Ratio Lower Upper  
104 100  
78 47.4 38.5 57.7  
103 53.9 43.1 64.7









#71  
Bromoform  
Concen: 60.747 ug/l  
RT: 12.576 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2012  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:152 Resp: 152136  
Ion Ratio Lower Upper  
152 100  
115 60.1 30.1 90.5  
150 176.5 0.0 345.0





#73

Isopropylbenzene

Concen: 53.511 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#74

N-amyl acetate

Concen: 53.765 ug/l

RT: 12.512 min Scan# 1795

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt Ion: 43 Resp: 208229


Ion Ratio Lower Upper




43 100

70 52.6 37.6 56.4

55 26.6 19.0 28.6

61 26.4 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 56.816 ug/l  
RT: 12.935 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#76  
1,2,3-Trichloropropane  
Concen: 49.771 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 179987  
Ion Ratio Lower Upper  
75 100  
77 187.3 79.3 237.9





#77

Bromobenzene

Concen: 57.512 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carbone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#78

n-propylbenzene

Concen: 53.234 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086940.D


Acq: 11 Jun 2025 11:32

Tgt Ion: 91 Resp: 717012

Ion Ratio Lower Upper

91 100

120 23.3 11.4 34.2



#79

2-Chlorotoluene

Concen: 54.413 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :



VSTDCCC050




Tgt Ion: 91 Resp: 43952  
Ion Ratio Lower Upper  
91 100  
126 33.4 16.2 48.5

**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#80  
1,3,5-Trimethylbenzene  
Concen: 53.628 ug/l  
RT: 13.171 min Scan# 1907  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32



Tgt Ion:105 Resp: 490727  
Ion Ratio Lower Upper  
105 100  
120 50.8 24.9 74.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 59.701 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument: MSVOA\_N  
ClientSampleId: VSTDCCC050


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#82  
4-Chlorotoluene  
Concen: 54.598 ug/l  
RT: 13.218 min Scan# 1915  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 91 Resp: 446237  
Ion Ratio Lower Upper  
91 100  
126 34.0 16.7 50.1





#83

tert-Butylbenzene

Concen: 53.498 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#84

1,2,4-Trimethylbenzene

Concen: 54.048 ug/l


RT: 13.482 min Scan# 1960




Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
105	100				
120	47.2	23.2	495936	69.6	









#85  
sec-Butylbenzene  
Concen: 50.788 ug/l  
RT: 13.612 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#86  
p-Isopropyltoluene  
Concen: 52.156 ug/l  
RT: 13.729 min Scan# 2002  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:119 Resp: 524440  
Ion Ratio Lower Upper  
119 100  
134 26.4 13.5 40.4  
91 21.8 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 56.062 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N





ClientSampleId :

VSTDCCC050

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#88

1,4-Dichlorobenzene

Concen: 56.315 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion:146 Resp: 287122

Ion Ratio Lower Upper

146 100

111 39.5 19.9 59.7

148 62.7 31.8 95.4



#89

n-Butylbenzene

Concen: 49.007 ug/l

RT: 14.053 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 91 Resp: 477300

Ion Ratio Lower Upper

91 100




92 54.9 27.5 82.4

134 26.6 13.0 39.0

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025





#90  
Hexachloroethane  
Concen: 52.738 ug/l  
RT: 14.329 min Scan# 2104  
Delta R.T. -0.006 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32




Tgt Ion:117 Resp: 89886

Ion Ratio Lower Upper

117 100

201 67.6 33.2 99.6









#91

1,2-Dichlorobenzene  
Concen: 55.766 ug/l  
RT: 14.106 min Scan# 2170  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Instrument : MSVOA\_N  
ClientSampleId : VSTDCCC050

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#92

1,2-Dibromo-3-Chloropropane  
Concen: 52.704 ug/l  
RT: 14.717 min Scan# 2170  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion: 75 Resp: 47326  
Ion Ratio Lower Upper  
75 100  
155 87.3 41.1 123.4  
157 114.9 52.7 158.1



#93

1,2,4-Trichlorobenzene

Concen: 52.161 ug/l

RT: 15.388 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086940.D




Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N

ClientSampleId :

VSTDCCC050







Tgt Ion:180 Resp: 160030  
Ion Ratio Lower Upper

180 100  
182 92.1 46.9 140.7  
145 31.7 16.0 47.9


**Manual Integrations**  
**APPROVED**







Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#94  
Hexachlorobutadiene  
Concen: 44.682 ug/l  
RT: 15.500 min Scan# 2303  
Delta R.T. -0.000 min  
Lab File: VN086940.D  
Acq: 11 Jun 2025 11:32

Tgt Ion:225 Resp: 51073  
Ion Ratio Lower Upper  
225 100  
223 65.0 33.5 100.4  
227 62.2 31.1 93.5





#95

Naphthalene

Concen: 54.679 ug/l

RT: 15.635 min Scan# 2360

Delta R.T. -0.006 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32

Instrument:

MSVOA\_N


ClientSampleId :

VSTDCCC050

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#96

1,2,3-Trichlorobenzene

Concen: 49.672 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN086940.D

Acq: 11 Jun 2025 11:32


Tgt Ion:180 Resp: 151411

Ion Ratio Lower Upper

180 100

182 94.5 47.0 141.0

145 33.0 16.6 49.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.499	0.538	-7.8	106	0.00
3 P	Chloromethane	0.644	0.555	13.8	92	0.00
4 C	Vinyl Chloride	0.664	0.701	-5.6#	108	0.00
5 T	Bromomethane	0.372	0.312	16.1	86	0.00
6 T	Chloroethane	0.429	0.457	-6.5	111	0.00
7 T	Trichlorofluoromethane	0.868	0.930	-7.1	110	0.00
8 T	Diethyl Ether	0.378	0.418	-10.6	114	0.00
9 T	1,1,2-Trichlorotrifluoroethane	0.545	0.582	-6.8	111	0.00
10 T	Methyl Iodide	0.707	0.565	20.1	82	0.01
11 T	Tert butyl alcohol	0.182	0.178	2.2	100	0.00
12 CM	1,1-Dichloroethene	0.557	0.593	-6.5#	110	0.00
13 T	Acrolein	0.057	0.066	-15.8	145	0.00
14 T	Allyl chloride	0.923	0.912	1.2	104	0.00
15 T	Acrylonitrile	0.425	0.446	-4.9	108	0.00
16 T	Acetone	0.355	0.384	-8.2	118	0.00
17 T	Carbon Disulfide	1.539	1.505	2.2	104	0.00
18 T	Methyl Acetate	1.035	1.085	-4.8	109	0.00
19 T	Methyl tert-butyl Ether	2.015	2.172	-7.8	111	0.00
20 T	Methylene Chloride	0.665	0.673	-1.2	110	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.633	-2.3	110	0.00
22 T	Diisopropyl ether	1.945	2.009	-3.3	107	0.00
23 T	Vinyl Acetate	1.644	1.722	-4.7	107	0.00
24 P	1,1-Dichloroethane	1.120	1.178	-5.2	110	0.00
25 T	2-Butanone	0.577	0.568	1.6	102	0.00
26 T	2,2-Dichloropropane	0.871	1.049	-20.4	115	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.793	-7.2	112	0.00
28 T	Bromochloromethane	0.550	0.578	-5.1	122	0.00
29 T	Tetrahydrofuran	0.376	0.371	1.3	102	0.00
30 C	Chloroform	1.118	1.176	-5.2#	110	0.00
31 T	Cyclohexane	1.086	1.017	6.4	100	0.00
32 T	1,1,1-Trichloroethane	0.951	0.976	-2.6	108	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.622	7.0	123	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.296	0.309	-4.4	133	0.00
36 T	1,1-Dichloropropene	0.442	0.483	-9.3	109	0.00
37 T	Ethyl Acetate	0.562	0.592	-5.3	104	0.00
38 T	Carbon Tetrachloride	0.433	0.472	-9.0	109	0.00
39 T	Methylcyclohexane	0.605	0.576	4.8	95	0.00
40 TM	Benzene	1.444	1.553	-7.5	109	0.00
41 T	Methacrylonitrile	0.316	0.322	-1.9	100	0.00
42 TM	1,2-Dichloroethane	0.438	0.472	-7.8	108	0.00
43 T	Isopropyl Acetate	0.902	0.946	-4.9	105	0.00
44 TM	Trichloroethene	0.342	0.382	-11.7	110	0.00
45 C	1,2-Dichloropropane	0.351	0.384	-9.4#	109	0.00
46 T	Dibromomethane	0.233	0.265	-13.7	113	0.00
47 T	Bromodichloromethane	0.480	0.532	-10.8	110	0.00
48 T	Methyl methacrylate	0.415	0.431	-3.9	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	103	0.00
50 S	Toluene-d8	1.173	1.181	-0.7	130	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.590	-8.7	104	0.00
52 CM	Toluene	0.882	0.978	-10.9#	111	0.00
53 T	t-1,3-Dichloropropene	0.537	0.609	-13.4	110	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.661	-15.2	113	0.00
55 T	1,1,2-Trichloroethane	0.340	0.383	-12.6	112	0.00
56 T	Ethyl methacrylate	0.541	0.647	-19.6	110	0.00
57 T	1,3-Dichloropropane	0.589	0.651	-10.5	110	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.372	-15.2	128	0.00
59 T	2-Hexanone	0.350	0.395	-12.9	101	0.00
60 T	Dibromochloromethane	0.354	0.414	-16.9	115	0.00
61 T	1,2-Dibromoethane	0.348	0.385	-10.6	111	0.00
62 S	4-Bromofluorobenzene	0.436	0.440	-0.9	128	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.316	0.337	-6.6	108	0.00
65 PM	Chlorobenzene	1.103	1.222	-10.8	112	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.400	-13.0	111	0.00
67 C	Ethyl Benzene	1.899	2.059	-8.4#	108	0.00
68 T	m/p-Xylenes	0.727	0.802	-10.3	107	0.00
69 T	o-Xylene	0.696	0.782	-12.4	109	0.00
70 T	Styrene	1.192	1.358	-13.9	110	0.00
71 P	Bromoform	0.263	0.319	-21.3	113	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.643	3.898	-7.0	103	0.00
74 T	N-amyl acetate	1.273	1.369	-7.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.402	-13.6	108	0.00
76 T	1,2,3-Trichloropropane	1.189	1.183	0.5	92	0.00
77 T	Bromobenzene	0.835	0.961	-15.1	110	0.00
78 T	n-propylbenzene	4.427	4.713	-6.5	102	0.00
79 T	2-Chlorotoluene	2.655	2.889	-8.8	105	0.00
80 T	1,3,5-Trimethylbenzene	3.007	3.226	-7.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.616	-19.4	118	0.00
82 T	4-Chlorotoluene	2.686	2.933	-9.2	105	0.00
83 T	tert-Butylbenzene	2.753	2.945	-7.0	102	0.00
84 T	1,2,4-Trimethylbenzene	3.016	3.260	-8.1	102	0.00
85 T	sec-Butylbenzene	3.998	4.061	-1.6	96	0.00
86 T	p-Isopropyltoluene	3.305	3.447	-4.3	98	0.00
87 T	1,3-Dichlorobenzene	1.644	1.843	-12.1	108	0.00
88 T	1,4-Dichlorobenzene	1.676	1.887	-12.6	109	0.00
89 T	n-Butylbenzene	3.201	3.137	2.0	93	0.00
90 T	Hexachloroethane	0.560	0.591	-5.5	100	0.00
91 T	1,2-Dichlorobenzene	1.579	1.761	-11.5	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.311	-5.4	104	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.052	-4.4	99	0.00
94 T	Hexachlorobutadiene	0.376	0.336	10.6	84	0.00
95 T	Naphthalene	3.753	4.104	-9.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
Data File : VN086940.D  
Acq On : 11 Jun 2025 11:32  
Operator : JC\MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	1.002	0.995	0.7	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	50.000	53.916	-7.8	106	0.00
3 P	Chloromethane	50.000	43.069	13.9	92	0.00
4 C	Vinyl Chloride	50.000	52.773	-5.5#	108	0.00
5 T	Bromomethane	50.000	41.928	16.1	86	0.00
6 T	Chloroethane	50.000	53.253	-6.5	111	0.00
7 T	Trichlorofluoromethane	50.000	53.570	-7.1	110	0.00
8 T	Diethyl Ether	50.000	55.281	-10.6	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.364	-6.7	111	0.00
10 T	Methyl Iodide	50.000	39.959	20.1	82	0.01
11 T	Tert butyl alcohol	250.000	245.667	1.7	100	0.00
12 CM	1,1-Dichloroethene	50.000	53.307	-6.6#	110	0.00
13 T	Acrolein	250.000	288.340	-15.3	145	0.00
14 T	Allyl chloride	50.000	49.420	1.2	104	0.00
15 T	Acrylonitrile	250.000	262.349	-4.9	108	0.00
16 T	Acetone	250.000	270.423	-8.2	118	0.00
17 T	Carbon Disulfide	50.000	48.889	2.2	104	0.00
18 T	Methyl Acetate	50.000	52.419	-4.8	109	0.00
19 T	Methyl tert-butyl Ether	50.000	53.894	-7.8	111	0.00
20 T	Methylene Chloride	50.000	50.660	-1.3	110	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.154	-2.3	110	0.00
22 T	Diisopropyl ether	50.000	51.639	-3.3	107	0.00
23 T	Vinyl Acetate	250.000	261.960	-4.8	107	0.00
24 P	1,1-Dichloroethane	50.000	52.602	-5.2	110	0.00
25 T	2-Butanone	250.000	246.018	1.6	102	0.00
26 T	2,2-Dichloropropane	50.000	60.206	-20.4	115	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.551	-7.1	112	0.00
28 T	Bromochloromethane	50.000	52.462	-4.9	122	0.00
29 T	Tetrahydrofuran	250.000	246.502	1.4	102	0.00
30 C	Chloroform	50.000	52.589	-5.2#	110	0.00
31 T	Cyclohexane	50.000	46.846	6.3	100	0.00
32 T	1,1,1-Trichloroethane	50.000	51.307	-2.6	108	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.478	7.0	123	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	52.151	-4.3	133	0.00
36 T	1,1-Dichloropropene	50.000	54.664	-9.3	109	0.00
37 T	Ethyl Acetate	50.000	52.670	-5.3	104	0.00
38 T	Carbon Tetrachloride	50.000	54.485	-9.0	109	0.00
39 T	Methylcyclohexane	50.000	47.613	4.8	95	0.00
40 TM	Benzene	50.000	53.798	-7.6	109	0.00
41 T	Methacrylonitrile	50.000	51.017	-2.0	100	0.00
42 TM	1,2-Dichloroethane	50.000	53.833	-7.7	108	0.00
43 T	Isopropyl Acetate	50.000	52.442	-4.9	105	0.00
44 TM	Trichloroethene	50.000	55.842	-11.7	110	0.00
45 C	1,2-Dichloropropane	50.000	54.626	-9.3#	109	0.00
46 T	Dibromomethane	50.000	56.697	-13.4	113	0.00
47 T	Bromodichloromethane	50.000	55.440	-10.9	110	0.00
48 T	Methyl methacrylate	50.000	51.898	-3.8	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086940.D  
 Acq On : 11 Jun 2025 11:32  
 Operator : JC\MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1128.409	-12.8	103	0.00
50 S	Toluene-d8	50.000	50.342	-0.7	130	0.00
51 T	4-Methyl-2-Pentanone	250.000	271.529	-8.6	104	0.00
52 CM	Toluene	50.000	55.435	-10.9#	111	0.00
53 T	t-1,3-Dichloropropene	50.000	56.714	-13.4	110	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.518	-15.0	113	0.00
55 T	1,1,2-Trichloroethane	50.000	56.454	-12.9	112	0.00
56 T	Ethyl methacrylate	50.000	59.778	-19.6	110	0.00
57 T	1,3-Dichloropropane	50.000	55.247	-10.5	110	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	288.579	-15.4	128	0.00
59 T	2-Hexanone	250.000	282.011	-12.8	101	0.00
60 T	Dibromochloromethane	50.000	58.520	-17.0	115	0.00
61 T	1,2-Dibromoethane	50.000	55.351	-10.7	111	0.00
62 S	4-Bromofluorobenzene	50.000	50.489	-1.0	128	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	53.190	-6.4	108	0.00
65 PM	Chlorobenzene	50.000	55.424	-10.8	112	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	56.463	-12.9	111	0.00
67 C	Ethyl Benzene	50.000	54.206	-8.4#	108	0.00
68 T	m/p-Xylenes	100.000	110.334	-10.3	107	0.00
69 T	o-Xylene	50.000	56.127	-12.3	109	0.00
70 T	Styrene	50.000	56.996	-14.0	110	0.00
71 P	Bromoform	50.000	60.747	-21.5	113	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	91	0.00
73 T	Isopropylbenzene	50.000	53.511	-7.0	103	0.00
74 T	N-amyl acetate	50.000	53.765	-7.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	56.816	-13.6	108	0.00
76 T	1,2,3-Trichloropropane	50.000	49.771	0.5	92	0.00
77 T	Bromobenzene	50.000	57.512	-15.0	110	0.00
78 T	n-propylbenzene	50.000	53.234	-6.5	102	0.00
79 T	2-Chlorotoluene	50.000	54.413	-8.8	105	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.628	-7.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	59.701	-19.4	118	0.00
82 T	4-Chlorotoluene	50.000	54.598	-9.2	105	0.00
83 T	tert-Butylbenzene	50.000	53.498	-7.0	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.048	-8.1	102	0.00
85 T	sec-Butylbenzene	50.000	50.788	-1.6	96	0.00
86 T	p-Isopropyltoluene	50.000	52.156	-4.3	98	0.00
87 T	1,3-Dichlorobenzene	50.000	56.062	-12.1	108	0.00
88 T	1,4-Dichlorobenzene	50.000	56.315	-12.6	109	0.00
89 T	n-Butylbenzene	50.000	49.007	2.0	93	0.00
90 T	Hexachloroethane	50.000	52.738	-5.5	100	0.00
91 T	1,2-Dichlorobenzene	50.000	55.766	-11.5	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.704	-5.4	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.161	-4.3	99	0.00
94 T	Hexachlorobutadiene	50.000	44.682	10.6	84	0.00
95 T	Naphthalene	50.000	54.679	-9.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
Data File : VN086940.D  
Acq On : 11 Jun 2025 11:32  
Operator : JC\MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Jun 12 01:28:53 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	50.000	49.672	0.7	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



# QC SAMPLE


# DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060625\  
 Data File : VN086861.D  
 Acq On : 06 Jun 2025 07:59  
 Operator : JC\MD  
 Sample : BFB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Title : SW846 8260  
 Last Update : Sat Jun 07 02:12:50 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	11610	PASS
75	95	30	60	48.1	32235	PASS
95	95	100	100	100.0	67037	PASS
96	95	5	9	6.4	4284	PASS
173	174	0.00	2	1.0	453	PASS
174	95	50	100	66.6	44632	PASS
175	174	5	9	7.1	3184	PASS
176	174	95	101	98.1	43805	PASS
177	176	5	9	6.8	2979	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086912.D  
 Acq On : 10 Jun 2025 08:39  
 Operator : JC\MD  
 Sample : BFB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Title : SW846 8260  
 Last Update : Sat Jun 07 02:12:50 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	12262	PASS
75	95	30	60	47.3	36032	PASS
95	95	100	100	100.0	76240	PASS
96	95	5	9	7.0	5316	PASS
173	174	0.00	2	0.8	431	PASS
174	95	50	100	73.1	55768	PASS
175	174	5	9	7.2	4025	PASS
176	174	95	101	97.5	54389	PASS
177	176	5	9	6.9	3729	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086939.D  
 Acq On : 11 Jun 2025 10:22  
 Operator : JC\MD  
 Sample : BFB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Title : SW846 8260  
 Last Update : Sat Jun 07 02:12:50 2025



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	10335	PASS
75	95	30	60	46.5	30525	PASS
95	95	100	100	100.0	65627	PASS
96	95	5	9	7.1	4656	PASS
173	174	0.00	2	1.0	446	PASS
174	95	50	100	71.1	46675	PASS
175	174	5	9	7.5	3520	PASS
176	174	95	101	96.8	45165	PASS
177	176	5	9	7.0	3156	PASS



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:
Client Sample ID:	VN0610WBL01		SDG No.:	Q2250
Lab Sample ID:	VN0610WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group3
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086915.D	1		06/10/25 10:09	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	46.7		70 (74) - 130 (125)	93%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		70 (75) - 130 (124)	98%	SPK: 50
2037-26-5	Toluene-d8	51.5		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.0		70 (77) - 130 (121)	98%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	241000	8.23			
540-36-3	1,4-Difluorobenzene	441000	9.106			
3114-55-4	Chlorobenzene-d5	386000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	183000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086915.D  
 Acq On : 10 Jun 2025 10:09  
 Operator : JC\MD  
 Sample : VN0610WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0610WBL01

Quant Time: Jun 11 01:48:33 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	241251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	440901	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	386425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	183380	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.589	65	150863	46.706	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.420%		
35) Dibromofluoromethane	8.177	113	127940	48.965	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.940%		
50) Toluene-d8	10.565	98	532983	51.527	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.060%		
62) 4-Bromofluorobenzene	12.847	95	188114	48.950	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.900%		


Target Compounds	Qvalue
(#= qualifier out of range (m) = manual integration (+) = signals summed	

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086915.D  
Acq On : 10 Jun 2025 10:09  
Operator : JC\MD  
Sample : VN0610WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0610WBL01


Quant Time: Jun 11 01:48:33 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration








#1  
 Pentafluorobenzene  
 Concen: 50.000 ug/l  
 RT: 8.230 min Scan# 1  
 Delta R.T. -0.006 min  
 Lab File: VN086915.D  
 Acq: 10 Jun 2025 10:09


Instrument : MSVOA\_N  
 ClientSampleId : VN0610WBL01





Tgt Ion:168 Resp: 241251  
 Ion Ratio Lower Upper  
 168 100  
 99 56.4 49.1 73.7




#33  
 1,2-Dichloroethane-d4  
 Concen: 46.706 ug/l  
 RT: 8.589 min Scan# 1128  
 Delta R.T. 0.000 min  
 Lab File: VN086915.D  
 Acq: 10 Jun 2025 10:09



Tgt Ion: 65 Resp: 150863  
 Ion Ratio Lower Upper  
 65 100  
 67 54.8 0.0 105.6





#34

1,4-Difluorobenzene


Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN086915.D

Acq: 10 Jun 2025 10:09

**Instrument:** MSVOA\_N  
**ClientSampleId :** VN0610WBL01




Tgt Ion:114 Resp: 440901

Ion Ratio Lower Upper

114 100

63 19.0 0.0 39.6

88 15.9 0.0 30.2



#35

Dibromofluoromethane


Concen: 48.965 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086915.D

Acq: 10 Jun 2025 10:09





Tgt Ion:113 Resp: 127940

Ion Ratio Lower Upper

113 100

111 103.3 84.2 126.2

192 18.1 14.2 21.4



Abundance Scan 1058 (8.177 min): VN086915.D\data.ms (-1)




Sub

50

0

40.0 60 80 100 120 140 160 180 200





m/z--&gt;



#50  
Toluene-d8  
Concen: 51.527 ug/l  
RT: 10.565 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086915.D  
Acq: 10 Jun 2025 10:09


Instrument : MSVOA\_N  
ClientSampleId : VN0610WBL01


Tgt Ion: 98 Resp: 532983  
Ion Ratio Lower Upper  
98 100  
100 66.7 53.4 80.0



#62  
4-Bromofluorobenzene  
Concen: 48.950 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. 0.000 min  
Lab File: VN086915.D  
Acq: 10 Jun 2025 10:09

Tgt Ion: 95 Resp: 188114  
Ion Ratio Lower Upper  
95 100  
174 71.2 0.0 141.8  
176 69.4 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VN086915.D  
Acq: 10 Jun 2025 10:09


Instrument : MSVOA\_N  
ClientSampleId : VN0610WBL01

Tgt Ion:117 Resp: 386425  
Ion Ratio Lower Upper  
117 100  
82 54.2 44.6 67.0  
119 31.7 25.5 38.3



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2012  
Delta R.T. -0.006 min  
Lab File: VN086915.D  
Acq: 10 Jun 2025 10:09

Tgt Ion:152 Resp: 183380  
Ion Ratio Lower Upper  
152 100  
115 60.4 30.1 90.5  
150 158.5 0.0 345.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:
Client Sample ID:	VN0611WBL01		SDG No.:	Q2250
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group3
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	48.3		70 (74) - 130 (125)	97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		70 (75) - 130 (124)	99%	SPK: 50
2037-26-5	Toluene-d8	51.7		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		70 (77) - 130 (121)	99%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	328000	8.23			
540-36-3	1,4-Difluorobenzene	619000	9.106			
3114-55-4	Chlorobenzene-d5	543000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	257000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086942.D  
 Acq On : 11 Jun 2025 12:28  
 Operator : JC\MD  
 Sample : VN0611WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**VN0611WBL01**

Quant Time: Jun 12 01:30:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	328120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	618651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	543433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	257257	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.583	65	212062	48.271	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.540%		
35) Dibromofluoromethane	8.177	113	180701	49.287	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.580%		
50) Toluene-d8	10.565	98	749967	51.673	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.340%		
62) 4-Bromofluorobenzene	12.847	95	266545	49.431	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.860%		


Target Compounds	Qvalue
(#= qualifier out of range (m) = manual integration (+) = signals summed	

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086942.D  
 Acq On : 11 Jun 2025 12:28  
 Operator : JC\MD  
 Sample : VN0611WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0611WBL01


Quant Time: Jun 12 01:30:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086942.D  
Acq: 11 Jun 2025 12:28


Instrument : MSVOA\_N  
ClientSampleId : VN0611WBL01





Tgt Ion:168 Resp: 328120  
Ion Ratio Lower Upper  
168 100  
99 58.0 49.1 73.7






#33  
1,2-Dichloroethane-d4  
Concen: 48.271 ug/l  
RT: 8.583 min Scan# 1127  
Delta R.T. -0.006 min  
Lab File: VN086942.D  
Acq: 11 Jun 2025 12:28



Tgt Ion: 65 Resp: 212062  
Ion Ratio Lower Upper  
65 100  
67 55.4 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086942.D

Acq: 11 Jun 2025 12:28

Instrument:

MSVOA\_N

ClientSampleId :

VN0611WBL01

Tgt Ion:114 Resp: 618651

Ion Ratio Lower Upper

114 100

63 19.9

88 16.5

0.0 39.6

0.0 30.2

Abundance

300000

200000

100000

0

9.00 9.10 9.20

Time--&gt;

#35

Dibromofluoromethane

Concen: 49.287 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN086942.D

Acq: 11 Jun 2025 12:28

Tgt Ion:113 Resp: 180701

Ion Ratio Lower Upper

113 100

111 103.1

192 18.0

84.2 126.2

14.2 21.4

Abundance

60000




40000




20000

0

8.10 8.20 8.30





Time--&gt;






#50  
Toluene-d8  
Concen: 51.673 ug/l  
RT: 10.565 min Scan# 1  
Instrument: MSVOA\_N  
Delta R.T. -0.006 min  
Lab File: VN086942.D  
Acq: 11 Jun 2025 12:28  
ClientSampleId : VN0611WBL01


Tgt Ion: 98 Resp: 749967  
Ion Ratio Lower Upper  
98 100  
100 66.0 53.4 80.0



#62  
4-Bromofluorobenzene  
Concen: 49.431 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086942.D  
Acq: 11 Jun 2025 12:28



Tgt Ion: 95 Resp: 266545  
Ion Ratio Lower Upper  
95 100  
174 72.3 0.0 141.8  
176 68.8 0.0 132.6






#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Instrument: MSVOA\_N  
Delta R.T. -0.000 min  
Lab File: VN086942.D  
Acq: 11 Jun 2025 12:28  
ClientSampleId : VN0611WBL01

Tgt Ion:117 Resp: 543433  
Ion Ratio Lower Upper  
117 100  
82 55.0 44.6 67.0  
119 31.9 25.5 38.3



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.788 min Scan# 2012  
Delta R.T. -0.006 min  
Lab File: VN086942.D  
Acq: 11 Jun 2025 12:28

Tgt Ion:152 Resp: 257257  
Ion Ratio Lower Upper  
152 100  
115 61.2 30.1 90.5  
150 157.9 0.0 345.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:
Client Sample ID:	VN0610WBS01		SDG No.:	Q2250
Lab Sample ID:	VN0610WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group3
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086916.D	1		06/10/25 10:30	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	18.0		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.7		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.2		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.7		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.1		0.20	1.00	ug/L
71-43-2	Benzene	18.5		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.7		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.5		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	17.9		0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	45.4		70 (74) - 130 (125)	91%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		70 (75) - 130 (124)	99%	SPK: 50
2037-26-5	Toluene-d8	48.6		70 (86) - 130 (113)	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.3		70 (77) - 130 (121)	97%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	250000	8.23			
540-36-3	1,4-Difluorobenzene	443000	9.106			
3114-55-4	Chlorobenzene-d5	389000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	190000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086916.D  
 Acq On : 10 Jun 2025 10:30  
 Operator : JC\MD  
 Sample : VN0610WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0610WBS01

Quant Time: Jun 11 01:48:54 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	249843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	443174	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	389297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	189620	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.582	65	151805	45.381	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	90.760%	
35) Dibromofluoromethane	8.177	113	129937	49.474	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	98.940%	
50) Toluene-d8	10.571	98	505508	48.620	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	97.240%	
62) 4-Bromofluorobenzene	12.847	95	186676	48.327	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	96.660%	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.153	85	47416	19.034	ug/l	98
3) Chloromethane	2.401	50	51246	15.931	ug/l	100
4) Vinyl Chloride	2.553	62	59638	17.973	ug/l	98
5) Bromomethane	3.000	94	36539	19.677	ug/l	96
6) Chloroethane	3.159	64	39443	18.402	ug/l	96
7) Trichlorofluoromethane	3.530	101	79526	18.343	ug/l	97
8) Diethyl Ether	3.983	74	35508	18.798	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.400	101	49596	18.208	ug/l	99
10) Methyl Iodide	4.612	142	51915	14.703	ug/l	95
11) Tert butyl alcohol	5.530	59	74605	82.194	ug/l	99
12) 1,1-Dichloroethene	4.365	96	52125	18.744	ug/l	90
13) Acrolein	4.200	56	22542	78.459	ug/l	98
14) Allyl chloride	5.047	41	78157	16.947	ug/l	97
15) Acrylonitrile	5.741	53	189727	89.429	ug/l	99
16) Acetone	4.447	43	142458	80.306	ug/l	97
17) Carbon Disulfide	4.736	76	131873	17.144	ug/l	100
18) Methyl Acetate	5.047	43	88551	17.129	ug/l	97
19) Methyl tert-butyl Ether	5.812	73	187833	18.655	ug/l	99
20) Methylene Chloride	5.300	84	57671	17.365	ug/l	95
21) trans-1,2-Dichloroethene	5.812	96	55486	17.934	ug/l	96
22) Diisopropyl ether	6.688	45	176809	18.188	ug/l	96
23) Vinyl Acetate	6.618	43	744861	90.687	ug/l	98
24) 1,1-Dichloroethane	6.583	63	101669	18.173	ug/l	98
25) 2-Butanone	7.494	43	242187	83.991	ug/l	95
26) 2,2-Dichloropropane	7.500	77	89631	20.596	ug/l	100
27) cis-1,2-Dichloroethene	7.494	96	69159	18.691	ug/l	99
28) Bromochloromethane	7.824	49	45748	16.632	ug/l	93
29) Tetrahydrofuran	7.847	42	160591	85.494	ug/l	97
30) Chloroform	7.977	83	101594	18.184	ug/l	97
31) Cyclohexane	8.265	56	89487	16.494	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	85981	18.094	ug/l	98
36) 1,1-Dichloropropene	8.382	75	74129	18.942	ug/l	100
37) Ethyl Acetate	7.571	43	88740	17.812	ug/l	99
38) Carbon Tetrachloride	8.371	117	72144	18.777	ug/l	97
39) Methylcyclohexane	9.606	83	85476	15.942	ug/l	96
40) Benzene	8.612	78	236213	18.458	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086916.D  
 Acq On : 10 Jun 2025 10:30  
 Operator : JC\MD  
 Sample : VN0610WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0610WBS01

Quant Time: Jun 11 01:48:54 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	48124	17.200	ug/1	93
42) 1,2-Dichloroethane	8.677	62	72690	18.727	ug/1	99
43) Isopropyl Acetate	8.694	43	144037	18.014	ug/1	97
44) Trichloroethene	9.353	130	58449	19.262	ug/1	97
45) 1,2-Dichloropropane	9.624	63	58046	18.644	ug/1	97
46) Dibromomethane	9.712	93	39229	18.970	ug/1	98
47) Bromodichloromethane	9.888	83	80034	18.810	ug/1	99
48) Methyl methacrylate	9.682	41	64650	17.567	ug/1	96
49) 1,4-Dioxane	9.700	88	23656	353.325	ug/1 #	98
51) 4-Methyl-2-Pentanone	10.447	43	440301	91.469	ug/1	98
52) Toluene	10.629	92	148340	18.967	ug/1	97
53) t-1,3-Dichloropropene	10.835	75	93038	19.555	ug/1	98
54) cis-1,3-Dichloropropene	10.312	75	98066	19.264	ug/1	99
55) 1,1,2-Trichloroethane	11.018	97	58629	19.483	ug/1	99
56) Ethyl methacrylate	10.882	69	93272	19.460	ug/1	96
57) 1,3-Dichloropropane	11.165	76	99132	18.990	ug/1	99
58) 2-Chloroethyl Vinyl ether	10.159	63	251485	87.967	ug/1	98
59) 2-Hexanone	11.206	43	271013	87.405	ug/1	91
60) Dibromochloromethane	11.359	129	61081	19.481	ug/1	99
61) 1,2-Dibromoethane	11.470	107	58347	18.916	ug/1	100
64) Tetrachloroethene	11.106	164	44151	17.920	ug/1	94
65) Chlorobenzene	11.888	112	162051	18.874	ug/1	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	52766	19.118	ug/1	99
67) Ethyl Benzene	11.965	91	270410	18.284	ug/1	98
68) m/p-Xylenes	12.070	106	213409	37.696	ug/1	96
69) o-Xylene	12.400	106	102597	18.922	ug/1	98
70) Styrene	12.412	104	176358	19.008	ug/1	99
71) Bromoform	12.576	173	41058	20.077	ug/1 #	98
73) Isopropylbenzene	12.694	105	249411	18.055	ug/1	100
74) N-amyl acetate	12.529	43	98634m	20.433	ug/1	
75) 1,1,2,2-Tetrachloroethane	12.941	83	90657	19.372	ug/1	99
76) 1,2,3-Trichloropropane	12.994	75	77098m	17.105	ug/1	
77) Bromobenzene	12.976	156	61956	19.557	ug/1	97
78) n-propylbenzene	13.035	91	298896	17.805	ug/1	100
79) 2-Chlorotoluene	13.123	91	183491	18.226	ug/1	96
80) 1,3,5-Trimethylbenzene	13.170	105	201924	17.705	ug/1	97
81) trans-1,4-Dichloro-2-b...	12.735	75	39145	19.992	ug/1	91
82) 4-Chlorotoluene	13.217	91	188207	18.476	ug/1	99
83) tert-Butylbenzene	13.435	119	186449	17.859	ug/1	97
84) 1,2,4-Trimethylbenzene	13.482	105	206670	18.071	ug/1	100
85) sec-Butylbenzene	13.612	105	255364	16.843	ug/1	99
86) p-Isopropyltoluene	13.729	119	212141	16.927	ug/1	99
87) 1,3-Dichlorobenzene	13.729	146	117325	18.823	ug/1	99
88) 1,4-Dichlorobenzene	13.812	146	118648	18.671	ug/1	99
89) n-Butylbenzene	14.053	91	197259	16.250	ug/1	99
90) Hexachloroethane	14.329	117	36492	17.178	ug/1	99
91) 1,2-Dichlorobenzene	14.106	146	112799	18.837	ug/1	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	20803	18.587	ug/1	99
93) 1,2,4-Trichlorobenzene	15.388	180	64838	16.956	ug/1	100
94) Hexachlorobutadiene	15.500	225	20732	14.552	ug/1	98
95) Naphthalene	15.635	128	257750	18.109	ug/1	100
96) 1,2,3-Trichlorobenzene	15.835	180	59536	15.670	ug/1	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086916.D  
 Acq On : 10 Jun 2025 10:30  
 Operator : JC\MD  
 Sample : VN0610WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VN0610WBS01

Quant Time: Jun 11 01:48:54 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
(#)						

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086916.D  
Acq On : 10 Jun 2025 10:30  
Operator : JC\MD  
Sample : VN0610WBS01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 5 Sample Multiplier: 1




Quant Time: Jun 11 01:48:54 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.230 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#2

Dichlorodifluoromethane

Concen: 19.034 ug/l


RT: 2.153 min Scan# 34







Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt Ion: 85 Resp: 47416  
 Ion Ratio Lower Upper  
 85 100  
 87 33.6 16.2 48.6





#3

Chloromethane

Concen: 15.931 ug/l

RT: 2.401 min Scan# 7

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#4

Vinyl Chloride

Concen: 17.973 ug/l


RT: 2.553 min Scan# 102


Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion: 62 Resp: 59638  
 Ion Ratio Lower Upper  
 62 100  
 64 33.6 26.0 39.0





#5  
Bromomethane  
Concen: 19.677 ug/l  
RT: 3.000 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30




Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01




Tgt Ion: 94 Resp: 36539  
Ion Ratio Lower Upper  
94 100  
96 93.4 77.9 116.9

**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#6  
Chloroethane  
Concen: 18.402 ug/l  
RT: 3.159 min Scan# 205  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30




Tgt Ion: 64 Resp: 39443  
Ion Ratio Lower Upper  
64 100  
66 32.9 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 18.343 ug/l  
RT: 3.530 min Scan# 2  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30




Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01




Tgt Ion:101 Resp: 79520  
Ion Ratio Lower Upper  
101 100  
103 67.6 52.2 78.2

Manual Integrations  
**APPROVED**



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025









#8  
Diethyl Ether  
Concen: 18.798 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30



Tgt Ion: 74 Resp: 35508  
Ion Ratio Lower Upper  
74 100  
45 86.0 45.5 136.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 18.208 ug/l

RT: 4.400 min Scan# 4

Instrument : MSVOA\_N

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

ClientSampleId : VN0610WBS01

Manual Integrations


APPROVED

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

Abundance

Time--&gt;



#10

Methyl Iodide

Concen: 14.703 ug/l

RT: 4.612 min Scan# 452

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt Ion:142 Resp: 51915

Ion Ratio Lower Upper


142 100




127 35.5 30.9 46.3

141 14.4 10.4 15.6

Abundance

Time--&gt;









#11

Tert butyl alcohol  
Concen: 82.194 ug/l  
RT: 5.530 min Scan# 6  
Delta R.T. -0.018 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#12





1,1-Dichloroethene  
Concen: 18.744 ug/l  
RT: 4.365 min Scan# 410  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 96 Resp: 52125  
Ion Ratio Lower Upper  
96 100  
61 144.3 125.8 188.6  
98 57.9 52.5 78.7





Abundance Scan 382 (4.200 min): VN086916.D\data.ms



#13

Acrolein

Concen: 78.459 ug/l

RT: 4.200 min Scan# 3

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#14

Allyl chloride

Concen: 16.947 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion: 41 Resp: 78157




Ion Ratio Lower Upper

41 100

39 62.2 51.3 76.9

76 41.3 31.4 47.0





#15

Acrylonitrile

Concen: 89.429 ug/l

RT: 5.741 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N




ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#16

Acetone

Concen: 80.306 ug/l

RT: 4.447 min Scan# 424

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion: 43 Resp: 142458




Ion Ratio Lower Upper

43 100

58 36.8 28.0 42.0

1 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16





#17

Carbon Disulfide

Concen: 17.144 ug/l

RT: 4.736 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#18

Methyl Acetate

Concen: 17.129 ug/l

RT: 5.047 min Scan# 526


Delta R.T. 0.006 min


Lab File: VN086916.D

Acq: 10 Jun 2025 10:30



Tgt Ion: 43 Resp: 88551

Ion	Ratio	Lower	Upper
43	100		
74	27.2	20.7	31.1






Abundance Scan 656 (5.812 min): VN086916.D\data.ms



Abundance Scan 569 (5.300 min): VN086916.D\data.ms



#19

Methyl tert-butyl Ether

Concen: 18.655 ug/l

RT: 5.812 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#20

Methylene Chloride

Concen: 17.365 ug/l

RT: 5.300 min Scan# 569

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt Ion: 84 Resp: 57671


Ion Ratio Lower Upper


84 100

49 106.7 90.5 135.7


51 34.0 28.5 42.7

86 63.8 48.1 72.1






Abundance Scan 656 (5.812 min): VN086916.D\data.ms



Abundance Scan 656 (5.812 min): VN086916.D\data.ms (-60)





#21

trans-1,2-Dichloroethene  
Concen: 17.934 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30


Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



Abundance Scan 805 (6.688 min): VN086916.D\data.ms




Abundance Scan 805 (6.688 min): VN086916.D\data.ms (-75)




#22

Diisopropyl ether  
Concen: 18.188 ug/l  
RT: 6.688 min Scan# 805  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 45 Resp: 176809  
Ion Ratio Lower Upper  
45 100  
43 49.8 41.9 62.9  
87 31.5 22.4 33.6  
59 12.1 9.8 14.6





#23

**Vinyl Acetate**

Concen: 90.687 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086916.D



Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N

ClientSampleId :

VN0610WBS01



Tgt Ion: 43 Resp: 74486

Ion Ratio Lower Upper



43 100

86 12.6 9.5 14.3

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#24

**1,1-Dichloroethane**

Concen: 18.173 ug/l

RT: 6.583 min Scan# 787

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30




Tgt Ion: 63 Resp: 101669




Ion Ratio Lower Upper

63 100

98 6.8 3.6 10.8

100 4.5 2.6 7.8





#25

2-Butanone

Concen: 83.991 ug/l

RT: 7.494 min Scan# 942

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#26

2,2-Dichloropropane

Concen: 20.596 ug/l


RT: 7.500 min Scan# 943




Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt Ion: 77 Resp: 89631  
 Ion Ratio Lower Upper  
 77 100  
 97 23.2 11.6 34.7









#27

cis-1,2-Dichloroethene  
Concen: 18.691 ug/l  
RT: 7.494 min Scan# 942  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025









#28

Bromochloromethane  
Concen: 16.632 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 49 Resp: 45748  
Ion Ratio Lower Upper  
49 100  
129 1.6 0.0 3.8  
130 83.6 62.2 93.2





#29

Tetrahydrofuran

Concen: 85.494 ug/l

RT: 7.847 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#30

Chloroform

Concen: 18.184 ug/l

RT: 7.977 min Scan# 1024

Delta R.T. -0.000 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30







Tgt Ion: 83 Resp: 101594

Ion Ratio Lower Upper

83 100

85 63.7 52.9 79.3





#31

Cyclohexane

Concen: 16.494 ug/l

RT: 8.265 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#32

1,1,1-Trichloroethane

Concen: 18.094 ug/l


RT: 8.177 min Scan# 1058




Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt	Ion	Resp:	
	97	85981	
	100		
	97		
	100		
	99	63.7	52.0
	61	43.5	35.9
			53.9









#33

1,2-Dichloroethane-d4  
Concen: 45.381 ug/l  
RT: 8.582 min Scan# 1216  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025









#34


1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion:114 Resp: 443174  
Ion Ratio Lower Upper  
114 100  
63 19.9 0.0 39.6  
88 15.9 0.0 30.2












#35

Dibromofluoromethane  
Concen: 49.474 ug/lRT: 8.177 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01**Manual Integrations  
APPROVED**Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025

#36  
1,1-Dichloropropene  
Concen: 18.942 ug/l  
RT: 8.382 min Scan# 1093  
Delta R.T. 0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 75 Resp: 74129  
Ion Ratio Lower Upper  
75 100  
110 35.1 17.6 52.9  
77 31.3 25.3 37.9





#37

**Ethyl Acetate**

Concen: 17.812 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N

ClientSampleId :

VN0610WBS01

Tgt Ion: 43 Resp: 88740

Ion Ratio Lower Upper

43 100


61 15.4 12.1 18.1

70 12.8 9.8 14.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#38

**Carbon Tetrachloride**

Concen: 18.777 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion:117 Resp: 72144


Ion Ratio Lower Upper

117 100

119 96.8 75.0 112.6

121 29.7 24.9 37.3





#39

Methylcyclohexane

Concen: 15.942 ug/l

RT: 9.606 min Scan# 1


Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N

ClientSampleId : VN0610WBS01



Tgt Ion: 83 Resp: 85470

Ion Ratio Lower Upper

83 100




55 70.6 61.0 91.6

98 48.3 38.9 58.3

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#40

Benzene

Concen: 18.458 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086916.D




Acq: 10 Jun 2025 10:30







Tgt Ion: 78 Resp: 236213

Ion Ratio Lower Upper

78 100

77 24.0 18.7 28.1





#41

Methacrylonitrile

Concen: 17.200 ug/l

RT: 7.788 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#42

1,2-Dichloroethane

Concen: 18.727 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30


Tgt Ion: 62 Resp: 72690

Ion Ratio Lower Upper

62 100

98 11.1 0.0 21.2





#43

Isopropyl Acetate

Concen: 18.014 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N

ClientSampleId :

VN0610WBS01



Tgt Ion: 43 Resp: 14403

Ion Ratio Lower Upper

43 100

61 29.0 22.1 33.1







87 15.9 11.8 17.6


Manual Integrations

APPROVED


Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#44  
Trichloroethene  
Concen: 19.262 ug/l  
RT: 9.353 min Scan# 1258  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30Tgt Ion:130 Resp: 58449  
Ion Ratio Lower Upper  
130 100  
95 95.4 0.0 196.2




Abundance Scan 1304 (9.624 min): VN086916.D\data.ms




Abundance Scan 1304 (9.624 min): VN086916.D\data.ms (-1)



Abundance Scan 1319 (9.712 min): VN086916.D\data.ms



Abundance Scan 1319 (9.712 min): VN086916.D\data.ms (-1)



#45

1,2-Dichloropropane

Concen: 18.644 ug/l

RT: 9.624 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N

ClientSampleId :

VN0610WBS01

Tgt Ion: 63 Resp: 58040

Ion Ratio Lower Upper

63 100


65 33.7 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#46

Dibromomethane

Concen: 18.970 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion: 93 Resp: 39229




Ion Ratio Lower Upper

93 100

95 84.5 67.1 100.7

174 96.7 74.2 111.4









#47

Bromodichloromethane  
Concen: 18.810 ug/l  
RT: 9.888 min Scan# 1349  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#48

Methyl methacrylate  
Concen: 17.567 ug/l  
RT: 9.682 min Scan# 1314  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 41 Resp: 64650  
Ion Ratio Lower Upper  
41 100  
69 98.8 74.6 112.0  
39 51.7 40.2 60.4





#49

1,4-Dioxane

Concen: 353.325 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N



ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**




Reviewed By :John Carlone 06/11/2025




Supervised By :Mahesh Dadoda 06/11/2025



#50  
Toluene-d8  
Concen: 48.620 ug/l  
RT: 10.571 min Scan# 1465  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 98 Resp: 505508  
Ion Ratio Lower Upper  
98 100  
100 66.9 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 91.469 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#52

Toluene

Concen: 18.967 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. -0.006 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30


Tgt Ion: 92 Resp: 148340

Ion Ratio Lower Upper


92 100

91 166.8 136.5 204.7








Abundance Scan 1510 (10.835 min): VN086916.D\data.ms (-)




Abundance Scan 1510 (10.835 min): VN086916.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086916.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086916.D\data.ms (-)



#53

t-1,3-Dichloropropene

Concen: 19.555 ug/l

RT: 10.835 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

### Manual Integrations APPROVED

Reviewed By :John Carbone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#54

cis-1,3-Dichloropropene

Concen: 19.264 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion: 75 Resp: 98066




Ion Ratio Lower Upper

75 100

77 32.1 26.2 39.4

39 39.4 32.3 48.5









#55

1,1,2-Trichloroethane  
Concen: 19.483 ug/l  
RT: 11.018 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#56

Ethyl methacrylate  
Concen: 19.460 ug/l  
RT: 10.882 min Scan# 1518  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 69 Resp: 93272  
Ion Ratio Lower Upper  
69 100  
41 56.3 47.2 70.8  
39 25.5 23.0 34.6









#57

1,3-Dichloropropane  
Concen: 18.990 ug/l  
RT: 11.165 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#58



2-Chloroethyl Vinyl ether  
Concen: 87.967 ug/l  
RT: 10.159 min Scan# 1395  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 63 Resp: 251485  
Ion Ratio Lower Upper  
63 100  
106 29.7 23.0 34.6







Abundance Scan 1573 (11.206 min): VN086916.D\data.ms



Abundance Scan 1599 (11.359 min): VN086916.D\data.ms



#59

2-Hexanone

Concen: 87.405 ug/l

RT: 11.206 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N


ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#60

Dibromochloromethane

Concen: 19.481 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30




Tgt Ion:129 Resp: 61081

Ion Ratio Lower Upper

129 100

127 77.9 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 18.916 ug/l  
RT: 11.470 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0610WBS01


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#62  
4-Bromofluorobenzene  
Concen: 48.327 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 95 Resp: 186676  
Ion Ratio Lower Upper  
95 100  
174 72.7 0.0 141.8  
176 69.2 0.0 132.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#64

Tetrachloroethene


Concen: 17.920 ug/l


RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30







#65  
Chlorobenzene  
Concen: 18.874 ug/l  
RT: 11.888 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

Tgt Ion:112 Resp: 16205  
Ion Ratio Lower Upper  
112 100  
114 31.4 26.2 39.2


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.118 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion:131 Resp: 52766  
Ion Ratio Lower Upper  
131 100  
133 96.0 48.3 144.8  
119 69.7 34.0 101.8





#67

Ethyl Benzene

Concen: 18.284 ug/l

RT: 11.965 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N

ClientSampleId :

VN0610WBS01

Tgt Ion: 91 Resp: 270410

Ion Ratio Lower Upper

91 100




106 30.7 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#68

m/p-Xylenes

Concen: 37.696 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30


Tgt Ion: 106 Resp: 213409

Ion Ratio Lower Upper

106 100

91 193.9 159.4 239.0







#69  
o-Xylene  
Concen: 18.922 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#70  
Styrene  
Concen: 19.008 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion:104 Resp: 176358  
Ion Ratio Lower Upper  
104 100  
78 48.0 38.5 57.7  
103 54.6 43.1 64.7





#71

Bromoform

Concen: 20.077 ug/l

RT: 12.576 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N



ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion:152 Resp: 189620




Ion Ratio Lower Upper

152 100

115 59.9 30.1 90.5

150 164.9 0.0 345.0





#73

Isopropylbenzene

Concen: 18.055 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#74

N-amyl acetate

Concen: 20.433 ug/l

RT: 12.529 min Scan# 1798

Delta R.T. 0.017 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt Ion: 43 Resp: 98634


Ion Ratio Lower Upper




43 100

70 40.2 37.6 56.4

55 18.8 19.0 28.6

61 22.8 20.8 31.2









#75  
1,1,2,2-Tetrachloroethane  
Concen: 19.372 ug/l  
RT: 12.941 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#76  
1,2,3-Trichloropropane  
Concen: 17.105 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 75 Resp: 77098  
Ion Ratio Lower Upper  
75 100  
77 185.1 79.3 237.9





#77

Bromobenzene

Concen: 19.557 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#78

n-propylbenzene

Concen: 17.805 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086916.D


Acq: 10 Jun 2025 10:30




Tgt Ion: 91 Resp: 298896

Ion Ratio Lower Upper

91 100

120 22.9 11.4 34.2





#79

2-Chlorotoluene

Concen: 18.226 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N




ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#80

1,3,5-Trimethylbenzene

Concen: 17.705 ug/l


RT: 13.170 min Scan# 1907




Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt	Ion:105	Resp:	201924
Ion	Ratio	Lower	Upper
105	100		
120	51.6	24.9	74.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 19.992 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#82  
4-Chlorotoluene  
Concen: 18.476 ug/l  
RT: 13.217 min Scan# 1915  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 91 Resp: 188207  
Ion Ratio Lower Upper  
91 100  
126 33.9 16.7 50.1





#83

tert-Butylbenzene

Concen: 17.859 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N




ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#84

1,2,4-Trimethylbenzene

Concen: 18.071 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion:105 Resp: 206670




Ion Ratio Lower Upper

105 100

120 46.5 23.2 69.6

Time--&gt; 13.35 13.40 13.45 13.50









#85  
sec-Butylbenzene  
Concen: 16.843 ug/l  
RT: 13.612 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01


1 Manual Integrations  
2 APPROVED




3 Reviewed By :John Carlone 06/11/2025  
4 Supervised By :Mahesh Dadoda 06/11/2025



#86  
p-Isopropyltoluene  
Concen: 16.927 ug/l  
RT: 13.729 min Scan# 2002  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion:119 Resp: 212141  
Ion Ratio Lower Upper  
119 100  
134 26.6 13.5 40.4  
91 22.5 11.5 34.4





#87

1,3-Dichlorobenzene

Concen: 18.823 ug/l

RT: 13.729 min Scan# 2

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#88

1,4-Dichlorobenzene

Concen: 18.671 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30


Tgt Ion:146 Resp: 118648




Ion Ratio Lower Upper

146 100

111 40.8 19.9 59.7

148 64.3 31.8 95.4





#89

n-Butylbenzene

Concen: 16.250 ug/l

RT: 14.053 min Scan# 2

Instrument:

Delta R.T. -0.000 min

MSVOA\_N

Lab File: VN086916.D

ClientSampleId :





Acq: 10 Jun 2025 10:30

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#90

Hexachloroethane

Concen: 17.178 ug/l


RT: 14.329 min Scan# 2104




Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt	Ion	Ion Ratio	Lower	Upper
117	100			
201	67.5	33.2	99.6	









#91

1,2-Dichlorobenzene  
Concen: 18.837 ug/l  
RT: 14.106 min Scan# 2170  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#92

1,2-Dibromo-3-Chloropropane  
Concen: 18.587 ug/l  
RT: 14.717 min Scan# 2170  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion: 75 Resp: 20803  
Ion Ratio Lower Upper  
75 100  
155 83.1 41.1 123.4  
157 104.1 52.7 158.1









#93

1,2,4-Trichlorobenzene  
Concen: 16.956 ug/l  
RT: 15.388 min Scan# 2303  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Instrument : MSVOA\_N  
ClientSampleId : VN0610WBS01

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#94

Hexachlorobutadiene  
Concen: 14.552 ug/l  
RT: 15.500 min Scan# 2303  
Delta R.T. -0.000 min  
Lab File: VN086916.D  
Acq: 10 Jun 2025 10:30

Tgt Ion:225 Resp: 20732  
Ion Ratio Lower Upper  
225 100  
223 64.7 33.5 100.4  
227 63.1 31.1 93.5





#95

Naphthalene

Concen: 18.109 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.006 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Instrument:

MSVOA\_N





ClientSampleId :

VN0610WBS01

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#96

1,2,3-Trichlorobenzene

Concen: 15.670 ug/l


RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN086916.D

Acq: 10 Jun 2025 10:30

Tgt	Ion:180	Resp:	59536
Ion	Ratio	Lower	Upper
180	100		
182	95.5	47.0	141.0
145	33.5	16.6	49.8





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:
Client Sample ID:	VN0611WBS02		SDG No.:	Q2250
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group3
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	18.9		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.5		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.0		0.20	1.00	ug/L
71-43-2	Benzene	19.3		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	47.1		70 (74) - 130 (125)	94%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		70 (75) - 130 (124)	102%	SPK: 50
2037-26-5	Toluene-d8	48.7		70 (86) - 130 (113)	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		70 (77) - 130 (121)	100%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	207000		8.235		
540-36-3	1,4-Difluorobenzene	371000		9.106		
3114-55-4	Chlorobenzene-d5	321000		11.865		
3855-82-1	1,4-Dichlorobenzene-d4	157000		13.788		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086944.D  
 Acq On : 11 Jun 2025 13:27  
 Operator : JC\MD  
 Sample : VN0611WBS02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0611WBS02

Quant Time: Jun 12 01:31:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.235	168	206799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	370524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157403	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	130342	47.075	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125			Recovery =	94.160%	
35) Dibromofluoromethane	8.177	113	112198	51.096	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124			Recovery =	102.200%	
50) Toluene-d8	10.571	98	423380	48.706	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113			Recovery =	97.420%	
62) 4-Bromofluorobenzene	12.847	95	161321	49.951	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121			Recovery =	99.900%	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.153	85	40672	19.725	ug/l	100
3) Chloromethane	2.401	50	42812	16.079	ug/l	97
4) Vinyl Chloride	2.553	62	51780	18.852	ug/l	95
5) Bromomethane	3.000	94	25577	16.641	ug/l	90
6) Chloroethane	3.153	64	34500	19.446	ug/l	99
7) Trichlorofluoromethane	3.530	101	68752	19.159	ug/l	99
8) Diethyl Ether	3.983	74	30914	19.772	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.400	101	44117	19.568	ug/l	99
10) Methyl Iodide	4.612	142	36924	12.634	ug/l	100
11) Tert butyl alcohol	5.541	59	69743	92.831	ug/l	99
12) 1,1-Dichloroethene	4.365	96	44269	19.233	ug/l	97
13) Acrolein	4.206	56	22290	93.730	ug/l	98
14) Allyl chloride	5.047	41	68027	17.821	ug/l	97
15) Acrylonitrile	5.741	53	166971	95.084	ug/l	100
16) Acetone	4.447	43	145322	98.972	ug/l	98
17) Carbon Disulfide	4.742	76	113132	17.769	ug/l	100
18) Methyl Acetate	5.047	43	81357	19.013	ug/l	99
19) Methyl tert-butyl Ether	5.812	73	163338	19.599	ug/l	98
20) Methylene Chloride	5.294	84	52196	18.988	ug/l	96
21) trans-1,2-Dichloroethene	5.812	96	47284	18.464	ug/l	95
22) Diisopropyl ether	6.688	45	155508	19.326	ug/l	97
23) Vinyl Acetate	6.618	43	653278	96.092	ug/l	98
24) 1,1-Dichloroethane	6.588	63	89930	19.421	ug/l	100
25) 2-Butanone	7.494	43	215760	90.401	ug/l	96
26) 2,2-Dichloropropane	7.500	77	78031	21.663	ug/l	98
27) cis-1,2-Dichloroethene	7.500	96	59578	19.453	ug/l	99
28) Bromochloromethane	7.824	49	45329	19.909	ug/l	95
29) Tetrahydrofuran	7.853	42	143866	92.532	ug/l	97
30) Chloroform	7.977	83	90272	19.521	ug/l	98
31) Cyclohexane	8.265	56	79050	17.603	ug/l	97
32) 1,1,1-Trichloroethane	8.182	97	74861	19.033	ug/l	97
36) 1,1-Dichloropropene	8.377	75	63577	19.431	ug/l	99
37) Ethyl Acetate	7.571	43	80415	19.306	ug/l	99
38) Carbon Tetrachloride	8.371	117	61747	19.222	ug/l	97
39) Methylcyclohexane	9.606	83	75971	16.947	ug/l	98
40) Benzene	8.612	78	206255	19.277	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086944.D  
 Acq On : 11 Jun 2025 13:27  
 Operator : JC\MD  
 Sample : VN0611WBS02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0611WBS02

Quant Time: Jun 12 01:31:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	42617	18.218	ug/1	93
42) 1,2-Dichloroethane	8.677	62	63379	19.530	ug/1	98
43) Isopropyl Acetate	8.694	43	126914	18.984	ug/1	98
44) Trichloroethene	9.359	130	50593	19.942	ug/1	99
45) 1,2-Dichloropropane	9.624	63	51091	19.628	ug/1	95
46) Dibromomethane	9.712	93	34741	20.093	ug/1	99
47) Bromodichloromethane	9.894	83	69845	19.634	ug/1	98
48) Methyl methacrylate	9.682	41	57757	18.771	ug/1	99
49) 1,4-Dioxane	9.706	88	22914	409.347	ug/1 #	97
51) 4-Methyl-2-Pentanone	10.447	43	394396	97.997	ug/1	99
52) Toluene	10.629	92	127563	19.509	ug/1	99
53) t-1,3-Dichloropropene	10.835	75	79224	19.917	ug/1	100
54) cis-1,3-Dichloropropene	10.318	75	85586	20.109	ug/1	99
55) 1,1,2-Trichloroethane	11.018	97	51421	20.438	ug/1	94
56) Ethyl methacrylate	10.882	69	80942	20.198	ug/1	94
57) 1,3-Dichloropropane	11.165	76	85602	19.613	ug/1	99
58) 2-Chloroethyl Vinyl ether	10.165	63	243584	101.910	ug/1	99
59) 2-Hexanone	11.206	43	231264	89.210	ug/1	95
60) Dibromochloromethane	11.359	129	53655	20.468	ug/1	99
61) 1,2-Dibromoethane	11.471	107	50832	19.711	ug/1	99
64) Tetrachloroethene	11.106	164	38926	19.150	ug/1	98
65) Chlorobenzene	11.894	112	141721	20.006	ug/1	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	46404	20.378	ug/1	99
67) Ethyl Benzene	11.965	91	239733	19.647	ug/1	100
68) m/p-Xylenes	12.070	106	185026	39.613	ug/1	99
69) o-Xylene	12.400	106	90240	20.172	ug/1	99
70) Styrene	12.412	104	154149	20.137	ug/1	99
71) Bromoform	12.582	173	35350	20.952	ug/1 #	96
73) Isopropylbenzene	12.694	105	221545	19.320	ug/1	99
74) N-amyl acetate	12.529	43	63289	15.794	ug/1 #	85
75) 1,1,2,2-Tetrachloroethane	12.941	83	81476	20.973	ug/1	100
76) 1,2,3-Trichloropropane	12.994	75	71582m	19.132	ug/1	
77) Bromobenzene	12.982	156	55176	20.981	ug/1	98
78) n-propylbenzene	13.035	91	267384	19.187	ug/1	100
79) 2-Chlorotoluene	13.123	91	164813	19.721	ug/1	99
80) 1,3,5-Trimethylbenzene	13.170	105	181773	19.200	ug/1	99
81) trans-1,4-Dichloro-2-b...	12.735	75	30121	18.532	ug/1 #	82
82) 4-Chlorotoluene	13.217	91	167991	19.866	ug/1	100
83) tert-Butylbenzene	13.435	119	161593	18.646	ug/1	99
84) 1,2,4-Trimethylbenzene	13.482	105	182431	19.216	ug/1	98
85) sec-Butylbenzene	13.617	105	228907	18.189	ug/1	100
86) p-Isopropyltoluene	13.729	119	192991	18.551	ug/1	99
87) 1,3-Dichlorobenzene	13.735	146	104883	20.271	ug/1	98
88) 1,4-Dichlorobenzene	13.812	146	107750	20.426	ug/1	99
89) n-Butylbenzene	14.053	91	177482	17.613	ug/1	100
90) Hexachloroethane	14.329	117	33456	18.972	ug/1	99
91) 1,2-Dichlorobenzene	14.106	146	99394	19.995	ug/1	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	17796	19.155	ug/1	98
93) 1,2,4-Trichlorobenzene	15.394	180	58509	18.433	ug/1	98
94) Hexachlorobutadiene	15.500	225	19592	16.567	ug/1	97
95) Naphthalene	15.635	128	232760	19.700	ug/1	100
96) 1,2,3-Trichlorobenzene	15.841	180	56144	17.802	ug/1	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086944.D  
 Acq On : 11 Jun 2025 13:27  
 Operator : JC\MD  
 Sample : VN0611WBS02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

**Instrument :**  
**MSVOA\_N**  
**ClientSampleId :**  
**VN0611WBS02**

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025

Quant Time: Jun 12 01:31:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061125\  
 Data File : VN086944.D  
 Acq On : 11 Jun 2025 13:27  
 Operator : JC\MD  
 Sample : VN0611WBS02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1




Quant Time: Jun 12 01:31:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0611WBS02

### Manual Integrations APPROVED

Reviewed By :John Carlane 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025









#1

Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.235 min Scan# 1  
Delta R.T. -0.001 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0611WBS02

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#2

Dichlorodifluoromethane  
Concen: 19.725 ug/l  
RT: 2.153 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 85 Resp: 40672  
Ion Ratio Lower Upper  
85 100  
87 32.5 16.2 48.6









#3  
Chloromethane  
Concen: 16.079 ug/l  
RT: 2.401 min Scan# 7  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#4  
Vinyl Chloride  
Concen: 18.852 ug/l  
RT: 2.553 min Scan# 102  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27


Tgt Ion: 62 Resp: 51780  
Ion Ratio Lower Upper  
62 100  
64 29.7 26.0 39.0





#5  
Bromomethane  
Concen: 16.641 ug/l  
RT: 3.000 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27




Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02




Tgt Ion: 94 Resp: 2557  
Ion Ratio Lower Upper  
94 100  
96 87.8 77.9 116.9

**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025









#6  
Chloroethane  
Concen: 19.446 ug/l  
RT: 3.153 min Scan# 204  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27



Tgt Ion: 64 Resp: 34500  
Ion Ratio Lower Upper  
64 100  
66 30.3 24.6 36.8






#7

Trichlorofluoromethane  
Concen: 19.159 ug/l  
RT: 3.530 min Scan# 2  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025




#8

Diethyl Ether  
Concen: 19.772 ug/l  
RT: 3.983 min Scan# 345  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 74 Resp: 30914  
Ion Ratio Lower Upper  
74 100  
45 85.5 45.5 136.5









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 19.568 ug/l  
RT: 4.400 min Scan# 4  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

Tgt Ion:101 Resp: 4411  
Ion Ratio Lower Upper  
101 100  
85 41.5 33.8 50.8  
151 74.0 58.7 88.1




Manual Integrations  
APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#10  
Methyl Iodide  
Concen: 12.634 ug/l  
RT: 4.612 min Scan# 452  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:142 Resp: 36924  
Ion Ratio Lower Upper  
142 100  
127 38.9 30.9 46.3  
141 12.9 10.4 15.6









#11

Tert butyl alcohol  
Concen: 92.831 ug/l  
RT: 5.541 min Scan# 6  
Delta R.T. -0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025









#12

1,1-Dichloroethene  
Concen: 19.233 ug/l  
RT: 4.365 min Scan# 410  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 96 Resp: 44269  
Ion Ratio Lower Upper  
96 100  
61 153.8 125.8 188.6  
98 61.2 52.5 78.7






#13

Acrolein  
Concen: 93.730 ug/l  
RT: 4.206 min Scan# 31  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025









#14

Allyl chloride  
Concen: 17.821 ug/l  
RT: 5.047 min Scan# 526  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 41 Resp: 68027  
Ion Ratio Lower Upper  
41 100  
39 62.5 51.3 76.9  
76 41.4 31.4 47.0





#15

Acrylonitrile

Concen: 95.084 ug/l

RT: 5.741 min Scan# 6

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025

Abundance

40000

30000

20000

10000

0

Time--&gt; 5.60 5.741 5.80 6.00

#16

Acetone

Concen: 98.972 ug/l

RT: 4.447 min Scan# 424

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 43 Resp: 145322

Ion Ratio Lower Upper

43 100

58 35.9 28.0 42.0

Abundance

40000




30000

20000

10000

0

Time--&gt; 4.40 4.447 4.60



#17

Carbon Disulfide

Concen: 17.769 ug/l

RT: 4.742 min Scan# 4

Delta R.T. 0.012 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#18

Methyl Acetate

Concen: 19.013 ug/l

RT: 5.047 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086944.D


Acq: 11 Jun 2025 13:27







Tgt Ion: 43 Resp: 81357

Ion Ratio Lower Upper

43 100

74 26.2 20.7 31.1






#19

Methyl tert-butyl Ether  
Concen: 19.599 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025




#20


Methylene Chloride  
Concen: 18.988 ug/l  
RT: 5.294 min Scan# 568  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 84 Resp: 52196  
Ion Ratio Lower Upper  
84 100  
49 108.4 90.5 135.7  
51 35.0 28.5 42.7  
86 64.6 48.1 72.1








Abundance Scan 656 (5.812 min): VN086944.D\data.ms




Abundance Scan 656 (5.812 min): VN086944.D\data.ms (-60)



Abundance Scan 805 (6.688 min): VN086944.D\data.ms



Abundance Scan 805 (6.688 min): VN086944.D\data.ms (-75)



#21

trans-1,2-Dichloroethene

Concen: 18.464 ug/l

RT: 5.812 min Scan# 6

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N


ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#22

Diisopropyl ether

Concen: 19.326 ug/l

RT: 6.688 min Scan# 805

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 45 Resp: 155508


Ion Ratio Lower Upper




45 100

43 50.5 41.9 62.9

87 29.8 22.4 33.6

59 12.1 9.8 14.6





#23

**Vinyl Acetate**

Concen: 96.092 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#24

**1,1-Dichloroethane**

Concen: 19.421 ug/l

RT: 6.588 min Scan# 788

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion: 63 Resp: 89930







Ion Ratio Lower Upper

63 100

98 7.3 3.6 10.8

100 5.3 2.6 7.8





#25

2-Butanone

Concen: 90.401 ug/l

RT: 7.494 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N





ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#26

2,2-Dichloropropane

Concen: 21.663 ug/l


RT: 7.500 min Scan# 943




Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 77 Resp: 78031  
 Ion Ratio Lower Upper  
 77 100  
 97 24.3 11.6 34.7









#27  
cis-1,2-Dichloroethene  
Concen: 19.453 ug/l  
RT: 7.500 min Scan# 943  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#28  
Bromochloromethane  
Concen: 19.909 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 49 Resp: 45329  
Ion Ratio Lower Upper  
49 100  
129 1.6 0.0 3.8  
130 82.2 62.2 93.2





#29

Tetrahydrofuran

Concen: 92.532 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N





ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#30

Chloroform

Concen: 19.521 ug/l


RT: 7.977 min Scan# 1024




Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 83 Resp: 90272  
 Ion Ratio Lower Upper  
 83 100  
 85 64.2 52.9 79.3





#31

Cyclohexane

Concen: 17.603 ug/l

RT: 8.265 min Scan# 1058

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#32

1,1,1-Trichloroethane

Concen: 19.033 ug/l


RT: 8.182 min Scan# 1059




Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt	Ion	Resp:	
			Lower
			Upper
	97	74861	
	100		
99	64.2	52.0	78.0
61	47.9	35.9	53.9








#33

1,2-Dichloroethane-d4  
Concen: 47.075 ug/l  
RT: 8.588 min Scan# 1128  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:114 Resp: 370524  
Ion Ratio Lower Upper  
114 100  
63 18.7 0.0 39.6  
88 16.4 0.0 30.2





#35

Dibromofluoromethane

Concen: 51.096 ug/l

RT: 8.177 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#36

1,1-Dichloropropene

Concen: 19.431 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion: 75 Resp: 63577




Ion Ratio Lower Upper

75 100

110 35.1 17.6 52.9

77 30.1 25.3 37.9





#37

**Ethyl Acetate**

Concen: 19.306 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :

VN0611WBS02

Tgt Ion: 43 Resp: 8041

Ion Ratio Lower Upper

43 100





61 15.4 12.1 18.1

70 12.3 9.8 14.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#38

**Carbon Tetrachloride**

Concen: 19.222 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion:117 Resp: 61747







Ion Ratio Lower Upper

117 100

119 97.1 75.0 112.6

121 32.6 24.9 37.3





#39

Methylcyclohexane

Concen: 16.947 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N


ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carbone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#40

Benzene

Concen: 19.277 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086944.D


Acq: 11 Jun 2025 13:27




Tgt Ion: 78 Resp: 206255

Ion Ratio Lower Upper

78 100

77 23.1 18.7 28.1





#41

Methacrylonitrile

Concen: 18.218 ug/l

RT: 7.788 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :





VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025

Abundance



#42

1,2-Dichloroethane

Concen: 19.530 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion: 62 Resp: 63379


Ion Ratio Lower Upper

62 100

98 11.4 0.0 21.2

Abundance





#43

Isopropyl Acetate



Concen: 18.984 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27



Tgt Ion: 43 Resp: 12691

Ion Ratio Lower Upper

43 100



61 28.5 22.1 33.1

87 15.6 11.8 17.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#44

Trichloroethene


Concen: 19.942 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27






Tgt Ion:130 Resp: 50593

Ion Ratio Lower Upper


130 100

95 99.6 0.0 196.2








Abundance Scan 1304 (9.624 min): VN086944.D\data.ms




Abundance Scan 1304 (9.624 min): VN086944.D\data.ms (-1)



Abundance Scan 1319 (9.712 min): VN086944.D\data.ms



Abundance Scan 1319 (9.712 min): VN086944.D\data.ms (-1)



#45

1,2-Dichloropropane

Concen: 19.628 ug/l

RT: 9.624 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N


ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#46

Dibromomethane

Concen: 20.093 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion: 93 Resp: 34741


Ion Ratio Lower Upper

93 100

95 85.1 67.1 100.7

174 94.0 74.2 111.4







#47

Bromodichloromethane  
Concen: 19.634 ug/l  
RT: 9.894 min Scan# 1350  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27



Instrument: MSVOA\_N  
Client SampleId: VN0611WBS02




Tgt Ion: 83 Resp: 69845  
Ion Ratio Lower Upper  
83 100  
85 65.0 50.9 76.3  
127 8.1 6.6 10.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#48  
Methyl methacrylate  
Concen: 18.771 ug/l  
RT: 9.682 min Scan# 1314  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27



Tgt Ion: 41 Resp: 57757  
Ion Ratio Lower Upper  
41 100  
69 93.3 74.6 112.0  
39 48.5 40.2 60.4









#49  
1,4-Dioxane  
Concen: 409.347 ug/l  
RT: 9.706 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#50  
Toluene-d8  
Concen: 48.706 ug/l  
RT: 10.571 min Scan# 1465  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 98 Resp: 423380  
Ion Ratio Lower Upper  
98 100  
100 67.3 53.4 80.0





#51

4-Methyl-2-Pentanone

Concen: 97.997 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N





ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#52

Toluene

Concen: 19.509 ug/l


RT: 10.629 min Scan# 1475




Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 92 Resp: 127563  
 Ion Ratio Lower Upper  
 92 100  
 91 171.7 136.5 204.7









#53  
t-1,3-Dichloropropene  
Concen: 19.917 ug/l  
RT: 10.835 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0611WBS02


1 Manual Integrations  
2 APPROVED




3 Reviewed By :John Carlone 06/12/2025  
4 Supervised By :Semsettin Yesilyurt 06/12/2025



#54  
cis-1,3-Dichloropropene  
Concen: 20.109 ug/l  
RT: 10.318 min Scan# 1422  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 85586  
Ion Ratio Lower Upper  
75 100  
77 33.6 26.2 39.4  
39 39.5 32.3 48.5









#55

1,1,2-Trichloroethane  
Concen: 20.438 ug/l  
RT: 11.018 min Scan# 1541  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#56

Ethyl methacrylate  
Concen: 20.198 ug/l  
RT: 10.882 min Scan# 1518  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 69 Resp: 80942  
Ion Ratio Lower Upper  
69 100  
41 55.5 47.2 70.8  
39 24.3 23.0 34.6









#57

1,3-Dichloropropane  
Concen: 19.613 ug/l  
RT: 11.165 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025






#58

2-Chloroethyl Vinyl ether  
Concen: 101.910 ug/l  
RT: 10.165 min Scan# 1396  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 63 Resp: 243584  
Ion Ratio Lower Upper  
63 100  
106 29.2 23.0 34.6









#59  
2-Hexanone  
Concen: 89.210 ug/l  
RT: 11.206 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0611WBS02


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#60  
Dibromochloromethane  
Concen: 20.468 ug/l  
RT: 11.359 min Scan# 1599  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:129 Resp: 53655  
Ion Ratio Lower Upper  
129 100  
127 76.3 38.6 115.8









#61  
1,2-Dibromoethane  
Concen: 19.711 ug/l  
RT: 11.471 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


### Manual Integrations APPROVED


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#62  
4-Bromofluorobenzene  
Concen: 49.951 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 95 Resp: 161321  
Ion Ratio Lower Upper  
95 100  
174 71.0 0.0 141.8  
176 67.2 0.0 132.6







#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.865 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#64  
Tetrachloroethene  
Concen: 19.150 ug/l  
RT: 11.106 min Scan# 1556  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:164 Resp: 38926  
Ion Ratio Lower Upper  
164 100  
166 128.0 101.8 152.8  
129 106.5 83.0 124.4  
131 100.0 78.2 117.2







#65  
Chlorobenzene  
Concen: 20.006 ug/l  
RT: 11.894 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 20.378 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:131 Resp: 46404  
Ion Ratio Lower Upper  
131 100  
133 94.2 48.3 144.8  
119 68.0 34.0 101.8



#67

Ethyl Benzene

Concen: 19.647 ug/l

RT: 11.965 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D


Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :




VN0611WBS02



Tgt Ion: 91 Resp: 23973  
 Ion Ratio Lower Upper  
 91 100  
 106 31.9 25.4 38.2





### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
 Supervised By :Semsettin Yesilyurt 06/12/2025



#68  
 m/p-Xylenes  
 Concen: 39.613 ug/l  
 RT: 12.070 min Scan# 1720  
 Delta R.T. -0.000 min  
 Lab File: VN086944.D  
 Acq: 11 Jun 2025 13:27

Tgt Ion:106 Resp: 185026  
 Ion Ratio Lower Upper  
 106 100  
 91 197.2 159.4 239.0









#69  
o-Xylene  
Concen: 20.172 ug/l  
RT: 12.400 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0611WBS02


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#70  
Styrene  
Concen: 20.137 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:104 Resp: 154149  
Ion Ratio Lower Upper  
104 100  
78 47.5 38.5 57.7  
103 54.3 43.1 64.7





#71

Bromoform

Concen: 20.952 ug/l

RT: 12.582 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion:152 Resp: 157403




Ion Ratio Lower Upper

152 100

115 61.4 30.1 90.5

150 164.6 0.0 345.0





#73

Isopropylbenzene

Concen: 19.320 ug/l

RT: 12.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#74

N-amyl acetate

Concen: 15.794 ug/l

RT: 12.529 min Scan# 1798

Delta R.T. 0.017 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt Ion: 43 Resp: 63289


Ion Ratio Lower Upper




43 100

70 34.9 37.6 56.4#

55 19.9 19.0 28.6

61 18.0 20.8 31.2#









#75  
1,1,2,2-Tetrachloroethane  
Concen: 20.973 ug/l  
RT: 12.941 min Scan# 1868  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#76  
1,2,3-Trichloropropane  
Concen: 19.132 ug/l  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 75 Resp: 71582  
Ion Ratio Lower Upper  
75 100  
77 177.6 79.3 237.9









#77  
Bromobenzene  
Concen: 20.981 ug/l  
RT: 12.982 min Scan# 1875  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


### Manual Integrations APPROVED

Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#78  
n-propylbenzene  
Concen: 19.187 ug/l  
RT: 13.035 min Scan# 1884  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 91 Resp: 267384  
Ion Ratio Lower Upper  
91 100  
120 22.9 11.4 34.2



#79

2-Chlorotoluene

Concen: 19.721 ug/l

RT: 13.123 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :

VN0611WBS02



Tgt Ion: 91 Resp: 16481

Ion Ratio Lower Upper

91 100





126 32.6 16.2 48.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025


#80  
1,3,5-Trimethylbenzene  
Concen: 19.200 ug/l  
RT: 13.170 min Scan# 1907  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27




Tgt Ion:105 Resp: 181773

Ion Ratio Lower Upper

105 100

120 50.4 24.9 74.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 18.532 ug/l  
RT: 12.735 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument : MSVOA\_N  
ClientSampleId : VN0611WBS02


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/12/2025  
Supervised By :Semsettin Yesilyurt 06/12/2025



#82  
4-Chlorotoluene  
Concen: 19.866 ug/l  
RT: 13.217 min Scan# 1915  
Delta R.T. -0.006 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion: 91 Resp: 167991  
Ion Ratio Lower Upper  
91 100  
126 33.4 16.7 50.1





#83

tert-Butylbenzene

Concen: 18.646 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#84

1,2,4-Trimethylbenzene

Concen: 19.216 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN086944.D


Acq: 11 Jun 2025 13:27




Tgt Ion:105 Resp: 182431

Ion Ratio Lower Upper

105 100

120 47.6 23.2 69.6









#85  
sec-Butylbenzene  
Concen: 18.189 ug/l  
RT: 13.617 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0611WBS02


1 Manual Integrations  
2 APPROVED

3 Reviewed By :John Carlone 06/12/2025  
4 Supervised By :Semsettin Yesilyurt 06/12/2025



#86  
p-Isopropyltoluene  
Concen: 18.551 ug/l  
RT: 13.729 min Scan# 2002  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:119 Resp: 192991  
Ion Ratio Lower Upper  
119 100  
134 26.2 13.5 40.4  
91 22.1 11.5 34.4



#87

1,3-Dichlorobenzene

Concen: 20.271 ug/l

RT: 13.735 min Scan# 2

Delta R.T. -0.000 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :

VN0611WBS02



Tgt Ion:146 Resp: 104883

Ion Ratio Lower Upper

146 100

111 41.3 20.9 62.8






148 62.6 32.0 96.2




Manual Integrations

APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025

#88  
1,4-Dichlorobenzene  
Concen: 20.426 ug/l  
RT: 13.812 min Scan# 2016  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27Tgt Ion:146 Resp: 107750  
Ion Ratio Lower Upper  
146 100  
111 40.7 19.9 59.7  
148 63.1 31.8 95.4



#89

n-Butylbenzene

Concen: 17.613 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N





ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#90

Hexachloroethane

Concen: 18.972 ug/l


RT: 14.329 min Scan# 2104




Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt	Ion:117	Resp:	33456
Ion	Ratio	Lower	Upper
117	100		
201	67.1	33.2	99.6





#91

1,2-Dichlorobenzene

Concen: 19.995 ug/l

RT: 14.106 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N




ClientSampleId :

VN0611WBS02

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.155 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27


Tgt Ion: 75 Resp: 17796

Ion Ratio Lower Upper

75 100

155 85.2 41.1 123.4

157 106.8 52.7 158.1



#93

1,2,4-Trichlorobenzene

Concen: 18.433 ug/l

RT: 15.394 min Scan# 2

Delta R.T. 0.006 min

Lab File: VN086944.D




Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N

ClientSampleId :

VN0611WBS02



Tgt Ion:180 Resp: 58509

Ion Ratio Lower Upper

180 100





182 96.7 46.9 140.7

145 32.3 16.0 47.9

**Manual Integrations  
APPROVED**


Reviewed By :John Carlone 06/12/2025







Supervised By :Semsettin Yesilyurt 06/12/2025



#94  
Hexachlorobutadiene  
Concen: 16.567 ug/l  
RT: 15.500 min Scan# 2303  
Delta R.T. -0.000 min  
Lab File: VN086944.D  
Acq: 11 Jun 2025 13:27

Tgt Ion:225 Resp: 19592  
Ion Ratio Lower Upper  
225 100  
223 64.0 33.5 100.4  
227 60.0 31.1 93.5





#95

Naphthalene

Concen: 19.700 ug/l

RT: 15.635 min Scan# 2327

Delta R.T. -0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Instrument:

MSVOA\_N


ClientSampleId :

VN0611WBS02

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/12/2025

Supervised By :Semsettin Yesilyurt 06/12/2025



#96

1,2,3-Trichlorobenzene

Concen: 17.802 ug/l


RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN086944.D

Acq: 11 Jun 2025 13:27

Tgt	Ion:180	Resp:	56144
Ion	Ratio	Lower	Upper
180	100		
182	96.0	47.0	141.0
145	33.6	16.6	49.8





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:	06/05/25	
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:	06/05/25	
Client Sample ID:	MW-11A-13.5-060525MS			SDG No.:	Q2250	
Lab Sample ID:	Q2250-02MS			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:	uL			Test:	VOCMS Group3	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086922.D	1		06/10/25 12:51	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	50.4		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	47.9		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	45.2		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1900	E	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	37.2		0.20	1.00	ug/L
71-43-2	Benzene	39.1		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	39.3		0.22	1.00	ug/L
79-01-6	Trichloroethene	2900	E	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	40.0		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	57.0		0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	49.9		70 (74) - 130 (125)	100%	SPK: 50
1868-53-7	Dibromofluoromethane	52.4		70 (75) - 130 (124)	105%	SPK: 50
2037-26-5	Toluene-d8	51.2		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		70 (77) - 130 (121)	102%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	174000	8.23			
540-36-3	1,4-Difluorobenzene	311000	9.106			
3114-55-4	Chlorobenzene-d5	271000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	133000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086922.D  
 Acq On : 10 Jun 2025 12:51  
 Operator : JC\MD  
 Sample : Q2250-02MS  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-11A-13.5-060525MS

Quant Time: Jun 11 01:52:35 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	174378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	310932	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132628	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.583	65	116443	49.874	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	99.740%	
35) Dibromofluoromethane	8.177	113	96506	52.373	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	104.740%	
50) Toluene-d8	10.571	98	373191	51.160	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	102.320%	
62) 4-Bromofluorobenzene	12.847	95	138034	50.932	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	101.860%	
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	68302	39.284	ug/l	96
3) Chloromethane	2.395	50	80842	36.007	ug/l	100
4) Vinyl Chloride	2.554	62	116824	50.442	ug/l	100
5) Bromomethane	3.001	94	51812	39.977	ug/l	94
6) Chloroethane	3.159	64	58869	39.351	ug/l	100
7) Trichlorofluoromethane	3.530	101	115932	38.313	ug/l	96
8) Diethyl Ether	3.989	74	51551	39.101	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.400	101	367438	193.276	ug/l	99
10) Methyl Iodide	4.618	142	79922	32.430	ug/l	97
11) Tert butyl alcohol	5.530	59	107683	169.980	ug/l	98
12) 1,1-Dichloroethene	4.371	96	92890	47.860	ug/l	93
13) Acrolein	4.200	56	33731	168.212	ug/l	98
14) Allyl chloride	5.047	41	117940	36.641	ug/l	96
15) Acrylonitrile	5.736	53	284483	192.124	ug/l	98
16) Acetone	4.448	43	215008	173.656	ug/l	99
17) Carbon Disulfide	4.742	76	196337	36.571	ug/l	98
18) Methyl Acetate	5.047	43	132769	36.798	ug/l	99
19) Methyl tert-butyl Ether	5.812	73	271783	38.674	ug/l	98
20) Methylene Chloride	5.300	84	87753	37.857	ug/l	96
21) trans-1,2-Dichloroethene	5.806	96	102714	47.566	ug/l	97
22) Diisopropyl ether	6.689	45	255658	37.680	ug/l	99
23) Vinyl Acetate	6.618	43	1062011	185.258	ug/l	99
24) 1,1-Dichloroethane	6.583	63	176376	45.171	ug/l	99
25) 2-Butanone	7.494	43	364801	181.265	ug/l	# 80
26) 2,2-Dichloropropane	7.506	77	128482	42.301	ug/l	# 1
27) cis-1,2-Dichloroethene	7.500	96	4853647	1879.430	ug/l	90
28) Bromochloromethane	7.824	49	78797	41.044	ug/l	93
29) Tetrahydrofuran	7.847	42	239404	182.609	ug/l	96
30) Chloroform	7.977	83	151067	38.741	ug/l	99
31) Cyclohexane	8.265	56	123047	32.495	ug/l	99
32) 1,1,1-Trichloroethane	8.177	97	123440	37.219	ug/l	98
36) 1,1-Dichloropropene	8.377	75	106081	38.635	ug/l	100
37) Ethyl Acetate	7.571	43	126676	36.242	ug/l	# 79
38) Carbon Tetrachloride	8.371	117	102158	37.897	ug/l	96
39) Methylcyclohexane	9.606	83	118983	31.630	ug/l	95
40) Benzene	8.612	78	351060	39.099	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086922.D  
 Acq On : 10 Jun 2025 12:51  
 Operator : JC\MD  
 Sample : Q2250-02MS  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-11A-13.5-060525MS

Quant Time: Jun 11 01:52:35 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	70459	35.893	ug/l	94
42) 1,2-Dichloroethane	8.677	62	106970	39.279	ug/l	100
43) Isopropyl Acetate	8.694	43	203030	36.190	ug/l	98
44) Trichloroethene	9.359	130	6264273	2942.351	ug/l	98
45) 1,2-Dichloropropane	9.624	63	84220	38.556	ug/l	99
46) Dibromomethane	9.712	93	58291	40.176	ug/l	99
47) Bromodichloromethane	9.888	83	115490	38.688	ug/l	98
48) Methyl methacrylate	9.682	41	95039	36.807	ug/l	97
49) 1,4-Dioxane	9.700	88	37276	793.544	ug/l	#
51) 4-Methyl-2-Pentanone	10.447	43	627690	185.856	ug/l	97
52) Toluene	10.629	92	211099	38.472	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	134072	40.165	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	141166	39.524	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	84553	40.047	ug/l	98
56) Ethyl methacrylate	10.882	69	138783	41.269	ug/l	95
57) 1,3-Dichloropropane	11.165	76	142808	38.991	ug/l	98
59) 2-Hexanone	11.200	43	410712	188.796	ug/l	94
60) Dibromochloromethane	11.359	129	88571	40.264	ug/l	98
61) 1,2-Dibromoethane	11.471	107	85449	39.485	ug/l	98
64) Tetrachloroethene	11.106	164	97793	57.032	ug/l	97
65) Chlorobenzene	11.894	112	236181	39.523	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	74715	38.896	ug/l	99
67) Ethyl Benzene	11.965	91	387237	37.621	ug/l	99
68) m/p-Xylenes	12.071	106	305789	77.607	ug/l	97
69) o-Xylene	12.394	106	147363	39.050	ug/l	98
70) Styrene	12.412	104	254157	39.358	ug/l	100
71) Bromoform	12.576	173	59843	42.046	ug/l	#
73) Isopropylbenzene	12.694	105	356009	36.846	ug/l	100
74) N-amyl acetate	12.512	43	142895m	42.322	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.935	83	131015	40.025	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	111129m	35.250	ug/l	
77) Bromobenzene	12.982	156	89530	40.405	ug/l	98
78) n-propylbenzene	13.035	91	421170	35.869	ug/l	99
79) 2-Chlorotoluene	13.123	91	262472	37.274	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	287840	36.083	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	56257	41.077	ug/l	97
82) 4-Chlorotoluene	13.218	91	272044	38.181	ug/l	100
83) tert-Butylbenzene	13.435	119	247652	33.915	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	294041	36.759	ug/l	100
85) sec-Butylbenzene	13.612	105	353284	33.315	ug/l	100
86) p-Isopropyltoluene	13.723	119	297463	33.934	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	165194	37.892	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	172677	38.850	ug/l	99
89) n-Butylbenzene	14.053	91	270209	31.824	ug/l	100
90) Hexachloroethane	14.329	117	52840	35.562	ug/l	96
91) 1,2-Dichlorobenzene	14.106	146	161871	38.647	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	28266	36.108	ug/l	96
93) 1,2,4-Trichlorobenzene	15.388	180	89607	33.503	ug/l	99
94) Hexachlorobutadiene	15.494	225	26250	26.343	ug/l	99
95) Naphthalene	15.635	128	365200	36.684	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	85176	32.053	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086922.D  
 Acq On : 10 Jun 2025 12:51  
 Operator : JC\MD  
 Sample : Q2250-02MS  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
MW-11A-13.5-060525MS

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Quant Time: Jun 11 01:52:35 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
(#)	= qualifier out of range	(m)	= manual integration	(+)	= signals summed	


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086922.D  
Acq On : 10 Jun 2025 12:51  
Operator : JC\MD  
Sample : Q2250-02MS  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 11 Sample Multiplier: 1




Quant Time: Jun 11 01:52:35 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
MW-11A-13.5-060525MSS

## Manual Integrations APPROVED

Reviewed By :John Caralone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.230 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#2

Dichlorodifluoromethane

Concen: 39.284 ug/l

RT: 2.154 min Scan# 34

Delta R.T. -0.000 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51




Tgt Ion: 85 Resp: 68302

Ion Ratio Lower Upper

85 100

87 30.1 16.2 48.6









#3  
Chloromethane  
Concen: 36.007 ug/l  
RT: 2.395 min Scan# 7  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#4  
Vinyl Chloride  
Concen: 50.442 ug/l  
RT: 2.554 min Scan# 102  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51


Tgt Ion: 62 Resp: 116824  
Ion Ratio Lower Upper  
62 100  
64 32.5 26.0 39.0





#5  
Bromomethane  
Concen: 39.977 ug/l  
RT: 3.001 min Scan# 1  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS




Tgt Ion: 94 Resp: 5181  
Ion Ratio Lower Upper  
94 100  
96 91.1 77.9 116.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#6  
Chloroethane  
Concen: 39.351 ug/l  
RT: 3.159 min Scan# 205  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Tgt Ion: 64 Resp: 58869  
Ion Ratio Lower Upper  
64 100  
66 30.9 24.6 36.8





#7  
Trichlorofluoromethane  
Concen: 38.313 ug/l  
RT: 3.530 min Scan# 2  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS




Tgt Ion:101 Resp: 11593  
Ion Ratio Lower Upper  
101 100  
103 61.9 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#8  
Diethyl Ether  
Concen: 39.101 ug/l  
RT: 3.989 min Scan# 346  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Tgt Ion: 74 Resp: 51551  
Ion Ratio Lower Upper  
74 100  
45 88.6 45.5 136.5





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 193.276 ug/l  
RT: 4.400 min Scan# 4  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS




Tgt Ion:101 Resp: 367438  
Ion Ratio Lower Upper  
101 100  
85 41.8 33.8 50.8  
151 74.1 58.7 88.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#10  
Methyl Iodide  
Concen: 32.430 ug/l  
RT: 4.618 min Scan# 453  
Delta R.T. 0.012 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51



Tgt Ion:142 Resp: 79922  
Ion Ratio Lower Upper  
142 100  
127 36.8 30.9 46.3  
141 14.3 10.4 15.6








#11

Tert butyl alcohol  
Concen: 169.980 ug/l  
RT: 5.530 min Scan# 6

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS  
Acq: 10 Jun 2025 12:51

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#12

1,1-Dichloroethene  
Concen: 47.860 ug/l  
RT: 4.371 min Scan# 411  
Delta R.T. 0.012 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51



Tgt Ion: 96 Resp: 92890  
Ion Ratio Lower Upper  
96 100  
61 147.5 125.8 188.6  
98 61.2 52.5 78.7





#13



Acrolein  
Concen: 168.212 ug/l  
RT: 4.200 min Scan# 382  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Tgt Ion: 56 Resp: 3373  
Ion Ratio Lower Upper  
56 100  
55 68.1 55.6 83.4

### Manual Integrations APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#14  
Allyl chloride  
Concen: 36.641 ug/l  
RT: 5.047 min Scan# 526  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51



Tgt Ion: 41 Resp: 117940  
Ion Ratio Lower Upper  
41 100  
39 60.5 51.3 76.9  
76 41.5 31.4 47.0





#15

Acrylonitrile

Concen: 192.124 ug/l

RT: 5.736 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086922.D





Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#16

Acetone

Concen: 173.656 ug/l


RT: 4.448 min Scan# 424




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 43 Resp: 215008  
 Ion Ratio Lower Upper  
 43 100  
 58 35.8 28.0 42.0





#17

Carbon Disulfide

Concen: 36.571 ug/l

RT: 4.742 min Scan# 4

Delta R.T. 0.012 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#18

Methyl Acetate

Concen: 36.798 ug/l


RT: 5.047 min Scan# 526




Delta R.T. 0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 43 Resp: 132769  
 Ion Ratio Lower Upper  
 43 100  
 74 25.5 20.7 31.1








#19

Methyl tert-butyl Ether  
Concen: 38.674 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument: MSVOA\_N  
ClientSampleId: MW-11A-13.5-060525MS



### Manual Integrations APPROVED





Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#20  
Methylene Chloride  
Concen: 37.857 ug/l  
RT: 5.300 min Scan# 569  
Delta R.T. 0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion: 84 Resp: 87753  
Ion Ratio Lower Upper  
84 100  
49 108.7 90.5 135.7  
51 33.5 28.5 42.7  
86 62.3 48.1 72.1






#21

trans-1,2-Dichloroethene  
Concen: 47.566 ug/l  
RT: 5.806 min Scan# 6

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS  
Acq: 10 Jun 2025 12:51

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#22

Diisopropyl ether  
Concen: 37.680 ug/l  
RT: 6.689 min Scan# 805  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion: 45 Resp: 255658  
Ion Ratio Lower Upper  
45 100  
43 51.5 41.9 62.9  
87 28.6 22.4 33.6  
59 12.9 9.8 14.6





#23

Vinyl Acetate

Concen: 185.258 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min





Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-11A-13.5-060525MS

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#24

1,1-Dichloroethane

Concen: 45.171 ug/l


RT: 6.583 min Scan# 787




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 63 Resp: 176376  
Ion Ratio Lower Upper  
63 100  
98 7.0 3.6 10.8  
100 4.7 2.6 7.8





#25

2-Butanone

Concen: 181.265 ug/l

RT: 7.494 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#26

2,2-Dichloropropane

Concen: 42.301 ug/l

RT: 7.506 min Scan# 944


Delta R.T. 0.006 min


Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 77 Resp: 128482


Ion	Ratio	Lower	Upper
77	100		
97	156.1	11.6	34.7





#27  
cis-1,2-Dichloroethene  
Concen: 1879.430 ug/l  
RT: 7.500 min Scan# 943  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51



Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS



Tgt Ion: 96 Resp: 485364  
Ion Ratio Lower Upper  
96 100  
61 121.8 0.0 278.0  
98 64.3 0.0 128.2



Manual Integrations  
APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#28  
Bromochloromethane  
Concen: 41.044 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion: 49 Resp: 78797  
Ion Ratio Lower Upper  
49 100  
129 2.1 0.0 3.8  
130 83.7 62.2 93.2





#29

Tetrahydrofuran

Concen: 182.609 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#30

Chloroform

Concen: 38.741 ug/l


RT: 7.977 min Scan# 1024




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 83 Resp: 151067  
 Ion Ratio Lower Upper  
 83 100  
 85 65.5 52.9 79.3





#31

Cyclohexane

Concen: 32.495 ug/l

RT: 8.265 min Scan# 1074

Delta R.T. -0.006 min





Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument: MSVOA\_N  
ClientSampleId: MW-11A-13.5-060525MS

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#32

1,1,1-Trichloroethane

Concen: 37.219 ug/l


RT: 8.177 min Scan# 1058







Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 97 Resp: 123440  
Ion Ratio Lower Upper  
97 100  
99 62.0 52.0 78.0  
61 44.9 35.9 53.9





#33

1,2-Dichloroethane-d4

Concen: 49.874 ug/l

RT: 8.583 min Scan# 1

Delta R.T. -0.006 min


Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Tgt Ion:114 Resp: 310932







Ion Ratio Lower Upper

114 100

63 20.0 0.0 39.6

88 16.3 0.0 30.2





#35

Dibromofluoromethane

Concen: 52.373 ug/l

RT: 8.177 min Scan# 1

Delta R.T. 0.006 min


Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#36

1,1-Dichloropropene

Concen: 38.635 ug/l


RT: 8.377 min Scan# 1092




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 75 Resp: 106081  
Ion Ratio Lower Upper  
75 100  
110 35.0 17.6 52.9  
77 31.4 25.3 37.9





#37

**Ethyl Acetate**

Concen: 36.242 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#38

**Carbon Tetrachloride**

Concen: 37.897 ug/l


RT: 8.371 min Scan# 1091




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt	Ion:117	Resp:	102158
Ion	Ratio	Lower	Upper
117	100		
119	98.8	75.0	112.6
121	31.2	24.9	37.3





#39

Methylcyclohexane

Concen: 31.630 ug/l

RT: 9.606 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51





Instrument: MSVOA\_N

ClientSampleId: MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#40

Benzene

Concen: 39.099 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51




Tgt Ion: 78 Resp: 351060

Ion Ratio Lower Upper

78 100

77 22.5 18.7 28.1





#41

Methacrylonitrile

Concen: 35.893 ug/l

RT: 7.788 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument: MSVOA\_N

ClientSampleId: MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

Abundance

40000

30000




20000

10000

0

7.788

Time--&gt;



#42

1,2-Dichloroethane

Concen: 39.279 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 62 Resp: 106970

Ion Ratio Lower Upper


62 100

98 10.4 0.0 21.2

Abundance

8.677

Time--&gt;



#43

Isopropyl Acetate

Concen: 36.190 ug/l


RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS



Tgt Ion: 43 Resp: 203030

Ion Ratio Lower Upper




43 100

61 28.8 22.1 33.1


87 15.8 11.8 17.6

**Manual Integrations**  
**APPROVED**



Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025




#44  
 Trichloroethene  
 Concen: 2942.351 ug/l  
 RT: 9.359 min Scan# 1259  
 Delta R.T. -0.000 min  
 Lab File: VN086922.D  
 Acq: 10 Jun 2025 12:51



Tgt Ion:130 Resp: 6264273  
 Ion Ratio Lower Upper  
 130 100  
 95 99.9 0.0 196.2





#45

1,2-Dichloropropane

Concen: 38.556 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.006 min



Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS

### Manual Integrations APPROVED

Reviewed By :John Carbone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#46

Dibromomethane

Concen: 40.176 ug/l


RT: 9.712 min Scan# 1319


Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 93 Resp: 58291  
 Ion Ratio Lower Upper  
 93 100  
 95 85.4 67.1 100.7  
 174 94.1 74.2 111.4






#47

Bromodichloromethane  
Concen: 38.688 ug/l  
RT: 9.888 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51




Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS




Tgt Ion: 83 Resp: 115490  
Ion Ratio Lower Upper  
83 100  
85 61.7 50.9 76.3  
127 8.4 6.6 10.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#48  
Methyl methacrylate  
Concen: 36.807 ug/l  
RT: 9.682 min Scan# 1314  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51



Tgt Ion: 41 Resp: 95039  
Ion Ratio Lower Upper  
41 100  
69 96.9 74.6 112.0  
39 49.8 40.2 60.4





#49

1,4-Dioxane

Concen: 793.544 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#50

Toluene-d8

Concen: 51.160 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. -0.000 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51


Tgt Ion: 98 Resp: 373191

Ion Ratio Lower Upper


98 100

100 67.5 53.4 80.0






Abundance Scan 1444 (10.447 min): VN086922.D\data.ms (-)



Abundance Scan 1444 (10.447 min): VN086922.D\data.ms (-)



#51

4-Methyl-2-Pentanone

Concen: 185.856 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min



Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



Abundance Scan 1475 (10.629 min): VN086922.D\data.ms (-)



Abundance Scan 1475 (10.629 min): VN086922.D\data.ms (-)



#52

Toluene

Concen: 38.472 ug/l


RT: 10.629 min Scan# 1475


Delta R.T. -0.006 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51

Tgt Ion: 92 Resp: 211099  
 Ion Ratio Lower Upper  
 92 100  
 91 169.0 136.5 204.7








Abundance Scan 1510 (10.835 min): VN086922.D\data.ms (-)





Abundance Scan 1510 (10.835 min): VN086922.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086922.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086922.D\data.ms (-)



#53

t-1,3-Dichloropropene

Concen: 40.165 ug/l

RT: 10.835 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N

ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**


Reviewed By :John Carbone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

Tgt Ion: 75 Resp: 13407:  
Ion Ratio Lower Upper

75 100

77 31.9 26.4 39.6



#54

cis-1,3-Dichloropropene

Concen: 39.524 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Tgt Ion: 75 Resp: 141166




Ion Ratio Lower Upper

75 100

77 31.8 26.2 39.4

39 38.6 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 40.047 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.000 min





Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#56

Ethyl methacrylate

Concen: 41.269 ug/l


RT: 10.882 min Scan# 1518




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion: 69 Resp: 138783  
 Ion Ratio Lower Upper  
 69 100  
 41 55.4 47.2 70.8  
 39 25.7 23.0 34.6





#57

1,3-Dichloropropane

Concen: 38.991 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51





Instrument: MSVOA\_N

ClientSampleId : MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#59

2-Hexanone

Concen: 188.796 ug/l

RT: 11.200 min Scan# 1572

Delta R.T. -0.000 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51




Tgt Ion: 43 Resp: 410712

Ion Ratio Lower Upper

43 100

58 65.8 30.4 91.3





#60

Dibromochloromethane

Concen: 40.264 ug/l

RT: 11.359 min Scan# 1

Delta R.T. -0.000 min





Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument: MSVOA\_N

ClientSampleId: MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

 Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025


#61

1,2-Dibromoethane


Concen: 39.485 ug/l




RT: 11.471 min Scan# 1618

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

 Tgt Ion:107 Resp: 85449  
 Ion Ratio Lower Upper  
 107 100  
 109 93.7 76.3 114.5




#62

4-Bromofluorobenzene

Concen: 50.932 ug/l

RT: 12.847 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51





Instrument: MSVOA\_N

ClientSampleId: MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min


Lab File: VN086922.D




Acq: 10 Jun 2025 12:51

Tgt Ion:117 Resp: 270945

Ion Ratio Lower Upper

117	100		
82	54.4	44.6	67.0
119	31.3	25.5	38.3





#64

Tetrachloroethene

Concen: 57.032 ug/l

RT: 11.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N




ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#65

Chlorobenzene

Concen: 39.523 ug/l

RT: 11.894 min Scan# 1690

Delta R.T. -0.000 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51


Tgt Ion:112 Resp: 236181

Ion Ratio Lower Upper

112 100

114 31.8 26.2 39.2







#66  
1,1,1,2-Tetrachloroethane  
Concen: 38.896 ug/l  
RT: 11.959 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS


**Manual Integrations**  
**APPROVED**




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#67  
Ethyl Benzene  
Concen: 37.621 ug/l  
RT: 11.965 min Scan# 1702  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion: 91 Resp: 387237  
Ion Ratio Lower Upper  
91 100  
106 32.1 25.4 38.2









#68  
m/p-Xylenes  
Concen: 77.607 ug/l  
RT: 12.071 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#69  
o-Xylene  
Concen: 39.050 ug/l  
RT: 12.394 min Scan# 1775  
Delta R.T. -0.006 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion:106 Resp: 147363  
Ion Ratio Lower Upper  
106 100  
91 208.6 106.1 318.1





#70

Styrene

Concen: 39.358 ug/l

RT: 12.412 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#71

Bromoform

Concen: 42.046 ug/l


RT: 12.576 min Scan# 1806



Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt	Ion:173	Resp:	59843
Ion	Ratio	Lower	Upper
173	100		
175	48.3	24.1	72.4
254	0.0	0.0	0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#73

Isopropylbenzene

Concen: 36.846 ug/l

RT: 12.694 min Scan# 1826


Delta R.T. -0.000 min




Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion:105 Resp: 356009

Ion	Ratio	Lower	Upper
105	100		
120	26.9	13.5	40.4





#74

N-amyl acetate

Concen: 42.322 ug/l m

RT: 12.512 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 40.025 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.006 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51




Tgt Ion: 83 Resp: 131015

Ion Ratio Lower Upper

	100		
83	100		
131	9.9	4.9	14.7

	64.2	32.0	96.0
85	64.2	32.0	96.0





#76

1,2,3-Trichloropropane

Concen: 35.250 ug/l m

RT: 12.994 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carbone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#77

Bromobenzene

Concen: 40.405 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Tgt Ion:156 Resp: 89530




Ion Ratio Lower Upper

156 100

77 229.9 117.7 353.1

158 96.2 48.1 144.4





#78

n-propylbenzene

Concen: 35.869 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#79

2-Chlorotoluene

Concen: 37.274 ug/l


RT: 13.123 min Scan# 1899




Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt	Ion	91	100	Resp:	262472
Ion Ratio	Lower	126	33.7	Upper	48.5
		91	100		
		126	33.7	16.2	48.5








#80

1,3,5-Trimethylbenzene  
Concen: 36.083 ug/l  
RT: 13.170 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS



### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#81  
trans-1,4-Dichloro-2-butene  
Concen: 41.077 ug/l  
RT: 12.735 min Scan# 1833  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion: 75 Resp: 56257  
Ion Ratio Lower Upper  
75 100  
53 82.3 66.6 99.8  
89 41.1 36.7 55.1





#82

4-Chlorotoluene

Concen: 38.181 ug/l

RT: 13.218 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N




ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#83

tert-Butylbenzene

Concen: 33.915 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Tgt Ion:119 Resp: 247652







Ion Ratio Lower Upper

119 100

91 64.8 32.1 96.5

134 27.4 13.4 40.1





#84

1,2,4-Trimethylbenzene

Concen: 36.759 ug/l

RT: 13.482 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#85

sec-Butylbenzene

Concen: 33.315 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.006 min

Lab File: VN086922.D


Acq: 10 Jun 2025 12:51




Tgt Ion:105 Resp: 353284

Ion Ratio Lower Upper

105 100

134 20.2 10.0 30.0





#86

p-Isopropyltoluene

Concen: 33.934 ug/l

RT: 13.723 min Scan# 2001





Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MS  
 Manual Integrations  
**APPROVED**

Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#87

1,3-Dichlorobenzene

Concen: 37.892 ug/l


RT: 13.729 min Scan# 2002




Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion:146 Resp: 165194  
 Ion Ratio Lower Upper  
 146 100  
 111 42.1 20.9 62.8  
 148 64.7 32.0 96.2





#88

1,4-Dichlorobenzene

Concen: 38.850 ug/l

RT: 13.812 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#89

n-Butylbenzene

Concen: 31.824 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Tgt Ion: 91 Resp: 270209







Ion Ratio Lower Upper

91 100

92 54.9 27.5 82.4

134 26.4 13.0 39.0





#90

Hexachloroethane

Concen: 35.562 ug/l

RT: 14.329 min Scan# 2

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#91

1,2-Dichlorobenzene

Concen: 38.647 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51


Tgt Ion:146 Resp: 161871




Ion Ratio Lower Upper

146 100

111 43.0 21.6 64.6

148 62.1 31.7 95.1








#92  
1,2-Dibromo-3-Chloropropane  
Concen: 36.108 ug/l  
RT: 14.717 min Scan# 2170  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#93  
1,2,4-Trichlorobenzene  
Concen: 33.503 ug/l  
RT: 15.388 min Scan# 2284  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Tgt Ion:180 Resp: 89607  
Ion Ratio Lower Upper  
180 100  
182 93.1 46.9 140.7  
145 32.0 16.0 47.9





#94

Hexachlorobutadiene

Concen: 26.343 ug/l

RT: 15.494 min Scan# 2

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Instrument:

MSVOA\_N

ClientSampleId :




MW-11A-13.5-060525MS

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

Abundance



#95

Naphthalene

Concen: 36.684 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.006 min

Lab File: VN086922.D

Acq: 10 Jun 2025 12:51

Tgt Ion:128 Resp: 365200


Ion Ratio Lower Upper

128 100


127 12.5 10.2 15.2

129 10.6 8.8 13.2

Abundance



VN086922.D 82N060625W.M




#96  
1,2,3-Trichlorobenzene  
Concen: 32.053 ug/l  
RT: 15.835 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086922.D  
Acq: 10 Jun 2025 12:51

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MS

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.			Date Collected:	06/05/25
Project:	Former Schlumberger STC PTC Site D3868221			Date Received:	06/05/25
Client Sample ID:	MW-11A-13.5-060525MSD			SDG No.:	Q2250
Lab Sample ID:	Q2250-03MSD			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL			Test:	VOCMS Group3
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086923.D	1		06/10/25 13:12	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	57.2		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	53.1		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	49.5		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2100	E	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	41.4		0.20	1.00	ug/L
71-43-2	Benzene	43.0		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	43.1		0.22	1.00	ug/L
79-01-6	Trichloroethene	3300	E	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	44.5		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	66.0		0.23	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	52.7		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	56.6		70 (75) - 130 (124)	113%	SPK: 50
2037-26-5	Toluene-d8	54.8		70 (86) - 130 (113)	110%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.3		70 (77) - 130 (121)	111%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	173000	8.23			
540-36-3	1,4-Difluorobenzene	310000	9.106			
3114-55-4	Chlorobenzene-d5	267000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	130000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086923.D  
 Acq On : 10 Jun 2025 13:12  
 Operator : JC\MD  
 Sample : Q2250-03MSD  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-11A-13.5-060525MSD

Quant Time: Jun 11 01:53:29 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
<b>Internal Standards</b>						
1) Pentafluorobenzene	8.230	168	172955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	309901	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	266637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129873	50.000	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.588	65	121937	52.657	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 105.320%		
35) Dibromofluoromethane	8.177	113	103950	56.601	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 113.200%		
50) Toluene-d8	10.565	98	398352	54.791	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 109.580%		
62) 4-Bromofluorobenzene	12.847	95	149297	55.271	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 110.540%		
<b>Target Compounds</b>						
				Qvalue		
2) Dichlorodifluoromethane	2.154	85	73929	42.871	ug/l	99
3) Chloromethane	2.395	50	84867	38.111	ug/l	97
4) Vinyl Chloride	2.554	62	131440	57.220	ug/l	98
5) Bromomethane	3.001	94	58237	45.304	ug/l	96
6) Chloroethane	3.159	64	65054	43.843	ug/l	99
7) Trichlorofluoromethane	3.530	101	127412	42.453	ug/l	99
8) Diethyl Ether	3.989	74	57896	44.275	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.401	101	438614	232.614	ug/l	99
10) Methyl Iodide	4.618	142	90414	36.989	ug/l	97
11) Tert butyl alcohol	5.536	59	118532	188.644	ug/l	99
12) 1,1-Dichloroethene	4.371	96	102158	53.068	ug/l	93
13) Acrolein	4.195	56	33620	169.038	ug/l	100
14) Allyl chloride	5.053	41	127951	40.078	ug/l	96
15) Acrylonitrile	5.736	53	308878	210.315	ug/l	100
16) Acetone	4.442	43	233959	190.517	ug/l	98
17) Carbon Disulfide	4.742	76	217837	40.910	ug/l	100
18) Methyl Acetate	5.048	43	144814	40.466	ug/l	99
19) Methyl tert-butyl Ether	5.812	73	300398	43.098	ug/l	99
20) Methylene Chloride	5.300	84	95461	41.521	ug/l	96
21) trans-1,2-Dichloroethene	5.812	96	108472	50.646	ug/l	95
22) Diisopropyl ether	6.689	45	277373	41.216	ug/l	97
23) Vinyl Acetate	6.618	43	1159679	203.959	ug/l	100
24) 1,1-Dichloroethane	6.589	63	191588	49.470	ug/l	98
25) 2-Butanone	7.494	43	391789	196.276	ug/l #	80
26) 2,2-Dichloropropane	7.506	77	139381	46.267	ug/l #	1
27) cis-1,2-Dichloroethene	7.500	96	5451059	2128.127	ug/l	90
28) Bromochloromethane	7.824	49	74118	38.924	ug/l	97
29) Tetrahydrofuran	7.847	42	263616	202.731	ug/l	96
30) Chloroform	7.977	83	162638	42.052	ug/l	97
31) Cyclohexane	8.265	56	133904	35.653	ug/l	98
32) 1,1,1-Trichloroethane	8.177	97	136050	41.359	ug/l	99
36) 1,1-Dichloropropene	8.377	75	117527	42.946	ug/l	99
37) Ethyl Acetate	7.571	43	139485	40.039	ug/l #	79
38) Carbon Tetrachloride	8.371	117	113739	42.334	ug/l	97
39) Methylcyclohexane	9.606	83	133181	35.522	ug/l	97
40) Benzene	8.612	78	385021	43.024	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086923.D  
 Acq On : 10 Jun 2025 13:12  
 Operator : JC\MD  
 Sample : Q2250-03MSD  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-11A-13.5-060525MSD

Quant Time: Jun 11 01:53:29 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlane 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.789	41	82542	42.189	ug/1	97
42) 1,2-Dichloroethane	8.677	62	116885	43.063	ug/1	100
43) Isopropyl Acetate	8.694	43	222284	39.754	ug/1	96
44) Trichloroethene	9.359	130	6931651	3266.653	ug/1	99
45) 1,2-Dichloropropane	9.624	63	92441	42.460	ug/1	99
46) Dibromomethane	9.712	93	64838	44.837	ug/1	99
47) Bromodichloromethane	9.888	83	126533	42.528	ug/1	100
48) Methyl methacrylate	9.682	41	101839	39.572	ug/1	95
49) 1,4-Dioxane	9.700	88	40909	873.782	ug/1	#
51) 4-Methyl-2-Pentanone	10.447	43	702657	208.746	ug/1	99
52) Toluene	10.630	92	233021	42.608	ug/1	99
53) t-1,3-Dichloropropene	10.835	75	146110	43.917	ug/1	97
54) cis-1,3-Dichloropropene	10.312	75	154243	43.329	ug/1	97
55) 1,1,2-Trichloroethane	11.018	97	93633	44.495	ug/1	99
56) Ethyl methacrylate	10.877	69	153861	45.905	ug/1	95
57) 1,3-Dichloropropane	11.165	76	155775	42.673	ug/1	97
59) 2-Hexanone	11.200	43	455196	209.941	ug/1	95
60) Dibromochloromethane	11.359	129	99583	45.420	ug/1	99
61) 1,2-Dibromoethane	11.471	107	92981	43.108	ug/1	99
64) Tetrachloroethene	11.106	164	111429	66.034	ug/1	95
65) Chlorobenzene	11.894	112	259599	44.144	ug/1	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	84686	44.799	ug/1	99
67) Ethyl Benzene	11.965	91	427727	42.226	ug/1	99
68) m/p-Xylenes	12.071	106	333950	86.123	ug/1	98
69) o-Xylene	12.400	106	162838	43.848	ug/1	97
70) Styrene	12.412	104	282053	44.384	ug/1	99
71) Bromoform	12.576	173	65807	46.983	ug/1	#
73) Isopropylbenzene	12.694	105	394852	41.733	ug/1	100
74) N-amyl acetate	12.518	43	143729	43.473	ug/1	96
75) 1,1,2,2-Tetrachloroethane	12.935	83	144127	44.965	ug/1	99
76) 1,2,3-Trichloropropane	12.994	75	118377m	38.345	ug/1	
77) Bromobenzene	12.976	156	99031	45.640	ug/1	98
78) n-propylbenzene	13.035	91	464253	40.377	ug/1	98
79) 2-Chlorotoluene	13.123	91	289730	42.017	ug/1	98
80) 1,3,5-Trimethylbenzene	13.171	105	321135	41.111	ug/1	99
81) trans-1,4-Dichloro-2-b...	12.735	75	62181	46.366	ug/1	98
82) 4-Chlorotoluene	13.218	91	298311	42.756	ug/1	99
83) tert-Butylbenzene	13.435	119	273596	38.263	ug/1	99
84) 1,2,4-Trimethylbenzene	13.482	105	326290	41.655	ug/1	100
85) sec-Butylbenzene	13.612	105	394374	37.979	ug/1	99
86) p-Isopropyltoluene	13.729	119	332552	38.742	ug/1	99
87) 1,3-Dichlorobenzene	13.729	146	184900	43.312	ug/1	99
88) 1,4-Dichlorobenzene	13.812	146	191420	43.980	ug/1	99
89) n-Butylbenzene	14.053	91	302661	36.403	ug/1	99
90) Hexachloroethane	14.335	117	57752	39.692	ug/1	98
91) 1,2-Dichlorobenzene	14.106	146	177603	43.302	ug/1	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	31210	40.715	ug/1	94
93) 1,2,4-Trichlorobenzene	15.388	180	98501	37.609	ug/1	98
94) Hexachlorobutadiene	15.494	225	28911	29.629	ug/1	98
95) Naphthalene	15.635	128	399267	40.957	ug/1	100
96) 1,2,3-Trichlorobenzene	15.835	180	93291	35.852	ug/1	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
 Data File : VN086923.D  
 Acq On : 10 Jun 2025 13:12  
 Operator : JC\MD  
 Sample : Q2250-03MSD  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
MW-11A-13.5-060525MSD

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025

Quant Time: Jun 11 01:53:29 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jun 07 02:12:50 2025  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
(#) = qualifier out of range (m) = manual integration (+) = signals summed						


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061025\  
Data File : VN086923.D  
Acq On : 10 Jun 2025 13:12  
Operator : JC\MD  
Sample : Q2250-03MSD  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 12 Sample Multiplier: 1




Quant Time: Jun 11 01:53:29 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060625W.M  
Quant Title : SW846 8260  
QLast Update : Sat Jun 07 02:12:50 2025  
Response via : Initial Calibration

**Instrument :** MSVOA\_N  
**ClientSampleId :** MW-11A-13.5-060525MSD

## Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.230 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#2

Dichlorodifluoromethane

Concen: 42.871 ug/l


RT: 2.154 min Scan# 34




Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 85 Resp: 73929  
 Ion Ratio Lower Upper  
 85 100  
 87 31.9 16.2 48.6









#3  
Chloromethane  
Concen: 38.111 ug/l  
RT: 2.395 min Scan# 7  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#4  
Vinyl Chloride  
Concen: 57.220 ug/l  
RT: 2.554 min Scan# 102  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion: 62 Resp: 131440  
Ion Ratio Lower Upper  
62 100  
64 31.5 26.0 39.0





#5

Bromomethane

Concen: 45.304 ug/l

RT: 3.001 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#6

Chloroethane

Concen: 43.843 ug/l

RT: 3.159 min Scan# 205


Delta R.T. 0.006 min


Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 64 Resp: 65054


Ion	Ratio	Lower	Upper
64	100		
66	30.2	24.6	36.8





#7  
Trichlorofluoromethane  
Concen: 42.453 ug/l  
RT: 3.530 min Scan# 2  
Delta R.T. 0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12




Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD




Tgt Ion:101 Resp: 12741:  
Ion Ratio Lower Upper  
101 100  
103 64.1 52.2 78.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#8  
Diethyl Ether  
Concen: 44.275 ug/l  
RT: 3.989 min Scan# 346  
Delta R.T. 0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12




Tgt Ion: 74 Resp: 57896  
Ion Ratio Lower Upper  
74 100  
45 85.5 45.5 136.5





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 232.614 ug/l  
RT: 4.401 min Scan# 4  
Delta R.T. 0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12



Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD




Tgt Ion:101 Resp: 438614  
Ion Ratio Lower Upper  
101 100  
85 41.1 33.8 50.8  
151 73.4 58.7 88.1

Manual Integrations  
APPROVED





Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#10  
Methyl Iodide  
Concen: 36.989 ug/l  
RT: 4.618 min Scan# 453  
Delta R.T. 0.012 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12



Tgt Ion:142 Resp: 90414  
Ion Ratio Lower Upper  
142 100  
127 36.1 30.9 46.3  
141 13.1 10.4 15.6









#11

Tert butyl alcohol  
Concen: 188.644 ug/l  
RT: 5.536 min Scan# 6  
Delta R.T. -0.012 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#12


1,1-Dichloroethene  
Concen: 53.068 ug/l  
RT: 4.371 min Scan# 411  
Delta R.T. 0.012 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion: 96 Resp: 102158  
Ion Ratio Lower Upper  
96 100  
61 148.5 125.8 188.6  
98 60.2 52.5 78.7








Abundance Scan 381 (4.195 min): VN086923.D\data.ms




Abundance Scan 381 (4.195 min): VN086923.D\data.ms (-33)



Abundance Scan 527 (5.053 min): VN086923.D\data.ms



Abundance Scan 527 (5.053 min): VN086923.D\data.ms (-47)



#13

Acrolein

Concen: 169.038 ug/l

RT: 4.195 min Scan# 3

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#14

Allyl chloride

Concen: 40.078 ug/l

RT: 5.053 min Scan# 527

Delta R.T. 0.012 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 41 Resp: 127951




Ion Ratio Lower Upper

41 100

39 61.5 51.3 76.9

76 42.0 31.4 47.0





#15

Acrylonitrile

Concen: 210.315 ug/l

RT: 5.736 min Scan# 6

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#16

Acetone

Concen: 190.517 ug/l


RT: 4.442 min Scan# 423




Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 43 Resp: 233959  
 Ion Ratio Lower Upper  
 43 100  
 58 34.0 28.0 42.0





#17

Carbon Disulfide

Concen: 40.910 ug/l

RT: 4.742 min Scan# 4

Delta R.T. 0.012 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#18

Methyl Acetate

Concen: 40.466 ug/l

RT: 5.048 min Scan# 526

Delta R.T. 0.006 min

Lab File: VN086923.D


Acq: 10 Jun 2025 13:12


Tgt Ion: 43 Resp: 144814

Ion Ratio Lower Upper

43 100

74 26.3 20.7 31.1






#19

Methyl tert-butyl Ether  
Concen: 43.098 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12



Instrument: MSVOA\_N  
ClientSampleId: MW-11A-13.5-060525MSD




Tgt Ion: 73 Resp: 30039  
Ion Ratio Lower Upper  
73 100  
57 21.2 17.4 26.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#20  
Methylene Chloride  
Concen: 41.521 ug/l  
RT: 5.300 min Scan# 569  
Delta R.T. 0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12




Tgt Ion: 84 Resp: 95461  
Ion Ratio Lower Upper  
84 100  
49 107.3 90.5 135.7  
51 34.3 28.5 42.7  
86 62.3 48.1 72.1






Abundance Scan 656 (5.812 min): VN086923.D\data.ms



Abundance Scan 656 (5.812 min): VN086923.D\data.ms (-60)





#21

trans-1,2-Dichloroethene  
Concen: 50.646 ug/l  
RT: 5.812 min Scan# 6  
Delta R.T. 0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12


Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



Abundance Scan 805 (6.689 min): VN086923.D\data.ms




Abundance Scan 805 (6.689 min): VN086923.D\data.ms (-75)






#22

Diisopropyl ether  
Concen: 41.216 ug/l  
RT: 6.689 min Scan# 805  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion: 45 Resp: 277373  
Ion Ratio Lower Upper  
45 100  
43 51.1 41.9 62.9  
87 31.4 22.4 33.6  
59 12.0 9.8 14.6





#23

**Vinyl Acetate**

Concen: 203.959 ug/l

RT: 6.618 min Scan# 7

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#24

**1,1-Dichloroethane**

Concen: 49.470 ug/l

RT: 6.589 min Scan# 788


Delta R.T. 0.006 min




Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 63 Resp: 191588

Ion	Ratio	Lower	Upper
63	100		
98	7.7	3.6	10.8
100	4.3	2.6	7.8





#25

2-Butanone

Concen: 196.276 ug/l

RT: 7.494 min Scan# 9

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N




ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#26

2,2-Dichloropropane

Concen: 46.267 ug/l

RT: 7.506 min Scan# 944

Delta R.T. 0.006 min

Lab File: VN086923.D


Acq: 10 Jun 2025 13:12






Tgt Ion: 77 Resp: 139381

Ion Ratio Lower Upper

77 100

97 162.1 11.6 34.7#





#27

cis-1,2-Dichloroethene

Concen: 2128.127 ug/l

RT: 7.500 min Scan# 9

Delta R.T. -0.000 min


Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#28  
Bromochloromethane  
Concen: 38.924 ug/l  
RT: 7.824 min Scan# 998  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion: 49 Resp: 74118  
Ion Ratio Lower Upper  
49 100  
129 1.9 0.0 3.8  
130 80.7 62.2 93.2





#29

Tetrahydrofuran

Concen: 202.731 ug/l

RT: 7.847 min Scan# 1002

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#30

Chloroform

Concen: 42.052 ug/l


RT: 7.977 min Scan# 1024




Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 83 Resp: 162638  
 Ion Ratio Lower Upper  
 83 100  
 85 63.9 52.9 79.3





#31

Cyclohexane

Concen: 35.653 ug/l

RT: 8.265 min Scan# 1058

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#32

1,1,1-Trichloroethane

Concen: 41.359 ug/l

RT: 8.177 min Scan# 1058


Delta R.T. -0.000 min


Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 97 Resp: 136050

Ion	Ratio	Lower	Upper
97	100		
99	65.9	52.0	78.0
61	45.7	35.9	53.9






#33

1,2-Dichloroethane-d4  
Concen: 52.657 ug/l  
RT: 8.588 min Scan# 121931  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12




Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD




Tgt Ion: 65 Resp: 121931  
Ion Ratio Lower Upper  
65 100  
67 55.5 0.0 105.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12



Tgt Ion:114 Resp: 309901  
Ion Ratio Lower Upper  
114 100  
63 19.0 0.0 39.6  
88 16.1 0.0 30.2





#35

Dibromofluoromethane



Concen: 56.601 ug/l

RT: 8.177 min Scan# 103950

Delta R.T. 0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12



Tgt Ion: 113 Resp: 103950

Ion Ratio Lower Upper

113 100






111 103.1 84.2 126.2




192 17.4 14.2 21.4

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

#36  
1,1-Dichloropropene  
Concen: 42.946 ug/l  
RT: 8.377 min Scan# 1092  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12Tgt Ion: 75 Resp: 117527  
Ion Ratio Lower Upper  
75 100  
110 35.8 17.6 52.9  
77 30.8 25.3 37.9



#37

Ethyl Acetate

Concen: 40.039 ug/l

RT: 7.571 min Scan# 9

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#38

Carbon Tetrachloride

Concen: 42.334 ug/l


RT: 8.371 min Scan# 1091


Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt	Ion	117	Resp:	113739
Ion	Ratio	Lower	Upper	
117	100			
119	96.2	75.0	112.6	
121	29.4	24.9	37.3	





#39

Methylcyclohexane

Concen: 35.522 ug/l

RT: 9.606 min Scan# 1



Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument: MSVOA\_N

ClientSampleId: MW-11A-13.5-060525MSD



Tgt Ion: 83 Resp: 13318

Ion Ratio Lower Upper

83 100



55 71.9 61.0 91.6

98 48.0 38.9 58.3

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#40

Benzene

Concen: 43.024 ug/l

RT: 8.612 min Scan# 1132

Delta R.T. -0.000 min

Lab File: VN086923.D




Acq: 10 Jun 2025 13:12




Tgt Ion: 78 Resp: 385021

Ion Ratio Lower Upper

78 100

77 23.1 18.7 28.1





#41

Methacrylonitrile

Concen: 42.189 ug/l

RT: 7.789 min Scan# 992

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N

ClientSampleId :





MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025

Abundance



#42

1,2-Dichloroethane

Concen: 43.063 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 62 Resp: 116885


Ion Ratio Lower Upper

62 100

98 10.7 0.0 21.2

Abundance





#43

Isopropyl Acetate



Concen: 39.754 ug/l

RT: 8.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12



Tgt Ion: 43 Resp: 222284

Ion Ratio Lower Upper

43 100



61 29.8 22.1 33.1

87 15.9 11.8 17.6

**Manual Integrations****APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#44

Trichloroethene


Concen: 3266.653 ug/l

RT: 9.359 min Scan# 1259

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12





Tgt Ion:130 Resp: 6931651

Ion Ratio Lower Upper



130 100

95 98.9 0.0 196.2






Abundance Scan 1304 (9.624 min): VN086923.D\data.ms



Abundance Scan 1319 (9.712 min): VN086923.D\data.ms



#45

1,2-Dichloropropane

Concen: 42.460 ug/l

RT: 9.624 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#46

Dibromomethane

Concen: 44.837 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 93 Resp: 64838




Ion Ratio Lower Upper

93 100

95 84.6 67.1 100.7

174 92.0 74.2 111.4





#47

Bromodichloromethane

Concen: 42.528 ug/l

RT: 9.888 min Scan# 1349





Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MSD  
 Manual Integrations  
**APPROVED**

Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#48

Methyl methacrylate

Concen: 39.572 ug/l


RT: 9.682 min Scan# 1314




Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion: 41 Resp: 101839  
 Ion Ratio Lower Upper  
 41 100  
 69 100.5 74.6 112.0  
 39 51.5 40.2 60.4





#49

1,4-Dioxane

Concen: 873.782 ug/l

RT: 9.700 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N




ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**



Reviewed By :John Carlone 06/11/2025


Supervised By :Mahesh Dadoda 06/11/2025




#50  
Toluene-d8  
Concen: 54.791 ug/l  
RT: 10.565 min Scan# 1464  
Delta R.T. -0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion: 98 Resp: 398352  
Ion Ratio Lower Upper  
98 100  
100 67.1 53.4 80.0







Abundance Scan 1444 (10.447 min): VN086923.D\data.ms




Abundance Scan 1444 (10.447 min): VN086923.D\data.ms (-)



Abundance Scan 1475 (10.630 min): VN086923.D\data.ms



Abundance Scan 1475 (10.630 min): VN086923.D\data.ms (-)



#51

4-Methyl-2-Pentanone

Concen: 208.746 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#52

Toluene

Concen: 42.608 ug/l

RT: 10.630 min Scan# 1475

Delta R.T. -0.006 min

Lab File: VN086923.D


Acq: 10 Jun 2025 13:12


Tgt Ion: 92 Resp: 233021

Ion Ratio Lower Upper



92 100

91 168.9 136.5 204.7






Abundance Scan 1510 (10.835 min): VN086923.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086923.D\data.ms (-)



Abundance Scan 1421 (10.312 min): VN086923.D\data.ms (-)

#53

t-1,3-Dichloropropene

Concen: 43.917 ug/l

RT: 10.835 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#54

cis-1,3-Dichloropropene

Concen: 43.329 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 75 Resp: 154243




Ion Ratio Lower Upper

75 100

77 30.2 26.2 39.4

39 39.6 32.3 48.5





#55

1,1,2-Trichloroethane

Concen: 44.495 ug/l

RT: 11.018 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#56

Ethyl methacrylate

Concen: 45.905 ug/l

RT: 10.877 min Scan# 1517

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 69 Resp: 153861




Ion Ratio Lower Upper

69 100

41 56.1 47.2 70.8

39 25.1 23.0 34.6





#57

1,3-Dichloropropane

Concen: 42.673 ug/l

RT: 11.165 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carbone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#59

2-Hexanone

Concen: 209.941 ug/l

RT: 11.200 min Scan# 1572

Delta R.T. -0.000 min

Lab File: VN086923.D


Acq: 10 Jun 2025 13:12




Tgt Ion: 43 Resp: 455196

Ion Ratio Lower Upper

43 100

58 64.6 30.4 91.3





#60

Dibromochloromethane

Concen: 45.420 ug/l

RT: 11.359 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N




ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#61

1,2-Dibromoethane

Concen: 43.108 ug/l

RT: 11.471 min Scan# 1618

Delta R.T. -0.000 min

Lab File: VN086923.D


Acq: 10 Jun 2025 13:12




Tgt Ion:107 Resp: 92981

Ion Ratio Lower Upper

107 100

109 96.1 76.3 114.5





#62

4-Bromofluorobenzene

Concen: 55.271 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min





Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MSD

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion:117 Resp: 266637


Ion Ratio Lower Upper

117 100

82 56.2 44.6 67.0

119 32.9 25.5 38.3





#64

Tetrachloroethene

Concen: 66.034 ug/l

RT: 11.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D



Acq: 10 Jun 2025 13:12

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-11A-13.5-060525MSD

**Manual Integrations**  
**APPROVED**


Reviewed By :John Carbone 06/11/2025


Supervised By :Mahesh Dadoda 06/11/2025







#65  
Chlorobenzene  
Concen: 44.144 ug/l  
RT: 11.894 min Scan# 1690  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion:112 Resp: 259599  
Ion Ratio Lower Upper  
112 100  
114 32.9 26.2 39.2





Abundance Scan 1701 (11.959 min): VN086923.D\data.ms (-)



#66

1,1,1,2-Tetrachloroethane

Concen: 44.799 ug/l

RT: 11.959 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N


ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#67

Ethyl Benzene

Concen: 42.226 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN086923.D


Acq: 10 Jun 2025 13:12




Tgt Ion: 91 Resp: 427727

Ion Ratio Lower Upper

91 100

106 32.3 25.4 38.2









#68  
m/p-Xylenes  
Concen: 86.123 ug/l  
RT: 12.071 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD


### Manual Integrations APPROVED




Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#69  
o-Xylene  
Concen: 43.848 ug/l  
RT: 12.400 min Scan# 1776  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion:106 Resp: 162838  
Ion Ratio Lower Upper  
106 100  
91 207.3 106.1 318.1





#70

Styrene

Concen: 44.384 ug/l

RT: 12.412 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#71

Bromoform

Concen: 46.983 ug/l


RT: 12.576 min Scan# 1806




Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt	Ion:173	Resp:	65807
Ion	Ratio	Lower	Upper
173	100		
175	49.1	24.1	72.4
254	0.0	0.0	0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#73

Isopropylbenzene

Concen: 41.733 ug/l


RT: 12.694 min Scan# 1826


Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion:105 Resp: 394852  
 Ion Ratio Lower Upper  
 105 100  
 120 27.1 13.5 40.4





#74

N-amyl acetate

Concen: 43.473 ug/l


RT: 12.518 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MSD



Tgt Ion: 43 Resp: 143729

Ion Ratio Lower Upper

43 100

70 50.9 37.6 56.4




55 25.6 19.0 28.6

61 27.0 20.8 31.2

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 44.965 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12



Tgt Ion: 83 Resp: 144127


Ion Ratio Lower Upper

83 100



131 10.1 4.9 14.7

85 64.9 32.0 96.0






Abundance Scan 1877 (12.994 min): VN086923.D\data.ms



Abundance Scan 1874 (12.976 min): VN086923.D\data.ms



Abundance Scan 1874 (12.976 min): VN086923.D\data.ms

#76

1,2,3-Trichloropropane

Concen: 38.345 ug/l m

RT: 12.994 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N



ClientSampleId :

MW-11A-13.5-060525MSD

### Manual Integrations APPROVED


Reviewed By :John Carbone 06/11/2025


Supervised By :Mahesh Dadoda 06/11/2025




#77  
Bromobenzene  
Concen: 45.640 ug/l  
RT: 12.976 min Scan# 1874  
Delta R.T. -0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion:156 Resp: 99031  
Ion Ratio Lower Upper  
156 100  
77 230.9 117.7 353.1  
158 95.9 48.1 144.4






Abundance Scan 1884 (13.035 min): VN086923.D\data.ms (-)



Abundance Scan 1884 (13.035 min): VN086923.D\data.ms (-)




#78  
n-propylbenzene  
Concen: 40.377 ug/l  
RT: 13.035 min Scan# 1  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12


Tgt Ion: 91 Resp: 464251  
Ion Ratio Lower Upper  
91 100  
120 23.6 11.4 34.2

1 Manual Integrations  
2 APPROVED


3 Reviewed By :John Carlone 06/11/2025  
4 Supervised By :Mahesh Dadoda 06/11/2025



Abundance Scan 1899 (13.123 min): VN086923.D\data.ms (-)







Abundance Scan 1899 (13.123 min): VN086923.D\data.ms (-)



#79  
2-Chlorotoluene  
Concen: 42.017 ug/l  
RT: 13.123 min Scan# 1899  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion: 91 Resp: 289730  
Ion Ratio Lower Upper  
91 100  
126 33.6 16.2 48.5





#80

1,3,5-Trimethylbenzene

Concen: 41.111 ug/l

RT: 13.171 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#81

trans-1,4-Dichloro-2-butene

Concen: 46.366 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 75 Resp: 62181




Ion Ratio Lower Upper

75 100

53 83.5 66.6 99.8

89 42.1 36.7 55.1





#82

4-Chlorotoluene

Concen: 42.756 ug/l

RT: 13.218 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12





Instrument:

MSVOA\_N

ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

 Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025


#83

tert-Butylbenzene


Concen: 38.263 ug/l




RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

 Tgt Ion:119 Resp: 273596  
 Ion Ratio Lower Upper  
 119 100  
 91 65.0 32.1 96.5  
 134 27.7 13.4 40.1








#84

1,2,4-Trimethylbenzene  
Concen: 41.655 ug/l  
RT: 13.482 min Scan# 1960  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD

### Manual Integrations APPROVED


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025






#85

sec-Butylbenzene  
Concen: 37.979 ug/l  
RT: 13.612 min Scan# 1982  
Delta R.T. -0.006 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion:105 Resp: 394374  
Ion Ratio Lower Upper  
105 100  
134 20.6 10.0 30.0





#86

p-Isopropyltoluene

Concen: 38.742 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#87

1,3-Dichlorobenzene

Concen: 43.312 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion:146 Resp: 184900




Ion Ratio Lower Upper

146 100

111 41.4 20.9 62.8

148 63.6 32.0 96.2





#88

1,4-Dichlorobenzene

Concen: 43.980 ug/l

RT: 13.812 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument:

MSVOA\_N





ClientSampleId :

MW-11A-13.5-060525MSD

**Manual Integrations  
APPROVED**

Reviewed By :John Carlone 06/11/2025

Supervised By :Mahesh Dadoda 06/11/2025



#89

n-Butylbenzene

Concen: 36.403 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12


Tgt Ion: 91 Resp: 302661




Ion Ratio Lower Upper

91 100

92 54.5 27.5 82.4

134 26.8 13.0 39.0









#90  
Hexachloroethane  
Concen: 39.692 ug/l  
RT: 14.335 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD


1 Manual Integrations  
2 APPROVED




3 Reviewed By :John Carlone 06/11/2025  
4 Supervised By :Mahesh Dadoda 06/11/2025



#91  
1,2-Dichlorobenzene  
Concen: 43.302 ug/l  
RT: 14.106 min Scan# 2066  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion:146 Resp: 177603  
Ion Ratio Lower Upper  
146 100  
111 43.1 21.6 64.6  
148 63.3 31.7 95.1









#92  
1,2-Dibromo-3-Chloropropane  
Concen: 40.715 ug/l  
RT: 14.717 min Scan# 2170  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD


**Manual Integrations**  
**APPROVED**


Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



#93  
1,2,4-Trichlorobenzene  
Concen: 37.609 ug/l  
RT: 15.388 min Scan# 2284  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Tgt Ion:180 Resp: 98501  
Ion Ratio Lower Upper  
180 100  
182 96.0 46.9 140.7  
145 32.7 16.0 47.9





#94

Hexachlorobutadiene

Concen: 29.629 ug/l

RT: 15.494 min Scan# 2

Delta R.T. -0.006 min



Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
 ClientSampleId : MW-11A-13.5-060525MSD

### Manual Integrations APPROVED

Reviewed By :John Carlone 06/11/2025  
 Supervised By :Mahesh Dadoda 06/11/2025



#95

Naphthalene

Concen: 40.957 ug/l


RT: 15.635 min Scan# 2326


Delta R.T. -0.006 min

Lab File: VN086923.D

Acq: 10 Jun 2025 13:12

Tgt Ion:128 Resp: 399267  
 Ion Ratio Lower Upper  
 128 100  
 127 12.8 10.2 15.2  
 129 11.0 8.8 13.2






#96  
1,2,3-Trichlorobenzene  
Concen: 35.852 ug/l  
RT: 15.835 min Scan# 2  
Delta R.T. -0.000 min  
Lab File: VN086923.D  
Acq: 10 Jun 2025 13:12

Instrument : MSVOA\_N  
ClientSampleId : MW-11A-13.5-060525MSD

**Manual Integrations**  
**APPROVED**

Reviewed By :John Carlone 06/11/2025  
Supervised By :Mahesh Dadoda 06/11/2025



## Manual Integration Report

Sequence:	vn060625	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN086862.D	1,1,2-Trichlorotrifluoroethane	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	1,4-Dichlorobenzene	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	2-Hexanone	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC001	VN086862.D	N-amyl acetate	JOHN	6/9/2025 8:02:09 AM	MMDadoda	6/9/2025 1:13:18 PM	Peak Integrated by Software
VSTDICC005	VN086863.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:13 AM	MMDadoda	6/9/2025 1:13:19 PM	Peak Integrated by Software
VSTDICC005	VN086863.D	N-amyl acetate	JOHN	6/9/2025 8:02:13 AM	MMDadoda	6/9/2025 1:13:19 PM	Peak Integrated by Software
VSTDICC020	VN086864.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:19 AM	MMDadoda	6/9/2025 1:13:21 PM	Peak Integrated by Software
VSTDICCC050	VN086865.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:25 AM	MMDadoda	6/9/2025 1:13:23 PM	Peak Integrated by Software
VSTDICC100	VN086866.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:29 AM	MMDadoda	6/9/2025 1:13:28 PM	Peak Integrated by Software
VSTDICC150	VN086867.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:34 AM	MMDadoda	6/9/2025 1:13:30 PM	Peak Integrated by Software
VSTDICV050	VN086869.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:38 AM	MMDadoda	6/9/2025 1:13:34 PM	Peak Integrated by Software
VSTDCCC050	VN086886.D	1,2,3-Trichloropropane	JOHN	6/9/2025 8:02:55 AM	MMDadoda	6/9/2025 1:13:44 PM	Peak Integrated by Software

 1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

## Manual Integration Report

Sequence:	vn060625	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16

### Manual Integration Report

Sequence:	VN061025	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN086913.D	1,2,3-Trichloropropane	JOHN	6/11/2025 9:51:44 AM	MMDadoda	6/11/2025 11:12:24 AM	Peak Integrated by Software
VSTDCCC050	VN086913.D	trans-1,4-Dichloro-2-butene	JOHN	6/11/2025 9:51:44 AM	MMDadoda	6/11/2025 11:12:24 AM	Peak Integrated by Software
VN0610WBS01	VN086916.D	1,2,3-Trichloropropane	JOHN	6/11/2025 9:51:48 AM	MMDadoda	6/11/2025 11:12:38 AM	Peak Integrated by Software
VN0610WBS01	VN086916.D	N-amyl acetate	JOHN	6/11/2025 9:51:48 AM	MMDadoda	6/11/2025 11:12:38 AM	Peak Integrated by Software
Q2250-02MS	VN086922.D	1,2,3-Trichloropropane	JOHN	6/11/2025 9:51:56 AM	MMDadoda	6/11/2025 11:12:42 AM	Peak Integrated by Software
Q2250-02MS	VN086922.D	N-amyl acetate	JOHN	6/11/2025 9:51:56 AM	MMDadoda	6/11/2025 11:12:42 AM	Peak Integrated by Software
Q2250-03MSD	VN086923.D	1,2,3-Trichloropropane	JOHN	6/11/2025 9:52:00 AM	MMDadoda	6/11/2025 11:12:43 AM	Peak Integrated by Software
VSTDCCC050	VN086938.D	1,2,3-Trichloropropane	JOHN	6/11/2025 9:52:32 AM	MMDadoda	6/11/2025 11:12:50 AM	Peak Integrated by Software
VSTDCCC050	VN086938.D	trans-1,4-Dichloro-2-butene	JOHN	6/11/2025 9:52:32 AM	MMDadoda	6/11/2025 11:12:50 AM	Peak Integrated by Software

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16

### Manual Integration Report

Sequence:	vn061125	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN086940.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:47:09 AM	Sam	6/12/2025 3:54:50 PM	Peak Integrated by Software
VN0611WBS02	VN086944.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:47:21 AM	Sam	6/12/2025 3:54:50 PM	Peak Integrated by Software
VSTDCCC050	VN086965.D	1,2,3-Trichloropropane	JOHN	6/12/2025 9:53:57 AM	Sam	6/12/2025 3:54:56 PM	Peak Integrated by Software

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN060625**

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Mahesh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134155 VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134156 VP134248		

Sr #	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086861.D	06 Jun 2025 07:59	JC\MD	Ok
2	VSTDICCC001	VN086862.D	06 Jun 2025 12:44	JC\MD	Ok,M
3	VSTDICCC005	VN086863.D	06 Jun 2025 13:17	JC\MD	Ok,M
4	VSTDICCC020	VN086864.D	06 Jun 2025 13:40	JC\MD	Ok,M
5	VSTDICCC050	VN086865.D	06 Jun 2025 14:03	JC\MD	Ok,M
6	VSTDICCC100	VN086866.D	06 Jun 2025 14:26	JC\MD	Ok,M
7	VSTDICCC150	VN086867.D	06 Jun 2025 14:49	JC\MD	Ok,M
8	IBLK	VN086868.D	06 Jun 2025 15:12	JC\MD	Ok
9	VSTDICV050	VN086869.D	06 Jun 2025 15:54	JC\MD	Ok,M
10	VN0606WBL01	VN086870.D	06 Jun 2025 16:47	JC\MD	Ok
11	VN0606WBL02	VN086871.D	06 Jun 2025 17:10	JC\MD	Ok
12	VN0606WBS01	VN086872.D	06 Jun 2025 17:33	JC\MD	Ok,M
13	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56	JC\MD	Ok,M
14	Q2254-01	VN086874.D	06 Jun 2025 18:19	JC\MD	Not Ok
15	Q2237-02	VN086875.D	06 Jun 2025 18:42	JC\MD	Ok
16	Q2216-02	VN086876.D	06 Jun 2025 19:05	JC\MD	Ok
17	Q2216-03	VN086877.D	06 Jun 2025 19:28	JC\MD	Ok
18	Q2216-04	VN086878.D	06 Jun 2025 19:51	JC\MD	Ok
19	Q2216-05	VN086879.D	06 Jun 2025 20:13	JC\MD	Not Ok
20	Q2216-06	VN086880.D	06 Jun 2025 20:36	JC\MD	Not Ok
21	Q2206-04	VN086881.D	06 Jun 2025 20:59	JC\MD	Not Ok

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN060625**

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Mahesh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134155 VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134156 VP134248		

22	Q2242-04	VN086882.D	06 Jun 2025 21:21	JC\MD	Not Ok
23	Q2192-01	VN086883.D	06 Jun 2025 21:44	JC\MD	Not Ok
24	Q2198-02	VN086884.D	06 Jun 2025 22:07	JC\MD	Not Ok
25	Q2198-04	VN086885.D	06 Jun 2025 22:29	JC\MD	Not Ok
26	VSTDCCC050	VN086886.D	06 Jun 2025 22:52	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN061025**

Review By	John Carlone	Review On	6/11/2025 9:59:27 AM
Supervise By	Mahesh Dadoda	Supervise On	6/11/2025 11:14:28 AM
SubDirectory	VN061025	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134203		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134204,VP134205		

Sr #	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086912.D	10 Jun 2025 08:39	JC\MD	Ok
2	VSTDCCC050	VN086913.D	10 Jun 2025 09:13	JC\MD	Ok,M
3	VN0610MBL01	VN086914.D	10 Jun 2025 09:47	JC\MD	Ok
4	VN0610WBL01	VN086915.D	10 Jun 2025 10:09	JC\MD	Ok
5	VN0610WBS01	VN086916.D	10 Jun 2025 10:30	JC\MD	Ok,M
6	VN0610WBSD01	VN086917.D	10 Jun 2025 11:04	JC\MD	Ok,M
7	IBLK	VN086918.D	10 Jun 2025 11:25	JC\MD	Ok
8	Q2250-05	VN086919.D	10 Jun 2025 11:47	JC\MD	Ok
9	Q2250-06	VN086920.D	10 Jun 2025 12:08	JC\MD	Ok
10	Q2250-01	VN086921.D	10 Jun 2025 12:30	JC\MD	Dilution
11	Q2250-02MS	VN086922.D	10 Jun 2025 12:51	JC\MD	Ok,M
12	Q2250-03MSD	VN086923.D	10 Jun 2025 13:12	JC\MD	Ok,M
13	IBLK	VN086924.D	10 Jun 2025 13:34	JC\MD	Ok
14	Q2251-01	VN086925.D	10 Jun 2025 13:55	JC\MD	Not Ok
15	Q2251-06	VN086926.D	10 Jun 2025 14:17	JC\MD	Not Ok
16	Q2251-05	VN086927.D	10 Jun 2025 14:38	JC\MD	ReRun
17	Q2251-02	VN086928.D	10 Jun 2025 15:00	JC\MD	Ok
18	Q2251-03	VN086929.D	10 Jun 2025 15:22	JC\MD	Ok
19	IBLK	VN086930.D	10 Jun 2025 15:43	JC\MD	Ok
20	IBLK	VN086931.D	10 Jun 2025 16:05	JC\MD	Ok
21	Q2251-01	VN086932.D	10 Jun 2025 16:26	JC\MD	Ok

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN061025**

Review By	John Carfone	Review On	6/11/2025 9:59:27 AM
Supervise By	Mahesh Dadoda	Supervise On	6/11/2025 11:14:28 AM
SubDirectory	VN061025	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134203		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134204,VP134205		

22	Q2251-05	VN086933.D	10 Jun 2025 16:48	JC\MD	Ok
23	Q2202-01	VN086934.D	10 Jun 2025 17:09	JC\MD	Ok
24	Q2202-02	VN086935.D	10 Jun 2025 17:31	JC\MD	Ok
25	Q2202-03	VN086936.D	10 Jun 2025 17:52	JC\MD	Dilution
26	Q2279-01	VN086937.D	10 Jun 2025 18:14	JC\MD	Ok
27	VSTDCCC050	VN086938.D	10 Jun 2025 18:57	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN061125**

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134284		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286		

Sr #	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086939.D	11 Jun 2025 10:22	JC\MD	Ok
2	VSTDCCC050	VN086940.D	11 Jun 2025 11:32	JC\MD	Ok,M
3	VN0611MBL01	VN086941.D	11 Jun 2025 12:06	JC\MD	Ok
4	VN0611WBL01	VN086942.D	11 Jun 2025 12:28	JC\MD	Ok
5	VN0611WBS01	VN086943.D	11 Jun 2025 12:50	JC\MD	Not Ok
6	VN0611WBS02	VN086944.D	11 Jun 2025 13:27	JC\MD	Ok,M
7	VN0611WBSD02	VN086945.D	11 Jun 2025 14:01	JC\MD	Ok,M
8	Q2251-06	VN086946.D	11 Jun 2025 14:22	JC\MD	Ok
9	Q2202-03DL	VN086947.D	11 Jun 2025 14:44	JC\MD	Ok,M
10	Q2250-01DL	VN086948.D	11 Jun 2025 15:06	JC\MD	Ok
11	Q2202-06	VN086949.D	11 Jun 2025 15:28	JC\MD	Ok
12	Q2189-01	VN086950.D	11 Jun 2025 15:50	JC\MD	Ok
13	Q2202-05	VN086951.D	11 Jun 2025 16:12	JC\MD	Ok
14	Q2202-04	VN086952.D	11 Jun 2025 16:35	JC\MD	Ok
15	Q2231-01	VN086953.D	11 Jun 2025 16:57	JC\MD	Ok
16	Q2231-02	VN086954.D	11 Jun 2025 17:19	JC\MD	Ok,M
17	Q2231-03	VN086955.D	11 Jun 2025 17:41	JC\MD	Ok
18	Q2231-04	VN086956.D	11 Jun 2025 18:04	JC\MD	Ok
19	Q2231-05	VN086957.D	11 Jun 2025 18:26	JC\MD	Ok
20	Q2231-06	VN086958.D	11 Jun 2025 18:48	JC\MD	Ok
21	Q2210-01	VN086959.D	11 Jun 2025 19:10	JC\MD	Ok

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN061125**

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134284		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286		

22	Q2209-01	VN086960.D	11 Jun 2025 19:33	JC\MD	Ok
23	IBLK	VN086961.D	11 Jun 2025 19:55	JC\MD	Ok
24	IBLK	VN086962.D	11 Jun 2025 20:17	JC\MD	Ok
25	Q2263-01	VN086963.D	11 Jun 2025 20:39	JC\MD	Ok
26	Q2263-02	VN086964.D	11 Jun 2025 21:01	JC\MD	Ok
27	VSTDCCC050	VN086965.D	11 Jun 2025 21:23	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QCBatch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Mahesh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134155		
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC	VP134156		
Internal Standard/PEM	VP134248		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr #	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086861.D	06 Jun 2025 07:59		JC\MD	Ok
2	VSTDICCC001	VSTDICCC001	VN086862.D	06 Jun 2025 12:44	Method failed for com.#13	JC\MD	Ok,M
3	VSTDICCC005	VSTDICCC005	VN086863.D	06 Jun 2025 13:17		JC\MD	Ok,M
4	VSTDICCC020	VSTDICCC020	VN086864.D	06 Jun 2025 13:40		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN086865.D	06 Jun 2025 14:03		JC\MD	Ok,M
6	VSTDICCC100	VSTDICCC100	VN086866.D	06 Jun 2025 14:26		JC\MD	Ok,M
7	VSTDICCC150	VSTDICCC150	VN086867.D	06 Jun 2025 14:49		JC\MD	Ok,M
8	IBLK	IBLK	VN086868.D	06 Jun 2025 15:12		JC\MD	Ok
9	VSTDICV050	ICVVN060625	VN086869.D	06 Jun 2025 15:54		JC\MD	Ok,M
10	VN0606WBL01	VN0606WBL01	VN086870.D	06 Jun 2025 16:47		JC\MD	Ok
11	VN0606WBL02	VN0606WBL02	VN086871.D	06 Jun 2025 17:10		JC\MD	Ok
12	VN0606WBS01	VN0606WBS01	VN086872.D	06 Jun 2025 17:33		JC\MD	Ok,M
13	VN0606WBSD01	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56		JC\MD	Ok,M
14	Q2254-01	BP-VPB-182-GW-810-8	VN086874.D	06 Jun 2025 18:19	vial A pH<2 endccc out of tune	JC\MD	Not Ok
15	Q2237-02	TW-WTS-10	VN086875.D	06 Jun 2025 18:42	vial A pH<2	JC\MD	Ok
16	Q2216-02	3887	VN086876.D	06 Jun 2025 19:05	vial A pH<2	JC\MD	Ok
17	Q2216-03	3888	VN086877.D	06 Jun 2025 19:28	vial A pH<2	JC\MD	Ok
18	Q2216-04	3864	VN086878.D	06 Jun 2025 19:51	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QCBatch ID # VN060625

Review By	John Caralone	Review On	6/9/2025 8:08:23 AM
Supervise By	Mahesh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134155		
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC	VP134156		
Internal Standard/PEM	VP134248		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2216-05	3865	VN086879.D	06 Jun 2025 20:13	vial A pH<2 Out of Tune	JC\MD	Not Ok
20	Q2216-06	3851	VN086880.D	06 Jun 2025 20:36	vial A pH<2 Out of Tune	JC\MD	Not Ok
21	Q2206-04	TP-1	VN086881.D	06 Jun 2025 20:59	vial A pH<2 Out of Tune	JC\MD	Not Ok
22	Q2242-04	TP09-MHJ	VN086882.D	06 Jun 2025 21:21	vial A pH<2 Out of Tune	JC\MD	Not Ok
23	Q2192-01	SB-1	VN086883.D	06 Jun 2025 21:44	vial A pH<2 Out of Tune	JC\MD	Not Ok
24	Q2198-02	B-202-SB02	VN086884.D	06 Jun 2025 22:07	vial A pH<2 Out of Tune	JC\MD	Not Ok
25	Q2198-04	B-207-SB02	VN086885.D	06 Jun 2025 22:29	vial A pH<2 Out of Tune	JC\MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VN086886.D	06 Jun 2025 22:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QCBatch ID # VN061025

Review By	John Carlone	Review On	6/11/2025 9:59:27 AM
Supervise By	Mahesh Dadoda	Supervise On	6/11/2025 11:14:28 AM
SubDirectory	VN061025	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134203		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134204,VP134205		

Sr #	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086912.D	10 Jun 2025 08:39		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086913.D	10 Jun 2025 09:13	pH#Lot#V12668	JC\MD	Ok,M
3	VN0610MBL01	VN0610MBL01	VN086914.D	10 Jun 2025 09:47		JC\MD	Ok
4	VN0610WBL01	VN0610WBL01	VN086915.D	10 Jun 2025 10:09		JC\MD	Ok
5	VN0610WBS01	VN0610WBS01	VN086916.D	10 Jun 2025 10:30		JC\MD	Ok,M
6	VN0610WBSD01	VN0610WBSD01	VN086917.D	10 Jun 2025 11:04		JC\MD	Ok,M
7	IBLK	IBLK	VN086918.D	10 Jun 2025 11:25		JC\MD	Ok
8	Q2250-05	EB02-060525	VN086919.D	10 Jun 2025 11:47	vial A pH<2 EB	JC\MD	Ok
9	Q2250-06	TB-01-060525	VN086920.D	10 Jun 2025 12:08	vial A pH<2 TB	JC\MD	Ok
10	Q2250-01	MW-11A-13.5-060525	VN086921.D	10 Jun 2025 12:30	vial A pH<2 Need 20X	JC\MD	Dilution
11	Q2250-02MS	MW-11A-13.5-060525M	VN086922.D	10 Jun 2025 12:51	vial A pH<2	JC\MD	Ok,M
12	Q2250-03MSD	MW-11A-13.5-060525M	VN086923.D	10 Jun 2025 13:12	vial A pH<2	JC\MD	Ok,M
13	IBLK	IBLK	VN086924.D	10 Jun 2025 13:34		JC\MD	Ok
14	Q2251-01	BP-VPB-182-TB-20250	VN086925.D	10 Jun 2025 13:55	Confirm concentration	JC\MD	Not Ok
15	Q2251-06	VPB182-HYD-2025060	VN086926.D	10 Jun 2025 14:17	Confirm concentration	JC\MD	Not Ok
16	Q2251-05	BP-VPB-182-EB-20250	VN086927.D	10 Jun 2025 14:38	vial A pH<2 EB;Hit of comp.#16,25	JC\MD	ReRun
17	Q2251-02	BP-VPB-182-GW-740-7	VN086928.D	10 Jun 2025 15:00	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QCBatch ID # VN061025**

Review By	John Carlone	Review On	6/11/2025 9:59:27 AM
Supervise By	Mahesh Dadoda	Supervise On	6/11/2025 11:14:28 AM
SubDirectory	VN061025	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds  CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134203  VP134204,VP134205		

18	Q2251-03	BP-VPB-182-GW-760-7	VN086929.D	10 Jun 2025 15:22	vial A pH<2	JC\MD	Ok
19	IBLK	IBLK	VN086930.D	10 Jun 2025 15:43		JC\MD	Ok
20	IBLK	IBLK	VN086931.D	10 Jun 2025 16:05		JC\MD	Ok
21	Q2251-01	BP-VPB-182-TB-20250	VN086932.D	10 Jun 2025 16:26	vial A pH<2 TB	JC\MD	Ok
22	Q2251-05	BP-VPB-182-EB-20250	VN086933.D	10 Jun 2025 16:48	vial B pH<2 EB	JC\MD	Ok
23	Q2202-01	MW-9-20250603	VN086934.D	10 Jun 2025 17:09	vial A pH<2	JC\MD	Ok
24	Q2202-02	MW-11-20250603	VN086935.D	10 Jun 2025 17:31	vial A pH<2	JC\MD	Ok
25	Q2202-03	MW-12-20250603	VN086936.D	10 Jun 2025 17:52	vial A pH<2 Need 5X	JC\MD	Dilution
26	Q2279-01	301-469-5TH-AVE	VN086937.D	10 Jun 2025 18:14	vial A pH<2	JC\MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VN086938.D	10 Jun 2025 18:57		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QCBatch ID # VN061125

Review By	John Carlone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134284		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134285,VP134286		

Sr #	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086939.D	11 Jun 2025 10:22		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086940.D	11 Jun 2025 11:32	pH#Lot#V12668	JC\MD	Ok,M
3	VN0611MBL01	VN0611MBL01	VN086941.D	11 Jun 2025 12:06		JC\MD	Ok
4	VN0611WBL01	VN0611WBL01	VN086942.D	11 Jun 2025 12:28		JC\MD	Ok
5	VN0611WBS01	VN0611WBS01	VN086943.D	11 Jun 2025 12:50	Recovery fail	JC\MD	Not Ok
6	VN0611WBS02	VN0611WBS02	VN086944.D	11 Jun 2025 13:27		JC\MD	Ok,M
7	VN0611WBSD02	VN0611WBSD02	VN086945.D	11 Jun 2025 14:01		JC\MD	Ok,M
8	Q2251-06	VPB182-HYD-2025060	VN086946.D	11 Jun 2025 14:22	vial A pH<2	JC\MD	Ok
9	Q2202-03DL	MW-12-20250603DL	VN086947.D	11 Jun 2025 14:44	vial B pH<2	JC\MD	Ok,M
10	Q2250-01DL	MW-11A-13.5-060525D	VN086948.D	11 Jun 2025 15:06	vial B pH<2	JC\MD	Ok
11	Q2202-06	TB-20250603	VN086949.D	11 Jun 2025 15:28	vial A pH<2 TB	JC\MD	Ok
12	Q2189-01	MW-7-20250602	VN086950.D	11 Jun 2025 15:50	vial B pH<2	JC\MD	Ok
13	Q2202-05	MW-1A-20250603	VN086951.D	11 Jun 2025 16:12	vial A pH<2	JC\MD	Ok
14	Q2202-04	MW-130-20250603	VN086952.D	11 Jun 2025 16:35	vial A pH<2	JC\MD	Ok
15	Q2231-01	MW-10D-20250604	VN086953.D	11 Jun 2025 16:57	vial A pH<2	JC\MD	Ok
16	Q2231-02	MW-14-20250604	VN086954.D	11 Jun 2025 17:19	vial A pH<2	JC\MD	Ok,M
17	Q2231-03	MW-15-20250604	VN086955.D	11 Jun 2025 17:41	vial A pH<2	JC\MD	Ok
18	Q2231-04	MW-16D-20250604	VN086956.D	11 Jun 2025 18:04	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA\_N

### Daily Analysis Runlog For Sequence/QCBatch ID # VN061125

Review By	John Caralone	Review On	6/12/2025 9:57:37 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/12/2025 3:55:33 PM
SubDirectory	VN061125	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds  CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134284  VP134285,VP134286		

19	Q2231-05	MW-17-20250604	VN086957.D	11 Jun 2025 18:26	vial A pH<2	JC\MD	Ok
20	Q2231-06	TB-20250604	VN086958.D	11 Jun 2025 18:48	vial A pH<2 TB	JC\MD	Ok
21	Q2210-01	TW1	VN086959.D	11 Jun 2025 19:10	vial A pH<2	JC\MD	Ok
22	Q2209-01	P01W	VN086960.D	11 Jun 2025 19:33	vial A pH<2	JC\MD	Ok
23	IBLK	IBLK	VN086961.D	11 Jun 2025 19:55		JC\MD	Ok
24	IBLK	IBLK	VN086962.D	11 Jun 2025 20:17		JC\MD	Ok
25	Q2263-01	RW9-MW01D3-202506	VN086963.D	11 Jun 2025 20:39	vial A pH<2	JC\MD	Ok
26	Q2263-02	RW9-MW01S-2025060	VN086964.D	11 Jun 2025 21:01	vial A pH<2	JC\MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VN086965.D	11 Jun 2025 21:23		JC\MD	Ok,M

M : Manual Integration

## Prep Standard - Chemical Standard Summary

**Order ID :** Q2250

**Test :** VOCMS Group3

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** VN061025,vn061125,

**Standard ID :**

VP133174,VP133887,VP133935,VP133953,VP133991,VP134142,VP134149,VP134203,VP134204,VP134205,VP134284,VP134285,VP134286,

**Chemical ID :**

V13391,V13706,V14290,V14432,V14435,V14503,V14504,V14525,V14526,V14613,V14620,V14626,V14636,V14637,V14638,V14639,V14668,V14671,V14673,V14675,V14711,V14717,V14718,V14721,V14749,V14750,V14811,V14812,V14843,V14921,V14929,V14944,V14945,V14946,V14947,W3112,

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	<a href="#">VP133174</a>	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	<a href="#">VP133887</a>	05/12/2025	06/23/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 05/14/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14432 + 1.00000ml of V14435 + 1.00000ml of V14503 + 1.00000ml of V14504 + 1.00000ml of V14525 + 1.00000ml of V14526 + 1.00000ml of V14711 + 1.00000ml of V14717 + 1.00000ml of V14718 + 1.00000ml of V14721 + 1.00000ml of V14749 + 1.00000ml of V14750 + 1.00000ml of V14811 + 1.00000ml of V14812 + 10.60000ml of V14921 = Final Quantity: 25.000 ml

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	<a href="#">VP133935</a>	05/16/2025	11/12/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 05/21/2025

FROM 0.25000ml of V14290 + 24.75000ml of V14921 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	<a href="#">VP133953</a>	05/19/2025	11/09/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 05/21/2025

FROM 0.25000ml of V13391 + 24.75000ml of V14626 = Final Quantity: 25.000 ml

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	<a href="#">VP133991</a>	05/22/2025	06/19/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 05/24/2025

**FROM** 1.00000ml of V14944 + 1.00000ml of V14945 + 1.00000ml of V14946 + 1.00000ml of V14947 + 21.00000ml of V14620 = Final  
Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	<a href="#">VP134142</a>	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025

**FROM** 1.00000ml of V14668 + 1.00000ml of V14671 + 1.00000ml of V14673 + 1.00000ml of V14675 + 16.00000ml of V14929 = Final  
Quantity: 20.000 ml

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	<a href="#">VP134149</a>	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025

FROM 1.00000ml of V14636 + 1.00000ml of V14637 + 1.00000ml of V14638 + 1.00000ml of V14639 + 46.00000ml of V14929 = Final  
Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	<a href="#">VP134203</a>	06/10/2025	06/11/2025	John Carlone	None	None	Mahesh Dadoda 06/10/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP133953 = Final Quantity: 40.000 ml

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	<a href="#">VP134204</a>	06/10/2025	06/11/2025	John Carlone	None	None	Mahesh Dadoda 06/10/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	<a href="#">VP134205</a>	06/10/2025	06/11/2025	John Carlone	None	None	Mahesh Dadoda 06/10/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	<a href="#">VP134284</a>	06/11/2025	06/12/2025	John Carlone	None	None	Mahesh Dadoda 06/12/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP133953 = Final Quantity: 40.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	<a href="#">VP134285</a>	06/11/2025	06/12/2025	John Carlone	None	None	Mahesh Dadoda 06/12/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml

## VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	<a href="#">VP134286</a>	06/11/2025	06/12/2025	John Carlone	None	None	Mahesh Dadoda 06/12/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml							

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16

### CHEMICAL RECEIPT LOG BOOK

<b>Supplier</b>	<b>ItemCode / ItemName</b>	<b>Lot #</b>	<b>Expiration Date</b>	<b>Date Opened / Opened By</b>	<b>Received Date / Received By</b>	<b>Chemtech Lot #</b>
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	05/12/2025 / SAM	08/15/2024 / SAM	V14432
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14503

### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14504
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14525
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14526
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	10/25/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14620
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14636
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14637
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14638
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14639
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14668
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14671

### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14673
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14675
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14711
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14717
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14718
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14721

### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/13/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14749
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14750
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0220471	11/12/2025	05/12/2025 / SAM	01/08/2025 / SAM	V14811
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0220471	06/30/2026	05/12/2025 / SAM	01/08/2025 / SAM	V14812
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	11/12/2025	05/12/2025 / SAM	05/09/2025 / SAM	V14921

### CHEMICAL RECEIPT LOG BOOK

<b>Supplier</b>	<b>ItemCode / ItemName</b>	<b>Lot #</b>	<b>Expiration Date</b>	<b>Date Opened / Opened By</b>	<b>Received Date / Received By</b>	<b>Chemtech Lot #</b>
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	12/06/2025	06/06/2025 / SAM	05/09/2025 / SAM	V14929
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14944
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14945
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14946
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14947
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis



Material No.: 9077-02  
Batch No.: 23I0762004  
Manufactured Date: 2023-08-11  
Expiration Date: 2026-08-10  
Revision No.: 0

## Certificate of Analysis


Test	Specification	Result
Assay (CH <sub>3</sub> OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titrable Acid (μeq/g)	≤ 0.3	0.2
Titrable Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein  
Sr. Manager, Quality Assurance

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis



Material No.: 9077-02  
Batch No.: 22L0562016  
Manufactured Date: 2022-10-26  
Expiration Date: 2025-10-25  
Revision No.: 0

## Certificate of Analysis


Test	Specification	Result
Assay (CH <sub>3</sub> OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (μeq/g)	≤ 0.3	0.2
Titrable Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier  
Vice President Global Quality

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis



Material No.: 9077-02  
Batch No.: 22L0562016  
Manufactured Date: 2022-10-26  
Expiration Date: 2025-10-25  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay (CH <sub>3</sub> OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (μeq/g)	≤ 0.3	0.2
Titrable Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier  
Vice President Global Quality

Ree 09/17/24

## CERTIFIED WEIGHT REPORT

Part Number: 95317  
 Lot Number: 021624  
 Description: Universal VOA Mixture  
 69 components

Solvent(s): Lot#  
 Methanol EG359-USQ12

Expiration Date: 02/2027  
 Recommended Storage: Freeze (0 °C)

Nominal Concentration (ug/mL): 2000NIST Test ID#: 8UTBWeight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Compound	(RM#)	Lot Number	Dil. Factor	Initial Vol. (mL)	Initial Conc. (ug/mL)	Nominal Conc. (ug/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc. (ug/mL)	Actual Conc. (ug/mL) (+/-) (ug/mL)	Expanded Uncertainty	SDS Information
	Part Number											CAS#	OSHA PEL (TWA)	LD50	
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m <sup>3</sup> /BH) orl-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102398	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m <sup>3</sup> /BH) orl-rat 700mg/kg
3. Carbon disulfide	(0600)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.8	8.1	75-15-0	4 ppm (12mg/m <sup>3</sup> ) (skin) orl-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1198)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1478-11-6	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A
6. Diethyl ether	(0153)	K18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20108	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m <sup>3</sup> /BH) (skin) orl-rat 14800mg/kg
9. 2-Methyl-1-propanol	(0445)	18241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20108	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m <sup>3</sup> /BH) orl-rat 240mg/kg
10. Methacrylonitrile	(0442)	0042ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m <sup>3</sup> /BH) (skin) orl-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m <sup>3</sup> /BH) (skin) orl-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKBW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-82-6	100 ppm (410mg/m <sup>3</sup> /BH) orl-rat 787mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-85-3	1 ppm (3mg/m <sup>3</sup> /BH) (skin) orl-rat 780mg/kg
14. 2-Nitropropane	(0481)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-48-9	10 ppm (35mg/m <sup>3</sup> /BH) orl-rat 720mg/kg
15. Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18590	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m <sup>3</sup> /BH) orl-rat 45mg/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1988.6	22.8	75-27-4	N/A
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm orl-rat 820mg/kg
22. 1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m <sup>3</sup> /BH) orl-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m <sup>3</sup> ) (skin) orl-rat 933mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	2 ppm (12.8mg/m <sup>3</sup> /BH) (CL) orl-rat 908mg/kg
25. Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	0.5 ppm (5mg/m <sup>3</sup> /BH) (skin) orl-rat 933mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-88-3	60 ppm (240mg/m <sup>3</sup> ) (CL) orl-rat 908mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	N/A
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	2002.0	22.9	78-75-7	75 ppm (350mg/m <sup>3</sup> /BH) orl-rat 1847mg/kg
29. Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	1999.8	20.4	59-20-7	N/A
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.8	127-18-4	25 ppm (170mg/m <sup>3</sup> /BH)(final) orl-rat 2620mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m <sup>3</sup> /BH) orl-rat 10300mg/kg
32. 1,2-Dibromopentane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.3	22.9	98-12-8	0.001 ppm orl-rat 170mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-03-4	20 ppm (BH) orl-rat 108mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (BH) orl-rat 670mg/kg
35. 1,2-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m <sup>3</sup> /BH) orl-mus 1847mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-29-9	N/A orl-mus 3600mg/kg
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.1	28.7	593-58-6	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10081-01-5	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.4	23.0	106-01-2	N/A
40. 1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-08-3	0.02 ppm (0.24mg/m <sup>3</sup> /BH) orl-rat 82mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A orl-rat 670mg/kg
42. 1,1,2-Tetrafluoroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (55mg/m <sup>3</sup> /BH) (skin) orl-rat 600mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m <sup>3</sup> /BH) (skin) orl-rat 833mg/kg
44. 1,2,2-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.9	22.9	78-01-6	50 ppm (270mg/m <sup>3</sup> /BH) orl-mus 240mg/kg
45. Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	98-18-4	10 ppm (60mg/m <sup>3</sup> /BH) orl-rat 149.8mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm orl-rat 480mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	N/A orl-rat 269mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A
49. p-Isopropyl tolue	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m <sup>3</sup> /BH) orl-rat >2000mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-57-6	N/A orl-rat 475mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m <sup>3</sup> /BH) orl-rat 490mg/kg
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm orl-rat 5000mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm orl-rat 5000mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m <sup>3</sup> ) orl-rat 756mg/kg
55. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A orl-rat 5g/kg
56. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-67-8	N/A orl-rat 5000mg/kg
57. <i>t</i> -Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m <sup>3</sup> /BH) orl-rat 5g/kg
58. <i>sec</i> -Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	N/A orl-rat 5g/kg
59. <i>n</i> -Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A orl-rat 2240mg/kg
60. Chlorobenzene	35163	101923	0												



Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.

Created: Sat, Feb 17, 2024 at 8:56:46 AM.

Sampled: Sequence "021624-GC5M1", Method "GC5-M1".

Analyzed using Method "GC5-M1".

#### Comments

GC5-M1 Analysis by Candice Warren

Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness

Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,


Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.

Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.).

Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.

FID Signal = Edaq Channel 1

Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	0.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.74
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,3-Dichloropropene	18.60
12	Methyl vinyl ketone/Methyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	19.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.68
18	1,2-Dichloropropane	24.26
19	Methyl methacrylate	24.52
20	Bromoacetylbenzene	25.13
21	Dibromoethane/1,2-Dibromoethane	25.48
22	cis-1,3-Dichloropropene	27.07
23	Toluene	28.03
24	Ethyl methacrylate/trans-1,3-Dichloropropene	29.73
25	1,1,2-Trichloroethane	29.94
26	Toluene/butanone/1,2-Dichloropropane	30.84
27	Dibromochloromethane	31.16
28	1,2-Dibromoethane	31.89
29	Chlorobutane	33.20
30	Ethyleneglycerol,1,1,2-Tetraacetylethane	33.60
31	m-Xylene/p-Xylene	33.89
32	o-Xylene	33.93
33	Glyrene	35.94
34	Isopropylbenzene/Bromoform	36.46
35	cis-1,4-Dichloro-1-butene	36.60
36	1,1,2,2-Tetrachloroethane	37.23
37	j,2,3-Trichloropropene	37.77
38	n-Propylbenzene	37.92
39	trans-1,3-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	2-Chlorotoluene	38.63
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	40.91
46	Pentachloroethane	40.17
47	tert-Butylbenzene	40.53
48	p Isopropyltoluene	41.03
49	1,3-Dichloropropane	41.13
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.62
52	1,2-Dichlorobenzene	42.16
53	1,2-Dibromo-3-chloropropane	46.17
54	Norbornane	46.49
55	1,3,4-Trimethylbenzene	49.26
56	Hexachlorobutadiene	49.32
57	Naphthalene	50.16
58	1,2,3-Trichlorobenzene	51.16



Boiling Point	65°C	Specific Gravity (H <sub>2</sub> O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

**Section X. STABILITY AND REACTIVITY**

Chemical stability: Stable under recommended storage conditions.  
 Possibility of hazardous reactions: Vapours may form explosive mixture with air.  
 Conditions to avoid: Heat, flames, sparks, extreme temperature and sunlight.  
 Materials to avoid: Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids  
 Hazardous decomposition products formed under fire conditions. - Carbon oxides

**Section XI. TOXICOLOGICAL INFORMATION**

LD50 Oral - rat - 5,628 mg/kg  
 LC50 Inhalation - rat - 4 h - 64000 ppm  
 LD50 Dermal - rabbit - 15,800 mg/kg  
 Toxic if absorbed through skin. Causes skin irritation.  
 Eye damage/eye irritation  
 Toxic if inhaled. Causes respiratory tract irritation.  
 Toxic if swallowed.

**Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.**

LC50 15,400 mg/l - 96 h  
 EC50 24,500.00 mg/l - 48 h  
 EC100 10,000.00 mg/l - 24 h

**Section XIII. DISPOSAL CONSIDERATIONS**

Dispose with normal Laboratory Solvent Waste.

**Section XIV. TRANSPORT INFORMATION**

DOT (US) IATA  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol

**Section XV. REGULATORY INFORMATION**

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant  
 SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

**Section XVI. Misc. INFORMATION**

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Ree 09/17/24

## CERTIFIED WEIGHT REPORT

Part Number: 95317  
 Lot Number: 021624  
 Description: Universal VOA Megamix  
 69 components

Solvent(s): Lot#  
 Methanol EG359-USQ12

Expiration Date: 02/2027  
 Recommended Storage: Freeze (0 °C)

Nominal Concentration (ug/mL): 2000NIST Test ID#: 8UTBWeight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

<u>P. Shant Chauhan</u>		021624
Formulated By:	Prashanti Chauhan	DATE
<u>Pedro L. Rentas</u>		021624
Reviewed By:	Pedro L. Rentas	DATE

Compound	(RM#)	Lot Number	Dil. Factor	Initial Vol. (mL)	Initial Conc. (ug/mL)	Nominal Conc. (ug/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc. (ug/mL)	Expanded Uncertainty (+/-) (ug/mL)	SDS Information		
														CAS#	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m <sup>3</sup> /BH)	orl-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102398	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m <sup>3</sup> /BH)	orl-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.8	8.1	75-15-0	4 ppm (12mg/m <sup>3</sup> ) (skin)	orl-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1198)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBPE041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	K18CA500K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	orl-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20108	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m <sup>3</sup> /BH) (skin)	orl-rat 76mg/kg
9. 2-Methyl-1-propanol	(0445)	18241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20108	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m <sup>3</sup> /BH)	orl-rat 240mg/kg
10. Methylacrylonitrile	(0442)	0042ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m <sup>3</sup> /BH) (skin)	orl-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m <sup>3</sup> /BH) (skin)	orl-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-82-6	100 ppm (410mg/m <sup>3</sup> /BH)	orl-rat 787mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-85-3	1 ppm (5mg/m <sup>3</sup> /BH) (skin)	orl-rat 780mg/kg
14. 2-Nitropropane	(0481)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-48-9	10 ppm (35mg/m <sup>3</sup> /BH)	orl-rat 720mg/kg
15. Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18590	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m <sup>3</sup> /BH)	orl-rat 45mg/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1988.6	22.8	75-27-4	N/A	orl-rat 161mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	orl-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	orl-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	75-09-2	500 ppm	orl-rat 820mg/kg
22. 1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m <sup>3</sup> /BH)	orl-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m <sup>3</sup> ) (skin)	orl-rat 933mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	2 ppm (12.8mg/m <sup>3</sup> /BH)	orl-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	10 ppm (410mg/m <sup>3</sup> /BH)	orl-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-88-3	60 ppm (240mg/m <sup>3</sup> ) (CL)	orl-rat 908mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	N/A	orl-rat 108mg/kg
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	orl-rat 725mg/kg
29. Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	1999.8	20.4	59-20-7	N/A	N/A
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.8	127-18-4	25 ppm (170mg/m <sup>3</sup> /BH)(final)	orl-rat 2620mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m <sup>3</sup> /BH)	orl-rat 10300mg/kg
32. 1,2-Dibromopentane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.3	22.9	98-12-8	0.001 ppm	orl-rat 170mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-03-4	20 ppm (BH)	orl-rat 108mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (BH)	orl-rat 670mg/kg
35. 1,2-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m <sup>3</sup> /BH)	orl-rat 1847mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-29-9	N/A	im-mus 3600mg/kg
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.1	28.7	583-58-6	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10081-01-5	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
40. 1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-08-3	0.02 ppm (0.24mg/m <sup>3</sup> /BH)	orl-rat 82mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	orl-rat 670mg/kg
42. 1,1,2-Tetrafluoroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.9	22.9	78-34-5	5 ppm (55mg/m <sup>3</sup> /BH) (skin)	orl-rat 600mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	1999.8	23.0	78-00-5	10 ppm (45mg/m <sup>3</sup> /BH) (skin)	orl-rat 836mg/kg
44. 1,2,2-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.9	22.9	78-12-8	0.001 ppm	orl-rat 170mg/kg
45. Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	98-18-4	10 ppm (60mg/m <sup>3</sup> /BH)	orl-rat 149.8mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	orl-rat 480mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	N/A	orl-rat 269mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
49. p-Isopropyl tolue	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m <sup>3</sup> /BH)	orl-rat >2000mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-57-6	N/A	orl-rat 475mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m <sup>3</sup> /BH)	orl-rat 490mg/kg
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	orl-rat 5000mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	orl-rat 5000mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m <sup>3</sup> )	orl-rat 756mg/kg
55. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	orl-rat 5g/kg
56. m-Xylene	35162	050823	0.05	5.00	40005.8											



Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.

Created: Sat, Feb 17, 2024 at 8:56:46 AM.

Sampled: Sequence "021624-GC5M1", Method "GC5-M1".

Analyzed using Method "GC5-M1".

#### Comments

GC5-M1 Analysis by Candice Warren

Column ID SPB-Vocel 105 meter X 0.53mm X 3.0µm film thickness

Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,


Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.

Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.).

Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.

FID Signal = Edaq Channel 1

Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	0.97
2	1,1,2-Trichloro-1,2,3-trifluoropropane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.74
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,3-Dichloropropene	18.60
12	Methylacrylonitrile/Methyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloropropane	19.91
14	1,1-Dichloropropane	20.79
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.68
18	1,2-Dichloropropane	24.26
19	Methyl methacrylate	24.52
20	Bromoacetylbenzene	25.13
21	Dibromoethane/1,2-Dibromoethane	25.48
22	cis-1,3-Dichloropropene	27.07
23	Toluene	28.03
24	Ethyl methacrylate/trans-1,3-Dichloropropene	29.73
25	1,1,2-Trichloroethane	29.94
26	Toluene/butanone/1,2-Dichloropropane	30.84
27	Dibromochloromethane	31.16
28	1,2-Dibromoethane	31.89
29	Chlorobutane	33.20
30	Ethylobenzenes/1,1,2-Tetrachloroethane	33.60
31	m-Xyleno/p-Xylene	33.89
32	o-Xylene	33.93
33	Glyrene	35.94
34	Isopropylbenzene/Bromoform	36.48
35	cis-1,4-Dichloro-1-butene	36.60
36	1,1,2,2-Tetrachloroethane	37.23
37	j,2,3-Trichloropropene	37.77
38	n-Propylbenzene	37.92
39	trans-1,3-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	2-Chlorotoluene	38.63
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	40.91
46	Pentachloroethane	40.17
47	tert-Butylbenzene	40.53
48	p Isopropyltoluene	41.03
49	1,3-Dichloropropane	41.13
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.62
52	1,2-Dichlorobenzene	42.16
53	1,2-Dibromo-3-chloropropane	46.17
54	Nitrobenzene	46.49
55	1,2,4-Trimethoxybenzene	49.26
56	Hexachlorobutadiene	49.32
57	Naphthalene	50.16
58	1,2,3-Trichlorobenzene	51.16



Boiling Point	65°C	Specific Gravity (H <sub>2</sub> O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

**Section X. STABILITY AND REACTIVITY**

Chemical stability: Stable under recommended storage conditions.  
 Possibility of hazardous reactions: Vapours may form explosive mixture with air.  
 Conditions to avoid: Heat, flames, sparks, extreme temperature and sunlight.  
 Materials to avoid: Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids  
 Hazardous decomposition products formed under fire conditions. - Carbon oxides

**Section XI. TOXICOLOGICAL INFORMATION**

LD50 Oral - rat - 5,628 mg/kg  
 LC50 Inhalation - rat - 4 h - 64000 ppm  
 LD50 Dermal - rabbit - 15,800 mg/kg  
 Toxic if absorbed through skin. Causes skin irritation.  
 Eye damage/eye irritation  
 Toxic if inhaled. Causes respiratory tract irritation.  
 Toxic if swallowed.

**Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.**

LC50 15,400 mg/l - 96 h  
 EC50 24,500.00 mg/l - 48 h  
 EC100 10,000.00 mg/l - 24 h

**Section XIII. DISPOSAL CONSIDERATIONS**

Dispose with normal Laboratory Solvent Waste.

**Section XIV. TRANSPORT INFORMATION**

DOT (US) IATA  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol

**Section XV. REGULATORY INFORMATION**

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant  
 SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

**Section XVI. Misc. INFORMATION**

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



## Certified Reference Material CRM

Dec 09/17/24

2 V1of

ANAB ISO 17034 Accredited  
AR-1539 Certificate Num:  
<https://Absolutestandards.co...>


## CERTIFIED WEIGHT REPORT

Part Number:	<u>91980</u>	Solvent(s):	Lot#
Lot Number:	<u>091424</u>	Water	072324Q
Description:	Acrolein		
Expiration Date:	101424		
Recommended Storage:	Refrigerate (4 °C)		
Nominal Concentration (µg/mL):	5000		
NIST Test ID#:	6UTB	5E-05 Balance Uncertainty	
Weight(s) shown below were combined and diluted to (mL):	10.0	0.001 Flask Uncertainty	

	091424
Formulated By: Justin Dippold	DATE
	091424
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity	Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
				(%)	Purity	Weight(g)	Conc (µg/mL)	(+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50	
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

**Method:** GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C ( Time 2 = 8.75 min.). **Rate** = 4°C/min., **Injector Temp.** = 200°C, **Detector Temp.** = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5 % of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



## Certified Reference Material CRM

Dec 09/17/24

2 Uvof

## CERTIFIED WEIGHT REPORT

Part Number: 91980  
 Lot Number: 091424  
 Description: Acrolein

Solvent(s): Water  
 Lot #: 072324Q

ANAB ISO 17034 Accredited  
 AR-1539 Certificate Num:  
<https://Absolutestandards.co...>


Expiration Date: 101424  
 Recommended Storage: Refrigerate (4 °C)  
 Nominal Concentration (µg/mL): 5000  
 NIST Test ID#: 6UTB  
 Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty  
 0.001 Flask Uncertainty

		091424
Formulated By:	Justin Dippold	DATE
		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

**Method:** GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C ( Time 2 = 8.75 min.). Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5 % of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Rec 12/16/24



CERTIFIED WEIGHT REPORT

Part Number: 95318  
Lot Number: 120524  
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#  
Methanol EJ143-US

Expiration Date: 120527  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration (µg/mL): 10000  
NIST Test ID#: 6UTB




✓ 14630 to  
✓ 14649

Weight(s) shown below were combined and diluted to (mL): 50.0      5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>	120524
Formulated By:	Prashant Chauhan
<i>Pedro L. Rentas</i>	120524
Reviewed By:	Pedro L. Rentas

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information			
										CAS#	Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	oral-rat 250mg/kg	

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,  
Injector B Temp.= 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified ( $\pm$ ) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Boiling Point		Specific Gravity (H <sub>2</sub> O = 1)	
Vapor Pressure (mm Hg)	65°C	Melting Point	0.79
Vapor Density (AIR = 1)	96	Evaporation rate (Butyl Acetate = 1)	-98°C
Solubility in Water	1.11		4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

**Section X. STABILITY AND REACTIVITY**

Chemical stability Stable under recommended storage conditions.  
 Possibility of hazardous reactions Vapours may form explosive mixture with air.  
 Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.  
 Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids  
 Hazardous decomposition products formed under fire conditions. - Carbon oxides

**Section XI. TOXICOLOGICAL INFORMATION**

LD50 Oral - rat - 5,628 mg/kg  
 LC50 Inhalation - rat - 4 h - 64000 ppm  
 LD50 Dermal - rabbit - 15,800 mg/kg  
 Toxic if absorbed through skin. Causes skin irritation.  
 Eye damage/eye irritation  
 Toxic if inhaled. Causes respiratory tract irritation.  
 Toxic if swallowed.

**Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.**

LC50 15,400 mg/l - 96 h  
 EC50 24,500.00 mg/l - 48 h  
 EC100 10,000.00 mg/l - 24 h

**Section XIII. DISPOSAL CONSIDERATIONS**

Dispose with normal Laboratory Solvent Waste.

**Section XIV. TRANSPORT INFORMATION**

DOT (US) IATA  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol

**Section XV. REGULATORY INFORMATION**

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant  
 SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

**Section XVI. Misc. INFORMATION**

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Rec 12/16/24



CERTIFIED WEIGHT REPORT

Part Number: 95318  
Lot Number: 120524  
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#  
Methanol EJ143-US

Expiration Date: 120527  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration ( $\mu\text{g/mL}$ ): 10000  
NIST Test ID#: 6UTB

V14630 to  
V14649



Weight(s) shown below were combined and diluted to (mL): 50.0

5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>	120524
Formulated By:	Prashant Chauhan
<i>Pedro L. Rentas</i>	120524
Reviewed By:	Pedro L. Rentas

Compound	RM#	Lot Number	Nominal Conc ( $\mu\text{g/mL}$ )	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc( $\mu\text{g/mL}$ )	Expanded Uncertainty (+/-) ( $\mu\text{g/mL}$ )	SDS Information			
										(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	oral-rat 250mg/kg	

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5  $\mu\text{m}$ ). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,  
Injector B Temp.= 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified ( $\pm$ ) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

## **Safety Data Sheet (SDS)      GHS/OSHA Compliant**

## **Section I Product and Company Identification**

**IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

## **Section II - Hazards Identification**

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



## **Signal Word: DANGER**

### **Section III - Composition**

Components (Specific Chemical Identity; Common Name(s)) % (optional)  
Methanol METHYL ALCOHOL CAS#: 67-56-1 > 97

**See Certified Weight Report For Other Analytes Present At Trace Quantities.**

**INTENDED USE: REFERENCE MATERIAL**

#### **Section IV. FIRST AID MEASURES**

**General advice** Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.  
**If inhaled** If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.  
**In case of skin contact** Wash with soap and water. Consult a physician.  
**In case of eye contact** Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.  
**If swallowed** Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

## **Section V. FIREFIGHTING MEASURES**

**Flammability** Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.  
**Suitable extinguishing media** Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.  
**Protective equipment for fire** Wear self contained breathing apparatus for fire fighting if necessary.

## **Section VI. ACCIDENTAL RELEASE MEASURES**

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

## **Section VII. HANDLING AND STORAGE**

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

## **Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION**

**Methanol** 67-56-1 TWA 200 ppm  
**Skin notation** TWA 200 ppm  
Potential for skin absorption, ingestion and inhalation.  
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection. Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

## **Section IX - Physical/Chemical Characteristics**

Boiling Point		Specific Gravity (H <sub>2</sub> O = 1)	
Vapor Pressure (mm Hg)	65°C	Melting Point	0.79
Vapor Density (AIR = 1)	96	Evaporation rate (Butyl Acetate = 1)	-98°C
Solubility in Water	1.11		4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

**Section X. STABILITY AND REACTIVITY**

Chemical stability Stable under recommended storage conditions.  
 Possibility of hazardous reactions Vapours may form explosive mixture with air.  
 Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.  
 Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids  
 Hazardous decomposition products formed under fire conditions. - Carbon oxides

**Section XI. TOXICOLOGICAL INFORMATION**

LD50 Oral - rat - 5,628 mg/kg  
 LC50 Inhalation - rat - 4 h - 64000 ppm  
 LD50 Dermal - rabbit - 15,800 mg/kg  
 Toxic if absorbed through skin. Causes skin irritation.  
 Eye damage/eye irritation  
 Toxic if inhaled. Causes respiratory tract irritation.  
 Toxic if swallowed.

**Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.**

LC50 15,400 mg/l - 96 h  
 EC50 24,500.00 mg/l - 48 h  
 EC100 10,000.00 mg/l - 24 h

**Section XIII. DISPOSAL CONSIDERATIONS**

Dispose with normal Laboratory Solvent Waste.

**Section XIV. TRANSPORT INFORMATION**

DOT (US) IATA  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol

**Section XV. REGULATORY INFORMATION**

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant  
 SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

**Section XVI. Misc. INFORMATION**

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Rec 12/16/24



CERTIFIED WEIGHT REPORT

Part Number: 95318  
Lot Number: 120524  
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#  
Methanol EJ143-US

Expiration Date: 120527  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration (µg/mL): 10000  
NIST Test ID#: 6UTB




✓ 14630 to  
✓ 14649

Weight(s) shown below were combined and diluted to (mL): 50.0      5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>	120524
Formulated By:	Prashant Chauhan
<i>Pedro L. Rentas</i>	120524
Reviewed By:	Pedro L. Rentas

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information			
										CAS#	Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	oral-rat 250mg/kg	

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,  
Injector B Temp.= 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified ( $\pm$ ) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

## Safety Data Sheet (SDS) GHS/OSHA Compliant

**Section I Product and Company Identification****IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name      ABSOLUTE STANDARDS INC      Emergency Telephone USA & CANADA      1-800-535-5053  
 Address      44 Rossotto Dr.      Emergency Telephone International      1-352-323-3500  
                  Hamden CT, 06514      Date Prepared/Revised      January 1, 2024

**Section II - Hazards Identification****GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)**

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water

**Signal Word: DANGER****Section III - Composition**

Components (Specific Chemical Identity; Common Name(s))      % (optional)  
 Methanol      METHYL ALCOHOL      > 97

**See Certified Weight Report For Other Analytes Present At Trace Quantities.****INTENDED USE: REFERENCE MATERIAL****Section IV. FIRST AID MEASURES**

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

**Section V. FIREFIGHTING MEASURES**

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.

Protective equipment for fire      Wear self contained breathing apparatus for fire fighting if necessary.

**Section VI. ACCIDENTAL RELEASE MEASURES**

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.

Clean up      Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

**Section VII. HANDLING AND STORAGE**

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

**Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION**

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

**Section IX - Physical/Chemical Characteristics**

Boiling Point		Specific Gravity (H <sub>2</sub> O = 1)	
Vapor Pressure (mm Hg)	65°C	Melting Point	0.79
Vapor Density (AIR = 1)	96	Evaporation rate (Butyl Acetate = 1)	-98°C
Solubility in Water	1.11		4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

**Section X. STABILITY AND REACTIVITY**

Chemical stability Stable under recommended storage conditions.  
 Possibility of hazardous reactions Vapours may form explosive mixture with air.  
 Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.  
 Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids  
 Hazardous decomposition products formed under fire conditions. - Carbon oxides

**Section XI. TOXICOLOGICAL INFORMATION**

LD50 Oral - rat - 5,628 mg/kg  
 LC50 Inhalation - rat - 4 h - 64000 ppm  
 LD50 Dermal - rabbit - 15,800 mg/kg  
 Toxic if absorbed through skin. Causes skin irritation.  
 Eye damage/eye irritation  
 Toxic if inhaled. Causes respiratory tract irritation.  
 Toxic if swallowed.

**Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.**

LC50 15,400 mg/l - 96 h  
 EC50 24,500.00 mg/l - 48 h  
 EC100 10,000.00 mg/l - 24 h

**Section XIII. DISPOSAL CONSIDERATIONS**

Dispose with normal Laboratory Solvent Waste.

**Section XIV. TRANSPORT INFORMATION**

DOT (US) IATA  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol

**Section XV. REGULATORY INFORMATION**

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant  
 SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

**Section XVI. Misc. INFORMATION**

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Rec 12/16/24



CERTIFIED WEIGHT REPORT

Part Number: 95318  
Lot Number: 120524  
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#  
Methanol EJ143-US

Expiration Date: 120527  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration (µg/mL): 10000  
NIST Test ID#: 6UTB




✓ 14630 to  
✓ 14649

Weight(s) shown below were combined and diluted to (mL): 50.0      5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>	120524
Formulated By:	Prashant Chauhan
<i>Pedro L. Rentas</i>	120524
Reviewed By:	Pedro L. Rentas

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information			
										CAS#	Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	oral-rat 250mg/kg	

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,  
Injector B Temp.= 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified ( $\pm$ ) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

## Safety Data Sheet (SDS) GHS/OSHA Compliant

**Section I Product and Company Identification****IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name      ABSOLUTE STANDARDS INC      Emergency Telephone USA & CANADA      **1-800-535-5053**  
 Address      44 Rossotto Dr.      Emergency Telephone International      **1-352-323-3500**  
                   Hamden CT, 06514      Date Prepared/Revised      January 1, 2024

**Section II - Hazards Identification****GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)**

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water

**Signal Word: DANGER****Section III - Composition**

Components (Specific Chemical Identity; Common Name(s))      % (optional)  
 Methanol      METHYL ALCOHOL      > 97

**See Certified Weight Report For Other Analytes Present At Trace Quantities.****INTENDED USE: REFERENCE MATERIAL****Section IV. FIRST AID MEASURES**

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

**Section V. FIREFIGHTING MEASURES**

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.

Protective equipment for fire      Wear self contained breathing apparatus for fire fighting if necessary.

**Section VI. ACCIDENTAL RELEASE MEASURES**

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.

Clean up      Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

**Section VII. HANDLING AND STORAGE**

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

**Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION**

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

**Section IX - Physical/Chemical Characteristics**

Boiling Point		Specific Gravity (H <sub>2</sub> O = 1)	
Vapor Pressure (mm Hg)	65°C	Melting Point	0.79
Vapor Density (AIR = 1)	96	Evaporation rate (Butyl Acetate = 1)	-98°C
Solubility in Water	1.11		4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

**Section X. STABILITY AND REACTIVITY**

Chemical stability Stable under recommended storage conditions.  
 Possibility of hazardous reactions Vapours may form explosive mixture with air.  
 Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.  
 Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids  
 Hazardous decomposition products formed under fire conditions. - Carbon oxides

**Section XI. TOXICOLOGICAL INFORMATION**

LD50 Oral - rat - 5,628 mg/kg  
 LC50 Inhalation - rat - 4 h - 64000 ppm  
 LD50 Dermal - rabbit - 15,800 mg/kg  
 Toxic if absorbed through skin. Causes skin irritation.  
 Eye damage/eye irritation  
 Toxic if inhaled. Causes respiratory tract irritation.  
 Toxic if swallowed.

**Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.**

LC50 15,400 mg/l - 96 h  
 EC50 24,500.00 mg/l - 48 h  
 EC100 10,000.00 mg/l - 24 h

**Section XIII. DISPOSAL CONSIDERATIONS**

Dispose with normal Laboratory Solvent Waste.

**Section XIV. TRANSPORT INFORMATION**

DOT (US) IATA  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol  
 UN number: 1230 Class: 3 Packing group: II  
 Proper shipping name: Methanol

**Section XV. REGULATORY INFORMATION**

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant  
 SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

**Section XVI. Misc. INFORMATION**

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL



# Certificate of Analysis

*chromatographic plus*

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 30067

**Lot No.:** A0191805

**Description :** 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500 $\mu$ g/mL, P&T Methanol,  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** November 30, 2027

**Storage:** 0°C or colder

**Ship:** Ambient

### C E R T I F I E D   V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 $\mu$ g/mL	+/- 139.5488

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

  
Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022      Balance Serial #: B251644995

  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

### Certificate of Analysis *gravimetric*



#### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555582

**Lot No.:** A0196865

**Description :** Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000 $\mu$ g/mL, P&T Methanol,  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** April 30, 2026

**Storage:** 10°C or colder

**Ship:** Ambient

#### C E R T I F I E D   V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 $\mu$ g/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 $\mu$ g/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 $\mu$ g/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 $\mu$ g/mL	+/- 1,417.4648

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

Russ Bookhamer - Operations Technician

Date Mixed: 11-Apr-2023 Balance: 1127510105

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ $\mu$ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*



**ILAC**  
ACCREDITED  
ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



**ILAC**  
ACCREDITED  
ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 30489

**Lot No.:** A0209618

**Description :** 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025

**Storage:** -20°C or colder

**Handling:** This product is photosensitive.

**Ship:** On Ice

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

### Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

## Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

*Samuel Moodier*  
Sam Moodier - Operations Tech I

Date Mixed: 28-Mar-2024      Balance Serial #: B707717271

*Dillan Murphy*  
Dillan Murphy - Operations Technician |

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ $\mu$ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

$k$  is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*



**ILAC**  
ACCREDITED  
ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



**ILAC**  
ACCREDITED  
ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 30489

**Lot No.:** A0209618

**Description :** 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025

**Storage:** -20°C or colder

**Handling:** This product is photosensitive.

**Ship:** On Ice

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

### Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

## Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

*Samuel Moodier*  
Sam Moodier - Operations Tech I

Date Mixed: 28-Mar-2024      Balance Serial #: B707717271

*Dillan Murphy*  
Dillan Murphy - Operations Technician |

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

### Certificate of Analysis *gravimetric*



#### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No.:** 555581

**Lot No.:** A0210184

**Description :** Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000 $\mu$ g/mL, P&T Methanol,  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** April 30, 2027


**Storage:** 10°C or colder

**Ship:** Ambient

#### C E R T I F I E D   V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 $\mu$ g/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 $\mu$ g/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 $\mu$ g/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 $\mu$ g/mL	+/- 1,426.0734

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

  
John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

APPROVED  
By Analyst Name: 1.80 mg/g for 10.00 g

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

Dec 12 (17) 24



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

# Certificate of Analysis

*chromatographic plus*

V14697-to-14726

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

**Catalog No. :** 30006

**Lot No.:** A0210618

**Description :** VOA Calibration Mix #1

VOA Calibration Mix #1 5,000 $\mu$ g/mL, P&T Methanol/Water(90:10),  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2027

**Storage:** 0°C or colder

**Ship:** Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 $\mu$ g/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 $\mu$ g/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 $\mu$ g/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 $\mu$ g/mL	+/- 173.2261

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol/Water (90:10)

**CAS #** 67-56-1/7732-18-5

**Purity** 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**


1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

  
Dakota Parson - Operations Technician I.

Date Mixed: 22-Apr-2024      Balance Serial #: B707717271

  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ $\mu$ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

Dec 12 (17) 24

30 v14

# Certificate of Analysis

chromatographic plus

V14697-to-14726



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

**Catalog No. :** 30006

**Lot No.:** A0210618

**Description :** VOA Calibration Mix #1

VOA Calibration Mix #1 5,000 $\mu$ g/mL, P&T Methanol/Water(90:10),  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2027

**Storage:** 0°C or colder

**Ship:** Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 $\mu$ g/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 $\mu$ g/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 $\mu$ g/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 $\mu$ g/mL	+/- 173.2261

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol/Water (90:10)

**CAS #** 67-56-1/7732-18-5

**Purity** 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**


1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

  
Dakota Parson - Operations Technician I.

Date Mixed: 22-Apr-2024      Balance Serial #: B707717271

  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ $\mu$ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

Dec 12 (17) 24

30 v14

# Certificate of Analysis

*chromatographic plus*

V14697-to-14726



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

**Catalog No. :** 30006

**Lot No.:** A0210618

**Description :** VOA Calibration Mix #1

VOA Calibration Mix #1 5,000 $\mu$ g/mL, P&T Methanol/Water(90:10),  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2027

**Storage:** 0°C or colder

**Ship:** Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 $\mu$ g/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 $\mu$ g/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 $\mu$ g/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 $\mu$ g/mL	+/- 173.2261

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol/Water (90:10)

**CAS #** 67-56-1/7732-18-5

**Purity** 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**


1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

  
Dakota Parson - Operations Technician I.

Date Mixed: 22-Apr-2024      Balance Serial #: B707717271

  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ $\mu$ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.





110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

Dec 12 (17) 24

30 v14

# Certificate of Analysis

chromatographic plus

V14697-to-14726



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

**Catalog No. :** 30006

**Lot No.:** A0210618

**Description :** VOA Calibration Mix #1

VOA Calibration Mix #1 5,000 $\mu$ g/mL, P&T Methanol/Water(90:10),  
1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2027

**Storage:** 0°C or colder

**Ship:** Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 $\mu$ g/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 $\mu$ g/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 $\mu$ g/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 $\mu$ g/mL	+/- 173.2261

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol/Water (90:10)

**CAS #** 67-56-1/7732-18-5

**Purity** 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**


1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

  
Dakota Parson - Operations Technician I.

Date Mixed: 22-Apr-2024      Balance Serial #: B707717271

  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ $\mu$ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.





110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

See: 12/09/24

## Certificate of Analysis

chromatographic plus

V14667+  
V14676



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225

Lot No.: A0214960

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : August 31, 2029

Storage: 0°C or colder

Ship: Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

\* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner  
Stacey Wanner - Operations Technician |

Date Mixed: 08-Aug-2024 Balance Serial #: 1127510105

Jennifer Pollino  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

See: 12/09/24

## Certificate of Analysis

chromatographic plus

V14667-5  
V14676



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225

Lot No.: A0214960

Description : Bromochloromethane Standard

Bromochloromethane 2000 $\mu$ g/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : August 31, 2029

Storage: 0°C or colder

Ship: Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 $\mu$ g/mL	+/- 113.0519

\* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner  
Stacey Wanner - Operations Technician |

Date Mixed: 08-Aug-2024 Balance Serial #: 1127510105

Jennifer Pollino  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

See: 12/09/24

## Certificate of Analysis

chromatographic plus

V14667+  
V14676



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225

Lot No.: A0214960

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : August 31, 2029

Storage: 0°C or colder

Ship: Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

\* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner  
Stacey Wanner - Operations Technician |

Date Mixed: 08-Aug-2024 Balance Serial #: 1127510105

Jennifer Pollino  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

See: 12/09/24

## Certificate of Analysis

chromatographic plus

V14667+  
V14676



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225

Lot No.: A0214960

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : August 31, 2029

Storage: 0°C or colder

Ship: Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

\* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner  
Stacey Wanner - Operations Technician |

Date Mixed: 08-Aug-2024 Balance Serial #: 1127510105

Jennifer Pollino  
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

*Rec 12/17/24*  
**CERTIFIED REFERENCE MATERIAL**

*30 μL*



**ILAC**  
ACCREDITED  
ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222-01



**ILAC**  
ACCREDITED  
ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222-02

**Certificate of Analysis**  
*chromatographic plus*

*V14727 +  
V14756*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 30042

**Lot No.:** A0216826

**Description :** 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000μg/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** May 31, 2031

**Storage:** 0°C or colder

**Ship:** Ambient

**C E R T I F I E D V A L U E S**

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 μg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 μg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 μg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 μg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 μg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 μg/mL	+/- 112.3992

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol

**CAS #** 67-56-1

**Purity** 99%

# Quality Confirmation Test

**Column:**

60m x 0.25mm x 1.4 $\mu$ m  
Rtx-502.2 (cat.#10916)

**Carrier Gas:**

helium-constant flow 2.0 mL/min.

**Temp. Program:**

40°C (hold 6 min.) to 100°C  
@ 6°C/min.

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


MSD

**Split Vent:**

Split ratio 10:1

**Inj. Vol**

1 $\mu$ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar Mix Technician

Date Mixed: 23-Sep-2024 Balance Serial #: B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

*Rev 12/17/24*  
**CERTIFIED REFERENCE MATERIAL**

*30 mL*



**ILAC**  
ACCREDITED  
ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222-01



**ILAC**  
ACCREDITED  
ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222-02

**Certificate of Analysis**  
*chromatographic plus*

*V14727 +  
V14756*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 30042

**Lot No.:** A0216826

**Description :** 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** May 31, 2031

**Storage:** 0°C or colder

**Ship:** Ambient

**C E R T I F I E D V A L U E S**

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol

**CAS #** 67-56-1

**Purity** 99%

# Quality Confirmation Test

**Column:**

60m x 0.25mm x 1.4 $\mu$ m  
Rtx-502.2 (cat.#10916)

**Carrier Gas:**

helium-constant flow 2.0 mL/min.

**Temp. Program:**

40°C (hold 6 min.) to 100°C  
@ 6°C/min.

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


MSD

**Split Vent:**

Split ratio 10:1

**Inj. Vol**

1 $\mu$ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar Mix Technician

Date Mixed: 23-Sep-2024 Balance Serial #: B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL



ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

## Certificate of Analysis

*chromatographic plus*

✓ 14842 to 14846

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No.:** 30470

**Lot No.:** A0217535

**Description :** tert-Butanol Standard

tert-Butanol Std 50,000 $\mu$ g/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** October 31, 2027

**Storage:** 0°C or colder

**Ship:** Ambient

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 $\mu$ g/mL	+/- 717.6137

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

*Aaron Enyart*  
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024      Balance Serial #: B251644995

*Brittany Federinko*  
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

2014 Dec 01 (08/21)



ILAC  
ACCREDITED  
ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ILAC  
ACCREDITED  
ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

## Certificate of Analysis

chromatographic

J14803 - J14822

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

**Catalog No.:** 555408-SL

**Lot No.:** A0220471

**Description :** Custom Vinyl Acetate Standard

Custom Vinyl Acetate Standard 8,000 $\mu$ g/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** June 30, 2026

**Storage:** -20°C or colder

**Handling:** This product is photosensitive.

**Ship:** On Ice

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 $\mu$ g/mL	+/- 278.7979

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol

**CAS #** 67-56-1

**Purity** 99%

### Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

*Ethan Winiarski*  
Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024 Balance Serial #: 1127510105

*Dillan Murphy*  
Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED  
By Jennifer Polson at 7:17 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: 1-814-353-1300  
Fax: 1-814-353-1309

[www.restek.com](http://www.restek.com)

## CERTIFIED REFERENCE MATERIAL

2014 Dec 01 (08/21)



ILAC  
ACCREDITED  
ISO 17034 Accredited  
Reference Material Producer  
Certificate #3222.01



ILAC  
ACCREDITED  
ISO/IEC 17025 Accredited  
Testing Laboratory  
Certificate #3222.02

# Certificate of Analysis

chromatographic

J14803 - J14822

### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

**Catalog No.:** 555408-SL

**Lot No.:** A0220471

**Description :** Custom Vinyl Acetate Standard

Custom Vinyl Acetate Standard 8,000 $\mu$ g/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL

**Pkg Amt:** > 1 mL

**Expiration Date :** June 30, 2026

**Storage:** -20°C or colder

**Handling:** This product is photosensitive.

**Ship:** On Ice

### C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 $\mu$ g/mL	+/- 278.7979

\* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** P&T Methanol  
**CAS #** 67-56-1  
**Purity** 99%

### Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

# Quality Confirmation Test

**Column:**

105m x 0.53mm x 3.0 $\mu$ m  
Rtx-502.2 (cat.#10910)

**Carrier Gas:**

hydrogen-constant pressure 11.0 psi.

**Temp. Program:**

40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**

200°C

**Det. Temp:**

250°C

**Det. Type:**


FID

**Split Vent:**

40 ml/min

**Inj. Vol**

1 $\mu$ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

*Ethan Winiarski*  
Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024      Balance Serial #: 1127510105

*Dillan Murphy*  
Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED  
By Jennifer Polson at 7:17 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

*k* is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis

Rev 05/09/25  
**avantor**™



V14921 to  
V14938

Material No.: 9077-02  
Batch No.: 24G0262002  
Manufactured Date: 2024-05-14  
Expiration Date: 2027-05-14  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay ( $\text{CH}_3\text{OH}$ ) (by GC, corrected for water)	$\geq 99.9 \%$	100.0 %
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.3 ppm
Titrable Acid ( $\mu\text{eq/g}$ )	$\leq 0.3$	0.3
Titrable Base ( $\mu\text{eq/g}$ )	$\leq 0.10$	0.03
Water (by KF, coulometric)	$\leq 0.08 \%$	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature of the name "Jamie Croak".

Jamie Croak  
Director Quality Operations, Bioscience Production

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis

Rev 05/09/25  
**avantor**™



V14921 to  
V14938

Material No.: 9077-02  
Batch No.: 24G0262002  
Manufactured Date: 2024-05-14  
Expiration Date: 2027-05-14  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay ( $\text{CH}_3\text{OH}$ ) (by GC, corrected for water)	$\geq 99.9 \%$	100.0 %
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.3 ppm
Titrable Acid ( $\mu\text{eq/g}$ )	$\leq 0.3$	0.3
Titrable Base ( $\mu\text{eq/g}$ )	$\leq 0.10$	0.03
Water (by KF, coulometric)	$\leq 0.08 \%$	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature of the name "Jamie Croak".

Jamie Croak  
Director Quality Operations, Bioscience Production



see US/21(25)



CERTIFIED WEIGHT REPORT

**Part Number:** 91980  
**Lot Number:** 051925  
**Description:** Acrolein

Solvent(s):      Lot#


Expiration Date: 061925  
Recommended Storage: Refrigerate (4 °C)  
Concentration (µg/mL): 5000  
NIST Test ID#: 61UTR

**Weight(s) shown below were combined and diluted to (mL):** 10.0 0.001 **Flask Uncertainty:**

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V105	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ext-est 46mg/kg

**Method:** GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5 $\mu$ m film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C ( Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Renias. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
  - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
  - Standards are certified ( $\pm 0.5\%$  of the stated value, unless otherwise stated).
  - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
  - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: 91980  
Lot Number: 051925  
Description: Acrolein

Solvent(s): Water  
Lot# 041725Q

5. v1g  
J14944-V14948

Expiration Date: 061925  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration ( $\mu\text{g/mL}$ ): 5000  
NIST Test ID#: 6UTB


Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

	051925
Formulated By:	Lawrence Barry
	051925
Reviewed By:	Pedro L. Rentas

Compound	RM#	Lot Number	Nominal Conc ( $\mu\text{g/mL}$ )	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc ( $\mu\text{g/mL}$ )	SDS Information			
									(+/-) ( $\mu\text{g/mL}$ )	CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5 $\mu\text{m}$  film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C ( Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
  - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
  - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
  - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
  - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: 91980  
Lot Number: 051925  
Description: Acrolein

Solvent(s): Water  
Lot# 041725Q

5. v1g  
J14944-V14948

Expiration Date: 061925  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration ( $\mu\text{g/mL}$ ): 5000  
NIST Test ID#: 6UTB


Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

	051925
Formulated By:	Lawrence Barry
	051925
Reviewed By:	Pedro L. Rentas

Compound	RM#	Lot Number	Nominal Conc ( $\mu\text{g/mL}$ )	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc ( $\mu\text{g/mL}$ )	Expanded Uncertainty (+/-) ( $\mu\text{g/mL}$ )	SDS Information		
										(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5 $\mu\text{m}$  film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C ( Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
  - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
  - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
  - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
  - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: 91980  
Lot Number: 051925  
Description: Acrolein

Solvent(s): Water  
Lot# 041725Q

5. v1g  
J14944-V14948

Expiration Date: 061925  
Recommended Storage: Refrigerate (4 °C)  
Nominal Concentration ( $\mu$ g/mL): 5000  
NIST Test ID#: 6UTB


Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty  
0.001 Flask Uncertainty

<i>Lawrence Barry</i>	051925
Formulated By: Lawrence Barry	DATE
<i>Pedro Rentas</i>	051925
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc ( $\mu$ g/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc ( $\mu$ g/mL)	Expanded Uncertainty (+/-) ( $\mu$ g/mL)	SDS Information		
										(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5 $\mu$ m film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C ( Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
  - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
  - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
  - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
  - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



# SHIPPING DOCUMENTS

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16



284 Sheffield Street, Mountainside, NJ 07092  
 (908) 789-8900 • Fax (908) 789-8922  
[www.chemtech.net](http://www.chemtech.net)

ALLIANCE PROJECT NO.

QUOTE NO.

Q2250

COC Number

2046478

**CLIENT INFORMATION**

REPORT TO BE SENT TO:

COMPANY: Jacobs

ADDRESS: 412 Mt. Kemble Ave., Suite 100

CITY Morristown STATE: NJ ZIP: 07960

ATTENTION: John Vinfate John.Vinfate@Jacobs.com

PHONE: FAX:

**CLIENT PROJECT INFORMATION**

PROJECT NAME: STC Princeton

PROJECT NO.: D3868221 LOCATION: Princeton Junction

PROJECT MANAGER: Mary Murphy

E-mail: Mary.Murphy@Jacobs.com

PHONE:

FAX:

**CLIENT BILLING INFORMATION**

BILL TO: Mary Murphy

PO#:

ADDRESS:

CITY STATE: ZIP:

ATTENTION: PHONE:

**ANALYSIS**

**DATA TURNAROUND INFORMATION**

FAX (RUSH) Standard TAT DAYS\*

HARDCOPY (DATA PACKAGE) DAYS\*

EDD: DAYS\*

\*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

**DATA DELIVERABLE INFORMATION**

- Level 1 (Results Only)  Level 4 (QC + Full Raw Data)  
 Level 2 (Results + QC)  NJ Reduced  US EPA CLP  
 Level 3 (Results + QC)  NYS ASP A  NYS ASP B  
 + Raw Data  Other  
 EDD FORMAT

1. SITE Specific VOC's (60 VOC's)  
 14-Dioxane (60 VOC's)  
 1,4-Dioxane (60 VOC's)  
 Benzene (60 VOC's)  
 Chloroform (60 VOC's)  
 Crude Oil (60 VOC's)  
 Ethylbenzene (60 VOC's)  
 Hexane (60 VOC's)  
 Methylbenzene (60 VOC's)  
 Toluene (60 VOC's)  
 Xylenes (60 VOC's)

Alkalinity (SM 2540C)  
 Dissolved Solids (SM 2540C)  
 TDS (SM 2540C)  
 Anions (9050)

1. 2. 3. 4. 5. 6. 7. 8. 9. Tr.

**PRESERVATIVES**

**COMMENTS**

← Specify Preservatives  
 A-HCl D-NaOH  
 B-HNO3 E-ICE  
 C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS						
			COMP	GRAB	DATE	TIME		A/E	E	B/E	B/E	E	E	E	1	2	3	4	5	6	7	8	9
1.	MW-11A-13.5-060525	Gw	X		6/5/25	1230	15	X	X														MS/MSD
2.	MW-06-6.5-060525	Gw	X		6/5/25	1310	8																
3.	ER02-060525	DI	X		6/5/25	0930	10	X	X	X	X	X	X	X									
4.	TB-01-060525	DI	X		6/5/25	1600	2	X															
5.																							
6.																							
7.																							
8.																							
9.																							
10.																							

**SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY**

RELINQUISHED BY SAMPLER:

1. *Jeanne J.*

RELINQUISHED BY SAMPLER:

2.

RELINQUISHED BY SAMPLER:

3. *JT*

DATE/TIME: 1550

6/5/25

DATE/TIME:

DATE/TIME: 1730

6.5.25

RECEIVED BY:

1. *JT*

RECEIVED BY:

2.

RECEIVED BY:

3.

Conditions of bottles or coolers at receipt:  COMPLIANT  NON COMPLIANT  COOLER TEMP 3.2 °C

Comments: See work order for list of site specific VOC's

PO# 148064511

Page 1 of 2

CLIENT:  Hand Delivered  Other

Shipment Complete  
 YES  NO

**Laboratory Certification**

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

## LOGIN REPORT/SAMPLE TRANSFER

Order ID :	Q2250	JAC005	Order Date :	6/5/2025 4:22:00 PM	Project Mgr :
Client Name :	JACOBS Engineering Grou		Project Name :	Former Schlumberger STC	Report Type :
Client Contact :	John Ynfante		Receive DateTime :	6/5/2025 12:00:00 AM <i>17:30</i>	EDD Type :
Invoice Name :	JACOBS Engineering Grou		Purchase Order :		Hard Copy Date :
Invoice Contact :	John Ynfante				Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2250-01	MW-11A-13.5-060525	Water	06/05/2025	12:30	VOCMS Group3		8260-Low	10 Bus. Days	
Q2250-02	Q2250-01MS	Water	06/05/2025	12:30	VOCMS Group3		8260-Low	10 Bus. Days	
Q2250-03	Q2250-01MSD	Water	06/05/2025	12:30	VOCMS Group3		8260-Low	10 Bus. Days	
Q2250-04	<del>MW-06-6.5-060525</del>	<del>Water</del>	<del>06/05/2025</del>	<del>13:10</del>	<del>VOCMS Group3</del>		<del>8260-Low</del>	<del>10 Bus. Days</del>	yg 06/11/25
Q2250-05	EB02-060525	Water	06/05/2025	09:30	VOCMS Group3		8260-Low	10 Bus. Days	
Q2250-06	TB-01-060525	Water	06/05/2025	16:00	VOCMS Group3		8260-Low	10 Bus. Days	

## LOGIN REPORT/SAMPLE TRANSFER

**Order ID :** Q2250      **JACO05**  
**Client Name :** JACOBS Engineering Grou  
**Client Contact :** John Ynfante  
**Invoice Name :** JACOBS Engineering Grou  
**Invoice Contact :** John Ynfante

**Order Date :** 6/5/2025 4:22:00 PM      **Project Mgr :**  
**Project Name :** Former Schlumberger STC      **Report Type :** Level 4  
**Receive DateTime :** 6/5/2025 +2:00:00 AM      **EDD Type :** CH2MHILL  
**Purchase Order :** 17:30      **Hard Copy Date :**  
      **Date Signoff :**

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Relinquished By :

Date / Time : 6/6/25 0810

Samples received on 6/5/25  
placed in SM-REF-2

Received By :

Date / Time :

6/6/25 8:10 AM

Storage Area : VOA Refrigerator Room