



SAMPLE DATA

A

B

C

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	06/05/25 09:30
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	06/05/25
Client Sample ID:	EB02-060525	SDG No.:	Q2250
Lab Sample ID:	Q2250-05	Matrix:	Water
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Alkalinity	1.00	U	1	1.00	2.00	mg/L		06/10/25 12:55	SM 2320 B-11
Chloride	0.19	U	1	0.19	0.60	mg/L		06/06/25 11:55	9056A
Nitrate	0.095	U	1	0.095	0.50	mg/L		06/06/25 11:55	9056A
Sulfate	0.48	J	1	0.46	3.00	mg/L		06/06/25 11:55	9056A
TDS	76.0		1	1.00	10.0	mg/L		06/06/25 12:30	SM 2540 C-15

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits



LAB CHRONICLE

OrderID:	Q2250	OrderDate:	6/5/2025 4:22:00 PM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger STC PTC Site D3868221
Contact:	John Ynfante	Location:	D22,N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2250-05	EB02-060525	WATER			06/05/25 09:30			06/05/25
			Alkalinity	SM2320 B			06/10/25 12:55	
			Anions Group1	9056A			06/06/25 11:55	
			TDS	SM2540 C			06/06/25 12:30	