

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2250

Project: Former Schlumberger STC PTC Site D3868221

RunNo.: LB136018

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	05/22/2025
Chloride	mg/L	3.1	3	103	90-110	05/22/2025
Fluoride	mg/L	2.1	2	105	90-110	05/22/2025
Nitrite	mg/L	3.1	3	103	90-110	05/22/2025
Nitrate	mg/L	2.6	2.5	104	90-110	05/22/2025
Sulfate	mg/L	15.1	15	101	90-110	05/22/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	05/22/2025
Sample ID: CCV1						
Bromide	mg/L	10.2	10	102	90-110	06/05/2025
Chloride	mg/L	3.1	3	103	90-110	06/05/2025
Fluoride	mg/L	2	2	100	90-110	06/05/2025
Nitrite	mg/L	3.1	3	103	90-110	06/05/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/05/2025
Sulfate	mg/L	15.1	15	101	90-110	06/05/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	06/05/2025
Sample ID: CCV2						
Bromide	mg/L	10.3	10	103	90-110	06/05/2025
Chloride	mg/L	3.1	3	103	90-110	06/05/2025
Fluoride	mg/L	2.1	2	105	90-110	06/05/2025
Nitrite	mg/L	3.1	3	103	90-110	06/05/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/05/2025
Sulfate	mg/L	15.2	15	101	90-110	06/05/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	06/05/2025
Sample ID: CCV3						
Bromide	mg/L	10.3	10	103	90-110	06/06/2025
Chloride	mg/L	3.1	3	103	90-110	06/06/2025
Fluoride	mg/L	2	2	100	90-110	06/06/2025
Nitrite	mg/L	3.1	3	103	90-110	06/06/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/06/2025
Sulfate	mg/L	15.2	15	101	90-110	06/06/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	06/06/2025
Sample ID: CCV4						
Bromide	mg/L	10.3	10	103	90-110	06/06/2025
Chloride	mg/L	3.1	3	103	90-110	06/06/2025
Fluoride	mg/L	2.1	2	105	90-110	06/06/2025
Nitrite	mg/L	3.1	3	103	90-110	06/06/2025
Nitrate	mg/L	2.6	2.5	104	90-110	06/06/2025
Sulfate	mg/L	15.3	15	102	90-110	06/06/2025

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2250

Project: Former Schlumberger STC PTC Site D3868221

RunNo.: LB136018

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Orthophosphate as P	mg/L	5.4	5	108	90-110	06/06/2025