

Laboratory Control Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q2250
Project:	Former Schlumberger STC PTC Site D3868221	Run No.:	LB136018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136018BSW							
Bromide	mg/L	10	10.3		103	1	90-110	06/05/2025
Chloride	mg/L	3	3.10		103	1	90-110	06/05/2025
Fluoride	mg/L	2	2.00		100	1	90-110	06/05/2025
Nitrite	mg/L	3	3.10		103	1	90-110	06/05/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	06/05/2025
Sulfate	mg/L	15	15.3		102	1	90-110	06/05/2025
Orthophosphate as P	mg/L	5	5.30		106	1	90-110	06/05/2025

Laboratory Control Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q2250
Project:	Former Schlumberger STC PTC Site D3868221	Run No.:	LB136018

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136018BSW2							
Bromide	mg/L	10	10.3		103	1	90-110	06/06/2025
Chloride	mg/L	3	3.10		103	1	90-110	06/06/2025
Fluoride	mg/L	2	2.00		100	1	90-110	06/06/2025
Nitrite	mg/L	3	3.10		103	1	90-110	06/06/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	06/06/2025
Sulfate	mg/L	15	15.4		103	1	90-110	06/06/2025
Orthophosphate as P	mg/L	5	5.40		108	1	90-110	06/06/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2250

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB136041

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136041BS							
TDS	mg/L	100	95.0		95	1	90-110	06/06/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2250

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB136111

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136111BS							
Alkalinity	mg/L	50	46.2		92	1	80-120	06/10/2025