

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2250 SAS No.: Q2250 SDG NO.: Q2250

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037204.D Time Analyzed: 23:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1688	7.589	4447	10.37	2457	14.23
UPPER LIMIT	3376	8.089	8894	10.872	4914	14.734
LOWER LIMIT	844	7.089	2223.5	9.872	1228.5	13.734
EPA SAMPLE NO.						
01 MW-06-6.5-060525	1975	7.59	5099	10.36	2655	14.23

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.