

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2250

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD	RPD Qual	Limits		RPD
					Rec	Qual			Low	High	
Lab Sample ID :	Q2250-03MSD	Client Sample ID :	MW-11A-13.5-060525MSD						Datafile :	VN086923.D	
Vinyl chloride	50	5.00	57.2	ug/L	104		13		70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	50	4.20	53.1	ug/L	98		12		70 (53)	130 (162)	20 (20)
1,1-Dichloroethane	50	2.80	49.5	ug/L	93		9		70 (78)	130 (122)	20 (20)
cis-1,2-Dichloroethene	50	850	2100	ug/L	2500	*	17		70 (82)	130 (124)	20 (20)
1,1,1-Trichloroethane	50	0	41.4	ug/L	83		11		70 (83)	130 (117)	20 (20)
Benzene	50	0.42	43.0	ug/L	85		10		70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	50	0	43.1	ug/L	86		9		70 (76)	130 (120)	20 (20)
Trichloroethene	50	1400	3300	ug/L	3800	*	24	*	70 (28)	130 (175)	20 (20)
1,1,2-Trichloroethane	50	0.50	44.5	ug/L	88		11		70 (27)	130 (175)	20 (20)
Tetrachloroethene	50	9.40	66.0	ug/L	113		17		70 (48)	130 (153)	20 (20)