

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05  
Lab Code: CHEM Case No.: Q2250 SAS No.: Q2250 SDG NO.: Q2250  
Lab File ID: VN086913.D Date Analyzed: 06/10/2025  
Instrument ID: MSVOA\_N Time Analyzed: 09:13  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                       | IS1<br>AREA # | RT # | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|-----------------------|---------------|------|---------------|-------|---------------|--------|
| 12 HOUR STD           | 239452        | 8.23 | 422119        | 9.11  | 362285        | 11.87  |
| UPPER LIMIT           | 478904        | 8.73 | 844238        | 9.606 | 724570        | 12.365 |
| LOWER LIMIT           | 119726        | 7.73 | 211060        | 8.606 | 181143        | 11.365 |
| EPA SAMPLE NO.        |               |      |               |       |               |        |
| MW-11A-13.5-060525    | 361832        | 8.23 | 672514        | 9.11  | 593044        | 11.87  |
| MW-11A-13.5-060525MS  | 174378        | 8.23 | 310932        | 9.11  | 270945        | 11.87  |
| MW-11A-13.5-060525MSD | 172955        | 8.23 | 309901        | 9.11  | 266637        | 11.87  |
| EB02-060525           | 232754        | 8.23 | 433385        | 9.11  | 381245        | 11.87  |
| TB-01-060525          | 350151        | 8.23 | 648560        | 9.11  | 563762        | 11.87  |
| VN0610WBL01           | 241251        | 8.23 | 440901        | 9.11  | 386425        | 11.87  |
| VN0610WBS01           | 249843        | 8.23 | 443174        | 9.11  | 389297        | 11.87  |

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.