

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2250 SAS No.: Q2250 SDG NO.: Q2250
 Lab File ID: VN086913.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	170987	13.788				
UPPER LIMIT	341974	14.288				
LOWER LIMIT	85493.5	13.288				
EPA SAMPLE NO.						
MW-11A-13.5-060525	281220	13.79				
MW-11A-13.5-060525MS	132628	13.79				
MW-11A-13.5-060525MSD	129873	13.79				
EB02-060525	180063	13.79				
TB-01-060525	266280	13.79				
VN0610WBL01	183380	13.79				
VN0610WBS01	189620	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.