

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG No.: Q2251

Instrument ID: MSVOA_N Calibration Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF001 = VN086862.D	RRF005 = VN086863.D	RRF020 = VN086864.D	RRF050 = VN086865.D	RRF100 = VN086866.D	RRF150 = VN086867.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Chloromethane	0.762	0.654	0.645	0.597	0.617	0.587	0.644	9.9
Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664	2.5
Bromomethane		0.375	0.380	0.357	0.379	0.368	0.372	2.6
Chloroethane	0.460	0.444	0.442	0.408	0.418	0.402	0.429	5.4
Trichlorofluoromethane	0.882	0.903	0.904	0.834	0.858	0.825	0.868	3.9
1,1,2-Trichlorotrifluoroethane	0.554	0.567	0.563	0.519	0.546	0.520	0.545	3.8
1,1-Dichloroethene	0.573	0.593	0.563	0.533	0.550	0.527	0.557	4.4
Acetone	0.426	0.366	0.366	0.322	0.334	0.316	0.355	11.5
Carbon Disulfide	1.718	1.622	1.542	1.426	1.496	1.433	1.539	7.4
Methyl tert-butyl Ether	2.120	2.038	2.051	1.933	2.021	1.926	2.015	3.7
Methylene Chloride	0.822	0.688	0.643	0.605	0.629	0.601	0.665	12.5
trans-1,2-Dichloroethene	0.700	0.674	0.621	0.567	0.591	0.561	0.619	9.3
1,1-Dichloroethane	1.192	1.153	1.156	1.063	1.110	1.043	1.120	5.2
2-Butanone	0.604	0.598	0.604	0.551	0.573	0.533	0.577	5.2
Carbon Tetrachloride	0.453	0.449	0.434	0.409	0.435	0.421	0.433	3.9
cis-1,2-Dichloroethene	0.786	0.766	0.762	0.699	0.729	0.701	0.740	4.9
Chloroform	1.235	1.152	1.145	1.061	1.085	1.030	1.118	6.7
1,1,1-Trichloroethane	1.029	0.995	0.969	0.895	0.925	0.893	0.951	5.9
Methylcyclohexane	0.633	0.645	0.588	0.570	0.603	0.589	0.605	4.8
Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444	6.2
1,2-Dichloroethane	0.473	0.456	0.444	0.411	0.430	0.413	0.438	5.6
Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.328	0.342	4.2
1,2-Dichloropropane	0.366	0.369	0.354	0.332	0.352	0.335	0.351	4.4
Bromodichloromethane	0.510	0.484	0.480	0.457	0.483	0.465	0.480	3.8
4-Methyl-2-Pentanone	0.505	0.549	0.576	0.538	0.562	0.528	0.543	4.6
Toluene	0.918	0.914	0.885	0.835	0.883	0.859	0.882	3.6
t-1,3-Dichloropropene	0.571	0.526	0.524	0.522	0.548	0.530	0.537	3.6
cis-1,3-Dichloropropene	0.609	0.577	0.564	0.551	0.584	0.561	0.574	3.6
1,1,2-Trichloroethane	0.359	0.355	0.342	0.322	0.335	0.323	0.340	4.7
2-Hexanone	0.312	0.282	0.363	0.368	0.397	0.376	0.350	12.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG No.: Q2251
Instrument ID: MSVOA_N Calibration Date(s): 06/06/2025 06/06/2025
Heated Purge: (Y/N) N Calibration Time(s): 12:44 14:49
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086862.D		RRF005 = VN086863.D		RRF020 = VN086864.D			
		RRF050 = VN086865.D		RRF100 = VN086866.D		RRF150 = VN086867.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dibromochloromethane	0.361	0.351	0.358	0.340	0.363	0.349	0.354	2.5	
Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.5	
Chlorobenzene	1.233	1.135	1.107	1.023	1.089	1.030	1.103	7	
Ethyl Benzene	1.989	1.975	1.907	1.796	1.913	1.816	1.900	4.2	
m/p-Xylenes	0.730	0.751	0.741	0.701	0.736	0.703	0.727	2.8	
o-Xylene	0.714	0.699	0.702	0.674	0.711	0.678	0.696	2.4	
Styrene	1.177	1.193	1.226	1.162	1.229	1.164	1.192	2.5	
Bromoform	0.217	0.266	0.276	0.265	0.286	0.267	0.263	9.1	
Isopropylbenzene	3.864	3.749	3.649	3.426	3.621	3.546	3.643	4.2	
1,1,2,2-Tetrachloroethane	1.292	1.299	1.273	1.178	1.205	1.157	1.234	5	
1,3-Dichlorobenzene	1.763	1.709	1.657	1.554	1.612	1.566	1.644	5	
1,4-Dichlorobenzene	1.820	1.786	1.657	1.572	1.642	1.576	1.676	6.3	
1,2-Dichlorobenzene	1.675	1.651	1.596	1.500	1.557	1.496	1.579	4.8	
1,2-Dichloroethane-d4		0.732	0.707	0.500	0.656	0.751	0.669	15.1	
Dibromofluoromethane		0.303	0.310	0.219	0.298	0.351	0.296	16.2	
Toluene-d8		1.245	1.203	0.861	1.178	1.377	1.173	16.2	
4-Bromofluorobenzene		0.441	0.446	0.325	0.446	0.521	0.436	16.2	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.