

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG No.: Q2251

Instrument ID: MSVOA_N Calibration Date/Time: 06/10/2025 09:13

Lab File ID: VN086913.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.644	0.552	0.1	-14.29	20
Vinyl Chloride	0.664	0.653		-1.66	20
Bromomethane	0.372	0.358		-3.76	20
Chloroethane	0.429	0.421		-1.87	20
Trichlorofluoromethane	0.868	0.848		-2.3	20
1,1,2-Trichlorotrifluoroethane	0.545	0.530		-2.75	20
1,1-Dichloroethene	0.557	0.548		-1.62	20
Acetone	0.355	0.348		-1.97	20
Carbon Disulfide	1.539	1.428		-7.21	20
Methyl tert-butyl Ether	2.015	1.999		-0.79	20
Methylene Chloride	0.665	0.617		-7.22	20
trans-1,2-Dichloroethene	0.619	0.587		-5.17	20
1,1-Dichloroethane	1.120	1.079	0.1	-3.66	20
2-Butanone	0.577	0.493		-14.56	20
Carbon Tetrachloride	0.433	0.429		-0.92	20
cis-1,2-Dichloroethene	0.740	0.721		-2.57	20
Chloroform	1.118	1.053		-5.81	20
1,1,1-Trichloroethane	0.951	0.894		-5.99	20
Methylcyclohexane	0.605	0.523		-13.55	20
Benzene	1.444	1.400		-3.05	20
1,2-Dichloroethane	0.438	0.426		-2.74	20
Trichloroethene	0.342	0.345		0.88	20
1,2-Dichloropropane	0.351	0.344		-1.99	20
Bromodichloromethane	0.480	0.479		-0.21	20
4-Methyl-2-Pentanone	0.543	0.513		-5.53	20
Toluene	0.882	0.871		-1.25	20
t-1,3-Dichloropropene	0.537	0.556		3.54	20
cis-1,3-Dichloropropene	0.574	0.592		3.14	20
1,1,2-Trichloroethane	0.340	0.338		-0.59	20
2-Hexanone	0.350	0.339		-3.14	20
Dibromochloromethane	0.354	0.363		2.54	20
Tetrachloroethene	0.316	0.308		-2.53	20
Chlorobenzene	1.103	1.117	0.3	1.27	20
Ethyl Benzene	1.900	1.844		-2.95	20
m/p-Xylenes	0.727	0.727		0	20
o-Xylene	0.696	0.714		2.59	20
Styrene	1.192	1.232		3.36	20
Bromoform	0.263	0.287	0.1	9.13	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG No.: Q2251

Instrument ID: MSVOA_N Calibration Date/Time: 06/10/2025 09:13

Lab File ID: VN086913.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.643	3.637		-0.19	20
1,1,2,2-Tetrachloroethane	1.234	1.268	0.3	2.76	20
1,3-Dichlorobenzene	1.644	1.678		2.07	20
1,4-Dichlorobenzene	1.676	1.702		1.55	20
1,2-Dichlorobenzene	1.579	1.605		1.65	20
1,2-Dichloroethane-d4	0.669	0.613		-8.37	20
Dibromofluoromethane	0.296	0.301		1.69	20
Toluene-d8	1.173	1.153		-1.71	20
4-Bromofluorobenzene	0.436	0.420		-3.67	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.